



Recommendation ITU-R SM.1837
(12/2007)

**Test procedure for measuring
the 3rd order intercept
point (IP₃) level of radio
monitoring receivers**

SM Series
Spectrum management

Foreword

The role of the Radiocommunication Sector is to ensure the rational, equitable, efficient and economical use of the radio-frequency spectrum by all radiocommunication services, including satellite services, and carry out studies without limit of frequency range on the basis of which Recommendations are adopted.

The regulatory and policy functions of the Radiocommunication Sector are performed by World and Regional Radiocommunication Conferences and Radiocommunication Assemblies supported by Study Groups.

Policy on Intellectual Property Right (IPR)

ITU-R policy on IPR is described in the Common Patent Policy for ITU-T/ITU-R/ISO/IEC referenced in Annex 1 of Resolution ITU-R 1. Forms to be used for the submission of patent statements and licensing declarations by patent holders are available from <http://www.itu.int/ITU-R/go/patents/en> where the Guidelines for Implementation of the Common Patent Policy for ITU-T/ITU-R/ISO/IEC and the ITU-R patent information database can also be found.

Series of ITU-R Recommendations

(Also available online at <http://www.itu.int/publ/R-REC/en>)

Series	Title
BO	Satellite delivery
BR	Recording for production, archival and play-out; film for television
BS	Broadcasting service (sound)
BT	Broadcasting service (television)
F	Fixed service
M	Mobile, radiodetermination, amateur and related satellite services
P	Radiowave propagation
RA	Radio astronomy
RS	Remote sensing systems
S	Fixed-satellite service
SA	Space applications and meteorology
SF	Frequency sharing and coordination between fixed-satellite and fixed service systems
SM	Spectrum management
SNG	Satellite news gathering
TF	Time signals and frequency standards emissions
V	Vocabulary and related subjects

Note: This ITU-R Recommendation was approved in English under the procedure detailed in Resolution ITU-R 1.

*Electronic Publication
Geneva, 2009*

© ITU 2009

All rights reserved. No part of this publication may be reproduced, by any means whatsoever, without written permission of ITU.

RECOMMENDATION ITU-R SM.1837*

Test procedure for measuring the 3rd order intercept point (IP₃) level of radio monitoring receivers

(2007)

Scope

This Recommendation belongs to a set of Recommendations describing the test methods to determine technical parameters of radio monitoring receivers that are important for the users of these receivers. When the described methods are followed by manufacturers, comparing different receivers is made easier. This Recommendation specifies the test procedure for the determination of the IP₃ of a monitoring receiver. This test procedure definition is recommended to all the manufacturers, with the advantage for the users of such receivers, that an easier and more objective assessment of product quality is possible.

The ITU Radiocommunication Assembly,

considering

- a) that ITU-R has published the Typical recommended specifications for analogue and digital monitoring receivers in the ITU Handbook on Spectrum Monitoring (2011), but that nothing is said about the test procedures behind such specifications;
- b) that the specification of the 3rd order intercept point (IP₃) strongly depends on the test procedures applied;
- c) that the IP₃ level specified in the data sheet of a receiver depends extremely on the test frequencies used, the spacing of the test signals, the levels of the test signals and the ambient temperature prevailing during the tests;
- d) that the IP₃ characteristic has a direct influence on the suitability of a receiver to fulfil certain monitoring tasks, specially under real environment conditions (high-level signals in the spectra near to the tuned frequency);
- e) that without a defined test procedure, the IP₃ specifications published by the manufacturers have to be made comparable by some kind of conversion and that this conversion might be complex or even impossible to perform;
- f) that a defined test procedure for IP₃ must be independent of the receiver design;
- g) that a well-defined test procedure for IP₃, if adopted by all the manufacturers of radio monitoring receivers, will have the advantage for the users of such receivers, that an easier and more objective assessment of products from different manufacturers is possible, avoiding ambiguities;
- h) that supplementary information about these IP₃ measurements can be found in the Report ITU-R SM.2125 – Parameters of and measurement procedures on H/V/UHF monitoring receivers and stations;
- j) that when comparing the performance between two receivers, the IP₃ and the noise figure or the sensitivity are considered concurrently,

* Radiocommunication Study Group 1 made editorial amendments to this Recommendation in 2010 in accordance with Resolution ITU-R 1-5.

recommends

- 1 that the measurement method in Annex 1 should be used to determine the 3rd order intercept point (IP₃);
- 2 that each IP₃ figure should be accompanied by the noise figure or sensitivity during this IP₃'s measurement condition.

Annex 1

Test procedure for measuring the 3rd order intercept point (IP₃) level of radio monitoring receivers

1 General aspects

The IP₃ level depends on:

- the RF spacing between the two test signals;
- the RF level of the two test signals;
- the frequency pool selected for the test;
- the temperature prevailing during the tests.

Furthermore, to correctly assess the IP₃ level:

- both intermodulation products generated by the receiver for each pair of test frequencies have to be measured and evaluated. If their levels are different, the higher one is to be considered;
- duplicated measurements (on so called “image frequencies”), as they have influence on the mean value of IP₃, should be avoided. Duplicated measurements result in similar IP₃ level measurements, that do not really contribute to assessing the quality of the receiver;
- a *minimum value* for the IP₃ level must be specified and published by the manufacturer in the data sheet over the receiver's whole operating range. The measuring conditions for each measurement (spacing, levels of the test signals, equivalent input noise floor, AGC “on” or “off”) must also be published in the data sheet. For each spacing, the value should be indicated as to the receivers test condition (defined in “definitions”). Since IP₃ values are frequency dependent the manufacturer can choose to additionally specify IP₃ for selected frequency bands or ranges;
- a *mean value* (the arithmetic mean of a number of test measurements) may also be indicated.

Definitions

The following definitions apply for this Recommendation:

- The receiver test condition is defined as “*condition 1*” if both test signals pass through the complete analogue signal path including A/D converters or detectors.
- The receiver test condition is defined as “*condition 2*” if both test signals pass through the analogue signal path to an analogue IF output.
- The receiver test condition is defined as “*condition 3*” if one or more of the test signals is removed at an intermediate stage of the receivers analogue signal path.

2 Basics of the concept “3rd order intercept point level (IP₃)”

Two unmodulated test signals of the same r.m.s.-power P_{in} each and the frequencies f_1 and f_2 ($f_1 < f_2$) are inserted into the antenna input of the receiver. Due to non-linearities, two intermodulation products of the frequency $f_3 = \{(2 \times f_1) - f_2\}$ and $f_4 = \{(2 \times f_2) - f_1\}$ may appear.

The IP₃ level is then calculated as follows:

$$IP_3 = P_{in} + a/2$$

where:

- P_{in} : r.m.s.-power measured for each of the two inserted test signals (dBm)
- a : difference (dB) between the level of the inserted test signals and the level of the intermodulation products at the measurement point. If the intermodulation products are of different level, the higher one is to be considered.

Variation of the receivers noise floor

The equivalent input noise floor of the receiver under test has to be measured during the whole IP₃ measurement with and without the test signals switched on. Details about the frequencies used for this test (f_5 and f_6) are in the paragraph “*measurement of the levels*”.

Test bench conditions

In making this measurement, it is important to have sufficient isolation between the two test sources to prevent IM products from being generated in the output stages of the sources at frequencies of f_3 and f_4 . It may be necessary to use devices such as isolators, fixed attenuators, isolation amplifiers, or high isolation Wilkinson power combiners. Low-pass filtering may also be required to attenuate the 2nd harmonics of the sources.

To verify the reliability of all these devices in the test set-up, an IP₃ measurement of the test bench may be performed without connecting the receiver under test. The test bench IP₃ should be at least 10 dB better than the expected receiver IP₃.

Measurement of the levels

The levels are measured at f_1, f_2, f_3, f_4 , and to ensure a constant equivalent input noise floor of the receiver also at f_5 and f_6 using the built in level indicator of the receiver. f_5 and f_6 represent the adjacent channels of the bandwidth BW of the receiver ($f_5 = f_3 - BW$ and $f_6 = f_4 + BW$) levels at f_5 and f_6 are measured with the test signals switched on and off. This built in level indicator needs to be calibrated over the amplitude range of the expected measured signal levels. The maximum allowed absolute error is ± 1 dB and the readout resolution of the level indicator need to be ≤ 0.1 dB. When the IF output is available it is also allowed to use a spectrum analyser or measurement receiver connected to the last IF of the receiver as a level indicator, the same calibration and accuracy requirements as for the receivers built in indicator apply here but a relatively narrow measurement filter can be used to cope with difficulties that arise when measuring the IP₃ of receivers with a narrow bandwidth.

Phase noise

If the levels of the IM products at f_3 and f_4 are near the noise floor or phase noise skirts, the measured levels will represent the IM product level plus noise. The actual IM product level can be obtained by subtracting out the noise contribution.

Frequency spacing of the test signals

The frequency spacing between the frequencies f_1 and f_2 is chosen according to § 3 below and in such a way that both test signals comply either with condition 1, condition 2 or condition 3.

The chosen bandwidth used during the measurement has to be specified.

At least two frequency pairs per octave evenly distributed over the full frequency range of the receiver are chosen as test signals. The published IP_3 values have to be valid over the entire rated temperature range indicated in the data sheet. Limitations, if any, are mentioned in the data sheet. If a variable input attenuator exists, it must be set to 0 dB attenuation during the tests. In any other case e.g. when the receiver design does not allow to change the attenuator setting manually, the attenuation used during the tests and the reason why another value than 0 dB was used must be specified. To more fully characterize a receiver, the measured performance at more than one attenuation setting may be provided. The receiver must be set up under normal operating conditions, and it must be indicated whether the AGC is switched on or off, depending on the measurement.

Configuration of the receiver under test

If a switchable preamplifier exists, the measurements must be done in the condition “preamplifier off”. Some receiver designs do not allow a physical amplifier switch off but use a combination of variable step attenuators. This combination of amplifier and attenuator should be set to 0 dB gain.

3 Definition of a test procedure for measuring the 3rd order intercept point level of receivers for the frequency ranges 9 kHz to 30 MHz and 20 MHz to 3 000 MHz

Parameters for the IP_3 test should be chosen with the typical use of the receiver in mind. For comparing receivers with an overlap in their fields of application this is sometimes difficult. The approach of this Recommendation is to test receivers under comparable conditions, and:

- the basic considerations in § 2 should be followed;
- the permitted level range of test signals f_1 and f_2 at the antenna input of the monitoring receiver is -30 dBm to 10 dBm;
- the IF filter bandwidth or resolution bandwidth (RBW) for the measurement can be chosen by the manufacturer but must be a realistic choice for the type of receiver and intended application. The chosen bandwidth (BW) for the frequency range 9 kHz to 30 MHz must be ≤ 5 kHz and for the frequency range 20 MHz to 3 000 MHz it must be ≤ 30 kHz;
- receivers should be tested with a range of frequency spacings for the test frequencies. A harmonically increasing list should be used starting at 1 Hz then increasing from 3 Hz, 10 Hz, 30 Hz, 100 Hz, 300 Hz, 1 kHz, 3 kHz up to 300 MHz. For many receivers and applications it is not useful to measure the whole range of spacings. Another frequency spacing for the first and last in the list can be chosen. It is however necessary to measure using all frequency spacings in between these start and stop frequencies;
- the tolerance of the frequency spacing should be $\leq 1\%$;
- measured values of IP_3 should be presented in a single table indicating condition 1, condition 2 and condition 3 measurements, or in separate tables, depending on the number of conditions. The table should contain for each measurement, the used frequency spacing, an indication under which condition (condition 1, 2 or 3) the measurement is performed, and the measured IP_3 ;
- each entry in the table should contain a note if the measurement performed represents a real life practical usage condition of the receiver. Additional information describing the measurements conditions can be added at the bottom of the table.