

International Telecommunication Union

ITU-T

TELECOMMUNICATION
STANDARDIZATION SECTOR
OF ITU

H.841

(01/2015)

SERIES H: AUDIOVISUAL AND MULTIMEDIA SYSTEMS

E-health multimedia services and applications –
Interoperability compliance testing of personal health
systems (HRN, PAN, LAN and WAN)

**Conformance of ITU-T H.810 personal health
devices: PAN/LAN/TAN interface Part 1:
Optimized exchange protocol: Agent**

Recommendation ITU-T H.841

ITU-T



ITU-T H-SERIES RECOMMENDATIONS
AUDIOVISUAL AND MULTIMEDIA SYSTEMS

CHARACTERISTICS OF VISUAL TELEPHONE SYSTEMS	H.100–H.199
INFRASTRUCTURE OF AUDIOVISUAL SERVICES	
General	H.200–H.219
Transmission multiplexing and synchronization	H.220–H.229
Systems aspects	H.230–H.239
Communication procedures	H.240–H.259
Coding of moving video	H.260–H.279
Related systems aspects	H.280–H.299
Systems and terminal equipment for audiovisual services	H.300–H.349
Directory services architecture for audiovisual and multimedia services	H.350–H.359
Quality of service architecture for audiovisual and multimedia services	H.360–H.369
Telepresence	H.420–H.429
Supplementary services for multimedia	H.450–H.499
MOBILITY AND COLLABORATION PROCEDURES	
Overview of Mobility and Collaboration, definitions, protocols and procedures	H.500–H.509
Mobility for H-Series multimedia systems and services	H.510–H.519
Mobile multimedia collaboration applications and services	H.520–H.529
Security for mobile multimedia systems and services	H.530–H.539
Security for mobile multimedia collaboration applications and services	H.540–H.549
Mobility interworking procedures	H.550–H.559
Mobile multimedia collaboration inter-working procedures	H.560–H.569
BROADBAND, TRIPLE-PLAY AND ADVANCED MULTIMEDIA SERVICES	
Broadband multimedia services over VDSL	H.610–H.619
Advanced multimedia services and applications	H.620–H.629
Ubiquitous sensor network applications and Internet of Things	H.640–H.649
IPTV MULTIMEDIA SERVICES AND APPLICATIONS FOR IPTV	
General aspects	H.700–H.719
IPTV terminal devices	H.720–H.729
IPTV middleware	H.730–H.739
IPTV application event handling	H.740–H.749
IPTV metadata	H.750–H.759
IPTV multimedia application frameworks	H.760–H.769
IPTV service discovery up to consumption	H.770–H.779
Digital Signage	H.780–H.789
E-HEALTH MULTIMEDIA SERVICES AND APPLICATIONS	
Interoperability compliance testing of personal health systems (HRN, PAN, LAN, TAN and WAN)	H.820–H.859
Multimedia e-health data exchange services	H.860–H.869

For further details, please refer to the list of ITU-T Recommendations.

Recommendation ITU-T H.841

Conformance of ITU-T H.810 personal health devices: PAN/LAN/TAN interface Part 1: Optimized exchange protocol: Agent

Summary

Recommendation ITU-T H.841 is a transposition of Continua Test Tool DG2013, Test Suite Structure & Test Purposes, PAN-LAN-TAN Interface; Part 1: Optimized Exchange Protocol. Agent (Version 1.6, 2014-01-24), that was developed by the Continua Health Alliance. A number of versions of this specification existed before transposition.

This Recommendation includes an electronic attachment with the protocol implementation conformance statements (PICS) and the protocol implementation extra information for testing (PIXIT) required for the implementation of Annex A.

History

Edition	Recommendation	Approval	Study Group	Unique ID*
1.0	ITU-T H.841	2015-01-13	16	11.1002/1000/12258

* To access the Recommendation, type the URL <http://handle.itu.int/> in the address field of your web browser, followed by the Recommendation's unique ID. For example, <http://handle.itu.int/11.1002/1000/11830-en>.

FOREWORD

The International Telecommunication Union (ITU) is the United Nations specialized agency in the field of telecommunications, information and communication technologies (ICTs). The ITU Telecommunication Standardization Sector (ITU-T) is a permanent organ of ITU. ITU-T is responsible for studying technical, operating and tariff questions and issuing Recommendations on them with a view to standardizing telecommunications on a worldwide basis.

The World Telecommunication Standardization Assembly (WTSA), which meets every four years, establishes the topics for study by the ITU-T study groups which, in turn, produce Recommendations on these topics.

The approval of ITU-T Recommendations is covered by the procedure laid down in WTSA Resolution 1.

In some areas of information technology which fall within ITU-T's purview, the necessary standards are prepared on a collaborative basis with ISO and IEC.

NOTE

In this Recommendation, the expression "Administration" is used for conciseness to indicate both a telecommunication administration and a recognized operating agency.

Compliance with this Recommendation is voluntary. However, the Recommendation may contain certain mandatory provisions (to ensure, e.g., interoperability or applicability) and compliance with the Recommendation is achieved when all of these mandatory provisions are met. The words "shall" or some other obligatory language such as "must" and the negative equivalents are used to express requirements. The use of such words does not suggest that compliance with the Recommendation is required of any party.

INTELLECTUAL PROPERTY RIGHTS

ITU draws attention to the possibility that the practice or implementation of this Recommendation may involve the use of a claimed Intellectual Property Right. ITU takes no position concerning the evidence, validity or applicability of claimed Intellectual Property Rights, whether asserted by ITU members or others outside of the Recommendation development process.

As of the date of approval of this Recommendation, ITU had not received notice of intellectual property, protected by patents, which may be required to implement this Recommendation. However, implementers are cautioned that this may not represent the latest information and are therefore strongly urged to consult the TSB patent database at <http://www.itu.int/ITU-T/ipr/>.

© ITU 2015

All rights reserved. No part of this publication may be reproduced, by any means whatsoever, without the prior written permission of ITU.

Table of Contents

	Page
1 Scope.....	1
2 References.....	2
3 Definitions	2
3.1 Terms defined elsewhere	2
3.2 Terms defined in this Recommendation.....	2
4 Abbreviations and acronyms	2
5 Conventions	3
6 Test suite structure (TSS)	4
7 Electronic attachment	6
Annex A – Test purposes (TPs).....	7
A.1 TP definition conventions.....	7
A.2 Subgroup 1.2.1 – PHD domain information model (DIM)	8
A.3 Subgroup 1.2.2 – PHD service model (SER)	78
A.4 Subgroup 1.2.3 – PHD communication model (COM).....	86
Bibliography.....	125

Electronic attachment: Protocol implementation conformance statements (PICS) and protocol implementation extra information for testing (PIXIT) required for the implementation of Annex A.

Introduction

This Recommendation is a transposition of Continua Test Tool DG2013, Test Suite Structure & Test Purposes, PAN-LAN-TAN Interface; Part 1: Optimized Exchange Protocol. Agent (Version 1.6, 2014-01-24), that was developed by the Continua Health Alliance. A number of versions of this specification existed before transposition and these can be found in the table below.

Version	Date	Revision history
1.4	2012-10-05	Initial release for Test Tool DG2011. This uses "TSS&TP_1.5_PAN-LAN_PART_1_v1.3.doc" as a baseline and adds new features included in [b-CDG 2011] (Person ID and Errata).
1.5	2013-05-24	Initial release for Test Tool DG2012. This uses "TSS&TP_DG2011_PAN-LAN_PART_1_v1.4.doc" as a baseline and adds new features included in [b-CDG 2012]: <ul style="list-style-type: none">• Adds glucose meter• Adds body composition analyser device specialization• Adds basic electrocardiograph device specialization
1.6	2014-01-24	Initial release for Test Tool DG2013. This uses "TSS&TP_DG2012_PAN-LAN_PART_1_v1.5.doc" as a baseline and adds new features included in [ITU-T H.810]: <ul style="list-style-type: none">• Adds glucose meter BLE• Adds BLE SSP support• Adds NFC new transport• Adds INR device specialization

Recommendation ITU-T H.841

Conformance of ITU-T H.810 personal health devices: PAN/LAN/TAN interface Part 1: Optimized exchange protocol: Agent

1 Scope

The scope of this Recommendation¹ is to provide a test suite structure and the test purposes (TSS & TP) for the PAN/LAN/TAN interface based on the requirements defined in the Continua Design Guidelines (CDG) [ITU-T H.810]. The objective of this test specification is to provide a high probability of air interface interoperability between different devices.

The TSS & TP for the PAN/LAN/TAN interface document have been divided into 10 parts. Each part is listed below:

- **Part 1:** Optimized exchange protocol [ISO/IEEE 11073-20601A] Agent
- **Part 2:** Optimized exchange protocol ([ISO/IEEE 11073-20601A] Manager
- **Part 3:** Continua design guidelines. Agent
- **Part 4:** Continua design guidelines. Manager
- **Part 5:** Device specializations. Agent. This document is divided into 14 subparts:
 - **Part 5A:** Weighing scales
 - **Part 5B:** Glucose meter
 - **Part 5C:** Pulse oximeter
 - **Part 5D:** Blood pressure monitor
 - **Part 5E:** Thermometer
 - **Part 5F:** Cardiovascular fitness and activity monitor
 - **Part 5G:** Strength fitness equipment
 - **Part 5H:** Independent living activity hub
 - **Part 5I:** Adherence monitor
 - **Part 5J:** Insulin pump (Future development)
 - **Part 5K:** Peak flow
 - **Part 5L:** Body composition analyser
 - **Part 5M:** Basic electrocardiograph
 - **Part 5N:** International normalized ratio monitor
- **Part 6:** Device specializations. Manager
- **Part 7:** Continua design guidelines. Agent BLE
- **Part 8:** Continua design guidelines. Manager BLE
- **Part 9:** Personal health devices transcoding whitepaper. Agent
- **Part 10:** Personal health devices transcoding whitepaper. Manager

¹ This Recommendation includes an electronic attachment with the protocol implementation conformance statements (PICS) and the protocol implementation extra information for testing (PIXIT) required for the implementation Annex A.

2 References

The following ITU-T Recommendations and other references contain provisions which, through reference in this text, constitute provisions of this Recommendation. At the time of publication, the editions indicated were valid. All Recommendations and other references are subject to revision; users of this Recommendation are therefore encouraged to investigate the possibility of applying the most recent edition of the Recommendations and other references listed below. A list of the currently valid ITU-T Recommendations is regularly published. The reference to a document within this Recommendation does not give it, as a stand-alone document, the status of a Recommendation.

- [ITU-T H.810] Recommendation ITU-T H.810 (2013), *Interoperability design guidelines for personal health systems*.
- [IEEE 11073-10406] IEEE 11073-10406-2011, *Health informatics – Personal health device communication Part 10406: Device specialization – Basic electrocardiograph (ECG) (1 to 3-lead ECG)*.
- [ISO/IEEE 11073-20601A] ISO/IEEE 11073-20601:2010, *Health informatics – Personal health device communication – Part 20601: Application profile – Optimized exchange protocol*, including ISO/IEEE 11073-20601:2010 Amd 1:2015.
<http://www.iso.org/iso/home/store/catalogue_tc/catalogue_detail.htm?csnumber=54331>
with
<http://www.iso.org/iso/home/store/catalogue_tc/catalogue_detail.htm?csnumber=63972>
- [ISO/IEEE 11073-104xx] ISO/IEEE 11073-104xx (in force), *Health informatics – Personal health device communication – Device specialization*.
NOTE – This is shorthand used to refer to the collection of device specialization standards that utilize [ISO/IEEE 11073-20601A], where xx can be any number from 01 to 99, inclusive.
- [ISO/IEEE 11073-10472] ISO/IEEE 11073-10472:2012, *Health informatics – Personal health device communication – Part 10472: Device specialization – Medication monitor*.

3 Definitions

3.1 Terms defined elsewhere

3.1.1 agent [ISO/IEEE 11073-20601A]: A node that collects and transmits personal health data to an associated manager.

3.1.2 manager [ISO/IEEE 11073-20601A]: A node receiving data from one or more agent systems. Some examples of managers include a cellular phone, health appliance, set top box, or a computer system.

3.2 Terms defined in this Recommendation

None.

4 Abbreviations and acronyms

This Recommendation uses the following abbreviations and acronyms:

ATS	Abstract Test Suite
DUT	Device Under Test
CDG	Continua Design Guidelines

GUI	Graphical User Interface
INR	International Normalized Ratio
IUT	Implementation Under Test
MDS	Medical Device System
NFC	Near Field Communication
PAN	Personal Area Network
PCT	Protocol Conformance Testing
PCO	Point of Control and Observation
PHD	Personal Healthcare Device
PHDC	Personal Healthcare Device Class
PHM	Personal Health Manager
PICS	Protocol Implementation Conformance Statement
PIXIT	Protocol Implementation extra Information for Testing
SDP	Service Discovery Protocol
SOAP	Simple Object Access Protocol
TCRL	Test Case Reference List
TCWG	Test and Certification Working Group
TP	Test Purpose
TSS	Test Suite Structure
USB	Universal Serial Bus
WDM	Windows Driver Model

5 Conventions

The key words "SHALL", "SHALL NOT", "SHOULD", "SHOULD NOT", "MAY", "MAY NOT" in this document are to be interpreted as in [b-ETSI SR 001 262].

- SHALL is equivalent to 'must' or 'it is required to'.
- SHALL NOT is equivalent to 'must not' or 'it is not allowed'.
- SHOULD is equivalent to 'it is recommended to'.
- SHOULD NOT is equivalent to 'it is not recommended to'.
- MAY is equivalent to 'is permitted'.
- MAY NOT is equivalent to 'it is not required that'.

NOTE – The above-mentioned key words are capitalized for illustrative purposes only and they do not appear capitalized within this Recommendation.

Reference is made in the ITU-T H.800-series of Recommendations to different versions of the Continua Design Guidelines (CDG) by a specific designation. The list of terms that may be used in this Recommendation is provided in Table 1.

Table 1 – List of designations associated with the various versions of the CDG

CDG name	Transposed as	Version	Description	Designation
2013 plus errata	ITU-T H.810	4.1	CDG 2013 plus errata noting all ratified bugs.	–
2013	–	4.0	Release 2013 of the CDG including maintenance updates of CDG 2012 and additional guidelines that cover new functionalities.	Endorphin
2012 plus errata	–	3.1	CDG 2012 plus errata noting all ratified bugs [b-CDG 2012].	–
2012	–	3.0	Release 2012 of the CDG including maintenance updates of CDG 2011 and additional guidelines that cover new functionalities.	Catalyst
2011 plus errata	–	2.1	CDG 2011 integrated with identified errata.	–
2011	–	2.0	Release 2011 of the CDG including maintenance updates of CDG 2010 and additional guidelines that cover new functionalities [b-CDG 2011].	Adrenaline
2010 plus errata	–	1.6	CDG 2010 integrated with identified errata.	–
2010	–	1.5	Release 2010 of the CDG with maintenance updates of CDG Version 1 and additional guidelines that cover new functionalities [b-CDG 2010].	1.5
1.0	–	1.0	First released version of the CDG [b-CDG 1.0].	–

6 Test suite structure (TSS)

The test purposes (TPs) for the PAN/LAN/TAN interface have been divided into the main subgroups specified below. Annex A describes the TPs for subgroups 1.2.1, 1.2.2 and 1.2.3 (shown in bold).

- Group 1: Agent (AG)
 - Group 1.1: Transport (TR)
 - Subgroup 1.1.1: Design guidelines: Common (DGC)
 - Subgroup 1.1.2: USB design guidelines (UDG)
 - Subgroup 1.1.3: Bluetooth design guidelines (BDG)
 - Subgroup 1.1.4: Pulse oximeter design guidelines (PODG)
 - Subgroup 1.1.5: Cardiovascular design guidelines (CVDG)
 - Subgroup 1.1.6: Activity hub design guidelines (HUBDG)
 - Subgroup 1.1.7: ZigBee design guidelines (ZDG)
 - Subgroup 1.1.8: Glucose meter design guidelines (GLDG)
 - Subgroup 1.1.9: Bluetooth low energy design guidelines (BLEDG)
 - Subgroup 1.1.10: Basic electrocardiograph design guidelines (ECGDG)
 - Subgroup 1.1.11: NFC design guidelines (NDG)

- Group 1.2: Optimized exchange protocol (OXP)
 - **Subgroup 1.2.1: PHD domain information model (DIM)**
 - **Subgroup 1.2.2: PHD service model (SER)**
 - **Subgroup 1.2.3: PHD communication model (COM)**
- Group 1.3: Devices class specializations (CLASS)
 - Subgroup 1.3.1: Weighing scales (WEG)
 - Subgroup 1.3.2: Glucose meter (GL)
 - Subgroup 1.3.3: Pulse oximeter (PO)
 - Subgroup 1.3.4: Blood pressure monitor (BPM)
 - Subgroup 1.3.5: Thermometer (TH)
 - Subgroup 1.3.6: Cardiovascular (CV)
 - Subgroup 1.3.7: Strength (ST)
 - Subgroup 1.3.8: Activity hub (HUB)
 - Subgroup 1.3.9: Adherence monitor (AM)
 - Subgroup 1.3.10: Insulin pump (IP) (Future development)
 - Subgroup 1.3.11: Peak flow (PF)
 - Subgroup 1.3.12: Body composition analyser (BCA)
 - Subgroup 1.3.13: Basic electrocardiograph (ECG)
 - Subgroup 1.3.14: International normalized ratio (INR)
- Group 1.4: Personal health device transcoding whitepaper (PHDTW)
 - Subgroup 1.4.1: Whitepaper general requirements (GEN)
 - Subgroup 1.4.2: Whitepaper thermometer requirements (TH)
 - Subgroup 1.4.3: Whitepaper blood pressure requirements (BPM)
 - Subgroup 1.4.4: Whitepaper heart rate requirements (HR)
 - Subgroup 1.4.5: Whitepaper glucose meter requirements (GL)
- Group 2: Manager (MAN)
 - Group 2.1: Transport (TR)
 - Subgroup 2.1.1: Design guidelines: common (DGC)
 - Subgroup 2.1.2: USB design guidelines (UDG)
 - Subgroup 2.1.3: Bluetooth design guidelines (BDG)
 - Subgroup 2.1.4: Cardiovascular design guidelines (CVDG)
 - Subgroup 2.1.5: Activity hub design guidelines (HUBDG)
 - Subgroup 2.1.6: ZigBee design guidelines (ZDG)
 - Subgroup 2.1.7: Bluetooth low energy design guidelines (BLEDG)
 - Subgroup 2.1.8: NFC design guidelines (NDG)
 - Group 2.2: 20601: Optimized exchange protocol (OXP)
 - Subgroup 2.2.1: General (GEN)
 - Subgroup 2.2.2: PHD domain information model (DIM)

- Subgroup 2.2.3: PHD service model (SER)
- Subgroup 2.2.4: PHD communication model (COM)
- Group 2.3: Devices class specializations (CLASS)
 - Subgroup 2.3.1: Weighing scales (WEG)
 - Subgroup 2.3.2: Glucose meter (GL)
 - Subgroup 2.3.3: Pulse oximeter (PO)
 - Subgroup 2.3.4: Blood pressure monitor (BPM)
 - Subgroup 2.3.5: Thermometer (TH)
 - Subgroup 2.3.6: Cardiovascular (CV)
 - Subgroup 2.3.7: Strength (ST)
 - Subgroup 2.3.8: Activity hub (HUB)
 - Subgroup 2.3.9: Adherence monitor (AM)
 - Subgroup 2.3.10: Insulin pump (IP) (Future development)
 - Subgroup 2.3.11: Peak flow (PF)
 - Subgroup 2.3.12: Body composition analyser (BCA)
 - Subgroup 2.3.13: Basic electrocardiograph (ECG)
 - Subgroup 2.3.14: International normalized ratio (INR)
- Group 2.4: Personal health device transcoding whitepaper (PHDTW)
 - Subgroup 2.4.1: Whitepaper general requirements (GEN)
 - Subgroup 2.4.2: Whitepaper thermometer requirements (TH)
 - Subgroup 2.4.3: Whitepaper blood pressure measurement requirements (BPM)
 - Subgroup 2.4.4: Whitepaper heart rate requirements (HR)
 - Subgroup 2.4.5: Whitepaper glucose meter requirements (GL)

7 Electronic attachment

The protocol implementation conformance statements (PICS) and the protocol implementation extra information for testing (PIXIT) required for the implementation of this Annex can be downloaded from <http://handle.itu.int/11.1002/2000/12067>.

In the electronic attachment, letters "C" and "I" in the column labelled "Mandatory" are used to distinguish between "PICS" and "PIXIT" respectively during testing. If the cell is empty, the corresponding PICS is "independent". If the field contains a "C", the corresponding PICS is dependent on other PICS, and the logical expression is detailed in the "SCR_Expression" field. The static conformance review (SCR) is used in the test tool to assert whether the PICS selection is consistent.

Annex A

Test purposes (TPs)

(This annex forms an integral part of this Recommendation.)

A.1 TP definition conventions

The test purposes are defined according to the following rules:

- **TP Id:** This is a unique identifier (TP/<TT>/<DUT>/<GR>/<SGR>/<XX> – <NNN>). It is specified according to the naming convention defined below:
 - Each test purpose identifier is introduced by the prefix "TP".
 - <TT>: This is the test tool that will be used in the test case:
 - PAN: Personal area network (Bluetooth or USB)
 - LAN: Local area network (ZigBee)
 - PAN-LAN: Personal area network (Bluetooth or USB) - Local area network (ZigBee)
 - LP-PAN: Low power personal area network (Bluetooth Low Energy)
 - TAN: Touch area network (NFC)
 - PLT: Personal area network (Bluetooth or USB) – Local area network (ZigBee) – Touch area network (NFC)
 - <DUT>: This is the device under test:
 - AG: PAN/LAN Agent
 - MAN: PAN/LAN Manager
 - <GR>: This identifies a group of test cases.
 - <SGR>: This identifies a subgroup of test cases.
 - <XX>: This identifies the type of testing:
 - BV: Valid behaviour test
 - BI: Invalid behaviour test
 - <NNN>: This is a sequential number that identifies the test purpose.
 - **TP label:** It is the TP's title.
 - **Coverage:** This contains the specification reference and clause to be checked by the TP:
 - Spec: This indicates the earliest version of the specification from which the testable items to be checked by the TP were included.
 - Testable item: This contains the testable items to be checked by the TP.
 - **Test purpose:** This is a description of the requirements to be tested.
 - **Applicability:** This contains the PICS items that define if the test case is applicable or not for a specific device. When a TP contains an "ALL" in this field it means that it applies to the device under test within that scope of the test (specialization, transport used, etc.).
 - **Initial condition:** This indicates the state to which the DUT needs to be moved at the beginning of TC execution.
 - **Test procedure:** This describes the steps to be followed in order to execute the test case.
 - **Pass/Fail criteria:** This provides criteria to decide whether the DUT passes or fails the test case.

A.2 Subgroup 1.2.1 – PHD domain information model (DIM)

TP Id	TP/PLT/AG/EXP/DIM/BV-000			
TP Label	MDS Object: Mandatory, Conditional and Optional Attributes			
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	MDSClass 1; M	MDSClass 2; M	MDSClassAttr 2; M
		MDSClassAttr 3; C	MDSClassAttr 4; M	MDSClassAttr 5; M
		MDSClassAttr 6; M	MDSClassAttr 7; C	MDSClassAttr 8; O
		MDSClassAttr 9; C	MDSClassAttr 10; C	MDSClassAttr 11; C
		MDSClassAttr 12; C	MDSClassAttr 13; C	MDSClassAttr 14; O
		MDSClassAttr 15; O	MDSClassAttr 16; O	MDSClassAttr 17; O
		MDSClassAttr 18; C	MDSClassAttr 19; O	MDSService 1; M
		MDSService 4; M	OperNormProc 2; M	OperNormProc 3; M
		ConfNormalProc 21; M	CommonCharac 3; M	ConfNormalProc 1; M
		OperNormProc 6; M	MDSMethod 7; O	ConfEventRep 28; M
		ConfEventRep 33; O	ConfEventRep 34; M	MDSMethod 6; M
	BaseTimOffset3; M			
Spec	[ITU-T H.810]			
Testable items	Regulatory 4; M	Regulatory 5; O		
	multi_funct_LAN 1; M			
	General 2; M			
Applicability	C_AG_EXP_000			
Initial condition	The simulated manager and the agent under test are in the unassociated state.			
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 4. Check that the MDS object and its attributes are not present in the Configuration Event Report. 5. The simulated manager issues "roiv-cmip-get" command with the handle set to 0 (to request an MDS object) and an empty attribute-id-list to indicate all attributes. Record the invoke-id of the message sent. 6. The agent responds with with a "rors-cmip-get" service message in which the attribute-list contains a list of all implemented attributes of the MDS object: Verify that the invoke-id is mirrored from the Get request. Verify that the DataApdu contains the SEQUENCE GetResultSimple (0x02 0x03) <ol style="list-style-type: none"> a. Verify that the GetResultSimple.obj-handle = 0x00 0x00 b. The GetResultSimple contains an AttributeList: <ul style="list-style-type: none"> <input type="checkbox"/> AttributeList.count = number of implemented attributes that are included in the GET response. (must be greater than 3) <input type="checkbox"/> AttributeList.length = the length of the remainder of the APDU Mandatory Attributes <ol style="list-style-type: none"> a. Mandatory attribute HANDLE shall not be present. b. Mandatory attribute System-Model <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_MODEL (0x09 0x28) <input type="checkbox"/> attribute-value.length = 0xXX 0xXX <input type="checkbox"/> attribute-value = SystemModel 			

- SystemModel.manufacturer = OCTET STRING
- length = 0xXX 0xXX, where X is even and equal to the length of the value
- value = <Check with PIXITS>
- SystemModel.model-number = OCTET STRING
- length = 0xXX 0xXX, where X is even and equal to the length of the value
- value = <Check with PIXITS>

c. Mandatory attribute System-Id

- attribute-id = MDC_ATTR_SYS_ID (0x09 0x84)
- attribute-type = OCTET STRING
- attribute-value.length = 10 bytes
- attribute-value = OCTET STRING(Size(8))
- size = 0x00 0x08
- value = <Check with PIXITS>

d. Mandatory attribute Dev-Configuration-Id

- attribute-id = MDC_ATTR_DEV_CONFIG_ID (0x0A 0x44)
- attribute-type = ConfigId
- attribute-value.length = 2 bytes
- attribute-value = <between 0x00 0x01 and 0x7F 0xFF>

Conditional and Optional Attributes

e. One and only one of [System-Type-Spec_List] or [System-Type] shall be present.

f. IF System-Type attribute is present:

- attribute-id = MDC_ATTR_SYS_TYPE (0x09 0x86)
- attribute-type = TYPE
- attribute-value.length = 4 bytes
- attribute-value = One of the supported specializations:
 - MDC_DEV_SPEC_PROFILE_PULSE_OXIM (0x10 0x04)
 - MDC_DEV_SPEC_PROFILE_BP (0x10 0x06)
 - MDC_DEV_SPEC_PROFILE_TEMP (0x10 0x07)
 - MDC_DEV_SPEC_PROFILE_SCALE (0x10 0x0F)
 - MDC_DEV_SPEC_PROFILE_GLUCOSE (0x10 0x11)
 - MDC_DEV_SPEC_PROFILE_HF_CARDIO (0x10 0x29)
 - MDC_DEV_SPEC_PROFILE_HF_STRENGTH (0x10 0x30)
 - MDC_DEV_SPEC_PROFILE_AI_ACTIVITY_HUB (0x10 0x47)
 - MDC_DEV_SPEC_PROFILE_AI_MED_MINDER (0x10 0x48)
 - MDC_DEV_SPEC_PROFILE_PEAK_FLOW (0x10 0x15)
 - MDC_DEV_SPEC_PROFILE_BCA (0x10 0x14)
 - MDC_DEV_SPEC_PROFILE_ECG (0x10 0x06)
 - MDC_DEV_SPEC_PROFILE_COAG (0x10 0x16)

g. IF System-Type-Spec-List attribute is present

- attribute-id = MDC_ATTR_SYS_TYPE_SPEC_LIST (0x0A 0x5A)
- attribute-type = TypeVerList
- attribute-value.count = N (record for next step)
- attribute-value.length = N*4 bytes
- attribute-value = N of the following supported specializations:

- MDC_DEV_SPEC_PROFILE_PULSE_OXIM (0x10 0x04)
- MDC_DEV_SPEC_PROFILE_BP (0x10 0x06)
- MDC_DEV_SPEC_PROFILE_TEMP (0x10 0x07)
- MDC_DEV_SPEC_PROFILE_SCALE (0x10 0x0F)
- MDC_DEV_SPEC_PROFILE_GLUCOSE (0x10 0x11)
- MDC_DEV_SPEC_PROFILE_HF_CARDIO (0x10 0x29)
- MDC_DEV_SPEC_PROFILE_HF_STRENGTH (0x10 0x30)
- MDC_DEV_SPEC_PROFILE_AI_ACTIVITY_HUB (0x10 0x47)
- MDC_DEV_SPEC_PROFILE_AI_MED_MINDER (0x10 0x48)
- MDC_DEV_SPEC_PROFILE_PEAK_FLOW (0x10 0x15)
- MDC_DEV_SPEC_PROFILE_BCA (0x10 0x14)
- MDC_DEV_SPEC_PROFILE_ECG (0x10 0x06)
- MDC_DEV_SPEC_PROFILE_COAG (0x10 0x16)
- Profiles for Cardiovascular fitness and monitor specialization: If the agent supports Step Counter profile THEN the agent shall support Cardiovascular specialization
 - MDC_DEV_SUB_SPEC_PROFILE_STEP_COUNTER (0x10 0x68)
- Profiles for Activity Hub specialization: If the agent supports any of the profiles defined for Activity Hub THEN the agent shall support Activity Hub specialization:
 - MDC_DEV_SUB_SPEC_PROFILE_FALL_SENSOR (0x10 0x75)
 - MDC_DEV_SUB_SPEC_PROFILE_PERS_SENSOR (0x10 0x76)
 - MDC_DEV_SUB_SPEC_PROFILE_SMOKE_SENSOR (0x10 0x77)
 - MDC_DEV_SUB_SPEC_PROFILE_CO_SENSOR (0x10 0x78)
 - MDC_DEV_SUB_SPEC_PROFILE_WATER_SENSOR (0x10 0x79)
 - MDC_DEV_SUB_SPEC_PROFILE_GAS_SENSOR (0x10 0x7A)
 - MDC_DEV_SUB_SPEC_PROFILE_MOTION_SENSOR (0x10 0x7B)
 - MDC_DEV_SUB_SPEC_PROFILE_PROPEXIT_SENSOR (0x10 0x7C)
 - MDC_DEV_SUB_SPEC_PROFILE_ENURESIS_SENSOR (0x10 0x7D)
 - MDC_DEV_SUB_SPEC_PROFILE_CONTACTCLOSURE_SENSOR (0x10 0x7E)
 - MDC_DEV_SUB_SPEC_PROFILE_USAGE_SENSOR (0x10 0x7F)
 - MDC_DEV_SUB_SPEC_PROFILE_SWITCH_SENSOR (0x10 0x80)
 - MDC_DEV_SUB_SPEC_PROFILE_DOSAGE_SENSOR (0x10 0x81)
 - MDC_DEV_SUB_SPEC_PROFILE_TEMP_SENSOR (0x10 0x82)
- Profiles for Basic Electrocardiograph specialization: If the agent supports any of the profiles defined for Basic Electrocardiograph THEN the agent shall support Basic Electrocardiograph specialization:
 - MDC_DEV_SUB_SPEC_PROFILE_ECG (0x10 0x8C)
 - MDC_DEV_SUB_SPEC_PROFILE_HR (0x10 0x8D)

h. IF Attribute-Value-Map is present

- attribute-id = MDC_ATTR_ATTRIBUTE_VAL_MAP (0X0A 0X55)
- attribute-type = AttrValMap

- ❑ attribute-value.count = M
- ❑ attribute-value.length = M*4 bytes
- ❑ attribute-value = <check that M attributes are defined here>

i. IF MDS TimeInfo is present

- ❑ attribute-id = MDC_ATTR_MDS_TIME_INFO (0X0A 0X45)
- ❑ attribute-type = MdsTimeInfo
- ❑ attribute-value.length = 16 bytes
- ❑ attribute-value =
 - mds-time-cap-state:
 - IF (C_AG_OXP_006 = TRUE) THEN mds-time-capab-real-time-clock = 1 ELSE mds-time-capab-real-time-clock = 0
 - IF (C_AG_OXP_008 = TRUE) THEN mds-time-capab-set-clock = 1 ELSE mds-time-capab-set-clock = 0
 - IF (C_AG_OXP_010 = TRUE) THEN mds-time-capab-relative-time = 1 ELSE mds-time-capab-relative-time = 0
 - IF (C_AG_OXP_011 = TRUE) THEN mds-time-capab-high-res-relative-time = 1 ELSE mds-time-capab-high-res-relative-time = 0
 - IF (C_AG_OXP_014 = TRUE) THEN mds-time-capab-bo-time = 1 ELSE mds-time-capab-bo-time = 0
 - IF (C_AG_OXP_015 = TRUE) THEN mds-time-bo-time-utc-aligned(14) = 1 ELSE mds-time-bo-time-utc-aligned(14) = 0
 - IF (C_AG_OXP_007 = TRUE AND C_AG_OXP_009 = TRUE) THEN mds-time-capab-sync-abs-time = 1 or 0 ELSE mds-time-capab-sync-abs-time = 0
 - IF (C_AG_OXP_007 = TRUE AND C_AG_OXP_010 = TRUE) THEN mds-time-capab-sync-rel-time = 1 or 0 ELSE mds-time-capab-sync-rel-time = 0
 - IF (C_AG_OXP_007 = TRUE AND C_AG_OXP_011 = TRUE) THEN mds-time-capab-sync-hi-res-relative-time = 1 or 0 ELSE mds-time-capab-sync-hi-res-relative-time = 0.
 - IF (C_AG_OXP_007 = TRUE AND C_AG_OXP_014 = TRUE) THEN mds-time-capab-sync-bo-time = 1 or 0 ELSE mds-time-capab-sync-bo-time = 0
 - IF (C_AG_OXP_007 = TRUE) THEN mds-time-capab-sync-abs-time = 1 OR mds-time-capab-sync-res-time = 1 OR mds-time-capab-sync-hi-res-relative-time = 1 OR mds-time-capab-sync-bo-time = 1
 - Only one of mds-time-capab-real-time-clock and mds-time-capab-bo-time bits shall be set to 1.
 - Only one of mds-time-capab-sync-abs-time and mds-time-capab-sync-bo-time bits shall be set to 1.
 - Only one of mds-time-state-abs-time-synced and mds-time-state-bo-time-synced shall be set to 1
 - Time-sync-protocol:
 - IF (C_AG_OXP_007 = FALSE) THEN time-sync-protocol = MDC_TIME_SYNC_NONE ELSE time-sync-protocol = (MDC_TIME_SYNC_NTPV3 or MDC_TIME_SYNC_NTPV4 or MDC_TIME_SYNC_SntpV4 or MDC_TIME_SYNC_SntpV4330 or MDC_TIME_SYNC_BTV1)
 - Time-sync-accuracy:
 - IF (C_AG_OXP_007 = FALSE) THEN time-sync-accuracy = 0xFFFFFFFF
 - Time-resolution-abs-time:

- IF (C_AG_OXP_009 = FALSE AND C_AG_OXP_014 = FALSE) THEN time-resolution-abs-time = 0x0000
 - Time-resolution-rel-time:
 - IF (C_AG_OXP_010 = FALSE) THEN time-resolution-rel-time = 0x0000
 - Time-resolution-high-res-time:
 - IF (C_AG_OXP_011 = FALSE) THEN time-resolution-high-res-time = 0x0000
- j. IF attribute Date-and-Time is present
- attribute-id = MDC_ATTR_TIME_ABS (0x09 0x87)
 - attribute-type = AbsoluteTime
 - attribute-value.length = 8 bytes
 - attribute-value =
 - century =
 - year ≤ 99
 - month ≤ 12
 - day ≤ 31
 - hour ≤ 24
 - minute ≤ 60
 - second ≤ 60
 - sec-fractions ≤ 100
 - If Date-and-Time is present THEN Base-Offset-Time shall not be present.
- k. IF (C_AG_OXP_014 = TRUE) THEN Base-Offset-Time attribute shall be present ELSE Base-Offset-Time attribute shall not be present
- attribute-id = MDC_ATTR_TIME_BO (0x0A 0x81)
 - attribute-type = BaseOffsetTime
 - attribute-value.length = 8 bytes
 - If Base-Offset-Time is present THEN Date-and-Time shall not be present.
- l. IF Relative-Time attribute is present
- attribute-id = MDC_ATTR_TIME_REL (0x09 0x8F)
 - attribute-type = RelativeTime
 - attribute-value.length = 4 bytes
 - Verify that C_AG_OXP_010 is set to True
- m. IF HiRes-Relative-Time attribute is present
- attribute-id = MDC_ATTR_TIME_REL_HI_RES (0x09 0xE9)
 - attribute-type = HighResRelativeTime
 - attribute-value.length = 8 bytes
- n. IF Date-and-Time-Adjustment attribute is present
- attribute-id = MDC_ATTR_TIME_ABS_ADJUST (0x0A 0X62)_
 - attribute-type = AbsoluteTimeAdjust
 - attribute-value.length = 6 bytes
 - attribute-value = 0
 - Note: If queried with Get MDS command, this attribute shall be not present or 0.
- o. IF Production-Specification attribute is present
- attribute-id = MDC_ATTR_ID_PROD_SPECN (0X09 0X2D)

- attribute-type = ProductionSpec
- attribute-value.length = <even>
- attribute-value = <Vendor specific>

p. IF Power-Status attribute is present

- attribute-id = MDC_ATTR_POWER_STAT (0X09 0X55)
- attribute-type = PowerStatus
- attribute-value.length = 2 bytes
- attribute-value =
 - ON_MAINS (0x8000) or ON_BATTERY(0x4000)

Only one of the following may be active:

- chargingFull(8),
- chargingTrickle(9),
- chargingOff(10).

q. IF Battery-Level attribute is present

- attribute-id = MDC_ATTR_VAL_BATT_CHARGE (0X09 0X9C)
- attribute-type = INT-U16
- attribute-value.length = 2 bytes
- attribute-value = <value between 0 and 100> If value >100, the meaning of the value is "undefined"

r. IF Remain-Battery-Time attribute is present

- attribute-id = MDC_ATTR_TIME_BATT_REMAIN (0X09 0X88)
- attribute-type = BatMeasure
- attribute-value.length = 6 bytes
- attribute-value = <4 bytes to define the value. 2 remaining bytes to define the units, which shall be set to one of: MDC_DIM_MIN (0x08 0xA0), MDC_DIM_HR (0x08 0xC0) or MDC_DIM_DAY (0x08 0xE0) >

s. IF attribute Reg Cert Data List is present

- attribute-id = MDC_ATTR_REG_CERT_DATA_LIST (0X0A 0X4B)
- attribute-type = RegCertDataList
- attribute-value.length = < Variable to be checked>
- attribute-value = <Depends on the Autorization body, checked on the Design GuideLines>

t. IF Confirm Timeout attribute is present:

- attribute-id = MDC_ATTR_CONFIRM-TIMEOUT (0x09 0x14)
- attribute-type = RelativeTime
- attribute-value.length = 4 bytes

Furthermore, if MDS Scan Event Reports are sent by the Agent to report data for MDS object:

7. Wait for a Scan Event Report from the agent.

- a. If the agent sends Fixed Format Event Report for an MDS object, an Attribute-Value-Map has to be received in the GET response.
- b. If the agent sends Variable Format Event Report for an MDS object, the attributes whose values can be reported will be the attributes defined as dynamic\observational: Attribute-Value-Map, Mds-Time-Info, Date-and-Time or Base-Offset-Time, Relative-Time, HiRes-Relative-Time, Date-and-Time-Adjustment, Power-Status, Battery-Level, Remaining-Battery-Time. Dynamic attribute Confirm-Timeout is recommended not to be present. If Static attributes are present, the value shall remain unchanged. Handle, System-Type, System-Model, System-Id, Dev-Configuration-Id, Production-Specification, Reg-Cert-Data-List,

System-Type-Spec-List.	
Pass/Fail criteria	<ul style="list-style-type: none"> • All checked values are as specified in the test procedure. • The total size of the response can not exceed the sum of the APDU sizes of the supported specializations (limited to an absolute limit of 64512 octets): <ul style="list-style-type: none"> ○ Pulse oximeter -> 9216 octets ○ Weighing scales -> 896 octets ○ Glucose meter -> 5120 octets or 64512 octets if agent supports PM-Store ○ Blood pressure -> 896 octets ○ Thermometer -> 896 octets ○ Independent activity hub -> 5120 octets ○ Cardiovascular -> 64512 octets or 6624 octets if it supports Step Counter Profile ○ Strength -> 64512 octets ○ Adherence monitor -> 1024 octets ○ Peak flow -> 2030 octets ○ Body composition analyser -> 7730 octets ○ Basic ECG/Simple ECG -> 7168 octets or 64512 octets if the agent supports PM-Store ○ Basic ECG/Heart rate -> 1280 octets or 64512 octets if the agent supports PM-Store ○ International normalized ratio -> 896 octets or 64512 if the agent supports PM-Store
Notes	

TP Id	TP/PLT/AG/OXP/DIM/BV-000_B		
TP label	MDS Object: Dev-Configuration-Id and System-Id semantic feature		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	MDSClassAttr 6; M	ConfNormalProc 2; M
			ConfNormalProc 7 ; M
Applicability	C_AG_OXP_000		
Initial condition	The simulated manager and the agent under test are in the unassociated state.		
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test with a dev-config-id and a system-id. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent responds with a roiv-cmip-confirmed-event report message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 4. Record the field config-report-id. 5. If the config-report-id is that of the configuration being tested, the simulated manager responds with a rors-cmip-confirmed-event-report with result "accepted", else the manager responds with result "unsupported-config". 6. Repeat step 5 until the config-report-id is set to the configuration being tested. 7. Once in the operating state, the simulated manager issues roiv-cmip-get command with the handle set to 0 (to request an MDS object) and the attribute-id-list set to 0 to indicate all attributes. 8. The agent responds with with a "rors-cmip-get" service message in which the attribute-list contains a list of all implemented attributes of the MDS object. 9. Disconnect the device and conect it again. The simulated manager receives an association request from the agent. Check the system-id. 		
Pass/Fail criteria	<ul style="list-style-type: none"> • Dev-Config-Id must be the same in step 1 and step 4 		

	<ul style="list-style-type: none"> • Dev-Config-Id must be the same in step 6 and in step 8 • System-Id must be the same in steps 1, 7 and 8
Notes	

TP Id	TP/PLT/AG/OXP/DIM/BV-000_C		
TP label	MDS Object: Confirm-Timeout attribute, semantic feature		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	MDSClassAttr 19; M	OperErrorCond 5; M
		TimeOutVar 1; C	
Applicability	(C_AG_OXP_182 OR C_AG_OXP_183 OR C_AG_OXP_184 OR C_AG_OXP_189) AND C_AG_OXP_053 AND C_AG_OXP_000		
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<ol style="list-style-type: none"> 1. Record the Confirm-Timeout value from the Get MDS operation. If the attribute is not present in the MDS its value shall be TO_{cer-mds} (3s). 2. Take a measurement with the agent under test that will provoke an MDS event report to be sent. 3. The agent sends a "Remote Operation Invoke Confirmed Event Report". 4. The simulated manager does not respond for at least the time specified in the field Confirm-Timeout or 3 s if Confirm-Timeout is not supported. 		
Pass/Fail criteria	The agent must wait the specified time before unassociating.		
Notes			

TP Id	TP/PLT/AG/OXP/DIM/BV-001_A			
TP label	Static Numeric attributes derived from Metrics class and Dynamic Numeric attributes			
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	MetricClassAttr 1; M	MetricClassAttr 2; M	MetricClassAttr 3; O
		MetricClassAttr 4; M	MetricClassAttr 5; O	MetricClassAttr 7; O
		MetricClassAttr 8; O	MetricClassAttr 9; C	MetricClassAttr 10; O
		MetricClassAttr 11; O	MetricClassAttr 12; C	MetricClassAttr 13; O
		MetricClassAttr 14; O	MetricClassAttr 15; O	MetricClassAttr 16; C
		MetricClassAttr 17; C	MetricClassAttr 18; C	MetricClassAttr 19; O
		NumClass 1; M	NumClass 2; M	NumClassAttr 1; M
		NumClassAttr 2; C	NumClassAttr 3; C	NumClassAttr 4; C
		NumClassAttr 5; C	NumClassAttr 6; C	NumClassAttr 7; C
		NumClassAttr 8; O	ConfNormalProc 1;M	ConfEventRep 29; M
		ConfEventRep 30; M	ConfEventRep 31; C	ConfEventRep 33; O
ConfEventRep 33; O				
Spec	[ITU-T H.810]			
Testable items	Communication 6; M	General 2; M		
Applicability	C_AG_OXP_040 AND C_AG_OXP_000			
Initial condition	The simulated manager and the agent under test are in the unassociated state.			
Test procedure	<ol style="list-style-type: none"> 1. The agent under test sends and the simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 			

3. The agent responds with a "Remote Operation Invoke | Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager.
4. Numeric class attributes must be (ConfigReport -> ConfigObject-> AttributeList):
 - a. Mandatory attribute Handle shall not be present.
 - attribute-type = HANDLE
 - attribute-length = 2 bytes
 - attribute-value = must be unique and non-zero. Actual value may be specified by the Device Specilization.
 - b. Mandatory attribute Type shall be present in ConfigReport
 - attribute-id = MDC_ATTR_ID_TYPE (0X09 0X2F)
 - attribute-type = TYPE
 - attribute-value = SEQUENCE OF (SIZE 6)
 - attribute-value = <Not relevant in this test>
 - c. Mandatory attribute Metric-Spec-Small should be present
 - attribute-id = MDC_ATTR_METRIC_SPEC_SMALL
 - attribute-type = MetricSpecSmall
 - attribute-value.length = 2 bytes
 - attribute-value =
 - IF C_AG_OXP_201=TRUE and C_AG_OXP_041=FALSE THEN mss-avail-stored-data(1)=1(There is at least one object that has mss-avail.stored-data(1)=1)
 - IF C_AG_OXP_201=TRUE and C_AG_OXP_041=TRUE THEN mss-avail-stored-data(1)=1 or 0
 - d. Only one attribute of Metric-Id and Metric-Id-List shall be present.
 - e. If Metric-Id-List attribute is supported, it should be present in ConfigReport
 - attribute-id = MDC_ATTR_ID_PHYSIO_LIST
 - attribute-type = MetricIdList
 - attribute-value.length = SEQUENCE OF (SIZE 2)
 - attribute-value =
 - The [Metric-Id-List] attribute shall be used if a compound observed value is used, which does not incorporate the Metric-Id directly. The order of the Metric-Id-List shall correspond to the order of the elements in the compound observed value.
 - IF the agent supports Metric-Id-List at least for a Numeric object (C_AG_OXP_190 =TRUE) THEN this attribute shall be present at least for one object, ELSE this attribute is not present.
 - f. IF Metric-Id attribute is supported, it should be present in ConfigReport
 - attribute-id = MDC_ATTR_ID_PHYSIO
 - attribute-type = OID-Type
 - attribute-value.length = 2 bytes
 - g. IF Attribute-Value-Map is supported, it should be present in ConfigReport
 - attribute-id = MDC_ATTR_ATTRIBUTE_VAL_MAP (0X0A 0X5A)
 - attribute-type = AttrValMap
 - attribute-count = n (record for next attribute field)
 - attribute-value.length = n*4 bytes
 - attribute-value = <n attributes will be declared, each one will have 4 bytes composed of:
 - attribute-id = 2 bytes (MDC_ATTR_*).

- Attribute-length = 0x00 0x02: (2 bytes to declare the length of the attribute, but the contents of the attribute in the event report is not these 2 bytes length)

IF Supplemental-Types attribute is supported, it shall be present in ConfigReport:

- attribute-id = MDC_ATTR_SUPPLEMENTAL_TYPES
- attribute-type = SupplementalTypeList
- attribute-value.length = SEQUENCE OF (SIZE (4))
- attribute-value = <Not relevant in this test>

h. IF attribute Metric-Structure-Small is supported, it should be present in ConfigReport:

- attribute-id = MDC_ATTR_METRIC_STRUCTURE_SMALL
- attribute-type = MetricStructureSmall
- attribute-value.length = 2 bytes
- attribute-value =
 - ms-struct = one of the following:
 - ms-struct-simple (0x00)
 - ms-struct-compound (0x01)
 - ms-struct-reserved (0x02)
 - ms-struct-compound-simple (0x03)
 - ms-compound-no = one of the following:
 - IF ms-struct = ms-struct-simple THEN = 0
 - ELSE = maximum number of components in a compound value

i. IF attribute Metric-Id-Partition is supported, it should be present in ConfigReport

- attribute-id = MDC_ATTR_METRIC_ID_PART
- attribute-type = NomPartition
- attribute-value.length = 2 bytes
- attribute-value = one of the next
 - nom-part-unspec (0x00 0x00)
 - nom-part-obj (0x00 0x01)
 - nom-part-metric (0x00 0x02)
 - nom-part-alert (0x00 0x03)
 - nom-part-dim (0x00 0x04)
 - nom-part-vattr (0x00 0x05)
 - nom-part-pgrp (0x00 0x06)
 - nom-part-sites (0x00 0x07)
 - nom-part-infrastruc (0x00 0x08)
 - nom-part-fef (0x00 0x09)
 - nom-part-ecg-extn (0x00 0x0A)
 - nom-part-phd-dm (0x00 0x80)
 - nom-part-phd-hf (0x00 0x81)
 - nom-part-phd-ai (0x00 0x82)
 - nom-part-ret-code(0x00 0xFF)
 - nom-part-ext-nom (0x01 0x00)
 - nom-part-priv (0x04 0x00)

- j. IF attribute Unit-Code is supported, it should be present in ConfigReport:
 - attribute-id = MDC_ATTR_UNIT_CODE
 - attribute-type = OID-Type
 - attribute-value.length = 2 bytes
 - attribute-value = One of MDC_PART_DIM (may be defined in the specialization)
- k. IF attribute Source-Handle-Reference is supported, it should be present in ConfigReport:
 - attribute-id = MDC_ATTR_SOURCE_HANDLE_REF
 - attribute-type = HANDLE
 - attribute-value.length = 2 bytes
 - attribute-value = < The value of an existing object's handle >
- l. IF attribute Label-String is supported, it should be present in ConfigReport:
 - attribute-id = MDC_ATTR_ID_LABEL_STRING
 - attribute-type = OCTET STRING
 - attribute-value.length =
 - attribute-value = <Textual representation of attribute Type, Printable ASCII>
- m. IF attribute Unit-Label-String is supported, it should be present in ConfigReport:
 - attribute-id = MDC_ATTR_UNIT_LABEL_STRING
 - attribute-type = OCTET STRING
 - attribute-value.length =
 - attribute-value = <Textual representation of attribute Unit-Code, Printable ASCII>
- n. IF attribute Accuracy is supported, it shall be present in ConfigReport:
 - attribute-id = MDC_ATTR_NU_ACCUR_MSMT
 - attribute-type = FLOAT-Type
 - attribute-value.length = 4 bytes
 - attribute-value = <Not relevant in this test>
- o. IF attribute Measure-Active-Period is supported, it should be present in ConfigReport:
 - attribute-id = MDC_ATTR_TIME_PD_MSMT_ACTIVE
 - attribute-type = FLOAT-Type
 - attribute-value.length = 4 bytes
 - attribute-value = <Not relevant in this test>
- p. Time-stamp attributes (Absolute-Time-Stamp, Base-Offset-Time, Relative-Time-Stamp and HiRes-Time-Stamp), observed values attributes (Simple-Nu-Observed-Value, Basic-Nu-Observed-Value, Nu-Observed-Value, Compound-Simple-Nu-Observed-Value, Compound-Basic-Nu-Observed-Value, Compound-Nu-Observed-Value) and Measurement-Status shall not be present if mss-cat-setting and/or mss-cat-manual bit of the Metric-Spec-Small attribute is set to 0. If any of these bits is set to 1, observational attributes may be present in ConfigReport.

Furthermore if MDS event reports are sent by the agent, (the agent sends fixed format value messages to report dynamic data for Numeric Objects or uses variable format event report, C_AG_OXP_182 = TRUE or C_AG_OXP_189 = TRUE):

- 5. Take a measurement with the agent
- 6. Wait for an event report from the agent
 - a. One of {Simple-Nu-Observed-Value, Basic-Nu-Observed-Value, Nu-Observed-Value, Compound-Nu-Observed-Value, Compound-Simple-Nu-Observed-Value, Compound-Basic-Nu-Observed-Value} shall be present

- b. IF attribute Simple-Nu-Observed-Value is present
 - attribute-id = MDC_ATTR_NU_VAL_OBS_SIMP
 - attribute-type = SimpleNuObsValue
 - attribute-value.length = 4 bytes
 - attribute-value = <Not relevant in this test>
- c. IF attribute Basic-Nu-Observed-Value is present
 - attribute-id = MDC_ATTR_NU_VAL_OBS_BASIC
 - attribute-type = BasicNuObsValue
 - attribute-value.length = 2 bytes
 - attribute-value = <Not relevant in this test>
- d. IF attribute Nu-Observed-Value is present
 - attribute-id = MDC_ATTR_NU_VAL_OBS
 - attribute-type = NuObsValue
 - attribute-value.length = 10 bytes
 - attribute-value = <Not relevant in this test>
- e. IF attribute Compound-Simple-Nu-Observed-Value is present
 - attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_SIMP
 - attribute-type = SimpleNuObsValueCmp
 - attribute-value.length = SEQUENCE OF (SIZE (4))
 - attribute-value = <Not relevant in this test>
- f. IF attribute Compound-Basic-Nu-Observed-Value is present
 - attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_BASIC
 - attribute-type = BasicNuObsValueCmp
 - attribute-value.length = SEQUENCE OF (SIZE(4))
 - attribute-value = <Not relevant in this test>
 IF attribute Compound-Nu-Observed-Value is present
 - attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS
 - attribute-type = NuObsValueCmp
 - attribute-value.length = SEQUENCE OF (SIZE (10))
 - attribute-value = <Not relevant in this test>
- g. IF attribute Measure-Active-Period is present
 - attribute-id = MDC_ATTR_TIME_PD_MSMT_ACTIVE
 - attribute-type = FLOAT-Type
 - attribute-value.length = 4 bytes
 - attribute-value = <Not relevant in this test>
- h. IF attribute Absolute-Time-Stamp is present
 - attribute-id = MDC_ATTR_TIME_STAMP_ABS
 - attribute-type = AbsoluteTime
 - attribute-value.length = 8 bytes
 - attribute-value = <Not relevant in this test>
 - century =
 - year ≤ 99
 - month ≤ 12
 - day ≤ 31

	<ul style="list-style-type: none"> ▪ hour ≤ 24 ▪ minute ≤ 60 ▪ second ≤ 60 ▪ sec-fractions ≤ 100 <ul style="list-style-type: none"> <input type="checkbox"/> If an agent stores data, it shall associate a time stamp with the data. <input type="checkbox"/> If an agent supports Absolute-Time-Stamp, Base-Offset-Time-Stamp attribute shall not be supported. <p>i. IF (C_AG_OXP_014 = TRUE) THEN Attribute Base-Offset-Time-Stamp may be present ELSE it shall not be present</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_STAMP_BO <input type="checkbox"/> attribute-type = BaseOffsetTime <input type="checkbox"/> attribute-value.length = 8 bytes <input type="checkbox"/> attribute-value = <Not relevant in this test> <input type="checkbox"/> If an agent stores data, it shall associate a time stamp with the data. <input type="checkbox"/> If an agent supports Base-Offset-Time-Stamp, Absolute-Time-Stamp attribute shall not be supported. <p>j. IF attribute Relative-Time-Stamp is present</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_STAMP_REL <input type="checkbox"/> attribute-type = RelativeTime <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant in this test> <input type="checkbox"/> If an Agent stores data, it shall associate a time stamp with the data <p>k. IF attribute HiRes-Time-Stamp is present</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_STAMP_REL_HI_RES <input type="checkbox"/> attribute-type = HighResRelativeTime <input type="checkbox"/> attribute-value.length = 8 bytes <input type="checkbox"/> attribute-value = <Not relevant in this test> <p>l. IF attribute Measurement Status is present</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_MSMT_STAT <input type="checkbox"/> attribute-type = MeasurementStatus <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <not relevant for this test> <p>m. Check dynamic attributes that can be present: Metric-Structure-Small, Metric-Id, Metric-Id-List, Metric-Id-Partition, Unit-Code, Source-Handle-Reference, Label-String, Unit-Label-String, Measure-Active-Period</p>
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	<p>Checking that "IF C_AG_OXP_201=FALSE THEN mss-avail-stored data(1)=0" for MetricSpecSmall attribute has been removed because [ISO/IEEE 11073-20601A] has specified (subsection A.11.3) that "The setting of bits 0 to 5 is primarily informational and shall be set if the condition is true but- a manager cannot assume that if they are set the behavior will be observed".</p> <p>According to Device Specialization spec, standard configurations (most of them) state the value for bit mss-avail-stored-data to 1, even if the agent can not store measurements.</p> <p>So, if the test procedure checks that if C_AG_OXP_201=FALSE then mss-avail-stored-data =0 the test case will give a FAIL verdict and it would not be correct according to what [ISO/IEEE 11073-20601A] states in clause A.11.3.</p>

TP Id	TP/PLT/AG/OSP/DIM/BV-001_B			
TP label	Static Enumerated attributes derived from Metrics class and Dynamic Enumerated attributes			
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	MetricClassAttr 1; M	MetricClassAttr 2; M	MetricClassAttr 3; O
		MetricClassAttr 4; M	MetricClassAttr 5; O	MetricClassAttr 7; O
		MetricClassAttr 8; O	MetricClassAttr 9; C	MetricClassAttr 10; O
		MetricClassAttr 11; O	MetricClassAttr 12; C	MetricClassAttr 13; O
		MetricClassAttr 14; O	MetricClassAttr 15; O	MetricClassAttr 16; C
		MetricClassAttr 17; C	MetricClassAttr 18; C	MetricClassAttr 19; O
		EnumClass 2; M	EnumClass 3; M	EnumClassAttr 1; M
		EnumClassAttr 2; C	EnumClassAttr 3; C	EnumClassAttr 4; C
		EnumClassAttr 5; C	EnumClassAttr 6; C	EnumClassAttr 7; O
		ConfNormalProc 1; M	ConfEventRep 29; M	ConfEventRep 30; M
	ConfEventRep 31; C	ConfEventRep 33; O		
Spec	[ITU-T H.810]			
Testable items	Communication 6; M	General 2; M		
Applicability	C_AG_OXP_043 AND C_AG_OXP_000			
Initial condition	The simulated manager and the agent under test are in the unassociated state.			
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 4. Enumeration object attributes must be (ConfigReport -> ConfigObject-> AttributeList): <ol style="list-style-type: none"> a. Mandatory attribute Handle shall not be present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-type = HANDLE <input type="checkbox"/> attribute-value = 2 bytes <input type="checkbox"/> attribute-value = must be unique and non-zero. Actual value may be specified by the Device Specilization. b. Mandatory attribute Type shall be present in ConfigReport: <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE (0X09 0X2F) <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = SEQUENCE OF (SIZE 6) <input type="checkbox"/> attribute-value = <Not relevant in this test> c. Mandatory attribute Metric-Spec-Small should be present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Checked in the specializations > <ul style="list-style-type: none"> ▪ IF C_AG_OXP_202=TRUE and C_AG_OXP_041=FALSE THEN mss-avail-stored-data(1)=1 (There is at least one object that has mss-avail-stored-data(1) = 1) ▪ IF C_AG_OXP_202=TRUE and C_AG_OXP_041=TRUE THEN mss-avail-stored-data(1)=1 or 0 d. Only one attribute of Metric-Id and Metric-Id-List shall be present. 			

- e. IF Metric-Id-List attribute is supported, it should be present in ConfigReport:
- attribute-id = MDC_ATTR_ID_PHYSIO_LIST
 - attribute-type = MetricIdList
 - attribute-value.length = SEQUENCE OF (SIZE 2)
 - attribute-value =
 - The [Metric-Id-List] attribute shall be used if a compound observed value is used, which does not incorporate the Metric-Id directly. The order of the Metric-Id-List shall correspond to the order of the elements in the compound observed value.
- f. IF Metric-Id attribute is supported, it should be present in ConfigReport:
- attribute-id = MDC_ATTR_ID_PHYSIO
 - attribute-type = OID-Type
 - attribute-value.length = 2 bytes
- g. IF Attribute-Value-Map is supported, it should be present in ConfigReport:
- attribute-id = MDC_ATTR_ATTRIBUTE_VAL_MAP (0X0A 0X55)
 - attribute-type = AttrValMap
 - attribute-count = n (record for next attribute field)
 - attribute-value.length = n*4 bytes
 - attribute-value = <n attributes will be declared, each one will have 4 bytes composed of:
 - attribute-id = 2 bytes (MDC_ATTR_*)
 - Attribute-length = 0x00 0x02: (2 bytes to declare the length of the attribute, but the contents of the attribute in the event report is not these 2 bytes length)

IF Supplemental-Types attribute is supported, it shall be present in ConfigReport:

 - attribute-id = MDC_ATTR_SUPPLEMENTAL_TYPES
 - attribute-type = SupplementalTypeList
 - attribute-value.length = SEQUENCE OF (SIZE (4))
 - attribute-value = <Not relevant in this test>
- h. IF attribute Metric-Structure-Small is supported, it should be present in ConfigReport:
- attribute-id = MDC_ATTR_METRIC_STRUCTURE_SMALL
 - attribute-type = MetricStructureSmall
 - attribute-value.length = 2 bytes
 - attribute-value =
 - ms-struct = one of the following:
 - ms-struct-simple (0x01)
 - ms-struct-compound (0x02)
 - ms-struct-reserved (0x03)
 - ms-struct-compound-simple (0x04)
 - ms-compound-no = one of the following:
 - IF ms-struct = ms-struct-simple THEN = 0
 - ELSE = maximum number of components in a compound value
- i. IF attribute Metric-Id-Partition is supported, it should be present in ConfigReport:
- attribute-id = MDC_ATTR_METRIC_ID_PART
 - attribute-type = NomPartition

- attribute-value.length = 2 bytes
- attribute-value = one of the next
 - nom-part-unspec (0x00 0x00)
 - nom-part-obj (0x00 0x01)
 - nom-part-metric (0x00 0x02)
 - nom-part-alert (0x00 0x03)
 - nom-part-dim (0x00 0x04)
 - nom-part-vattr (0x00 0x05)
 - nom-part-pgrp (0x00 0x06)
 - nom-part-sites (0x00 0x07)
 - nom-part-infrac (0x00 0x08)
 - nom-part-feb (0x00 0x09)
 - nom-part-ecg-extn (0x00 0x0A)
 - nom-part-phd-dm (0x00 0x80)
 - nom-part-phd-hf (0x00 0x81)
 - nom-part-phd-ai (0x00 0x82)
 - nom-part-ret-code(0x00 0xFF)
 - nom-part-ext-nom (0x01 0x00)
 - nom-part-priv (0x04 0x00)
- j. IF attribute Unit-Code is supported, it should be present in ConfigReport:
 - attribute-id = MDC_ATTR_UNIT_CODE
 - attribute-type = OID-Type
 - attribute-value.length = 2 bytes
 - attribute-value = One of MDC_PART_DIM (may be defined in the specialization)
- k. IF attribute Source-Handle-Reference is supported, it should be present in ConfigReport:
 - attribute-id = MDC_ATTR_SOURCE_HANDLE_REF
 - attribute-type = HANDLE
 - attribute-value.length = 2 bytes
 - attribute-value = < The value of an existing object's handle >
- l. IF attribute Label-String is supported, it should be present in ConfigReport
 - attribute-id = MDC_ATTR_ID_LABEL_STRING
 - attribute-type = OCTET STRING
 - attribute-value.length =
 - attribute-value = <Textual representation of attribute Type, Printable ASCII>
- m. IF attribute Unit-Label-String is supported, it should be present in ConfigReport:
 - attribute-id = MDC_ATTR_UNIT_LABEL_STRING
 - attribute-type = OCTET STRING
 - attribute-value.length =
 - attribute-value = <Textual representation of attribute Unit-Code, Printable ASCII>
- n. IF attribute Measure-Active-Period is supported, it should be present in ConfigReport:
 - attribute-id = MDC_ATTR_TIME_PD_MSMT_ACTIVE

- attribute-type = FLOAT-Type
 - attribute-value.length = 4 bytes
 - attribute-value =<Not relevant for this test>
- o. IF attribute Enum-Observed-Value-Partition is supported it should be present in ConfigReport:
- attribute-id = MDC_ATTR_ENUM_OBS_VAL_PART
 - attribute-type = NomPartition
 - attribute-value.length = 2 bytes
 - attribute-value = one of the next
 - nom-part-unspec (0x00 0x00)
 - nom-part-obj (0x00 0x01)
 - nom-part-metric (0x00 0x02)
 - nom-part-alert (0x00 0x03)
 - nom-part-dim (0x00 0x04)
 - nom-part-vattr (0x00 0x05)
 - nom-part-pgrp (0x00 0x06)
 - nom-part-sites (0x00 0x07)
 - nom-part-infrastruc (0x00 0x08)
 - nom-part-feb (0x00 0x09)
 - nom-part-ecg-extn (0x00 0x0A)
 - nom-part-phd-dm (0x00 0x80)
 - nom-part-phd-hf (0x00 0x81)
 - nom-part-phd-ai (0x00 0x82)
 - nom-part-ret-code(0x00 0xFF)
 - nom-part-ext-nom (0x01 0x00)
 - nom-part-priv (0x04 0x00)
- p. Time-stamp attributes (Absolute-Time-Stamp, Base-Offset-Time, Relative-Time-Stamp and HiRes-Time-Stamp), observed values attributes (Enum-Observed-Value-Simple-OID, Enum-Observed-Value-Simple-Bit-Str, Enum-Observed-Value-Simple-Str, Enum-Observed-Value-Basic-Bit-Str, Enum-Observed-Value) and Measurement-Status shall not be present if mss-cat-setting and/or mss-cat-manual bit of the Metric-Spec-Small attribute is set to 0. If any of these bits is set to 1, observational attributes may be present in ConfigReport.

Furthermore if MDS event reports are sent by the agent (the agent sends fixed format value messages to report dynamic data for Enumeration Objects or uses variable format event report, C_AG_OXP_183=TRUE or C_AG_OXP_189=TRUE):

5. Take a measurement with the agent.
6. Wait for an event report from the agent:
 - a. One of Enum-Observed-Value-Simple-OID, Enum-Observed-Value-Simple-Bit-Str, Enum-Observed-Value-Simple-Str, Enum-Observed-Value-Basic-Bit-Str, Enum-Observed-Value shall be present.
 - b. IF attribute Measure-Active-Period is present
 - attribute-id = MDC_ATTR_TIME_PD_MSMT_ACTIVE
 - attribute-type = FLOAT-Type
 - attribute-value.length = 4 bytes
 - attribute-value = <Not relevant for this test>
 - c. IF attribute Enum-Observed-Value-Simple-OID is present
 - attribute-id = MDC_ATTR_ENUM_OBS_VAL_SIMP_OID

- attribute-type = OID-Type
- attribute-value.length = 2 bytes
- attribute-value = <Not relevant in this test>
- d. IF attribute Enum-Observed-Value-Simple-Bit-Str is present
 - attribute-id = MDC_ATTR_ENUM_OBS_VAL_SIMP_BIT_STR
 - attribute-type = BITS-32
 - attribute-value.length = 4 bytes
 - attribute-value = <Not relevant in this test>
- e. IF attribute Enum-Observed-Value-Basic-Bit-Str is present
 - attribute-id = MDC_ATTR_ENUM_OBS_VAL_BASIC_BIT_STR
 - attribute-type = BITS-16
 - attribute-value.length = 2 bytes
 - attribute-value = <Not relevant in this test>
- f. IF attribute Enum-Observed-Value-Simple-Str is present
 - attribute-id = MDC_ATTR_ENUM_OBS_VAL_SIMP_STR
 - attribute-type = EnumPrintableString
 - attribute-value.length = <Variable>
 - attribute-value = <ASCII printable>
- g. IF attribute Enum-Observed-Value is present
 - attribute-id = MDC_ATTR_VAL_ENUM_OBS
 - attribute-type = EnumObsValue
 - attribute-value.length = <Variable>
 - attribute-value = <Checked in each specialization>
- h. IF attribute Enum-Observed-Value-Partition is present
 - attribute-id = MDC_ATTR_ENUM_OBS_VAL_PART
 - attribute-type = NomPartition
 - attribute-value.length = 2 bytes
 - attribute-value = one of the next
 - nom-part-unspec (0x00 0x00)
 - nom-part-obj (0x00 0x01)
 - nom-part-metric (0x00 0x02)
 - nom-part-alert (0x00 0x03)
 - nom-part-dim (0x00 0x04)
 - nom-part-vattr (0x00 0x05)
 - nom-part-pgrp (0x00 0x06)
 - nom-part-sites (0x00 0x07)
 - nom-part-infrastruc (0x00 0x08)
 - nom-part-fef (0x00 0x09)
 - nom-part-ecg-extn (0x00 0x0A)
 - nom-part-phd-dm (0x00 0x80)
 - nom-part-phd-hf (0x00 0x81)
 - nom-part-phd-ai (0x00 0x82)
 - nom-part-ret-code (0x00 0xFF)
 - nom-part-ext-nom (0x01 0x00)

- nom-part-priv (0x04 0x00)
- i. IF attribute Absolute-Time-Stamp is present
 - ❑ attribute-id = MDC_ATTR_TIME_STAMP_ABS
 - ❑ attribute-type = AbsoluteTime
 - ❑ attribute-value.length = 8 bytes
 - ❑ attribute-value =
 - century =
 - year ≤ 99
 - month ≤ 12
 - day ≤ 31
 - hour ≤ 24
 - minute ≤ 60
 - second ≤ 60
 - sec-fractions ≤ 100
 - ❑ If an agent stores data, it shall associate a time stamp with the data (This case will be tested in TP/PLT/AG/SER/BV-007)
 - ❑ If an agent supports Absolute-Time-Stamp, Base-Offset-Time-Stamp attribute shall not be supported.
- j. IF (C_AG_OXP_014 = TRUE) THEN Attribute Base-Offset-Time-Stamp may be present ELSE it shall not be present
 - ❑ attribute-id = MDC_ATTR_TIME_STAMP_BO
 - ❑ attribute-type = BaseOffsetTime
 - ❑ attribute-value.length = 8 bytes
 - ❑ attribute-value = <Not relevant in this test>
 - ❑ If an Agent stores data, it shall associate a time stamp with the data.
 - ❑ If an Agent supports Base-Offset-Time-Stamp, Absolute-Time-Stamp attribute shall not be supported.
- k. IF attribute Relative-Time-Stamp is present
 - ❑ attribute-id = MDC_ATTR_TIME_STAMP_REL
 - ❑ attribute-type = RelativeTime
 - ❑ attribute-value.length = 4 bytes
 - ❑ attribute-value = <Not relevant in this test>
- l. If an Agent stores data, it shall associate a time stamp with the data
- m. IF attribute HiResRelative-Time-Stamp is present
 - ❑ attribute-id = MDC_ATTR_TIME_STAMP_REL_HI_RES
 - ❑ attribute-type = HighResRelativeTime
 - ❑ attribute-value.length = 8 bytes
 - ❑ attribute-value = <Not relevant in this test>
- n. IF attribute Measurement Status is present
 - ❑ attribute-id = MDC_ATTR_MSMT_STAT
 - ❑ attribute-type = MeasurementStatus
 - ❑ attribute-value.length = 2 bytes
 - ❑ attribute-value = <not relevant for this test>
- o. Check dynamic attributes that can be present: Metric-Structure-Small, Metric-Id, Metric-Id-List, Metric-Id-Partition, Unit-Code, Source-Handle-Reference, Label-String, Unit-Label-String, Measure-Active-Period

Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	<p>Checking that "IF C_AG_OXP_202=FALSE THEN mss-avail-stored data(1)=0" for MetricSpecSmall attribute has been removed because [ISO/IEEE 11073-20601A] has specified (clause A.11.3) that "The setting of bits 0 to 5 is primarily informational and shall be set if the condition is true but-- a manager cannot assume that if they are set the behavior will be observed".</p> <p>According to the Device Specialization spec, standard configurations (most of them) state the value for bit mss-avail-stored-data to 1, even if the agent can not store measurements.</p> <p>So if test procedure checks that if C_AG_OXP_202=FALSE then mss-avail-stored-data =0 the test case will give a FAIL verdict, and it would not be correct according to what [ISO/IEEE 11073-20601A] states (clause A.11.3).</p>

TP Id	TP/PLT/AG/OXP/DIM/BV-001_C			
TP label	Static RealTime-SA attributes derived from Metrics class and and Dynamic RealTime-SA attributes.			
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	MetricClassAttr 1; M	MetricClassAttr 2; M	MetricClassAttr 3; O
		MetricClassAttr 4; M	MetricClassAttr 5; O	MetricClassAttr 7; O
		MetricClassAttr 8; O	MetricClassAttr 9; C	MetricClassAttr 10; O
		MetricClassAttr 11; O	MetricClassAttr 12; C	MetricClassAttr 13; O
		MetricClassAttr 14; O	MetricClassAttr 15; O	MetricClassAttr 16; C
		MetricClassAttr 17; C	MetricClassAttr 18; C	MetricClassAttr 19; O
		ArrayClass 1; M	ArrayClass 2; M	ArrayClassAttr 1; M
		ArrayClassAttr 2; M	ArrayClassAttr 3; M	ArrayClassAttr 4; M
		ConfNormalProc 1; M	ConfEventRep 29; M	ConfEventRep 30; M
		ConfEventRep 31; C	ConfEventRep 33; O	
Spec	[ITU-T H.810]			
Testable items	Communication 6; M	General 2; M		
Applicability	C_AG_OXP_042 AND C_AG_OXP_000			
Initial condition	The simulated manager and the agent under test are in the unassociated state.			
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 4. RealTime-SA object attributes must be(ConfigReport -> ConfigObject-> AttributeList): <ol style="list-style-type: none"> a. Mandatory attribute Handle shall not be present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-type = HANDLE <input type="checkbox"/> attribute-length = 2 bytes <input type="checkbox"/> attribute-value = must be unique and non-zero. Actual value may be specified by the Device Specilization. b. Mandatory attribute Type shall be present in ConfigReport: <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE (0X09 0X2F) <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = SEQUENCE OF (SIZE 6) <input type="checkbox"/> attribute-value = <Not relevant in this test> c. Mandatory attribute Metric-Spec-Small should be present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL 			

- attribute-type = MetricSpecSmall
 - attribute-value.length = 2 bytes
 - attribute-value = <Checked in the specializations>
 - Bit mss-avail-store-data must be set to 0
- d. Only one attribute of Metric-Id and Metric-Id-List shall be present.
- e. If Metric-Id-List attribute is supported, it should be present in ConfigReport:
- attribute-id = MDC_ATTR_ID_PHYSIO_LIST
 - attribute-type = MetricIdList
 - attribute-value.length = SEQUENCE OF (SIZE 2)
 - attribute-value =
 - The [Metric-Id-List] attribute shall be used if a compound observed value is used, which does not incorporate the Metric-Id directly. The order of the Metric-Id-List shall correspond to the order of the elements in the compound observed value.
- f. IF Metric-Id attribute is supported, it should be present in ConfigReport:
- attribute-id = MDC_ATTR_ID_PHYSIO
 - attribute-type = OID-Type
 - attribute-value.length = 2 bytes
- g. IF Attribute-Value-Map is supported, it should be present in ConfigReport:
- attribute-id = MDC_ATTR_ATTRIBUTE_VAL_MAP (0X0A 0X55)
 - attribute-type = AttrValMap
 - attribute-count = n (record for next attribute field)
 - attribute-value.length = n*4 bytes
 - attribute-value = <n attributes will be declared, each one will have 4 bytes composed of:
 - attribute-id = 2 bytes (MDC_ATTR_*)
 - Attribute-length = 0x00 0x02: (2 bytes to declare the length of the attribute, but the contents of the attribute in the event report is not these 2 bytes length

IF Supplemental-Types attribute is supported, it shall be present in ConfigReport:

 - attribute-id = MDC_ATTR_SUPPLEMENTAL_TYPES
 - attribute-type = SupplementalTypeList
 - attribute-value.length = SEQUENCE OF (SIZE (4))
 - attribute-value = <Not relevant in this test>
- h. IF attribute Metric-Structure-Small is supported, it should be present in ConfigReport:
- attribute-id = MDC_ATTR_METRIC_STRUCTURE_SMALL
 - attribute-type = MetricStructureSmall
 - attribute-value.length = 2 bytes
 - attribute-value =
 - ms-struct = one of the following:
 - ms-struct-simple (0x01)
 - ms-struct-compound (0x02)
 - ms-struct-reserved (0x03)
 - ms-struct-compound-simple (0x04)
 - ms-compound-no = one of the following:

- IF ms-struct = ms-struct-simple THEN = 0
 - ELSE = maximum number of components in a compound value
- i. IF attribute Metric-Id-Partition is supported, it should be present in ConfigReport:
- attribute-id = MDC_ATTR_METRIC_ID_PART
 - attribute-type = NomPartition
 - attribute-value.length = 2 bytes
 - attribute-value = one of the next
 - nom-part-unspec (0x00 0x00)
 - nom-part-obj (0x00 0x01)
 - nom-part-metric (0x00 0x02)
 - nom-part-alert (0x00 0x03)
 - nom-part-dim (0x00 0x04)
 - nom-part-vattr (0x00 0x05)
 - nom-part-pgrp (0x00 0x06)
 - nom-part-sites (0x00 0x07)
 - nom-part-infrac (0x00 0x08)
 - nom-part-fef (0x00 0x09)
 - nom-part-ecg-extn (0x00 0x0A)
 - nom-part-phd-dm (0x00 0x80)
 - nom-part-phd-hf (0x00 0x81)
 - nom-part-phd-ai (0x00 0x82)
 - nom-part-ret-code(0x00 0xFF)
 - nom-part-ext-nom (0x01 0x00)
 - nom-part-priv (0x04 0x00)
- j. IF attribute Unit-Code is supported, it should be present in ConfigReport:
- attribute-id = MDC_ATTR_UNIT_CODE
 - attribute-type = OID-Type
 - attribute-value.length = 2 bytes
 - attribute-value = One of MDC_PART_DIM (may be defined in the specialization)
- k. IF attribute Source-Handle-Reference is supported, it should be present in ConfigReport:
- attribute-id = MDC_ATTR_SOURCE_HANDLE_REF
 - attribute-type = HANDLE
 - attribute-value.length = 2 bytes
 - attribute-value = < The value of an existing object's handle >
- l. IF attribute Label-String is supported, it should be present in ConfigReport:
- attribute-id = MDC_ATTR_ID_LABEL_STRING
 - attribute-type = OCTET STRING
 - attribute-value.length =
 - attribute-value = <Textual representation of attribute Type, Printable ASCII>
- m. IF attribute Unit-Label-String is supported, it should be present in ConfigReport:
- attribute-id = MDC_ATTR_UNIT_LABEL_STRING
 - attribute-type = OCTET STRING
 - attribute-value.length =

- attribute-value = <Textual representation of attribute Unit-Code, Printable ASCII>
- n. IF attribute Measure-Active-Period is supported, it should be present in ConfigReport:
 - attribute-id = MDC_ATTR_TIME_PD_MSMT_ACTIVE
 - attribute-type = FLOAT-Type
 - attribute-value.length = 4 bytes
 - attribute-value = <Not relevant in this test>
- o. Mandatory attribute Sample-Period shall be present in ConfigReport:
 - attribute-id = MDC_ATTR_TIME_PD_SAMP
 - attribute-type = RelativeTime
 - attribute-value.length = 4 bytes
 - attribute-value = <Not relevant in this test>
- p. Mandatory attribute Scale-and-Range-Specification shall be present in ConfigReport:
 - attribute-id = MDC_ATTR_SCALE_SPECN_I8; MDC_ATTR_SCALE_SPECN_I16; MDC_ATTR_SCALE_SPECN_I32
 - attribute-type = ScaleRangeSpec8 OR ScaleRangeSpec16 OR ScaleRangeSpec32
 - attribute-value.length = 1, 2 OR 4 bytes, depending of the type
 - attribute-value = <Not relevant in this test>
- q. Mandatory attribute Sa-Specification shall be present in ConfigReport:
 - attribute-id = MDC_ATTR_SA_SPECN
 - attribute-type = SaSpec
 - attribute-value.length = 6 bytes
 - attribute-value = <Not relevant in this test>
- r. Time-stamp attributes (Absolute-Time-Stamp, Base-Offset-Time, Relative-Time-Stamp and HiRes-Time-Stamp), observed values attributes (Simple-Sa-Observed-Value) and Measurement-Status shall not be present if mss-cat-setting and/or mss-cat-manual bit of the Metric-Spec-Small attribute is set to 0. If any of these bits is set to 1, observational attributes may be present in ConfigReport.

Furthermore, if MDS event reports are sent by the agent (the agent sends fixed format value messages to report dynamic data for RT-SA Objects or uses variable format event report, C_AG_OXP_184=TRUE or C_AG_OXP_189=TRUE):

5. Take a measurement with the agent
6. Wait for an event report from the agent
 - a. IF attribute Absolute-Time-Stamp is present
 - attribute-id = MDC_ATTR_TIME_STAMP_ABS
 - attribute-type = AbsoluteTime
 - attribute-value.length = 8 bytes
 - attribute-value = <Not relevant in this test>
 - century =
 - year ≤ 99
 - month ≤ 12
 - day ≤ 31
 - hour ≤ 24
 - minute ≤ 60
 - second ≤ 60

	<ul style="list-style-type: none"> ▪ $\text{sec-fractions} \leq 100$ <input type="checkbox"/> If an agent stores data, it shall associate a time stamp with the data. <input type="checkbox"/> If an agent supports Absolute-Time-Stamp, Base-Offset-Time-Stamp attribute shall not be supported. b. IF (C_AG_OXP_014 = TRUE) THEN Attribute Base-Offset-Time-Stamp may be present ELSE it shall not be present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_STAMP_BO <input type="checkbox"/> attribute-type = BaseOffsetTime <input type="checkbox"/> attribute-value.length = 8 bytes <input type="checkbox"/> attribute-value = <Not relevant in this test> <input type="checkbox"/> If an Agent stores data, it shall associate a time stamp with the data. <input type="checkbox"/> If an Agent supports Base-Offset-Time-Stamp, Absolute-Time-Stamp attribute shall not be supported. c. IF attribute Measure-Active-Period is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_PD_MSMT_ACTIVE <input type="checkbox"/> attribute-type = FLOAT-Type <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant in this test> d. IF attribute Relative-Time-Stamp is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_STAMP_REL <input type="checkbox"/> attribute-type = RelativeTime <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant in this test> e. IF an Agent stores data, it shall associate a time stamp with the data f. IF attribute HiResRelative-Time-Stamp is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_STAMP_REL_HI_RES <input type="checkbox"/> attribute-type = HighResRelativeTime <input type="checkbox"/> attribute-value.length = 8 bytes <input type="checkbox"/> attribute-value = <Not relevant in this test> g. IF attribute Measurement Status is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_MSMT_STAT <input type="checkbox"/> attribute-type = MeasurementStatus <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <not relevant for this test> h. Mandatory attribute Simple-Sa-Observed-Value <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SIMP_SA_OBS_VAL <input type="checkbox"/> attribute-type = OCTET STRING <input type="checkbox"/> attribute-value.length = The length shall be even with padding bytes at the end. <input type="checkbox"/> attribute-value = <Not relevant in this test> i. Check dynamic attributes that may be present: Metric-Structure-Small, Metric-Id, Metric-Id-List, Metric-Id-Partition, Unit-Code,Source-Handle-Reference, Label-String, Unit-Label-String, Measure-Active-Period
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id		TP/PLT/AG/OXP/DIM/BV-001_D		
TP label		RealTime-SA: Sa-Specification semantic		
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	ArrayClassAttr 4; M		
Applicability		C_AG_OXP_042 AND C_AG_OXP_000		
Initial condition		The simulated manager and the agent under test are in the operating state.		
Test procedure		<ol style="list-style-type: none"> 1. Take a measurement with the agent of a value stored in a RT-SA object. 2. Wait until the manager receives an event report, the attributes of interest are: <ol style="list-style-type: none"> a. Scale-and-Range-Specification: <ul style="list-style-type: none"> <input type="checkbox"/> Attribute-value = See below <input type="checkbox"/> lower-*-value = <Record for comparison> <input type="checkbox"/> upper-*-value = <Record for comparison> b. Simple-Sa-Observed-Value <ul style="list-style-type: none"> <input type="checkbox"/> Attribute-value = <Record for comparison> 		
Pass/Fail criteria		<ul style="list-style-type: none"> • Upper values must be \geq than the lower values • Measurement Value must be in the allowed range. 		
Notes				

TP Id		TP/PLT/AG/OXP/DIM/BV-001_E		
TP label		Numeric attributes: Metric-Id-List Semantic		
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	MetricClassAttr 9; C		
Applicability		C_AG_OXP_190 AND C_AG_OXP_000		
Initial condition		The simulated manager and the agent under test are in the unassociated state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 4. Make the agent send measurements for objects that support Metric-Id-List. 5. Wait for event reports. 6. The Compound Simple and Basic Numeric observed value will be taken from the Event Report received in step 4 and this information and the Metric-Id-List will be showed to the operator using a pop-up. 		
Pass/Fail criteria		The operator checks in step 6 that the order of the Metric-Id-List corresponds to the order of the elements in the compound observed value.		
Notes		If the agent does not send the Metric-Id-List value in the ConfigReport, the agent will send it in an MDS-Event Report before sending an observation (compound value). SE is allowed for the Metric-Id-List.		

TP Id		TP/PLT/AG/OXP/DIM/BV-002_A		
TP label		PM-Store Object: Mandatory, Conditional and Optional Attributes 1		
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	StoreClass 1; M	StoreClassAttr 1; M	StoreClassAttr 2; M
		StoreClassAttr 3; M	StoreClassAttr 4; O	StoreClassAttr 5; O

		StoreClassAttr 6; M	StoreClassAttr 7; O	StoreClassAttr 8; C
		StoreClassAttr 9; M	StoreClassAttr 11; M	PM-StoreService 1; M
		PersStoreMtrDatTransf 25; M		
	Spec	[ITU-T H.810]		
	Testable items	Communication 6; M		
	Spec	[ISO/IEEE 11073-10472]		
	Testable items	MM_PMStoreAttr6; C		
	Spec	[IEEE 11073-10406]		
	Testable items	PerPMStoreAtt2; M	AperPMStoreAtt2; M	PMStoreServ1; M
		PMStoreServ2; M		
Applicability	C_AG_OXP_041 AND C_AG_OXP_000			
Initial condition	The simulated manager and the agent under test are in the unassociated state.			
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 4. Record the handle for the PM-Store object. 5. The simulated manager shall send a Get request for the PM-Store object with an attribute-id-list set to 0 to indicate all PM-Store attributes. 6. The agent issues a GET response with the PM-Store attributes it supports: Verify the invoke-id is mirrored from the Get request. <ol style="list-style-type: none"> a. Mandatory attribute Handle shall not be present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-type = HANDLE <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant in this test case> b. Mandatory attribute PM-Store-Capab <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_PM_STORE_CAPAB <input type="checkbox"/> attribute-type = PmStoreCapab <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = one or more of the following bits may be set: <ul style="list-style-type: none"> ▪ pmsc-var-no-of-segm (bit 0) ▪ IF C_AG_OXP_187 then bit 4 (pmsc-epi-seg-entries) must be set ▪ IF C_AG_OXP_188 then bit 5 (pmsc-peri-seg-entries) must be set ▪ pmsc-abs-time-select (bit 6) ▪ pmsc-clear-segm-by-list-sup (bit 7) ▪ pmsc-clear-segm-by-time-sup (bit 8) ▪ pmsc-clear-segm-remove (bit 9) ▪ pmsc-multi-person (bit 12) ▪ All other bits shall be set to zero c. Mandatory attribute Store-Sample-Algorithm <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_STORE_SAMPLE_ALG <input type="checkbox"/> attribute-type = StoSampleAlg <input type="checkbox"/> attribute-value.length = 2 bytes 			

- attribute-value = One of the next
 - st-alg-nos(0x00 0x00)
 - st-alg-moving-average(0x00 0x01)
 - st-alg-recursive(0x00 0x02)
 - st-alg-min-pick(0x00 0x03)
 - st-alg-max-pick(0x00 0x04)
 - st-alg-median(0x00 0x05)
 - st-alg-trended(0x02 0x00)
 - st-alg-no-downsampling(0x04 0x00)
 - st-alg-manuf-specific-start(0xF0 0x0061440)
 - st-alg-manuf-specific-end(0xFF 0xFF)
- d. Mandatory attribute Operational-State
 - attribute-id = MDC_ATTR_OP_STAT
 - attribute-type = OperationalState
 - attribute-value.length = 2 bytes
 - attribute-value = One of the next
 - disabled (0x00 0x00)
 - enabled (0x00 0x01)
 - notAvailable (0x00 0x02)
- e. Mandatory attribute Number-Of-Segments
 - attribute-id = MDC_ATTR_NUM_SEG
 - attribute-type = INT-U16
 - attribute-value.length = 2 bytes
 - attribute-value = <not relevant for this test>
- f. IF attribute Clear-Timeout
 - attribute-id = MDC_ATTR_CLEAR_TIMEOUT
 - attribute-type = RelativeTime
 - attribute-value.length = 4 bytes
 - attribute-value = <Not relevant in this test>
 - IF agent supports the clear segment action, Clear-Timeout attribute is mandatory.
- g. IF attribute Sample-Period is present
 - attribute-id = MDC_ATTR_TIME_PD_SAMP
 - attribute-type = RelativeTime
 - attribute-value.length = 4 bytes
 - attribute-value = <Not relevant in this test>
- h. IF Storage-Capacity-Count is present
 - attribute-id = MDC_ATTR_METRIC_STORE_CAPAC_CNT
 - attribute-type = INT-U32
 - attribute-value.length = 4 bytes
 - attribute-value = See relation with next attribute
- i. IF Storage-Usage-Count is present
 - attribute-id = MDC_ATTR_METRIC_STORE_USAGE_CNT
 - attribute-type = INT-U32

	<ul style="list-style-type: none"> <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = always ≤ than Storage-Capacity-Count <p>j. IF attribute PM-Store-Label</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_PM_STORE_LABEL_STRING <input type="checkbox"/> attribute-type = OCTET STRING <input type="checkbox"/> attribute-value.length = <input type="checkbox"/> attribute-value = Printable ASCII <p>7. The simulated manager shall send a Get-Segment-Info object action for the PM-Store object with SegmSelection set to all-segments.</p> <p>8. The agent issues a response (rors-cmip-confirmed-action) with the PM-Segment attributes it supports in the SegmentInfoList structure.</p>
Pass/Fail criteria	<ul style="list-style-type: none"> • All checked values are as specified in the test procedure. • IF in step 6.f the Sample-Period was not present it must be present in each PM-Segment.
Notes	

TP Id	TP/PLT/AG/OXP/DIM/BV-002_B			
TP label	PM-Store Object: Mandatory, Conditional and Optional Attributes 2			
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	StoreClassAttr 9; M	StoreClassAttr 2; M	StoreClassAttr 5; M
	Spec	[ISO/IEEE 11073-10472]		
	Testable items	MM_PMStoreAttr5; C		
Applicability	C_AG_OXP_041 AND C_AG_OXP_000			
Initial condition	The simulated manager and the agent under test are in the operating state.			
Test procedure	<p>1. Make sure there are no measurements being taken.</p> <p>2. The simulated manager shall send a Get request for the PM-Store object with an attribute-id-list set to 0 to indicate all PM-Store attributes.</p> <p>3. The agent issues a GET response with the PM-Store attributes. The attributes of interest are:</p> <p>a. Mandatory attribute PM-Store-Capab</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_PM_STORE_CAPAB <input type="checkbox"/> attribute-type = PmStoreCapab <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <ul style="list-style-type: none"> ▪ pmsc-var-no-of-segm. Record state for later comparison <p>b. Mandatory attribute Number-Of-Segments</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NUM_SEG <input type="checkbox"/> attribute-type = INT-U16 <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <record for later comparison> <p>4. The simulated manager shall send a Get-Segment-Info object action for the PM-Store object with SegmSelection set to all-segments:</p> <p>a. Data APDU</p> <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action, 			

	<ul style="list-style-type: none"> <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_GET_INFO <input type="checkbox"/> SegmSelection = all-segments <p>5. The agent issues a response (rors-cmip-confirmed-action) with the PM-Segment attributes it supports in the SegmentInfoList structure:</p> <ul style="list-style-type: none"> a. Verify the invoke-id is mirrored from the Get request a. b. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Response Confirmed Action, <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_GET_INFO <input type="checkbox"/> SegmentInfoList = <Attributes of the segments> <p>6. Record the number of existing Segments.</p> <p>7. If the agent can record measurements in PM-Store while it is connected then take measurements whose values are stored in a PM-Segment.</p> <p>8. Repeat steps 2 through 5.</p>
Pass/Fail criteria	<ul style="list-style-type: none"> ▪ In step 2.a, if bit pmsc-var-no-of-segm is not set, the number of segments stated in step 2.b and checked in step 5.b must remain unchanged. ▪ The PM-Store attribute Number-Of-Segments value must contain the exact number of segments recorded in step 6.
Notes	

TP Id	TP/PLT/AG/OXP/DIM/BV-002_C		
TP label	PM-Store Object: Clear-Timeout Semantics		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	StoreClassAttr 11; M	
Applicability	C_AG_OXP_041 AND C_AG_OXP_000 AND C_AG_OXP_071		
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<p>1. Take a measurement that would be placed in the PM-Store.</p> <p>2. The simulated manager shall send a Get request for the PM-Store object with an attribute-id-list set to 0 to indicate all PM-Store attributes.</p> <p>3. The agent issues a GET response with the PM-Store attributes.</p> <p>4. Record the value of the Clear-Timeout Attribute (time in seconds = attribute-value*125/1e6).</p> <p>5. The simulated manager sends a Segment Clear to one of the PM-Segments:</p> <ul style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action, <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_CLEAR <input type="checkbox"/> SegmSelection = all-segments <p>6. The agent under test operation response:</p> <p>Verify the invoke-id is mirrored from the Get request.</p> <ul style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Response Confirmed Action, <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_CLEAR 		

Pass/Fail criteria	The manager has to receive the confirmation in less than the value specified in the Clear-Timeout attribute.
Notes	

TP Id	TP/PLT/AG/OXP/DIM/BV-002_D		
TP label	PM-Store Object: Episodic Semantics		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	StoreClassAttr 2; M	
Applicability	C_AG_OXP_041AND C_AG_OXP_187 AND C_AG_OXP_000		
Initial condition	The simulated manager and the agent under test are in the unassociated state.		
Test procedure	<ol style="list-style-type: none"> 1. Make sure there are no measurements being taken. 2. The simulated manager receives an association request from the agent under test. 3. The simulated manager responds with a result = accepted-unknown-config. 4. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message: <ol style="list-style-type: none"> a. Event-type=MDC_NOTI_CONFIG 5. Check that the PM-Store-Capab attribute has the pmsc-epi-seg-entries bit set. 6. The simulated manager shall send a Get-Segment-Info object action for the PM-Store object with SegmSelection set to all-segments. 7. The agent shall respond to the Get-Segment-Info, indicating the attributes of the PM-Segment. 8. Check the PM-Segment-Entry-Map to make sure that a Time-Stamp is associated with the measurement data. 9. Take measurements with the agent under test. 10. The simulated manager sends a request for the PM-Segment Data to one of the PM-Segments that contains data (sends the Action MDC_ACT_SEG_TRIG_XFER). 11. The agent issues an action response. 12. The agent under test starts Data transfer: <ol style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Invoke CfmEventReport <input type="checkbox"/> Action = MDC_NOTI_SEGMENT_DATA <input type="checkbox"/> SegmentDataEvent 13. The simulated manager responds to transferred data APDU's. 		
Pass/Fail criteria	The PM-Segment-Entry-Map contains a Time-Stamp associated with measurement data and it has the correct format in the SegmentDataEvent received.		
Notes			

TP Id	TP/PLT/AG/OXP/DIM/BV-002_E			
TP Label	PM-Store Object: Mandatory, Conditional and Optional Attributes 3 Configuration			
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	StoreClass 1; M	StoreClassAttr 1; M	StoreClassAttr 2; M
		StoreClassAttr 3; M	StoreClassAttr 4; O	StoreClassAttr 5; O
		StoreClassAttr 6; M	StoreClassAttr 7; O	StoreClassAttr 8; C
		StoreClassAttr 9; M	StoreClassAttr 11; M	ConfEventRep 29; M
ConfEventRep 30; M		ConfEventRep 31; C	ConfEventRep 33; O	

		PM-StoreService 3; O		
	Spec	[ITU-T H.810]		
	Testable items	Communication 6; M		
Applicability		C_AG_OXP_041 AND C_AG_OXP_000		
Initial condition		The simulated manager and the agent under test are in the operating state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 4. The PM-Store object attributes must be(ConfigReport -> ConfigObject-> AttributeList): <ol style="list-style-type: none"> a. Mandatory attribute Handle shall not be present. <ul style="list-style-type: none"> <input type="checkbox"/> attribute-type = HANDLE <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = must be unique and non-zero. Actual value may be specified by the Device Specilization. b. Mandatory attribute PM-Store-Capab shall be present in ConfigReport: <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_PM_STORE_CAPAB <input type="checkbox"/> attribute-type = PmStoreCapab <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant in this test> c. Mandatory attribute Store-Sample-Algorithm shall be present in ConfigReport: <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_STORE_SAMPLE_ALG <input type="checkbox"/> attribute-type = StoSampleAlg <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant in this test> d. IF Storage-Capacity-Count is supported, it shall be present in ConfigReport: <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_STORE_CAPAC_CNT <input type="checkbox"/> attribute-type = INT-U32 <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant in this test> e. IF Storage-Usage-Count is supported, it should be present in ConfigReport: <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_STORE_USAGE_CNT <input type="checkbox"/> attribute-type = INT-U32 <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant in this test> f. Mandatory attribute Operational-State should be present in ConfigReport: <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_OP_STAT <input type="checkbox"/> attribute-type = OperationalState <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant in this test> g. IF attribute PM-Store-Label is supported, it shall be present in ConfigReport: <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_PM_STORE_LABEL_STRING <input type="checkbox"/> attribute-type = OCTET STRING <input type="checkbox"/> attribute-value.length = 		

	<ul style="list-style-type: none"> <input type="checkbox"/> attribute-value = <Not relevant in this test> <p>h. IF attribute Sample-Period is supported, it shall be present in ConfigReport:</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_PD_SAMP <input type="checkbox"/> attribute-type = RelativeTime <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant in this test> <p>i. Mandatory attribute Number-Of-Segments should be present in ConfigReport:</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NUM_SEG <input type="checkbox"/> attribute-type = INT-U16 <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <not relevant for this test> <p>j. If attribute Clear-Timeout is supported, it should be present in ConfigReport:</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_CLEAR_TIMEOUT <input type="checkbox"/> attribute-type = RelativeTime <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant in this test> <p>Furthermore if MDS event reports are sent by the agent:</p> <ol style="list-style-type: none"> 5. Take a measurement with the agent. 6. Wait for a variable format event report from the agent, check that dynamic attributes for PM-Store may be reported (Store-Usage-Count, Operational-State, Number-Of-Segments, Clear-Timeout).
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/AG/OXP/DIM/BV-003_A			
TP label	PM-Segment Object: Mandatory, Conditional and Optional Attributes			
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	PM-SegmAttr 1; M	PM-SegmAttr 2; M	PM-SegmAttr 3; C
		PM-SegmAttr 5; C	PM-SegmAttr 6; O	PM-SegmAttr 7; O
		PM-SegmAttr 8; O	PM-SegmAttr 9; O	PM-SegmAttr 10; O
		PM-SegmAttr 11; O	PM-StoreMeth 10; M	PM-StoreMeth 11; M
		PM-StoreMeth 1; M	PM-SegmAttr 19; C	PM-SegmAttr 20; C
	Spec	[IEEE 11073-10406]		
	Testable items	PerPMStoreAtt12; C	PMStoreObjMeth2; M	PerPMSegObj1; M
		PerPMSegObj2; M	PerPMSegObj3; C	PerPMSegObj4; M
		PerPMSegObj5; C	PerPMSegObj6; O	PerPMSegObj7; C
		PerPMSegObj8; C	PerPMSegObj9; C	PerPMSegObj10; C
		PerPMSegObj12; M	PerPMSegObj13; O	PerPMSegObj14; M
		PerPMSegObj15; O	PerPMSegObj16; M	PerPMSegObj17; C
		AperPMSegObj1; M	AperPMSegObj2; M	AperPMSegObj3; C
AperPMSegObj4; M		AperPMSegObj5; O	AperPMSegObj6; O	
AperPMSegObj7; M		AperPMSegObj8; M	AperPMSegObj9; C	
AperPMSegObj10; C		AperPMSegObj11; C	AperPMSegObj12; M	
AperPMSegObj13; O	AperPMSegObj14; M	AperPMSegObj15; O		

		AperPMSegObj16; M		
Coverage	Spec	[ITU-T H.810]		
	Testable items	General 2; M		
Applicability		C_AG_OXP_041 AND C_AG_OXP_000		
Initial condition		The simulated manager and the agent under test are in the operating state.		
Test procedure		<p>1. The simulated manager shall send a Get request for the PM-Store object with an attribute-id-list set to 0 to indicate all PM-Store attributes.</p> <p>2. The agent issues a GET response with the PM-Store attributes. Check for the existence of:</p> <ul style="list-style-type: none"> a. attribute Sample-Period is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_PD_SAMP <input type="checkbox"/> attribute-type = RelativeTime <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant in this test> <p>3. The simulated manager shall send a Get-Segment-Info object action for the PM-Segment object with SegmSelection = all-segments to indicate the PM-Segment attributes of all available PM-Segments.</p> <p>4. The agent issues a "rors-cmip-confirmed-action" response with the PM-Segment attributes it supports:</p> <p>Verify the invoke-id is mirrored from the Get request.</p> <ul style="list-style-type: none"> a. Mandatory attribute Instance-Number <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_INSTNO <input type="checkbox"/> attribute-type = InstNumber <input type="checkbox"/> attribute-length = 2 bytes <input type="checkbox"/> attribute-value = unique in its PM-Store (This is why we ask for all the attributes of all the PM-Segment) b. Mandatory attribute PM-Segment-Entry-Map <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_PM_SEG_MAP <input type="checkbox"/> attribute-type = PmSegmentEntryMap <input type="checkbox"/> attribute-value = SEQUENCE, it must match the entries c. Mandatory attribute Operational-State <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_OP_STAT <input type="checkbox"/> attribute-type = OperationalState <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = {disabled(0), enabled(1), notAvailable(2)} d. Mandatory attribute Transfer-Timeout <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TRANSFER_TIMEOUT <input type="checkbox"/> attribute-type = RelativeTime <input type="checkbox"/> attribute-value.length = 4 bytes e. IF attribute PM-Seg-Person-Id is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_PM_SEG_PERSON_ID <input type="checkbox"/> attribute-type = PersonId <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <input type="checkbox"/> If the PM-Store is able to store data for multiple persons it shall set the pmsc-multi-person bit in the PM-Store-Capab attribute. If this bit is set, all PM- 		

Segment instances contained in the PM-Store shall support the PM-Seg-Person-Id attribute, check with the attributes obtained in step 2.

- f. IF attribute Sample-Period is present
 - attribute-id = MDC_ATTR_TIME_PD_SAMP
 - attribute-type = RelativeTime
 - attribute-value.length = 4 bytes
 - attribute-value =
 - IF in step 2 Sample-Period was not present and values are sampled periodically this attribute must be present either in the PM-Store or alternatively in each PM-Segment. If values are sampled, then the pmsc-peri-seg-entries bit in the PM-Store-Capab attribute shall be set, check with attributes obtained in step 2 The [Sample-Period] attribute must be present in either the PM-Store or alternatively in each of the PM-Segments if values are sampled periodically - so the time difference for 2 entries in the Fixed-Segment-Data is constant (the pmsc-peri-seg-entries bit in the Pm-Store-Capab attribute is set).
- g. IF attribute Date-and-Time-Adjustment is present
 - attribute-id = MDC_ATTR_TIME_ABS_ADJUST (0x0A 0x62)
 - attribute-type = AbsoluteTimeAdjust
 - attribute-value.length = 6 bytes
 - attribute-value =
 - If the agent ever adjusts the Date-and-Time, this attribute reports the time adjustment.
- h. IF attribute Segment-Label is present
 - attribute-id = MDC_ATTR_PM_SEG_LABEL_STRING
 - attribute-type = OCTET STRING
 - attribute-value.length = consistent with value
 - attribute-value = <printable ASCII>
- i. IF(C_AG_OXP_009 = TRUE) THEN attribute Segment-Start-Abs-Time may be present ELSE it shall not be present
 - attribute-id = MDC_ATTR_TIME_START_SEG
 - attribute-type = AbsoluteTime
 - attribute-value.length = 8 bytes
 - attribute-value =
 - century =
 - year ≤ 99
 - month ≤ 12
 - day ≤ 31
 - hour ≤ 24
 - minute ≤ 60
 - second ≤ 60
 - sec-fractions ≤ 100
 - Note: This attribute is required if the agent supports actions on the segment by time (i.e., the pmsc-abs-time-select and/or the pmsc-clear-segm-bytime-sup bits are set) this attribute shall be present.
 - If this attribute is used, the Segment-Start-BO-Time shall not be used.
- j. IF (C_AG_OXP_009 = TRUE) THAN attribute Segment-End-Abs-Time may be present ELSE it shall not be present
 - attribute-id = MDC_ATTR_TIME_END_SEG
 - attribute-type = AbsoluteTime

	<ul style="list-style-type: none"> <input type="checkbox"/> attribute-value.length = 8 bytes <input type="checkbox"/> attribute-value = <ul style="list-style-type: none"> ▪ century = ▪ year ≤ 99 ▪ month ≤ 12 ▪ day ≤ 31 ▪ hour ≤ 24 ▪ minute ≤ 60 ▪ second ≤ 60 ▪ sec-fractions ≤ 100 <input type="checkbox"/> Note: This attribute is required if the agent supports actions on the segment by time (i.e., the pmsc-abs-time-select and/or the pmsc-clear-segm-bytime-sup bits are set) this attribute shall be present. <input type="checkbox"/> If this attribute is used, the Segment-End-BO-Time shall not be used. k. IF (C_AG_OXP_014 = TRUE) THEN attribute Segment-Start-BO-Time may be present ELSE it shall not be present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_START_SEG_BO <input type="checkbox"/> attribute-type = BaseOffsetTime <input type="checkbox"/> attribute-value.length = 8 bytes <input type="checkbox"/> attribute-value = <Not relevant in this test> <input type="checkbox"/> If this attribute is used, the Segment-Start-Abs-Time shall not be used. l. IF (C_AG_OXP_014 = TRUE) THEN attribute Segment-End-BO-Time may be present ELSE it shall not be present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_START_SEG_BO <input type="checkbox"/> attribute-type = BaseOffsetTime <input type="checkbox"/> attribute-value.length = 8 bytes <input type="checkbox"/> attribute-value = <Not relevant in this test> <input type="checkbox"/> If this attribute is used, the Segment-End-Abs-Time shall not be used m. IF attribute Segment-Usage-Count is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SEG_USAGE_CNT <input type="checkbox"/> attribute-type = INT-U32 <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = check that it has the number of stored entries n. IF attribute Segment-Statistics <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SEG_STATS <input type="checkbox"/> attribute-type = SegmentStatistics <input type="checkbox"/> attribute-value.length = must be consistent with EntryMap <input type="checkbox"/> attribute-value = o. IF attribute Confirm-Timeout is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_CONFIRM_TIMEOUT <input type="checkbox"/> attribute-type = RelativeTime <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <p>5. Repeat steps 3 and 4 for every Segment.</p>
Pass/Fail criteria	All checked values are as specified in the test procedure.

Notes	
--------------	--

TP Id	TP/PLT/AG/OXP/DIM/BV-003_B		
TP label	PM-Segment Object: Semantic of Segment Statistic attribute		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	PM-SegmAttr 11; O	
Applicability	C_AG_OXP_041 AND C_AG_OXP_074 AND C_AG_OXP_000		
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager sends a "roiv-cmip-confirmed-action", action-type MDC_ACT_SEG_GET_INFO, with SegmSelection (all-segments). 2. The agent responds with a "rors-cmip-confirmed-action", action-type MDC_ACT_SEG_GET_INFO, giving information about the attributes of every PM-Segment. For every segment, the Segment-Statistic attribute is recorded if it is supported by the agent. 3. The simulated manager sends a "roiv-cmip-confirmed-action", action-type MDC_ACT_SEG_TRIG_XFER for a PM-Segment that supports the Segment-Statistic attribute. 4. The agent sends a "rors-cmip-confirmed-action", action-type MDC_ACT_SEG_TRIG_XFER, with TrigSegmDataXferRsp "tsxr-successful". 5. The agent sends a "roiv-cmip-confirmed-event-report", action-type MDC_NOTI_SEGMENT_DATA. 6. Record the segment-data-event-entries in step 5, calculate the min, max or mean for every entry and compare it with the Segment-Statistic attribute value. 		
Pass/Fail criteria	<ul style="list-style-type: none"> • The maximum of every entry recorded in step 5 is not higher than the max defined in the SegmentStatistic attribute recorded in step 2 if SegStatType is "segm-stat-type-maximum". • The minimum of every entry recorded in step 5 is not lower than the min defined in the SegmentStatistic attribute recorded in step 2 if SegStatType is "segm-stat-type-minimum". • The mean of every entry recorded in step 5 matches the mean defined in the SegmentStatistic attribute recorded in step 2 if SegStatType is "segm-stat-type-average". 		
Notes			

TP Id	TP/PLT/AG/OXP/DIM/BV-003_C		
TP label	PM-Segment Object: Semantic of PM-Seg-Person-Id attribute		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	PM-SegmAttr 3; O	
	Spec	[ITU-T H.810]	
	Testable items	General 6; C	
Applicability	C_AG_OXP_041 AND C_AG_OXP_035 AND C_AG_OXP_000		
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager shall send a Get request for the PM-Store object with an attribute-id-list set to 0 to indicate all PM-Store attributes. 2. The agent issues a GET response with the PM-Store attributes. 3. The simulated manager shall send a Get-Segment-Info object action for the PM-Store object with SegmSelection set to all-segments to indicate all PM-Segments attributes. 4. The agent issues a response with the PM-Segment attributes it supports. 		

	<ol style="list-style-type: none"> 5. The simulated manager sends a request for the PM-Segment Data. 6. The agent issues an action response (action: MDC_ACT_SEG_TRIG_XFER, action-info-args: TrigSegmDataXferRsp). 7. The agent under test sends a Segment-Data-Event message.
Pass/Fail criteria	<p>The pmsc-multi-person bit in the PM-Store-Capab attribute must be set and all PM-Segment instances contained in the PM-Store must contain the PM-Seg-Person-Id attribute.</p> <p>In step 7, measurements stored in the PM-Store have to be assigned correctly to every person.</p>
Notes	

TP Id	TP/PLT/AG/OSP/DIM/BV-003_D		
TP label	PM-Segment Object: semantic of data-and-Time Adjustment attribute		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	PM-SegmAttr 9; O	AbsTime 15; C
Applicability	C_AG_OXP_041 AND C_AG_OXP_012 AND C_AG_OXP_000 AND C_AG_OXP_016		
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<ol style="list-style-type: none"> 1. Take a measurement with agent. 2. Make a noticeable change in change in the Date or Time of the agent. 3. Take a new measurement. 4. The simulated manager shall send a Get-Segment-Info object action for the PM-Store object with an attribute-id-list set to 0 to indicate all PM-Segments attributes. 5. The agent issues a response with the PM-Segment attributes it supports, the attribute of interest: <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_ABS_ADJUST <input type="checkbox"/> attribute-type = absolute-time-adjust <input type="checkbox"/> attribute-length = 6 bytes <input type="checkbox"/> attribute-value = <Must contain the adjustment (+-44505 years)> 		
Pass/Fail criteria	The PM-Segment attribute Date-and-Time-Adjustment must inform of the change.		
Notes			

TP Id	TP/PLT/AG/OSP/DIM/BV-003_E		
TP label	PM-Segment Object: semantic of data-and-Time Adjustment attribute. Disconnected		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	PM-SegmAttr 9; O	AbsTime 15; C
Applicability	C_AG_OXP_041 AND C_AG_OXP_012 AND C_AG_OXP_000		
Initial condition	The simulated manager and the agent under test are in the unassociated state.		
Test procedure	<ol style="list-style-type: none"> 1. Take measurements with agent that are stored in a segment. 2. Make a noticeable change in change in the Date or Time of the agent. 3. Take a new measurement. 4. The simulated manager receives an association request from the agent under test. 5. The simulated manager responds with a result = accepted-unknown-config. 6. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 7. Once in the operating state, the simulated manager shall send a Get-Segment-Info 		

	<p>object action for the PM-Store object with an attribute-id-list set to 0 to indicate all PM-Segment attributes.</p> <p>8. The agent issues a response with the PM-Segment attributes it supports, the attribute of interest:</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_ABS_ADJUST <input type="checkbox"/> attribute-type = absolute-time-adjust <input type="checkbox"/> attribute-length = 6 bytes <input type="checkbox"/> attribute-value = <Must contain the adjustment (+-44505 years)>
Pass/Fail criteria	The PM-Segment attribute Date-and-Time-Adjustment must inform of the change.
Notes	

TP Id	TP/PLT/AG/OSP/DIM/BV-005		
TP label	PM-Segment Object. Confirm Timeout		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	PM-SegmAttr 13; O	OperErrorCond 5; M
		TimeOutVar 2; C	
Applicability	C_AG_OXP_041 AND C_AG_OXP_000		
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager shall send a Get-Segment-Info object action for the PM-Segment object with SegmSelection set to all-segments to indicate all PM-Segments attributes. 2. Record the Confirm-Timeout value from the the Get PM-Segment operation. If the attribute is not present its value shall be 3 s($TO_{cer-pms}$). 3. The simulated manager sends a request for the PM-Segment Data with SegmSelection = Segment-id-list. 4. The agent issues a response with the PM-Segments attributes (action: MDC_ACT_SEG_GET_INFO). 5. The simulated manager sends a request for the PM-Segment Data. 6. The agent issues an action response (action: MDC_ACT_SEG_TRIG_XFER, action-info-args: TrigSegmDataXferRsp). 7. The agent under test sends a Segment-Data-Event message. 8. The simulated manager does not respond for at least the time specified in the field Confirm-Timeout. 9. The agent waits the Confirm-Timeout time and then must send an abort message to the manager and move to the unassociated state. 		
Pass/Fail criteria	The agent waits $TO_{cer-pms}$ time and then must send an abort message to the manager and changes to the unassociated state.		
Notes			

TP Id	TP/PLT/AG/OSP/DIM/BV-008			
TP label	EpiCfgScanner Object: Mandatory, Conditional and Optional Attributes			
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	ScanClassAttr 1; M	ScanClassAttr 2; M	ScanClassAttr 4; C
		ScanClassAttr 6; C	CfgScanAttr 1; M	CfgScanAttr 2; C
		CfgScanAttr 5; O	EpiCfgScanAttr 1; O	EpiCfgScanClass 3; M
		ConfNormalProc 1; M	ConfEventRep 29; M	ConfEventRep 30; M
		ConfEventRep 31; C	ConfEventRep 33; O	
Spec	[IEEE 11073-10406]			

	Testable items	EpiScanObjAttr1; M	EpiScanObjAttr2; M	EpiScanObjAttr3; C
		EpiScanObjAttr4; C	EpiScanObjAttr5; M	EpiScanObjAttr6; O
		EpiScanObjAttr7; O		
	Spec	[ITU-T H.810]		
Testable items	Communication 6; M			
Applicability	C_AG_OXP_047 AND C_AG_OXP_000			
Initial condition	The simulated manager and the agent under test have been associated, but the agent configuration is unknown for simulated manager, so the agent and the simulated manager will be in the configuring state.			
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 4. The Configurable Episodic Scanner object (ConfigReport -> ConfigObject-> AttributeList) must have: <ol style="list-style-type: none"> a. Mandatory attribute Handle shall not be present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-type = HANDLE <input type="checkbox"/> attribute-value = 2 bytes <input type="checkbox"/> attribute-value = <Must be unique but not relevant in this test> b. Mandatory attribute Operational-State should be present un ConfigReport: <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_OP_STAT <input type="checkbox"/> attribute-type = OperationalState <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = 0 at start c. IF attribute Scan-Handle-List is supported, it should be present in ConfigReport: <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SCAN_HANDLE_LIST <input type="checkbox"/> attribute-type = HANDLEList <input type="checkbox"/> attribute-value.length = <input type="checkbox"/> attribute-value = <Not Relevant for this test> d. IF attribute Scan-Handle-Attr-Val-Map is supported, it should be present in ConfigReport: <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SCAN_HANDLE_ATTR_VAL_MAP <input type="checkbox"/> attribute-type = HANDLEAttrValMap <input type="checkbox"/> attribute-value.count = N <input type="checkbox"/> attribute-value.length = <Variable> <input type="checkbox"/> attribute-value = <Not Relevant for this test> e. Mandatory attribute Confirm-Mode should be present in ConfigReport: <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_CONFIRM_MODE <input type="checkbox"/> attribute-type = ConfirmMode <input type="checkbox"/> attribute-value = One of: <ul style="list-style-type: none"> ▪ unconfirmed (0x00 0x00) ▪ confirmed (0x00 0x01) f. Optional Confirm-Timeout should be present in ConfigReport: <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_CONFIRM_TIMEOUT <input type="checkbox"/> attribute-type = RelativeTime 			

	<ul style="list-style-type: none"> <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> <p>g. IF attribute Transmit-Window is supported, it should be present in ConfigReport:</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TX_WIND <input type="checkbox"/> attribute-type = INT-U16 <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = 1 <p>h. Optional attribute Min-Reporting-Interval should be present in ConfigReport:</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SCAN_REP_PD_MIN <input type="checkbox"/> attribute-type = RelativeTime <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <defined by vendor> IF agent supports Min-Reporting-Interval attribute (C_AG_OXP_144) THEN at least there is a scanner object that supports this attribute, ELSE, no scanner objects support this attribute. <p>5. Furthermore check if Variable MDS Scan Event Reports are sent by the agent for the Scanner object:</p> <p>a. Wait for a Scan Event Report from the agent.</p> <ul style="list-style-type: none"> • Attributes whose values may be reported will be the attributes defined as dynamic: Operational-State, Scan-Handle-List, Scan-Handle-Attr-Val-Map, Confirm-Mode, Confirm-Timeout, Transmit-Window and Min-Reporting-Interval. <p>b. Set to enable the Operational-State for the Episodic Scanner object to make the Scanner object send event reports:</p> <ul style="list-style-type: none"> • If the agent sends Unbuf-Scan-Report-Fixed or Variable, Scan-Handle-List attributes shall be received previously. • If the agent sends Unbuf-Scan-Report-Grouped, Scan-Handle-Attr-Val-Map attributes shall be received previously. <p>c. Set to disable the Operation-State for the Episodic Scanner object and repeat step 5b for every episodic scanner object.</p>
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/AG/OXP/DIM/BV-009			
TP label	PeriCfgScanner Object: Mandatory, Conditional and Optional Attributes			
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	ScanClassAttr 1; M	ScanClassAttr 2; M	ScanClassAttr 4; C
		ScanClassAttr 6; C	CfgScanAttr 1; M	CfgScanAttr 2; C
		CfgScanAttr 5; O	PeriCfgScanClass 2; M	PeriCfgScanAttr 1; M
		ConfNormalProc 1; M	ConfEventRep 29; M	ConfEventRep 30; M
		ConfEventRep 31; C	ConfEventRep 33; O	
	Spec	[IEEE 11073-10406]		
	Testable items	PerScanObjAttr1; M	PerScanObjAttr2; M	PerScanObjAttr3; C
		PerScanObjAttr4;	PerScanObjAttr5; M	PerScanObjAttr6; O
		PerScanObjAttr7; O	PerScanObjAttr8; M	
Spec	[ITU-T H.810]			
Testable items	Communication 6; M			

Applicability	C_AG_OXP_046 AND C_AG_OXP_000
Initial condition	<ul style="list-style-type: none"> The simulated manager and the agent under test have been associated, but the agent configuration is unknown for the simulated manager, so the agent and the simulated manager will be in the configuring state.
Test procedure	<ol style="list-style-type: none"> The simulated manager receives an association request from the agent under test. The simulated manager responds with a result = accepted-unknown-config. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. The Configurable Periodic Scanner object (ConfigReport -> ConfigObject->AttributeList) must have: <ol style="list-style-type: none"> Mandatory attribute Handle shall not be present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-type = HANDLE <input type="checkbox"/> attribute-value = 2 bytes <input type="checkbox"/> attribute-value = must be unique<Not relevant in this test> Mandatory attribute Operational-State should be present in ConfigReport: <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_OP_STAT <input type="checkbox"/> attribute-type = OperationalState <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = 0 at start IF Attribute Scan-Handle-List is supported, it should be present in ConfigReport: <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SCAN_HANDLE_LIST <input type="checkbox"/> attribute-type = HANDLEList <input type="checkbox"/> attribute-value.length = <input type="checkbox"/> attribute-value = <Not relevant for this test> IF attribute Scan-Handle-Attr-Val-Map is supported, it should be present in ConfigReport: <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SCAN_HANDLE_ATTR_VAL_MAP <input type="checkbox"/> attribute-type = HANDLEAttrValMap <input type="checkbox"/> attribute-value.count = N <input type="checkbox"/> attribute-value.length = <Variable> <input type="checkbox"/> attribute-value = N metric-derived object must be specified here, verify the correct format of the object and that the handle points to the object. Mandatory attribute Confirm-Mode should be present in ConfigReport: <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_CONFIRM_MODE <input type="checkbox"/> attribute-type = ConfirmMode <input type="checkbox"/> attribute-value = One of: <ul style="list-style-type: none"> unconfirmed (0x00 0x00) confirmed (0x00 0x01) Optional Confirm-Timeout should be present in ConfigReport: <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_CONFIRM_TIMEOUT <input type="checkbox"/> attribute-type = RelativeTime <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test>. IF attribute Transmit-Window is supported, it should be present in ConfigReport: <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TX_WIND <input type="checkbox"/> attribute-type = INT-U16

	<ul style="list-style-type: none"> <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = 1 <p>h. Mandatory attribute Reporting-Interval should be present in ConfigReport:</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SCAN_REP_PD <input type="checkbox"/> attribute-type = RelativeTime <input type="checkbox"/> attribute-length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> <p>5. Furthermore check if the Variable MDS Scan Event Reports are sent by the agent for the Scanner object:</p> <p>a. Wait for a Scan Event Report from the agent.</p> <ul style="list-style-type: none"> • Attributes whose values may be reported will be the attributes defined as dynamic: Operational-State, Scan-Handle-List, Scan-Handle-Attr-Val-Map, Confirm-Mode, Confirm-Timeout, Transmit-Window and Reporting-Interval. <p>b. Set to enable Operational-State for the Periodic Scanner object to make the Scanner object send event reports:</p> <ul style="list-style-type: none"> • If the agent sends a Buf-Scan-Report-Fixed or Variable, Scan-Handle-List attributes shall be received previously. • If the agent sends a Buf-Scan-Report-Grouped, Scan-Handle-Attr-Val-Map attributes shall be received previously. <p>c. Set to disable Operation-State for Periodic Scanner object and repeat step 5b for every periodic scanner object.</p>
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/AG/OXP/DIM/BV-010		
TP label	MDS objects methods and events. Agent data transmission		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	MDSEvent 2; C	MDSEvent 3; C
		MDSEvent 5; C	CommonCharac 3; M
	Spec	[ITU-T H.810]	
Testable items	General 7; C		
Applicability	C_AG_OXP_000 AND (C_AG_OXP_182 OR C_AG_OXP_183 OR C_AG_OXP_184 OR C_AG_OXP_189)		
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<ol style="list-style-type: none"> 1. Take some measurements with the agent under test. 2. Wait until the agent sends event reports with the data: <ol style="list-style-type: none"> a. APDU Type = 0xE7 0x00 b. Invoke-Id <ul style="list-style-type: none"> <input type="checkbox"/> Length = 2 bytes <input type="checkbox"/> Value = <Not relevant for this Test Case> c. CHOICE <ul style="list-style-type: none"> <input type="checkbox"/> Length = 2 bytes <input type="checkbox"/> Value = 0x01 0x00 (Unconfirmed) OR 0x01 0x01 (Confirmed) d. Obj-Handle <ul style="list-style-type: none"> <input type="checkbox"/> Length = 2 bytes <input type="checkbox"/> Value = 0 (MDS object) 		

	<p>e. Event-Time</p> <ul style="list-style-type: none"> <input type="checkbox"/> Length = 4 bytes <input type="checkbox"/> Value = <0xFF 0xFF 0xFF 0xFF> If NOT C_AG_OXP_010 <p>f. IF the data is from one person and uses a variable format event reporting, it must be:</p> <ul style="list-style-type: none"> <input type="checkbox"/> Event-type = MDC_NOTI_SCAN_REPORT_VAR <input type="checkbox"/> Event-info parameter = ScanReportInfoVar <p>g. IF the data is from one person and uses a fixed format event reporting, it must be:</p> <ul style="list-style-type: none"> <input type="checkbox"/> Event-type = MDC_NOTI_SCAN_REPORT_FIXED <input type="checkbox"/> Event-info parameter = ScanReportInfoFixed <p>h. IF the data is from multiple persons and uses a variable format event reporting, it must be:</p> <ul style="list-style-type: none"> <input type="checkbox"/> Event-type = MDC_NOTI_SCAN_REPORT_MP_VAR <input type="checkbox"/> Event-info parameter = ScanReportInfoMPVar <p>i. IF the data is from multiple persons and uses a fixed format event reporting, it must be:</p> <ul style="list-style-type: none"> <input type="checkbox"/> Event-type = MDC_NOTI_SCAN_REPORT_MP_FIXED <input type="checkbox"/> Event-info parameter = ScanReportInfoMPFixed
Pass/Fail criteria	<ul style="list-style-type: none"> • All checked values are as specified in the test procedure. • The total size of the event report can not exceed the maximum APDU size established by the specialization. • If the agent does not support confirmed event reports (C_AG_OXP_053= FALSE), the agent can not send confirmed event reports. • If the agent does not use variable event report (C_AG_OXP_189= FALSE), the agent can not send variable event reports. • If the agent supports multi-person event reports for one or more metric object (ScanReportPer*) (C_AG_OXP_031= TRUE), the agent has to send multi-person event reports. • If C_AG_OXP_031= TRUE and MP event reports have been received, a pop-up will show the received measurements to make the operator identify if measurements have been correctly assigned to every person.
Notes	

TP Id	TP/PLT/AG/OXP/DIM/BV-011		
TP label	MDS objects methods. Agent real-time clock (RTC). Absolute-Time		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	MDSMethod 4; M	AbsTime 1;C
		MDSMethod 6; M	BaseTimOffset2 ; M
Applicability	(C_AG_OXP_007 OR C_AG_OXP_008) AND C_AG_OXP_000		
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager sends a Get request for the MDS object with an attribute-id-list set to 0 read all the attributes. 2. Check the Value of the MdsTimeCapab bits in the MDS-Time-Info-Attribute <ul style="list-style-type: none"> <input type="checkbox"/> IF C_AG_OXP_009 = TRUE THEN check that mds-time-capab-real-time-clock(0) is set to TRUE indicating support of an RTC, ELSE this bit is set to FALSE. <input type="checkbox"/> IF C_AG_OXP_008 = TRUE THEN check that mds-time-capab-set-clock(1) is set to TRUE indicating support of the Set Time Action ELSE this bit is set to FALSE. <input type="checkbox"/> IF C_AG_OXP_014 = TRUE THEN check that mds-time-capab-bo-time(7) is set to TRUE indicating support of Base-Offset-Time and record the value of Base-Offset- 		

	<p>Time attribute (MDC_ATTR_TIME_BO) ELSE this bit is set to FALSE indicating agent does not support of a Base-Offset-Time.</p> <ol style="list-style-type: none"> 3. IF Set Time Action is supported and C_AG_OXP_009 = TRUE: <ol style="list-style-type: none"> a. The simulated manager sends a SET action: <ul style="list-style-type: none"> <input type="checkbox"/> CHOICE = SetTimeInvoke <input type="checkbox"/> action-type = MDC_ACT_SET_TIME <input type="checkbox"/> the action-info-args are SetTimeInvoke <ul style="list-style-type: none"> • date-time = century ≤ 99, year ≤ 99, month ≤ 12, day ≤ 31, hour ≤ 24, minute ≤ 60, second ≤ 60, sec-fractions ≤ 100 • accuracy = 0 b. The agent under test response must be a rors-cmip-confirmed-action but the action-info-args shall be empty in this response. c. The simulated manager sends a Get request for the MDS object with an attribute-id-list set to 0 read all the attributes. The Date-and-Time attribute value matches with the Absolute Time set in step 3.a. 4. IF Set Time Action is supported and C_AG_OXP_014 = TRUE: <ol style="list-style-type: none"> a. The simulated manager sends a SET action: <ul style="list-style-type: none"> <input type="checkbox"/> CHOICE = SetBOTimeInvoke <input type="checkbox"/> action-type = MDC_ACT_SET_BO_TIME <input type="checkbox"/> the action-info-args are SetBOTimeInvoke <ul style="list-style-type: none"> • date-time = bo-seconds = 0x00 0x00 0x00 0x00, bo-fractions = 0x00 0x00, bo-time-offset = <original bo-time-offest> + 60 b. The agent under test response must be a rors-cmip-confirmed-action but the action-info-args shall be empty in this response. c. The simulated manager sends a Get request for the MDS object with an attribute-id-list set to 0 read all the attributes. The Base-Offset-Time attribute value matches with Base-Offset-Time set in step 4.a.
Pass/Fail criteria	<ul style="list-style-type: none"> • All checked values are as specified in the test procedure and check that the time has been set correctly in step 3.c or 4.c if that action was possible. • If Set Time Action and Base-Offset-Time is supported, then the value of bo-seconds and bo-fractions in step 2 and step 4.c shall be the same and only bo-time-offset is set.
Notes	

TP Id	TP/PLT/AG/OXP/DIM/BV-012		
TP label	MDS object events. Agent configuration event		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	MDSEvent 1; M	MetricClassAttr 1; M
		ScanClassAttr 1; M	StoreClassAttr 1; M
Applicability	C_AG_OXP_000		
Initial condition	The simulated manager and the agent under test are in the unassociated state.		
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent responds with a "Remote Operation Invoke Confirmed Event Report" (roiv-cmip-confirmed-event-report) message: <ol style="list-style-type: none"> a. APDU Type <ul style="list-style-type: none"> <input type="checkbox"/> field-length = 2 bytes <input type="checkbox"/> field-value = 0xE7 0x00 <input type="checkbox"/> This value is for association request "prst" (PrstApdu). 		

- b. invoke-id
 - field- type = InvokeIDType
 - field-length = 2 bytes
 - field- value = This value identifies the message; the confirmed response that will be sent by the simulated manager shall have the same invoke-id.
- c. obj-handle (EventReportArgumentSimple)
 - field- type = HANDLE
 - field-length = 2 bytes
 - field- value = 0x00 0x00
 - This obj-handle represents MDS-Object.
- d. event-time (EventReportArgumentSimple)
 - field- type = Relative Time
 - field-length = 4 bytes
 - If the agent does not support relative time clock:
 - field-value =
 - IF NOT C_AG_OXP_010 THEN = 0xFF 0xFF 0xFF 0xFF
- e. event-type (EventReportArgumentSimple)
 - field- type = OID-Type
 - field-length =2 bytes
 - field- value = 0x0D 0x1C (MDC_NOTI_CONFIG)
- f. config-report-id (ConfigReport)
 - field- type = ConfigId
 - field-length = 2 bytes
 - field- value = <between 0x00 0x01 and 0x7F 0xFF>
- g. obj-class (ConfigReport → ConfigObjectList (ConfigObject))
 - field- type = OID-Type
 - field-length = 2 bytes
 - field- value = 0x00 0x06 (MDC_MOC_VMO_METRIC_NU) or 0x00 0x09 (MDC_MOC_VMO_METRIC_SA-RT) or 0x00 0x05 (MDC_MOC_VMO_METRIC_ENUM) or 0x00 0x12 (MDC_MOC_SCAN_CFG_EPI) 0x00 0x13 (MDC_MOC_SCAN_CFG_PERI) or 0x00 0x3D (MDC_MOC_VMO_PMSTORE)) or a value between 0xF000 and 0xFBFF
 - IF the agent supports at least one numeric object (C_AG_OXP_040=TRUE) then MDC_MOC_VMO_METRIC_NU shall be present, ELSE no numeric object is present.
 - IF the agent supports at least one PM-Store object (C_AG_OXP_041=TRUE) then MDC_MOC_VMO_PMSTORE shall be present, ELSE no PM-Store object is present.
 - IF the agent supports at least one RT-SA object (C_AG_OXP_042=TRUE) then MDC_MOC_VMO_METRIC_SA-RT shall be present, ELSE no RT-SA object is present.
 - IF the agent supports at least one enumerated object (C_AG_OXP_043=TRUE) then MDC_MOC_VMO_METRIC_ENUM shall be present, ELSE no enumerated object is present.
 - IF the agent supports at least one periodic scanner object (C_AG_OXP_046=TRUE) then MDC_MOC_SCAN_CFG_PERI shall be present, ELSE no periodic scanner object is present.
 - IF the agent supports at least one episodic scanner object (C_AG_OXP_047=TRUE) then MDC_MOC_SCAN_CFG_EPI shall be

	<p>present, ELSE no episodic scanner object is present.</p> <p>h. obj-handle (ConfigReport → ConfigObjectList (ConfigObject))</p> <ul style="list-style-type: none"> <input type="checkbox"/> field- type = HANDLE <input type="checkbox"/> field-length = 2 bytes <input type="checkbox"/> field-value = <Check that each object have a unique identifier and non -zero> <p>i. attribute-id (ConfigReport → ConfigObjectList (ConfigObject)→Attribute List)</p> <ul style="list-style-type: none"> <input type="checkbox"/> field- type = OID-Type <input type="checkbox"/> field-length = 2 bytes <input type="checkbox"/> field-value = <between 0x0913 (2323) and 0x0A77 (2679)> or <between 0xF000(61440) and 0xFBFF(64511)>
Pass/Fail criteria	<ul style="list-style-type: none"> • All checked values are as specified in the test procedure. • The total size of the response can not exceed the sum of the APDU sizes of the supported specializations (limited to an absolute limit of 64512 octets): <ul style="list-style-type: none"> ○ Pulse oximeter -> 9216 octets ○ Weighing scales -> 896 octets ○ Glucose meter -> 5210 octets or 64512 octets if the agent supports PM-Store ○ Blood pressure -> 896 octets ○ Thermometer -> 896 octets ○ Independent activity hub -> 5120 octets ○ Cardiovascular -> 64512 octets or 6624 octets if the agent supports Step Counter Profile ○ Strength -> 64512 octets ○ Adherence monitor -> 1024 octets ○ Peak flow -> 2030 octets ○ Body composition analyser -> 7730 octets ○ Basic ECG/Simple ECG -> 7168 octets or 64512 octets if the agent supports PM-Store ○ Basic ECG/Heart rate -> 1280 octets or 64512 octets if the agent supports PM-Store ○ International Normalized Ratio -> 896 octets or 64512 if the agent supports PM-Store
Notes	

TP Id	TP/PLT/AG/OXP/DIM/BV-013			
TP label	PM-Store object methods. Clear-Segments method			
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	PM-StoreMeth 1; O	PM-StoreMeth 2; C	
		PM-StoreMeth 8; O	PersStoreMtrDatTransf 20; M	PersStoreMtrDatTransf 21; M
		PersStoreMtrDatTransf 22; O	PersStoreMtrDatTransf 23; M	PM-StoreMeth 6; M
		PM-StoreMeth 20; C	PM-StoreMeth 21; C	PM-StoreMeth 7; M
		PM-StoreMeth 30; O	PM-StoreMeth 31; M	
Spec	[ITU-T H.810]			
Testable	Communication 1; M			

	items			
Applicability	C_AG_OXP_041 AND C_AG_OXP_071 AND C_AG_OXP_000			
Initial condition	The simulated manager and the agent under test are in the operating state and the agent has at least one PM-Segment with data stored.			
Test procedure	<ol style="list-style-type: none"> 1. Make sure the agent under test is not taking measurements which are stored in PM-Segments. 2. The simulated manager shall send a Get request for the PM-Store object with an attribute-id-list set to 0 to indicate all PM-Store attributes. 3. The agent under test issues a GET response with the PM-Store attributes. Record the values of the PM-Store-Capab attribute. <ol style="list-style-type: none"> a. PM-Store-Capab: <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_PM_STORE_CAPAB <input type="checkbox"/> attribute-type = PmStoreCapab <input type="checkbox"/> attribute-value = Record the value of bit 10 (Indicates that PM-Segments in the SegmSelection data type can be cleared by segment selection –all segments). This bit shall be set to 1. 4. The simulated manager shall send a Get-Segment-Info object action for the PM-Store object with SegmSelection set to all-segments. 5. The agent issues a response (rors-cmip-confirmed-action) with the PM-Segment attributes it supports. 6. The simulated manager sends a Clear-Segment: <ol style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action, <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_CLEAR <input type="checkbox"/> SegmSelection = all-segments 7. If the agent does not protect all segments, the agent under test operation response will be: <ol style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Response Confirmed Action <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_CLEAR <input type="checkbox"/> Check the invoke-id of the response is mirrored from the request. 8. If the agent does protect all segments, the agent under test operation response will be: <ol style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Roer <input type="checkbox"/> ErrorResult = no-allowed-by-object (24) and return code shall be MDC_RET_CODE_UNKNOWN. <input type="checkbox"/> Check the invoke-id of the response is mirrored from the request 9. Delay 10. If the agent has sent the confirmation in step 7, the simulated manager sends a request for the PM-Segment Data to obtain all the segments: <ol style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_TRIG_XFER <input type="checkbox"/> SegmSelection = <Instance number of the selected PM-Segment that contained data before the clear-segment action> 11. The agent under test issues an action response with the Data: 			

	<ul style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Response Confirmed Action <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_TRIG_XFER <input type="checkbox"/> TrigSegmXferRsp = <ul style="list-style-type: none"> • IF pmsc-clear-segm-remove of the PM-Store-Capab attribute is NOT set then <ul style="list-style-type: none"> <input type="checkbox"/> TrigSegmXferRsp = tsxr-fail-segm-empty • ELSE then <ul style="list-style-type: none"> <input type="checkbox"/> TrigSegmXferRsp = tsxr-fail-no-such-segment
Pass/Fail criteria	<ul style="list-style-type: none"> • In step 7, the agent must send a confirmation if the agent does not protect any segments, otherwise the agent shall send a roer message (step 8). • If the agent sends the confirmation in step 7, the agent shall send the response specified in step 11 at least for a segment. • After APDU received by the simulated manager in step 11, the agent does not send any message of type "Segment-data-event" with data stored.
Notes	

TP Id	TP/PLT/AG/OXP/DIM/BV-013_A			
TP label	PM-Store object methods. Clear-Segments List method			
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	PM-StoreMeth 1; O	PM-StoreMeth 2; C	PM-StoreMeth 6; M
		PM-StoreMeth 8; O	PersStoreMtrDatTransf 20; M	PersStoreMtrDatTransf 21; M
		PersStoreMtrDatTransf 22; O	PersStoreMtrDatTransf 23; M	PM-StoreMeth 22; C
PersStoreMtrDatTransf 24; M				
Applicability	C_AG_OXP_041 AND C_AG_OXP_071 AND C_AG_OXP_000			
Initial condition	The simulated manager and the agent under test are in the operating state and the agent has at least two PM-Segments with data stored.			
Test procedure	<ol style="list-style-type: none"> 1. Make sure the agent is not taking measures which are stored in PM-Segments. 2. The simulated manager shall send a Get request for the PM-Store object with an attribute-id-list set to 0 to indicate all PM-Store attributes. 3. The agent issues a GET response with the PM-Store attributes, check the values of the PM-Store-Capab attribute: <ol style="list-style-type: none"> a. PM-Store-Capab: <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_PM_STORE_CAPAB <input type="checkbox"/> attribute-type = PmStoreCapab <input type="checkbox"/> attribute-value = Record the value of bit 7 (Indicates that PM-Segments in the SegmSelection data type can be cleared by defining a list) 4. The simulated manager shall send a Get-Segment-Info object action for the PM-Store object with SegmSelection set to all-segments. 5. The agent issues a response (rors-cmip-confirmed-action) with the PM-Segment attributes it supports. IF bit 7 of PmStoreCapab was set: 6. The simulated manager sends a Clear-Segment: <ol style="list-style-type: none"> a. Data APDU 			

	<ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action, <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_CLEAR <input type="checkbox"/> SegmSelection = segm-id-list (list of integers containing 2 of the instance numbers obtained in step 5) <p>7. The agent under test operation response:</p> <p>a. Data APDU</p> <ul style="list-style-type: none"> <input type="checkbox"/> Type = Response Confirmed Action, <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_CLEAR <p>8. Delay.</p> <p>9. The simulated manager sends a request for the PM-Segment Data of one of the cleared PM-Segments:</p> <p>a. Data APDU</p> <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action, <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_TRIG_XFER <input type="checkbox"/> SegmSelection = <Instance number of the selected PM-Segment that contained data before the clear-segment action> <p>10. The agent issues an action response with the Data</p> <p>a. Data APDU</p> <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action, <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_TRIG_XFER <input type="checkbox"/> TrigSegmXferRsp = <ul style="list-style-type: none"> ▪ IF pmsc-clear-segm-remove is NOT set then <ul style="list-style-type: none"> o TrigSegmXferRsp = tsxr-fail-segm-empty ▪ ELSE then <ul style="list-style-type: none"> o TrigSegmXferRsp = tsxr-fail-no-such-segm <p>IF bit 7 of PMStoreCapab was NOT set</p> <p>11. The simulated manager sends a Clear-Segment:</p> <p>a. Data APDU</p> <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action, <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_CLEAR <input type="checkbox"/> SegmSelection = segm-id-list (list of integers containing 2 of the instance numbers obtained in step 5) <p>12. The agent under test operation response:</p> <p>a. Data APDU</p> <ul style="list-style-type: none"> <input type="checkbox"/> Type = Roer <input type="checkbox"/> ErrorResult = not-allowed-by-object (24)
Pass/Fail criteria	<ul style="list-style-type: none"> • In step 7, the agent must send a confirmation • The last APDU received by the simulated manager has no data
Notes	

TP Id	TP/PLT/AG/OSP/DIM/BV-013_B			
TP label	PM-Store object methods. Clear-Segments Time Range method 1			
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	PM-StoreMeth 1; O	PM-StoreMeth 2; C	PM-StoreMeth 6; M
		PM-StoreMeth 8; O	PersStoreMtrDatTransf 20; M	PersStoreMtrDatTransf 21; M
	PersStoreMtrDatTransf 22; O	PersStoreMtrDatTransf 23; M	PM-StoreMeth 19; M	
Applicability	C_AG_OXP_041 AND C_AG_OXP_071 AND C_AG_OXP_072 AND C_AG_OXP_009 AND C_AG_OXP_000			
Initial condition	The simulated manager and the agent under test are in the operating state and the agent supports at least one PM-Segment with data stored.			
Test procedure	<ol style="list-style-type: none"> 1. Make sure the agent under test is not taking measurements which are stored in PM-Segments. 2. The simulated manager shall send a Get request for the PM-Store object with an attribute-id-list set to 0 to indicate all PM-Store attributes. 3. The agent under test issues a GET response with the PM-Store attributes, record the values of the PM-Store-Capab attribute: <ol style="list-style-type: none"> a. PM-Store-Capab: <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_PM_STORE_CAPAB <input type="checkbox"/> attribute-type = PmStoreCapab <input type="checkbox"/> attribute-value = Record the value of bit 8 (Indicates that PM-Segments in the SegmSelection data type can be cleared by defining an AbsTimeRange) 4. The simulated manager shall send a Get-Segment-Info object action for the PM-Store object with SegmSelection set to all-segments. 5. The agent issues a response (rors-cmip-confirmed-action) with the PM-Segment attributes it supports, record the attributes "Segment-Start-Abs-Time" and "Segment-End-Abs-Time" of every PM-Segment. IF bit 8 of PMStoreCapab was set: 6. The simulated manager sends a Clear-Segment: <ol style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action, <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_CLEAR <input type="checkbox"/> SegmSelection = abs-time-range, selecting a range with its boundaries set to an earlier date of any of the existing segments. 7. The agent under test operation response: <ol style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Roer <input type="checkbox"/> ErrorResult = no-such-action (9) 			
Pass/Fail criteria	In step 7 the agent must send the specified error.			
Notes	<p>Error code was not clearly defined in the spec.</p> <p>In the new edition of [ISO/IEEE 11073-20601A], the clear-segment using time range has been clarified. "For PM-segments cleared using the by time method, only PM-segments having Segment-Start-Abs-Time and Segment-End-Abs-Time fields entirely within the specified time period are cleared.", but if the Manager sends a Clear-Segment but the segment has not a Segment-Start-Abs-Time and Segment-End-Abs-Time within the specified time-period, the agent will send a Roer message.</p> <p>At this point, it would be up to the agent what error code (Roer message) to send (No-such-action, not-allowed-by-object, or both). If one wants to clear the segment due to all the</p>			

	internal timestamps that were saved in the segment as falling inside the given time period, then the agent could do that. Returning an error is also possible.
--	--

TP Id		TP/PLT/AG/OXP/DIM/BV-013_C		
TP label		PM-Store object methods. Clear-Segments Time Range method 2		
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	PM-StoreMeth 1; O	PM-StoreMeth 2; C	PM-StoreMeth 6; M
		PM-StoreMeth 8; O	PersStoreMtrDatTransf 20; M	PersStoreMtrDatTransf 21; M
		PersStoreMtrDatTransf 22; O	PersStoreMtrDatTransf 23; M	PM-StoreMeth 19; M
Applicability		C_AG_OXP_041 AND C_AG_OXP_071 AND C_AG_OXP_072 AND C_AG_OXP_009 AND C_AG_OXP_000		
Initial condition		The simulated manager and the agent under test are in the operating state and the agent has at least one PM-Segment with data stored.		
Test procedure		<ol style="list-style-type: none"> 1. Make sure the agent is not taking measures which are stored in PM-Segments. 2. The simulated manager shall send a Get request for the PM-Store object with an attribute-id-list set to 0 to indicate all PM-Store attributes. 3. The agent issues a GET response with the PM-Store attributes, record the values of the PM-Store-Capab attribute: <ol style="list-style-type: none"> a. PM-Store-Capab: <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_PM_STORE_CAPAB <input type="checkbox"/> attribute-type = PmStoreCapab <input type="checkbox"/> attribute-value = Record the value of bit 8 (Indicates that PM-Segments in the SegmSelection data type can be cleared by defining an AbsTimeRange) 4. The simulated manager shall send a Get-Segment-Info object action for the PM-Store object with SegmSelection set to all-segments. 5. The agent issues a response (rors-cmip-confirmed-action) with the PM-Segment attributes it supports, record the attributes "Segment-Start-Abs-Time" and "Segment-End-Abs-Time" of every PM-Segment. IF bit 8 of PMStoreCapab was set: <ol style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action, <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_CLEAR <input type="checkbox"/> SegmSelection = abs-time-range, selecting a range with its boundaries set to a later date of any of the existing segments. 7. The agent under test operation response: <ol style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Roer <input type="checkbox"/> ErrorResult = no-such-action (9) 		
Pass/Fail criteria		In step 7 the agent must send the specified error.		
Notes		See Note for test case TP/PLT/AG/OXP/DIM/BV-013_B.		

TP Id		TP/PLT/AG/OXP/DIM/BV-013_D		
TP label		PM-Store object methods. Clear-Segments Time Range method 3		
Coverage	Spec	[ISO/IEEE 11073-20601A]		

	Testable items	PM-StoreMeth 1; O	PM-StoreMeth 2; C	PM-StoreMeth 6; M
		PM-StoreMeth 8; O	PersStoreMtrDatTransf 20; M	PersStoreMtrDatTransf 21; M
		PersStoreMtrDatTransf 22; O	PersStoreMtrDatTransf 23; M	PM-StoreMeth 19; M
Applicability	C_AG_OXP_041 AND C_AG_OXP_071 AND C_AG_OXP_072 AND C_AG_OXP_009 AND C_AG_OXP_000			
Initial condition	The simulated manager and the agent under test are in the operating state and the agent has at least one PM-Segment with data stored.			
Test procedure	<ol style="list-style-type: none"> 1. Make sure the agent is not taking measures which are stored in PM-Segments. 2. The simulated manager shall send a Get request for the PM-Store object with an attribute-id-list set to 0 to indicate all PM-Store attributes. 3. The agent issues a GET response with the PM-Store attributes, record the values of the PM-Store-Capab attribute: <ol style="list-style-type: none"> a. PM-Store-Capab: <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_PM_STORE_CAPAB <input type="checkbox"/> attribute-type = PmStoreCapab <input type="checkbox"/> attribute-value = Record the value of bit 8 (Indicates that PM-Segments in the SegmSelection data type can be cleared by defining an AbsTimeRange) 4. The simulated manager shall send a Get-Segment-Info object action for the PM-Store object with SegmSelection set to all-segments. 5. The agent issues a response (rors-cmip-confirmed-action) with the PM-Segment attributes it supports, record the attributes "Segment-Start-Abs-Time" and "Segment-End-Abs-Time" of every PM-Segment. IF bit 8 off PMStoreCapab was set: 6. The simulated manager sends a Clear-Segment: <ol style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action, <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_CLEAR <input type="checkbox"/> SegmSelection = abs-time-range, selecting a range with one of its boundaries set to an earlier date of any of the existing segments and the other set to date contained between Segment-Start-Abs-Time and Segment-End-Abs-Time of one of the PM-Segments 7. The agent under test operation response: <ol style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Roer <input type="checkbox"/> ErrorResult = no-such-action (9) 			
Pass/Fail criteria	In step 7 the agent must send the specified error.			
Notes	See Note for test case TP/PLT/AG/OXP/DIM/BV-013_B.			

TP Id		TP/PLT/AG/OXP/DIM/BV-013_E		
TP label		PM-Store object methods. Clear-Segments Time Range method 4		
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	PM-StoreMeth 1; O	PM-StoreMeth 2; C	PM-StoreMeth 6; M
		PM-StoreMeth 8; O	PersStoreMtrDatTransf 20; M	PersStoreMtrDatTransf 21; M
	PersStoreMtrDatTransf 22; O	PersStoreMtrDatTransf 23; M	PM-StoreMeth 19; M	
Applicability		C_AG_OXP_041 AND C_AG_OXP_071 AND C_AG_OXP_072 AND C_AG_OXP_009 AND C_AG_OXP_000		
Initial condition		The simulated manager and the agent under test are in the operating state and the agent has at least one PM-Segment with data stored.		
Test procedure		<ol style="list-style-type: none"> 1. Make sure the agent is not taking measures which are stored in PM-Segments. 2. The simulated manager shall send a Get request for the PM-Store object with an attribute-id-list set to 0 to indicate all PM-Store attributes. 3. The agent issues a GET response with the PM-Store attributes, record the values of the PM-Store-Capab attribute: <ol style="list-style-type: none"> a. PM-Store-Capab: <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_PM_STORE_CAPAB <input type="checkbox"/> attribute-type = PmStoreCapab <input type="checkbox"/> attribute-value = Record the value of bit 8 (Indicates that PM-Segments in the SegmSelection data type can be cleared by defining an AbsTimeRange) 4. The simulated manager shall send a Get-Segment-Info object action for the PM-Store object with SegmSelection set to all-segments. 5. The agent issues a response (rors-cmip-confirmed-action) with the PM-Segment attributes it supports, record the attributes "Segment-Start-Abs-Time" and "Segment-End-Abs-Time" of every PM-Segment. IF bit 8 of PMStoreCapab was set: 6. The simulated manager sends a Clear-Segment: <ol style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action, <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_CLEAR <input type="checkbox"/> SegmSelection = abs-time-range, selecting a range with one of its boundaries set between Segment-Start-Abs-Time and Segment-End-Abs-Time of one of the PM-Segments and the other set to a later date of any of the existing segments 7. The agent under test operation response: <ol style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Roer <input type="checkbox"/> ErrorResult = no-such-action (9) 		
Pass/Fail criteria		In step 7 the agent must send the specified error.		
Notes		See Note for test case TP/PLT/AG/OXP/DIM/BV-013_B.		

TP Id		TP/PLT/AG/OXP/DIM/BV-013_F		
TP label		PM-Store object methods. Clear-Segments Time Range method 5		
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable	PM-StoreMeth 1; O	PM-StoreMeth 2; C	PM-StoreMeth 6; M

	items	PM-StoreMeth 8; O	PersStoreMtrDatTransf 20; M	PersStoreMtrDatTransf 21; M
		PersStoreMtrDatTransf 22; O	PersStoreMtrDatTransf 23; M	PM-StoreMeth 19; M
Applicability		C_AG_OXP_041 AND C_AG_OXP_071 AND C_AG_OXP_072 AND C_AG_OXP_009 AND C_AG_OXP_000		
Initial condition		The simulated manager and the agent under test are in the operating state and the agent has at least one PM-Segment with data stored.		
Test procedure		<ol style="list-style-type: none"> 1. Make sure the agent is not taking measures which are stored in PM-Segments. 2. The simulated manager shall send a Get request for the PM-Store object with an attribute-id-list set to 0 to indicate all PM-Store attributes. 3. The agent issues a GET response with the PM-Store attributes, record the values of the PM-Store-Capab attribute: <ol style="list-style-type: none"> a. PM-Store-Capab: <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_PM_STORE_CAPAB <input type="checkbox"/> attribute-type = PmStoreCapab <input type="checkbox"/> attribute-value = Record the value of bit 8 (Indicates that PM-Segments in the SegmSelection data type can be cleared by defining an AbsTimeRange) 4. The simulated manager shall send a Get-Segment-Info object action for the PM-Store object with SegmSelection set to all-segments. 5. The agent issues a response (rors-cmip-confirmed-action) with the PM-Segment attributes it supports, record the attributes "Segment-Start-Abs-Time" and "Segment-End-Abs-Time" of every PM-Segment. <p>IF bit 8 of PMStoreCapab was set:</p> <ol style="list-style-type: none"> 6. The simulated manager sends a Clear-Segment: <ol style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action, <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_CLEAR <input type="checkbox"/> SegmSelection = abs-time-range, selecting a range with its boundaries set to Segment-Start-Abs-Time and Segment-End-Abs-Time of one of the PM-Segments 7. The agent under test operation response: <ol style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Response Confirmed Action, <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_CLEAR 8. Delay. 9. The simulated manager sends a request for the PM-Segment Data of one of the cleared PM-Segments: <ol style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action, <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_TRIG_XFER <input type="checkbox"/> SegmSelection = <Instance number of the selected PM-Segment that contained data before the clear-segment action in step 6> 10. The agent issues an action response with the Data: <ol style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action, 		

	<ul style="list-style-type: none"> <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_TRIG_XFER <input type="checkbox"/> TrigSegmXferRsp = <ul style="list-style-type: none"> ▪ IF pmsc-clear-segm-remove is NOT set then <ul style="list-style-type: none"> o TrigSegmXferRsp = tsxr-fail-segm-empty ▪ ELSE then <ul style="list-style-type: none"> o TrigSegmXferRsp = tsxr-fail-no-such-segment
Pass/Fail criteria	<ul style="list-style-type: none"> • In step 7 the agent must send a confirmation • In step 10 the TrigSemgXferRsp must be the specified
Notes	See Note for test case TP/PLT/AG/OSP/DIM/BV-013_B.

TP Id	TP/PLT/AG/OSP/DIM/BV-013_G			
TP label	PM-Store object methods. Clear-Segments Time Range method 6			
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	PM-StoreMeth 1; O	PM-StoreMeth 2; C	PM-StoreMeth 6; M
		PM-StoreMeth 8; O	PersStoreMtrDatTransf 20; M	PersStoreMtrDatTransf 21; M
		PersStoreMtrDatTransf 22; O	PersStoreMtrDatTransf 23; M	PM-StoreMeth 19; M
PM-StoreMeth 23; C		PersStoreMtrDatTransf 24; M		
Applicability	C_AG_OXP_041 AND C_AG_OXP_071 AND C_AG_OXP_009 AND C_AG_OXP_000			
Initial condition	The simulated manager and the agent under test are in the operating state and the agent has at least one PM-Segment with data stored.			
Test procedure	<ol style="list-style-type: none"> 1. Make sure the agent is not taking measures which are stored in PM-Segments. 2. The simulated manager shall send a Get request for the PM-Store object with an attribute-id-list set to 0 to indicate all PM-Store attributes. 3. The agent issues a GET response with the PM-Store attributes, record the values of the PM-Store-Capab attribute: <ol style="list-style-type: none"> a. PM-Store-Capab: <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_PM_STORE_CAPAB <input type="checkbox"/> attribute-type = PmStoreCapab <input type="checkbox"/> attribute-value = Record the value of bit 8 (Indicates that PM-Segments in the SegmSelection data type can be cleared by defining an AbsTimeRange) <p>IF bit 8 of PMStoreCapab was set:</p> 4. The simulated manager shall send a Get-Segment-Info object action for the PM-Store object with SegmSelection set to all-segments. 5. The agent issues a response (rors-cmip-confirmed-action) with the PM-Segment attributes it supports, record the attributes "Segment-Start-Abs-Time" and "Segment-End-Abs-Time" of every PM-Segment. 6. The simulated manager sends a Clear-Segment: <ol style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action, <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_CLEAR <input type="checkbox"/> SegmSelection = abs-time-range, selecting a range with its boundaries set to include inside from Segment-Start-Abs-Time to Segment-End-Abs-Time of one of the PM-Segments 			

	<p>7. The agent under test operation response:</p> <p>a. Data APDU</p> <ul style="list-style-type: none"> <input type="checkbox"/> Type = Response Confirmed Action, <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_CLEAR <p>8. Delay.</p> <p>9. The simulated manager sends a request for the PM-Segment Data of one of the cleared PM-Segments:</p> <p>a. Data APDU</p> <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action, <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_TRIG_XFER <input type="checkbox"/> SegmSelection = Instance number of the selected PM-Segment that contained data before the clear-segment action in step 6 <p>10. The agent issues an action response with the Data</p> <p>a. Data APDU</p> <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action, <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_TRIG_XFER <input type="checkbox"/> TrigSegmXferRsp = <ul style="list-style-type: none"> ▪ IF pmsc-clear-segm-remove is NOT set then <ul style="list-style-type: none"> o TrigSegmXferRsp = tsxr-fail-segm-empty ▪ ELSE then <ul style="list-style-type: none"> o TrigSegmXferRsp = tsxr-fail-no-such-segm <p>IF bit 8 of PMStoreCapab was NOT set</p> <p>11. Simulated Manager sends a Clear-Segment:</p> <p>a. Data APDU</p> <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action, <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_CLEAR <input type="checkbox"/> SegmSelection = abs-time-range, selecting a range with its boundaries set to the absolut minimum of Abosulte-Time type and to the absolute maximun of the Absolute-Time type <p>12. The agent under test operation response:</p> <p>a. Data APDU</p> <ul style="list-style-type: none"> <input type="checkbox"/> Type = Roer <input type="checkbox"/> ErrorResult = no-allowed-by-object (24)
Pass/Fail criteria	<ul style="list-style-type: none"> • In step 7 the agent must send a confirmation • In step 10 the TrigSemgXferRsp must be the specified • If the agent does not support Clear-Segment by time, the agent must send a roer (not-allowed-by -object) message.
Notes	

TP Id	TP/PLT/AG/OXP/DIM/BV-014
TP label	PM-Store object methods. Clear-Segments method 1

Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	PM-StoreMeth 7; M	StoreClassAttr 6; M	PM-StoreMeth 29; M
Applicability		C_AG_OXP_041 AND C_AG_OXP_000 AND C_AG_OXP_071 AND C_AG_OXP_018		
Initial condition		The simulated manager and the agent under test are in the operating state.		
Test procedure		<ol style="list-style-type: none"> 1. Take measurements with the agent of a value that is stored on a PM-Segment. 2. The simulated manager shall send a Get request for the PM-Store object with an attribute-id-list set to 0 to indicate all PM-Store attributes. 3. The agent issues a GET response with the PM-Store attributes. 4. The simulated manager shall send a Get-Segment-Info object action with segmSelection set to all-segments to check what Segments are in use. 5. The simulated manager sends a Clear-Segment to all segments: <ol style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action, <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_CLEAR <input type="checkbox"/> SegmSelection = all-segments 6. The agent under test operation response: <ol style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = roer <input type="checkbox"/> value = not-allowed-by-object <input type="checkbox"/> value-returncode = MDC_RET_CODE_OBJ_BUSY 		
Pass/Fail criteria		The agent must respond with the specified error.		
Notes		The purpose of this test is to check that it is not possible to clear a segment that is in use, i.e. operational-state set to 1 by the agent.		

TP Id		TP/PLT/AG/OXP/DIM/BV-016		
TP label		PM-Store object methods. Trig-Segment-Data-Xfer method 1		
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	PM-StoreMeth 13; M	PM-StoreMeth 14; M	PM-StoreEvent 1; M
		PM-StoreEvent 2; M	CommonCharac 3; M	PM-StoreMeth 1; M
	Spec	[IEEE 11073-10406]		
Testable items	PMStoreObjMeth3; M	PMStoreObjEvent1; M	PMStoreObjEvent2; M	
Applicability		C_AG_OXP_041 AND C_AG_OXP_000		
Initial condition		The simulated manager and the agent under test are in the operating state and the agent has at least one PM-Segment with more data loaded than the maximum allowed by specialization.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated manager issues a GET for the PM-Store object. 2. The agent under test responds with the attributes of the PM-Store. 3. The simulated manager issues a Get-Segment-Info with SegmSelection set to all-segments. 4. The simulated manager sends a request for the PM-Segment Data to one of the PM-Segments that contains data: <ol style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action, <input type="checkbox"/> HANDLE = obj-handle 		

	<ul style="list-style-type: none"> <input type="checkbox"/> Action = MDC_ACT_SEG_TRIG_XFER <input type="checkbox"/> TrigSegmDataXferReq = <Instance number of the selected PM-Segment that contains the data> <p>5. The agent issues an action response:</p> <ul style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action, <input type="checkbox"/> HANDLE = obj-handle b. Action = MDC_ACT_SEG_TRIG_XFER <ul style="list-style-type: none"> <input type="checkbox"/> TrigSegmDataXferRsp = <Same Instance number> tsxr-succesful (0x00 0x00) <p>6. The agent under test starts Data transfer:</p> <ul style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Invoke CfmEventReport <input type="checkbox"/> Action = MDC_NOTI_SEGMENT_DATA <input type="checkbox"/> SegmentDataEvent <p>7. The simulated manager response to transferred data APDU's:</p> <ul style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action,, <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_NOTI_SEGMENT_DATA <input type="checkbox"/> SegmentDataResult <p>8. The agent under test repeats steps 6 and 7 until all the data is transferred.</p>
Pass/Fail criteria	<ul style="list-style-type: none"> • All checked values are as specified in the test procedure • Data is transferred • The total size of the response can not exceed the sum of the APDU sizes of the supported specializations (limited to an absolute limit of 64512 octets): <ul style="list-style-type: none"> ○ Pulse oximeter -> 9216 octets ○ Weighing scales -> 896 octets ○ Glucose meter -> 5120 octets or 64512 octets if agent supports PM-Store ○ Blood pressure -> 896 octets ○ Thermometer -> 896 octets ○ Independent activity hub -> 5120 octets ○ Cardiovascular -> 64512 octets or 6624 octets if it supports Step Counter Profile ○ Strength -> 64512 octets ○ Adherence monitor -> 1024 octets ○ Peak Flow -> 2030 octets ○ Body Composition Analyzer -> 7730 octets ○ Basic ECG/Simple ECG -> 7168 octets or 64512 octets if the agent supports PM-Store ○ Basic ECG/Heart rate -> 1280 octets or 64512 octets if the agent supports PM-Store ○ International normalized ratio -> 896 octets or 64512 octets if the agent supports PM-Store
Notes	

TP Id		TP/PLT/AG/OXP/DIM/BV-017	
TP label		PM-Store object methods. Trig-Segment-Data-Xfer method 2	
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	PM-StoreMeth 16; M	PM-SegmAttr 4; M
Applicability		C_AG_OXP_041 AND C_AG_OXP_000 AND C_AG_OXP_018	
Initial condition		The simulated manager and the agent under test are in the operating state and the agent has at least one PM-Segment.	
Test procedure		<ol style="list-style-type: none"> 1. The simulated manager issues a GET for the PM-Store object. 2. The agent under test responds with the attributes of the PM-Store. 3. The simulated manager issues a Get-Segment-Info with SegmSelection set to all-segments. 4. The simulated manager sends a request for the PM-Segment Data to one of the PM-Segments that is being used (OperationalState bit enabled): <ol style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action, <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_TRIG_XFER <input type="checkbox"/> TrigSegmDataXferReq = <Instance number of the selected PM-Segment that contains the data> 5. The agent issues a "roer" message with reason = not-allowed-by-object (24) and return code = MDC_RET_CODE_OBJ_BUSY. 	
Pass/Fail criteria		The agent must respond with the specified error.	
Notes			

TP Id		TP/PLT/AG/OXP/DIM/BV-020	
TP label		Scanner object services. SET Operational-State service	
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	ScanClassServ 1; M	
	Spec	[IEEE 11073-10406]	
	Testable items	PerScanObjAttr9; M	EpiScanObjAttr9; M
Applicability		(C_AG_OXP_046 OR C_AG_OXP_047) AND C_AG_OXP_000	
Initial condition		The simulated manager and the agent under test are in the operating state.	
Test procedure		<ol style="list-style-type: none"> 1. The simulated manager sends the scanner a SET Operational-State attribute: <ol style="list-style-type: none"> a. If C_AG_OXP_180 THEN APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Remote Operation Invoke Confirmed Event Report <input type="checkbox"/> roiv-cmip-confirmed-set <input type="checkbox"/> attribute = OperationalState <input type="checkbox"/> value = 0 b. If not C_AG_OXP_180 THEN APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Remote Operation Invoke Event Report <input type="checkbox"/> roiv-cmip-set <input type="checkbox"/> attribute = OperationalState <input type="checkbox"/> value = 0 	

	<p>2. If C_AG_OXP_180 the agent under test must respond with a confirmation ELSE no response for roiv-cmip-set will be received.</p> <p>If C_AG_OXP_180, verify the invoke-id is mirrored from the Set request:</p> <p>a. APDU</p> <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action <input type="checkbox"/> result = accepted
Pass/Fail criteria	The procedure is executed without errors.
Notes	The semantics of the Operational-State Attribute are tested in TP/PLT/AG/OXP/COM/BV-056.

TP Id	TP/PLT/AG/OXP/DIM/BV-021		
TP label	CfgScanner object Attributes. Confirm-Timeout operation		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	CfgScanAttr 3; C	CfgScanAttr 4; C
		OperErrorCond 6; M	TimeOutVar 3; C
Applicability	(C_AG_OXP_046 OR C_AG_OXP_047) AND C_AG_OXP_053 AND C_AG_OXP_000		
Initial condition	The simulated manager and the agent under test are in the unassociated state.		
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with a MDC_NOTI_CONFIG event to send its configuration to the manager, record the Scanner attribute Confirm-Timeout, if it is not present the simulated manager will use 3s as a default value. 4. Wait for the agent under test and the simulated manager to reach the operating state. 5. Take some measurements in the agent. 6. The simulated manager sets the operational state of the scanner to 1: <ol style="list-style-type: none"> a. If C_AG_OXP_180 THEN APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Remote Operation Invoke Confirmed Event Report <input type="checkbox"/> roiv-cmip-confirmed-set <input type="checkbox"/> attribute = OperationalState <input type="checkbox"/> value = 1 b. If not C_AG_OXP_180 THEN APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Remote Operation Invoke Event Report <input type="checkbox"/> roiv-cmip-set <input type="checkbox"/> attribute = OperationalState <input type="checkbox"/> value = 1 7. If C_AG_OXP_180 the agent under test must respond with a confirmation <ol style="list-style-type: none"> a. APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action <input type="checkbox"/> result = accepted 8. Wait until the agent under test starts to send data. 9. The manager must not respond for at least the Confirm-Timeout time. 		
Pass/Fail criteria	The agent must wait for a Confirmed Event Report Response message for a Confirm-TimeOut period. If the time expires, the agent must send an abort to the manager.		
Notes			

TP Id		TP/PLT/AG/OSP/DIM/BV-023		
TP label		EpiCfScanner object. Reports		
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	EpiCfScannerClass 1; M	EpiCfScannerClass 2; M	EpiCfScannerEvent 1; C
		ScanClassConcept8; C	ScanClassConcept9; C	EpiCfScannerEvent 28; M
Applicability		C_AG_OXP_047 AND C_AG_OXP_000		
Initial condition		The simulated manager and the agent under test are in the unassociated state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 4. The Configurable Episodic Scanner object attribute of interest for this test is: <ol style="list-style-type: none"> a. Attribute Min-Interval-Reporting <ul style="list-style-type: none"> <input type="checkbox"/> TP/PLT/AG/OSP/DIM/BV-001_B <input type="checkbox"/> attribute-type = RelativeTime <input type="checkbox"/> attribute-length = 4 bytes <input type="checkbox"/> attribute-value = < Record for later comparison > 5. The simulated manager sets the operational state of the scanner to 1. 6. Take a measurement with the agent under test. 7. Check that the simulated manager receives the Event sent by the agent with the changed value and reports it with a grouped type event: <ol style="list-style-type: none"> a. PrstApdu <ul style="list-style-type: none"> <input type="checkbox"/> Remote Operation Invoke Confirmed Event Report <input type="checkbox"/> Type = MDC_NOTI_UNBUF_SCAN_REPORT_GROUPED (0x0D 0x24) or MDC_NOTI_UNBUF_SCAN_REPORT_VAR (0x0D 0x22) or MDC_NOTI_UNBUF_SCAN_REPORT_FIXED (0x0D 0x23) or MDC_NOTI_UNBUF_SCAN_REPORT_MP_GROUPED (0x0D 0x27) or MDC_NOTI_UNBUF_SCAN_REPORT_MP_VAR (0x0D 0x25) or MDC_NOTI_UNBUF_SCAN_REPORT_MP_FIXED (0x0D 0x26) 8. Take measurements faster than the Reporting Interval recorded in step 4. 9. Wait for the next event report. 10. If it is possible, force the agent not to change the values that are collected by the scanner object. 11. Wait for the next event report. 12. Check that no scanner event report is sent. 		
Pass/Fail criteria		<ul style="list-style-type: none"> • The agent sends an event report when the attribute changes • The received events are of grouped, variable or fixed type • The event reports are not sent at a rate faster than the minimum reporting interval 		
Notes				

TP Id		TP/PLT/AG/OSP/DIM/BV-027		
TP label		EpiCfScanner object events. Unbuf-Scan-Report		
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	EpiCfScannerEvent 10; C	EpiCfScannerEvent 11; C	EpiCfScannerEvent 13; C
		EpiCfScannerEvent 30; C	EpiCfScannerEvent 22; C	EpiCfScannerEvent 23; C
		EpiCfScannerEvent 25; C	EpiCfScannerEvent 33; C	EpiCfScannerEvent 2; C

		EpiCfgScanEvent 3; C	EpiCfgScanEvent 5; C	EpiCfgScanEvent 6; C
		EpiCfgScanEvent 7; C	EpiCfgScanEvent 9; C	EpiCfgScanEvent 14; C
		EpiCfgScanEvent 15; C	EpiCfgScanEvent 17; C	EpiCfgScanEvent 18; C
		EpiCfgScanEvent 19; C	EpiCfgScanEvent 21; C	EpiCfgScanEvent 28; C
		EpiCfgScanEvent 29; C	EpiCfgScanEvent 31; C	EpiCfgScanEvent 32; C
		ScannerGeneral1; O		
	Spec	[IEEE 11073-10406]		
	Testable items	EpiScanObjEv1; M	EpiScanObjEv3; M	EpiScanObjEv4; M
		EpiScanObjEv5; M	EpiScanObjEv6; M	EpiScanObjEv6; M
		EpiScanObjEv7; M	EpiScanObjEv8; M	ObjAccServ2; O
	Spec	[ITU-T H.810]		
	Testable items	General 7; C		
Applicability	C_AG_OXP_047 AND C_AG_OXP_000			
Initial condition	The simulated manager and the agent under test are in the operating state.			
Test procedure	<ol style="list-style-type: none"> 1. Take some measurements with the agent under test. 2. The simulated manager sets the operational state of the scanner to 1. 3. Wait until the agent under test starts to send its data. 4. Check that the agent uses the ScanReportInfoGrouped Event-info parameter, whenever data values change: <ol style="list-style-type: none"> a. PrstApdu <ul style="list-style-type: none"> <input type="checkbox"/> Remote Operation Invoke Confirmed Event Report OR Event Report <input type="checkbox"/> Event-Type = MDC_NOTI_UNBUF_SCAN_REPORT_GROUPED (0x0D 0x24) <ul style="list-style-type: none"> <input type="checkbox"/> scanReportInfoGrouped:SEQUENCE of: <ul style="list-style-type: none"> ▪ data-req-id = <Not relevant for this test> ▪ scan-report-no = <counter for detection of missin scan reports> ▪ obs-scan-grouped = SEQUENCE OF octect strings <input type="checkbox"/> Or MDC_NOTI_UNBUF_SCAN_REPORT_MP_GROUPED (0x0D 0x27) <ul style="list-style-type: none"> <input type="checkbox"/> ScanReportInfoMPGrouped.scan-per-grouped = SEQUENCE of: <ul style="list-style-type: none"> ▪ person-id.value = <record for comparison> ▪ obs-scan-grouped = <Not relevant for this Test> <input type="checkbox"/> Or MDC_NOTI_UNBUF_SCAN_REPORT_VAR (0x0D 0x22) <ul style="list-style-type: none"> <input type="checkbox"/> ScanReportInfoVar= SEQUENCE of: <ul style="list-style-type: none"> ▪ data-req-id = <Not relevant for this test> ▪ scan-report-no = <counter for detection of missin scan reports> ▪ obs-scan-var = SEQUENCE OF ObservationScan <input type="checkbox"/> Or MDC_NOTI_UNBUF_SCAN_REPORT_MP_VAR (0x0D 0x25) <ul style="list-style-type: none"> <input type="checkbox"/> ScanReportInfoMPVar.scan-per-var = SEQUENCE of: <ul style="list-style-type: none"> ▪ person-id.value = <record for comparison> ▪ obs-scan-var = <Not relevant for this Test> <input type="checkbox"/> Or MDC_NOTI_UNBUF_SCAN_REPORT_FIXED (0x0D 0x23) <ul style="list-style-type: none"> <input type="checkbox"/> ScanReportInfoFixed= SEQUENCE of: <ul style="list-style-type: none"> ▪ data-req-id = <Not relevant for this test> ▪ scan-report-no = <counter for detection of missin scan reports> 			

	<ul style="list-style-type: none"> ▪ obs-scan-fixed = SEQUENCE OF ObservationScanFixed ❑ Or MDC_NOTI_UNBUF_SCAN_REPORT_MP_FIXED (0x0D 0x26) ❑ ScanReportInfoMPFixed.scan-per-fixed = SEQUENCE of: <ul style="list-style-type: none"> ▪ person-id.value = <record for comparison> ▪ obs-scan-fixed = <Not relevant for this Test>
Pass/Fail criteria	<ul style="list-style-type: none"> • The agent sends data using grouped, variable or fixed event reports. • If the agent supports multi-person event reports for one or more episodic scanner object (C_AG_OXP_033= TRUE) THEN the agent uses MP Unbuf Event report, and check that every person-id is different from each other or "unkown-person-id" (65535). • If C_AG_OXP_033= TRUE and MP event reports have been received, a pop-up will show the received measurements to make the operator identify if measurements have been correctly assigned to every person.
Notes	

TP Id	TP/PLT/AG/OXP/DIM/BV-032_A			
TP label	PeriCfgScanner object Attribute. Reporting interval attribute 1			
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	PeriCfgScanClass 1; M	PeriCfgScanAttr 2; M	PeriCfgScanEvent 26; M
		PeriCfgScanClass 3; R	ScanClassConcep8; C	ScanClassConcep9; C
Applicability	C_AG_OXP_046 AND C_AG_OXP_000			
Initial condition	The simulated manager and the agent under test have been associated, but the agent configuration is unknown for simulated manager, so the agent and the simulated manager will be in the configuring state.			
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 4. The Configurable Periodic Scanner object attribute of interest for this test is: <ol style="list-style-type: none"> a. Mandatory attribute Reporting-Interval <ul style="list-style-type: none"> ❑ attribute-id = MDC_ATTR_SCAN_REP_PD ❑ attribute-type = RelativeTime ❑ attribute-length = 4 bytes ❑ attribute-value = < Record for later comparison > 5. The simulated manager sets the operational state of the scanner to 1. 6. Take several measurements. 7. Wait until the agent under test starts to send its data. 8. Wait for the next event report. 9. If it is possible, force the agent not to change the values that are collected by scanner object. 10. Wait for the next event report. 11. Check that an empty event report is sent. 			
Pass/Fail criteria	<p>The Event reports must arrive periodically with a period the same as the time defined in Reporting-Interval.</p> <p>The time between the scanner being enabled (step 5)and the sending of the first scan report (step 6) should not exceed the reporting interval plus 15 seconds.</p> <p>An empty Scan Event Report is sent by the agent under test when the value has not changed (step 11).</p>			
Notes				

TP Id		TP/PLT/AG/OSP/DIM/BV-032_B		
TP label		PeriCfgScanner object Attribute. Reporting interval attribute 2		
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	PeriCfgScanClass 1; M	PeriCfgScanAttr 2; M	PeriCfgScanEvent 26; M
Applicability		C_AG_OXP_046 AND C_AG_OXP_000		
Initial condition		The simulated manager and the agent under test have been associated, but the agent configuration is unknown for the simulated manager, so the agent and the simulated manager will be in the configuring state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 4. The Configurable Periodic Scanner object attribute of interest for this test is: <ol style="list-style-type: none"> a. Mandatory attribute Reporting-Interval <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SCAN_REP_PD <input type="checkbox"/> attribute-type = RelativeTime <input type="checkbox"/> attribute-length = 4 bytes <input type="checkbox"/> attribute-value = < Record for later comparison > 5. The simulated manager sets the operational state of the scanner to 1. 6. Wait until the agent under test sends two event reports. 7. Take measurements faster than the Reporting Interval recorded in step 4. 8. Wait for the next event report. 		
Pass/Fail criteria		<ul style="list-style-type: none"> • In step 6 verify that the received observed value is the same for the two events (same objects and attributes, but not attribute value). • Verify that in step 8 the received event contains a number of measurements higher than the number of measurements received in step 6. 		
Notes		<p>In last paragraph of clause 6.3.9.5.1 there is an example where states that it must send ALL the measurements, not only the last change. It has to send all the registered observations.</p> <p>Example: A Periodic Configurable Scanner is set up to 'scan' two Metric objects with a Reporting-Interval of 1 sec. The two objects update their corresponding observed value periodically with an interval of 1 sec and ½ sec respectively. The Periodic Configurable Scanner then issues Event Reports every second containing one observation scan of Metric object #1 and two observation scans of Metric object #2.</p>		

TP Id		TP/PLT/AG/OSP/DIM/BV-036		
TP label		PeriCfgScanner object events. Buf-Scan-Report		
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	PeriCfgScanEvent 10; C	PeriCfgScanEvent 11; C	PeriCfgScanEvent 13; C
		PeriCfgScanEvent 1; C	PeriCfgScanEvent 22; C	PeriCfgScanEvent 23; C
		PeriCfgScanEvent 25; C	PeriCfgScanEvent 2; C	PeriCfgScanEvent 3; C
		PeriCfgScanEvent 5; C	PeriCfgScanEvent 6; C	PeriCfgScanEvent 7; C
		PeriCfgScanEvent 9; C	PeriCfgScanEvent 14; C	PeriCfgScanEvent 15; C
		PeriCfgScanEvent 17; C	PeriCfgScanEvent 18; C	PeriCfgScanEvent 19; C
		PeriCfgScanEvent 21; C	ScannerGeneral1; O	
Spec	[IEEE 11073-10406]			

	Testable items	PerScanObjEv1; M	PerScanObjEv3; M	PerScanObjEv4; M
		PerScanObjEv5; M	PerScanObjEv6; M	PerScanObjEv7; M
		PerScanObjEv8; M	ObjAccServ2; O	
	Spec	[ITU-T H.810]		
Testable items	General 7; C			
Applicability	C_AG_OXP_046 AND C_AG_OXP_000			
Initial condition	The simulated manager and the agent under test are in the operating state.			
Test procedure	<ol style="list-style-type: none"> 1. Make a change to one of the observed values by the agent under test. 2. The simulated manager sets the operational state of the scanner to 1. 3. Wait until the agent under test starts to send its data. 4. Check that the simulated manager receives the Event send by the agent with the changed value and reports it with a grouped type event: <ol style="list-style-type: none"> a. DataApdu <ul style="list-style-type: none"> <input type="checkbox"/> Remote Operation Invoke Confirmed Event Report or Event Report <input type="checkbox"/> Event-Type = MDC_NOTI_BUF_SCAN_REPORT_GROUPED (0x0D 0x2A) <ul style="list-style-type: none"> <input type="checkbox"/> ScanReportInfoGrouped:SEQUENCE of: <ul style="list-style-type: none"> ▪ data-req-id = <Not relevant for this test> ▪ scan-report-no = <counter for detection of missin scan reports> ▪ obs-scan-grouped = SEQUENCE OF octect strings <input type="checkbox"/> Or MDC_NOTI_BUF_SCAN_REPORT_MP_GROUPED (0x0D 0x2D) <ul style="list-style-type: none"> <input type="checkbox"/> ScanReportInfoMPGrouped.scan-per-grouped = SEQUENCE of: <ul style="list-style-type: none"> ▪ person-id.value = <record for comparison> ▪ obs-scan-grouped = <Not relevant for this Test> <input type="checkbox"/> Or MDC_NOTI_BUF_SCAN_REPORT_VAR (0x0D 0x28) <ul style="list-style-type: none"> <input type="checkbox"/> ScanReportInfoVar= SEQUENCE of: <ul style="list-style-type: none"> ▪ data-req-id = <Not relevant for this test> ▪ scan-report-no = <counter for detection of missin scan reports> ▪ obs-scan-var = SEQUENCE OF ObservationScan <input type="checkbox"/> Or MDC_NOTI_BUF_SCAN_REPORT_MP_VAR (0x0D 0x2B) <ul style="list-style-type: none"> <input type="checkbox"/> ScanReportInfoMPVar.scan-per-var = SEQUENCE of: <ul style="list-style-type: none"> ▪ person-id.value = <record for comparison> ▪ obs-scan-var = <Not relevant for this Test> <input type="checkbox"/> Or MDC_NOTI_BUF_SCAN_REPORT_FIXED (0x0D 0x29) <ul style="list-style-type: none"> <input type="checkbox"/> ScanReportInfoFixed= SEQUENCE of: <ul style="list-style-type: none"> ▪ data-req-id = <Not relevant for this test> ▪ scan-report-no = <counter for detection of missin scan reports> ▪ obs-scan-fixed = SEQUENCE OF ObservationScanFixed <input type="checkbox"/> Or MDC_NOTI_BUF_SCAN_REPORT_MP_FIXED (0x0D 0x2C) <ul style="list-style-type: none"> <input type="checkbox"/> ScanReportInfoMPFixed.scan-per-fixed = SEQUENCE of: <ul style="list-style-type: none"> ▪ person-id.value = <record for comparison> ▪ obs-scan-fixed = <Not relevant for this Test> 			
Pass/Fail criteria	<ul style="list-style-type: none"> • The agent sends grouped, variable or fixed format event reports. 			

	<ul style="list-style-type: none"> • If the agent supports multi-person event reports for one or more periodic scanner object (C_AG_OXP_034= TRUE) THEN the agent uses MP Buf Event report, check that every person-id is different from each other or "unkown-person-id" (65535). • If C_AG_OXP_034 = TRUE and MP event reports have been received, a pop-up will show the received measurements to make the operator identify if the measurements have been correctly assigned to every person.
Notes	

TP Id	TP/PLT/AG/OXP/DIM/BV-042		
TP label	CfgScanner object. Confirm-Mode attribute		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable Items	CfgScanAttr 1;M	
Applicability	(C_AG_OXP_046 OR C_AG_OXP_047) AND C_AG_OXP_000		
Initial condition	The simulated manager and the agent under test are in the unassociated state.		
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager, record the Scanner attribute Confirm-Mode. 4. Wait for the agent under test and the simulated manager to reach the operating state. 5. The simulated manager sets the operational state of the scanner to 1. 6. Wait until the agent under test starts to send data. 7. Check that the simulated manager receives the Event sent by the agent: IF the Confirm-Mode recorded in step 3, is "confirmed": <ol style="list-style-type: none"> a. PrstApdu <ul style="list-style-type: none"> <input type="checkbox"/> Remote Operation Invoke Confirmed Event Report IF the Confirm-Mode recorded in step 3, is "unconfirmed": <ol style="list-style-type: none"> a. PrstApdu <ul style="list-style-type: none"> <input type="checkbox"/> Remote Operation Invoke Event Report 		
Pass/Fail criteria	IF the Confirm-Mode value is confirmed THEN the agent sends a Confirmed Event Report. IF the Confirm-mode value is unconfirmed THEN the agent sends an Unconfirmed Event Report.		
Notes			

TP Id	TP/PLT/AG/OXP/DIM/BV-043		
TP label	PM-Store object. Change Unit Code attribute		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	StoreClassGen 1; M	
Applicability	(C_AG_OXP_073) AND C_AG_OXP_000 AND C_AG_DGC_018		
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<ol style="list-style-type: none"> 1. Make a change to the contextual attribute Unit-Code for an object that is stored in the PM-Store. 2. The simulated manager sends a request (Get-Segment-Info) for the PM-Segment attributes with SegmSelection = 1 to obtain all the segments for the PM-Store: <ol style="list-style-type: none"> a. Data APDU 		

	<ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action, <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_GET_INFO <input type="checkbox"/> SegmSelection = all-segments <p>3. The agent issues a response with the PM-Segments attributes:</p> <p>a. Data APDU</p> <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action, <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_GET_INFO <input type="checkbox"/> SegmentInfoList: Record value for PM-Segment-Entry-Map attribute <p>4. Repeat steps 1 and 2 for every PM-Store.</p>
Pass/Fail criteria	In step 2, there is at least one segment that stores Unit-code attribute (PM-Segment-Entry-Map).
Notes	

TP Id	TP/PLT/AG/OXP/DIM/BV-044		
TP label	PeriCfgScanner object. Reporting interval and FIFO		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	ScanClassConcep4; M	ScanClassConcep7; M
Applicability	C_AG_OXP_046 AND C_AG_OXP_000		
Initial condition	The simulated manager and the agent under test are in the unassociated state.		
Test procedure	<p>1. The simulated manager receives an association request from the agent under test.</p> <p>2. The simulated manager responds with a result = accepted-unknown-config.</p> <p>3. The agent under test responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager.</p> <p>4. Record attribute for Periodic Scanner Object:</p> <p>a. Mandatory attribute Reporting-Interval</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SCAN_REP_PD <input type="checkbox"/> attribute-type = RelativeTime <input type="checkbox"/> attribute-length = 4 bytes <input type="checkbox"/> attribute-value = < Record for later comparison > <p>5. Wait for 4 * Reporting Interval (the reasonable time for the scanner being enabled and the sending of the first scan report) or 4*15 seconds, whichever is greater.</p> <p>6. Set the Operational State to 1 for the Periodic Scanner object.</p> <p>7. Wait until the agent under test starts to send its data and record it.</p> <p>8. Set the Operational State to 0 for the Periodic Scanner object.</p> <p>9. Wait for 4 * Reporting Interval or 4*15 seconds, whichever is greater.</p> <p>10. Set the Operational State to 1 for the Periodic Scanner object.</p> <p>11. Wait until the agent under test starts to send its data and record it.</p>		
Pass/Fail criteria	In steps 7 and 11 the same number of observations must be received from the agent. In step 11, check that measurements have been received following a FIFO sequence.		
Notes			

TP Id		TP/PLT/AG/OSP/DIM/BV-045		
TP label		PM-Store object methods. Clear-Segments Base-Offset-Time Range method 1		
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	PM-StoreMeth 1; O	PM-StoreMeth 2; C	PM-StoreMeth 6; M
		PM-StoreMeth 8; O	PersStoreMtrDatTransf 20; M	PersStoreMtrDatTransf 21; M
		PersStoreMtrDatTransf 22; O	PersStoreMtrDatTransf 23; M	PM-StoreMeth 32; M
PM-StoreMeth 33; M				
Applicability		C_AG_OXP_041 AND C_AG_OXP_071 AND C_AG_OXP_072 AND C_AG_OXP_014 AND C_AG_OXP_000		
Initial condition		The simulated manager and the agent under test are in the operating state and the agent supports at least one PM-Segment with data stored.		
Test procedure		<ol style="list-style-type: none"> 1. Make sure the agent under test is not taking measurements which are stored in PM-Segments. 2. The simulated manager shall send a Get request for the PM-Store object with an attribute-id-list set to 0 to indicate all PM-Store attributes. 3. The agent under test issues a GET response with the PM-Store attributes, record the values of the PM-Store-Capab attribute: <ol style="list-style-type: none"> a. PM-Store-Capab: <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_PM_STORE_CAPAB <input type="checkbox"/> attribute-type = PmStoreCapab <input type="checkbox"/> attribute-value = Record the value of bit 8 (Indicates that PM-Segments in the SegmSelection data type can be cleared by defining a time range) 4. The simulated manager shall send a Get-Segment-Info object action for the PM-Store object with SegmSelection set to all-segments. 5. The agent issues a response (rors-cmip-confirmed-action) with the PM-Segment attributes it supports, record the attributes "Segment-Start-BO-Time" and "Segment-End-BO-Time" of every PM-Segment. <p>IF bit 8 of PMStoreCapab was set:</p> <ol style="list-style-type: none"> 6. The simulated manager sends a Clear-Segment: <ol style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action, <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_CLEAR <input type="checkbox"/> SegmSelection = bo-time-range, selecting a range with its boundaries that is not within the Segment-Start-BO-Time and Segment-End-BO-Time 7. The agent under test operation response: <ol style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Roer <input type="checkbox"/> ErrorResult = no-such-action (9) 		
Pass/Fail criteria		In step 7 the agent must send the specified error.		
Notes		<p>Error code was not clearly defined in the spec.</p> <p>In the new edition of [ISO/IEEE 11073-20601A], the clear-segment using time range has been clarified. "For PM-segments cleared using the by time method, only PM-segments having Segment-Start-Abs-Time and Segment-End-Abs-Time fields entirely within the specified time period are cleared.", but if the manager sends a Clear-Segment but the segment has not a Segment-Start-Abs-Time and Segment-End-Abs-Time within the specified time-period, the agent will send a Roer message.</p>		

	At this point, it would be up to the agent what error code (Roer message) to send (No-such-action, not-allowed-by-object, or both). If one wants to clear the segment due to all the internal timestamps that were saved in the segment as falling inside the given time period, then the agent could do that. Returning an error is also possible.
--	---

TP Id		TP/PLT/AG/OXP/DIM/BV-046		
TP label		PM-Store object methods. Clear-Segments Base-Offset-Time Range method 2		
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	PM-StoreMeth 1; O	PM-StoreMeth 2; C	PM-StoreMeth 6; M
		PM-StoreMeth 8; O	PersStoreMtrDatTransf 20; M	PersStoreMtrDatTransf 21; M
		PersStoreMtrDatTransf 22; O	PersStoreMtrDatTransf 23; M	PM-StoreMeth 32; M
PM-StoreMeth 33; M		PM-StoreMeth 23; C	PersStoreMtrDatTransf 24; M	
Applicability		C_AG_OXP_041 AND C_AG_OXP_071 AND C_AG_OXP_014 AND C_AG_OXP_000		
Initial condition		The simulated manager and the agent under test are in the operating state and the agent has at least one PM-Segment with data stored.		
Test procedure		<ol style="list-style-type: none"> 1. Make sure the agent is not taking measures which are stored in PM-Segments. 2. The simulated manager shall send a Get request for the PM-Store object with an attribute-id-list set to 0 to indicate all PM-Store attributes. 3. The agent issues a GET response with the PM-Store attributes, record the values of the PM-Store-Capab attribute: <ol style="list-style-type: none"> a. PM-Store-Capab: <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_PM_STORE_CAPAB <input type="checkbox"/> attribute-type = PmStoreCapab <input type="checkbox"/> attribute-value = Record the value of bit 8 (Indicates that PM-Segments in the SegmSelection data type can be cleared by defining a TimeRange) <p>IF bit 8 of PMStoreCapab was set:</p> <ol style="list-style-type: none"> 4. The simulated manager shall send a Get-Segment-Info object action for the PM-Store object with SegmSelection set to all-segments. 5. The agent issues a response (rors-cmip-confirmed-action) with the PM-Segment attributes it supports, record the attributes "Segment-Start-BO-Time" and "Segment-BO-Abs-Time" of every PM-Segment. 6. The simulated manager sends a Clear-Segment: <ol style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action, <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_CLEAR <input type="checkbox"/> SegmSelection = bo-time-range, selecting a range with its boundaries set to Segment-Start-BO-Time and to Segment-End-BO-Time of one of the PM-Segments. 7. The agent under test operation response: <ol style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Response Confirmed Action, <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_CLEAR 8. Delay. 9. The simulated manager sends a request for the PM-Segment Data of one of the cleared PM-Segments: 		

	<p>a. Data APDU</p> <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action, <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_TRIG_XFER <input type="checkbox"/> SegmSelection = Instance number of the selected PM-Segment that contained data before the clear-segment action in step 6 <p>10. The agent issues an action response with the Data</p> <p>a. Data APDU</p> <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action, <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_TRIG_XFER <input type="checkbox"/> TrigSegmXferRsp = <ul style="list-style-type: none"> ▪ IF pmsc-clear-segm-remove(9) = 0 THEN TrigSegmXferRsp = tsxr-fail-segm-empty ELSE TrigSegmXferRsp = tsxr-fail-no-such-segment <p>IF bit 8 of PMStoreCapab was NOT set</p> <p>11. Simulated Manager sends a Clear-Segment:</p> <p>a. Data APDU</p> <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action, <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_CLEAR <input type="checkbox"/> SegmSelection = bo-time-range, selecting a range with its boundaries set to the a minimum of Base-Offset-Time type and to the absolute of the Base-Offset-Time type <p>12. The agent under test operation response:</p> <p>a. Data APDU</p> <ul style="list-style-type: none"> <input type="checkbox"/> Type = Roer <input type="checkbox"/> ErrorResult = no-allowed-by-object (24)
Pass/Fail criteria	<ul style="list-style-type: none"> • In step 7 the agent must send a confirmation • In step 10 the TrigSemgXferRsp must be the specified • If the agent does not support Clear-Segment by time, the agent must send roer (not-allowed-by -object)
Notes	

A.3 Subgroup 1.2.2 – PHD service model (SER)

TP Id		TP/PLT/AG/OXP/SER/BV-000		
TP label		Object Access Services: No-Such-Action Error		
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	ObjAccessServ 3; M	DataTrans 2; O	MDSMethod 4; M
		PersStoreMtrDatTransf 24; M	MDSMethod 6; M	PM-StoreMeth 34; C
	Spec	[ITU-T H.810]		
Testable items	General 2; M			
Applicability		C_AG_OXP_000		
Initial condition		The simulated manager and the agent under test are in the operating state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated manager sends a roiv-cmip-confirmed-action with action-type = MDC_ACT_DATA_REQUEST. 2. The agent under test shall reply with an error. The expected fields sent by the agent are: <ol style="list-style-type: none"> a. Error-value <ul style="list-style-type: none"> <input type="checkbox"/> field-type = INT-U16 <input type="checkbox"/> field-length = 2 bytes <input type="checkbox"/> field- value = no-such-action(9) 3. IF the agent under test does not support Set Time (i.e., C_AG_OXP_008 = FALSE) OR the agent under test supports Set Time for Base-Offset-Time (i.e., C_AG_OXP_008 = TRUE and C_AG_OXP_014 = TRUE) THEN the simulated manager sends a roiv-cmip-confirmed-action with action-type = MDC_ACT_SET_TIME using Absolute-Time (SetTimeInvoke). 4. The agent under test shall reply with an error. The expected fields sent by the agent are: <ol style="list-style-type: none"> a. Error-value <ul style="list-style-type: none"> <input type="checkbox"/> field-type = INT-U16 <input type="checkbox"/> field-length = 2 bytes <input type="checkbox"/> field- value = no-such-action(9) 5. IF the agent under test does not support Set Time (i.e., C_AG_OXP_008 = FALSE) OR the agent under test supports Set Time for Absolute-Time Time (i.e., C_AG_OXP_008 = TRUE and C_AG_OXP_009 = TRUE) THEN the manager sends a roiv-cmip-confirmed-action with action-type = MDC_ACT_SET_BO_TIME using Base-Offset-Time (SetBOTimeInvoke). 6. The agent under test shall reply with an error. The expected fields sent by the agent are: <ol style="list-style-type: none"> a. Error-value <ul style="list-style-type: none"> <input type="checkbox"/> field-type = INT-U16 <input type="checkbox"/> field-length = 2 bytes <input type="checkbox"/> field- value = no-such-action(9) 7. IF C_AG_OXP_041 and does not support the Clear-Segment action then the simulated manager sends a roiv-cmip-confirmed-action with action-type = MDC_ACT_SEG_CLR. 8. The agent under test shall reply with an error. The expected fields sent by the agent are: <ol style="list-style-type: none"> a. Error-value <ul style="list-style-type: none"> <input type="checkbox"/> field-type = INT-U16 <input type="checkbox"/> field-length = 2 bytes 		

	<ul style="list-style-type: none"> <input type="checkbox"/> field- value = no-such-action(9) <p>9. IF C_AG_OXP_041 and it does not support selection by time range THEN the simulated manager sends a roiv-cmip-confirmed-action with action-type = MDC_ACT_SEG_GET_INFO and SegmSelection = abs-time-range (if the agent supports Absolute Time) or SegmSelection = bo-time-range (if the agent supports Base Offset Time)</p> <p>10. The agent under test shall reply with an error. The expected fields sent by the agent are</p> <ul style="list-style-type: none"> a. Error-value <ul style="list-style-type: none"> <input type="checkbox"/> field-type = INT-U16 <input type="checkbox"/> field-length = 2 bytes <input type="checkbox"/> field- value = no-such-action(9)
Pass/Fail criteria	The agent under test sends a No-Such-Action Error and the invoke-id is mirrored from the roiv-cmip-* messages.
Notes	

TP Id	TP/PLT/AG/OXP/SER/BV-001			
TP label	Configuration event report: dev-configuration-id is locally unique			
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	ConfEventRep 2; C	ConfEventRep 6; M	ConfEventRep 19; M
		ConfEventRep 24; M	ConfEventRep 25; R	ConfNormalProc 13; C
		ConfNormalProc 14;M	ConfExitCond 4; C	ConfEventRep 27; M
		AgentStateMach 79; M	ConfNormalProc 25; R	ConfEventRep 37; R
Applicability	C_AG_OXP_000			
Initial condition	The simulated manager and the agent under test are in the unassociated state.			
Test procedure	<ol style="list-style-type: none"> 1. The agent under test sends an Association Request to the simulated manager. The expected fields sent by the agent are: <ol style="list-style-type: none"> a. dev-config-id <ul style="list-style-type: none"> <input type="checkbox"/> field-type = ConfigId <input type="checkbox"/> field-length = 2 bytes <input type="checkbox"/> field- value = Record it for comparison b. Data-Req-Mode-Capab: <ul style="list-style-type: none"> <input type="checkbox"/> field-length = 4 bytes <input type="checkbox"/> field- value = 0xXX 0xXX 0x01 0xXX (Agent initiated) 2. The simulated manager responds with an accepted-unknown-config. 3. The agent sends a configuration event report, with the following fields: dev-config-id <ul style="list-style-type: none"> <input type="checkbox"/> field-type = ConfigId <input type="checkbox"/> field-length = 2 bytes <input type="checkbox"/> field- value = <Record this Value> 4. The simulated manager responds with an unsupported-configuration. 5. The agent sends a new configuration event report with a new configuration (if it has more). 6. Repeat the last two steps recording all the ConfigId-values until the agent sends a ReleaseRequest with the reason "no-more-configurations" and the agent moves to the disassociating state. 7. The simulated manager responds with a Release Response message. 8. Wait for the agent under test to send an Association Request. 			

	9. Repeat steps 2 to 5 until the agent under test sends a Release Request with the reason "no-more-configurations".
Pass/Fail criteria	<ul style="list-style-type: none"> • Dev-config-id is the same for the two first messages and every other ConfigId must be different from all others before the Release Request of step 6 • The agent shall send a Release-request (no-more-configurations) in step 6 • Verify that the agent should use the same Dev-Config-Id in steps 2 – 5 and that in step 9 and in every received configuration in step 9 it is the same as one of the received in the configuration messages in step 3
Notes	

TP Id	TP/PLT/AG/OXP/SER/BV-001_A		
TP label	Configuration event report: Maximum Size		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	CommonCharac 3; M	
Applicability	C_AG_OXP_000		
Initial condition	The simulated manager and the agent under test are in the unassociated state.		
Test procedure	<ol style="list-style-type: none"> 1. The agent under test sends an Association Request to the simulated manager. 2. The simulated manager responds with an "accepted-unkown". 3. The agent under test sends its configuration with an event report. Record the size of the event report. 4. The simulated manager responds with an unsupported-configuration. 5. The agent under test sends a new configuration event report with a new configuration (if it has more). Record the size of the event report. 6. Repeat the last two steps recording all the ConfigId-values until the agent sends a ReleaseRequest with the reason "no-more-configurations". 		
Pass/Fail criteria	<ul style="list-style-type: none"> • Total size of the response can not exceed the sum of the APDU sizes of the supported specializations (limited to an absolute limit of 64512 octets): <ul style="list-style-type: none"> ○ Pulse oximeter -> 9216 octets ○ Weighing scales -> 896 octets ○ Glucose meter -> 5120 octets or 64512 octets if the agent supports PM-Store ○ Blood pressure -> 896 octets ○ Thermometer -> 896 octets ○ Independent activity hub -> 5120 octets ○ Cardiovascular -> 64512 octets or 6624 octets if the agent supports Step Counter Profile ○ Strength -> 64512 octets ○ Adherence monitor -> 1024 octets ○ Peak flow -> 2030 octets ○ Body composition analyser -> 7730 octets ○ Basic ECG/Simple ECG -> 7168 octets or 64512 octets if the agent supports PM-Store ○ Basic ECG/Heart rate -> 1280 octets or 64512 octets if the agent supports PM-Store ○ International normalized ratio -> 896 octets or 64512 if the agent supports PM-Store 		
Notes			

TP Id		TP/PLT/AG/OXP/SER/BV-002		
TP label		Configuration event report: Change attributes values		
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	ConfEventRep 10; O	ConfEventRep 13; M	ConfEventRep 14; M
Applicability		C_AG_OXP_098 AND C_AG_OXP_000		
Initial condition		The simulated manager and the agent under test are in the unassociated state.		
Test procedure		<ol style="list-style-type: none"> 1. The agent under test sends an Association Request to the simulated manager. 2. The simulated manager responds with an accepted-unknown-config. 3. The agent responds with a roiv-cmip-confirmed-event report message with a MDC_NOTI_CONFIG event to send its configuration to the manager. 4. If ConfigId (ConfigReport) matches the tested configuration, the simulated manager responds with "accepted-config" and records the ConfigReport received in step 3. 5. Wait for the agent under test to reach operating state. 6. Make a change to the attribute or add it. 7. Check that the event report informing about the attribute change or addition is received. 8. Send a release-request to the agent under test with reason normal (0). 9. Make the agent try to re-associate. 10. The simulated manager responds with a accepted-unkown-config. 11. Check the attribute that has been changed or added is not present when the agent sends the ConfigReport. 		
Pass/Fail criteria		Changes made to the attribute must not be present in the second association.		
Notes		The attribute that is changed in step 6 must be an attribute whose "initial" value is defined in the ConfigReport.		

TP Id		TP/PLT/AG/OXP/SER/BV-004		
TP label		Agent transmits data in a fixed format Event Report		
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	FormatEventRep 1; M	FormatEventRep 4; O	PersonEventRep 2; O
		FormatEventRep 6; M	FormatEventRep 7; M	
Applicability		(C_AG_OXP_182 OR C_AG_OXP_183 OR C_AG_OXP_184) AND C_AG_OXP_000		
Initial condition		The simulated manager and the agent under test are in the configuring state.		
Test procedure		<ol style="list-style-type: none"> 1. The agent under test sends its configuration to the simulated manager. Save the number of measurement objects (config-obj-list.count=n). 2. Every measurement object has an obj-handle and one or more attributes. The objects of interest for this test cases are all the metric derived ones (Obj-class=MDC_MOC_VMO_METRIC_ENUM 0x00 0x05, Obj-class=MDC_MOC_VMO_METRIC_NU 0x00 0x06, or Obj-class=MDC_MOC_VMO_METRIC_SA_RT 0x00 0x09). For each of them, check: <ol style="list-style-type: none"> a. Attribute Attribute-Val-Map <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ATTRIBUTE_VAL_MAP (0x0A 0x55) <input type="checkbox"/> AttrValMap.count = N (number of object for this measurement object) <input type="checkbox"/> AttrValMap.length = L b. For each attribute (of the N present) check the ID and the length at which it will be transmitted <ul style="list-style-type: none"> <input type="checkbox"/> field-type=MDC_ATTR_* 		

	<ul style="list-style-type: none"> <input type="checkbox"/> field-value=0xXX 0xXX, where the length will be declared (K). <ol style="list-style-type: none"> 3. Wait for the agent under test to reach the operating state and take some measurements. 4. When the agent under test sends an event report to the simulated manager with measurement observations, check the following: <ul style="list-style-type: none"> <input type="checkbox"/> event-type= MDC_NOTI_SCAN_REPORT_FIXED (0X0D 0X1D) <input type="checkbox"/> ObservationScanFixed.count= P (where P<=N, and it is the number of objects reported in this event report) <input type="checkbox"/> obj-handle = It has to be the same that obj-handle of the Measurement object sent in the agent's configuration <input type="checkbox"/> obs-val-data.length= 0xXX 0xXX, where the value is the length "K". <input type="checkbox"/> The actual observed measurement value will come in the next field, but this value is of no interest for this test case <p>IF the Absolute-Time attribute is present in the Attribute-Val-Map THEN</p> <ol style="list-style-type: none"> 5. Record the value of the received Time Stamp 6. The simulated manager issues a "roiv-cmip-get" command with the handle set to 0 (to request an MDS object) and an empty attribute-id-list to indicate all attributes. 7. The agent responds with with a "rors-cmip-get" service message in which the attribute-list contains a list of all implemented attributes of the MDS object, record the Date-and-Time value.
Pass/Fail criteria	<ul style="list-style-type: none"> • The Fixed Event report contains the same list of attributes (and sizes) in the same order that was declared in the configuration message • The Time Stamp and Date-and-Time values are coherent • The Attr-Val-Map for an object has to be received prior to the measurement (Config Report or MDS Event Report)
Notes	

TP Id	TP/PLT/AG/OXP/SER/BV-004_A		
TP label	Agent transmits data in variable format Event Report		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	FormatEventRep 4; O	PersonEventRep 2; O
Applicability	C_AG_OXP_189 AND C_AG_OXP_000		
Initial condition	The simulated manager and the agent under test are in the unassociated state.		
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager, record the attribute Attribute-Value-Map. 4. Once in the operating state take a measurement with the agent under test. 5. Wait for the event report from the agent under test. 6. When the agent under test sends an event report to the simulated manager with measurement observations, check the following: <ol style="list-style-type: none"> a. If the data is from a single person <ul style="list-style-type: none"> <input type="checkbox"/> Event-type = MDC_NOTI_SCAN_REPORT_VAR <input type="checkbox"/> Event-info parameter = ScanReportInfoVar <input type="checkbox"/> attribute.identification-field = <Variable> <input type="checkbox"/> Value-length = <Variable> <input type="checkbox"/> Value = <Not relevant for this Test> 		

	<ul style="list-style-type: none"> b. If the data is from multiple persons <ul style="list-style-type: none"> <input type="checkbox"/> Event-type = MDC_NOTI_SCAN_REPORT_MP_VAR <input type="checkbox"/> Event-info parameter = ScanReportInfoMPVar <input type="checkbox"/> attribute.identification-field = <Variable> <input type="checkbox"/> Value-length = <Variable> <input type="checkbox"/> Value = <Not relevant for this Test>
Pass/Fail criteria	<ul style="list-style-type: none"> • Variable format event report contains the listed fields.
Notes	

TP Id	TP/PLT/AG/OXP/SER/BV-005		
TP label	Agent transmits data in a grouped format Event Report (Scanner Objects)		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	FormatEventRep 2; M	PersonEventRep 2; O
		ConfNormalProc 4; M	FormatEventRep 8; M
Applicability	(C_AG_OXP_046 OR C_AG_OXP_047) AND C_AG_OXP_048 AND C_AG_OXP_000		
Initial condition	The simulated manager and the agent under test are in the configuring state.		
Test procedure	<ol style="list-style-type: none"> 1. When the agent under test sends its configuration to the simulated manager the number of the measurement object (config-obj-list.count=n) must be saved. 2. Every measurement object has an obj-handle and one or more attributes: <ol style="list-style-type: none"> a. Attribute Scan-Handle-Value-Map <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SCAN_HANDLE_ATTR_VAL_MAP (0x0A 0x53) <input type="checkbox"/> HandleAttrValMap.count = N (number of object for this measurement object) <input type="checkbox"/> HandleAttrValMap.length = L b. For each attribute (of the L present) its length is needed: <ul style="list-style-type: none"> <input type="checkbox"/> HandleAttrValMap = <One of an actual metric derived object> <input type="checkbox"/> HandleAttrValMap.count = K (number attributes of this object) <input type="checkbox"/> HandleAttrValMap.length = M 3. The sum of the lengths values is the total length of the measurement data for this object, in this case, the sum of all the M's 4. When the agent under test sends an event report to the simulated manager with a measurement observation, the format of this message is: <ol style="list-style-type: none"> a. PrstA pdu (0xE7 0x00) <ul style="list-style-type: none"> <input type="checkbox"/> obj-handle = <It has to be the same that obj-handle of the Measurement object sent in the agent's configuration> <input type="checkbox"/> event-type = MDC_NOTI_BUF_SCAN_REPORT_GROUPED or MDC_NOTI_UNBUF_SCAN_REPORT_GROUPED <input type="checkbox"/> grouped-length = L This length has to be the same as the total length of the measurement data for this object, it is the sum calculated when the agent sent its configuration 		
Pass/Fail criteria	<ul style="list-style-type: none"> • The metric derived objects must appear in the event report in the same order as were declared on the configuration report. The length of the event report must match the length indicated by the Handle-Attribute-Value-Map and can not exceed the maximum APDU size established by the specialization: <ul style="list-style-type: none"> ○ Pulse oximeter -> 9216 octets ○ Weighing scales -> 896 octets ○ Glucose meter -> 5120 octets or 64512 octets if the agent supports PM-Store ○ Blood pressure -> 896 octets 		

	<ul style="list-style-type: none"> ○ Thermometer -> 896 octets ○ Independent activity hub -> 5120 octets ○ Cardiovascular -> 64512 octets or 6624 octets if it supports Step Counter Profile ○ Strength -> 64512 octets ○ Adherence monitor -> 1024 octets ○ Peak Flow -> 2030 octets ○ Body composition analyser -> 7730 octets ○ Basic ECG/Simple ECG -> 7168 octets or 64512 octets if the agent supports PM-Store ○ Basic ECG/Heart rate -> 1280 octets or 64512 octets if the agent supports PM-Store ○ International normalized ratio -> 896 octets or 64512 if the agent supports PM-Store
Notes	

TP Id	TP/PLT/AG/OSP/SER/BV-007		
TP Label	Temporarily Stored Measurements		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable Items	TempStored 1; O	TempStored 2; C
		TempStored 7; R	TempStored 8; M
Applicability	C_AG_OXP_032 AND C_AG_OXP_000		
Initial Condition	The simulated manager and the agent under test are in the unassociated state.		
Test Procedure	<ol style="list-style-type: none"> 1. The agent under test takes more than 25 measurements before connecting to the simulated manager. 2. The agent under test gets connected to the simulated manager. The Metric-Spec-Small attribute sent in ConfigReport is recorded for numeric and enumeration objects. 3. Once in the operating state we check that: <ol style="list-style-type: none"> a. No more than 25 Temporary Stored Measurements are sent in the same event report. b. That the event reports used to transmit the measurements should be confirmed c. That every Temporary Stored Measurement sent has a Time Stamp attribute (Date-and-Time, Relative-Time or HIRes-Relative-Time) d. Data sent is Enumerated or Numeric. e. Metric-Spec-Small – mss-avail-stored-data bit is set. 		
Pass/Fail Criteria	<ul style="list-style-type: none"> • The conditions in step 3 are met • The total size of the response can not exceed the sum of the APDU sizes of the supported specializations (limited to an absolute limit of 64512 octets): <ul style="list-style-type: none"> ○ Pulse oximeter -> 9216 octets ○ Weighing scales -> 896 octets ○ Glucose meter -> 5120 octets or 64512 octets if the agent supports PM-Store ○ Blood pressure -> 896 octets ○ Thermometer -> 896 octets ○ Independent activity hub -> 5120 octets ○ Cardiovascular -> 64512 octets or 6624 octets if the agent supports Step Counter Profile ○ Strength -> 64512 octets 		

	<ul style="list-style-type: none"> ○ Adherence monitor -> 1024 octets ○ Peak flow -> 2030 octets ○ Body composition analyser -> 7730 octets ○ Basic ECG/Simple ECG -> 7168 octets or 64512 octets if the agent supports PM-Store ○ Basic ECG/Heart rate -> 1280 octets or 64512 octets if the agent supports PM-Store ○ International normalized ratio -> 896 octets or 64512 if the agent supports PM-Store
Notes	<p>It is possible that the agent supports Temporarily Stored Measurements (TSM) for a set of objects and that it does not support TSM for other set of objects.</p> <p>The test tool identifies the objects that support TSM using the Metric-Spec-Small attribute – mss-avail-stored-data(1) bit. When this bit is set to 1 in one object, the test tool assumes that this object stores TSM, although it is not totally sure because this bit is "informational".</p> <p>From [ISO/IEEE 11073-20601A] clause A.11.3:</p> <ul style="list-style-type: none"> • IF object stores TSM THEN mss-avail-stored-data(1) SHALL be set to 1 • IF object does not store TSM THEN mss-avail-stored-data(1) MAY be set to 1 or 0 <p>If agent under test sets mss-avail-stored-data(1) bit to 1 for one object but this object does not store TSM and the agent does not include the time stamp in event report, the test tool gives a FAIL verdict since the test tool identifies that this object stores TSM. When the vendor implements this behaviour in a device, a waiver may be required to complete the Certification. See Bugzilla #840 and contact Continua TOM for further details.</p>

A.4 Subgroup 1.2.3 – PHD communication model (COM)

TP Id		TP/PLT/AG/OXP/COM/BV-003_A		
TP label		Communication Characteristics: Reliable virtual channel 1		
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	CommuCharac 2; M	CommuCharac 3; M	
Applicability		C_AG_OXP_000		
Initial condition		The simulated manager and the agent under test are in the unassociated state.		
Test procedure		<ol style="list-style-type: none"> 1. The agent under test sends an Association Request to the simulated manager. 2. The simulated manager sends an Association Response with result = accepted-unknown-config. 3. The agent under test responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 4. Wait until the agent under test reaches the operating state. 5. The simulated manager issues a "roiv-cmip-get" command with the handle set to 0 (to request an MDS object) and an empty attribute-id-list to indicate all attributes. 6. The agent responds with with a "rors-cmip-get". 7. IF C_AG_OXP_041 THEN <ol style="list-style-type: none"> a. The simulated manager sends a Get request for the PM-Store with an attribute-id-list set to 0 to indicate all PM-Store attributes. b. The agent under test issues a GET response. c. The simulated manager shall send a Get-Segment-Info object action for the PM-Segment object with SegmSelection = all-segments to indicate the PM-Segments attributes of all available PM-Segments. d. The agent under test issues a "rors-cmip-confirmed-action" response. 8. IF (C_AG_OXP_046 OR C_AG_OXP_047) THEN <ol style="list-style-type: none"> a. The simulated manager sends a "roiv-cmip-confirmed-set" to set the OperationalState of the scanner object to 1. b. The agent under test responds with a "rors-cmip-set". 9. The simulated manager sends a Release Request to the agent under test with reason = normal(0). 10. The agent under test responds with a Release Response. 		
Pass/Fail criteria		The "reliable" virtual channel must be used in steps 1, 3, 6, 7.b, 7.d, 8.b and 10.		
Notes				

TP Id		TP/PLT/AG/OXP/COM/BV-003_B		
TP label		Communication Characteristics: Reliable virtual channel 2		
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	CommuCharac 2; M		
Applicability		C_AG_OXP_000		
Initial condition		The simulated manager and the agent under test are in the operating state.		
Test procedure		<ol style="list-style-type: none"> 1. The agent under test sends an Association Request from the agent under test. 2. The simulated manager sends an Association Response with result = accepted-unknown-config. 3. The agent under test responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the simulated manager. 		

	<ol style="list-style-type: none"> 4. The simulated manager responds with a "unsupported-config", waits for a new configuration and keeps responding with "unsupported-config" to every new configuration. 5. The agent under test sends a Release Request.
Pass/Fail criteria	The "reliable" virtual channel must be used for the Release Request.
Notes	

TP Id	TP/PLT/AG/OXP/COM/BV-003_C		
TP label	Communication Characteristics: Reliable virtual channel 3		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	CommuCharac 2; M	
Applicability	C_AG_OXP_000		
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager sends a Release Response to the agent under test. 2. The agent responds with an Abort message. 		
Pass/Fail criteria	The "reliable" virtual channel must be used for the Abort message.		
Notes			

TP Id	TP/PLT/AG/OXP/COM/BV-003_D		
TP label	Communication Characteristics: Reliable virtual channel 4		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	CommuCharac 4; M	
Applicability	C_AG_OXP_000		
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager sends a roiv-cmip-confirmed-action with action-type = MDC_ACT_DATA_REQUEST. 2. The agent under test shall reply with a "roer" with reason = no-such-action (9). 		
Pass/Fail criteria	The "reliable" virtual channel must be used for the "roer" message.		
Notes			

TP Id	TP/PLT/AG/OXP/COM/BV-006_A		
TP label	Agent State machine. Accepted known configuration		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	AgentStateMach 21; M	
Applicability	C_AG_OXP_000		
Initial condition	Accepted known configuration of the agent state machine.		
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test (the agent passes to the associating state). 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent under test responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the simulated manager. The configuration report is reported. 4. The simulated manager sends an abort message. 		

	<ol style="list-style-type: none"> 5. The simulated manager and the Agent move to the unassociated state. 6. The simulated manager receives an association request from the agent under test. 7. The simulated manager responds with a result = accepted. 8. The agent must change to the operating state. <ul style="list-style-type: none"> • IF the agent supports the Scanner object: The simulated manager sends a Set command for the Scanner object and the agent shall reply. • ELSE IF the agent under test supports PM-Store, the simulated manager sends a Get Segment Info action and the agent shall reply. • ELSE the simulated manager waits to receive measurements from the agent.
Pass/Fail criteria	The agent under test has passed to the operating state after the last step.
Notes	

TP Id	TP/PLT/AG/OXP/COM/BV-006_B		
TP label	Agent State machine		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	AgentStateMach 64; M	ObjAccessServ 4; C
Applicability	C_AG_OXP_000		
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager sends a "roiv-cmip-get" to the agent, to get all the attributes for an MDS object. 2. The agent responds with a "rors-cmip-get" message. 3. IF C_AG_OXP_180 THEN the simulated manager sends a "roiv-cmip-confirmed-set", setting to default value the attribute Operational State for a scanner object. IF NOT C_AG_OXP_180 THEN the simulated manager sends a roiv-cmip-set, setting to default value the attribute Operational State for a scanner object. 4. IF C_AG_OXP_180 THEN the agent responds with a rors-cmip-confirmed-set if it supports a scanner object, otherwise it responds with a roer-* or rorj-*. IF NOT C_AG_OXP_180 the agent does not respond if it supports a scanner object, otherwise it responds with a roer-* or rorj-*. 5. The simulated manager sends a "roiv-cmip-confirmed-action", action-type MDC_ACT_DATA_REQUEST. 6. The agent responds with a "rors-cmip-confirmed-action", "roer-*" or "rorj-*". 7. The simulated manager sends a "roiv-cmip-confirmed-action", action-type MDC_ACT_SET_TIME. 8. The agent responds with a "rors-cmip-confirmed-action", "roer-*" or "rorj-*". 9. The simulated manager sends a "roiv-cmip-confirmed-action", action-type MDC_ACT_SEG_GET_INFO. 10. The agent responds with a "rors-cmip-confirmed-action" if the PM-Store object is supported by the agent, otherwise it responds with a "roer-*" or "rorj-*". 11. The simulated manager sends a "roiv-cmip-confirmed-action", action-type MDC_ACT_SEG_TRIG_XFER. 12. The agent responds with a "rors-cmip-confirmed-action" if the PM-Store object is supported by the agent, otherwise, a "roer-*" or "rorj-*". 13. The simulated manager sends a "roiv-cmip-confirmed-action", action-type MDC_ACT_SEG_CLR (all-segments). If the agent supports Clear-Segment action, THEN the test tool performs a GET request to read the Clear-Timeout attribute for Pm-Store. 14. The agent responds with a "rors-cmip-confirmed-action" if the PM-Store object is supported by the agent and it supports Clear-Segment action, otherwise, a "roer-*" or "rorj-*". 		

Pass/Fail criteria	<ul style="list-style-type: none"> The agent replies with messages specified in steps 2, 4, 6, 8, 10, 12 and 14 of the test procedure. If the agent sends a roer message, check that the error value is correct and that a parameter may be included.
Notes	

TP Id	TP/PLT/AG/OXP/COM/BV-007		
TP label	Agent State machine. Accepted unknown configuration		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	AgentStateMach 22; M	AgentStateMach 53; M
		ConfProc 1; M	ConfExitCond 3; M
Applicability	C_AG_OXP_000		
Initial condition	The simulated manager and the agent under test are in the unassociated state		
Test procedure	<ol style="list-style-type: none"> The simulated manager receives an association request from the agent under test (the agent passes to the associating state). The simulated manager responds with a result = accepted-unknown-config. The agent under test shall go to the "Sending Config" substate, and responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager (the agent shall go to the "waiting-approval" substate). The simulated manager sends a "Remote Operation Response Confirmed Event Report" with result "accepted-config". The agent under test shall move to the operating state. <ul style="list-style-type: none"> IF the agent under test supports the Scanner object: The simulated manager sends a Set command for the Scanner object and the agent under test shall reply. ELSE IF the agent under test supports PM-Store the simulated manager sends a Get Segment Info action and the Agent shall reply. ELSE the simulated manager waits for receiving measurements from the Agent under test. 		
Pass/Fail criteria	The agent under test has passed to the operating state after the last step.		
Notes			

TP Id	TP/PLT/AG/OXP/COM/BV-009		
TP label	Agent State machine. Leaving the Operating State 1		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	AgentStateMach 59; M	
Applicability	C_AG_OXP_000		
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<ol style="list-style-type: none"> The simulated manager sends an Association Request to the agent under test. The agent under test responds with an Abort message (abrt) with reason undefined. 		
Pass/Fail criteria	<ul style="list-style-type: none"> The agent transmits correctly the Abort message (abrt) with reason undefined and changes to the unassociated state. The simulated manager must not receive any message other than an Association Request after step 2. 		
Notes			

TP Id		TP/PLT/AG/OXP/COM/BV-010		
TP label		Agent State machine. Leaving the Operating State 2		
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	AgentStateMach 60; M		
Applicability		C_AG_OXP_000		
Initial condition		The simulated manager and the agent under test are in the operating state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated manager sends an Association Response. 2. The agent responds with an Abort message (abrt) with reason undefined. 		
Pass/Fail criteria		<ul style="list-style-type: none"> • The agent transmits correctly the Abort message (abrt) with reason undefined and changes to the unassociated state. • The simulated manager must not receive any message other than an Association Request after step 2. 		
Notes				

TP Id		TP/PLT/AG/OXP/COM/BV-011		
TP label		Agent State machine. Leaving the Operating State 3		
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	AgentStateMach 61; M		
Applicability		C_AG_OXP_000		
Initial condition		The simulated manager and the agent under test are in the operating state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated manager sends an Association Release Request (RIrq) message to the agent under test, with reason =0 (normal). 2. The agent under test shall respond with an Association Release Response (RIre) message with reason =0 (normal) and shall go to the unassociated state. 		
Pass/Fail criteria		The agent under test transmits correctly the RIre message.		
Notes				

TP Id		TP/PLT/AG/OXP/COM/BV-012		
TP label		Agent State machine. Leaving the Operating State 4		
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	AgentStateMach 62; M		
Applicability		C_AG_OXP_000		
Initial condition		The simulated manager and the agent under test are in the operating state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated manager sends an Association Release Response to the agent under test, with reason =0 (normal) 2. The agent under test responds with an Abort message (abrt) with reason undefined. 		
Pass/Fail criteria		<ul style="list-style-type: none"> • The agent under test transmits correctly the Abort message (abrt) reason undefined and changes to the unassociated state. • The simulated manager must not receive any message other than an Association Request after step 2. 		
Notes				

TP Id		TP/PLT/AG/OXP/COM/BV-013		
TP label		Agent State machine. Association timeout		
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	AgentStateMach 2; M	AgentStateMach 16; M	AgentStateMach 17; M
		AssocErrorCond 1; M	AssocErrorCond 2; M	AssocErrorCond 3; M
Applicability		C_AG_OXP_000		
Initial condition		The simulated manager and the agent under test are in the unassociated state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test (and agent under test passes to the associating state). 2. The simulated manager does NOT respond with any message. 3. The agent under test shall wait for the TO_{assoc} timer to expire (10 seconds) and retransmit a new association request. 4. Steps 2-3 shall be repeated until the Retry Count has reached (=3). 5. As the manager has not answered to any of the 4 messages, the agent under test shall send an abort message (abrt) with reason response-timeout to the manager and shall pass to the unassociated state. 		
Pass/Fail criteria		The TO _{assoc} timer and the RC values are properly implemented and in the last step the agent under test shall transmit an Abort message (abrt) with reason response-timeout.		
Notes				

TP Id		TP/PLT/AG/OXP/COM/BV-020		
TP label		Agent State machine. Connected Associating 1		
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	AgentStateMach 20; M		
Applicability		C_AG_OXP_000		
Initial condition		The agent under test is in the connected associating state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated manager issues an Association Request 2. The agent under test sends an Association Response message to the manager: <ul style="list-style-type: none"> • reason = rejected-permanent(1), • data-proto-id=data-proto-id-empty • data-proto-info=omit 		
Pass/Fail criteria		<ul style="list-style-type: none"> • The agent sends the detailed AARE message and changes to the unassociated state • The simulated manager must not receive any message other than an Association Request after step 2 		
Notes				

TP Id		TP/PLT/AG/OXP/COM/BV-021		
TP label		Agent State machine. Connected Associating 2		
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	AgentStateMach 24; M		
Applicability		C_AG_OXP_000		
Initial condition		The agent under test is in the connected associating state.		

Test procedure	<ol style="list-style-type: none"> 1. The simulated manager issues a Release Request. 2. The agent under test sends an abort message (abrt) reason undefined to the manager and shall pass to the unassociated state.
Pass/Fail criteria	<ul style="list-style-type: none"> • The agent under test sends the Abort message (abrt) with reason undefined and changes to the unassociated state • The simulated manager must not receive any message other than an Association Request after step 2
Notes	

TP Id	TP/PLT/AG/OXP/COM/BV-022		
TP label	Agent State machine. Connected Associating 2		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	AgentStateMach 25; M	
Applicability	C_AG_OXP_000		
Initial condition	The agent under test is in the connected associating state.		
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager issues a Release Response with reason = normal(0). 2. The agent under test sends an abort message (Abort-reason undefined) to the manager and shall pass to the unassociated state. 		
Pass/Fail criteria	<ul style="list-style-type: none"> • The agent sends the Abort message (Abort-reason undefined) and changes to the unassociated state • The simulated manager must not receive any message other than an Association Request after step 2 		
Notes			

TP Id	TP/PLT/AG/OXP/COM/BV-023		
TP label	Agent State machine. Connected Associating 4		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	AgentStateMach 27; M	
Applicability	C_AG_OXP_000		
Initial condition	The agent under test is in the connected associating state.		
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager issues a "roiv-cmip-get" command with the handle set to 0 (to request the MDS object) and the attribute-idlist set to "all-attributes". 2. The agent under test sends an abort message (abrt) reason undefined to the manager and shall pass to the unassociated state. 		
Pass/Fail criteria	<ul style="list-style-type: none"> • The agent sends the Abort message (abrt) with reason undefined and changes to the unassociated state • The simulated manager must not receive any message other than an Association Request after step 2 		
Notes			

TP Id	TP/PLT/AG/OXP/COM/BV-023_A		
TP label	Agent State machine. Connected Associating 5		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	AgentStateMach 27; M	

Applicability	C_AG_OXP_000
Initial condition	The agent under test is in the connected associating state.
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager sends a badly formatted message. 2. The agent under test sends an abort message (abrt) with reason undefined to the manager and shall pass to the unassociated state.
Pass/Fail criteria	<ul style="list-style-type: none"> • The agent sends the Abort message (abrt) with reason undefined and changes to the unassociated state • The simulated manager must not receive any message other than an Association Request after step 2
Notes	

TP Id	TP/PLT/AG/OXP/COM/BV-030		
TP label	Agent State machine. Connected Associated Configuring Waiting Approval 1		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	AgentStateMach 45; M	
Applicability	C_AG_OXP_000		
Initial condition	The agent under test is in the waiting approval state.		
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager issues an Association Request. 2. The agent under test sends an abort message (abrt) with reason undefined to the manager and shall pass to the unassociated state. 		
Pass/Fail criteria	<ul style="list-style-type: none"> • The agent sends the Abort message (abrt) with reason undefined and changes to the unassociated state • The simulated manager must not receive any message other than an Association Request after step 2 		
Notes			

TP Id	TP/PLT/AG/OXP/COM/BV-031		
TP label	Agent State machine. Connected Associated Configuring Waiting Approval 2		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	AgentStateMach 46; M	
Applicability	C_AG_OXP_000		
Initial condition	The agent under test is in the waiting approval state.		
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager issues an Association Response with reason = accepted(0). 2. The agent under test sends an abort message (abrt) with reason undefined to the manager and shall pass to the unassociated state. 		
Pass/Fail criteria	<ul style="list-style-type: none"> • The agent sends the Abort message (abrt) with reason undefined and changes to the unassociated state • The simulated manager must not receive any message other than an Association Request after step 2 		
Notes			

TP Id		TP/PLT/AG/OXP/COM/BV-032		
TP label		Agent State machine. Connected Associated Configuring Waiting Approval 3		
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	AgentStateMach 47; M		
Applicability		C_AG_OXP_000		
Initial condition		The agent under test is in the waiting approval state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated manager issues a Release Request. 2. The agent under test sends a Release Response to the manager and shall pass to the unassociated state. 		
Pass/Fail criteria		<ul style="list-style-type: none"> • The agent under test sends the Release Response message and changes to the unassociated state • The simulated manager must not receive any message other than an Association Request after step 2 		
Notes				

TP Id		TP/PLT/AG/OXP/COM/BV-033		
TP label		Agent State machine. Connected Associated Configuring Waiting Approval 4		
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	AgentStateMach 48; M		
Applicability		C_AG_OXP_000		
Initial condition		The agent under test is in the waiting approval state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated manager issues a Release Response with reason = normal(0). 2. The agent under test sends an abort message (abrt) with reason undefined to the manager and shall pass to the unassociated state. 		
Pass/Fail criteria		<ul style="list-style-type: none"> • The agent under test sends the Abort message (abrt) with reason undefined and changes to the unassociated state • Simulated manager must not receive any message other than an Association Request after step 2 		
Notes				

TP Id		TP/PLT/AG/OXP/COM/BV-034		
TP label		Agent State machine. Connected Associated Configuring Waiting Approval 5		
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	AgentStateMach 50; M	OperNormProc 5; R	
Applicability		C_AG_OXP_000		
Initial condition		The agent under test is in the connected associated configuring waiting approval state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated manager issues a "roiv-cmip-get" command with the handle set to 0 (to request the MDS object) and the attribute-id list set to MDC_ATTR_SYS_ID. 2. The agent under test responds with a "rors-cmip-get" message or a roer message (not-allowed-by-object). 3. The agent under test remains in Waiting Approval state. 4. The simulated manager responds with a rors-cmip-confirmed-event-report with result "unsupported-config". 		

	5. The agent responds with a roiv-cmip-confirmed-event report message with an MDC_NOTI_CONFIG event to send its configuration to the manager or Rlrq (no-more-configurations).
Pass/Fail criteria	The process detailed above must be successfully completed.
Notes	

TP Id	TP/PLT/AG/OXP/COM/BV-034_A		
TP label	Agent State machine. Get Request Sending Configuring		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	AgentStateMach 37; M	
Applicability	C_AG_OXP_000		
Initial condition	The simulated manager and the agent under test are in the unassociated state.		
Test procedure	<ol style="list-style-type: none"> 1. The agent under test sends an Association Request to the simulated manager. 2. The simulated manager responses with an accepted-unknown-config. 3. The simulated manager issues "roiv-cmip-get" command with the handle set to 0 (to request the MDS object) and the attribute-idlist set to "all-attributes". 4. The agent under test responds with a "rors-cmip-get" message in which the attribute=list contains a list of all implemented attributes of the MDS object. 5. The agent under test must send its configuration. 		
Pass/Fail criteria	The process detailed above must be completed.		
Notes			

TP Id	TP/PLT/AG/OXP/COM/BV-035		
TP label	Agent State machine. Connected Associated Configuring Waiting Approval 6		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	AgentStateMach 51; M	
Applicability	C_AG_OXP_000		
Initial condition	The agent under test is in the connected associated configuring waiting approval state.		
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager issues a GET with handle = 1. 2. The agent under test must send a "roer" with reason = no-such-object-instance(1). 3. The agent under test remains in the Waiting Approval state. 4. The simulated manager responds with a rors-cmip-confirmed-event-report with result "unsupported-config". 5. The agent responds with a roiv-cmip-confirmed-event report message with an MDC_NOTI_CONFIG event to send its configuration to the manager or Rlrq (no-more-configurations). 		
Pass/Fail criteria	The process detailed above must be successfully completed.		
Notes			

TP Id	TP/PLT/AG/OXP/COM/BV-036_A		
TP label	Agent State machine. Connected Associated Configuring Waiting Approval 7		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	AgentStateMach 54; M	
Applicability	C_AG_OXP_000		

Initial condition	The agent under test is in the connected associated configuring waiting approval state.
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager issues a Prst message, rors-cmip- get. 2. The agent under test sends an abort message (abrt) with reason undefined to the manager and shall pass to the unassociated state.
Pass/Fail criteria	<ul style="list-style-type: none"> • The agent under test sends the Abort message (abrt) with reason undefined and changes to the unassociated state • The simulated manager must not receive any message other than an Association Request after step 2
Notes	

TP Id	TP/PLT/AG/OSP/COM/BV-036_B		
TP label	Agent State machine. Connected Associated Configuring Waiting Approval 8		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	AgentStateMach 54; M	
Applicability	C_AG_OXP_000		
Initial condition	The agent under test is in the connected associated configuring waiting approval state.		
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager issues a roer message. 2. The agent under test sends an abort message (Abort-reason undefined) to the manager and shall pass to the unassociated state. 		
Pass/Fail criteria	<ul style="list-style-type: none"> • The agent under test sends the abort message (abrt) with reason undefined and changes to the unassociated state • The simulated manager must not receive any message other than an Association Request after step 2 		
Notes			

TP Id	TP/PLT/AG/OSP/COM/BV-036_C		
TP label	Agent State machine. Connected Associated Configuring Waiting Approval 9		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	AgentStateMach 54; M	
Applicability	C_AG_OXP_000		
Initial condition	The agent under test is in the connected associated configuring waiting approval state.		
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager issues a rorj message. 2. The agent under test sends an abort message (abrt) with reason undefined to the manager and shall pass to the unassociated state. 		
Pass/Fail criteria	<ul style="list-style-type: none"> • The agent under test sends the Abort message (abrt) with reason undefined and changes to unassociated state • The simulated manager must not receive any message other than an Association Request after step 2 		
Notes			

TP Id	TP/PLT/AG/OSP/COM/BV-037		
TP label	Agent State machine. Connected Disassociation 1		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	AgentStateMach 70; M	

Applicability	C_AG_OXP_000
Initial condition	The agent is in the unassociated state.
Test procedure	<ol style="list-style-type: none"> 1. The agent under test sends an Association Request to the simulated manager. 2. The simulated manager responds with an accepted-unknown-config. 3. The agent sends a configuration event report. 4. The simulated manager responds with a unsupported-configuration. 5. The agent sends a new configuration event report with a new configuration (if it has more). 6. Repeat the last two steps recording all the ConfigId-values until the agent sends a ReleaseRequest with reason "no-more-configurations". The agent moves to the disassociating state. 7. The simulated manager sends an AARQ message. 8. The agent responds with an Abort message (abrt) with reason undefined. 9. The agent and the manager move to the unassociated state.
Pass/Fail criteria	<ul style="list-style-type: none"> • The agent sends the Abort message (abrt) with reason undefined and changes to the unassociated state • The simulated manager must not receive any message other than an Association Request after step 9
Notes	

TP Id	TP/PLT/AG/OXP/COM/BV-038		
TP label	Agent State machine. Connected Disassociation 2		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	AgentStateMach 71; M	
Applicability	C_AG_OXP_000		
Initial condition	The agent is in the unassociated state.		
Test procedure	<ol style="list-style-type: none"> 1. The agent under test sends an Association Request to the simulated manager. 2. The simulated manager responds with an accepted-unknown-config. 3. The agent sends a configuration event report. 4. The simulated manager responds with an unsupported-configuration. 5. The agent sends a new configuration event report with a new configuration (if it has more). 6. Repeat the last two steps recording all the ConfigId-values until the agent sends a ReleaseRequest with reason "no-more-configurations". The agent moves to the disassociating state. 7. The simulated manager sends an AARE message. 8. Agent responses with an Abort message (abrt) with reason undefined. 9. The agent and the manager move to the unassociated state. 		
Pass/Fail criteria	<ul style="list-style-type: none"> • The agent sends the Abort message (abrt) with reason undefined and changes to unassociated state • The simulated manager must not receive any message other than an Association Request after step 9 		
Notes			

TP Id		TP/PLT/AG/OXP/COM/BV-039		
TP label		Agent State machine. Connected Disassociation 3		
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	AgentStateMach 72; M		
Applicability		C_AG_OXP_000		
Initial condition		The agent is in the unassociated state.		
Test procedure		<ol style="list-style-type: none"> 1. The agent under test sends an Association Request to the simulated manager. 2. The simulated manager responses with an accepted-unknown-config. 3. The agent sends a configuration event report. 4. The simulated manager responds with an unsupported-configuration. 5. The agent sends a new configuration event report with a new configuration (if it has more). 6. Repeat the last two steps recording all the ConfigId-values until the agent sends a ReleaseRequest with reason "no-more-configurations". The agent moves to the disassociating state. 7. The simulated manager sends a Rlrq message (reason=normal). 8. The agent responses with a Rlre message. 9. The agent and manager remain in the same state. 		
Pass/Fail criteria		<ul style="list-style-type: none"> • The agent sends the Rlre message and remain in the same state • The simulated manager must not receive any message other than an Association Request after step 9 		
Notes				

TP Id		TP/PLT/AG/OXP/COM/BV-040_A		
TP label		Agent State machine. Connected Disassociation 4		
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	AgentStateMach 76; M		
Applicability		C_AG_OXP_000		
Initial condition		The agent is in the unassociated state.		
Test procedure		<ol style="list-style-type: none"> 1. The agent under test sends an Association Request to the simulated manager. 2. The simulated manager responses with an accepted-unknown-config. 3. The agent sends a configuration event report. 4. The simulated manager responds with an unsupported-configuration. 5. The agent sends a new configuration event report with a new configuration (if it has more). 6. Repeat the last two steps recording all the ConfigId-values until the agent sends a ReleaseRequest with reason "no-more-configurations". The agent moves to the disassociating state. 7. The simulated manager sends a Roer message. 8. The agent responds with an Abort message (reason undefined). 9. The agent and the manager move to the unassociated state. 		
Pass/Fail criteria		<ul style="list-style-type: none"> • The agent sends the Abort (reason undefined) message and changes to the unassociated state • The simulated manager must not receive any message other than an Association Request after step 9 		

Notes	
--------------	--

TP Id	TP/PLT/AG/OXP/COM/BV-040_B		
TP label	Agent State machine. Connected Disassociation 5		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	AgentStateMach 76; M	
Applicability	C_AG_OXP_000		
Initial condition	The agent is in the disassociating state.		
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager sends an Rorj message. 2. The agent responses with an Abort message (reason undefined). 3. The agent and the manager move to the unassociated state. 		
Pass/Fail criteria	<ul style="list-style-type: none"> • The agent sends the Abort message (reason undefined) and changes to the unassociated state • The simulated manager must not receive any message other than an Association Request after step 9 		
Notes			

TP Id	TP/PLT/AG/OXP/COM/BV-042		
TP label	Association request format		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	AssocRequest 3; C	AssocRequest 4; C
		AssocRequest 6; M	MessageEncod 1;M
	Spec	[ITU-T H.810]	
Testable items	General 3; M	Dev.Information 9; M	
Applicability	C_AG_OXP_000		
Initial condition	The simulated manager and the agent under test are in the unassociated state.		
Test procedure	<ol style="list-style-type: none"> 1. The agent under test sends an AARQ message to the simulated manager. The expected fields sent by the agent are: <ol style="list-style-type: none"> a. APDU Type <ul style="list-style-type: none"> <input type="checkbox"/> field-length =2 bytes <input type="checkbox"/> field-value =0xE2 0x00 (AareApdu) b. The following two bytes indicate the length of the message. c. assoc-version <ul style="list-style-type: none"> <input type="checkbox"/> field-type = AssociationVersion <input type="checkbox"/> field-length =BITS-32 <input type="checkbox"/> field-value = <ul style="list-style-type: none"> ▪ Only one bit can be set d. The following four bytes indicate: <ul style="list-style-type: none"> <input type="checkbox"/> data-proto-list.count (two bytes) = At most there is a data-proto-id set to data-proto-id-20601 <input type="checkbox"/> Length of the message (two bytes) e. data-proto-id <ul style="list-style-type: none"> <input type="checkbox"/> field-type = DataProtold 		

- field-length =INT-U16
 - field-value = 0x50 0x79 (20601)
 - data-proto-id=20601 indicates exchange protocol follows this standard, and data-proto-info shall contain PhdAssociationInformation.
- f. The DataProto.Info field must contain two bytes and indicates the data-proto-info.length
- g. protocol-version
- field-type = Protocol Version
 - field-length =BITS-32
 - IF the agent supports basic electrocardiograph (ECG) or international normalized ratio (INR) device specialization (C_AG_OXP_165 = TRUE OR C_AG_OXP_164 = TRUE OR C_AG_OXP_163 = TRUE) THEN
 - field-value = At least bit protocol-version2(1) is set to 1 (0x40 0x00 0x00 0x00 OR 0xC0 0x00 0x00 0x00)
 - This value shows that version 2 of the data exchange protocol is supported (protocol-version2(1)=1).
 - ELSE
 - field-value = 0x80 0x00 0x00 0x00
 - This value shows that version 1 of the data exchange protocol is supported (protocol-version1(0)=1).
- h. encoding rules
- field-type = EncodingRules
 - field-length = BITS-16
 - field-value= depends on the encoding rules supported/selected.
 - Bit 0 (mder) must always be set
 - and xer(1) or/and per(2) may be set (optional).
- i. nomenclature version
- field-type = NomenclatureVersion
 - field-length =BITS-32
 - field-value = 0x80 0x00 0x00 0x00
 - This value indicates version 1 is supported (nom-version1(0) is set).
- j. functional-units
- field-type = FunctionalUnits
 - field-length = BITS-32
 - field-value =
 - Bit 0 must be 0
 - Bits 1 and 2 may be set
 - The rest of the bits must not be set
- k. system type
- field-type = SystemType
 - field-length = BITS-32
 - field-value = 0x00 0x80 0x00 0x00 (sys-type-agent)
- l. system-id
- field-type = OCTET STRING
 - field-length = 0x00 0x0A
 - field-value = 0x00 0x0X 0xXX 0xXX 0xXX 0xXX 0xXX 0xXX 0xXX 0xXX (octet string length = 8 | EUI-64 manufacturer and device)

	<ul style="list-style-type: none"> <input type="checkbox"/> This value will be System Id attribute of MDS Object. m. dev-config-id <ul style="list-style-type: none"> <input type="checkbox"/> field-type = ConfigId <input type="checkbox"/> field-length = INT-U16 <input type="checkbox"/> field-value = <Not relevant for this Test> n. Data-Req-Mode-Capab: <ul style="list-style-type: none"> <input type="checkbox"/> field-type = DataReqModeCapab <input type="checkbox"/> field-length = INT-U16 <input type="checkbox"/> field-value = SEQUENCE { <ul style="list-style-type: none"> ▪ data-req-mode-flags DataReqModeFlags, ▪ data-req-init-agent-count INT-U8, -- maximum number of parallel Agent initiated ▪ data-req-init-manager-count INT-U8, -- maximum number of parallel manager initiated data requests o. option-list: <ul style="list-style-type: none"> <input type="checkbox"/> field-type: AttributeList
Pass/Fail criteria	The structure and values of the association request message is correct.
Notes	

TP Id	TP/PLT/AG/OXP/COM/BV-049		
TP label	Configuring Procedure 4		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	ConfErrorCond 1; M	ConfErrorCond 2; M
Applicability	C_AG_OXP_000		
Initial condition	The simulated manager and the agent under test are in the unassociated state.		
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test (the agent passes to the associating state). 2. The simulated manager responds with an Association Response with result = "accepted-unknown-config". 3. The agent under test sends a configuration event report. 4. The simulated manager does not respond to the configuration event report for more than TO_{config} time. 		
Pass/Fail criteria	The agent must wait for a TO _{config} . If the time expires, the agent must send an abort to the manager.		
Notes			

TP Id	TP/PLT/AG/OXP/COM/BV-051		
TP label	Non-standard configuration, Dev-config-id		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	ConfNormalProc 20; C	ConfEventRep 21; M
Applicability	C_AG_OXP_181 AND C_AG_OXP_000		
Initial condition	The simulated manager and the agent under test are in the unassociated state.		
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test with a dev-config-id and a system-id. 2. The simulated manager responds with a result = accepted-unknown-config. 		

	<ol style="list-style-type: none"> 3. The agent responds with a roiv-cmip-confirmed-event report message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 4. The simulated manager responds with a rors-cmip-confirmed-event-report with result "accepted" if the Config-Id received in step 3 is the configuration that is being tested, in this case ConfigReport is recorded. Otherwise, the manager responds with a rors-cmip-confirmed-event-report with result "unsupported-config" and step 3 is repeated. 5. Check the config-report-id in the ConfigReport recorded in step 4.
Pass/Fail criteria	The config-report-id value checked in step 5 is in the range between the extended-config-start (16384) and the extended-config-end (32767), inclusive.
Notes	

TP Id	TP/PLT/AG/OXP/COM/BV-052_B		
TP label	Operating procedures. Specific Attributes request		
Coverage	Spec	[ITU-T H.810]	
	Testable items	OperNormProc 4; O	OperNormProc 5; R
Applicability	C_AG_OXP_000		
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager issues a "Remote Operation Invoke Get" command with: <ol style="list-style-type: none"> a. Obj-handle set to 0 (to request an MDS object) b. attribute-id-list.count=1 and a single AVA_Type MDC_ATTR_DEV_CONFIG_ID (0X0A 0X44) to retrieve the mandatory "Dev-Configuration-Id" attribute 2. The agent under test responds with: <ul style="list-style-type: none"> • IF C_AG_OXP_100 THEN: with a "rors-cmip-get" service message which contains the "Dev-Configuration-Id" • ELSE: with a "roer" service message with error-value set to not-allowed-by-object (24) 3. The simulated manager issues a "Remote Operation Invoke Get" command with: <ol style="list-style-type: none"> a. Obj-handle set to 0 (to request an MDS object) b. attribute-id-list empty to request all the attributes of MDS 4. The agent responds with with a "rors-cmip-get" service message which contains all the supported attributes of the MDS. 5. The simulated manager issues a "Remote Operation Invoke Get" command with: <ol style="list-style-type: none"> a. Obj-handle set to 0 b. attribute-id-list set to an attribute NOT supported by the agent 6. The agent responds with a "rors-cmip-get" service message: <ul style="list-style-type: none"> • IF C_AG_OXP_100 THEN: attribute-list must be empty • ELSE: with with a "roer" service message with error-value set to not-allowed-by-object (24) 7. The simulated manager issues a "Remote Operation Invoke Get" command with: <ol style="list-style-type: none"> a. Obj-handle set to 0 b. attribute-id-list contains one supported attribute and one unsupported attribute 8. The agent responds with a "rors-cmip-get" service message: <ul style="list-style-type: none"> • IF C_AG_OXP_100 THEN: attribute-list must contained the supported attribute • ELSE: with with a "roer" service message with error-value set to not-allowed-by-object (24) 		
Pass/Fail criteria	<ul style="list-style-type: none"> • In step 2 the agent properly sends the requested attribute or the error (not-allowed-by-object) 		

	<ul style="list-style-type: none"> In steps 6 and 8 the received attribute list must be empty if NOT C_AG_OXP_100 or the roer if the action is not supported
Notes	

TP Id	TP/PLT/AG/OXP/COM/BV-053		
TP label	Operating procedures. Agent-initiated transmission 1		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	MeasureDataTransf 4; C	MeasureDataTransf 5; C
Applicability	C_AG_OXP_000 AND (C_AG_OXP_182 OR C_AG_OXP_183 OR C_AG_OXP_184 OR C_AG_OXP_189)		
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<ol style="list-style-type: none"> The agent under test must send an Association Request to the simulated manager which contains the DataReqModeFlags field (of the DataReqModeCapab attribute). Check the value of the bit 15 (data-req-supp-init-agent). Check that if the bit is not set, there is at least one Scanner object in the agent under test. Once the device is in the operating state take a measurement and check that, if the bit was set, the agent under test sends the measurement value to the simulated manager without the manager requesting it using a "Remote Operation Invoke Confirmed Event Report" or a "Remote Operation Invoke Event Report" message with one of the MDC_NOTI_SCAN_REPORT_* event-types. Record the scan-report-no for later comparison and check data-req-id. Take another measurement, record the scan-report-no of the event and check data-req-id. 		
Pass/Fail criteria	The agent is able to send agent-initiated measurement reports, uses a correct event-type for doing so and the scan-report-no of the second event has increased once unit Data-req-id is set to data-req-id-agent-initiated (61440).		
Notes			

TP Id	TP/PLT/AG/OXP/COM/BV-053_A		
TP label	Operating procedures. Invoke-id		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	AgentStateMach 50;M	
Applicability	C_AG_OXP_000		
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<ol style="list-style-type: none"> The simulated manager issues a "roiv-cmip-get" command with the handle set to 0, an empty attribute-id-list to indicate all attributes and invoke-id =20. Record the invoke-id of the message sent. The agent responds with with a "rors-cmip-get" service and the invoke id is 20. The simulated manager issues a "roiv-cmip-get" command with the handle set to 0, an empty attribute-id-list to indicate all attributes and invoke-id =15. Record the invoke-id of the message sent. The agent responds with with a "rors-cmip-get" service and the invoke id is 15. The simulated manager issues a "roiv-cmip-get" command with the handle set to 0, an empty attribute-id-list to indicate all attributes and invoke-id =30. Record the invoke-id of the message sent. The agent responds with with a "rors-cmip-get" service and the invoke id is 30. The simulated manager issues a "roiv-cmip-get" command with the handle set to 0, an empty attribute-id-list to indicate all attributes and invoke-id =20. Record the invoke-id of the message sent. 		

	8. The agent responds with with a "rors-cmip-get" service and the invoke id is 20.
Pass/Fail criteria	In steps 2, 4, 6 and 8, the invoke-id has the correct value.
Notes	

TP Id	TP/PLT/AG/OXP/COM/BV-054		
TP label	Agent-initiated transmission.Scan-report-no		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	MeasureDataTransf 46; M	
Applicability	C_AG_OXP_000 AND (C_AG_OXP_182 OR C_AG_OXP_183 OR C_AG_OXP_184 OR C_AG_OXP_189 OR C_AG_OXP_046 OR C_AG_OXP_047)		
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<ol style="list-style-type: none"> 1. If the agent supports the scanner object, the simulated manager sends a Set action to set the Operational-State of the scanner to 1 (enabled). 2. Check that the first MDS-Event-Report, for the Metric object or Scanner object, scan-report-no starts at 0. 		
Pass/Fail criteria	In step 2, the scan-report-no shall be 0.		
Notes			

TP Id	TP/PLT/AG/OXP/COM/BV-056		
TP label	Operating procedures. Agent-initiated transmission. Scanner objects		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	MeasureDataTransf 9; M	MeasureDataTransf 10; C
		ScanClassAttr 3; M	MeasureDataTransf 47; M
Applicability	(C_AG_OXP_046 OR C_AG_OXP_047) AND C_AG_OXP_000		
Initial condition	The simulated manager is in the waiting config state and the agent under test is in the sending config state.		
Test procedure	<ol style="list-style-type: none"> 1. The agent under test must send its configuration to the manager. The scanner object must have the Operational-State set to 0. 2. The simulated manager sends a Set action to set the Operational-State of the scanner to 1 (enabled): <ol style="list-style-type: none"> a. APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Remote Operation Invoke Confirmed Event Report <input type="checkbox"/> roiv-cmip-confirmed-set <input type="checkbox"/> attribute = OperationalState <input type="checkbox"/> value = 1 3. Several measurements are taken with the agent under test. The Data-req-id field will be checked 4. Once the agent under test starts to transmit its data, the manager sets the Operational-State to 0: <ol style="list-style-type: none"> a. APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Remote Operation Invoke Confirmed set <input type="checkbox"/> roiv-cmip-confirmed-set <input type="checkbox"/> attribute = OperationalState <input type="checkbox"/> value = 0 5. The agent must stop sending its data. Record the last scan-report-no. 6. The simulated manager resumes the agent data transmission by setting the 		

	<p>Operational-State back to 1:</p> <p>a. APDU</p> <ul style="list-style-type: none"> <input type="checkbox"/> Type = Remote Operation Invoke Confirmed set <input type="checkbox"/> roiv-cmip-confirmed-set <input type="checkbox"/> attribute = OperationalState <input type="checkbox"/> value = 1 <p>7. Several measurements are taken with the agent under test. Check that the scan-report-no starts counting where it halted before (step 4).</p>
Pass/Fail criteria	<ul style="list-style-type: none"> • In step 3, the agent has to start to transmit data and the data-req-id field is set to data-req-id-agent-initiated • In step 5, the agent has to stop to transmit data • In step 7, the agent has to start again to transmit data and scan-report-no has to start counting where it was halted in step 4
Notes	

TP Id	TP/PLT/AG/EXP/COM/BV-072		
TP label	Operating procedures. PM-Store		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	PersStoreMtrDatTransf 3; M	PersStoreMtrDatTransf 4; M
		PM-StoreMeth 27; O	PM-StoreMeth 26; M
Applicability	C_AG_EXP_041 AND C_AG_EXP_000		
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager shall send a Get request for the PM-Store object with an attribute-id-list set to 0 to indicate all PM-Store attributes. 2. The agent under test issues a GET response with the PM-Store attributes it supports, check the values of the PM-Store-Capab attribute: <ol style="list-style-type: none"> a. PM-Store-Capab: <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_PM_STORE_CAPAB <input type="checkbox"/> attribute-type = PmStoreCapab <input type="checkbox"/> attribute-value = Record the value of bit 6 (Indicates that PM-Segments in the SegmSelection data type can be selected by defining an abs-time-range) and bit 3 (indicates that PM-Segments in the SegmSelection data type can be selected by defining a list of segment identifiers) 3. The simulated manager sends a request for the PM-Segment Data with SegmSelection = 1 to obtain all the segments: <ol style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action, <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_GET_INFO <input type="checkbox"/> SegmSelection = all-segments 4. The agent under test issues a response with the PM-Segments attributes <ol style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action, <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_GET_INFO <input type="checkbox"/> SegmentInfoList <p>IF bit 3 of PmStoreCapab was set:</p>		

5. The simulated manager sends a request for the PM-Segment Data with SegmSelection = segm-id-list which is known because in the previous phase the information of all the segments was retrieved:
 - a. Data APDU
 - Type = Invoke | Confirmed Action,
 - HANDLE = obj-handle
 - Action = MDC_ACT_SEG_GET_INFO
 - SegmSelection = segm-id-list (List of integers with the instance numbers of the selected Segments)
6. The agent under test issues a response with the required PM-Segments attributes:
 - a. Data APDU
 - Type = Invoke | Confirmed Action,
 - HANDLE = obj-handle
 - Action = MDC_ACT_SEG_GET_INFO
 - SegmentInfoList

IF bit 3 of PMStoreCapab was NOT set:
7. The simulated manager sends a Get-Segment-Info:
 - a. Data APDU
 - Type = Invoke | Confirmed Action,
 - HANDLE = obj-handle
 - Action = MDC_ACT_SEG_GET_INFO
 - SegmSelection = segm-id-list (List of integers with the instance numbers of the selected Segments)
8. The agent under test operation response:
 - a. Data APDU
 - Type = Roer
 - ErrorResult = no-such-action (9) or not-allowed-by-object (24)

IF bit 6 of PmStoreCapab was set AND the agent reports absolute-time:
9. The simulated manager sends a Get-Segment-Info:
 - a. Data APDU
 - Type = Invoke | Confirmed Action,
 - HANDLE = obj-handle
 - Action = MDC_ACT_SEG_GET_INFO
 - SegmSelection = abs-time-range, selecting a range with its boundaries set to an earlier date of any of the existing segments.
10. The agent under test operation response:
 - a. Data APDU
 - Type = Roer
 - ErrorResult = no-such-action (9)
11. The simulated manager sends a Get-Segment-Info:
 - a. Data APDU
 - Type = Invoke | Confirmed Action,
 - HANDLE = obj-handle
 - Action = MDC_ACT_SEG_GET_INFO
 - SegmSelection = abs-time-range, selecting a range with its boundaries set to a later date than any of the existing segments

12. The agent under test operation response:
 - a. Data APDU
 - Type = Roer
 - ErrorResult = no-such-action (9)
13. The simulated manager sends a Get-Segment-Info:
 - a. Data APDU
 - Type = Invoke | Confirmed Action,
 - HANDLE = obj-handle
 - Action = MDC_ACT_SEG_GET_INFO
 - SegmSelection = abs-time-range, selecting a range with one of its boundaries set to an earlier date than any of the existing segments and the other set to a date contained between Segment-Start-Abs-Time and Segment-End-Abs-Time of one of the PM-Segments
14. The agent under test operation response:
 - a. Data APDU
 - Type = Roer
 - ErrorResult = no-such-action (9)
15. The simulated manager sends a Get-Segment-Info:
 - a. Data APDU
 - Type = Invoke | Confirmed Action,
 - HANDLE = obj-handle
 - Action = MDC_ACT_SEG_GET_INFO
 - SegmSelection = abs-time-range, selecting a range with one of its boundaries set to a date contained between Segment-Start-Abs-Time and Segment-End-Abs-Time of one of the PM-Segments and the other set to a date later than any of the existing segments
16. The agent under test operation response:
 - a. Data APDU
 - Type = Roer
 - ErrorResult = no-such-action (9)
17. The simulated manager sends a Get-Segment-Info:
 - a. Data APDU
 - Type = Invoke | Confirmed Action,
 - HANDLE = obj-handle
 - Action = MDC_ACT_SEG_GET_INFO
 - SegmSelection = abs-time-range, selecting a range with its boundaries set to Segment-Start-Abs-Time and Segment-End-Abs-Time of one of the PM-Segments
18. The agent under test operation response:
 - a. Data APDU
 - Type = Invoke | Confirmed Action,
 - HANDLE = obj-handle
 - Action = MDC_ACT_SEG_GET_INFO
 - SegmentInfoList = Containing the attributes of the selected Segments
19. The simulated manager sends a Get-Segment-Info:
 - a. Data APDU
 - Type = Invoke | Confirmed Action,

- HANDLE = obj-handle
- Action = MDC_ACT_SEG_GET_INFO
- SegmSelection = abs-time-range, selecting a range with its boundaries set to include inside from Segment-Start-Abs-Time to Segment-End-Abs-Time one of the PM-Segments

20. The agent under test operation response:

a. Data APDU

- Type = Invoke | Confirmed Action,
- HANDLE = obj-handle
- Action = MDC_ACT_SEG_GET_INFO
- SegmentInfoList = Containing the attributes of the selected Segments

IF bit 6 of PmStoreCapab was set AND the agent reports the base-offset-time:

21. The simulated manager sends a Get-Segment-Info:

a. Data APDU

- Type = Invoke | Confirmed Action,
- HANDLE = obj-handle
- Action = MDC_ACT_SEG_GET_INFO
- SegmSelection = bo-time-range, selecting a range with its boundaries set to an earlier date than any of the existing segments.

22. The agent under test operation response:

a. Data APDU

- Type = Roer
- ErrorResult = no-such-action (9)

23. The simulated manager sends a Get-Segment-Info:

a. Data APDU

- Type = Invoke | Confirmed Action,
- HANDLE = obj-handle
- Action = MDC_ACT_SEG_GET_INFO
- SegmSelection = bo-time-range, selecting a range with its boundaries set to a later date than any of the existing segments.

24. The agent under test operation response:

a. Data APDU

- Type = Roer
- ErrorResult = no-such-action (9)

25. The simulated manager sends a Get-Segment-Info:

a. Data APDU

- Type = Invoke | Confirmed Action,
- HANDLE = obj-handle
- Action = MDC_ACT_SEG_GET_INFO
- SegmSelection = bo-time-range, selecting a range with one of its boundaries set to an earlier date than any of the existing segments and the other set to a date contained between Segment-Start-Bo-Time and Segment-End-Bo-Time of one of the PM-Segments

26. The agent under test operation response:

a. Data APDU

- Type = Roer
- ErrorResult = no-such-action (9)

27. The simulated manager sends a Get-Segment-Info:
- a. Data APDU
 - Type = Invoke | Confirmed Action,
 - HANDLE = obj-handle
 - Action = MDC_ACT_SEG_GET_INFO
 - SegmSelection = bo-time-range, selecting a range with one of its boundaries set to a date contained between Segment-Start-Bo-Time and Segment-End-Bo-Time of one of the PM-Segments and the other set to a later date than any of the existing segments
28. The agent under test operation response:
- a. Data APDU
 - Type = Roer
 - ErrorResult = no-such-action (9)
29. The simulated manager sends a Get-Segment-Info:
- a. Data APDU
 - Type = Invoke | Confirmed Action,
 - HANDLE = obj-handle
 - Action = MDC_ACT_SEG_GET_INFO
 - SegmSelection = bo-time-range, selecting a range with its boundaries set to Segment-Start-Bo-Time and Segment-End-Bo-Time of one of the PM-Segments
30. The agent under test operation response:
- a. Data APDU
 - Type = Invoke | Confirmed Action,
 - HANDLE = obj-handle
 - Action = MDC_ACT_SEG_GET_INFO
 - SegmentInfoList = Containing the attributes of the selected Segments
31. The simulated manager sends a Get-Segment-Info:
- a. Data APDU
 - Type = Invoke | Confirmed Action,
 - HANDLE = obj-handle
 - Action = MDC_ACT_SEG_GET_INFO
 - SegmSelection = bo-time-range, selecting a range with its boundaries set to include inside from Segment-Start-Bo-Time to Segment-End-Bo-Time of one of the PM-Segments
32. The agent under test operation response:
- a. Data APDU
 - Type = Invoke | Confirmed Action,
 - HANDLE = obj-handle
 - Action = MDC_ACT_SEG_GET_INFO
 - SegmentInfoList = Containing the attributes of the selected Segments
- IF bit 6 of PMStoreCapab was NOT set:
33. The simulated manager sends a Get-Segment-Info:
- a. Data APDU
 - Type = Invoke | Confirmed Action,
 - HANDLE = obj-handle
 - Action = MDC_ACT_SEG_GET_INFO

	<ul style="list-style-type: none"> <input type="checkbox"/> SegmSelection = abs-time-range, selecting a range with its boundaries set to the absolute minimum of Absolut-Time type and to the absolute maximum of the Absolute-Time type <p>34. The agent under test operation response:</p> <p>a. Data APDU</p> <ul style="list-style-type: none"> <input type="checkbox"/> Type = Roer <input type="checkbox"/> ErrorResult = no-such-action (9) or not-allowed-by-object (24)
Pass/Fail criteria	The agent properly sends the required PM-Segment attributes in all cases (all-segments, specific segments and time range selected segments) or the specified error.
Notes	For Clear-Segment, [ISO/IEEE 11073-20601A] has defined the error code not-allowed-by-object when the agent does not support the particular action (list of segments or range of segments), but for Get- Segment-Info any error code is defined. For this reason both error codes have been added to the test procedure because the error code used previously is not clearly defined in the change request.

TP Id	TP/PLT/AG/OXP/COM/BV-073		
TP label	Operating procedures. Error Code		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	PersStoreMtrDatTransf 6; M	
Applicability	C_AG_OXP_041 AND C_AG_OXP_000		
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager shall send a Get request for the PM-Store object with an attribute-id-list set to 0 to indicate all PM-Store attributes. 2. The simulated manager sends a request for the PM-Segment Data with SegmSelection = all-segments. 3. The simulated manager sends a request for the PM-Store Data to a non-existent Segment: <ol style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action, <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_TRIG_XFER <input type="checkbox"/> TrigSegmDataXferReq 4. The agent issues a response: <ol style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action, <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_TRIG_XFER <input type="checkbox"/> TrigSegmDataXferRsp = tsxr-fail-no-such-segment(1) 		
Pass/Fail criteria	The response from agent under test must be of type tsxr-fail-no-such-segment(1).		
Notes			

TP Id	TP/PLT/AG/OXP/COM/BV-073_A		
TP label	Operating procedures. Transfer PM-Segment content		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	PersStoreMtrDatTransf 6; M	
Applicability	C_AG_OXP_041 AND C_AG_OXP_000		

Initial condition	The simulated manager and the agent under test are in the operating state.
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager shall send a Get request for the PM-Store object with an attribute-id-list set to 0 to indicate all PM-Store attributes. 2. The simulated manager sends a request for the PM-Segment Data with SegmSelection = all-segments. 3. The simulated manager sends a request for the PM-Store Data to a Segment: <ol style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action, <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_TRIG_XFER <input type="checkbox"/> TrigSegmDataXferReq 4. The agent issues a response: <ol style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action, <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_TRIG_XFER <input type="checkbox"/> TrigSegmDataXferRsp <input type="checkbox"/> Check that the invoke-id of the response is mirrored from the request
Pass/Fail criteria	TrigSegmDataXferRsp must be one of: <ul style="list-style-type: none"> • tsxr-successful(0) • tsxr-fail-clear-in-process(2) • tsxr-fail-segm-empty(3) • tsxr-fail-not-otherwise-specified(512)
Notes	

TP Id	TP/PLT/AG/OXP/COM/BV-074			
TP label	Operating procedures. Segment Data Event			
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	PersStoreMtrDatTransf 7; M	PersStoreMtrDatTransf 8; M	PersStoreMtrDatTransf 9; M
		PersStoreMtrDatTransf 10; M	PersStoreMtrDatTransf 12; M	CommonCharac 3; M
Applicability	C_AG_OXP_041 AND C_AG_OXP_000			
Initial condition	The simulated manager and the agent under test are in the operating state.			
Test procedure	<ol style="list-style-type: none"> 1. Take some measurements with the agent under test. 2. The simulated manager shall send a Get request for the PM-Store object with an attribute-id-list set to 0 to indicate all PM-Store attributes. 3. The agent issues a GET response with the PM-Store attributes it supports. 4. The simulated manager sends a request for the PM-Segment info with SegmSelection = 1 to obtain all the segments: <ol style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action, <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_GET_INFO <input type="checkbox"/> SegmSelection = all-segments 			

	<ol style="list-style-type: none"> 5. The agent issues a response with the PM-Segments attributes: <ol style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action, <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_GET_INFO <input type="checkbox"/> SegmentInfoList 6. The simulated manager sends a request for PM-Segment Data: <ol style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action, <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_TRIG_XFER <input type="checkbox"/> TrigSegmDataXferReq 7. The agent issues a response: <ol style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action, <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_TRIG_XFER <input type="checkbox"/> TrigSegmDataXferRsp 8. The agent under test starts a Data transfer: <ol style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Invoke CfmEventReport <input type="checkbox"/> Action = MDC_NOTI_SEGMENT_DATA <input type="checkbox"/> SegmentDataEvent <input type="checkbox"/> Segm-data-event-entries = Data 9. The simulated manager responds to transferred data APDU's: <ol style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_NOTI_SEGMENT_DATA <input type="checkbox"/> SegmentDataResult 10. Steps 8 and 9 are repeated until all the data has been sent.
<p>Pass/Fail criteria</p>	<ul style="list-style-type: none"> • The agent replies to the Get request with the requested Data and sevtsta-manager-* bits to 0 • In the first Data event sent sevtsta-first-entry bit must be set by the agent • In the last data event sent the sevtsta-last-entry bit must be set by the agent • In step 7 the total size of the message can not exceed the maximum APDU size established by the specialization: <ul style="list-style-type: none"> ○ Pulse oximeter -> 9216 octets ○ Weighing scales -> 896 octets ○ Glucose meter -> 5120 octets or 64512 octets if the agent supports PM-Store ○ Blood pressure -> 896 octets ○ Thermometer -> 896 octets ○ Independent activity hub -> 5120 octets ○ Cardiovascular -> 64512 octets or 6624 octets if the agent supports Step Counter Profile

	<ul style="list-style-type: none"> ○ Strength -> 64512 octets ○ Adherence monitor -> 1024 octets ○ Peak flow -> 2030 octets ○ Body composition analyser -> 7730 octets ○ Basic ECG/Simple ECG -> 7168 octets or 64512 octets if the agent supports PM-Store ○ Basic ECG/Heart rate -> 1280 octets or 64512 octets if the agent supports PM-Store ○ International normalized ratio -> 896 octets or 64512 if the agent supports PM-Store
Notes	

TP Id	TP/PLT/AG/OXP/COM/BV-076		
TP label	Operating procedures. PM-Segment structure		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	PersStoreMtrDatTransf 15; M	
Applicability	C_AG_OXP_041 AND C_AG_OXP_000		
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager shall send a Get request for the PM-Store object with an attribute-id-list set to 0 to indicate all PM-Store attributes. 2. The agent issues a GET response with the PM-Store attributes. 3. The simulated manager issues a Get-Segment-Info action with SemgSelection set to all-segments, for this test we are interested in: <ol style="list-style-type: none"> a. Mandatory attribute PM-Segment-Entry-Map <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_PM_SEG_MAP <input type="checkbox"/> attribute-type = PmSegmentEntryMap <input type="checkbox"/> attribute-value = 4. The simulated manager sends a request for the PM-Segment that contains data: <ol style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action, <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_TRIG_XFER <input type="checkbox"/> TrigSegmDataXferReq 5. The agent issues an action response: <ol style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action, <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_TRIG_XFER <input type="checkbox"/> TrigSegmDataXferRsp 6. The agent under test starts Data transfer: <ol style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Invoke CfmEventReport <input type="checkbox"/> Action = MDC_NOTI_SEGMENT_DATA <input type="checkbox"/> SegmentDataEvent: <ul style="list-style-type: none"> - sevtsta-first-entry(0)=1 		

	<ul style="list-style-type: none"> - segm-data-event-entries=Data <p>7. The simulated manager responds to transferred data APDU's with an abort transfer:</p> <p>b. Data APDU</p> <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_NOTI_SEGMENT_DATA <input type="checkbox"/> SegmentDataResult <ul style="list-style-type: none"> - sevtsta-manager-abort(12)=1
Pass/Fail criteria	The format of the data has to coincide with the format expressed in the PmSegmentEntryMap field and the agent does not send any SegmentDataEvent after step 7.
Notes	

TP Id	TP/PLT/AG/OXP/COM/BV-080		
TP label	Release Request. Outstanding invoke-id		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	DisassocProc 3;M	DisassocProc 4;M
Applicability	C_AG_OXP_000		
Initial condition	The simulated manager and the agent under test are in the unassociated state.		
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an Associating Request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent responds with a roiv-cmip-confirmed-event report message with a MDC_NOTI_CONFIG event to send its configuration to the manager. TOconfig is started. 4. The simulated manager sends a Release Request with Reason = 0 "normal". 5. The agent under test responds with a Release Response and changes to the unassociated state. 6. Wait for a time equal to TOconfig. 		
Pass/Fail criteria	During the period of step 6 the agent does not send any abort message. After that point, the only message that may be received by the manager is a new Association request.		
Notes			

TP Id	TP/PLT/AG/OXP/COM/BV-081_A		
TP label	Disassociating procedure. Release Request Reason 1		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	DisassocProc 2; M	
Applicability	C_AG_OXP_186 AND C_AG_OXP_000		
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<ol style="list-style-type: none"> 1. Force the agent to send a Release Request. 2. Check that the manager receives a Release Request with reason = normal (0). 		
Pass/Fail criteria	The Association Release Request contains a ReleaseRequestReason to indicate the reason for releasing the association. The Reason code shall be the one described in step 2.		
Notes			

TP Id		TP/PLT/AG/OSP/COM/BV-081_B		
TP label		Disassociating procedure. Release Request Reason 2		
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	DisassocProc 2; M		
Applicability		C_AG_OXP_191 AND C_AG_OXP_000		
Initial condition		The simulated manager and the agent under test are in the operating state.		
Test procedure		<ol style="list-style-type: none"> 1. Change the configuration of the agent under test (adding or removing objects from the DIM). 2. Check that the simulated manager receives a Release Request with reason = configuration-changed (2). 		
Pass/Fail criteria		The Association Release Request contains a ReleaseRequestReason to indicate the reason for releasing the association. The Reason code shall be the one described in step 2.		
Notes				

TP Id		TP/PLT/AG/OSP/COM/BV-081_C		
TP label		Disassociating procedure. Release Request Reason 3		
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	DisassocProc 2; M		
Applicability		C_AG_OXP_000		
Initial condition		The simulated manager and the agent under test are in the operating state.		
Test procedure		<ol style="list-style-type: none"> 1. Disconnect and again connect the agent under test to the simulated manager. 2. The simulated manager receives an association request from the the agent under test. 3. The simulated manager responds with a result = accepted-unknown-config. 4. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 5. The simulated manager responds with an "unsupported-config". 6. Repeat steps 4 and 5 until the agent sends a Release Request. 		
Pass/Fail criteria		The Association Release Request contains a ReleaseRequestReason = no-more-configurations (1).		
Notes				

TP Id		TP/PLT/AG/OSP/COM/BV-083		
TP label		Disassociating procedure. Association Release Response		
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	DisassocProc 8; M	DisassocProc 9; M	
Applicability		C_AG_OXP_000		
Initial condition		The simulated manager and the agent under test are in the operating state.		
Test procedure		<ol style="list-style-type: none"> 1. IF the agent under test can be forced by the user to send a release request, send it. If not, follow this procedure: <ol style="list-style-type: none"> a. Disconnect and again connect the agent under test to the simulated manager. b. The simulated manager receives an association request from the agent under test. c. The simulated manager responds with a result = accepted-unknown-config. 		

	<ol style="list-style-type: none"> d. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. e. The simulated manager responds with an "unsupported-config". f. Repeat steps d and e until the agent sends a Release Request. <ol style="list-style-type: none"> 2. Once the agent under test has sent a Release Request: <ol style="list-style-type: none"> a. The simulated manager does not respond to the request for at least TO_{Release} (3 seg.).
Pass/Fail criteria	The agent waits the TO _{Release} time and then it must send an abort message to the manager.
Notes	

TP Id	TP/PLT/AG/EXP/COM/BV-084		
TP label	Absolute time 1		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	TimeCoord 1; C	AbsTime 1; C
		AbsTime 3; C	AbsTime 5; C
Applicability	C_AG_EXP_009 AND C_AG_EXP_013 AND C_AG_EXP_000		
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager issues "Remote Operation Invoke Get" command with the handle set to 0 (to request an MDS object) and the attribute-id-list set to 0 to indicate all attributes. 2. The agent responds with with a "rors-cmip-get" service message in which the attribute-list contains a list of all implemented attributes of the MDS object. The attribute of interest of this test is MDSTimeInfo: <ol style="list-style-type: none"> a. Mds-Time-Info: <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_MDS_TIME_INFO (0X0A 0X45) <input type="checkbox"/> attribute-type = MdsTimeInfo <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> mds-time-capab-real-time-clock must be set b. IF the agent can synchronize its absolute time then: <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_MDS_TIME_INFO (0X0A 0X45) <input type="checkbox"/> attribute-type = MdsTimeInfo <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> mds-time-capab-sync-abs-time must be set <input type="checkbox"/> time-sync-protocol field must indicate what protocol is used for synchronization 3. IF mds-time-capab-sync-abs-time = 1 THEN: <ol style="list-style-type: none"> a. Ask the test operator to connect the external source that is going to be used to synchronize the agent AbsoluteTime b. The simulated manager issues a "Remote Operation Invoke Get" command with the handle set to 0 (to request an MDS object) and the attribute-id-list set to 0 to indicate all attributes. c. The agent responds with a "rors-cmip-get" service message in which the attribute-list contains a list of all implemented attributes of the MDS object: <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_MDS_TIME_INFO (0X0A 0X45) <input type="checkbox"/> attribute-type = MdsTimeInfo <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> mds-time-capab-sync-abs-time must be set 		

	<input type="checkbox"/> mds-time-state -abs-time-synced must be set
Pass/Fail criteria	Check that the attribute mds-time-cap-state has correct values.
Notes	

TP Id	TP/PLT/AG/OXP/COM/BV-085_A		
TP label	Absolute time 2		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	AbsTime 11; C	MDSService 7; O
Applicability	C_AG_OXP_006 AND C_AG_OXP_009 AND C_AG_OXP_012 AND C_AG_OXP_000 AND C_AG_OXP_016		
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<ol style="list-style-type: none"> 1. Take a measurement with the agent. 2. Make a noticeable change in the Date or Time of the agent. 3. Take a new measurement. 4. Wait for a roiv-cmip-event-report OR a roiv-cmip-confirmed-event-report from the agent. 5. Verify that the device sends a variable format event report to update the Date-and-Time attribute on the MDS before it sends any measurement updates and that every measure taken before the time change is sent in the same event report. 		
Pass/Fail criteria	The agent transmitted data comes from the same unbroken timeline which means that every measure taken before the time change has a date-and-time-adjustment.		
Notes			

TP Id	TP/PLT/AG/OXP/COM/BV-085_B		
TP label	Absolute time 2: PM-Store		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	AbsTime 14; C	
Applicability	C_AG_OXP_012 AND C_AG_OXP_041 AND C_AG_OXP_000 AND C_AG_OXP_016		
Initial condition	The simulated manager and the agent under test are in the unassociated state.		
Test procedure	<ol style="list-style-type: none"> 1. While the agent is disconnected, make it store measurements in PM-Segments of every PM-Store, after doing this, connect the agent. 2. The simulated manager receives an association request from the agent under test. 3. The simulated manager responds with a result = accepted-unknown-config. 4. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message: <ul style="list-style-type: none"> o Event-type=MDC_NOTI_CONFIG 5. Record the PM-Store handle, PM-Store-Capab and Number-Of-Segments of every PM-Store object. 6. The simulated manager shall send a Get-Segment-Info object action for the PM-Store object with SegmSelection set to all-segments. 7. The agent shall respond to the Get-Segment-Info, indicating the attributes of the PM-Segment. 8. Make a noticeable change in the Date or Time of the agent. 9. Take a new measurement. 10. The simulated manager shall send a Get-command for every PM-Store. 11. The agent shall respond to the Get command, indicating the attributes of the PM-Store. 		

	<p>Record.</p> <p>12. The simulated manager shall send a Get-Segment-Info object action for the PM-Store object with SegmSelection set to all-segments.</p> <p>13. The agent shall respond to the Get-Segment-Info, indicating the attributes of the PM-Segment:</p> <ul style="list-style-type: none"> ○ The Date-and-Time adjustment attribute is present ○ If the pmsc-var-no-of-segm is set to 1 (PM-Store-Capab) then verify that the Pm-Store has created a new segment
Pass/Fail criteria	<ul style="list-style-type: none"> • If the pmsc-var-no-of-segm is set to 1, the number of segments recorded in step 5 has increased in step 11 and the Date and Time Adjustment is present for every new added segment • If the pmsc-var-no-of-segm is set to 0, the Date and Time Adjustment is present at least for one segment
Notes	

TP Id	TP/PLT/AG/OXP/COM/BV-085_C		
TP label	Absolute time 2: Store and Forward		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	AbsTime 12; C	AbsTime 13; C
Applicability	C_AG_OXP_009 AND C_AG_OXP_012 AND C_AG_OXP_032 AND C_AG_OXP_000		
Initial condition	The simulated manager and the agent under test are in the disconnected state.		
Test procedure	<ol style="list-style-type: none"> 1. Take some measurements with the agent under test. 2. Make a change in the Date or Time of the agent by advancing its clock 60 minutes. 3. Take new measurements. 4. Connect the agent under test to the simulated manager. 5. Once in the operating state the manager has to receive a variable event report containing the Date-and-Time-Adjustment attribute with the value = 360000. 		
Pass/Fail criteria	The agent recorded data before the Time change must be in a different segment than those recorded after the time change.		
Notes			

TP Id	TP/PLT/AG/OXP/COM/BV-086		
TP label	Relative time 1		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	RelativeTime 2; C	RelativeTime 5; C
Applicability	C_AG_OXP_010 AND C_AG_OXP_000		
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager issues a "Remote Operation Invoke Get" command with the handle set to 0 (to request an MDS object) and the attribute-id-list set to 0 to indicate all attributes. 2. The agent responds with with a "rors-cmip-get" service message in which the attribute-list contains a list of all implemented attributes of the MDS object: <ol style="list-style-type: none"> a. Mds-Time-Info shall be present: <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_MDS_TIME_INFO (0X0A 0X45) <input type="checkbox"/> attribute-type = MdsTimeInfo <input type="checkbox"/> attribute-value.length = 2 bytes 		

	<ul style="list-style-type: none"> <input type="checkbox"/> mds-time-capab-relative-time must be set <p>b. IF the agent can synchronize its relative timer then:</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_MDS_TIME_INFO (0X0A 0X45) <input type="checkbox"/> attribute-type = MdsTimeInfo <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> mds-time-capab-sync-rel-time must be set <input type="checkbox"/> time-sync-protocol field must indicate what protocol is used for synchronization <p>3. IF the mds-time-capab-sync-rel-time = 1 THEN:</p> <p>a. Ask the test operator to connect the external source that is going to be used to synchronize the Agent Relative-Time.</p> <p>b. The simulated manager issues a "Remote Operation Invoke Get" command with the handle set to 0 (to request an MDS object) and the attribute-id-list set to 0 to indicate all attributes.</p> <p>c. The agent responds with a "rors-cmip-get" service message in which the attribute-list contains a list of all implemented attributes of the MDS object:</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_MDS_TIME_INFO (0X0A 0X45) <input type="checkbox"/> attribute-type = MdsTimeInfo <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> mds-time-capab-sync-rel-time must be set <input type="checkbox"/> mds-time-state -rel-time-synced must be set
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id		TP/PLT/AG/OXP/COM/BV-087		
TP label		High-resolution Relative time		
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	Hi-resRelativeTime 1; C	Hi-resRelativeTime 4; C	Hi-resRelativeTime 5; C
		Hi-resRelativeTime 6; C		
Applicability		C_AG_OXP_011 AND C_AG_OXP_000		
Initial condition		The simulated manager and the agent under test are in the operating state.		
Test procedure		<p>1. The simulated manager issues a "Remote Operation Invoke Get" command with the handle set to 0 (to request an MDS object) and the attribute-id-list set to 0 to indicate all attributes.</p> <p>2. The agent responds with with a "rors-cmip-get" service message in which the attribute-list contains a list of all implemented attributes of the MDS object:</p> <p>a. To support Hires-Relative Time:</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_MDS_TIME_INFO (0X0A 0X45) <input type="checkbox"/> attribute-type = MdsTimeInfo <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> mds-time-capab-high-res-relative-time must be set <p>b. IF the agent can synchronize its High Resolution Relative timer then:</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_MDS_TIME_INFO (0X0A 0X45) <input type="checkbox"/> attribute-type = MdsTimeInfo <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> mds-time-capab-sync-hi-res-relative-time must be set <input type="checkbox"/> time-sync-protocol field must indicate what protocol is used for synchronization 		

	<p>3. IF the mds-time-capab-sync-hi-res-relative-time = 1 THEN:</p> <p>c. Ask the test operator to connect the external source that is going to be used to synchronize the agent Hi-Resolution-Relative-Time.</p> <p>d. The simulated manager issues a "Remote Operation Invoke Get" command with the handle set to 0 (to request an MDS object) and the attribute-id-list set to 0 to indicate all attributes.</p> <p>e. The agent responds with a "rors-cmip-get" service message in which the attribute-list contains a list of all implemented attributes of the MDS object:</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_MDS_TIME_INFO (0X0A 0X45) <input type="checkbox"/> attribute-type = MdsTimeInfo <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> mds-time-capab-sync-hi-res-relative-time must be set <input type="checkbox"/> mds-time-state-hi-resrelative-time-synced must be set <input type="checkbox"/> time-sync-accuracy will be recorded. <p>f. Ask the test operator to disconnect the external source that has been used to synchronize the agent Hi-Resolution-Relative-Time.</p> <p>g. Wait a time interval longer than the time specified in time-sync-accuracy, if it is undefined the test operator has to wait a time that he thinks is enough to exceed the accuracy of the clock synchronization.</p> <p>h. The simulated manager issues a "Remote Operation Invoke Get" command with the handle set to 0 (to request an MDS object) and the attribute-id-list set to 0 to indicate all attributes.</p> <p>i. The agent responds with a "rors-cmip-get" service message in which the attribute-list contains a list of all implemented attributes of the MDS object:</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_MDS_TIME_INFO (0X0A 0X45) <input type="checkbox"/> attribute-type = MdsTimeInfo <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> mds-time-state-hi-resrelative-time-synced must be clear
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/AG/OXP/COM/BV-088		
TP label	Base-Offset-Time 1		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	BaseTimOffset1; M	
Applicability	C_AG_OXP_014 AND C_AG_OXP_013 AND C_OXP_000		
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<p>1. The simulated manager issues a "Remote Operation Invoke Get" command with the handle set to 0 (to request an MDS object) and the attribute-id-list set to 0 to indicate all attributes.</p> <p>2. The agent responds with with a "rors-cmip-get" service message in which the attribute-list contains a list of all implemented attributes of the MDS object. The attribute of interest of this test is MDSTimeInfo:</p> <p>a. Mds-Time-Info:</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_MDS_TIME_INFO (0X0A 0X45) <input type="checkbox"/> attribute-type = MdsTimeInfo <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> mds-time-capab-bo-time must be set 		

	<p>b. IF the agent can synchronize its base-offset-time (i.e., C_AG_OXP_007 = TRUE) THEN:</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_MDS_TIME_INFO (0X0A 0X45) <input type="checkbox"/> attribute-type = MdsTimeInfo <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> mds-time-capab-sync-bo-time must be set <input type="checkbox"/> time-sync-protocol = MDC_TIME_SYNC_NTPV3 or MDC_TIME_SYNC_NTPV4 or MDC_TIME_SYNC_SNTPV4 or MDC_TIME_SYNC_SNTPV4330 or MDC_TIME_SYNC_BTV1 <p>3. IF mds-time-capab-sync-bo-time = 1 THEN:</p> <p>a. Ask the test operator to connect the external source that is going to be used to synchronize the agent Base-Offset time.</p> <p>b. The simulated manager issues a "Remote Operation Invoke Get" command with the handle set to 0 (to request an MDS object) and the attribute-id-list set to 0 to indicate all attributes.</p> <p>c. The agent responds with a "rors-cmip-get" service message in which the attribute-list contains a list of all implemented attributes of the MDS object:</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_MDS_TIME_INFO (0X0A 0X45) <input type="checkbox"/> attribute-type = MdsTimeInfo <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> mds-time-capab-sync-bo-time must be set <input type="checkbox"/> mds-time-state-bo-time-synced must be set
Pass/Fail criteria	Check the attribute mds-time-cap-state has correct values.
Notes	

TP Id	TP/PLT/AG/OXP/COM/BV-092		
TP label	Supported Standard Configuration		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	ConfNormalProc 16; M	
Applicability	C_AG_OXP_001 AND C_AG_OXP_002 AND C_AG_OXP_000		
Initial condition	The simulated manager and the agent under test are in the unassociated state.		
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 4. The simulated manager responds with a "unsupported-config". 5. Repeat steps 3 and 4 until the agent sends a Release Request with reason = "no-more-configurations". 		
Pass/Fail criteria	The agent under test must send at least one configuration with dev-config-id = <between 0x00 0x01 and 0x3F 0xFF > (standard range values) as a fallback if the extended configurations are not supported by the simulated manager.		
Notes			

TP Id	TP/PLT/AG/OXP/COM/BV-093		
TP label	Operating State. Abort message		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable	AgentStateMach 63; M	

	items			
Applicability		C_AG_OXP_000		
Initial condition		The simulated manager and the agent are in the operating state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated manager sends an Abort message to the agent under test. 2. Wait for an event report for the agent. 		
Pass/Fail criteria		The simulated manager must not receive any message other than an Association Request.		
Notes				

TP Id		TP/PLT/AG/OXP/COM/BV-096_A		
TP label		Agent State machine. Connected Associated Configuring Waiting Approval 11		
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	AgentStateMach 78; M		
Applicability		C_AG_OXP_000		
Initial condition		The agent under test is in the waiting approval state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated manager sends a badly formatted message. 2. The agent under test sends an abort message (abrt) with reason undefined to the manager and shall pass to the unassociated state. 		
Pass/Fail criteria		<ul style="list-style-type: none"> • The agent sends the Abort message (abrt) with reason undefined and changes to the unassociated state • The simulated manager must not receive any message other than an Association Request after step 2 		
Notes				

TP Id		TP/PLT/AG/OXP/COM/BV-097		
TP label		Agent State machine. Leaving Operating State 5		
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	AgentStateMach 80; M		
Applicability		C_AG_OXP_000		
Initial condition		The agent under test is in the operating state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated manager issues a Prst message, rors-cmip- get. 2. The agent under test sends an abort message (abrt) with reason undefined to the manager and shall pass to the unassociated state. 		
Pass/Fail criteria		<ul style="list-style-type: none"> • The agent under test sends the Abort message (abrt) with reason undefined and changes to the unassociated state • The simulated manager must not receive any message other than an Association Request after step 2 		
Notes				

TP Id		TP/PLT/AG/OXP/COM/BV-097_A		
TP label		Agent State machine. Leaving Operating State 6		
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	AgentStateMach 80; M		
Applicability		C_AG_OXP_000		

Initial condition	The agent under test is in the operating state.
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager sends a badly formatted message. 2. The agent under test sends an abort message (abrt) with reason undefined to the manager and shall pass to the unassociated state.
Pass/Fail criteria	<ul style="list-style-type: none"> • The agent under test sends the Abort message (abrt) with reason undefined and changes to the unassociated state • The simulated manager must not receive any message other than an Association Request after step 2
Notes	

TP Id	TP/PLT/AG/OSP/COM/BV-098_A		
TP label	Agent State machine. Connected Disassociation 7		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	AgentStateMach 81; M	
Applicability	C_AG_OXP_000		
Initial condition	The agent is in the unassociated state.		
Test procedure	<ol style="list-style-type: none"> 1. The agent under test sends an Association Request to the simulated manager. 2. The simulated manager responds with an accepted-unknown-config. 3. The agent under test sends a configuration event report. 4. The simulated manager responds with an unsupported-configuration. 5. The agent sends a new configuration event report with a new configuration (if it has more). 6. Repeat the last two steps recording all the ConfigId-values until the agent sends a Release Request with the reason "no-more-configurations". The agent moves to the disassociating state. 7. The simulated manager sends a badly formatted message. 8. The agent responds with an Abort message (abrt) with reason undefined. 9. The agent and the manager move to the unassociated state. 		
Pass/Fail criteria	<ul style="list-style-type: none"> • The agent under test sends the abort message (abrt) with reason undefined and changes to the unassociated state • The simulated manager must not receive any message other than an Association Request after step 9. 		
Notes			

TP Id	TP/PLT/AG/OSP/COM/BV-099		
TP label	Operating procedures. PM-Store Specific Attributes request		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	PersStoreMtrDatTransf 2; C	
Applicability	C_AG_OXP_000 AND C_AG_OXP_041		
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager issues a "Remote Operation Invoke Get" command with: <ol style="list-style-type: none"> a. the Obj-handle set to PM-Store object handle (to request its attributes) b. the attribute-id-list.count=1 and a single AVA_Type MDC_ATTR_PM_STORE_CAPAB (0X0A 0X4D) to retrieve the mandatory "PM-Store-Capab" attribute 2. The agent under test responds with: 		

	<ul style="list-style-type: none"> • IF C_AG_OXP_101 THEN: with a "rors-cmip-get" service message which contains the "PM-Store-Capab" • ELSE: with a "roer" service message with the error-value set to not-allowed-by-object (24) <ol style="list-style-type: none"> 3. The simulated manager issues a "Remote Operation Invoke Get" command with: <ol style="list-style-type: none"> a. the Obj-handle set to PM-Store object handle (to request its attributes) b. the attribute-id-list empty to request all the attributes of PM-Store object 4. The agent responds with with a "rors-cmip-get" service message which contains all the supported attributes of the PM-Store object. 5. The simulated manager issues a "Remote Operation Invoke Get" command with: <ol style="list-style-type: none"> a. the Obj-handle set to PM-Store object handle (to request its attributes) b. the attribute-id-list set to an attribute NOT supported by the PM-Store object 6. The agent responds with a "rors-cmip-get" service message: <ul style="list-style-type: none"> • IF C_AG_OXP_101 THEN: the attribute-list must be empty • ELSE: with with a "roer" service message with the error-value set to not-allowed-by-object (24) 7. The simulated manager issues a "Remote Operation Invoke Get" command with: <ol style="list-style-type: none"> a. the Obj-handle set to PM-Store object handle (to request its attributes) b. the attribute-id-list contains one supported attribute and one unsupported attribute 8. The agent responds with a "rors-cmip-get" service message: <ul style="list-style-type: none"> • IF C_AG_OXP_101 THEN: the attribute-list must contain the supported attribute • ELSE: with with a "roer" service message with the error-value set to not-allowed-by-object (24)
Pass/Fail criteria	<ul style="list-style-type: none"> • In step 2 the agent properly sends the requested attribute or the error • In steps 6 and 8 the received attribute list must be empty if NOT C_AG_OXP_101 or roer if the action is not supported
Notes	

Bibliography

- [b-CDG 1.0] Continua Health Alliance, Continua Design Guidelines v1.0 . (2008), *Continua Design Guidelines*.
- [b-CDG 2010] Continua Health Alliance, Continua Design Guidelines v1.5 (2010), *Continua Design Guidelines*.
- [b-CDG 2011] Continua Health Alliance, Continua Design Guidelines (2011), "Adrenaline", *Continua Design Guidelines*.
- [b-CDG 2012] Continua Health Alliance CDG, Continua Design Guidelines (2012), "Catalyst", *Continua Design Guidelines*.
- [b-ETSI 300 406] ETSI ETS 300 406, *Methods for Testing and Specifications (MTS); Protocol and profile conformance testing specifications; Standardization methodology*.
- [b-ETSI SR 001 262] ETSI SR 001 262 v1.8.1 (2003), *ETSI drafting rules*.
- [b-ISO/IEC 9646-1] ISO/IEC 9646-1, *Information Technology – Open Systems Interconnection – Conformance testing methodology and framework – Part 1: General concepts*.
- [b-ISO/IEC 9646-7] ISO/IEC 9646-7, *Information technology – Open Systems Interconnection – Conformance testing methodology and framework – Part 7: Implementation Conformance Statements*.

SERIES OF ITU-T RECOMMENDATIONS

Series A	Organization of the work of ITU-T
Series D	General tariff principles
Series E	Overall network operation, telephone service, service operation and human factors
Series F	Non-telephone telecommunication services
Series G	Transmission systems and media, digital systems and networks
Series H	Audiovisual and multimedia systems
Series I	Integrated services digital network
Series J	Cable networks and transmission of television, sound programme and other multimedia signals
Series K	Protection against interference
Series L	Construction, installation and protection of cables and other elements of outside plant
Series M	Telecommunication management, including TMN and network maintenance
Series N	Maintenance: international sound programme and television transmission circuits
Series O	Specifications of measuring equipment
Series P	Terminals and subjective and objective assessment methods
Series Q	Switching and signalling
Series R	Telegraph transmission
Series S	Telegraph services terminal equipment
Series T	Terminals for telematic services
Series U	Telegraph switching
Series V	Data communication over the telephone network
Series X	Data networks, open system communications and security
Series Y	Global information infrastructure, Internet protocol aspects and next-generation networks
Series Z	Languages and general software aspects for telecommunication systems