

ITU-T

TELECOMMUNICATION
STANDARDIZATION SECTOR
OF ITU

H.845.1

(01/2015)

SERIES H: AUDIOVISUAL AND MULTIMEDIA SYSTEMS

E-health multimedia services and applications –
Interoperability compliance testing of personal health
systems (HRN, PAN, LAN, TAN and WAN)

**Conformance of ITU-T H.810 personal health
devices: PAN/LAN/TAN interface Part 5A:
Weighing scales: Agent**

Recommendation ITU-T H.845.1

ITU-T H-SERIES RECOMMENDATIONS
AUDIOVISUAL AND MULTIMEDIA SYSTEMS

CHARACTERISTICS OF VISUAL TELEPHONE SYSTEMS	H.100–H.199
INFRASTRUCTURE OF AUDIOVISUAL SERVICES	
General	H.200–H.219
Transmission multiplexing and synchronization	H.220–H.229
Systems aspects	H.230–H.239
Communication procedures	H.240–H.259
Coding of moving video	H.260–H.279
Related systems aspects	H.280–H.299
Systems and terminal equipment for audiovisual services	H.300–H.349
Directory services architecture for audiovisual and multimedia services	H.350–H.359
Quality of service architecture for audiovisual and multimedia services	H.360–H.369
Telepresence	H.420–H.429
Supplementary services for multimedia	H.450–H.499
MOBILITY AND COLLABORATION PROCEDURES	
Overview of Mobility and Collaboration, definitions, protocols and procedures	H.500–H.509
Mobility for H-Series multimedia systems and services	H.510–H.519
Mobile multimedia collaboration applications and services	H.520–H.529
Security for mobile multimedia systems and services	H.530–H.539
Security for mobile multimedia collaboration applications and services	H.540–H.549
Mobility interworking procedures	H.550–H.559
Mobile multimedia collaboration inter-working procedures	H.560–H.569
BROADBAND, TRIPLE-PLAY AND ADVANCED MULTIMEDIA SERVICES	
Broadband multimedia services over VDSL	H.610–H.619
Advanced multimedia services and applications	H.620–H.629
Ubiquitous sensor network applications and Internet of Things	H.640–H.649
IPTV MULTIMEDIA SERVICES AND APPLICATIONS FOR IPTV	
General aspects	H.700–H.719
IPTV terminal devices	H.720–H.729
IPTV middleware	H.730–H.739
IPTV application event handling	H.740–H.749
IPTV metadata	H.750–H.759
IPTV multimedia application frameworks	H.760–H.769
IPTV service discovery up to consumption	H.770–H.779
Digital Signage	H.780–H.789
E-HEALTH MULTIMEDIA SERVICES AND APPLICATIONS	
Interoperability compliance testing of personal health systems (HRN, PAN, LAN, TAN and WAN)	H.820–H.859
Multimedia e-health data exchange services	H.860–H.869

For further details, please refer to the list of ITU-T Recommendations.

Recommendation ITU-T H.845.1

Conformance of ITU-T H.810 personal health devices: PAN/LAN/TAN interface Part 5A: Weighing scales: Agent

Summary

Recommendation ITU-T H.845.1 is a transposition of Continua Test Tool DG2013, Test Suite Structure & Test Purposes, PAN-LAN-TAN Interface; Part 5A: Device Specializations. Agent (Weighing Scale) (Version 1.4, 2014-01-24), that was developed by the Continua Health Alliance. A number of versions of this specification existed before transposition.

This Recommendation includes an electronic attachment with the protocol implementation conformance statements (PICS) and the protocol implementation extra information for testing (PIXIT) required for the implementation of Annex A.

History

Edition	Recommendation	Approval	Study Group	Unique ID*
1.0	ITU-T H.845.1	2015-01-13	16	11.1002/1000/12262

* To access the Recommendation, type the URL <http://handle.itu.int/> in the address field of your web browser, followed by the Recommendation's unique ID. For example, <http://handle.itu.int/11.1002/1000/11830-en>.

FOREWORD

The International Telecommunication Union (ITU) is the United Nations specialized agency in the field of telecommunications, information and communication technologies (ICTs). The ITU Telecommunication Standardization Sector (ITU-T) is a permanent organ of ITU. ITU-T is responsible for studying technical, operating and tariff questions and issuing Recommendations on them with a view to standardizing telecommunications on a worldwide basis.

The World Telecommunication Standardization Assembly (WTSA), which meets every four years, establishes the topics for study by the ITU-T study groups which, in turn, produce Recommendations on these topics.

The approval of ITU-T Recommendations is covered by the procedure laid down in WTSA Resolution 1.

In some areas of information technology which fall within ITU-T's purview, the necessary standards are prepared on a collaborative basis with ISO and IEC.

NOTE

In this Recommendation, the expression "Administration" is used for conciseness to indicate both a telecommunication administration and a recognized operating agency.

Compliance with this Recommendation is voluntary. However, the Recommendation may contain certain mandatory provisions (to ensure, e.g., interoperability or applicability) and compliance with the Recommendation is achieved when all of these mandatory provisions are met. The words "shall" or some other obligatory language such as "must" and the negative equivalents are used to express requirements. The use of such words does not suggest that compliance with the Recommendation is required of any party.

INTELLECTUAL PROPERTY RIGHTS

ITU draws attention to the possibility that the practice or implementation of this Recommendation may involve the use of a claimed Intellectual Property Right. ITU takes no position concerning the evidence, validity or applicability of claimed Intellectual Property Rights, whether asserted by ITU members or others outside of the Recommendation development process.

As of the date of approval of this Recommendation, ITU had not received notice of intellectual property, protected by patents, which may be required to implement this Recommendation. However, implementers are cautioned that this may not represent the latest information and are therefore strongly urged to consult the TSB patent database at <http://www.itu.int/ITU-T/ipr/>.

© ITU 2015

All rights reserved. No part of this publication may be reproduced, by any means whatsoever, without the prior written permission of ITU.

Table of Contents

	Page
1 Scope.....	1
2 References.....	2
3 Definitions	2
3.1 Terms defined elsewhere	2
3.2 Terms defined in this Recommendation.....	2
4 Abbreviations and acronyms	2
5 Conventions	3
6 Test suite structure (TSS)	4
7 Electronic attachment	6
Annex A – Test purposes (TPs).....	7
A.1 TP definition conventions.....	7
A.2 Subgroup 1.3.1: Weighing scales (WEG)	8
Bibliography.....	25

Electronic attachment: Protocol implementation conformance statements (PICS) and the protocol implementation extra information for testing (PIXIT) required for the implementation of Annex A.

Introduction

This Recommendation is a transposition of Continua Test Tool DG2013, Test Suite Structure & Test Purposes, PAN-LAN-TAN Interface; Part 5A: Device Specializations. Agent (Weighing Scale) (Version 1.4, 2014-01-24), that was developed by the Continua Health Alliance. A number of versions of this specification existed before transposition and these can be found in the table below.

Version	Date	Revision history
1.2	2012-10-05	Initial release for Test Tool DG2011. It is the same version as "TSS&TP_1.5_PAN-LAN_PART_5A_v1.2.doc" because new features included in [b-CDG 2011] do not affect the test procedures specified in this document.
1.3	2013-05-24	Initial release for Test Tool DG2012. It uses "TSS&TP_DG2011_PAN-LAN_PART_5A_v1.2.doc" as a baseline and it adds new features included in [b-CDG 2012] (max APDU size for GM, BCA and ECG).
1.4	2014-01-24	Initial release for Test Tool DG2013. It uses "TSS&TP_DG2012_PAN-LAN_PART_5A_v1.3.doc" as a baseline and it adds new features included in [ITU-T H.810]: <ul style="list-style-type: none">– Add Glucose Meter BLE– Add BLE SSP support– Add NFC new transport– Add INR Device Specialization

Recommendation ITU-T H.845.1

Conformance of ITU-T H.810 personal health devices: PAN/LAN/TAN interface Part 5A: Weighing scales: Agent

1 Scope

The scope of this Recommendation¹ is to provide a test suite structure and the test purposes (TSS & TP) for the PAN/LAN/TAN interface based on the requirements defined in Continua specifications. The objective of this test specification is to provide a high probability of air interface interoperability between different devices.

The TSS and TP for the PAN/LAN/TAN interface document have been divided into ten parts. Each part is listed below:

- **Part 1:** Optimized exchange protocol [IEEE 11073-20601A] Agent
- **Part 2:** Optimized exchange protocol [IEEE 11073-20601A] Manager
- **Part 3:** Continua design guidelines. Agent
- **Part 4:** Continua design guidelines. Manager
- **Part 5:** Device specializations. Agent. This document is divided into 14 subparts:
 - **Part 5A:** Weighing scales
 - **Part 5B:** Glucose meter
 - **Part 5C:** Pulse oximeter
 - **Part 5D:** Blood pressure monitor
 - **Part 5E:** Thermometer
 - **Part 5F:** Cardiovascular fitness and activity monitor
 - **Part 5G:** Strength fitness equipment
 - **Part 5H:** Independent living activity hub
 - **Part 5I:** Adherence monitor
 - **Part 5J:** Insulin pump (Future development)
 - **Part 5K:** Peak flow
 - **Part 5L:** Body composition analyser
 - **Part 5M:** Basic electrocardiograph
 - **Part 5N:** International normalized ratio monitor
- **Part 6:** Device specializations. Manager
- **Part 7:** Continua design guidelines. Agent BLE
- **Part 8:** Continua design guidelines. Manager BLE
- **Part 9:** Personal health devices transcoding white paper. Agent
- **Part 10:** Personal health devices transcoding white paper. Manager

¹ This Recommendation includes an electronic attachment with the protocol implementation extra information for testing (PIXIT) required for the implementation of Annex A.

2 References

The following ITU-T Recommendations and other references contain provisions which, through reference in this text, constitute provisions of this Recommendation. At the time of publication, the editions indicated were valid. All Recommendations and other references are subject to revision; users of this Recommendation are therefore encouraged to investigate the possibility of applying the most recent edition of the Recommendations and other references listed below. A list of the currently valid ITU-T Recommendations is regularly published. The reference to a document within this Recommendation does not give it, as a stand-alone document, the status of a Recommendation.

- [ITU-T H.810] Recommendation ITU-T H.810 (2013), *Interoperability design guidelines for personal health systems*.
- [IEEE 11073-20601A] IEEE 11073-20601A-2010, *IEEE Health informatics – Personal health device communication – Part 20601: Application profile – Optimized Exchange Protocol Amendment 1*.
- [ISO/IEEE 11073-104xx] ISO/IEEE 11073-104xx (in force), *Health informatics – Personal health device communication – Device specialization*.
NOTE – Shorthand to refer to the collection of device specialization standards that utilize [IEEE 11073-20601A], where xx can be any number from 01 to 99 inclusive.
- [ISO/IEEE 11073-10415] ISO/IEEE 11073-10415-2010, *Health Informatics – Personal health device communication – Part 10415: Device specialization – Weighing scale*.
- [ISO/IEEE 11073-20601] ISO/IEEE 11073-20601:2010, *Health informatics – Personal health device communication – Part 20601: Application profile – Optimized exchange protocol*.

3 Definitions

3.1 Terms defined elsewhere

This Recommendation uses the following terms defined elsewhere:

3.1.1 agent [IEEE 11073-20601A]: A node that collects and transmits personal health data to an associated manager.

3.1.2 manager [IEEE 11073-20601A]: A node receiving data from one or more agent systems. Some examples of managers include a cellular phone, health appliance, set top box, or a computer system.

3.2 Terms defined in this Recommendation

None.

4 Abbreviations and acronyms

This Recommendation uses the following abbreviations and acronyms:

ATS	Abstract Test Suite
CDG	Continua Design Guidelines
DUT	Device Under Test
GUI	Graphical User Interface
INR	International Normalized Ratio

IUT	Implementation Under Test
MDS	Medical Device System
NFC	Near Field Communication
PAN	Personal Area Network
PCO	Point of Control and Observation
PCT	Protocol Conformance Testing
PHD	Personal Healthcare Device
PHDC	Personal Healthcare Device Class
PHM	Personal Health Manager
PICS	Protocol Implementation Conformance Statement
PIXIT	Protocol Implementation extra Information for Testing
SDP	Service Discovery Protocol
SOAP	Simple Object Access Protocol
TCRL	Test Case Reference List
TCWG	Test and Certification Working Group
TP	Test Purpose
TSS	Test Suite Structure
USB	Universal Serial Bus

5 Conventions

The key words "SHALL", "SHALL NOT", "SHOULD", "SHOULD NOT", "MAY", "MAY NOT" in this document are to be interpreted as in [b-ETSI SR 001 262].

- SHALL is equivalent to 'must' or 'it is required to'.
- SHALL NOT is equivalent to 'must not' or 'it is not allowed'.
- SHOULD is equivalent to 'it is recommended to'.
- SHOULD NOT is equivalent to 'it is not recommended to'.
- MAY is equivalent to 'is permitted'.
- MAY NOT is equivalent to 'it is not required that'.

NOTE – The above-mentioned key words are capitalized for illustrative purposes only and they do not appear capitalized within this Recommendation.

Reference is made in the ITU-T H.800-series of Recommendations to different versions of the Continua Design Guidelines (CDG) by a specific designation. The list of terms that may be used in this Recommendation is provided in Table 1.

Table 1 – List of designations associated to the various versions of the CDG

CDG name	Transposed as	Version	Description	Designation
2013 plus errata	ITU-T H.810	4.1	CDG 2013 plus errata noting all ratified bugs.	–
2013	–	4.0	Release 2013 of the CDG including maintenance updates of CDG 2012 and additional guidelines that cover new functionalities.	Endorphin
2012 plus errata	–	3.1	CDG 2012 plus errata noting all ratified bugs [b-CDG 2012].	–
2012	–	3.0	Release 2012 of the CDG including maintenance updates of CDG 2011 and additional guidelines that cover new functionalities.	Catalyst
2011 plus errata	–	2.1	CDG 2011 integrated with identified errata.	–
2011	–	2.0	Release 2011 of the CDG including maintenance updates of CDG 2010 and additional guidelines that cover new functionalities [b-CDG 2011].	Adrenaline
2010 plus errata	–	1.6	CDG 2010 integrated with identified errata.	-
2010	–	1.5	Release 2010 of the CDG with maintenance updates of CDG Version 1 and additional guidelines that cover new functionalities [b-CDG 2010].	1.5
1.0	–	1.0	First released version of the CDG [b-CDG 1.0].	–

6 Test suite structure (TSS)

The test purposes (TPs) for the PAN/LAN/TAN interface have been divided into the main subgroups specified below. Annex A describes the TPs for subgroup 1.3.1 (shown in bold).

- Group 1: Agent (AG)
 - Group 1.1: Transport (TR)
 - Subgroup 1.1.1: Design guidelines: Common (DGC)
 - Subgroup 1.1.2: USB design guidelines (UDG)
 - Subgroup 1.1.3: Bluetooth design guidelines (BDG)
 - Subgroup 1.1.4: Pulse oximeter design guidelines (PODG)
 - Subgroup 1.1.5: Cardiovascular design guidelines (CVDG)
 - Subgroup 1.1.6: Activity hub design guidelines (HUBDG)
 - Subgroup 1.1.7: ZigBee design guidelines (ZDG)
 - Subgroup 1.1.8: Glucose meter design guidelines (GLDG)
 - Subgroup 1.1.9: Bluetooth low energy design guidelines (BLEDG)
 - Subgroup 1.1.10: Basic electrocardiograph design guidelines (ECGDG)
 - Subgroup 1.1.11: NFC design guidelines (NDG)

- Group 1.2: 20601: Optimized exchange protocol (OXP)
 - Subgroup 1.2.1: PHD domain information model (DIM)
 - Subgroup 1.2.2: PHD service model (SER)
 - Subgroup 1.2.3: PHD communication model (COM)
- Group 1.3: Devices class specializations (CLASS)
 - **Subgroup 1.3.1:** Weighing scales (WEG)
 - Subgroup 1.3.2: Glucose meter (GL)
 - Subgroup 1.3.3: Pulse oximeter (PO)
 - Subgroup 1.3.4: Blood pressure monitor (BPM)
 - Subgroup 1.3.5: Thermometer (TH)
 - Subgroup 1.3.6: Cardiovascular (CV)
 - Subgroup 1.3.7: Strength (ST)
 - Subgroup 1.3.8: Activity hub (HUB)
 - Subgroup 1.3.9: Adherence monitor (AM)
 - Subgroup 1.3.10: Insulin pump (IP) (Future development)
 - Subgroup 1.3.11: Peak flow (PF)
 - Subgroup 1.3.12: Body composition analyser (BCA)
 - Subgroup 1.3.13: Basic electrocardiograph (ECG)
 - Subgroup 1.3.14: International normalized ratio (INR)
- Group 1.4: Personal health device transcoding whitepaper (PHDTW)
 - Subgroup 1.4.1: General requirements (GEN)
 - Subgroup 1.4.2: Thermometer requirements (TH)
 - Subgroup 1.4.3: Blood pressure requirements (BPM)
 - Subgroup 1.4.4: Heart rate requirements (HR)
 - Subgroup 1.4.5: Glucose meter requirements (GL)
- Group 2: Manager (MAN)
 - Group 2.1: Transport (TR)
 - Subgroup 2.1.1: Design guidelines: Common (DGC)
 - Subgroup 2.1.2: USB design guidelines (UDG)
 - Subgroup 2.1.3: Bluetooth design guidelines (BDG)
 - Subgroup 2.1.4: Cardiovascular design guidelines (CVDG)
 - Subgroup 2.1.5: Activity hub design guidelines (HUBDG)
 - Subgroup 2.1.6: ZigBee design guidelines (ZDG)
 - Subgroup 2.1.7: Bluetooth low energy design guidelines (BLEDG)
 - Subgroup 2.1.8: NFC design guidelines (NDG)
 - Group 2.2: 20601: Optimized exchange protocol (OXP)
 - Subgroup 2.2.1: General (GEN)
 - Subgroup 2.2.2: PHD domain information model (DIM)
 - Subgroup 2.2.3: PHD service model (SER)
 - Subgroup 2.2.4: PHD communication model (COM)
 - Group 2.3: Devices class specializations (CLASS)

- Subgroup 2.3.1: Weighing scales (WEG)
- Subgroup 2.3.2: Glucose meter (GL)
- Subgroup 2.3.3: Pulse oximeter (PO)
- Subgroup 2.3.4: Blood pressure monitor (BPM)
- Subgroup 2.3.5: Thermometer (TH)
- Subgroup 2.3.6: Cardiovascular (CV)
- Subgroup 2.3.7: Strength (ST)
- Subgroup 2.3.8: Activity hub (HUB)
- Subgroup 2.3.9: Adherence monitor (AM)
- Subgroup 2.3.10: Insulin pump (IP) (Future development)
- Subgroup 2.3.11: Peak flow (PF)
- Subgroup 2.3.12: Body composition analyser (BCA)
- Subgroup 2.3.13: Basic electrocardiograph (ECG)
- Subgroup 2.3.14: International normalized ratio (INR)
- Group 2.4: Personal health device transcoding whitepaper (PHDTW)
 - Subgroup 2.4.1: General requirements (GEN)
 - Subgroup 2.4.2: Thermometer requirements (TH)
 - Subgroup 2.4.3: Blood pressure requirements (BPM)
 - Subgroup 2.4.4: Heart rate requirements (HR)
 - Subgroup 2.4.5: Glucose meter requirements (GL)

7 Electronic attachment

The protocol implementation conformance statements (PICS) and the protocol implementation extra information for testing (PIXIT) required for the implementation of Annex A can be downloaded from <http://handle.itu.int/11.1002/2000/12067>.

In the electronic attachment, letters "C" and "I" in the column labelled "Mandatory" are used to distinguish between "PICS" and "PIXIT" respectively during testing. If the cell is empty, the corresponding PICS is "independent". If the field contains a "C", the corresponding PICS is dependent on other PICS, and the logical expression is detailed in the "SCR_Expression" field. The static conformance review (SCR) is used in the test tool to assert whether the PICS selection is consistent.

Annex A

Test purposes (TPs)

(This annex forms an integral part of this Recommendation.)

A.1 TP definition conventions

The test purposes are defined according to the following rules:

- **TP Id:** This is a unique identifier (TP/<TT>/<DUT>/<GR>/<SGR>/<XX> – <NNN>). It is specified according to the naming convention defined below:
 - Each test purpose identifier is introduced by the prefix "TP".
 - <TT>: This is the test tool that will be used in the test case.
 - PAN: Personal area network (Bluetooth or USB)
 - LAN: Local area network (ZigBee)
 - PAN-LAN: Personal area network (Bluetooth or USB) - Local area network (ZigBee)
 - LP-PAN: Low power personal area network (Bluetooth low energy)
 - TAN: Touch area network (NFC)
 - PLT: Personal area network (Bluetooth or USB) – Local area network (ZigBee) – Touch area network (NFC)
 - <DUT>: This is the device under test.
 - AG: PAN/LAN Agent
 - MAN: PAN/LAN Manager
 - <GR>: This identifies a group of test cases.
 - <SGR>: This identifies a subgroup of test cases.
 - <XX>: This identifies the type of testing.
 - BV: Valid Behaviour Test
 - BI: Invalid Behaviour Test
 - <NNN>: This is a sequential number that identifies the test purpose.
- **TP label:** This is the TP's title.
- **Coverage:** This contains the specification reference and clause to be checked by the TP.
 - Spec: This indicates the earliest version of the specification from which the testable items to be checked by the TP were included.
 - Testable item: This contains testable items to be checked by the TP.
- **Test purpose:** This is a description of the requirements to be tested.
- **Applicability:** This contains the PICS items that define if the test case is applicable or not for a specific device. When a TP contains an "ALL" in this field it means that it applies to the device under test within that scope of the test (specialization, transport used, etc.).
- **Initial condition:** This indicates the state to which the DUT needs to be moved at the beginning of TC execution.
- **Test procedure:** This describes the steps to be followed in order to execute the test case.
- **Pass/Fail criteria:** This provides criteria to decide whether the DUT passes or fails the test case.

A.2 Subgroup 1.3.1: Weighing scales (WEG)

TP Id		TP/PLT/AG/CLASS/WEG/BV-000		
TP label		MDS Object for Standard - Extended Configuration		
Coverage	Spec	[ISO/IEEE 11073-10415]		
	Testable items	MDSClassAttr 1; M	MDSClassAttr 2; M	MDSClassAttr 3; M
		MDSClassAttr 4; M	MDSClassAttr 5; R	MDSClassAttr 6; R
		MDSClassAttr 7; R	MDSClassAttr 8; M	MDSServices 1; M
Weighing.InfoExt 3; C		Weighing.Oper 1; M		
Applicability		C_AG_OXP_174 AND C_AG_OXP_000		
Initial condition		The simulated manager and the agent under test are in the operating state.		
Test procedure		<p>1. The simulated manager issues a "roiv-cmip-get" command with the handle set to 0 (to request the MDS object) and the attribute-id-list set to 0 to indicate all attributes.</p> <p>2. The agent responds with a "rors-cmip-get" service message in which the attribute-list contains a list of all implemented attributes of the MDS object:</p> <p style="margin-left: 40px;">MDS attributes:</p> <p style="margin-left: 40px;">a. Mandatory attribute Dev-Configuration-Id</p> <p style="margin-left: 80px;"><input type="checkbox"/> If NOT C_AG_OXP_181 then attribute-value = 0x05DC (1500)</p> <p style="margin-left: 80px;"><input type="checkbox"/> If C_AG_OXP_181 then attribute-value = < between 0x4000 and 0x7FFF ></p> <p style="margin-left: 40px;">b. Attribute System-Type must not be present.</p> <p style="margin-left: 40px;">c. Mandatory attribute System-model</p> <p style="margin-left: 80px;"><input type="checkbox"/> attribute-id = MDC_ATTR_ID_MODEL (0x09 0x28)</p> <p style="margin-left: 80px;"><input type="checkbox"/> attribute-type = SystemModel</p> <p style="margin-left: 80px;"><input type="checkbox"/> attribute-value.length = <Variable></p> <p style="margin-left: 80px;"><input type="checkbox"/> attribute-value = {Manufacturer, Model}</p> <p style="margin-left: 40px;">d. If recommended Power-Status attribute is present:</p> <p style="margin-left: 80px;"><input type="checkbox"/> attribute-id = MDC_ATTR_POWER_STAT</p> <p style="margin-left: 80px;"><input type="checkbox"/> attribute-type = PowerStatus</p> <p style="margin-left: 80px;"><input type="checkbox"/> attribute-value.length = 2 bytes</p> <p style="margin-left: 80px;"><input type="checkbox"/> attribute-value =</p> <p style="margin-left: 120px;">▪ ON_MAINS (0x8000) or ON_BATTERY(0x4000)</p> <p style="margin-left: 80px;">Only one of the following may be active:</p> <p style="margin-left: 120px;">▪ chargingFull(8),</p> <p style="margin-left: 120px;">▪ chargingTrickle(9),</p> <p style="margin-left: 120px;">▪ chargingOff(10).</p> <p style="margin-left: 40px;">e. If recommended Battery-Level attribute is present:</p> <p style="margin-left: 80px;"><input type="checkbox"/> attribute-id = MDC_ATTR_VAL_BATT_CHARGE</p> <p style="margin-left: 80px;"><input type="checkbox"/> attribute-type = BITS-16</p> <p style="margin-left: 80px;"><input type="checkbox"/> attribute-value.length = 2 bytes</p> <p style="margin-left: 80px;"><input type="checkbox"/> attribute-value = <value between 0 and 100> If value >100, the meaning of the value is "undefined"</p>		

	<p>f. If recommended Remaining-Battery-Time attribute is present:</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_BATT_REMAIN <input type="checkbox"/> attribute-type = BatMeasure <input type="checkbox"/> attribute-value.length = 6 bytes <input type="checkbox"/> attribute-value = <4 bytes to define the value. 2 remaining bytes to define the units, which shall be set to one of: MDC_DIM_MIN (0x08 0xA0), MDC_DIM_HR (0x08 0xC0), MDC_DIM_DAY (0x08 0xE0) > <p>g. Mandatory attribute System-Type-Spec-List</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SYS_TYPE_SPEC_LIST <input type="checkbox"/> attribute-type = TypeVerList <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = {MDC_DEV_SPEC_PROFILE_SCALE (0x10 0x0F), 1}
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/AG/CLASS/WEG/BV-001			
TP label	Body Weight Object for Standard Configuration			
Coverage	Spec	[ISO/IEEE 11073-10415]		
	Testable items	Concepts 2; M	WeightNumClass 2; M	WeightNumClass 3; M
		WeightNumClass 4; M	WeightNumClass 6; R	WeightNumClass 8; M
		WeightNumClass 10; R	WeightNumClass 12; O	WeightNumClass 14; R
		WeightNumClass 16; R	WeightNumClass 18; R	WeightNumClass 20; M
		WeightNumClass 22; M	WeightNumClass 23; R	WeightNumClass 25; R
		WeightNumClass 27; C	WeightNumClass 28; R	WeightNumClass 30; C
Applicability	C_AG_OXP_174 AND (NOT C_AG_OXP_181) AND C_AG_OXP_000			
Initial condition	The simulated manager and the agent under test have been associated, but the agent configuration is unknown for the simulated manager, so the agent and the simulated manager will be in the configuring state.			
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 4. Check that the field Dev-Config-Id is set to 0x05DC (1500); if it is not, the manager responds with an "unsupported-config" and wait for a new configuration. 5. Once the agent under test sends a standard configuration, do the next steps. 6. Check that the only object present in the configuration shall be the Body Weight. 7. Body Weight object attributes must be: <ol style="list-style-type: none"> a. Mandatory attribute Handle 			

	<ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_HANDLE <input type="checkbox"/> attribute-type = HANDLE <input type="checkbox"/> attribute-value = 1 b. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = 0x00 0x02(MDC_PART_SCADA) , 0xE1 0x40(MDC_MASS_BODY_ACTUAL 57664) c. Mandatory attribute Metric-Spec-Small (for standard and extended configuration) <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall (BITS-16) <input type="checkbox"/> attribute-value ≠ 0x00 0x00 <ul style="list-style-type: none"> ▪ bit 0 (mss-avail-intermittent(0)) must be set ▪ bit 1 (mss-avail-stored-data(1)) must be set ▪ bit 2 (mss-upd-aperiodic(2)) must be set ▪ bit 3 (mss-msmt-aperiodic(3)) must be set ▪ bit 9 (mss-acc-agent-initiated) must be set d. Mandatory attribute Unit-Code <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE <input type="checkbox"/> attribute-type = OID-Type <input type="checkbox"/> attribute-value.length = INT-U16 <input type="checkbox"/> attribute-value = MDC_DIM_KILO_G e. Mandatory attribute Attribute-Value-Map <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ATTRIBUTE_VAL_MAP <input type="checkbox"/> attribute-type = AttrValMap (sequence of attribute-id(OID-Type) and attribute-length(INT-U16)) <input type="checkbox"/> attribute-value map.length = 8 bytes <input type="checkbox"/> attribute-value = MDC_ATTR_NU_VAL_OBS_SIMPLE MDC_ATTR_TIME_STAMP_ABS <p>8. Check that there are no more present attributes in the initial configuration.</p>
Pass/Fail criteria	All checked values are as specified in the test procedure. Besides, the only object present in the configuration shall be the Body Weight.
Notes	

TP Id	TP/PLT/AG/CLASS/WEG/BV-002			
TP label	Body Weight Object for Extended Configuration			
Coverage	Spec	[ISO/IEEE 11073-10415]		
	Testable items	Concepts 2; M	WeightNumClass 2; M	WeightNumClass 5; M
		WeightNumClass 7; R	WeightNumClass 9; M	WeightNumClass 11; R
		WeightNumClass 13; R	WeightNumClass 15; R	WeightNumClass 17; R
		WeightNumClass 19; R	WeightNumClass 21; M	WeightNumClass 24; R
		WeightNumClass 26; R	WeightNumClass 29; R	

Applicability	C_AG_OXP_174 AND C_AG_OXP_181 AND C_AG_OXP_000
Initial condition	The simulated manager and the agent under test have been associated, but the agent configuration is unknown for the simulated manager, so the agent and the simulated manager will be in the configuring state.
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 4. Check that the field Dev-Config-Id is in the extended range. If it is not, the simulated manager must respond with an "unsupported-config" and wait for a new configuration. Repeat this step until a Dev-config-Id in the extended range is received. 5. Once the agent under test sends an extended configuration, Body Weight object attributes must be: <ol style="list-style-type: none"> a. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = 0x00 0x02(MDC_PART_SCADA) , 0xE1 0x40(MDC_MASS_BODY_ACTUAL 57664) b. If Supplemental-Types is present: <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SUPPLEMENTAL_TYPES <input type="checkbox"/> attribute-type = SupplementalTypeList <input type="checkbox"/> attribute-value.length = Sequence of TYPE (TYPE.length= 4 bytes) <input type="checkbox"/> attribute-value = <Not relevant for this test> c. Mandatory attribute Metric-Spec_Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall (BITS-16) <input type="checkbox"/> attribute-value ≠ 0x00 0x00 <ul style="list-style-type: none"> ▪ Bit 0 (mss-avail-intermittent(0)) must be set. ▪ Bit 1 (mss-avail-stored-data(1)) must be set. ▪ Bit 2 (mss-upd-aperiodic(2)) must be set. ▪ Bit 3 (mss-msmt-aperiodic(3)) must be set. ▪ Bit 9 (mss-acc-agent-initiated(9)) must be set. d. If attribute Metric-Structure-Small is present: <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_STRUCTURE_SMALL <input type="checkbox"/> attribute-type = MetricStructureSmall <input type="checkbox"/> attribute-value.length = Sequence of (ms-struct.length =1byte(INT-U8) + ms-comp-no =1byte(INT-U8)) <input type="checkbox"/> attribute-value = <Not relevant for this test> e. If attribute Measurement-Status is present : <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_MSMT_STAT <input type="checkbox"/> attribute-type = MeasurementStatus <input type="checkbox"/> attribute-value.length = 2 bytes.(BITS-16) <input type="checkbox"/> attribute-value = <Not relevant for this test> f. If attribute Metric-Id is present: <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_PHYSIO <input type="checkbox"/> attribute-type = OID-Type

	<p><input type="checkbox"/> attribute-value.length =2 bytes</p> <p>g. Only one attribute of Metric-Id and Metric-Id-List shall be present.</p> <p>h. If attribute Metric-Id-List is present:</p> <p><input type="checkbox"/> attribute-id = MDC_ATTR_ID_PHYSIO_LIS</p> <p><input type="checkbox"/> attribute-type = MetricIdList</p> <p><input type="checkbox"/> attribute-value.length= SEQUENCE OF OID-Type (INT-U16)</p> <p><input type="checkbox"/> attribute-value = <Not relevant for this test></p> <p><input type="checkbox"/> The [Metric-Id-List] attribute shall be used if a compound observed value is used, which does not incorporate the Metric-Id directly. The order of the Metric-Id-List shall correspond to the order of the elements in the compound observed value. Only one attribute of Metric-Id and Metric-Id-List shall be present.</p> <p>i. If attribute Metric-Id-Part is present:</p> <p><input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_ID_PART</p> <p><input type="checkbox"/> attribute-type = NomPartition</p> <p><input type="checkbox"/> attribute-value.length = INT-U16</p> <p><input type="checkbox"/> attribute-value = <Not relevant for this test></p> <p>j. Mandatory attribute Unit-Code</p> <p><input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE</p> <p><input type="checkbox"/> attribute-type = OID-Type</p> <p><input type="checkbox"/> attribute-value.length = INT-U16</p> <p><input type="checkbox"/> attribute-value = MDC_DIM_KILO_G OR MDC_DIM_LB</p> <p>k. If attribute Source-Handle-Reference is present:</p> <p><input type="checkbox"/> attribute-id = MDC_ATTR_SOURCE_HANDLE_REF</p> <p><input type="checkbox"/> attribute-type = HANDLE</p> <p><input type="checkbox"/> attribute-value.length = INT-U16</p> <p><input type="checkbox"/> attribute-value = <Not relevant for this test></p> <p>l. If attribute Measure-Active-Period is present:</p> <p><input type="checkbox"/> attribute-id = MDC_ATTR_TIME_PD_MSMT_ACTIVE</p> <p><input type="checkbox"/> attribute-type = FLOAT-Type</p> <p><input type="checkbox"/> attribute-value.length = INT-U32</p> <p><input type="checkbox"/> attribute-value = <Not relevant for this test></p> <p>m. If attribute Accuracy is present:</p> <p><input type="checkbox"/> attribute-id = MDC_ATTR_NU_ACCUR_MSMT</p> <p><input type="checkbox"/> attribute-type = FLOAT-Type (INT-U32)</p> <p><input type="checkbox"/> attribute-value.length = 4 bytes</p> <p><input type="checkbox"/> attribute-value = <Not relevant for this test></p>
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id		TP/PLT/AG/CLASS/WEG/BV-003		
TP label		Body Height Object		
Coverage	Spec	[ISO/IEEE 11073-10415]		
	Testable items	Concepts 4; M	HeightNumClass 1; O	HeightNumClass 4; M
		HeightNumClass 5; R	HeightNumClass 6; M	HeightNumClass 7; R
		HeightNumClass 8; R	HeightNumClass 9; R	HeightNumClass 10; R
		HeightNumClass 11; R	HeightNumClass 12; M	HeightNumClass 13; R
		HeightNumClass 14; R	HeightNumClass 15; R	
Applicability		C_AG_OXP_174 AND C_AG_OXP_181 AND C_AG_WEG_059 AND C_AG_OXP_000		
Initial condition		The simulated manager and the agent under test have been associated, but the agent configuration is unknown for the simulated manager, so the agent and the simulated manager will be in the configuring state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 4. Check that the field Dev-Config-Id is in the extended range; if it is not, the simulated manager must respond with an "unsupported-config" and wait for a new configuration. 5. Once the agent under test sends an extended configuration, check that the Height object attributes are: <ol style="list-style-type: none"> a. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value=0x00 0x02(MDC_PART_SCADA) , 0xE1 0x44(MDC_LEN_BODY_ACTUAL 57668) b. If not recommended attribute Supplemental-Types is present: <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SUPPLEMENTAL_TYPES <input type="checkbox"/> attribute-type = SupplementalTypeList <input type="checkbox"/> attribute-value.length = SEQUENCE OF (SIZE 4) <input type="checkbox"/> attribute-value = <Not relevant for this test> c. Mandatory attribute Metric-Spec_Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall (BITS-16) <input type="checkbox"/> attribute-value ≠ 0x00 0x00 <ul style="list-style-type: none"> ▪ Bit 0 (mss-avail-intermittent(0)) must be set. ▪ Bit 1 (mss-avail-stored-data(1)) must be set. ▪ Bit 2 (mss-upd-aperiodic(2)) must be set. ▪ Bit 3 (mss-msmt-aperiodic(3)) must be set. ▪ Bit 9 (mss-acc-agent-initiated(9)) must be set ▪ Bit 12 (mss-cat-manual(12)) Must be set, the metric is acquired manually. 		

- d. If not recommended attribute Metric-Structure-Small is present:
- attribute-id = MDC_ATTR_METRIC_STRUCTURE_SMALL
 - attribute-type = MetricStructureSmall
 - attribute-value.length = 2 bytes
 - attribute-value =
 - ms-struct = one of the following:
 - ms-struct-simple (0x01)
 - ms-struct-compound (0x02)
 - ms-struct-reserved (0x03)
 - ms-struct-compound-simple (0x04)
 - ms-compound-no = one of the following:
 - If ms-struct = ms-struct-simple THEN = 0
 - ELSE = maximum number of components in a compound value
- e. If recommended attribute Measurement-Status is present:
- attribute-id = MDC_ATTR_MSMT_STAT
 - attribute-type = MeasurementStatus
 - attribute-value.length = 2 bytes
 - attribute-value = <Not relevant for this test>
- f. Only one attribute of Metric-Id and Metric-Id-List shall be present.
- g. If not recommended attribute Metric-Id is present:
- attribute-id = MDC_ATTR_ID_PHYSIO
 - attribute-type = OID-Type
 - attribute-value.length = INT-U16
- h. If not recommended attribute Metric-Id-List is present:
- attribute-id = MDC_ATTR_ID_PHYSIO_LIS
 - attribute-type = MetricIdList
 - attribute-value.length = SEQUENCE OF OID-Type (INT-U16)
 - attribute-value =
 - The [Metric-Id-List] attribute shall be used if a compound observed value is used, which does not incorporate the Metric-Id directly. The order of the Metric-Id-List shall correspond to the order of the elements in the compound observed value. Only one attribute of Metric-Id and Metric-Id-List shall be present.
- i. If not recommended attribute Metric-Id-Partition is present:
- attribute-id = MDC_ATTR_METRIC_ID_PART
 - attribute-type = NomPartition
 - attribute-value.length = INT-U16
- j. Mandatory attribute Unit-Code
- attribute-id = MDC_ATTR_UNIT_CODE
 - attribute-type = OID-Type
 - attribute-value.length = INT-U16
 - attribute-value = MDC_DIM_CENTI_M or MDC_DIM_INCH
- k. If not recommended attribute Source-Handle-Reference is present:
- attribute-id = MDC_ATTR_SOURCE_HANDLE_REF
 - attribute-type = HANDLE

	<ul style="list-style-type: none"> <input type="checkbox"/> attribute-value.length = INT-U16 l. If not recommended attribute Measurement-Active-Period: <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_PD_MSMT_ACTIVE <input type="checkbox"/> attribute-type = FLOAT-Type <input type="checkbox"/> attribute-value.length = INT-U32 m. If recommended attribute Accuracy is present: <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_ACCUR_MSMT <input type="checkbox"/> attribute-type = FLOAT-Type (INT-U32) <input type="checkbox"/> attribute-value.length = FLOAT-Type (INT-U32)
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/AG/CLASS/WEG/BV-004			
TP label	Body Mass Index Object			
coverage	Spec	[ISO/IEEE 11073-10415]		
	Testable items	MassNumClass 1; O	MassNumClass 3; M	MassNumClass 4; R
		MassNumClass 5; M	MassNumClass 6; R	MassNumClass 7; R
		MassNumClass 8; R	MassNumClass 9; R	MassNumClass 10; R
		MassNumClass 11; M	MassNumClass 12; M	MassNumClass 13; R
		MassNumClass 14; R		
Applicability	C_AG_OXP_174 AND C_AG_WEG_056 AND C_AG_OXP_181 AND C_AG_OXP_000			
Initial condition	The simulated manager and the agent under test have been associated, but the agent configuration is unknown for the simulated manager, so the agent and the simulated manager will be in the configuring state.			
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 4. Check that the field Dev-Config-Id is in the extended range; if it is not, the simulated manager must respond with an "unsupported-config" and wait for a new configuration. 5. Once the agent under test sends an extended configuration, check that Body Mass Index object attributes are: <ol style="list-style-type: none"> a. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value=0x00 0x02(MDC_PART_SCADA) , 0xE1 0x50(MDC_RATIO_MASS_BODY_LEN_SQ 57680) b. If not recommended attribute Supplemental-Types: <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SUPPLEMENTAL_TYPES <input type="checkbox"/> attribute-type = SupplementalTypeList 			

- ❑ attribute-value.length = Sequence of TYPE (TYPE.length= 4 bytes)
- c. Mandatory attribute Metric-Spec_Small
- ❑ attribute-id = MDC_ATTR_METRIC_SPEC_SMALL
 - ❑ attribute-type = MetricSpecSmall (BITS-16)
 - ❑ attribute-value ≠ 0x00 0x00
 - Bit 0 (mss-avail-intermittent(0)) must be set.
 - Bit 1 (mss-avail-stored-data(1)) must be set.
 - Bit 2 (mss-upd-aperiodic(2)) must be set.
 - Bit 3 (mss-msmt-aperiodic(3)) must be set.
 - Bit 9 (mss-acc-agent-initiated(9)) must be set.
 - If bit 14 is set(mss-cat_calculation(14)) is set, the metric represents a calculated value.
- d. If not recommended attribute Metric-Structure-Small:
- ❑ attribute-id = MDC_ATTR_METRIC_STRUCTURE_SMALL
 - ❑ attribute-type = MetricStructureSmall
 - ❑ attribute-value.length = 2 bytes
 - ❑ attribute-value
 - ms-struct = one of the following:
 - ms-struct-simple (0x01)
 - ms-struct-compound (0x02)
 - ms-struct-reserved (0x03)
 - ms-struct-compound-simple (0x04)
 - ms-compound-no = one of the following:
 - If ms-struct = ms-struct-simple THEN = 0
 - ELSE = maximum number of components in a compound value
- e. If recommended attribute Measurement-Status is present:
- ❑ attribute-id = MDC_ATTR_MSMT_STAT
 - ❑ attribute-type = MeasurementStatus
 - ❑ attribute-value.length = 2 bytes
 - ❑ attribute-value = <Not relevant for this test>
- f. Only one attribute of Metric-Id and Metric-Id-List shall be present.
- g. If not recommended attribute Metric-Id is present:
- ❑ attribute-id = MDC_ATTR_ID_PHYSIO
 - ❑ attribute-type = OID-Type
 - ❑ attribute-value.length = INT-U16
 - ❑ attribute-value =
- h. If not recommended attribute Metric-Id-List is present:
- ❑ attribute-id = MDC_ATTR_ID_PHYSIO_LIS
 - ❑ attribute-type = MetricIdList
 - ❑ attribute-value.length = SEQUENCE OF OID-Type (INT-U16)
 - ❑ attribute-value =
 - ❑ The [Metric-Id-List] attribute shall be used if a compound observed value is used, which does not incorporate the Metric-Id directly. The order of the Metric-Id-List shall correspond to the order of the elements in the

compound observed value. Only one attribute of Metric-Id and Metric-Id-List shall be present.

- i. If not recommended attribute Metric-Id -Partition is present:
 - attribute-id = MDC_ATTR_METRIC_ID_PART
 - attribute-type = NomPartition
 - attribute-value.length = INT-U16
 - attribute-value = one of the next
 - nom-part-unspec (0x00 0x00)
 - nom-part-obj (0x00 0x01)
 - nom-part-metric (0x00 0x02)
 - nom-part-alert (0x00 0x03)
 - nom-part-dim (0x00 0x04)
 - nom-part-vattr (0x00 0x05)
 - nom-part-pgrp (0x00 0x06)
 - nom-part-sites (0x00 0x07)
 - nom-part-infrac (0x00 0x08)
 - nom-part-fef (0x00 0x09)
 - nom-part-ecg-extn (0x00 0x0A)
 - nom-part-phd-dm (0x00 0x80)
 - nom-part-phd-hf (0x00 0x81)
 - nom-part-phd-ai (0x00 0x82)
 - nom-part-ret-code(0x00 0xFF)
 - nom-part-ext-nom (0x01 0x00)
 - nom-part-priv (0x04 0x00)
- j. Mandatory attribute Unit-Code
 - attribute-id = MDC_ATTR_UNIT_CODE
 - attribute-type = OID-Type
 - attribute-value.length = INT-U16
 - attribute-value = MDC_DIM_KG_PER_M_SQ
- k. If not recommended attribute Measure-Active-Period is present
 - attribute-id = MDC_ATTR_TIME_PD_MSMT_ACTIVE
 - attribute-type = FLOAT-Type
 - attribute-value.length = INT-U32
 - attribute-value =
- l. Mandatory attribute Source-Handle-Reference
 - attribute-id = MDC_ATTR_SOURCE_HANDLE_REF
 - attribute-type = HANDLE
 - attribute-value.length = INT-U16
 - attribute-value = It must be equal to the handle of another metric object in the configuration and it must point to an object that has a type of MDC_MASS_BODY_ACTUAL.
- m. If recommended attribute Accuracy is present:
 - attribute-id = MDC_ATTR_NU_ACCUR_MSMT
 - attribute-type = FLOAT-Type (INT-U32)
 - attribute-value.length = FLOAT-Type (INT-U32)

Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/AG/CLASS/WEG/BV-005		
TP label	Events for Update Data		
Coverage	Spec	[ISO/IEEE 11073-10415]	
	Testable items	MDSEvents 3; M	MDSEvents 4; M
MDSEvents 6; M		Weighing.Service 1; M	Weighing.Service 2; M
Applicability	C_AG_OXP_174 AND C_AG_OXP_000 AND (C_AG_OXP_182 OR C_AG_OXP_183 OR C_AG_OXP_184 OR C_AG_OXP_189)		
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<ol style="list-style-type: none"> 1. Take measurements for every supported object in the agent under test. 2. Wait to receive every event report and check: <ol style="list-style-type: none"> a. message <ul style="list-style-type: none"> ▪ field- type = Event Report ▪ field-length = 2 bytes ▪ field- value=0x01 0x01 (EventReportArgumentSimple, confirmed) <p>This field identifies the type of message sent by the agent; for the confirmed event configuration, roiv-cmip-confirmed-event-report.</p> 		
Pass/Fail criteria	<p>Check that every received report is a one of the following Data APDU and that it is confirmed.</p> <ul style="list-style-type: none"> • MDC_NOTI_SCAN_REPORT_FIXED • MDC_NOTI_SCAN_REPORT_MP_FIXED • MDC_NOTI_SCAN_REPORT_VAR • MDC_NOTI_SCAN_REPORT_MP_VAR 		
Notes			

TP Id	TP/PLT/AG/CLASS/WEG/BV-011		
TP label	Communication Model: Association Procedure		
Coverage	Spec	[ISO/IEEE 11073-10415]	
	Testable items	Weighing.Methods 5; M	Weighing.Association 1; M
Weighing.Association 5; M		Weighing.Association 10; M	Weighing.Association 11; M
Applicability	C_AG_OXP_174 AND C_AG_OXP_000		
Initial condition	The simulated manager and the agent under test are in the unassociated state.		

Test procedure	<p>1. The simulated manager receives an association request from the agent under test, the relevant field for this test:</p> <ul style="list-style-type: none"> a. functional – units <ul style="list-style-type: none"> <input type="checkbox"/> field- type = FunctionalUnits <input type="checkbox"/> field-length = 4 bytes <ul style="list-style-type: none"> ▪ Bit 0 must be 0. ▪ Bits 1 and 2 may be set ▪ The rest of the bits must not be set a. dev-config-id <ul style="list-style-type: none"> <input type="checkbox"/> field- type = ConfigId <input type="checkbox"/> field-length = INT-U16 <input type="checkbox"/> field- value = <ul style="list-style-type: none"> ▪ 0x05DC (1500) for standard configuration. ▪ Between 0x40 0x00 and 0x7F 0xFF for extended configuration. b. data-req-mode-flags (DataReqModeCapab) <ul style="list-style-type: none"> <input type="checkbox"/> field- type = DataReqModeFlags <input type="checkbox"/> field-length = BITS-16 <p>If the agent implements only the weighing scales specialization</p> <ul style="list-style-type: none"> <input type="checkbox"/> field- value = data-req-supp-init-agent will be set. c. data-req-init-agent-count (DataReqModeCapab) <ul style="list-style-type: none"> <input type="checkbox"/> field- type = INT-U8 <input type="checkbox"/> field-length = = INT-U8 <p>If the agent implements only this Device Specialization:</p> <ul style="list-style-type: none"> <input type="checkbox"/> field- value = 0x01
Pass/Fail criteria	All checked attributes have proper values.
Notes	

TP Id	TP/PLT/AG/CLASS/WEG/BV-012		
TP label	Configuring Procedure		
Coverage	Spec	[ISO/IEEE 11073-10415]	
	Testable items	Concepts 3; C	MDSEvents 1; M
Applicability	C_AG_OXP_174 AND C_AG_OXP_000		
Initial condition	The simulated manager and the agent under test are in the unassociated state.		

Test procedure

1. The simulated manager receives an association request from the agent under test.
2. The simulated manager responds with a result = accepted-unknown-config.
3. The agent responds with a "Remote Operation Invoke | Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager.
 - a. APDU Type
 - field- type = PrstApdu
 - field-length =2 bytes
 - field-value =0xE7 0x00
 - This value is for presentation APDU "prst" (PrstApdu).
 - b. The following two bytes indicates the length of the message (could be helpful to analyse the fields).
 - c. The following two bytes indicate the length of the OCTET STRING that contains the DataApdu (could be helpful to analyse the fields).
 - d. invoke-id
 - field- type = InvokeIDType
 - field-length =INT-U16
 - field- value=
 - This value identifies the message, the confirmed response that will be sent by the simulated manager shall have the same invoke-id.
 - e. Message
 - field- type =
 - field-length =two bytes
 - field- value=0x01 0x01 (EventReportArgumentSimple)
 - This field identifies the type of message sent by the agent; for the confirmed event configuration, roiv-cmip-confirmed-event-report.
 - f. The following two bytes indicate the length of the fields that make up the EventReportArgumentSimple.
 - g. obj-handle (EventReportArgumentSimple)
 - field- type = HANDLE
 - field-length =INT-U16
 - h. event-time (EventReportArgumentSimple)
 - field- type = Relative Time
 - field-length =INT-U32
 - If the agent does not support relative time:
 - field- value=0x FF 0x FF 0x FF 0x FF
 - i. event-type (EventReportArgumentSimple)
 - field- type = OID-Type
 - field-length =INT-U16
 - field- value=0x 0D 0x 1C (MDC_NOTI_CONFIG)
 - j. The following two bytes indicate the length for event-info (ConfigReport). This value shall not be 0. It is the start of ConfigReport.
 - k. config-report-id (ConfigReport)
 - field- type = ConfigId
 - field-length = INT-U16
 - field- value=
 - 0x 05DC for standard configuration.

	<ul style="list-style-type: none"> ▪ <between 0x40 0x00 and 0x7F 0xFF> for extended configuration. l. The following two bytes indicate the number of ConfigObjectList; this value shall not be 0. (The agent will have at least one ObjectList) m. The following two bytes indicate the length for ConfigObjectList, this value shall not be 0. n. obj-class (ConfigReport = ConfigObjectList (ConfigObject)) <ul style="list-style-type: none"> ❑ field- type = OID-Type ❑ field-length = INT-U16 ❑ field- value= 0 x00 0x06 (MDC_MOC_VMO_METRIC_NU --6) OR 0x00 0x25 (MDC_MOC_VMS_MDS_SIMP --37) ❑ MDS- Object and Numeric object are only required for the weighing scales agent, and for this reason only two classes identify codes that are possible. o. obj-handle (ConfigReport = ConfigObjectList (ConfigObject)) <ul style="list-style-type: none"> ❑ field- type = HANDLE ❑ field-length = INT-U16 ❑ field- value= p. The following two bytes indicate the number of Attributes; this value shall not be 0. (ConfigReport = ConfigObjectList (ConfigObject) = AttributeList) q. The following two bytes indicates the length (bytes) for the Attributes List; this value shall not be 0. r. attribute-id (ConfigReport = ConfigObjectList (ConfigObject) = Attribute List) <ul style="list-style-type: none"> ❑ field- type = OID-Type ❑ field-length = INT-U16 s. The following two bytes indicate the attribute-value.length, (i.e., length for the attribute value); this value shall not be 0. t. attribute-value(ConfigReport = ConfigObjectList (ConfigObject) Attribute List); this value depends on the attribute type. u. The last three fields will repeat until the number of bytes for Attributes List is completed. v. If Configuration is Standard then the body height numeric and body mass numeric object cannot be present. <ul style="list-style-type: none"> ❑ For every obj-class we have analysed in the message <ul style="list-style-type: none"> ▪ attribute-id = MDC_ATTR_ID_TYPE ▪ attribute-value !=MDC_PART_SCADA MDC_LEN_BODY_ACTUAL OR MDC_PART_SCADA MDC_RATIO_MASS_BODY_LEN_SQ
Pass/Fail criteria	Attributes of the configuration are properly formatted.
Notes	

TP Id	TP/PLT/AG/CLASS/WEG/BV-016		
TP label	Body mass and body height numeric objects		
Coverage	Spec	[ISO/IEEE 11073-10415]	
	Testable items	Concepts 5; C	Concepts 6; M
Applicability	C_AG_OXP_174 AND C_AG_WEG_056 AND C_AG_OXP_000		

Initial condition	The simulated manager and the agent under test are in the unassociated State.
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 4. Take weight and height measurements with the agent under test. 5. Wait for the simulated manager to receive an event report with the weight, height and BMI.
Pass/Fail criteria	<ul style="list-style-type: none"> ▪ Check PICS to make sure that if the BMI PIC is set, the Height PIC must also be set. ▪ If the BMI object is present, the Height object must be present too. ▪ Check that the BMI values are coherent with the height and weight values according to the formula defined in the specification.
Notes	

TP Id	TP/PLT/AG/CLASS/WEG/BV-017		
TP label	Config Changes Service. Contextual Attribute.		
Coverage	Spec	[ISO/IEEE 11073-10415]	
	Testable items	WeightNumClass 1;M	
	Spec	[ITU-T H.810]	
	Testable items	Communication 8; M	
Applicability	C_AG_OXP_174 AND C_AG_WEG_060 AND C_AG_OXP_000		
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<ol style="list-style-type: none"> 1. If an attribute that is going to be changed is reported in the fixed format event report, take some measurements with the agent under test. 2. Make a change to the contextual attribute Unit-Code for Body Weight Object (Pounds to kg or kg to pounds.) 3. The agent shall send an MDS event report indicating the new contextual attribute value. 4. Take some more measurements. 5. Wait for the manager to receive new event reports from the agent which reports the measurements from step 4. 		
Pass/Fail criteria	<ul style="list-style-type: none"> ▪ The agent sends an MDS event report to inform of the contextual attribute that has been changed. ▪ Data has changed in accordance with the new contextual attribute. 		
Notes			

TP Id		TP/PLT/AG/CLASS/WEG/BV-018		
TP label		Config Changes Service. Height object Contextual Attribute.		
Coverage	Spec	[ISO/IEEE 11073-10415]		
	Testable items	WeightNumClass 1;M		
Applicability		C_AG_OXP_174 AND C_AG_WEG_059 AND C_AG_WEG_061 AND C_AG_OXP_000		
Initial condition		The simulated manager and the agent under test are in the operating state.		
Test procedure		<ol style="list-style-type: none"> 1. If an attribute that is going to be changed is reported in the fixed format event report, take some measurements with the agent under test. 2. Make a change to the contextual attribute Unit-Code for Body Height Object (centimetres to inches or inches to centimetres). 3. The agent shall send an MDS event report indicating the new contextual attribute value. 4. Take some more measurements. 5. Wait for the manager to receive new event reports from the agent which reports the measurements from step 4. 		
Pass/Fail criteria		<ul style="list-style-type: none"> ▪ The agent sends an MDS event report to inform of the contextual attribute that has been changed. ▪ Data has changed in accordance with the new contextual attribute. 		
Notes				

TP Id		TP/PLT/AG/CLASS/WEG/BV-019		
TP label		Operating State. Manager to Agent Maximum APDU Size		
Coverage	Spec	[ISO/IEEE 11073-20601]		
	Testable items	CommonCharac 3; M		
	Spec	[ISO/IEEE 11073-10415]		
	Testable items	Weighing.CommModel 2; M		
Applicability		C_AG_OXP_000 AND C_AG_OXP_174		
Initial condition		The simulated manager and the agent are in the operating state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated manager issues the "Remote Operation Invoke Get" command with: <ol style="list-style-type: none"> a. Obj-handle set to 0 (to request the MDS object) b. attribute-id-list.count = 103 c. attribute-id-list: (MDC_ATTR_ID_MODEL, MDC_ATTR_SYS_ID, MDC_ATTR_DEV_CONFIG_ID) repeated 34 times followed by an additional MDC_ATTR_ID_MODEL 2. Check the response of the agent. 3. The simulated manager issues the "Remote Operation Invoke Get" command with the handle set to 0 (to request the MDS object) and an empty attribute-id-list to indicate all attributes. 4. Check the response of the agent. 		

<p>Pass/Fail criteria</p>	<ul style="list-style-type: none"> • In step 2, the agent under test may respond with a rors-cmip-get listing all the requested attributes, or with a roer message. If PICS C_AG_OXP_100 =TRUE and the agent does not respond with a rors-cmip-get message, it will respond with a roer message or rorj(resource-limitation) message; a WARNING will appear. <ul style="list-style-type: none"> ○ If the response is a get response, the total size of the response cannot exceed the sum of the APDU sizes of the supported specializations (limited to an absolute limit of 64512 octets): <ul style="list-style-type: none"> ▪ Pulse oximeter -> 9216 octets ▪ Weighing scales -> 896 octets ▪ Glucose meter -> 5120 octets or 64512 octets if agent supports PM-Store ▪ Blood pressure -> 896 octets ▪ Thermometer -> 896 octets ▪ Independent activity hub -> 5120 octets ▪ Cardiovascular -> 64512 octets or 6624 octets agent under test only supports Step Counter Profile ▪ Strength -> 64512 octets: ▪ Adherence monitor -> 1024 octets ▪ Peak flow -> 2030 octets ▪ Body composition analyser -> 7730 octets ▪ Basic ECG/Simple ECG -> 7168 octets or 64512 octets if agent supports PM-Store ▪ Basic ECG/Heart rate -> 1280 octets or 64512 octets if agent supports PM-Store ▪ International normalized ratio -> 896 octets or 64512 if agent supports PM-Store ○ In case it responds with a roer, the reason must not be protocol-violation (code value 23) • In step 4, the agent must respond with a rors-cmip-get message.
<p>Notes</p>	

Bibliography

- [b-CDG 1.0] Continua Health Alliance, Continua Design Guidelines v1.0 (2008), *Continua Design Guidelines*.
- [b-CDG 2010] Continua Health Alliance, Continua Design Guidelines v1.5 (2010), *Continua Design Guidelines*.
- [b-CDG 2011] Continua Health Alliance, Continua Design Guidelines (2011), "Adrenaline", *Continua Design Guidelines*.
- [b-CDG 2012] Continua Health Alliance CDG, Continua Design Guidelines (2012), "Catalyst", *Continua Design Guidelines*.
- [b-ETSI SR 001 262] ETSI SR 001 262 v1.8.1 (2003), *ETSI drafting rules*.

SERIES OF ITU-T RECOMMENDATIONS

Series A	Organization of the work of ITU-T
Series D	General tariff principles
Series E	Overall network operation, telephone service, service operation and human factors
Series F	Non-telephone telecommunication services
Series G	Transmission systems and media, digital systems and networks
Series H	Audiovisual and multimedia systems
Series I	Integrated services digital network
Series J	Cable networks and transmission of television, sound programme and other multimedia signals
Series K	Protection against interference
Series L	Construction, installation and protection of cables and other elements of outside plant
Series M	Telecommunication management, including TMN and network maintenance
Series N	Maintenance: international sound programme and television transmission circuits
Series O	Specifications of measuring equipment
Series P	Terminals and subjective and objective assessment methods
Series Q	Switching and signalling
Series R	Telegraph transmission
Series S	Telegraph services terminal equipment
Series T	Terminals for telematic services
Series U	Telegraph switching
Series V	Data communication over the telephone network
Series X	Data networks, open system communications and security
Series Y	Global information infrastructure, Internet protocol aspects and next-generation networks
Series Z	Languages and general software aspects for telecommunication systems