

ITU-T

TELECOMMUNICATION
STANDARDIZATION SECTOR
OF ITU

H.845.11

(01/2015)

SERIES H: AUDIOVISUAL AND MULTIMEDIA SYSTEMS

E-health multimedia services and applications –
Interoperability compliance testing of personal health
systems (HRN, PAN, LAN, TAN and WAN)

**Conformance of ITU-T H.810 personal health
devices: PAN/LAN/TAN interface Part 5K: Peak
expiratory flow monitor: Agent**

Recommendation ITU-T H.845.11

ITU-T H-SERIES RECOMMENDATIONS
AUDIOVISUAL AND MULTIMEDIA SYSTEMS

CHARACTERISTICS OF VISUAL TELEPHONE SYSTEMS	H.100–H.199
INFRASTRUCTURE OF AUDIOVISUAL SERVICES	
General	H.200–H.219
Transmission multiplexing and synchronization	H.220–H.229
Systems aspects	H.230–H.239
Communication procedures	H.240–H.259
Coding of moving video	H.260–H.279
Related systems aspects	H.280–H.299
Systems and terminal equipment for audiovisual services	H.300–H.349
Directory services architecture for audiovisual and multimedia services	H.350–H.359
Quality of service architecture for audiovisual and multimedia services	H.360–H.369
Telepresence	H.420–H.429
Supplementary services for multimedia	H.450–H.499
MOBILITY AND COLLABORATION PROCEDURES	
Overview of Mobility and Collaboration, definitions, protocols and procedures	H.500–H.509
Mobility for H-Series multimedia systems and services	H.510–H.519
Mobile multimedia collaboration applications and services	H.520–H.529
Security for mobile multimedia systems and services	H.530–H.539
Security for mobile multimedia collaboration applications and services	H.540–H.549
Mobility interworking procedures	H.550–H.559
Mobile multimedia collaboration inter-working procedures	H.560–H.569
BROADBAND, TRIPLE-PLAY AND ADVANCED MULTIMEDIA SERVICES	
Broadband multimedia services over VDSL	H.610–H.619
Advanced multimedia services and applications	H.620–H.629
Ubiquitous sensor network applications and Internet of Things	H.640–H.649
IPTV MULTIMEDIA SERVICES AND APPLICATIONS FOR IPTV	
General aspects	H.700–H.719
IPTV terminal devices	H.720–H.729
IPTV middleware	H.730–H.739
IPTV application event handling	H.740–H.749
IPTV metadata	H.750–H.759
IPTV multimedia application frameworks	H.760–H.769
IPTV service discovery up to consumption	H.770–H.779
Digital Signage	H.780–H.789
E-HEALTH MULTIMEDIA SERVICES AND APPLICATIONS	
Interoperability compliance testing of personal health systems (HRN, PAN, LAN, TAN and WAN)	H.820–H.859
Multimedia e-health data exchange services	H.860–H.869

For further details, please refer to the list of ITU-T Recommendations.

Recommendation ITU-T H.845.11

Conformance of ITU-T H.810 personal health devices: PAN/LAN/TAN interface Part 5K: Peak expiratory flow monitor: Agent

Summary

Recommendation ITU-T H.845.11 is a transposition of Continua Test Tool DG2013, Test Suite Structure & Test Purposes, PAN-LAN-TAN Interface; Part 5K: Device Specializations. Agent (Peak expiratory flow monitor) (Version 1.4, 2014-01-24), that was developed by the Continua Health Alliance. A number of versions of this specification existed before transposition.

This Recommendation includes an electronic attachment with the protocol implementation conformance statements (PICS) and the protocol implementation extra information for testing (PIXIT) required for the implementation of Annex A.

History

Edition	Recommendation	Approval	Study Group	Unique ID*
1.0	ITU-T H.845.11	2015-01-13	16	11.1002/1000/12271

* To access the Recommendation, type the URL <http://handle.itu.int/> in the address field of your web browser, followed by the Recommendation's unique ID. For example, <http://handle.itu.int/11.1002/1000/11830-en>.

FOREWORD

The International Telecommunication Union (ITU) is the United Nations specialized agency in the field of telecommunications, information and communication technologies (ICTs). The ITU Telecommunication Standardization Sector (ITU-T) is a permanent organ of ITU. ITU-T is responsible for studying technical, operating and tariff questions and issuing Recommendations on them with a view to standardizing telecommunications on a worldwide basis.

The World Telecommunication Standardization Assembly (WTSA), which meets every four years, establishes the topics for study by the ITU-T study groups which, in turn, produce Recommendations on these topics.

The approval of ITU-T Recommendations is covered by the procedure laid down in WTSA Resolution 1.

In some areas of information technology which fall within ITU-T's purview, the necessary standards are prepared on a collaborative basis with ISO and IEC.

NOTE

In this Recommendation, the expression "Administration" is used for conciseness to indicate both a telecommunication administration and a recognized operating agency.

Compliance with this Recommendation is voluntary. However, the Recommendation may contain certain mandatory provisions (to ensure, e.g., interoperability or applicability) and compliance with the Recommendation is achieved when all of these mandatory provisions are met. The words "shall" or some other obligatory language such as "must" and the negative equivalents are used to express requirements. The use of such words does not suggest that compliance with the Recommendation is required of any party.

INTELLECTUAL PROPERTY RIGHTS

ITU draws attention to the possibility that the practice or implementation of this Recommendation may involve the use of a claimed Intellectual Property Right. ITU takes no position concerning the evidence, validity or applicability of claimed Intellectual Property Rights, whether asserted by ITU members or others outside of the Recommendation development process.

As of the date of approval of this Recommendation, ITU had not received notice of intellectual property, protected by patents, which may be required to implement this Recommendation. However, implementers are cautioned that this may not represent the latest information and are therefore strongly urged to consult the TSB patent database at <http://www.itu.int/ITU-T/ipr/>.

© ITU 2015

All rights reserved. No part of this publication may be reproduced, by any means whatsoever, without the prior written permission of ITU.

Table of Contents

	Page
1 Scope.....	1
2 References.....	2
3 Definitions	2
3.1 Terms defined elsewhere	2
3.2 Terms defined in this Recommendation.....	2
4 Abbreviations and acronyms	2
5 Conventions	3
6 Test suite structure (TSS)	4
7 Electronic attachment	6
Annex A – Test purposes (TPs)	7
A.1 TP definition conventions.....	7
A.2 Subgroup 1.3.11: Peak expiratory flow monitor (PF).....	8
Bibliography.....	35

Electronic attachment: Protocol implementation conformance statements (PICS) and protocol implementation extra information for testing (PIXIT) required for the implementation of Annex A.

Introduction

This Recommendation is the transposition of Continua Test Tool DG2013, Test Suite Structure & Test Purposes, PAN-LAN-TAN Interface; Part 5K: Device Specializations. Agent (Peak expiratory flow monitor) (Version 1.4, 2014-01-24), that was developed by the Continua Health Alliance. A number of versions of this specification existed before transposition and these can be found in the table below.

Version	Date	Revision history
1.2	2012-10-05	Initial release for Test Tool DG2011. This is the same version as "TSS&TP_1.5_PAN-LAN_PART_5K_v1.2.doc" because new features included in [b-CDG 2011] do not affect the test procedures specified in this document.
1.3	2013-05-24	Initial release for Test Tool DG2012. This uses "TSS&TP_DG2011_PAN-LAN_PART_5K_v1.2.doc" as a baseline and adds new features included in [b-CDG 2012]: Max APDU size for GM, BCA and ECG.
1.4	2014-01-24	Initial release for Test Tool DG2013. This uses "TSS&TP_DG2012_PAN-LAN_PART_5K_v1.3.doc" as a baseline and adds new features included in [ITU-T H.810]: <ul style="list-style-type: none">• Adds glucose meter BLE• Adds BLE SSP support• Adds NFC new transport• Adds INR device specialization

Recommendation ITU-T H.845.11

Conformance of ITU-T H.810 personal health devices: PAN/LAN/TAN interface Part 5K: Peak expiratory flow monitor: Agent

1 Scope

The scope of this Recommendation¹ is to provide a test suite structure and the test purposes (TSS & TP) for the PAN/LAN/TAN interface based on the requirements defined in the Continua Design Guidelines (CDG) [ITU-T H.810]. The objective of this test specification is to provide a high probability of air interface interoperability between different devices.

The TSS and TP for the PAN/LAN/TAN interface document have been divided into 10 parts. Each part is listed below:

- **Part 1:** Optimized exchange protocol [ISO/IEEE 11073-20601A] Agent
- **Part 2:** Optimized exchange protocol [ISO/IEEE 11073-20601A] Manager
- **Part 3:** Continua design guidelines. Agent
- **Part 4:** Continua design guidelines. Manager
- **Part 5:** Device specializations. Agent. This document is divided into 12 subparts:
 - **Part 5A:** Weighing scales
 - **Part 5B:** Glucose meter
 - **Part 5C:** Pulse oximeter
 - **Part 5D:** Blood pressure monitor
 - **Part 5E:** Thermometer
 - **Part 5F:** Cardiovascular fitness and activity monitor
 - **Part 5G:** Strength fitness equipment
 - **Part 5H:** Independent living activity hub
 - **Part 5I:** Adherence monitor
 - **Part 5J:** Insulin pump (Future development)
 - **Part 5K:** Peak flow
 - **Part 5L:** Body composition analyser
 - **Part 5M:** Basic electrocardiograph
 - **Part 5N:** International normalized ratio monitor
- **Part 6:** Device specializations. Manager
- **Part 7:** Continua design guidelines. Agent BLE
- **Part 8:** Continua design guidelines. Manager BLE
- **Part 9:** Personal health devices transcoding whitepaper. Agent
- **Part 10:** Personal health devices transcoding whitepaper. Manager

¹ This Recommendation includes an electronic attachment with the protocol implementation conformance statements (PICS) and the protocol implementation extra information for testing (PIXIT) required for the implementation of Annex A.

2 References

The following ITU-T Recommendations and other references contain provisions which, through reference in this text, constitute provisions of this Recommendation. At the time of publication, the editions indicated were valid. All Recommendations and other references are subject to revision; users of this Recommendation are therefore encouraged to investigate the possibility of applying the most recent edition of the Recommendations and other references listed below. A list of the currently valid ITU-T Recommendations is regularly published. The reference to a document within this Recommendation does not give it, as a stand-alone document, the status of a Recommendation.

- [ITU-T H.810] Recommendation ITU-T H.810 (2013), *Interoperability design guidelines for personal health systems*.
- [ISO/IEEE 11073-20601A] ISO/IEEE 11073-20601:2010, *Health informatics – Personal health device communication – Part 20601: Application profile – Optimized exchange protocol*, including ISO/IEEE 11073-20601:2010 Amd 1:2015.
<http://www.iso.org/iso/home/store/catalogue_tc/catalogue_detail.htm?csnumber=54331>
with
<http://www.iso.org/iso/home/store/catalogue_tc/catalogue_detail.htm?csnumber=63972>
- [ISO/IEEE 11073-104xx] ISO/IEEE 11073-104xx (in force), *Health informatics – Personal health device communication – Device specialization*.
NOTE – Shorthand is used to refer to the collection of device specialization standards that utilize [ISO/IEEE 11073-20601A], where xx can be any number from 01 to 99, inclusive.
- [ISO/IEEE 11073-10421] ISO/IEEE 11073-10421-2010, *Health informatics – Personal health device communication – Part 10421: Device specialization – Peak expiratory flow monitor (peak flow)*.

3 Definitions

3.1 Terms defined elsewhere

This Recommendation uses the following terms defined elsewhere:

3.1.1 agent [ISO/IEEE 11073-20601A]: A node that collects and transmits personal health data to an associated manager.

3.1.2 manager [ISO/IEEE 11073-20601A]: A node receiving data from one or more agent systems. Some examples of managers include a cellular phone, health appliance, set top box, or a computer system.

3.2 Terms defined in this Recommendation

None.

4 Abbreviations and acronyms

This Recommendation uses the following abbreviations and acronyms:

ATS	Abstract Test Suite
DUT	Device Under Test
CDG	Continua Design Guidelines
GUI	Graphical User Interface

INR	International Normalized Ratio
IUT	Implementation Under Test
MDS	Medical Device System
NFC	Near Field Communication
PAN	Personal Area Network
PCT	Protocol Conformance Testing
PCO	Point of Control and Observation
PHD	Personal Healthcare Device
PHDC	Personal Healthcare Device Class
PHM	Personal Health Manager
PICS	Protocol Implementation Conformance Statement
PIXIT	Protocol Implementation extra Information for Testing
SDP	Service Discovery Protocol
SOAP	Simple Object Access Protocol
TCRL	Test Case Reference List
TCWG	Test and Certification Working Group
TP	Test Purpose
TSS	Test Suite Structure
USB	Universal Serial Bus
WDM	Windows Driver Model

5 Conventions

The key words "SHALL", "SHALL NOT", "SHOULD", "SHOULD NOT", "MAY", "MAY NOT" in this document are to be interpreted as in [b-ETSI SR 001 262].

- SHALL is equivalent to 'must' or 'it is required to'.
- SHALL NOT is equivalent to 'must not' or 'it is not allowed'.
- SHOULD is equivalent to 'it is recommended to'.
- SHOULD NOT is equivalent to 'it is not recommended to'.
- MAY is equivalent to 'is permitted'.
- MAY NOT is equivalent to 'it is not required that'.

NOTE – The above-mentioned key words are capitalized for illustrative purposes only and they do not appear capitalized within this Recommendation.

Reference is made in the ITU-T H.800-series of Recommendations to different versions of the Continua Design Guidelines (CDG) by a specific designation. The list of terms that may be used in this Recommendation is provided in Table 1.

Table 1 – List of designations associated with the various versions of the CDG

CDG name	Transposed as	Version	Description	Designation
2013 plus errata	ITU-T H.810	4.1	CDG 2013 plus errata noting all ratified bugs.	–
2013	–	4.0	Release 2013 of the CDG including maintenance updates of CDG 2012 and additional guidelines that cover new functionalities.	Endorphin
2012 plus errata	–	3.1	CDG 2012 plus errata noting all ratified bugs [b-CDG 2012].	–
2012	–	3.0	Release 2012 of the CDG including maintenance updates of CDG 2011 and additional guidelines that cover new functionalities.	Catalyst
2011 plus errata	–	2.1	CDG 2011 integrated with identified errata.	–
2011	–	2.0	Release 2011 of the CDG including maintenance updates of CDG 2010 and additional guidelines that cover new functionalities [b-CDG 2011].	Adrenaline
2010 plus errata	–	1.6	CDG 2010 integrated with identified errata.	–
2010	–	1.5	Release 2010 of the CDG with maintenance updates of CDG Version 1 and additional guidelines that cover new functionalities [b-CDG 2010].	1.5
1.0	–	1.0	First released version of the CDG [b-CDG 1.0].	–

6 Test suite structure (TSS)

The test purposes (TPs) for the PAN/LAN/TAN interface have been divided into the main subgroups specified below. Annex A describes the TPs for subgroup 1.3.11 (shown in bold).

- Group 1: Agent (AG)
 - Group 1.1: Transport (TR)
 - Subgroup 1.1.1: Design guidelines: Common (DGC)
 - Subgroup 1.1.2: USB design guidelines (UDG)
 - Subgroup 1.1.3: Bluetooth design guidelines (BDG)
 - Subgroup 1.1.4: Pulse oximeter design guidelines (PODG)
 - Subgroup 1.1.5: Cardiovascular design guidelines (CVDG)
 - Subgroup 1.1.6: Activity hub design guidelines (HUBDG)
 - Subgroup 1.1.7: ZigBee design guidelines (ZDG)
 - Subgroup 1.1.8: Glucose meter design guidelines (GLDG)
 - Subgroup 1.1.9: Bluetooth low energy design guidelines (BLEDG)
 - Subgroup 1.1.10: Basic electrocardiograph design guidelines (ECGDG)
 - Subgroup 1.1.11: NFC design guidelines (NDG)
 - Group 1.2: Optimized exchange protocol (OXP)
 - Subgroup 1.2.1: PHD domain information model (DIM)

- Subgroup 1.2.2: PHD service model (SER)
 - Subgroup 1.2.3: PHD communication model (COM)
 - Group 1.3: Devices class specializations (CLASS)
 - Subgroup 1.3.1: Weighing scales (WEG)
 - Subgroup 1.3.2: Glucose meter (GL)
 - Subgroup 1.3.3: Pulse oximeter (PO)
 - Subgroup 1.3.4: Blood pressure monitor (BPM)
 - Subgroup 1.3.5: Thermometer (TH)
 - Subgroup 1.3.6: Cardiovascular (CV)
 - Subgroup 1.3.7: Strength (ST)
 - Subgroup 1.3.8: Activity hub (HUB)
 - Subgroup 1.3.9: Adherence monitor (AM)
 - Subgroup 1.3.10: Insulin pump (IP) (Future development)
 - **Subgroup 1.3.11: Peak flow (PF)**
 - Subgroup 1.3.12: Body composition analyser (BCA)
 - Subgroup 1.3.13: Basic electrocardiograph (ECG)
 - Subgroup 1.3.14: International normalized ratio (INR)
 - Group 1.4: Personal health device transcoding whitepaper (PHDTW)
 - Subgroup 1.4.1: Whitepaper general requirements (GEN)
 - Subgroup 1.4.2: Whitepaper thermometer requirements (TH)
 - Subgroup 1.4.3: Whitepaper blood pressure requirements (BPM)
 - Subgroup 1.4.4: Whitepaper heart rate requirements (HR)
 - Subgroup 1.4.5: Whitepaper glucose meter requirements (GL)
- Group 2: Manager (MAN)
- Group 2.1: Transport (TR)
 - Subgroup 2.1.1: Design guidelines: common (DGC)
 - Subgroup 2.1.2: USB design guidelines (UDG)
 - Subgroup 2.1.3: Bluetooth design guidelines (BDG)
 - Subgroup 2.1.4: Cardiovascular design guidelines (CVDG)
 - Subgroup 2.1.5: Activity hub design guidelines (HUBDG)
 - Subgroup 2.1.6: ZigBee design guidelines (ZDG)
 - Subgroup 2.1.7: Bluetooth low energy design guidelines (BLEDG)
 - Subgroup 2.1.8: NFC design guidelines (NDG)
 - Group 2.2: 20601: Optimized exchange protocol (OXP)
 - Subgroup 2.2.1: General (GEN)
 - Subgroup 2.2.2: PHD domain information model (DIM)
 - Subgroup 2.2.3: PHD service model (SER)
 - Subgroup 2.2.4: PHD communication model (COM)
 - Group 2.3: Devices class specializations (CLASS)
 - Subgroup 2.3.1: Weighing scales (WEG)

- Subgroup 2.3.2: Glucose meter (GL)
- Subgroup 2.3.3: Pulse oximeter (PO)
- Subgroup 2.3.4: Blood pressure monitor (BPM)
- Subgroup 2.3.5: Thermometer (TH)
- Subgroup 2.3.6: Cardiovascular (CV)
- Subgroup 2.3.7: Strength (ST)
- Subgroup 2.3.8: Activity hub (HUB)
- Subgroup 2.3.9: Adherence monitor (AM)
- Subgroup 2.3.10: Insulin pump (IP) (Future development)
- Subgroup 2.3.11: Peak flow (PF)
- Subgroup 2.3.12: Body composition analyser (BCA)
- Subgroup 2.3.13: Basic electrocardiograph (ECG)
- Subgroup 2.3.14: International normalized ratio (INR)
- Group 2.4: Personal health device transcoding whitepaper (PHDTW)
 - Subgroup 2.4.1: Whitepaper general requirements (GEN)
 - Subgroup 2.4.2: Whitepaper thermometer requirements (TH)
 - Subgroup 2.4.3: Whitepaper blood pressure measurement requirements (BPM)
 - Subgroup 2.4.4: Whitepaper heart rate requirements (HR)
 - Subgroup 2.4.5: Whitepaper glucose meter requirements (GL)

7 Electronic attachment

The protocol implementation conformance statements (PICS) and the protocol implementation extra information for testing (PIXIT) required for the implementation of this Annex can be downloaded from <http://handle.itu.int/11.1002/2000/12067>.

In the electronic attachment, letters "C" and "I" in the column labelled "Mandatory" are used to distinguish between "PICS" and "PIXIT" respectively during testing. If the cell is empty, the corresponding PICS is "independent". If the field contains a "C", the corresponding PICS is dependent on other PICS, and the logical expression is detailed in the "SCR_Expression" field. The static conformance review (SCR) is used in the test tool to assert whether the PICS selection is consistent.

Annex A

Test purposes (TPs)

(This annex forms an integral part of this Recommendation.)

A.1 TP definition conventions

The test purposes are defined according to the following rules:

- **TP Id:** This is a unique identifier (TP/<TT>/<DUT>/<GR>/<SGR>/<XX> – <NNN>). It is specified according to the naming convention defined below:
 - Each test purpose identifier is introduced by the prefix "TP".
 - <TT>: This is the test tool that will be used in the test case:
 - PAN: Personal area network (Bluetooth or USB)
 - LAN: Local area network (ZigBee)
 - PAN-LAN: Personal area network (Bluetooth or USB) - Local area network (ZigBee)
 - LP-PAN: Low power personal area network (Bluetooth low energy)
 - TAN: Touch area network (NFC)
 - PLT: Personal area network (Bluetooth or USB) – Local area network (ZigBee) – Touch area network (NFC)
 - <DUT>: It is the device under test:
 - AG: PAN/LAN Agent
 - MAN: PAN/LAN Manager
 - <GR>: This identifies a group of test cases.
 - <SGR>: This identifies a subgroup of test cases.
 - <XX>: This identifies the type of testing:
 - BV: Valid behaviour test
 - BI: Invalid behaviour test
 - <NNN>: This is a sequential number that identifies the test purpose.
- **TP label:** This is the TP's title.
- **Coverage:** This contains the specification reference and clause to be checked by the TP.
 - Spec: This indicates the earliest version of the specification from which the testable items to be checked by the TP were included.
 - Testable item: This contains testable items to be checked by the TP.
- **Test purpose:** This is a description of the requirements to be tested.
- **Applicability:** This contains the PICS items that define if the test case is applicable or not for a specific device. When a TP contains an "ALL" in this field it means that it applies to the device under test within that scope of the test (specialization, transport used, etc.).
- **Initial condition:** This indicates the state to which the DUT needs to be moved at the beginning of TC execution.
- **Test procedure:** This describes the steps to be followed in order to execute the test case.
- **Pass/Fail criteria:** This provides criteria to decide whether the DUT passes or fails the test case.

A.2 Subgroup 1.3.11: Peak expiratory flow monitor (PF)

TP Id		TP/PLT/AG/CLASS/PF/BV-000		
TP label		Get MDS Object for peak expiratory flow monitor specialization: Mandatory, Conditional and Optional Attributes		
Coverage	Spec	[ISO/IEEE 11073-10421]		
	Testable items	PF_MDSAttr1; C	PF_MDSAttr2; M	PF_MDSAttr3; M
		PF_MDSAttr4; M	PF_MDSAttr5; O	PF_MDSAttr6; M
		PF_GETServ1; M	PF_GETServ3; M	PF_OperProc2; M
Applicability		C_AG_OXP_170 AND C_AG_OXP_000		
Initial condition		The simulated manager and the agent under test are in the operating state.		
Test procedure		<p>1. The simulated manager issues a "roiv-cmip-get" command with the handle set to 0 (to request for an MDS object) and the attribute-id-list set to 0 to indicate all attributes.</p> <p>2. The agent under test responds with a "rors-cmip-get" service message in which the attribute-list contains a list of all implemented attributes of the MDS object:</p> <p>MDS Attributes:</p> <ul style="list-style-type: none"> a. Conditional attribute System-Type shall not be present. <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SYS_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value =<not relevant > b. Mandatory attribute System-Type-Spec_List <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SYS_TYPE_SPEC_LIST <input type="checkbox"/> attribute-type = TypeVerList <input type="checkbox"/> attribute-value.length = 4 bytes attribute-value = MDC_DEV_SPEC_PROFILE_PEFM, 1 c. Mandatory attribute System-model <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_MODEL <input type="checkbox"/> attribute-type = SystemModel <input type="checkbox"/> attribute-value.length =<variable> <input type="checkbox"/> attribute-value = {Manufacturer, Model} d. Mandatory attribute Dev-Configuration-Id <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_DEV_CONFIG_ID <input type="checkbox"/> attribute-type = ConfigId <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <ul style="list-style-type: none"> - IF NOT C_AG_OXP_181 then attribute-value = 0x08 0x34 - ELSE attribute-value = < between 0x4000 and 0x7FFF> e. Optional attribute Power-Status <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_POWER_STAT <input type="checkbox"/> attribute-type = PowerStatus (BITS-16) <input type="checkbox"/> attribute-value.length = 2 bytes 		

	<ul style="list-style-type: none"> <input type="checkbox"/> attribute-value = ON_MAINS (0x8000) or ON_BATTERY(0x4000) <p>Only one of the following may be active:</p> <ul style="list-style-type: none"> ▪ chargingFull(8), ▪ chargingTrickle(9), ▪ chargingOff(10). <p>The rest of the bits must not be set.</p>
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/AG/CLASS/PF/BV-001			
TP label	MDS Configuration objects events for peak expiratory flow monitor agent			
Coverage	Spec	[ISO/IEEE 11073-10421]		
	Testable items	PF_MDSEvent1; M	PF_GenNumObj1; O	PEF1; M
		PersBest1; M	FEV1S1;M	FEV6S1; O
		ReadStatus1;M	PF_ExtRules2; M	PF_ConfProc1;M
Applicability	C_AG_OXP_170 AND C_AG_OXP_000			
Initial condition	The simulated manager and the agent under test are in the unassociated state.			
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager: <ol style="list-style-type: none"> a. APDU Type <ul style="list-style-type: none"> <input type="checkbox"/> field- type = PrstApdu <input type="checkbox"/> field-length =2 bytes <input type="checkbox"/> field-value =0xE7 0x00 b. invoke-id <ul style="list-style-type: none"> <input type="checkbox"/> field- type = InvokeIDType <input type="checkbox"/> field-length =INT-U16 <input type="checkbox"/> field- value=<Not relevant for this test> c. message <ul style="list-style-type: none"> <input type="checkbox"/> field- type = roiv-cmip-confirmed-event-report <input type="checkbox"/> field-length =two bytes <input type="checkbox"/> field- value=0x01 0x01 (EventReportArgumentSimple) d. obj-handle (EventReportArgumentSimple) <ul style="list-style-type: none"> <input type="checkbox"/> field- type = HANDLE <input type="checkbox"/> field-length =INT-U16 e. event-time (EventReportArgumentSimple) <ul style="list-style-type: none"> <input type="checkbox"/> field- type = Relative Time <input type="checkbox"/> field-length =INT-U32 			

	<ul style="list-style-type: none"> <input type="checkbox"/> field-value = <ul style="list-style-type: none"> • IF NOT C_AG_OXP_010 THEN value = 0xFF 0xFF 0xFF 0xFF f. event-type (EventReportArgumentSimple) <ul style="list-style-type: none"> <input type="checkbox"/> field- type = OID-Type <input type="checkbox"/> field-length =INT-U16 <input type="checkbox"/> field- value=0x 0D 0x 1C (MDC_NOTI_CONFIG) g. config-report-id (ConfigReport) <ul style="list-style-type: none"> <input type="checkbox"/> field- type = ConfigId <input type="checkbox"/> field-length = INT-U16 <input type="checkbox"/> field- value = <ul style="list-style-type: none"> – IF NOT C_AG_OXP_181 then attribute-value = 0x08 0x34 – ELSE attribute-value = < between 0x4000 and 0x7FFF > h. obj-class (ConfigReport → ConfigObjectList (ConfigObject)). To check the objects that are supported by the Agent, Type Attribute will be checked in AttributeList. <ul style="list-style-type: none"> <input type="checkbox"/> field- type = OID-Type <input type="checkbox"/> field-length = INT-U16 <input type="checkbox"/> field- value = <ul style="list-style-type: none"> – Three simple numeric objects for PEF, Personal Best and FEV1 shall be present. – One enumeration object, Reading status shall be present. – IF NOT C_AG_OXP_181 and C_AG_PF_ 001 FEV6 shall be present ELSE FEV6 shall not be present.
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/AG/CLASS/PF/BV-002			
TP label	MDS objects events Peak expiratory flow monitor			
Coverage	Spec	[ISO/IEEE 11073-10421]		
	Testable items	PF_MDSEvent3; M	PF_MDSEvent4; M	PF_MDSEvent5; M
		PF_MDSEvent6; M	PF_EventRepServ1; M	PF_EventRepServ2; M
		PF_EventRepServ3; O	PF_OperProc4; M	PF_OperProc8; M
PF_OperProc9; O				
Applicability	C_AG_OXP_170 AND C_AG_OXP_000 AND (C_AG_OXP_182 OR C_AG_OXP_183 OR C_AG_OXP_184 OR C_AG_OXP_189)			
Initial condition	The simulated manager and the agent under test are in the unassociated state.			
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent under test responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 			

	<p>4. Check that the field Dev-Config-Id is set to the tested configuration. If it is not, the manager responds with an "unsupported-config" and waits for a new configuration. Repeat this step until a Dev-config-Id equal to tested configuration is received.</p> <p>5. Record the agent configuration.</p> <p>6. Take Measurements for every supported object in the agent under test.</p> <p>7. Wait to receive every event report and check:</p> <ul style="list-style-type: none"> <input type="checkbox"/> field- type = Event Report <input type="checkbox"/> field-length = 2 bytes <input type="checkbox"/> field- value=0x01 0x01 (EventReportArgumentSimple, confirmed). This field identifies the type of message sent by the agent, for the confirmed event configuration, roiv-cmip-confirmed-event-report.
Pass/Fail criteria	<ul style="list-style-type: none"> • Check that every received MDS Event report is one of the following Data APDU and that it is confirmed. • For Standard Configuration (NOT C_AG_OXP_181): An MDS Event Report is sent by the agent under test to report measurements for every object: <ul style="list-style-type: none"> <input type="checkbox"/> MDC_NOTI_SCAN_REPORT_FIXED <input type="checkbox"/> MDC_NOTI_SCAN_REPORT_MP_FIXED • For Extended Configuration, an MDS Event Report is sent by the agent under test to report measurements for every object: <ul style="list-style-type: none"> <input type="checkbox"/> MDC_NOTI_SCAN_REPORT_FIXED <input type="checkbox"/> MDC_NOTI_SCAN_REPORT_MP_FIXED <input type="checkbox"/> MDC_NOTI_SCAN_REPORT_VAR <input type="checkbox"/> MDC_NOTI_SCAN_REPORT_MP_VAR
Notes	

TP Id	TP/PLT/AG/CLASS/PF/BV-003			
TP label	PEF Object for Standard Configuration (0x0834)			
Coverage	Spec	[ISO/IEEE 11073-10421]		
	Testable items	PEF2; M	PEF3; M	PEF4; R
		PEF5; M	PEF6; R	PEF7; R
		PEF8; R	PEF9; R	PEF10; R
		PEF11; M	PEF12; M	PEF13; R
		PEF14; O	PEF15; O	PEF16; M
		PEF17; C	PEF18; C	PEF19; R
		PEF20; C	PEF21; R	PEF22; R
		PEF23; R	PEF24; R	PEF25; R
		PEF26; R	PEF45; M	PF_ConfProc2; M
Applicability	C_AG_OXP_170 AND (NOT C_AG_OXP_181) AND C_AG_OXP_000			
Initial condition	The simulated manager and the agent under test are in the unassociated state.			

Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 4. Check that the field Dev-Config-Id is set to 0x0834. If it is not, the manager responds with an "unsupported-config" and waits for a new configuration. Repeat this step until a Dev-config-Id equal to 0x0834 is received. 5. Once the agent under test sends a standard configuration, check the PEF object. 6. The PEF object contents shall be: <ol style="list-style-type: none"> a. Mandatory attribute Handle <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_HANDLE <input type="checkbox"/> attribute-type = HANDLE <input type="checkbox"/> attribute-value = 0x00 0x01 b. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = MDC_PART_SCADA, MDC_FLOW_AWAY_EXP_FORCED_PEAK c. Mandatory attribute Metric-Spec-Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = 0xD0 0x40 <ul style="list-style-type: none"> • Bit 0 (mss-avail-intermittent(0)) is set. • Bit 1 (mss-avail-stored-data(1)) is set. • Bit 3 (mss-msmt-aperiodic(3)) is set. • Bit 9 (mss-acc-agent-initiated(9)) is set. d. Mandatory attribute Unit-Code <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE <input type="checkbox"/> attribute-type = OID-Type <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_DIM_X_L_PER_MIN e. Mandatory attribute Attribute-Value-Map <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ATTRIBUTE_VAL_MAP <input type="checkbox"/> attribute-type = AttrValMap <input type="checkbox"/> attribute-count = 2 <input type="checkbox"/> attribute-value = (MDC_ATTR_NU_VAL_OBS_SIMP,4 MDC_ATTR_TIME_STAMP_ABS, 8) 7. Check that no other attributes are present in the initial configuration.
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id		TP/PLT/AG/CLASS/PF/BV-004		
TP label		PEF Object for Extended Configuration		
Coverage	Spec	[ISO/IEEE 11073-10421]		
	Testable items	PEF27; M	PEF28; R	PEF29; M
		PEF30; R	PEF31; R	PEF32; R
		PEF33; R	PEF34; R	PEF35; M
		PEF37; R	PEF38; R	PEF39; R
		PEF40; R	PEF41; R	PEF42; R
		PEF43; R	PEF44; R	
Applicability		C_AG_OXP_170 AND C_AG_OXP_181 AND C_AG_OXP_000		
Initial condition		The simulated manager and the agent under test are in the unassociated state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent under test responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 4. Check that the field Dev-Config-Id is set to the tested extended configuration. If it is not, the manager responds with an "unsupported-config" and waits for a new configuration. Repeat this step until a Dev-config-Id equal to the extended configuration is received. 5. Once the agent under test sends the tested configuration, check the PEF object. 6. The PEF object contents shall be: <ol style="list-style-type: none"> a. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = MDC_PART_SCADA, MDC_FLOW_AWAY_EXP_FORCED_PEAK b. IF Not Recommended attribute Supplemental-Types <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SPPLEMENTAL_TYPES <input type="checkbox"/> attribute-type = SupplementalTypeList <input type="checkbox"/> attribute-value.length = <variable>Sequence of TYPE (TYPE.length= 4 bytes) <input type="checkbox"/> attribute-value = <Not relevant for this test> c. Mandatory attribute Metric-Spec-Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value =0xD0 0x40 <ul style="list-style-type: none"> • Bit 0 (mss-avail-intermittent(0)) is set. • Bit 1 (mss-avail-stored-data(1)) is set. • Bit 3 (mss-msmt-aperiodic(3)) is set. • Bit 9 (mss-acc-agent-initiated(9)) is set. 		

- d. IF Not recommended attribute Metric-Structure-Small is present
 - attribute-id = MDC_ATTR_METRIC_STRUCTURE_SMALL
 - attribute-type = MetricStructureSmall
 - attribute-length = 2 bytes
 - attribute-value = <Not relevant for this test>
- e. IF Not recommended attribute Measurement-Status is present
 - attribute-id = MDC_ATTR_MSMT_STAT
 - attribute-type = MeasurementStatus
 - attribute-value.length = 2 bytes
 - attribute-value = <Not relevant for this test>
- f. IF Not recommended attribute Metric-Id is present
 - attribute-id = MDC_ATTR_ID_PHYSIO
 - attribute-type = OID-Type(INT-U16)
 - attribute-value.length = 2 bytes
 - attribute-value = <Not relevant for this test>
- g. IF Not Recommended attribute Metric-Id-List is present
 - attribute-id = MDC_ATTR_ID_PHYSIO_LIS
 - attribute-type = MetricIdList
 - attribute-value = <Not relevant for this test>
- h. IF Not recommended attribute Metric-Id-Partition is present
 - attribute-id = MDC_ATTR_METRIC_ID_PART
 - attribute-type = NomPartition(INT-U16)
 - attribute-value.length = 2 bytes
 - attribute-value = <Not relevant for this test>
- i. Mandatory attribute Unit-Code
 - attribute-id = MDC_ATTR_UNIT_CODE
 - attribute-type = OID-Type
 - attribute-value.length = 2 bytes
 - attribute-value = MDC_DIM_X_L_PER_MIN
- j. IF Not recommended attribute Source-Handle-Reference is present
 - attribute-id = MDC_ATTR_SOURCE_HANDLE_REF
 - attribute-type = HANDLE(INT-U16)
 - attribute-value.length = 2 bytes
 - attribute-value = <Not relevant for this test>
- k. IF Not recommended attribute Measure-Active-Period
 - attribute-id = MDC_ATTR_TIME_PD_MSMT_ACTIVE
 - attribute-type = FLOAT-Type (INT-U32)
 - attribute-value.length = 4 bytes
 - attribute-value = <Not relevant for this test>
- l. IF Not recommended Compound-Simple-Nu-Observed-Value is present
 - attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_SIMP
 - attribute-type = SimpleNuObsValueCmp
 - attribute-value.length = <variable>
 - attribute-value = <Not relevant for this test>

	<ul style="list-style-type: none"> m. IF Not recommended attribute Basic-Nu-Observed-Value is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_VAL_OBS_BASIC <input type="checkbox"/> attribute-type = BasicNuObsValue <input type="checkbox"/> attribute-value.length = 2bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> n. IF Not recommended attribute Compound-Basic-Nu-Observed-Value is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_BASIC <input type="checkbox"/> attribute-type = BasicNuObsValueCmp <input type="checkbox"/> attribute-value.length = <variable> <input type="checkbox"/> attribute-value = <Not relevant for this test> o. IF Not recommended attribute Nu-Observed-Value is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_VAL_OBS <input type="checkbox"/> attribute-type = NuObsValue <input type="checkbox"/> attribute-value.length = 10bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> p. Not recommended attribute Compound-Nu-Observed-Value <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_SIMP <input type="checkbox"/> attribute-type = NuObsValueCmp <input type="checkbox"/> attribute-value.length = <variable> <input type="checkbox"/> attribute-value = <Not relevant for this test> q. IF Not Recommended attribute Accuracy is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_ACCUR_MSMT <input type="checkbox"/> attribute-type = FLOAT-Type (INT-U32) <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test>
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id		TP/PLT/AG/CLASS/PF/BV-005		
TP label		Personal Best Object for Standard Configuration (0x0834)		
Coverage	Spec	[ISO/IEEE 11073-10421]		
	Testable items	PersBest2; M	PersBest3; M	PersBest4; R
		PersBest5; M	PersBest6; R	PersBest7; R
		PersBest8; R	PersBest9; R	PersBest10; R
		PersBest11; M	PersBest12; M	PersBest13; R
		PersBest14; O	PersBest15; O	PersBest16; C
		PersBest17; R	PersBest18; C	PersBest19; R
		PersBest20; C	PersBest21; R	PersBest22; R
		PersBest23; R	PersBest24; R	PersBest25; R
		PersBest26; R	PersBest40; M	PF_ConfProc2; M
Applicability		C_AG_OXP_170 AND (NOT C_AG_OXP_181) AND C_AG_OXP_000		
Initial condition		The simulated manager and the agent under test are in the unassociated state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 4. Check that the field Dev-Config-Id is set to 0x0834. If it is not, the manager responds with an "unsupported-config" and waits for a new configuration. Repeat this step until a Dev-config-Id equal to 0x0834 is received. 5. Once the agent under test sends a standard configuration, check the Personal Best object. 6. The Personal Best object contents shall be: <ol style="list-style-type: none"> a. Mandatory attribute Handle <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_HANDLE <input type="checkbox"/> attribute-type = HANDLE <input type="checkbox"/> attribute-value = 0x00 0x02 b. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = MDC_PART_SCADA, MDC_FLOW_AWAY_EXP_FORCED_PEAK_PB c. Mandatory attribute Metric-Spec-Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = 0xC0 0x44 <ul style="list-style-type: none"> • Bit 0 (mss-avail-intermittent(0)) is set. • Bit 1 (mss-avail-stored-data(1)) is set. 		

	<ul style="list-style-type: none"> • Bit 9 (mss-acc-agent-initiated(9)) is set. • Bit 13 (mss-cat-setting (13)) is set. <p>d. Mandatory attribute Unit-Code</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE <input type="checkbox"/> attribute-type = OID-Type <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_DIM_X_L_PER_MIN <p>e. Mandatory attribute Attribute-Value-Map</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ATTRIBUTE_VAL_MAP <input type="checkbox"/> attribute-type = AttrValMap <input type="checkbox"/> attribute-count = 2 <input type="checkbox"/> attribute-value = (MDC_ATTR_NU_VAL_OBS_SIMP,4 MDC_ATTR_TIME_STAMP_ABS, 8) <p>7. Check that no other attributes are present in the initial configuration.</p>
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id		TP/PLT/AG/CLASS/PF/BV-006		
TP label		Personal Best Object for Extended Configuration		
Coverage	Spec	[ISO/IEEE 11073-10421]		
	Testable items	PersBest27; M	PersBest28; R	PersBest29; M
		PersBest30; R	PersBest31; R	PersBest32; R
		PersBest33; R	PersBest34; R	PersBest35; M
		PersBest37; R	PersBest38; R	PersBest39; R
Applicability		C_AG_OXP_170 AND C_AG_OXP_181 AND C_AG_OXP_000		
Initial condition		The simulated manager and the agent under test are in the unassociated state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 4. Check that the field Dev-Config-Id is set to the tested extended configuration. If it is not, the manager responds with an "unsupported-config" and waits for a new configuration. Repeat this step until a Dev-config-Id equal to the extended configuration is received. 5. Once the agent under test sends the tested configuration, check Personal Best object. 6. The Personal Best object contents shall be: <ol style="list-style-type: none"> a. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE 		

attribute-value = MDC_PART_SCADA,
MDC_FLOW_AWAY_EXP_FORCED_PEAK_PB

- b. IF Not Recommended attribute Supplemental-Types
 - attribute-id = MDC_ATTR_SPPLEMENTAL_TYPES
 - attribute-type = SupplementalTypeList
 - attribute-value.length = <variable>Sequence of TYPE (TYPE.length= 4 bytes)
 - attribute-value = <Not relevant for this test>
- c. Mandatory attribute Metric-Spec-Small
 - attribute-id = MDC_ATTR_METRIC_SPEC_SMALL
 - attribute-type = MetricSpecSmall
 - attribute-value.length = 2 bytes
 - attribute-value = 0xC0 0x44
 - Bit 0 (mss-avail-intermittent(0)) is set.
 - Bit 1 (mss-avail-stored-data(1)) is set.
 - Bit 9 (mss-acc-agent-initiated(9)) is set.
 - Bit 13 (mss-cat-setting (13)) is set.
- d. IF Not recommended attribute Metric-Structure-Small is present
 - attribute-id = MDC_ATTR_METRIC_STRUCTURE_SMALL
 - attribute-type = MetricStructureSmall
 - attribute-length = 2 bytes
 - attribute-value = <Not relevant for this test>
- e. IF Not recommended attribute Measurement-Status is present
 - attribute-id = MDC_ATTR_MSMT_STAT
 - attribute-type = MeasurementStatus
 - attribute-value.length = 2 bytes
 - attribute-value = <Not relevant for this test>
- f. IF Not recommended attribute Metric-Id is present
 - attribute-id = MDC_ATTR_ID_PHYSIO
 - attribute-type = OID-Type(INT-U16)
 - attribute-value.length = 2 bytes
 - attribute-value = <Not relevant for this test>
- g. IF Not Recommended attribute Metric-Id-List is present
 - attribute-id = MDC_ATTR_ID_PHYSIO_LIS
 - attribute-type = MetricIdList
 - attribute-value = <Not relevant for this test>
- h. IF Not recommended attribute Metric-Id-Partition is present
 - attribute-id = MDC_ATTR_METRIC_ID_PART
 - attribute-type = NomPartition(INT-U16)
 - attribute-value.length = 2 bytes
 - attribute-value = <Not relevant for this test>
- i. Mandatory recommended attribute Unit-Code
 - attribute-id = MDC_ATTR_UNIT_CODE
 - attribute-type = OID-Type(INT-U16)

	<ul style="list-style-type: none"> <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_DIM_X_L_PER_MIN j. IF Not recommended attribute Source-Handle-Reference is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SOURCE_HANDLE_REF <input type="checkbox"/> attribute-type = HANDLE(INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> k. IF Not recommended attribute Measure-Active-Period <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_PD_MSMT_ACTIVE <input type="checkbox"/> attribute-type = FLOAT-Type (INT-U32) <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> l. IF Not Recommended attribute Accuracy is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_ACCUR_MSMT <input type="checkbox"/> attribute-type = FLOAT-Type (INT-U32) <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test>
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/AG/CLASS/PF/BV-007			
TP label	FEV1 for Standard Configuration (0x0834)			
Coverage	Spec	[ISO/IEEE 11073-10421]		
	Testable items	FEV1S2; M	FEV1S3; M	FEV1S4; R
		FEV1S5; M	FEV1S6; R	FEV1S7; R
		FEV1S8; R	FEV1S9; R	FEV1S10; R
		FEV1S11; M	FEV1S12; M	FEV1S13; R
		FEV1S14; O	FEV1S15; O	FEV1S16; C
		FEV1S17; C	FEV1S18; C	FEV1S19; R
		FEV1S20; C	FEV1S21; R	FEV1S22; R
		FEV1S23; R	FEV1S24; R	FEV1S25; R
		FEV1S26; R	FEV1S45; M	PF_ConfProc2; M
Applicability	C_AG_OXP_170 AND (NOT C_AG_OXP_181) AND C_AG_OXP_000			
Initial condition	The simulated manager and the agent under test are in the unassociated state.			
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 			

	<p>4. Check that the field Dev-Config-Id is set to 0x0834. If it is not, the manager responds with an "unsupported-config" and waits for a new configuration. Repeat this step until a Dev-config-Id equal to 0x0834 is received.</p> <p>5. Once the agent under test sends a standard configuration, check the FEV1 object.</p> <p>6. The FEV1 contents shall be:</p> <ul style="list-style-type: none"> a. Mandatory attribute Handle <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_HANDLE <input type="checkbox"/> attribute-type = HANDLE <input type="checkbox"/> attribute-value = 0x00 0x03 b. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = MDC_PART_SCADA MDC_VOL_AWAY_EXP_FORCED_1S c. Mandatory attribute Metric-Spec-Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = 0xD0 0x40 <ul style="list-style-type: none"> • Bit 0 (mss-avail-intermittent(0)) is set. • Bit 1 (mss-avail-stored-data(1)) is set. • Bit 3 (mss-msmt-aperiodic(3)) is set. • Bit 9 (mss-acc-agent-initiated(9)) is set. d. Mandatory recommended attribute Unit-Code <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE <input type="checkbox"/> attribute-type = OID-Type(INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_DIM_X_L e. Mandatory attribute Attribute-Value-Map <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ATTRIBUTE_VAL_MAP <input type="checkbox"/> attribute-type = AttrValMap <input type="checkbox"/> attribute-count = 2 <input type="checkbox"/> attribute-value = (MDC_ATTR_NU_VAL_OBS_SIMP,4 MDC_ATTR_TIME_STAMP_ABS, 8) <p>7. Check that no other attributes are present in the initial configuration.</p>
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/AG/CLASS/PF/BV-008		
TP label	FEV1 Object for Extended Configuration		
Coverage	Spec	[ISO/IEEE 11073-10421]	
	Testable	FEV1S27; M	FEV1S28; R
	FEV1S29; R		

	items	FEV1S30; M	FEV1S31; R	FEV1S32; R
		FEV1S33; R	FEV1S34; R	FEV1S35; M
		FEV1S37; R	FEV1S38; R	FEV1S39; R
		FEV1S40; R	FEV1S41; R	FEV1S42; R
		FEV1S43; R	FEV1S44; R	
Applicability	C_AG_OXP_170 AND C_AG_OXP_181 AND C_AG_OXP_000			
Initial condition	The simulated manager and the agent under test are in the unassociated state.			
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 4. Check that the field Dev-Config-Id is set to the tested extended configuration. If it is not, the manager responds with an "unsupported-config" and waits for a new configuration. Repeat this step until a Dev-config-Id equal to the extended configuration. 5. Once the agent under test sends the tested configuration, check the FEV1 object. 6. The FEV1 object contents shall be: <ol style="list-style-type: none"> a. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = MDC_PART_SCADA MDC_VOL_AWAY_EXP_FORCED_1S b. IF Not Recommended attribute Supplemental-Types <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SPPLEMENTAL_TYPES <input type="checkbox"/> attribute-type = SupplementalTypeList <input type="checkbox"/> attribute-value.length = <variable>Sequence of TYPE (TYPE.length= 4 bytes) <input type="checkbox"/> attribute-value = <Not relevant for this test> c. Mandatory attribute Metric-Spec-Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value =0xD0 0x40 <ul style="list-style-type: none"> • Bit 0 (mss-avail-intermittent(0)) is set. • Bit 1 (mss-avail-stored-data(1)) is set. • Bit 3 (mss-msmt-aperiodic(3)) is set. • Bit 9 (mss-acc-agent-initiated(9)) is set. d. IF Not recommended attribute Metric-Structure-Small is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_STRUCTURE_SMALL <input type="checkbox"/> attribute-type = MetricStructureSmall <input type="checkbox"/> attribute-length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> e. IF Not recommended attribute Measurement-Status is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_MSMT_STAT <input type="checkbox"/> attribute-type = MeasurementStatus 			

- attribute-value.length = 2 bytes
- attribute-value = <Not relevant for this test>
- f. IF Not recommended attribute Metric-Id is present
 - attribute-id = MDC_ATTR_ID_PHYSIO
 - attribute-type = OID-Type(INT-U16)
 - attribute-value.length =2 bytes
 - attribute-value = <Not relevant for this test>
- g. IF Not Recommended attribute Metric-Id-List is present
 - attribute-id = MDC_ATTR_ID_PHYSIO_LIS
 - attribute-type = MetricIdList
 - attribute-value = <Not relevant for this test>
- h. IF Not recommended attribute Metric-Id-Partition is present
 - attribute-id = MDC_ATTR_METRIC_ID_PART
 - attribute-type = NomPartition(INT-U16)
 - attribute-value.length = 2 bytes
 - attribute-value = <Not relevant for this test>
- i. Mandatory attribute Unit-Code
 - attribute-id = MDC_ATTR_UNIT_CODE
 - attribute-type = OID-Type
 - attribute-value.length = 2 bytes
 - attribute-value = MDC_DIM_X_L
- j. IF Not recommended attribute Source-Handle-Reference is present
 - attribute-id = MDC_ATTR_SOURCE_HANDLE_REF
 - attribute-type = HANDLE(INT-U16)
 - attribute-value.length = 2 bytes
 - attribute-value = <Not relevant for this test>
- k. IF Not recommended attribute Measure-Active-Period
 - attribute-id = MDC_ATTR_TIME_PD_MSMT_ACTIVE
 - attribute-type = FLOAT-Type (INT-U32)
 - attribute-value.length = 4 bytes
 - attribute-value = <Not relevant for this test>
- l. IF Not recommended Compound-Simple-Nu-Observed-Value is present
 - attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_SIMP
 - attribute-type = SimpleNuObsValueCmp
 - attribute-value.length =<variable>
 - attribute-value = <Not relevant for this test>
- m. IF Not recommended attribute Basic-Nu-Observed-Value is present
 - attribute-id = MDC_ATTR_NU_VAL_OBS_BASIC
 - attribute-type = BasicNuObsValue
 - attribute-value.length = 2bytes
 - attribute-value = <Not relevant for this test>
- n. IF Not recommended attribute Compound-Basic-Nu-Observed-Value is present
 - attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_BASIC
 - attribute-type = BasicNuObsValueCmp

	<ul style="list-style-type: none"> <input type="checkbox"/> attribute-value.length = <variable> <input type="checkbox"/> attribute-value = <Not relevant for this test> o. IF Not recommended attribute Nu-Observed-Value is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_VAL_OBS <input type="checkbox"/> attribute-type = NuObsValue <input type="checkbox"/> attribute-value.length = 10bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> p. Not recommended attribute Compound-Nu-Observed-Value <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_SIMP <input type="checkbox"/> attribute-type = NuObsValueCmp <input type="checkbox"/> attribute-value.length = <variable> <input type="checkbox"/> attribute-value = <Not relevant for this test> q. IF Not Recommended attribute Accuracy is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_ACCUR_MSMT <input type="checkbox"/> attribute-type = FLOAT-Type (INT-U32) <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test>
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/AG/CLASS/PF/BV-009			
TP label	FEV6 Object for Extended Configuration			
Coverage	Spec	[ISO/IEEE 11073-10421]		
	Testable items	FEV6S2; M	FEV6S3; R	FEV6S4; R
		FEV6S5; M	FEV6S6; R	FEV6S7; R
		FEV6S8; R	FEV6S9; R	FEV6S10; M
		FEV6S12; R	FEV6S14; R	FEV6S15; R
		FEV6S16; R	FEV6S17; R	FEV6S18; R
		FEV6S19; R	FEV6S20; R	
Applicability	C_AG_OXP_170 AND C_AG_OXP_181 AND C_AG_PF_001 AND C_AG_OXP_000			
Initial condition	The simulated manager and the agent under test are in the unassociated state.			
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 4. Check that the field Dev-Config-Id is set to the tested extended configuration. If it is not, the manager responds with an "unsupported-config" and waits for a new configuration. Repeat this step until a Dev-config-Id equal to the extended configuration is received. 5. Once the agent under test sends the tested configuration, check FEV6 object. 6. The FEV6 object contents shall be: 			

- a. Mandatory attribute Type
 - attribute-id = MDC_ATTR_ID_TYPE
 - attribute-type = TYPE
 - attribute-value = MDC_PART_SCADA | MDC_VOL_AWAY_EXP_FORCED_6S
- b. IF Not Recommended attribute Supplemental-Types
 - attribute-id = MDC_ATTR_SPPLEMENTAL_TYPES
 - attribute-type = SupplementalTypeList
 - attribute-value.length = <variable>Sequence of TYPE (TYPE.length= 4 bytes)
 - attribute-value = <Not relevant for this test>
- c. Mandatory attribute Metric-Spec-Small
 - attribute-id = MDC_ATTR_METRIC_SPEC_SMALL
 - attribute-type = MetricSpecSmall
 - attribute-value.length = 2 bytes
 - attribute-value = 0xD0 0x40
 - Bit 0 (mss-avail-intermittent(0)) is set.
 - Bit 1 (mss-avail-stored-data(1)) is set.
 - Bit 3 (mss-msmt-aperiodic(3)) is set.
 - Bit 9 (mss-acc-agent-initiated(9)) is set.
- d. IF Not recommended attribute Metric-Structure-Small is present
 - attribute-id = MDC_ATTR_METRIC_STRUCTURE_SMALL
 - attribute-type = MetricStructureSmall
 - attribute-length = 2 bytes
 - attribute-value = <Not relevant for this test>
- e. IF Not recommended attribute Measurement-Status is present
 - attribute-id = MDC_ATTR_MSMT_STAT
 - attribute-type = MeasurementStatus
 - attribute-value.length = 2 bytes
 - attribute-value = <Not relevant for this test>
- f. IF Not recommended attribute Metric-Id is present
 - attribute-id = MDC_ATTR_ID_PHYSIO
 - attribute-type = OID-Type(INT-U16)
 - attribute-value.length = 2 bytes
 - attribute-value = <Not relevant for this test>
- g. IF Not Recommended attribute Metric-Id-List is present
 - attribute-id = MDC_ATTR_ID_PHYSIO_LIS
 - attribute-type = MetricIdList
 - attribute-value = <Not relevant for this test>
- h. IF Not recommended attribute Metric-Id-Partition is present
 - attribute-id = MDC_ATTR_METRIC_ID_PART
 - attribute-type = NomPartition(INT-U16)
 - attribute-value.length = 2 bytes
 - attribute-value = <Not relevant for this test>

- i. Mandatory attribute Unit-Code
 - attribute-id = MDC_ATTR_UNIT_CODE
 - attribute-type = OID-Type
 - attribute-value.length = 2 bytes
 - attribute-value = MDC_DIM_X_L
- j. IF Not recommended attribute Source-Handle-Reference is present
 - attribute-id = MDC_ATTR_SOURCE_HANDLE_REF
 - attribute-type = HANDLE(INT-U16)
 - attribute-value.length = 2 bytes
 - attribute-value = <Not relevant for this test>
- k. IF Not recommended attribute Measure-Active-Period
 - attribute-id = MDC_ATTR_TIME_PD_MSMT_ACTIVE
 - attribute-type = FLOAT-Type (INT-U32)
 - attribute-value.length = 4 bytes
 - attribute-value = <Not relevant for this test>
- l. IF Not recommended attribute Compound-Simple-Nu-Observed-Value is present
 - attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_SIMP
 - attribute-type = SimpleNuObsValueCmp
 - attribute-value.length = <variable>
 - attribute-value = <Not relevant for this test>
- m. IF Not recommended attribute Basic-Nu-Observed-Value is present
 - attribute-id = MDC_ATTR_NU_VAL_OBS_BASIC
 - attribute-type = BasicNuObsValue
 - attribute-value.length = 2bytes
 - attribute-value = <Not relevant for this test>
- n. IF Not recommended attribute Compound-Basic-Nu-Observed-Value is present
 - attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_BASIC
 - attribute-type = BasicNuObsValueCmp
 - attribute-value.length = <variable>
 - attribute-value = <Not relevant for this test>
- o. IF Not recommended attribute Nu-Observed-Value is present
 - attribute-id = MDC_ATTR_NU_VAL_OBS
 - attribute-type = NuObsValue
 - attribute-value.length = 10bytes
 - attribute-value = <Not relevant for this test>
- p. Not recommended attribute Compound-Nu-Observed-Value
 - attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_SIMP
 - attribute-type = NuObsValueCmp
 - attribute-value.length = <variable>
 - attribute-value = <Not relevant for this test>
- q. IF Not Recommended attribute Accuracy is present
 - attribute-id = MDC_ATTR_NU_ACCUR_MSMT
 - attribute-type = FLOAT-Type (INT-U32)
 - attribute-value.length = 4 bytes

	<input type="checkbox"/> attribute-value = <Not relevant for this test>
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/AG/CLASS/PF/BV-010			
TP label	Reading status Object for Standard Configuration (0x0834)			
Coverage	Spec	[ISO/IEEE 11073-10421]		
	Testable items	ReadStatus2; M	ReadStatus3; M	ReadStatus4; R
		ReadStatus5; M	ReadStatus6; R	ReadStatus7; R
		ReadStatus8; R	ReadStatus9; R	ReadStatus10; R
		ReadStatus11; R	ReadStatus12; M	ReadStatus13; R
		ReadStatus14; O	ReadStatus15; O	ReadStatus16; C
		ReadStatus17; R	ReadStatus18; C	ReadStatus19; O
		ReadStatus20; R	ReadStatus21; M	ReadStatus22; R
		ReadStatus23; R	ReadStatus24; R	ReadStatus41; M
	PF_ConfProc2; M			
Applicability	C_AG_OXP_170 AND (NOT C_AG_OXP_181) AND C_AG_OXP_000			
Initial condition	The simulated manager and the agent under test are in the unassociated state.			
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 4. Check that the field Dev-Config-Id is set to 0x0834. If it is not, the manager responds with an "unsupported-config" and waits for a new configuration. Repeat this step until a Dev-config-Id equal to 0x0834 is received. 5. Once the agent under test sends a standard configuration, check reading status object. 6. The Reading status object contents shall be: <ol style="list-style-type: none"> a. Mandatory attribute Handle <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_HANDLE <input type="checkbox"/> attribute-type = HANDLE <input type="checkbox"/> attribute-value = 0x00 0x05 b. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = MDC_PART_PHD_DM, MDC_PEF_READING_STATUS c. Mandatory attribute Metric-Spec-Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall (BITS-16) <input type="checkbox"/> attribute-value = 0xD0 0x40 			

	<ul style="list-style-type: none"> • Bit 0 (mss-avail-intermittent(0)) is set. • Bit 1 (mss-avail-stored-data(1)) is set. • Bit 3 (mss-msmt-aperiodic(3)) is set. • Bit 9 (mss-acc-agent-initiated(9)) is set. <p>d. Mandatory attribute Attribute-Value-Map</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ATTRIBUTE_VAL_MAP <input type="checkbox"/> attribute-type = AttrValMap <input type="checkbox"/> attribute-count = 2 <input type="checkbox"/> attribute-value= (MDC_ATTR_ENUM_VAL_OBS_BASIC_BIT_STRING, 2 MDC_ATTR_TIME_STAMP_ABS, 8) <p>7. Check that no other attributes are present in the initial configuration.</p> <p>8. Take a measurement with the agent.</p> <p>9. Wait for the agent to send an event report and check:</p> <p>a. Mandatory attribute Enum-Observed-Value-Basic-Bit-Str</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id= MDC_ATTR_ENUM_OBS_VAL_SIM_OID <input type="checkbox"/> attribute-type = OID-Type <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value= One of the following bits may be active: <ul style="list-style-type: none"> <input type="checkbox"/> pefm-read-stat-post-medication(0) <input type="checkbox"/> pefm-read-stat-cough(1) <input type="checkbox"/> pefm-read-stat-short-effort(2) <input type="checkbox"/> pefm-read-stat-long-time-to-peak(3) <input type="checkbox"/> The rest of the bits must not be set
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/AG/CLASS/PF/BV-011			
TP label	Reading status Object for Extended Configuration			
Coverage	Spec	[ISO/IEEE 11073-10421]		
	Testable items	ReadStatus25; M	ReadStatus26; R	ReadStatus27; R
		ReadStatus28; M	ReadStatus29; R	ReadStatus30; R
		ReadStatus31; R	ReadStatus32; R	ReadStatus33; R
		ReadStatus34; R	ReadStatus35; O	ReadStatus36; R
		ReadStatus37; M	ReadStatus38; R	ReadStatus39; R
		ReadStatus40; R	ReadStatus41; M	
Applicability	C_AG_OXP_170 AND C_AG_OXP_181 AND AND C_AG_OXP_000			
Initial condition	The simulated manager and the agent under test are in the unassociated state.			

<p>Test procedure</p>	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 4. Check that the field Dev-Config-Id is set to extended configuration. If it is not, the manager responds with an "unsupported-config" and waits for a new configuration. Repeat this step until a Dev-config-Id equal to tested extended configuration is received. 5. Once the agent under test sends the tested configuration, check reading status object. 6. The Reading status object contents shall be: <ol style="list-style-type: none"> a. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = MDC_PART_PHD_DM, MDC_PEF_READING_STATUS b. IF Not Recommended attribute Supplemental-Types <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SPPLEMENTAL_TYPES <input type="checkbox"/> attribute-type = SupplementalTypeList <input type="checkbox"/> attribute-value.length = <variable>Sequence of TYPE (TYPE.length= 4 bytes) <input type="checkbox"/> attribute-value = <Not relevant for this test> c. Mandatory attribute Metric-Spec-Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall (BITS-16) <input type="checkbox"/> attribute-value =0xD0 0x40 <ul style="list-style-type: none"> • Bit 0 (mss-avail-intermittent(0)) is set. • Bit 1 (mss-avail-stored-data(1)) is set. • Bit 3 (mss-msmt-aperiodic(3)) is set. • Bit 9 (mss-acc-agent-initiated(9)) is set. d. IF Not recommended attribute Metric-Structure-Small is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_STRUCTURE_SMALL <input type="checkbox"/> attribute-type = MetricStructureSmall <input type="checkbox"/> attribute-length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> e. IF Not recommended attribute Measurement-Status is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_MSMT_STAT <input type="checkbox"/> attribute-type = MeasurementStatus <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> f. IF Not recommended attribute Metric-Id is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_PHYSIO <input type="checkbox"/> attribute-type = OID-Type(INT-U16) <input type="checkbox"/> attribute-value.length =2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> g. IF Not Recommended attribute Metric-Id-List is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_PHYSIO_LIS <input type="checkbox"/> attribute-type = MetricIdList <input type="checkbox"/> attribute-value = <Not relevant for this test>
------------------------------	--

- g. IF Not recommended attribute Metric-Id-Partition is present
 - attribute-id = MDC_ATTR_METRIC_ID_PART
 - attribute-type = NomPartition(INT-U16)
 - attribute-value.length = 2 bytes
 - attribute-value = <Not relevant for this test>
- h. IF Not recommended attribute Unit-Code is present
 - attribute-id = MDC_ATTR_UNIT_CODE
 - attribute-type = OID-Type(INT-U16)
 - attribute-value.length = 2 bytes
 - attribute-value = <Not relevant for this test>
- i. IF Not recommended attribute Source-Handle-Reference is present
 - attribute-id = MDC_ATTR_SOURCE_HANDLE_REF
 - attribute-type = HANDLE(INT-U16)
 - attribute-value.length = 2 bytes
 - attribute-value = <Not relevant for this test>
- j. IF Optional attribute Enum-Observed-Value-Simple-OID is present
 - attribute-id= MDC_ATTR_ENUM_OBS_VAL_SIM_OID
 - attribute-type = OID-Type (INT-U16)
 - attribute-value.length = 2 bytes
 - attribute-value = <Not relevant for this test>
- k. IF Not Recommended attribute Enum-Observed-Value-Simple-Bit-Str is present
 - attribute-id= MDC_ATTR_ENUM_OBS_VAL_SIM_BIT_STR
 - attribute-type = BITS-32
 - attribute-value.length = BITS-32
 - attribute-value= <Not relevant for this test>
- l. Mandatory attribute Enum-Observed-Value-Basic-Bit-Str is present
 - attribute-id= MDC_ATTR_ENUM_OBS_VAL_BASIC_BIT_STR
 - attribute-type = BITS-16
 - attribute-value.length = 2 bytes
 - attribute-value = One of the following bits may be active:
 - pefm-read-stat-post-medication(0)
 - pefm-read-stat-cough(1)
 - pefm-read-stat-short-effort(2)
 - pefm-read-stat-long-time-to-peak(3)
 - The rest of the bits must not be set
- m. IF Not Recommended attribute Enum-Observed-Value-Simple-Str is present
 - attribute-id= MDC_ATTR_ENUM_OBS_VAL_SIM_STR
 - attribute-type = EnumPrintableString
 - attribute-value.length = <variable>
 - attribute-value = <Not relevant for this test>
- n. IF Not Recommended attribute Enum-Observed-Value is present
 - attribute-id= MDC_ATTR_VAL_ENUM_OBS
 - attribute-type = EnumObsValue

	<ul style="list-style-type: none"> <input type="checkbox"/> attribute-value.length = <variable> <input type="checkbox"/> attribute-value = <Not relevant for this test> <p>o. IF Not recommended attribute Enum-Observed-Value-Partition is present</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id= MDC_ATTR_ENUM_OBS_VAL_PART <input type="checkbox"/> attribute-type = NomPartition (INT-U16) <input type="checkbox"/> attribute-value-length=2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> <p>7. Take a measurement with the Agent</p> <p>8. Wait for the Agent to send an event report and check:</p> <p>b. Mandatory attribute Enum-Observed-Value-Basic-Bit-Str</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id= MDC_ATTR_ENUM_OBS_VAL_BASIC_BIT_STR <input type="checkbox"/> attribute-type = OID-Type <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value= One of the following bits may be active: <ul style="list-style-type: none"> • pefm-read-stat-post-medication(0) • pefm-read-stat-cough(1) • pefm-read-stat-short-effort(2) • pefm-read-stat-long-time-to-peak(3) • The rest of the bits must not be set
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/AG/CLASS/PF/BV-012			
TP label	Association Peak expiratory flow monitor Agent			
Coverage	Spec	[ISO/IEEE 11073-10421]		
	Testable items	PF_AssocReq1; M	PF_AssocReq2; M	PF_AssocReq3; M
		PF_AssocReq4; M	PF_AssocReq5; M	PF_AssocReq6; M
		PF_AssocReq7; M	PF_AssocReq8; M	PF_AssocReq9; M
		PF_AssocReq10; M	PF_AssocReq11; M	PF_AssocReq12; M
		PF_MDSMethod4; M		
Applicability	C_AG_OXP_170 AND C_AG_OXP_000			
Initial condition	The simulated manager and the agent under test are in the unassociated state.			
Test procedure	<p>1. The agent sends a message to associate to the simulated manager, the expected fields sent by the agent are:</p> <p>a. APDU Type</p> <ul style="list-style-type: none"> <input type="checkbox"/> field- type = AarqApdu <input type="checkbox"/> field-length =2 bytes <input type="checkbox"/> field-value =0xE2 0x00. <p>b. assoc-version</p>			

- field- type = AssociationVersion
- field-length =BITS-32
- field- value=0x80 0x00 0x00 0x00
- c. data-proto-id
 - field- type = DataProtold(INT-U16)
 - field-length =2 bytes
 - field- value=0x50 0x79 (20601)
- d. protocol-version
 - field- type = Protocol Version
 - field-length = 4 bytes
 - field- value=0x80 0x00 0x00 0x00
- e. encoding rules
 - field- type = EncodingRules
 - field-length = 2 bytes
 - field- value=
 - Bit 0 must be set (support MDER)
 - Bits 1 and 2 may be set
 - The rest of the bits must be 0
- f. nomenclature version
 - field- type = NomenclatureVersion
 - field-length = 4 bytes
 - field- value=0x80 0x00 0x00 0x00
 - This value indicates version1 is supported (nom-version1(0) is set).
- g. functional–units
 - field- type = FunctionalUnits
 - field-length = 4 bytes
 - field-value =
 - Bit 0 must no be set, only bit 1 or 2 may be set to 1.
- h. System type
 - field- type = SystemType
 - field-length = 4 bytes
 - field- value = 0x00 0x80 0x00 0x00 (sys-type-agent)
- i. System-Id
 - field- type = OCTET STRING
 - field-length = 8 bytes
 - field- value = 0xXX 0xXX 0xXX 0xXX 0xXX 0xXX 0xXX 0xXX (octet string length = 8 | UI-64 manufacturer and device)
 - This value will be the System Id attribute of the MDS object and the received value will be compared with the value defined in PIXIT I_AG_OXP_001 and I_AG_OXP_002.
- j. dev-config-id
 - field- type = ConfigId(INT-U16)
 - field-length = 2 bytes
 - field- value =
 - <0x0834> for standard configuration

	<ul style="list-style-type: none"> ▪ <between 0x40 0x00 and 0x7F 0xFF > for extended configuration. <p>k. data-req-mode-flags (DataReqModeCapab)</p> <ul style="list-style-type: none"> <input type="checkbox"/> field- type = DataReqModeFlags <input type="checkbox"/> field-length = 2 bytes <input type="checkbox"/> If Agent supports only Peak expiratory flow monitor specialization → Bit 15 is set (data-req-supp-init-agent(15)) <p>l. data-req-init-agent-count (DataReqModeCapab)</p> <ul style="list-style-type: none"> <input type="checkbox"/> field- type = INT-U8 <input type="checkbox"/> field-length = 2 bytes <input type="checkbox"/> field.value = 0x01 <p>m. data-req-init-manager-count (DataReqModeCapab)</p> <ul style="list-style-type: none"> <input type="checkbox"/> field- type = INT-U8 <input type="checkbox"/> field-length = 2 bytes <input type="checkbox"/> field.value = 0x00
Pass/Fail criteria	All checked attributes have proper values.
Notes	

TP Id	TP/PLT/AG/CLASS/PF/BV-013		
TP label	Set Time Peak expiratory flow monitor Agent		
Coverage	Spec	[ISO/IEEE 11073-10421]	
	Testable items	PF_MDSMethod2; C	
Applicability	C_AG_OXP_170 AND C_AG_OXP_000 AND C_AG_OXP_009		
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<p>1. The simulated manager sends a SET action:</p> <ul style="list-style-type: none"> <input type="checkbox"/> CHOICE = SetTimeInvoke <input type="checkbox"/> action-type = MDC_ACT_SET_TIME <input type="checkbox"/> the action-info-args are SetTimeInvoke <ul style="list-style-type: none"> ▪ date-time = <century, year ≤ 99 month ≤ 12 day ≤ 31 hour ≤ 24 minute ≤ 60 s ≤ 60 sec-fractions ≤ 100> ▪ accuracy = 0 <p>2. The agent under test response shall be a rors-cmip-confirmed-action:</p> <ul style="list-style-type: none"> <input type="checkbox"/> action-type = MDC_ACT_SET_TIME <input type="checkbox"/> action-info-args shall be empty. 		
Pass/Fail criteria	All checked values are as specified in the test procedure.		
Notes			

TP Id		TP/PLT/AG/CLASS/PF/BV-014		
TP label		Operating State. Manager to Agent Maximum APDU Size		
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	CommonCharac 3; M		
	Spec	[ISO/IEEE 11073-10421]		
	Testable items	PF_ComModel1;M	PF_ComModel2;M	
Applicability		C_AG_OXP_000 AND C_AG_OXP_170		
Initial condition		The simulated manager and the agent are in the operating state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated manager issues "Remote Operation Invoke Get" command with: <ol style="list-style-type: none"> a. Obj-handle set to 0 (to request for MDS object) b. attribute-id-list.count = 103 c. attribute-id-list: (MDC_ATTR_ID_MODEL, MDC_ATTR_SYS_ID, MDC_ATTR_DEV_CONFIG_ID) repeated 34 times followed by an additional MDC_ATTR_ID_MODEL 2. Check the response of the agent. 3. The simulated manager issues a "Remote Operation Invoke Get" command with the handle set to 0 (to request for an MDS object) and an empty attribute-id-list to indicate all attributes. 4. Check the response of the agent. 		
Pass/Fail criteria		<ul style="list-style-type: none"> • In step 2, the agent under test may respond with a rors-cmip-get listing all the requested attributes, or with a roer message. If PICS C_AG_OXP_100 =TRUE and the agent does not respond with a rors-cmip-get message, it responds with a roer message or rorj (resource-limitation) message, a WARNING will appear. <ul style="list-style-type: none"> ○ If the response is a get response, the total size of the response cannot exceed the sum of the APDU sizes of the supported specializations (limited to an absolute limit of 64 512 octets): <ul style="list-style-type: none"> ▪ Pulse oximeter -> 9216 octets ▪ Weighing scales -> 896 octets ▪ Glucose meter -> 5120 octets or 64 512 octets if the agent supports PM-Store ▪ Blood pressure -> 896 octets ▪ Thermometer -> 896 octets ▪ Independent activity hub -> 5120 octets ▪ Cardiovascular -> 64 512 octets or 6624 octets if the agent under test only supports Step Counter Profile ▪ Strength -> 64 512 octets: ▪ Adherence monitor -> 1024 octets ▪ Peak flow -> 2030 octets ▪ Body composition analyser -> 7730 octets ▪ Basic ECG/Simple ECG -> 7168 octets or 64 512 octets if the agent supports PM-Store ▪ Basic ECG/Heart rate -> 1280 octets or 64 512 octets if the agent supports PM-Store ▪ International normalized ratio -> 896 octets or 64 512 if the agent supports PM-Store 		

	<ul style="list-style-type: none">○ In the case where it responds with a roer, the reason must not be protocol-violation (23).• In step 4, the agent must respond with a rors-cmip-get message.
Notes	

Bibliography

- [b-CDG 1.0] Continua Health Alliance, Continua Design Guidelines v1.0 (2008), *Continua Design Guidelines*.
- [b-CDG 2010] Continua Health Alliance, Continua Design Guidelines v1.5 (2010), *Continua Design Guidelines*.
- [b-CDG 2011] Continua Health Alliance, Continua Design Guidelines (2011), "Adrenaline", *Continua Design Guidelines*.
- [b-CDG 2012] Continua Health Alliance, Continua Design Guidelines (2012), "Catalyst", *Continua Design Guidelines*.
- [b-ETSI SR 001 262] ETSI SR 001 262 v1.8.1 (2003-12), *ETSI drafting rules*.

SERIES OF ITU-T RECOMMENDATIONS

Series A	Organization of the work of ITU-T
Series D	General tariff principles
Series E	Overall network operation, telephone service, service operation and human factors
Series F	Non-telephone telecommunication services
Series G	Transmission systems and media, digital systems and networks
Series H	Audiovisual and multimedia systems
Series I	Integrated services digital network
Series J	Cable networks and transmission of television, sound programme and other multimedia signals
Series K	Protection against interference
Series L	Construction, installation and protection of cables and other elements of outside plant
Series M	Telecommunication management, including TMN and network maintenance
Series N	Maintenance: international sound programme and television transmission circuits
Series O	Specifications of measuring equipment
Series P	Terminals and subjective and objective assessment methods
Series Q	Switching and signalling
Series R	Telegraph transmission
Series S	Telegraph services terminal equipment
Series T	Terminals for telematic services
Series U	Telegraph switching
Series V	Data communication over the telephone network
Series X	Data networks, open system communications and security
Series Y	Global information infrastructure, Internet protocol aspects and next-generation networks
Series Z	Languages and general software aspects for telecommunication systems