

International Telecommunication Union

ITU-T

TELECOMMUNICATION
STANDARDIZATION SECTOR
OF ITU

H.845.12

(04/2017)

SERIES H: AUDIOVISUAL AND MULTIMEDIA SYSTEMS

E-health multimedia services and applications –
Interoperability compliance testing of personal health
systems (HRN, PAN, LAN, TAN and WAN)

**Conformance of ITU-T H.810 personal health
system: Personal Health Devices interface
Part 5L: Body composition analyser**

Recommendation ITU-T H.845.12



ITU-T H-SERIES RECOMMENDATIONS
AUDIOVISUAL AND MULTIMEDIA SYSTEMS

CHARACTERISTICS OF VISUAL TELEPHONE SYSTEMS	H.100–H.199
INFRASTRUCTURE OF AUDIOVISUAL SERVICES	
General	H.200–H.219
Transmission multiplexing and synchronization	H.220–H.229
Systems aspects	H.230–H.239
Communication procedures	H.240–H.259
Coding of moving video	H.260–H.279
Related systems aspects	H.280–H.299
Systems and terminal equipment for audiovisual services	H.300–H.349
Directory services architecture for audiovisual and multimedia services	H.350–H.359
Quality of service architecture for audiovisual and multimedia services	H.360–H.369
Telepresence	H.420–H.429
Supplementary services for multimedia	H.450–H.499
MOBILITY AND COLLABORATION PROCEDURES	
Overview of Mobility and Collaboration, definitions, protocols and procedures	H.500–H.509
Mobility for H-Series multimedia systems and services	H.510–H.519
Mobile multimedia collaboration applications and services	H.520–H.529
Security for mobile multimedia systems and services	H.530–H.539
Security for mobile multimedia collaboration applications and services	H.540–H.549
Mobility interworking procedures	H.550–H.559
Mobile multimedia collaboration inter-working procedures	H.560–H.569
BROADBAND, TRIPLE-PLAY AND ADVANCED MULTIMEDIA SERVICES	
Broadband multimedia services over VDSL	H.610–H.619
Advanced multimedia services and applications	H.620–H.629
Ubiquitous sensor network applications and Internet of Things	H.640–H.649
IPTV MULTIMEDIA SERVICES AND APPLICATIONS FOR IPTV	
General aspects	H.700–H.719
IPTV terminal devices	H.720–H.729
IPTV middleware	H.730–H.739
IPTV application event handling	H.740–H.749
IPTV metadata	H.750–H.759
IPTV multimedia application frameworks	H.760–H.769
IPTV service discovery up to consumption	H.770–H.779
Digital Signage	H.780–H.789
E-HEALTH MULTIMEDIA SERVICES AND APPLICATIONS	
Personal health systems	H.810–H.819
Interoperability compliance testing of personal health systems (HRN, PAN, LAN, TAN and WAN)	H.820–H.859
Multimedia e-health data exchange services	H.860–H.869

For further details, please refer to the list of ITU-T Recommendations.

Recommendation ITU-T H.845.12

Conformance of ITU-T H.810 personal health system: Personal Health Devices interface Part 5L: Body composition analyser

Summary

Recommendation ITU-T H.845.12 provides a test suite structure (TSS) and the test purposes (TP) for body composition analysers in the Personal Health Devices (PHD) interface, based on the requirements defined in the Recommendations of the ITU-T H.810 sub-series, of which Recommendation ITU-T H.810 (2016) is the base Recommendation. The objective of this test specification is to provide a high probability of interoperability at this interface.

Recommendation ITU-T H.845.12 is a transposition of Continua Test Tool DG2016, Test Suite Structure & Test Purposes, Personal Health Devices Interface; Part 5L: Device Specializations. Personal Health Device (Body Composition Analyzer) (Version 1.3, 2016-09-20), that was developed by the Personal Connected Health Alliance. A number of versions of this specification existed before transposition.

This Recommendation includes an electronic attachment with the protocol implementation conformance statements (PICS) and the protocol implementation extra information for testing (PIXIT) required for the implementation of Annex A.

History

Edition	Recommendation	Approval	Study Group	Unique ID*
1.0	ITU-T H.845.12	2015-01-13	16	11.1002/1000/12272
2.0	ITU-T H.845.12	2016-07-14	16	11.1002/1000/12949
3.0	ITU-T H.845.12	2017-04-13	16	11.1002/1000/13229

Keywords

Body composition analyser, conformance testing, Continua Design Guidelines, e-health, IEEE 11073 device specialization, ITU-T H.810, personal area network, personal connected health devices, Personal Health Devices interface, touch area network.

* To access the Recommendation, type the URL <http://handle.itu.int/> in the address field of your web browser, followed by the Recommendation's unique ID. For example, <http://handle.itu.int/11.1002/1000/11830-en>.

FOREWORD

The International Telecommunication Union (ITU) is the United Nations specialized agency in the field of telecommunications, information and communication technologies (ICTs). The ITU Telecommunication Standardization Sector (ITU-T) is a permanent organ of ITU. ITU-T is responsible for studying technical, operating and tariff questions and issuing Recommendations on them with a view to standardizing telecommunications on a worldwide basis.

The World Telecommunication Standardization Assembly (WTSA), which meets every four years, establishes the topics for study by the ITU-T study groups which, in turn, produce Recommendations on these topics.

The approval of ITU-T Recommendations is covered by the procedure laid down in WTSA Resolution 1.

In some areas of information technology which fall within ITU-T's purview, the necessary standards are prepared on a collaborative basis with ISO and IEC.

NOTE

In this Recommendation, the expression "Administration" is used for conciseness to indicate both a telecommunication administration and a recognized operating agency.

Compliance with this Recommendation is voluntary. However, the Recommendation may contain certain mandatory provisions (to ensure, e.g., interoperability or applicability) and compliance with the Recommendation is achieved when all of these mandatory provisions are met. The words "shall" or some other obligatory language such as "must" and the negative equivalents are used to express requirements. The use of such words does not suggest that compliance with the Recommendation is required of any party.

INTELLECTUAL PROPERTY RIGHTS

ITU draws attention to the possibility that the practice or implementation of this Recommendation may involve the use of a claimed Intellectual Property Right. ITU takes no position concerning the evidence, validity or applicability of claimed Intellectual Property Rights, whether asserted by ITU members or others outside of the Recommendation development process.

As of the date of approval of this Recommendation, ITU had not received notice of intellectual property, protected by patents, which may be required to implement this Recommendation. However, implementers are cautioned that this may not represent the latest information and are therefore strongly urged to consult the TSB patent database at <http://www.itu.int/ITU-T/ipr/>.

© ITU 2017

All rights reserved. No part of this publication may be reproduced, by any means whatsoever, without the prior written permission of ITU.

Table of Contents

	Page
1	Scope..... 1
2	References..... 2
3	Definitions 2
3.1	Terms defined elsewhere 2
3.2	Terms defined in this Recommendation 2
4	Abbreviations and acronyms 2
5	Conventions 3
6	Test suite structure (TSS) 5
7	Electronic attachment 7
Annex A	Test purposes 8
A.1	TP definition conventions..... 8
A.2	Subgroup 1.3.12 – Body composition analyser (BCA)..... 10
Bibliography 42

Electronic attachment: This Recommendation includes an electronic attachment with the protocol implementation conformance statements (PICS) and the protocol implementation extra information for testing (PIXIT) required for the implementation of Annex A.

Introduction

This Recommendation is a transposition of Continua Test Tool DG2016, Test Suite Structure & Test Purposes, Personal Health Devices Interface; Part 5L: Device Specializations. Personal Health Device (Body Composition Analyzer) (Version 1.3, 2016-09-20), that was developed by the Personal Connected Health Alliance. The table below shows the revision history of this test specification; it may contain versions that existed before transposition.

Version	Date	Revision history
1.0	2013-05-24	Initial release for Test Tool DG2012 based on the requirements in [b-CDG 2012].
1.1	2014-01-24	Initial release for Test Tool DG2013. This uses "TSS&TP_DG2012_PAN-LAN_PART_5L_v1.0.doc" as a baseline and adds new features included in [b-ITU-T H.810 (2013)]/[b-CDG 2013]: <ul style="list-style-type: none">• Adds glucose meter BLE• Adds BLE SSP support• Adds NFC new transport• Adds INR device specialization
1.2	2014-04-24	TM Lite & Doc Enhancements (Test Tool v4.0 Maintenance Release 1). It uses "TSS&TP_DG2013_PLT_PART_5L_v1.1.doc" as a baseline and adds new features included in Documentation Enhancements: <ul style="list-style-type: none">• "Other PICS" row added
1.2	2015-07-01	Initial release for Test Tool DG2015. It is the same version as "TSS&TP_DG2013_PLT_PART_5L_v1.1.doc" because new features included in [b-ITU-T H.810 (2015)]/[b-CDG 2015] do not affect the test procedures specified in this document.
1.3	2016-09-20	Initial release for Test Tool DG2016. It uses "TSS&TP_DG2015_PLT_PART_5L_v1.2.doc" as a baseline and adds new features included in [ITU-T H.810 (2016)]/[b-CDG 2016]

Recommendation ITU-T H.845.12

Conformance of ITU-T H.810 personal health system: Personal Health Devices interface Part 5L: Body composition analyser

1 Scope

The scope of this Recommendation¹ is to provide a test suite structure (TSS) and the test purposes (TP) for the Personal Health Devices interface based on the requirements defined in the Continua Design Guidelines (CDG) [ITU-T H.810 (2016)]. The objective of this test specification is to provide a high probability of interoperability at this interface.

The TSS and TP for the Personal Health Devices interface have been divided into the parts specified below. This Recommendation covers Part 5, subpart 5L.

- Part 1: Optimized exchange protocol. Personal Health Device
- Part 2: Optimized exchange protocol. Personal Health Gateway
- Part 3: Continua design guidelines. Personal Health Device
- Part 4: Continua design guidelines. Personal Health Gateway
- Part 5: Device specializations. Personal Health Devices interface. This document is divided into the following subparts:
 - Part 5A: Weighing scales
 - Part 5B: Glucose meter
 - Part 5C: Pulse oximeter
 - Part 5D: Blood pressure monitor
 - Part 5E: Thermometer
 - Part 5F: Cardiovascular fitness and activity monitor
 - Part 5G: Strength fitness equipment
 - Part 5H: Independent living activity hub
 - Part 5I: Adherence monitor
 - Part 5J: Insulin pump
 - Part 5K: Peak expiratory flow monitor
 - **Part 5L: Body composition analyser**
 - Part 5M: Basic electrocardiograph
 - Part 5N: International normalized ratio monitor
 - Part 5O: Sleep apnoea breathing therapy equipment (SABTE)
 - Part 5P: Continuous glucose monitor (CGM)
- Part 6: Device specializations. Personal Health Gateway
- Part 7: Continua Design Guidelines. BLE Personal Health Device

¹ This Recommendation includes an electronic attachment with the protocol implementation conformance statements (PICS) and the protocol implementation extra information for testing (PIXIT) required for the implementation of Annex A.

- Part 8: Continua Design Guidelines. BLE Personal Health Gateway
- Part 9: Personal Health Devices Transcoding Whitepaper. Personal Health Devices
- Part 10: Personal Health Devices Transcoding Whitepaper. Personal Health Gateway

2 References

The following ITU-T Recommendations and other references contain provisions which, through reference in this text, constitute provisions of this Recommendation. At the time of publication, the editions indicated were valid. All Recommendations and other references are subject to revision; users of this Recommendation are therefore encouraged to investigate the possibility of applying the most recent edition of the Recommendations and other references listed below. A list of the currently valid ITU-T Recommendations is regularly published. The reference to a document within this Recommendation does not give it, as a stand-alone document, the status of a Recommendation.

- | | |
|------------------------------|--|
| [ITU-T H.810 (2016)] | Recommendation ITU-T H.810 (2016), <i>Interoperability design guidelines for personal health systems</i> . |
| [ISO/IEEE 11073-10420] | ISO/IEEE 11073-10420-2012, <i>Health informatics – Personal health device communication – Part 10420: Device specialization – Body composition analyzer</i> .
https://www.iso.org/standard/61055.html |
| [ISO/IEEE 11073-20601-2015A] | ISO/IEEE 11073-20601:2010, <i>Health informatics – Personal health device communication – Part 20601: Application profile – Optimized exchange protocol</i> , including ISO/IEEE 11073-20601:2010 Amd 1:2015.
https://www.iso.org/standard/54331.html with
https://www.iso.org/standard/63972.html |
| [ISO/IEEE 11073-20601-2016C] | ISO/IEEE 11073-20601:2016, <i>Health informatics – Personal health device communication – Part 20601: Application profile – Optimized exchange protocol</i> , including ISO/IEEE 11073-20601:2016/Cor.1:2016.
https://www.iso.org/standard/66717.html with
https://www.iso.org/standard/71886.html |

3 Definitions

3.1 Terms defined elsewhere

This Recommendation uses the following terms defined elsewhere:

3.1.1 agent [ISO/IEEE 11073-20601-2016C]: A node that collects and transmits personal health data to an associated manager.

3.1.2 manager [ISO/IEEE 11073-20601-2016C]: A node receiving data from one or more agent systems. Some examples of managers include a cellular phone, health appliance, set top box, or a computer system.

3.2 Terms defined in this Recommendation

None.

4 Abbreviations and acronyms

This Recommendation uses the following abbreviations and acronyms:

ATS Abstract Test Suite

DUT	Device Under Test
CDG	Continua Design Guidelines
CGM	Continuous Glucose Monitor
GUI	Graphical User Interface
INR	International Normalized Ratio
IP	Insulin Pump
IUT	Implementation Under Test
MDS	Medical Device System
NFC	Near Field Communication
PAN	Personal Area Network
PCT	Protocol Conformance Testing
PCO	Point of Control and Observation
PHD	Personal Health Device
PHDC	Personal Healthcare Device Class
PHG	Personal Health Gateway
PICS	Protocol Implementation Conformance Statement
PIXIT	Protocol Implementation extra Information for Testing
SABTE	Sleep Apnoea Breathing Therapy Equipment
SCR	Static Conformance Review
SDP	Service Discovery Protocol
SOAP	Simple Object Access Protocol
TCRL	Test Case Reference List
TCWG	Test and Certification Working Group
TP	Test Purpose
TSS	Test Suite Structure
USB	Universal Serial Bus
WDM	Windows Driver Model

5 Conventions

The key words "SHALL", "SHALL NOT", "SHOULD", "SHOULD NOT", "MAY", "MAY NOT" in this Recommendation are to be interpreted as in [b-ETSI SR 001 262].

- SHALL is equivalent to 'must' or 'it is required to'.
- SHALL NOT is equivalent to 'must not' or 'it is not allowed'.
- SHOULD is equivalent to 'it is recommended to'.
- SHOULD NOT is equivalent to 'it is not recommended to'.
- MAY is equivalent to 'is permitted'.
- MAY NOT is equivalent to 'it is not required that'.

NOTE – The above-mentioned key words are capitalized for illustrative purposes only and they do not appear capitalized within this Recommendation.

Reference is made in the ITU-T H.800-series of Recommendations to different versions of the Continua Design Guidelines (CDG) by a specific designation. The list of terms that may be used in this Recommendation is provided in Table 1.

Table 1 – List of designations associated with the various versions of the CDG

CDG release	Transposed as	Version	Description	Designation
2016 plus errata	[ITU-T H.810 (2016)]	6.1	Release 2016 plus errata noting all ratified bugs [b-CDG 2016].	–
2016	–	6.0	Release 2016 of the CDG including maintenance updates of the CDG 2015 and additional guidelines that cover new functionalities.	Iris
2015 plus errata	[b-ITU-T H.810 (2015)]	5.1	Release 2015 plus errata noting all ratified bugs [b-CDG 2015]. The 2013 edition of H.810 is split into eight parts in the H.810-series.	–
2015	–	5.0	Release 2015 of the CDG including maintenance updates of the CDG 2013 and additional guidelines that cover new functionalities.	Genome
2013 plus errata	[b-ITU-T H.810 (2013)]	4.1	Release 2013 plus errata noting all ratified bugs [b-CDG 2013].	–
2013	–	4.0	Release 2013 of the CDG including maintenance updates of the CDG 2012 and additional guidelines that cover new functionalities.	Endorphin
2012 plus errata	–	3.1	Release 2012 plus errata noting all ratified bugs [b-CDG 2012].	–
2012	–	3.0	Release 2012 of the CDG including maintenance updates of the CDG 2011 and additional guidelines that cover new functionalities.	Catalyst
2011 plus errata	–	2.1	CDG 2011 integrated with identified errata.	–
2011	–	2.0	Release 2011 of the CDG including maintenance updates of the CDG 2010 and additional guidelines that cover new functionalities [b-CDG 2011].	Adrenaline
2010 plus errata	–	1.6	CDG 2010 integrated with identified errata	–
2010	–	1.5	Release 2010 of the CDG with maintenance updates of the CDG Version 1 and additional guidelines that cover new functionalities [b-CDG 2010].	1.5
1.0	–	1.0	First released version of the CDG [b-CDG 1.0].	–

6 Test suite structure (TSS)

The test purposes (TPs) for the Personal Health Devices interface have been divided into the main subgroups specified below. Annex A describes the TPs for subgroup 1.3.12 (shown in bold).

- Group 1: Personal Health Device (PHD)
 - Group 1.1: Transport (TR)
 - Subgroup 1.1.1: Design guidelines: Common (DGC)
 - Subgroup 1.1.2: USB design guidelines (UDG)
 - Subgroup 1.1.3: Bluetooth design guidelines (BDG)
 - Subgroup 1.1.4: Pulse oximeter design guidelines (PODG)
 - Subgroup 1.1.5: Cardiovascular design guidelines (CVDG)
 - Subgroup 1.1.6: Activity hub design guidelines (HUBDG)
 - Subgroup 1.1.7: ZigBee design guidelines (ZDG)
 - Subgroup 1.1.8: Glucose meter design guidelines (GLDG)
 - Subgroup 1.1.9: Bluetooth low energy design guidelines (BLEDG)
 - Subgroup 1.1.10: Basic electrocardiograph design guidelines (ECGDG)
 - Subgroup 1.1.11: NFC design guidelines (NDG)
 - Group 1.2: IEEE 20601 Optimized exchange protocol (OXP)
 - Subgroup 1.2.1: PHD domain information model (DIM)
 - Subgroup 1.2.2: PHD service model (SER)
 - Subgroup 1.2.3: PHD communication model (COM)
 - Group 1.3: Devices class specializations (CLASS)
 - Subgroup 1.3.1: Weighing scales (WEG)
 - Subgroup 1.3.2: Glucose meter (GL)
 - Subgroup 1.3.3: Pulse oximeter (PO)
 - Subgroup 1.3.4: Blood pressure monitor (BPM)
 - Subgroup 1.3.5: Thermometer (TH)
 - Subgroup 1.3.6: Cardiovascular (CV)
 - Subgroup 1.3.7: Strength (ST)
 - Subgroup 1.3.8: Activity hub (HUB)
 - Subgroup 1.3.9: Adherence monitor (AM)
 - Subgroup 1.3.10: Insulin pump (IP)
 - Subgroup 1.3.11: Peak flow (PF)
 - **Subgroup 1.3.12: Body composition analyser (BCA)**
 - Subgroup 1.3.13: Basic electrocardiograph (ECG)
 - Subgroup 1.3.14: International normalized ratio (INR)
 - Subgroup 1.3.15: Sleep apnoea breathing therapy equipment (SABTE)
 - Subgroup 1.3.16: Continuous glucose monitor (CGM)
 - Group 1.4: Personal health device transcoding whitepaper (PHDTW)
 - Subgroup 1.4.1: Whitepaper general requirements (GEN)
 - Subgroup 1.4.2: Whitepaper thermometer requirements (TH)
 - Subgroup 1.4.3: Whitepaper blood pressure requirements (BPM)

- Subgroup 1.4.4: Whitepaper heart rate requirements (HR)
 - Subgroup 1.4.5: Whitepaper glucose meter requirements (GL)
 - Subgroup 1.4.6: Whitepaper weight scale requirements (WS)
 - Subgroup 1.4.7: Whitepaper pulse oximeter requirements (PLX)
 - Subgroup 1.4.8: Whitepaper continuous glucose monitoring requirements (CGM)
- Group 2: Personal Health Gateway (PHG)
- Group 2.1: Transport (TR)
 - Subgroup 2.1.1: Design guidelines: Common (DGC)
 - Subgroup 2.1.2: USB design guidelines (UDG)
 - Subgroup 2.1.3: Bluetooth design guidelines (BDG)
 - Subgroup 2.1.4: Cardiovascular design guidelines (CVDG)
 - Subgroup 2.1.5: Activity hub design guidelines (HUBDG)
 - Subgroup 2.1.6: ZigBee design guidelines (ZDG)
 - Subgroup 2.1.7: Bluetooth low energy design guidelines (BLEDDG)
 - Subgroup 2.1.8: NFC design guidelines (NDG)
 - Group 2.2: IEEE 20601 Optimized exchange protocol (OXP)
 - Subgroup 2.2.1: General (GEN)
 - Subgroup 2.2.2: PHD domain information model (DIM)
 - Subgroup 2.2.3: PHD service model (SER)
 - Subgroup 2.2.4: PHD communication model (COM)
 - Group 2.3: Devices class specializations (CLASS)
 - Subgroup 2.3.1: Weighing scales (WEG)
 - Subgroup 2.3.2: Glucose meter (GL)
 - Subgroup 2.3.3: Pulse oximeter (PO)
 - Subgroup 2.3.4: Blood pressure monitor (BPM)
 - Subgroup 2.3.5: Thermometer (TH)
 - Subgroup 2.3.6: Cardiovascular (CV)
 - Subgroup 2.3.7: Strength (ST)
 - Subgroup 2.3.8: Activity hub (HUB)
 - Subgroup 2.3.9: Adherence monitor (AM)
 - Subgroup 2.3.10: Insulin pump (IP)
 - Subgroup 2.3.11: Peak flow (PF)
 - Subgroup 2.3.12: Body composition analyser (BCA)
 - Subgroup 2.3.13: Basic electrocardiograph (ECG)
 - Subgroup 2.3.14: International normalized ratio (INR)
 - Subgroup 2.3.15: Sleep apnoea breathing therapy equipment (SABTE)
 - Subgroup 2.3.16: Continuous glucose monitor (CGM)
 - Group 2.4: Personal health device transcoding whitepaper (PHDTW)
 - Subgroup 2.4.1: Whitepaper general requirements (GEN)
 - Subgroup 2.4.2: Whitepaper thermometer requirements (TH)
 - Subgroup 2.4.3: Whitepaper blood pressure measurement requirements (BPM)

- Subgroup 2.4.4: Whitepaper heart rate requirements (HR)
- Subgroup 2.4.5: Whitepaper glucose meter requirements (GL)
- Subgroup 2.4.6: Whitepaper weight scale requirements (WS)
- Subgroup 2.4.7: Whitepaper pulse oximeter requirements (PLX)
- Subgroup 2.4.8: Whitepaper continuous glucose monitoring requirements (CGM)

7 Electronic attachment

The protocol implementation conformance statements (PICS) and the protocol implementation extra information for testing (PIXIT) required for the implementation of Annex A can be downloaded from <http://handle.itu.int/11.1002/2000/12067>.

In the electronic attachment, letters "C" and "I" in the column labelled "Mandatory" are used to distinguish between "PICS" and "PIXIT" respectively during testing. If the cell is empty, the corresponding PICS is "independent". If the field contains a "C", the corresponding PICS is dependent on other PICS, and the logical expression is detailed in the "SCR_Expression" field. The static conformance review (SCR) is used in the test tool to assert whether the PICS selection is consistent.

Annex A

Test purposes

(This annex forms an integral part of this Recommendation.)

A.1 TP definition conventions

The test purposes (TPs) are defined according to the following rules:

- **TP Id:** This is a unique identifier (TP/<TT>/<DUT>/<GR>/<SGR>/<XX> – <NNN>). It is specified according to the naming convention defined below:
 - Each test purpose identifier is introduced by the prefix "TP".
 - <TT>: This is the test tool that will be used in the test case:
 - PAN: Personal area network (Bluetooth or USB)
 - LAN: Local area network (ZigBee)
 - PAN-LAN: Personal area network (Bluetooth or USB) – Local area network (ZigBee)
 - LP-PAN: Low power personal area network (Bluetooth Low Energy)
 - TAN: Touch area network (NFC)
 - PLT: Personal area network (Bluetooth or USB) – Local area network (ZigBee) – Touch area network (NFC)
 - <DUT>: This is the device under test:
 - PHD: Personal Health Device
 - PHG: Personal Health Gateway
 - <GR>: This identifies a group of test cases.
 - <SGR>: This identifies a subgroup of test cases.
 - <XX>: This identifies the type of testing:
 - BV: Valid behaviour test
 - BI: Invalid behaviour test
 - <NNN>: This is a sequential number that identifies the test purpose.
- **TP label:** This is the TP's title.
- **Coverage:** This contains the specification reference and clause to be checked by the TP:
 - Spec: This indicates the earliest version of the specification from which the testable items to be checked by the TP were included.
 - Testable item: This contains the testable items to be checked by the TP.
- **Test purpose:** This is a description of the requirements to be tested.
- **Applicability:** This contains the PICS items that define if the test case is applicable or not for a specific device. When a TP contains an "ALL" in this field it means that it applies to the device under test within that scope of the test (specialization, transport used, etc.).
- **Other PICS:** This contains additional PICS items (apart from the PICS specified in the Applicability row) which are used within the test case implementation and can modify the final verdict. When this row is empty, it means that only the PICS specified in the Applicability row are used within the test case implementation.

- **Initial condition:** This indicates the state to which the DUT needs to be moved at the beginning of TC execution.
- **Test procedure:** This describes the steps to be followed in order to execute the test case.
- **Pass/Fail criteria:** This provides criteria to decide whether the DUT passes or fails the test case.

A.2 Subgroup 1.3.12 – Body composition analyser (BCA)

TP Id		TP/PLT/PHD/CLASS/BCA/BV-000		
TP label		Get MDS Object for body composition analyser specialization: Mandatory, Conditional and Optional Attributes		
Coverage	Spec	[ISO/IEEE 11073-10420]		
	Testable items	MDSClassAttr 1; M	MDSClassAttr 2; M	MDSClassAttr 3; M
		MDSClassAttr 4; M	MDSClassAttr 5; M	MDSClassAttr 6; M
		MDSClassAttr 7; M	MDSClassAttr 8; R	MDSClassAttr 9; R
		MDSClassAttr 10; R	MDSServices 1; M	MDSServices 3; M
OperProc2; M				
Test purpose		<p>Check that:</p> <p>The PHD supports a Get command that requests all attributes</p> <p>[AND]</p> <p>The MDS Object contains the attributes specified for a body composition analyser PHD.</p>		
Applicability		C_AG_OXP_167 AND C_AG_OXP_000		
Other PICS		C_AG_OXP_181		
Initial condition		The simulated PHG and the PHD under test are in the Operating state.		
Test procedure		<ol style="list-style-type: none"> The simulated PHG issues a "roiv-cmip-get" command with the handle set to 0 (to request for an MDS object) and the attribute-id-list set to 0 to indicate all attributes. The PHD under test responds with a "rors-cmip-get" service message in which the attribute-list contains a list of all implemented attributes of the MDS object: <ul style="list-style-type: none"> MDS Attributes: <ol style="list-style-type: none"> Conditional attribute System-Type shall not be present. <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SYS_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <not relevant > Mandatory attribute System-Type-Spec_List <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SYS_TYPE_SPEC_LIST <input type="checkbox"/> attribute-type = TypeVerList <input type="checkbox"/> attribute-value.length = 4 bytes attribute-value = MDC_DEV_SPEC_PROFILE_BCA, 1 Mandatory attribute System-model <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_MODEL <input type="checkbox"/> attribute-type = SystemModel <input type="checkbox"/> attribute-value.length = <variable> <input type="checkbox"/> attribute-value = {Manufacturer, Model} Mandatory attribute Dev-Configuration-Id <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_DEV_CONFIG_ID <input type="checkbox"/> attribute-type = ConfigId <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <ul style="list-style-type: none"> – IF NOT C_AG_OXP_181 then attribute-value = 0x07D0 – ELSE attribute-value = < between 0x4000 and 0x7FFF> 		

	<p>e. Recommended attribute Power-Status</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_POWER_STAT <input type="checkbox"/> attribute-type = PowerStatus (BITS-16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = ON_MAINS (0x8000) or ON_BATTERY(0x4000) <p>Only one of the following may be active:</p> <ul style="list-style-type: none"> ▪ chargingFull(8), ▪ chargingTrickle(9), ▪ chargingOff(10). ▪ The rest of the bits must not be set <p>f. Recommended attribute Remain-Battery-Time</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_BATT_REMAIN (0X09 0X88) <input type="checkbox"/> attribute-type = BatMeasure <input type="checkbox"/> attribute-value.length = 6 bytes <input type="checkbox"/> attribute-value = <4 bytes to define the value. 2 remaining bytes to define the units, which shall be set to one of: MDC_DIM_MIN (0x08 0xA0), MDC_DIM_HR (0x08 0xC0) or MDC_DIM_DAY (0x08 0xE0) >
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/PHD/CLASS/BCA/BV-001			
TP label	MDS Configuration objects events for body composition analyser PHD			
Coverage	Spec	[ISO/IEEE 11073-10420]		
	Testable items	MDSEvents 1; M	BCA_NumGen1; M	BodyFat1; M
		BodyHeight1; M	WeightNumClass 1;M	BodyMassIndex1; O
		FatFreeMass1; O	SoftLeanMass1; O	BodyWater1; O
BCAExtRules3; M		ConfigProc1; M		
Test purpose	<p>Check that:</p> <p>A body composition analyser PHD shall send the [MDS-Configuration-Event] using a [Confirmed] event report. The [MDS-Configuration-Event] shall include the event-info [ConfigReport].</p> <p>[AND]</p> <p>Check objects supported by the PHD (standard /extended configuration)</p>			
Applicability	C_AG_OXP_167 AND C_AG_OXP_000			
Other PICS	C_AG_OXP_010, C_AG_OXP_181			
Initial condition	The simulated PHG and the PHD under test are in the Unassociated state.			
Test procedure	<ol style="list-style-type: none"> 1. The simulated PHG receives an association request from the PHD under test. 2. The simulated PHG responds with a result = accepted-unknown-config. 3. The PHD responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the PHG: <ol style="list-style-type: none"> a. APDU Type <ul style="list-style-type: none"> <input type="checkbox"/> field- type = PrstApdu <input type="checkbox"/> field-length = 2 bytes <input type="checkbox"/> field-value = 0xE7 0x00 			

	<ul style="list-style-type: none"> b. invoke-id <ul style="list-style-type: none"> <input type="checkbox"/> field- type = InvokeIDType <input type="checkbox"/> field-length = INT-U16 <input type="checkbox"/> field- value = <Not relevant for this test> c. message <ul style="list-style-type: none"> <input type="checkbox"/> field- type = roiv-cmip-confirmed-event-report <input type="checkbox"/> field-length = two bytes <input type="checkbox"/> field- value = 0x01 0x01 (EventReportArgumentSimple) d. obj-handle (EventReportArgumentSimple) <ul style="list-style-type: none"> <input type="checkbox"/> field- type = HANDLE <input type="checkbox"/> field-length = INT-U16 e. event-time (EventReportArgumentSimple) <ul style="list-style-type: none"> <input type="checkbox"/> field- type = Relative Time <input type="checkbox"/> field-length = INT-U32 <input type="checkbox"/> field-value = <ul style="list-style-type: none"> • IF NOT C_AG_OXP_010 THEN value = 0xFF 0xFF 0xFF 0xFF f. event-type (EventReportArgumentSimple) <ul style="list-style-type: none"> <input type="checkbox"/> field- type = OID-Type <input type="checkbox"/> field-length = INT-U16 <input type="checkbox"/> field- value = 0x0D 0x1C (MDC_NOTI_CONFIG) g. config-report-id (ConfigReport) <ul style="list-style-type: none"> <input type="checkbox"/> field- type = ConfigId <input type="checkbox"/> field-length = INT-U16 <input type="checkbox"/> field- value = <ul style="list-style-type: none"> – IF NOT C_AG_OXP_181 then attribute-value = 0x07D0 – ELSE attribute-value = < between 0x4000 and 0x7FFF > h. obj-class (ConfigReport → ConfigObjectList (ConfigObject)). To check the objects that are supported by the PHD, Type Attribute will be checked in AttributeList. <ul style="list-style-type: none"> <input type="checkbox"/> field- type = OID-Type <input type="checkbox"/> field-length = INT-U16 <input type="checkbox"/> field- value = <ul style="list-style-type: none"> – Three mandatory numeric objects for body fat, body height and body weight. – Four optional numeric objects for body mass index, fat free mass, soft lean mass and body water.
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/PHD/CLASS/BCA/BV-002			
TP label	MDS objects events Body composition analyser			
Coverage	Spec	[ISO/IEEE 11073-10420]		
	Testable items	MDSEvents 3; M	MDSEvents 4; M	MDSEvents 5; M
		MDSEvents 6; M	ObjAccServ1; M	ObjAccServ2; M
	ObjAccServ3; M	ObjAccServ6; O		
Test purpose	Check that:			

	<p>The PHD sends the MDS-Dynamic-Data-Update-Fixed using a confirmed event report and it includes the event-info ScanReportInfoFixed.</p> <p>[AND/OR]</p> <p>The PHD sends the MDS-Dynamic-Data-Update-Var using a confirmed event report and it includes the event-info ScanReportInfoVar.</p> <p>[AND]</p> <p>Event reports shall be used in confirmed mode.</p> <p>[AND]</p> <p>Agent-initiated mode shall be supported for measurement data transmission.</p> <p>[AND]</p> <p>A body composition analyser PHD may support either one or both single-person and multi-person event reports</p> <p>[AND]</p> <p>A body composition analyser PHD with standard configuration shall use the fixed format data update messages method for transmitting measurement data</p> <p>[AND]</p> <p>A body composition analyser PHD with extended configuration may use either fixed or variable format data update messages for transmitting measurement data.</p>
Applicability	C_AG_OXP_167 AND C_AG_OXP_000 AND (C_AG_OXP_182 OR C_AG_OXP_183 OR C_AG_OXP_184 OR C_AG_OXP_189)
Other PICS	C_AG_OXP_009, C_AG_OXP_014, C_AG_OXP_181, C_AG_OXP_293
Initial condition	The simulated PHG and the PHD under test are in the Unassociated state.
Test procedure	<ol style="list-style-type: none"> 1. The simulated PHG receives an association request from the PHD under test. 2. The simulated PHG responds with a result = accepted-unknown-config. 3. The PHD under test responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the PHG. 4. Check that the field Dev-Config-Id is set to the tested configuration. If it is not, the PHG responds with an "unsupported-config" and waits for a new configuration. Repeat this step until a Dev-config-Id equal to the tested configuration is received. 5. Record the PHD configuration. 6. IF C_AG_OXP_293: <ol style="list-style-type: none"> a. Once in Configuring/Sending GetMDS substate simulated PHG issues roiv-cmip-get command with handle set to 0 (to request for MDS object) and attribute-id-list set to 0 to indicate all attributes. b. The PHD responds with a rors-cmip-get service message in which the attribute-list contains a list of all implemented attributes of the MDS object. c. IF the mds-time-mgr-set-time bit is set: <ol style="list-style-type: none"> <input type="checkbox"/> The PHG moves to Configuring/Sending Set Time substate and: <ul style="list-style-type: none"> • IF C_AG_OXP_009 it issues the Set-Time action command. • IF C_AG_OXP_014 it issues the Set-Base-Offset-Time action command. <input type="checkbox"/> Once its internal time setting operation is completed, the PHD responds to the PHG. 7. Take Measurements for every supported object in the PHD under test. 8. Wait to receive every event report and check: <ol style="list-style-type: none"> <input type="checkbox"/> field- type = Event Report <input type="checkbox"/> field-length = 2 bytes <input type="checkbox"/> field- value = 0x01 0x01 (EventReportArgumentSimple, confirmed) This field identifies the type of message sent by the PHD, for the confirmed event configuration, roiv-cmip-confirmed-event-report.

Pass/Fail criteria	<ul style="list-style-type: none"> • Check that every received MDS Event report is one of the following Data APDU and that it is confirmed. • For Standard Configuration (NOT C_AG_OXP_181): An MDS Event Report is sent by the PHD under test to report measurements for every object. <ul style="list-style-type: none"> • MDC_NOTI_SCAN_REPORT_FIXED • MDC_NOTI_SCAN_REPORT_MP_FIXED • For Extended Configuration, an MDS Event Report is sent by the PHD under test to report measurements for every object: <ul style="list-style-type: none"> • MDC_NOTI_SCAN_REPORT_FIXED • MDC_NOTI_SCAN_REPORT_MP_FIXED • MDC_NOTI_SCAN_REPORT_VAR • MDC_NOTI_SCAN_REPORT_MP_VAR
Notes	

TP Id	TP/PLT/PHD/CLASS/BCA/BV-003		
TP label	Body Weight Object for Standard Configuration (0x07D0)		
Coverage	Spec	[ISO/IEEE 11073-10420]	
	Testable items	WeightNumClass 1; M	WeightNumClass 2; M
		WeightNumClass 5; R	WeightNumClass 7; M
		WeightNumClass 11; O	WeightNumClass 13; R
		WeightNumClass 17; R	WeightNumClass 19; M
		WeightNumClass 22; R	WeightNumClass 24; R
		WeightNumClass 27; R	WeightNumClass 29; C
		Concepts 4; M	
Test purpose	Check that: Body Weight Numeric Object contains the attributes specified for Standard Configuration (0x07D0)		
Applicability	C_AG_OXP_167 AND (NOT C_AG_OXP_181) AND C_AG_OXP_000		
Other PICS			
Initial condition	The simulated PHG and the PHD under test are in the Unassociated state.		
Test procedure	<ol style="list-style-type: none"> 1. The simulated PHG receives an association request from the PHD under test. 2. The simulated PHG responds with a result = accepted-unknown-config. 3. The PHD responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the PHG. 4. Check that the field Dev-Config-Id is set to 0x07D0. If it is not, the PHG responds with an "unsupported-config" and waits for a new configuration. Repeat this step until a Dev-config-Id equal to 0x07D0 is received. 5. Once the PHD under test sends a standard configuration, check the Body Weight object: 6. The Body Weight object contents shall be: <ol style="list-style-type: none"> a. Mandatory attribute Handle <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_HANDLE <input type="checkbox"/> attribute-type = HANDLE <input type="checkbox"/> attribute-value = 0x00 0x01 b. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE 		

	<ul style="list-style-type: none"> <input type="checkbox"/> attribute-value = 0x00 0x02(MDC_PART_SCADA), 0xE1 0x40(MDC_MASS_BODY_ACTUAL 57664) <p>c. Mandatory attribute Metric-Spec-Small</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value ≠ 0x00 0x00 <ul style="list-style-type: none"> • Bit 0 (mss-avail-intermittent(0)) is set. • Bit 1 (mss-avail-stored-data(1)) is set. • Bit 2 (mss-upd-aperiodic(2)) is set. • Bit 3 (mss-msmt-aperiodic(3)) is set. • Bit 9 (mss-acc-agent-initiated(9)) is set. • Bit 12 (mss-cat-manual(12)) is set. <p>d. Mandatory attribute Unit-Code</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE <input type="checkbox"/> attribute-type = OID-Type <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_DIM_KILO_G <p>e. Mandatory attribute Attribute-Value-Map</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ATTRIBUTE_VAL_MAP <input type="checkbox"/> attribute-type = AttrValMap <input type="checkbox"/> attribute-count = 2 <input type="checkbox"/> attribute-value = (MDC_ATTR_NU_VAL_OBS_SIMP,4 MDC_ATTR_TIME_STAMP_ABS,8) <p>7. Check that no other attributes are present in the initial configuration.</p>
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/PHD/CLASS/BCA/BV-004			
TP label	Body Weight Object for Extended Configuration			
Coverage	Spec	[ISO/IEEE 11073-10420]		
	Testable items	WeightNumClass 1; M	WeightNumClass 4; M	WeightNumClass 6; R
		WeightNumClass 8; M	WeightNumClass 10; R	WeightNumClass 12; R
		WeightNumClass 14; R	WeightNumClass 16; R	WeightNumClass 18; R
		WeightNumClass 20; M	WeightNumClass 23; R	WeightNumClass 25; R
		WeightNumClass 28; R	Concepts 4; M	
Test purpose	Check that: Body Weight Numeric Object contains the attributes specified for Extended Configuration			
Applicability	C_AG_OXP_167 AND C_AG_OXP_181 AND C_AG_OXP_000			
Other PICS				
Initial condition	The simulated PHG and the PHD under test are in the Unassociated state.			
Test procedure	<ol style="list-style-type: none"> 1. The simulated PHG receives an association request from the PHD under test. 2. The simulated PHG responds with a result = accepted-unknown-config. 3. The PHD under test responds with a "Remote Operation Invoke Confirmed Event 			

	<p>Report" message with an MDC_NOTI_CONFIG event to send its configuration to the PHG.</p> <ol style="list-style-type: none"> 4. Check that the field Dev-Config-Id is set to the tested extended configuration. If it is not, the PHG responds with an "unsupported-config" and waits for a new configuration. Repeat this step until a Dev-config-Id equal to the extended configuration is received. 5. Once the PHD under test sends the tested configuration, check the Body Weight object: 6. The Body Weight object contents shall be: <ol style="list-style-type: none"> a. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE attribute-value = 0x00 0x02(MDC_PART_SCADA), 0xE1 0x40(MDC_MASS_BODY_ACTUAL 57664) b. IF Not Recommended attribute Supplemental-Types <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SPPLEMENTAL_TYPES <input type="checkbox"/> attribute-type = SupplementalTypeList <input type="checkbox"/> attribute-value.length = <variable>Sequence of TYPE (TYPE.length = 4 bytes) <input type="checkbox"/> attribute-value = <Not relevant for this test> c. Mandatory attribute Metric-Spec-Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value ≠ 0x00 0x00 <ul style="list-style-type: none"> • Bit 0 (mss-avail-intermittent(0)) is set. • Bit 1 (mss-avail-stored-data(1)) is set. • Bit 2 (mss-upd-aperiodic(2)) is set. • Bit 3 (mss-msmt-aperiodic(3)) is set. • Bit 9 (mss-acc-agent-initiated(9)) is set. d. IF Not recommended attribute Metric-Structure-Small is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_STRUCTURE_SMALL <input type="checkbox"/> attribute-type = MetricStructureSmall <input type="checkbox"/> attribute-length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> e. IF Recommended attribute Measurement-Status is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_MSMT_STAT <input type="checkbox"/> attribute-type = MeasurementStatus <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> f. IF Not recommended attribute Metric-Id is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_PHYSIO <input type="checkbox"/> attribute-type = OID-Type(INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> g. IF Not Recommended attribute Metric-Id-List is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_PHYSIO_LIS <input type="checkbox"/> attribute-type = MetricIdList <input type="checkbox"/> attribute-value = <Not relevant for this test>
--	---

	<ul style="list-style-type: none"> h. IF Not recommended attribute Metric-Id-Partition is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_ID_PART <input type="checkbox"/> attribute-type = NomPartition(INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> i. Mandatory attribute Unit-Code <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE <input type="checkbox"/> attribute-type = OID-Type <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_DIM_KILO_G OR MDC_DIM_LB j. IF Not recommended attribute Source-Handle-Reference is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SOURCE_HANDLE_REF <input type="checkbox"/> attribute-type = HANDLE(INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> k. IF Not recommended attribute Measure-Active-Period <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_PD_MSMT_ACTIVE <input type="checkbox"/> attribute-type = FLOAT-Type (INT-U32) <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> l. IF Recommended attribute Accuracy is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_ACCUR_MSMT <input type="checkbox"/> attribute-type = FLOAT-Type (INT-U32) <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test>
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/PHD/CLASS/BCA/BV-005			
TP label	Body Height Object for Standard Configuration (0x07D0)			
Coverage	Spec	[ISO/IEEE 11073-10420]		
	Testable items	BodyHeight1; M	BodyHeight2; M	BodyHeight4; M
		BodyHeight6; R	BodyHeight8; M	BodyHeight10; R
		BodyHeight12; O	BodyHeight14; R	BodyHeight16; R
		BodyHeight18; R	BodyHeight20; M	BodyHeight22; M
		BodyHeight24; R	BodyHeight26; O	BodyHeight28; O
		BodyHeight30; C	BodyHeight32; C	BodyHeight34; C
		BodyHeight36; R	BodyHeight38; C	BodyHeight40; C
		BodyHeight42; C	BodyHeight44; C	BodyHeight46; C
		BodyHeight48; C	BodyHeight50; R	ConfigProc2; M
	Concepts 3; M			
Test purpose	Check that: Body Height Numeric Object contains the attributes specified for Standard Configuration (0x07D0)			

Applicability	C_AG_OXP_167 AND (NOT C_AG_OXP_181) AND C_AG_OXP_000
Other PICS	
Initial condition	The simulated PHG and the PHD under test are in the Unassociated state.
Test procedure	<ol style="list-style-type: none"> 1. The simulated PHG receives an association request from the PHD under test. 2. The simulated PHG responds with a result = accepted-unknown-config. 3. The PHD responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the PHG. 4. Check that the field Dev-Config-Id is set to 0x07D0. If it is not, the PHG responds with an "unsupported-config" and waits for a new configuration. Repeat this step until a Dev-config-Id equal to 0x07D0 is received. 5. Once the PHD under test sends a standard configuration, check the Body Height object: 6. The Body Height object contents shall be: <ol style="list-style-type: none"> a. Mandatory attribute Handle <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_HANDLE <input type="checkbox"/> attribute-type = HANDLE <input type="checkbox"/> attribute-value = 0x00 0x02 b. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = 0x00 0x02(MDC_PART_SCADA), 0xE1 0x44(MDC_LEN_BODY_ACTUAL) c. Mandatory attribute Metric-Spec-Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value ≠ 0x00 0x00 <ul style="list-style-type: none"> • Bit 0 (mss-avail-intermittent(0)) is set. • Bit 1 (mss-avail-stored-data(1)) is set. • Bit 2 (mss-upd-aperiodic(2)) is set. • Bit 3 (mss-msmt-aperiodic(3)) is set. • Bit 9 (mss-acc-agent-initiated(9)) is set. • Bit 12 (mss-cat-manual(12)) is set. d. Mandatory attribute Unit-Code <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE <input type="checkbox"/> attribute-type = OID-Type <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_DIM_CENTI_M e. Mandatory attribute Attribute-Value-Map <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ATTRIBUTE_VAL_MAP <input type="checkbox"/> attribute-type = AttrValMap <input type="checkbox"/> attribute-count = 2 <input type="checkbox"/> attribute-value = (MDC_ATTR_NU_VAL_OBS_SIMP,4 MDC_ATTR_TIME_STAMP_ABS,8) 7. Check that no other attributes are present in the initial configuration.
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id		TP/PLT/PHD/CLASS/BCA/BV-006		
TP label		Body Height Object for Extended Configuration		
Coverage	Spec	[ISO/IEEE 11073-10420]		
	Testable items	BodyHeight1; M	BodyHeight3; M	BodyHeight5; M
		BodyHeight7; R	BodyHeight9; M	BodyHeight11; R
		BodyHeight13; R	BodyHeight15; R	BodyHeight17; R
		BodyHeight19; R	BodyHeight21; M	BodyHeight23; C
		BodyHeight25; R	BodyHeight27; O	BodyHeight29; O
		BodyHeight31; C	BodyHeight33; C	BodyHeight35; C
		BodyHeight37; R	BodyHeight39; C	BodyHeight41; C
		BodyHeight43; C	BodyHeight45; C	BodyHeight47; C
		BodyHeight49; C	BodyHeight51; R	Concepts 3; M
Test purpose		<p>Check that:</p> <p>Body Height Numeric Object contains the attributes specified for Extended Configuration</p>		
Applicability		C_AG_OXP_167 AND C_AG_OXP_181 AND C_AG_OXP_000		
Other PICS				
Initial condition		The simulated PHG and the PHD under test are in the Unassociated state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated PHG receives an association request from the PHD under test. 2. The simulated PHG responds with a result = accepted-unknown-config. 3. The PHD responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the PHG. 4. Check that the field Dev-Config-Id is set to the tested extended configuration. If it is not, the PHG responds with an "unsupported-config" and waits for a new configuration. Repeat this step until a Dev-config-Id equal to the extended configuration is received. 5. Once the PHD under test sends the tested configuration, check the Body Height object: 6. The Body Height object contents shall be: <ol style="list-style-type: none"> a. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE attribute-value = 0x00 0x02(MDC_PART_SCADA), 0xE1 0x44(MDC_LEN_BODY_ACTUAL) b. IF Not Recommended attribute Supplemental-Types <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SPPLEMENTAL_TYPES <input type="checkbox"/> attribute-type = SupplementalTypeList <input type="checkbox"/> attribute-value.length = <variable>Sequence of TYPE (TYPE.length = 4 bytes) <input type="checkbox"/> attribute-value = <Not relevant for this test> c. Mandatory attribute Metric-Spec-Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value ≠ 0x00 0x00 <ul style="list-style-type: none"> • Bit 0 (mss-avail-intermittent(0)) is set. • Bit 1 (mss-avail-stored-data(1)) is set. • Bit 2 (mss-upd-aperiodic(2)) is set. 		

- Bit 3 (mss-msmt-aperiodic(3)) is set.
 - Bit 9 (mss-acc-agent-initiated(9)) is set.
 - Bit 12 (mss-cat-manual(12)) is set.
- d. IF Not recommended attribute Metric-Structure-Small is present
- attribute-id = MDC_ATTR_METRIC_STRUCTURE_SMALL
 - attribute-type = MetricStructureSmall
 - attribute-length = 2 bytes
 - attribute-value = <Not relevant for this test>
- e. IF Recommended attribute Measurement-Status is present
- attribute-id = MDC_ATTR_MSMT_STAT
 - attribute-type = MeasurementStatus
 - attribute-value.length = 2 bytes
 - attribute-value = <Not relevant for this test>
- f. IF Not recommended attribute Metric-Id is present
- attribute-id = MDC_ATTR_ID_PHYSIO
 - attribute-type = OID-Type(INT-U16)
 - attribute-value.length = 2 bytes
 - attribute-value = <Not relevant for this test>
- g. IF Not Recommended attribute Metric-Id-List is present
- attribute-id = MDC_ATTR_ID_PHYSIO_LIS
 - attribute-type = MetricIdList
 - attribute-value = <Not relevant for this test>
- h. IF Not recommended attribute Metric-Id-Partition is present
- attribute-id = MDC_ATTR_METRIC_ID_PART
 - attribute-type = NomPartition(INT-U16)
 - attribute-value.length = 2 bytes
 - attribute-value = <Not relevant for this test>
- i. Mandatory recommended attribute Unit-Code
- attribute-id = MDC_ATTR_UNIT_CODE
 - attribute-type = OID-Type(INT-U16)
 - attribute-value.length = 2 bytes
 - attribute-value = MDC_DIM_CENTI_M or MDC_DIM_INCH
- j. IF Not recommended attribute Source-Handle-Reference is present
- attribute-id = MDC_ATTR_SOURCE_HANDLE_REF
 - attribute-type = HANDLE(INT-U16)
 - attribute-value.length = 2 bytes
 - attribute-value = <Not relevant for this test>
- k. IF Not recommended attribute Measure-Active-Period
- attribute-id = MDC_ATTR_TIME_PD_MSMT_ACTIVE
 - attribute-type = FLOAT-Type (INT-U32)
 - attribute-value.length = 4 bytes
 - attribute-value = <Not relevant for this test>
- l. IF Recommended attribute Accuracy is present
- attribute-id = MDC_ATTR_NU_ACCUR_MSMT

	<input type="checkbox"/> attribute-type = FLOAT-Type (INT-U32) <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test>
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/PHD/CLASS/BCA/BV-007			
TP label	Body Fat Object for Standard Configuration (0x07D0)			
Coverage	Spec	[ISO/IEEE 11073-10420]		
	Testable items	BodyFat1; M	BodyFat2; O	BodyFat3; M
		BodyFat5; M	BodyFat7; R	BodyFat9; M
		BodyFat11; R	BodyFat13; O	BodyFat15; R
		BodyFat17; R	BodyFat19; R	BodyFat21; M
		BodyFat23; M	BodyFat25; R	BodyFat27; O
		BodyFat29; O	BodyFat31; C	BodyFat33; C
		BodyFat35; C	BodyFat37; R	BodyFat39; C
		BodyFat41; C	BodyFat43; C	BodyFat45; C
		BodyFat47; C	BodyFat49; C	BodyFat51; R
BodyFat53; M		ConfigProc2; M	Concepts 2; M	
Test purpose	Check that: Body Fat Numeric Object contains the attributes specified for Standard Configuration (0x07D0)			
Applicability	C_AG_OXP_167 AND (NOT C_AG_OXP_181) AND C_AG_OXP_000			
Other PICS				
Initial condition	The simulated PHG and the PHD under test are in the Unassociated state.			
Test procedure	<ol style="list-style-type: none"> 1. The simulated PHG receives an association request from the PHD under test. 2. The simulated PHG responds with a result = accepted-unknown-config. 3. The PHD responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the PHG. 4. Check that the field Dev-Config-Id is set to 0x07D0. If it is not, the PHG responds with an "unsupported-config" and waits for a new configuration. Repeat this step until a Dev-config-Id equal to 0x07D0 is received. 5. Once the PHD under test sends a standard configuration, check the Body Fat object. 6. The Body Fat contents shall be: <ol style="list-style-type: none"> a. Mandatory attribute Handle <input type="checkbox"/> attribute-id = MDC_ATTR_ID_HANDLE <input type="checkbox"/> attribute-type = HANDLE <input type="checkbox"/> attribute-value = 0x00 0x03 b. Mandatory attribute Type <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = MDC_PART_SCADA MDC_BODY_FAT c. Mandatory attribute Metric-Spec-Small <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall 			

	<ul style="list-style-type: none"> <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value ≠ 0x00 0x00 <ul style="list-style-type: none"> • Bit 0 (mss-avail-intermittent(0)) is set. • Bit 1 (mss-avail-stored-data(1)) is set. • Bit 2 (mss-upd-aperiodic(2)) is set. • Bit 3 (mss-msmt-aperiodic(3)) is set. • Bit 9 (mss-acc-agent-initiated(9)) is set. • Bit 14 (mss-cat-calculation(14)) is set. <p>d. Mandatory recommended attribute Unit-Code</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE <input type="checkbox"/> attribute-type = OID-Type(INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_DIM_PERCENT. <p>e. Mandatory attribute Attribute-Value-Map</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ATTRIBUTE_VAL_MAP <input type="checkbox"/> attribute-type = AttrValMap <input type="checkbox"/> attribute-count = 2 <input type="checkbox"/> attribute-value = (MDC_ATTR_NU_VAL_OBS_SIMP,4 MDC_ATTR_TIME_STAMP_ABS,8) <p>f. IF Recommended attribute Accuracy is present</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_ACCUR_MSMT <input type="checkbox"/> attribute-type = FLOAT-Type (INT-U32) <input type="checkbox"/> attribute-value.length = FLOAT-Type (INT-U32) <p>7. Check that no other attributes are present in the initial configuration.</p>
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/PHD/CLASS/BCA/BV-008			
TP label	Body Fat Object for Extended Configuration			
Coverage	Spec	[ISO/IEEE 11073-10420]		
	Testable items	BodyFat2; O	BodyFat4; M	BodyFat6; M
		BodyFat8; R	BodyFat10; M	BodyFat12; R
		BodyFat14; R	BodyFat16; C	BodyFat18; R
		BodyFat20; C	BodyFat22; M	BodyFat24; C
		BodyFat26; R	BodyFat28; O	BodyFat30; O
		BodyFat32; C	BodyFat34; C	BodyFat36; C
		BodyFat38; R	BodyFat40; C	BodyFat42; C
		BodyFat44; C	BodyFat46; C	BodyFat48; C
		BodyFat50; C	BodyFat52; R	Concepts 2; M
Test purpose	Check that: Body Fat Numeric Object contains the attributes specified for Extended Configuration			
Applicability	C_AG_OXP_167 AND C_AG_OXP_181 AND C_AG_OXP_000			
Other PICS				
Initial condition	The simulated PHG and the PHD under test are in the Unassociated state.			

<p>Test procedure</p>	<ol style="list-style-type: none"> 1. The simulated PHG receives an association request from the PHD under test. 2. The simulated PHG responds with a result = accepted-unknown-config. 3. The PHD responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the PHG. 4. Check that the field Dev-Config-Id is set to the tested extended configuration. If it is not, the PHG responds with an "unsupported-config" and waits for a new configuration. Repeat this step until a Dev-config-Id equal to the extended configuration is received. 5. Once the PHD under test sends the tested configuration, check the Body Fat object. 6. The Body Fat object contents shall be: <ol style="list-style-type: none"> a. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = MDC_PART_SCADA MDC_BODY_FAT b. IF Not Recommended attribute Supplemental-Types <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SPPLEMENTAL_TYPES <input type="checkbox"/> attribute-type = SupplementalTypeList <input type="checkbox"/> attribute-value.length = <variable>Sequence of TYPE (TYPE.length = 4 bytes) <input type="checkbox"/> attribute-value = <Not relevant for this test> c. Mandatory attribute Metric-Spec-Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> d. IF Not recommended attribute Metric-Structure-Small is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_STRUCTURE_SMALL <input type="checkbox"/> attribute-type = MetricStructureSmall <input type="checkbox"/> attribute-length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> e. IF Recommended attribute Measurement-Status is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_MSMT_STAT <input type="checkbox"/> attribute-type = MeasurementStatus <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> f. IF Conditional attribute Metric-Id is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_PHYSIO <input type="checkbox"/> attribute-type = OID-Type(INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> g. IF Not Recommended attribute Metric-Id-List is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_PHYSIO_LIS <input type="checkbox"/> attribute-type = MetricIdList <input type="checkbox"/> attribute-value = <Not relevant for this test> h. IF Conditional attribute Metric-Id-Partition is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_ID_PART <input type="checkbox"/> attribute-type = NomPartition(INT-U16)
------------------------------	---

	<ul style="list-style-type: none"> <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> i. Mandatory attribute Unit-Code <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE <input type="checkbox"/> attribute-type = OID-Type <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_DIM_PERCENT or MDC_DIM_KILO_G or MDC_DIM_LB j. IF Not recommended attribute Source-Handle-Reference is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SOURCE_HANDLE_REF <input type="checkbox"/> attribute-type = HANDLE(INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> k. IF Not recommended attribute Measure-Active-Period <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_PD_MSMT_ACTIVE <input type="checkbox"/> attribute-type = FLOAT-Type (INT-U32) <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> l. IF Not recommended attribute Compound-Simple-Nu-Observed-Value is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_SIMP <input type="checkbox"/> attribute-type = SimpleNuObsValueCmp <input type="checkbox"/> attribute-value.length = <variable> <input type="checkbox"/> attribute-value = <Not relevant for this test> m. IF Not recommended attribute Basic-Nu-Observed-Value is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_VAL_OBS_BASIC <input type="checkbox"/> attribute-type = BasicNuObsValue <input type="checkbox"/> attribute-value.length = 2bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> n. IF Not recommended attribute Compound-Basic-Nu-Observed-Value is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_BASIC <input type="checkbox"/> attribute-type = BasicNuObsValueCmp <input type="checkbox"/> attribute-value.length = <variable> <input type="checkbox"/> attribute-value = <Not relevant for this test> o. IF Not recommended attribute Nu-Observed-Value is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_VAL_OBS <input type="checkbox"/> attribute-type = NuObsValue <input type="checkbox"/> attribute-value.length = 10bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> p. Not recommended attribute Compound-Nu-Observed-Value <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_SIMP <input type="checkbox"/> attribute-type = NuObsValueCmp <input type="checkbox"/> attribute-value.length = <variable> <input type="checkbox"/> attribute-value = <Not relevant for this test> q. IF Recommended attribute Accuracy is present
--	---

	<input type="checkbox"/> attribute-id = MDC_ATTR_NU_ACCUR_MSMT <input type="checkbox"/> attribute-type = FLOAT-Type (INT-U32) <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test>
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/PHD/CLASS/BCA/BV-009			
TP label	Body Mass Index Object for Extended Configuration			
Coverage	Spec	[ISO/IEEE 11073-10420]		
	Testable items	BodyMassIndex1; O	BodyMassIndex2; M	BodyMassIndex3; M
		BodyMassIndex4; M	BodyMassIndex5; R	BodyMassIndex6; M
		BodyMassIndex7; R	BodyMassIndex8; R	BodyMassIndex9; R
		BodyMassIndex10; R	BodyMassIndex11; R	BodyMassIndex12; M
		BodyMassIndex13; C	BodyMassIndex14; M	BodyMassIndex15; O
		BodyMassIndex16; O	BodyMassIndex17; C	BodyMassIndex18; C
		BodyMassIndex19; C	BodyMassIndex20; R	BodyMassIndex21; C
		BodyMassIndex22; C	BodyMassIndex23; C	BodyMassIndex24; C
		BodyMassIndex25; C	BodyMassIndex26; C	BodyMassIndex27; R
	Concepts 5; O			
Test purpose	Check that: Body Mass Index Numeric Object contains the attributes specified for Extended Configuration			
Applicability	C_AG_OXP_167 AND C_AG_OXP_181 AND C_AG_BCA_001 AND C_AG_OXP_000			
Other PICS				
Initial condition	The simulated PHG and the PHD under test are in the Unassociated state.			
Test procedure	<ol style="list-style-type: none"> 1. The simulated PHG receives an association request from the PHD under test. 2. The simulated PHG responds with a result = accepted-unknown-config. 3. The PHD responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the PHG. 4. Check that the field Dev-Config-Id is set to the tested extended configuration. If it is not, the PHG responds with an "unsupported-config" and waits for a new configuration. Repeat this step until a Dev-config-Id equal to the extended configuration is received. 5. Once the PHD under test sends the tested configuration, check the Body Mass Index object. 6. The Body Mass Index object contents shall be: <ol style="list-style-type: none"> a. Mandatory attribute Type <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = MDC_PART_SCADA MDC_RATIO_MASS_BODY_LEN_SQ b. IF Not Recommended attribute Supplemental-Types <input type="checkbox"/> attribute-id = MDC_ATTR_SPPLEMENTAL_TYPES <input type="checkbox"/> attribute-type = SupplementalTypeList <input type="checkbox"/> attribute-value.length = <variable>Sequence of TYPE (TYPE.length = 4 bytes) <input type="checkbox"/> attribute-value = <Not relevant for this test> 			

	<p>c. Mandatory attribute Metric-Spec-Small</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value ≠ 0x00 0x00 <ul style="list-style-type: none"> ▪ Bit 0 (mss-avail-intermittent(0)) must be set. ▪ Bit 1 (mss-avail-stored-data(1)) must be set. ▪ Bit 2 (mss-upd-aperiodic(2)) must be set. ▪ Bit 3 (mss-msmt-aperiodic(3)) must be set. ▪ Bit 9 (mss-acc-agent-initiated(9)) must be set. ▪ Bit 14 (mss-cat_calculation(14)) must be set is set <p>d. IF Not recommended attribute Metric-Structure-Small is present</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_STRUCTURE_SMALL <input type="checkbox"/> attribute-type = MetricStructureSmall <input type="checkbox"/> attribute-length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> <p>e. IF Recommended attribute Measurement-Status is present</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_MSMT_STAT <input type="checkbox"/> attribute-type = MeasurementStatus <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> <p>f. IF Not recommended attribute Metric-Id is present</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_PHYSIO <input type="checkbox"/> attribute-type = OID-Type(INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> <p>g. IF Not Recommended attribute Metric-Id-List is present</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_PHYSIO_LIS <input type="checkbox"/> attribute-type = MetricIdList <input type="checkbox"/> attribute-value = <Not relevant for this test> <p>h. IF Not recommended attribute Metric-Id-Partition is present</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_ID_PART <input type="checkbox"/> attribute-type = NomPartition(INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> <p>i. Mandatory attribute Unit-Code</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE <input type="checkbox"/> attribute-type = OID-Type <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_DIM_KG_PER_M_SQ <p>j. Conditional attribute Attribute-Value-Map</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ATTRIBUTE_VAL_MAP <input type="checkbox"/> attribute-type = AttrValMap <input type="checkbox"/> attribute-count = 2
--	---

	<ul style="list-style-type: none"> <input type="checkbox"/> attribute-value = <Not relevant for this test> k. Mandatory attribute Source-Handle-Reference is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SOURCE_HANDLE_REF <input type="checkbox"/> attribute-type = HANDLE(INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = It must be equal to the handle of another metric object in the configuration and it must point to an object that has a type of MDC_MASS_BODY_ACTUAL. l. IF Not recommended attribute Measure-Active-Period <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_PD_MSMT_ACTIVE <input type="checkbox"/> attribute-type = FLOAT-Type (INT-U32) <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> m. IF Not recommended attribute Compound-Simple-Nu-Observed-Value is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_SIMP <input type="checkbox"/> attribute-type = SimpleNuObsValueCmp <input type="checkbox"/> attribute-value.length = <variable> <input type="checkbox"/> attribute-value = <Not relevant for this test> n. IF Not recommended attribute Basic-Nu-Observed-Value is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_VAL_OBS_BASIC <input type="checkbox"/> attribute-type = BasicNuObsValue <input type="checkbox"/> attribute-value.length = 2bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> o. IF Not recommended attribute Compound-Basic-Nu-Observed-Value is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_BASIC <input type="checkbox"/> attribute-type = BasicNuObsValueCmp <input type="checkbox"/> attribute-value.length = <variable> <input type="checkbox"/> attribute-value = <Not relevant for this test> p. IF Not recommended attribute Nu-Observed-Value is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_VAL_OBS <input type="checkbox"/> attribute-type = NuObsValue <input type="checkbox"/> attribute-value.length = 10bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> q. Not recommended attribute Compound-Nu-Observed-Value <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_SIMP <input type="checkbox"/> attribute-type = NuObsValueCmp <input type="checkbox"/> attribute-value.length = <variable> <input type="checkbox"/> attribute-value = <Not relevant for this test> r. IF Recommended attribute Accuracy is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_ACCUR_MSMT <input type="checkbox"/> attribute-type = FLOAT-Type (INT-U32) <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test>
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/PHD/CLASS/BCA/BV-010			
TP label	Fat Free Mass Object for Extended Configuration			
Coverage	Spec	[ISO/IEEE 11073-10420]		
	Testable items	FatFreeMass1; O	FatFreeMass2; M	FatFreeMass3; M
		FatFreeMass4; M	FatFreeMass5; R	FatFreeMass6; M
		FatFreeMass7; R	FatFreeMass8; R	FatFreeMass9; R
		FatFreeMass10; R	FatFreeMass11; R	FatFreeMass12; M
		FatFreeMass13; C	FatFreeMass14; R	FatFreeMass15; O
		FatFreeMass16; O	FatFreeMass17; C	FatFreeMass18; C
		FatFreeMass19; C	FatFreeMass20; R	FatFreeMass21; C
		FatFreeMass22; C	FatFreeMass23; C	FatFreeMass24; C
		FatFreeMass25; C	FatFreeMass26; C	FatFreeMass27; R
	Concepts 6; O			
Test purpose	Check that: Fat Free Mass Numeric Object contains the attributes specified for Extended Configuration			
Applicability	C_AG_OXP_167 AND C_AG_OXP_181 AND C_AG_BCA_004 AND C_AG_OXP_000			
Other PICS				
Initial condition	The simulated PHG and the PHD under test are in the Unassociated state.			
Test procedure	<ol style="list-style-type: none"> 1. The simulated PHG receives an association request from the PHD under test. 2. The simulated PHG responds with a result = accepted-unknown-config. 3. The PHD responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the PHG. 4. Check that the field Dev-Config-Id is set to the tested extended configuration. If it is not, the PHG responds with an "unsupported-config" and waits for a new configuration. Repeat this step until a Dev-config-Id equal to the extended configuration is received. 5. Once the PHD under test sends the tested configuration, check the Fat Free Mass object. 6. The Fat Free Mass object contents shall be: <ol style="list-style-type: none"> a. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = MDC_PART_SCADA MDC_MASS_BODY_FAT_FREE b. IF Not Recommended attribute Supplemental-Types <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SPPLEMENTAL_TYPES <input type="checkbox"/> attribute-type = SupplementalTypeList <input type="checkbox"/> attribute-value.length = <variable>Sequence of TYPE (TYPE.length = 4 bytes) <input type="checkbox"/> attribute-value = <Not relevant for this test> c. Mandatory attribute Metric-Spec-Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> d. IF Not recommended attribute Metric-Structure-Small is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_STRUCTURE_SMALL 			

	<ul style="list-style-type: none"> <input type="checkbox"/> attribute-type = MetricStructureSmall <input type="checkbox"/> attribute-length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> e. IF Recommended attribute Measurement-Status is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_MSMT_STAT <input type="checkbox"/> attribute-type = MeasurementStatus <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> f. IF Not recommended attribute Metric-Id is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_PHYSIO <input type="checkbox"/> attribute-type = OID-Type(INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> g. IF Not Recommended attribute Metric-Id-List is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_PHYSIO_LIS <input type="checkbox"/> attribute-type = MetricIdList <input type="checkbox"/> attribute-value = <Not relevant for this test> h. IF Not recommended attribute Metric-Id-Partition is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_ID_PART <input type="checkbox"/> attribute-type = NomPartition(INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> i. Mandatory attribute Unit-Code <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE <input type="checkbox"/> attribute-type = OID-Type <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_DIM_KILO_G or MDC_DIM_LB j. Conditional attribute Attribute-Value-Map <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ATTRIBUTE_VAL_MAP <input type="checkbox"/> attribute-type = AttrValMap <input type="checkbox"/> attribute-count = 2 <input type="checkbox"/> attribute-value = <Not relevant for this test> k. IF Not recommended attribute Source-Handle-Reference is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SOURCE_HANDLE_REF <input type="checkbox"/> attribute-type = HANDLE(INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> l. IF Not recommended attribute Measure-Active-Period <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_PD_MSMT_ACTIVE <input type="checkbox"/> attribute-type = FLOAT-Type (INT-U32) <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> m. IF Not recommended Compound-Simple-Nu-Observed-Value is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_SIMP
--	---

	<ul style="list-style-type: none"> <input type="checkbox"/> attribute-type = SimpleNuObsValueCmp <input type="checkbox"/> attribute-value.length = <variable> <input type="checkbox"/> attribute-value = <Not relevant for this test> n. IF Not recommended attribute Basic-Nu-Observed-Value is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_VAL_OBS_BASIC <input type="checkbox"/> attribute-type = BasicNuObsValue <input type="checkbox"/> attribute-value.length = 2bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> o. IF Not recommended attribute Compound-Basic-Nu-Observed-Value is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_BASIC <input type="checkbox"/> attribute-type = BasicNuObsValueCmp <input type="checkbox"/> attribute-value.length = <variable> <input type="checkbox"/> attribute-value = <Not relevant for this test> p. IF Not recommended attribute Nu-Observed-Value is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_VAL_OBS <input type="checkbox"/> attribute-type = NuObsValue <input type="checkbox"/> attribute-value.length = 10bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> q. Not recommended attribute Compound-Nu-Observed-Value <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_SIMP <input type="checkbox"/> attribute-type = NuObsValueCmp <input type="checkbox"/> attribute-value.length = <variable> <input type="checkbox"/> attribute-value = <Not relevant for this test> r. IF Recommended attribute Accuracy is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_ACCUR_MSMT <input type="checkbox"/> attribute-type = FLOAT-Type (INT-U32) <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test>
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/PHD/CLASS/BCA/BV-011			
TP label	Soft Lean Mass Object for Extended Configuration			
Coverage	Spec	[ISO/IEEE 11073-10420]		
	Testable items	SoftLeanMass1; O	SoftLeanMass2; M	SoftLeanMass3; M
		SoftLeanMass4; M	SoftLeanMass5; R	SoftLeanMass6; M
		SoftLeanMass7; R	SoftLeanMass8; R	SoftLeanMass9; R
		SoftLeanMass10; R	SoftLeanMass11; R	SoftLeanMass12; M
		SoftLeanMass13; C	SoftLeanMass14; R	SoftLeanMass15; O
		SoftLeanMass16; O	SoftLeanMass17; C	SoftLeanMass18; C
		SoftLeanMass19; C	SoftLeanMass20; R	SoftLeanMass21; C
		SoftLeanMass22; C	SoftLeanMass23; C	SoftLeanMass24; C
		SoftLeanMass25; C	SoftLeanMass26; C	SoftLeanMass27; R
	Concepts 7; O			

Test purpose	Check that: Soft Lean Mass Numeric Object contains the attributes specified for Extended Configuration
Applicability	C_AG_OXP_167 AND C_AG_OXP_181 AND C_AG_BCA_003 AND C_AG_OXP_000
Other PICS	
Initial condition	The simulated PHG and the PHD under test are in the Unassociated state.
Test procedure	<ol style="list-style-type: none"> 1. The simulated PHG receives an association request from the PHD under test. 2. The simulated PHG responds with a result = accepted-unknown-config. 3. The PHD responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the PHG. 4. Check that the field Dev-Config-Id is set to the tested extended configuration. If it is not, the PHG responds with an "unsupported-config" and waits for a new configuration. Repeat this step until a Dev-config-Id equal to the extended configuration is received. 5. Once the PHD under test sends the tested configuration, check the Soft Lean Mass object. 6. The Soft Lean Mass object contents shall be: <ol style="list-style-type: none"> a. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = MDC_PART_SCADA MDC_MASS_BODY_SOFT_LEAN b. IF Not Recommended attribute Supplemental-Types <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SPPLEMENTAL_TYPES <input type="checkbox"/> attribute-type = SupplementalTypeList <input type="checkbox"/> attribute-value.length = <variable>Sequence of TYPE (TYPE.length = 4 bytes) <input type="checkbox"/> attribute-value = <Not relevant for this test> c. Mandatory attribute Metric-Spec-Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> d. IF Not recommended attribute Metric-Structure-Small is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_STRUCTURE_SMALL <input type="checkbox"/> attribute-type = MetricStructureSmall <input type="checkbox"/> attribute-length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> e. IF Recommended attribute Measurement-Status is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_MSMT_STAT <input type="checkbox"/> attribute-type = MeasurementStatus <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> f. IF Not recommended attribute Metric-Id is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_PHYSIO <input type="checkbox"/> attribute-type = OID-Type(INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> g. IF Not Recommended attribute Metric-Id-List is present

	<ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_PHYSIO_LIS <input type="checkbox"/> attribute-type = MetricIdList <input type="checkbox"/> attribute-value = <Not relevant for this test> <p>h. IF Not recommended attribute Metric-Id-Partition is present</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_ID_PART <input type="checkbox"/> attribute-type = NomPartition(INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> <p>i. Mandatory attribute Unit-Code</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE <input type="checkbox"/> attribute-type = OID-Type <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_DIM_KILO_G or MDC_DIM_LB <p>j. Conditional attribute Attribute-Value-Map</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ATTRIBUTE_VAL_MAP <input type="checkbox"/> attribute-type = AttrValMap <input type="checkbox"/> attribute-count = 2 <input type="checkbox"/> attribute-value = <Not relevant for this test> attribute-value = <Not relevant for this test> <p>k. IF Not recommended attribute Source-Handle-Reference is present</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SOURCE_HANDLE_REF <input type="checkbox"/> attribute-type = HANDLE(INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> <p>l. IF Not recommended attribute Measure-Active-Period</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_PD_MSMT_ACTIVE <input type="checkbox"/> attribute-type = FLOAT-Type (INT-U32) <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> <p>m. IF Not recommended attribute Compound-Simple-Nu-Observed-Value is present</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_SIMP <input type="checkbox"/> attribute-type = SimpleNuObsValueCmp <input type="checkbox"/> attribute-value.length = <variable> <input type="checkbox"/> attribute-value = <Not relevant for this test> <p>n. IF Not recommended attribute Basic-Nu-Observed-Value is present</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_VAL_OBS_BASIC <input type="checkbox"/> attribute-type = BasicNuObsValue <input type="checkbox"/> attribute-value.length = 2bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> <p>o. IF Not recommended attribute Compound-Basic-Nu-Observed-Value is present</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_BASIC <input type="checkbox"/> attribute-type = BasicNuObsValueCmp <input type="checkbox"/> attribute-value.length = <variable> <input type="checkbox"/> attribute-value = <Not relevant for this test>
--	---

	<p>p. IF Not recommended attribute Nu-Observed-Value is present</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_VAL_OBS <input type="checkbox"/> attribute-type = NuObsValue <input type="checkbox"/> attribute-value.length = 10bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> <p>q. Not recommended attribute Compound-Nu-Observed-Value</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_SIMP <input type="checkbox"/> attribute-type = NuObsValueCmp <input type="checkbox"/> attribute-value.length = <variable> <input type="checkbox"/> attribute-value = <Not relevant for this test> <p>r. IF Recommended attribute Accuracy is present</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_ACCUR_MSMT <input type="checkbox"/> attribute-type = FLOAT-Type (INT-U32) <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test>
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/PHD/CLASS/BCA/BV-012			
TP label	Body Water Object for Extended Configuration			
Coverage	Spec	[ISO/IEEE 11073-10420]		
	Testable items	BodyWater1; O	BodyWater2; M	BodyWater3; M
		BodyWater4; M	BodyWater5; R	BodyWater6; M
		BodyWater7; R	BodyWater8; R	BodyWater9; R
		BodyWater10; R	BodyWater11; R	BodyWater12; M
		BodyWater13; C	BodyWater14; R	BodyWater15; O
		BodyWater16; O	BodyWater17; C	BodyWater18; C
		BodyWater19; C	BodyWater20; R	BodyWater21; C
		BodyWater22; C	BodyWater23; C	BodyWater24; C
		BodyWater25; C	BodyWater26; C	BodyWater27; R
BodyWater28; O		Concepts 8; O		
Test purpose	Check that: Body Water Numeric Object contains the attributes specified for Extended Configuration			
Applicability	C_AG_OXP_167 AND C_AG_OXP_181 AND C_AG_BCA_002 AND C_AG_OXP_000			
Other PICS				
Initial condition	The simulated PHG and the PHD under test are in the Unassociated state.			
Test procedure	<ol style="list-style-type: none"> 1. The simulated PHG receives an association request from the PHD under test. 2. The simulated PHG responds with a result = accepted-unknown-config. 3. The PHD responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the PHG. 4. Check that the field Dev-Config-Id is set to the tested extended configuration. If it is not, the PHG responds with an "unsupported-config" and waits for a new configuration. Repeat this step until a Dev-config-Id equal to the extended configuration is received. 5. Once the PHD under test sends the tested configuration, check the Body Water object. 6. The Body Water object contents shall be: 			

	<ul style="list-style-type: none"> a. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = MDC_PART_SCADA MDC_BODY_WATER b. IF Not Recommended attribute Supplemental-Types <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SPPLEMENTAL_TYPES <input type="checkbox"/> attribute-type = SupplementalTypeList <input type="checkbox"/> attribute-value.length = <variable>Sequence of TYPE (TYPE.length = 4 bytes) <input type="checkbox"/> attribute-value = <Not relevant for this test> c. Mandatory attribute Metric-Spec-Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> d. IF Not recommended attribute Metric-Structure-Small is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_STRUCTURE_SMALL <input type="checkbox"/> attribute-type = MetricStructureSmall <input type="checkbox"/> attribute-length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> e. IF Recommended attribute Measurement-Status is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_MSMT_STAT <input type="checkbox"/> attribute-type = MeasurementStatus <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> f. IF Not recommended attribute Metric-Id is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_PHYSIO <input type="checkbox"/> attribute-type = OID-Type(INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> g. IF Not Recommended attribute Metric-Id-List is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_PHYSIO_LIS <input type="checkbox"/> attribute-type = MetricIdList <input type="checkbox"/> attribute-value = <Not relevant for this test> h. IF Not recommended attribute Metric-Id-Partition is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_ID_PART <input type="checkbox"/> attribute-type = NomPartition(INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> i. Mandatory attribute Unit-Code <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE <input type="checkbox"/> attribute-type = OID-Type <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_DIM_KILO_G or MDC_DIM_LB or MDC_DIM_PERCENT.
--	--

	<ul style="list-style-type: none"> <input type="checkbox"/> The PHD is allowed to report two body water objects, one in kilograms (kg) and the other in percent (%). j. Conditional attribute Attribute-Value-Map <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ATTRIBUTE_VAL_MAP <input type="checkbox"/> attribute-type = AttrValMap <input type="checkbox"/> attribute-count = 2 <input type="checkbox"/> attribute-value = <Not relevant for this test> attribute-value = <Not relevant for this test> k. IF Not recommended attribute Source-Handle-Reference is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SOURCE_HANDLE_REF <input type="checkbox"/> attribute-type = HANDLE(INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> l. IF Not recommended attribute Measure-Active-Period <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_PD_MSMT_ACTIVE <input type="checkbox"/> attribute-type = FLOAT-Type (INT-U32) <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> m. IF Not recommended attribute Compound-Simple-Nu-Observed-Value is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_SIMP <input type="checkbox"/> attribute-type = SimpleNuObsValueCmp <input type="checkbox"/> attribute-value.length = <variable> <input type="checkbox"/> attribute-value = <Not relevant for this test> n. IF Not recommended attribute Basic-Nu-Observed-Value is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_VAL_OBS_BASIC <input type="checkbox"/> attribute-type = BasicNuObsValue <input type="checkbox"/> attribute-value.length = 2bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> o. IF Not recommended attribute Compound-Basic-Nu-Observed-Value is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_BASIC <input type="checkbox"/> attribute-type = BasicNuObsValueCmp <input type="checkbox"/> attribute-value.length = <variable> <input type="checkbox"/> attribute-value = <Not relevant for this test> p. IF Not recommended attribute Nu-Observed-Value is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_VAL_OBS <input type="checkbox"/> attribute-type = NuObsValue <input type="checkbox"/> attribute-value.length = 10bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> q. Not recommended attribute Compound-Nu-Observed-Value <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_SIMP <input type="checkbox"/> attribute-type = NuObsValueCmp <input type="checkbox"/> attribute-value.length = <variable> <input type="checkbox"/> attribute-value = <Not relevant for this test> r. IF Recommended attribute Accuracy is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_ACCUR_MSMT
--	---

	<input type="checkbox"/> attribute-type = FLOAT-Type (INT-U32) <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test>
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/PHD/CLASS/BCA/BV-013		
TP label	Operating State. PHG to PHD Maximum APDU Size		
Coverage	Spec	[ISO/IEEE 11073-20601-2015A] and [ISO/IEEE 11073-20601-2016C]	
	Testable items	CommonCharac 3; M	
	Spec	[ISO/IEEE 11073-10420]	
	Testable items	CommChar1;M	CommChar2;M
Test purpose	<p>Check that:</p> <p>The total size of the response does not exceed the maximum APDU size established by the specialization</p> <p>[AND]</p> <p>A body composition analyser PHD implementing only this device specialization shall be capable of receiving any APDU up to a size of Nr_x.</p> <p>For this standard, Nr_x shall be 1230 octets.</p>		
Applicability	C_AG_OXP_000 AND C_AG_OXP_167		
Other PICS	C_AG_OXP_041, C_AG_OXP_100		
Initial condition	The simulated PHG and the PHD under test are in the Operating state		
Test procedure	<ol style="list-style-type: none"> 1. The simulated PHG issues a "Remote Operation Invoke Get" command with: <ol style="list-style-type: none"> a. Obj-handle set to 0 (to request for an MDS object) b. attribute-id-list.count = 606 c. attribute-id-list: (MDC_ATTR_ID_MODEL, MDC_ATTR_SYS_ID, MDC_ATTR_DEV_CONFIG_ID) repeated 202 times 2. Check the response of the PHD. 3. The simulated PHG issues a "Remote Operation Invoke Get" command with the handle set to 0 (to request for an MDS object) and an empty attribute-id-list to indicate all attributes. 4. Check the response of the PHD. 		
Pass/Fail criteria	<ul style="list-style-type: none"> • In step 2, the PHD under test may respond with a rors-cmip-get listing all the requested attributes, or with a roer message. If PICS C_AG_OXP_100=TRUE and the PHD does not respond with a rors-cmip-get message, it responds with a roer message or rorj (resource-limitation) message, a WARNING will appear. <ul style="list-style-type: none"> ○ If the response is a get response, the total size of the response cannot exceed the sum of the APDU sizes of the supported specializations (limited to an absolute limit of 64512 octets): <ul style="list-style-type: none"> ▪ Pulse oximeter → 9216 octets ▪ Weighing scales → 896 octets ▪ Glucose meter → 5120 octets or 64512 octets if the PHD supports PM-Store ▪ Blood pressure → 896 octets ▪ Thermometer → 896 octets ▪ Independent activity hub → 5120 octets ▪ Cardiovascular → 64512 octets or 6624 octets if the PHD under test only 		

	<p>supports Step Counter Profile</p> <ul style="list-style-type: none"> ▪ Strength → 64512 octets: ▪ Adherence monitor → 1024 octets ▪ Peak flow → 2030 octets ▪ Body composition Analyser → 7730 octets ▪ Body composition Analyser → 7730 octets ▪ Basic ECG/Simple ECG → 7168 octets or 64512 octets if the PHD supports PM-Store ▪ Basic ECG/Heart rate → 1280 octets or 64512 octets if the PHD supports PM-Store ▪ International normalized ratio → 896 octets or 64512 if the PHD supports PM-Store <ul style="list-style-type: none"> ○ In the case where it responds with a roer, the reason must not be protocol-violation (23). <ul style="list-style-type: none"> • In step 4, the PHD must respond with a rors-cmip-get message.
Notes	

TP Id	TP/PLT/PHD/CLASS/BCA/BV-014			
TP label	Association Body composition analyser PHD			
Coverage	Spec	[ISO/IEEE 11073-10420]		
	Testable items	AgProcAsReq1; M	AgProcAsReq2; M	AgProcAsReq3; M
		AgProcAsReq4; M	AgProcAsReq5; M	AgProcAsReq6; M
		AgProcAsReq7; M	AgProcAsReq8; M	AgProcAsReq9; M
		AgProcAsReq10; M	AgProcAsReq11; M	AgProcAsReq12; M
	MDSMethods 4; M			
Test purpose	<p>Check that:</p> <p>During the association procedure, the Body composition analyser PHD sends the correct association request to the simulated PHG</p>			
Applicability	C_AG_OXP_167 AND C_AG_OXP_000			
Other PICS	C_AG_OXP_002, C_AG_OXP_017			
Initial condition	The simulated PHG and the PHD under test are in the Unassociated state.			
Test procedure	<p>1. The PHD sends a message to associate to the simulated PHG, the expected fields sent by the PHD are:</p> <ul style="list-style-type: none"> a. APDU Type <ul style="list-style-type: none"> <input type="checkbox"/> field- type = AarqApdu <input type="checkbox"/> field-length = 2 bytes <input type="checkbox"/> field-value = 0xE2 0x00. b. assoc-version <ul style="list-style-type: none"> <input type="checkbox"/> field- type = AssociationVersion <input type="checkbox"/> field-length = BITS-32 <input type="checkbox"/> field- value = 0x80 0x00 0x00 0x00 c. data-protoid <ul style="list-style-type: none"> <input type="checkbox"/> field- type = DataProtoid(INT-U16) <input type="checkbox"/> field-length = 2 bytes <input type="checkbox"/> field- value = 0x50 0x79 (20601) d. protocol-version 			

	<ul style="list-style-type: none"> <input type="checkbox"/> field- type = Protocol Version <input type="checkbox"/> field-length = 4 bytes <input type="checkbox"/> field- value = 0x80 0x00 0x00 0x00 <p>e. encoding rules</p> <ul style="list-style-type: none"> <input type="checkbox"/> field- type = EncodingRules <input type="checkbox"/> field-length = 2 bytes <input type="checkbox"/> field- value = <ul style="list-style-type: none"> ▪ Bit 0 must be set (support MDER) ▪ Bits 1 and 2 may be set ▪ The rest of the bits must be 0 <p>f. nomenclature version</p> <ul style="list-style-type: none"> <input type="checkbox"/> field- type = NomenclatureVersion <input type="checkbox"/> field-length = 4 bytes <input type="checkbox"/> field- value = 0x80 0x00 0x00 0x00 <input type="checkbox"/> This value indicates version1 is supported (nom-version1(0) is set). <p>g. functional-units</p> <ul style="list-style-type: none"> <input type="checkbox"/> field- type = FunctionalUnits <input type="checkbox"/> field-length = 4 bytes <input type="checkbox"/> field-value = <ul style="list-style-type: none"> ▪ Bit 0 must not be set, only bit 1 or 2 may be set to 1. <p>h. System type</p> <ul style="list-style-type: none"> <input type="checkbox"/> field- type = SystemType <input type="checkbox"/> field-length = 4 bytes <input type="checkbox"/> field- value = 0x00 0x80 0x00 0x00 (sys-type-agent) <p>i. System-Id</p> <ul style="list-style-type: none"> <input type="checkbox"/> field- type = OCTET STRING <input type="checkbox"/> field-length = 8 bytes <input type="checkbox"/> field- value = 0xXX 0xXX 0xXX 0xXX 0xXX 0xXX 0xXX 0xXX (octet string length = 8 UI-64 manufacturer and device) <input type="checkbox"/> This value will be the System Id attribute of the MDS Object and the received value will be compared with the value defined in PIXIT I_AG_OXP_001 and I_AG_OXP_002. <p>j. dev-config-id</p> <ul style="list-style-type: none"> <input type="checkbox"/> field- type = ConfigId(INT-U16) <input type="checkbox"/> field-length = 2 bytes <input type="checkbox"/> field- value = <ul style="list-style-type: none"> ▪ <0x07D0> for standard configuration ▪ <between 0x40 0x00 and 0x7F 0xFF > for extended configuration. <p>k. data-req-mode-flags (DataReqModeCapab)</p> <ul style="list-style-type: none"> <input type="checkbox"/> field- type = DataReqModeFlags <input type="checkbox"/> field-length = 2 bytes <input type="checkbox"/> If the PHD supports only Body composition analyser specialization → Bit 15 is set (data-req-supp-init-agent(15)) <p>l. data-req-init-agent-count (DataReqModeCapab)</p> <ul style="list-style-type: none"> <input type="checkbox"/> field- type = INT-U8
--	---

	<ul style="list-style-type: none"> <input type="checkbox"/> field-length = 2 bytes <input type="checkbox"/> field.value = 0x01 <p>m. data-req-init-manager-count (DataReqModeCapab)</p> <ul style="list-style-type: none"> <input type="checkbox"/> field- type = INT-U8 <input type="checkbox"/> field-length = 2 bytes <input type="checkbox"/> field.value = 0x00
Pass/Fail criteria	All checked attributes have proper values.
Notes	

TP Id	TP/PLT/PHD/CLASS/BCA/BV-015		
TP label	Set Time Body composition analyser PHD		
Coverage	Spec	[ISO/IEEE 11073-10420]	
	Testable items	MDSMethods 1; C	
Test purpose	<p>Check that:</p> <p>If the PHD supports the Absolute-Time-Stamp attribute, this method (Set Time) shall be implemented</p>		
Applicability	C_AG_OXP_167 AND C_AG_OXP_000 AND C_AG_OXP_009		
Other PICS			
Initial condition	The simulated PHG and the PHD under test are in the Operating state.		
Test procedure	<p>1. The simulated PHG sends a SET action:</p> <ul style="list-style-type: none"> <input type="checkbox"/> CHOICE = SetTimeInvoke <input type="checkbox"/> action-type = MDC_ACT_SET_TIME <input type="checkbox"/> the action-info-args are SetTimeInvoke <ul style="list-style-type: none"> ▪ date-time = <century, year ≤ 99 month ≤ 12 day ≤ 31 hour ≤ 24 minute ≤ 60 second ≤ 60 sec-fractions ≤ 100> ▪ accuracy = 0 <p>2. The PHD under test response shall be a rors-cmip-confirmed-action:</p> <ul style="list-style-type: none"> <input type="checkbox"/> action-type = MDC_ACT_SET_TIME <input type="checkbox"/> action-info-args shall be empty. 		
Pass/Fail criteria	All checked values are as specified in the test procedure.		
Notes			

TP Id	TP/PLT/PHD/CLASS/BCA/BV-016		
TP label	Config Changes Service. Contextual Attribute.		
Coverage	Spec	[ISO/IEEE 11073-10420]	
	Testable items	BCA_NumGen3; M	
Test purpose	<p>Check that:</p> <p>Whenever a contextual attribute changes, the PHD shall report these changes to the PHG using an MDS object event prior to reporting any of the dependent values</p> <p>[AND]</p> <p>Service component reports configuration changes to future measurements only</p>		
Applicability	C_AG_OXP_174 AND C_AG_BCA_005 AND C_AG_OXP_000		
Other PICS			

Initial condition	The simulated PHG and the PHD under test are in the Operating state.
Test procedure	<ol style="list-style-type: none"> 1. If the attribute that is going to be changed is reported in a Fixed format event report, take some measurements with the PHD under test. 2. Make a change to the contextual attribute Unit-Code for Body Weight object (Pounds to kg or kg to pounds). 3. The PHD shall send an MDS event report indicating the new contextual attribute value. 4. Take some more measurements. 5. Wait for the PHG to receive new event reports from the PHD, which report the measurements from step 4.
Pass/Fail criteria	<ul style="list-style-type: none"> ▪ The PHD sends an MDS event report to inform about the contextual attribute that has been changed. ▪ Data has changed accordingly to a new contextual attribute.
Notes	

TP Id	TP/PLT/PHD/CLASS/BCA/BV-017		
TP label	Config Changes Service. Body Height object Contextual Attribute.		
Coverage	Spec	[ISO/IEEE 11073-10420]	
	Testable items	BCA_NumGen3; M	
Test purpose	<p>Check that:</p> <p>Whenever a contextual attribute changes, the PHD shall report these changes to the PHG using an MDS object event prior to reporting any of the dependent values</p>		
Applicability	C_AG_OXP_174 AND C_AG_BCA_006 AND C_AG_OXP_000		
Other PICS			
Initial condition	The simulated PHG and the PHD under test are in the Operating state.		
Test procedure	<ol style="list-style-type: none"> 1. If the attribute that is going to be changed is reported in a Fixed format event report, take some measurements with the PHD under test. 2. Make a change to the contextual attribute Unit-Code for Body Height object (centimetres to inches or inches to centimetres). 3. The PHD shall send an MDS event report indicating the new contextual attribute value. 4. Take some more measurements. 5. Wait for the PHG to receive new event reports from the PHD, which report the measurements from step 4. 		
Pass/Fail criteria	<ul style="list-style-type: none"> ▪ The PHD sends an MDS event report to inform about the contextual attribute that has been changed. ▪ Data has changed accordingly to a new contextual attribute. 		
Notes			

TP Id	TP/PLT/PHD/CLASS/BCA/BV-018		
TP label	Config Changes Service. Body Fat object Contextual Attribute.		
Coverage	Spec	[ISO/IEEE 11073-10420]	
	Testable items	BCA_NumGen3; M	
Test purpose	<p>Check that:</p> <p>Whenever a contextual attribute changes, the PHD shall report these changes to the PHG using an MDS object event prior to reporting any of the dependent values</p> <p>[AND]</p> <p>Service component reports configuration changes to future measurements only</p>		

Applicability	C_AG_OXP_174 AND C_AG_BCA_007 AND C_AG_OXP_000
Other PICS	
Initial condition	The simulated PHG and the PHD under test are in the Operating state.
Test procedure	<ol style="list-style-type: none"> 1. If the attribute that is going to be changed is reported in a Fixed format event report, take some measurements with the PHD under test. 2. Make a change to the contextual attribute Unit-Code for Body Weight object (Pounds to kg or kg to pounds). 3. The PHD shall send an MDS event report indicating the new contextual attribute value. 4. Take some more measurements. 5. Wait for the PHG to receive new event reports from the PHD, which report the measurements from step 4.
Pass/Fail criteria	<ul style="list-style-type: none"> ▪ The PHD sends an MDS event report to inform about the contextual attribute that has been changed. ▪ Data has changed accordingly to a new contextual attribute.
Notes	

Bibliography

- [b-ITU-T H.810 (2013)] Recommendation ITU-T H.810 (2013), *Interoperability design guidelines for personal health systems*.
- [b-ITU-T H.810 (2015)] Recommendation ITU-T H.810 (2015), *Interoperability design guidelines for personal health systems*.
- [b-CDG 1.0] Continua Health Alliance, Continua Design Guidelines v1.0. (2008), *Continua Design Guidelines*.
- [b-CDG 2010] Continua Health Alliance, Continua Design Guidelines v1.5 (2010), *Continua Design Guidelines*.
- [b-CDG 2011] Continua Health Alliance, Continua Design Guidelines (2011), "Adrenaline", *Continua Design Guidelines*.
- [b-CDG 2012] Continua Health Alliance, Continua Design Guidelines (2012), "Catalyst ", *Continua Design Guidelines*.
- [b-CDG 2013] Continua Health Alliance, Continua Design Guidelines (2013), "Endorphin", *Continua Design Guidelines*.
- [b-CDG 2015] Continua Health Alliance, Continua Design Guidelines (2015), "Genome", *Continua Design Guidelines*.
- [b-CDG 2016] Personal Connected Health Alliance, Continua Design Guidelines (2016), "Iris", *Continua Design Guidelines*.
- [b-ETSI SR 001 262] ETSI SR 001 262 v1.8.1 (2003-12): *ETSI drafting rules*.
- [b-PHD PICS & PIXIT] Personal Health Device DG2016 PICS and PIXIT excel sheet v1.11.
<http://handle.itu.int/11.1002/2000/12067>
- [b-PHG PICS & PIXIT] Personal Health Gateway DG2016 PICS and PIXIT excel sheet v1.9.
<http://handle.itu.int/11.1002/2000/12067>
- [b-TCRL] Test Case Reference List_DG2016_v1.11.
<http://handle.itu.int/11.1002/2000/12067>
- [b-TI] Continua DG2016 PHD Testable items excel sheet v1.8.
<http://handle.itu.int/11.1002/2000/12067>

SERIES OF ITU-T RECOMMENDATIONS

Series A	Organization of the work of ITU-T
Series D	Tariff and accounting principles and international telecommunication/ICT economic and policy issues
Series E	Overall network operation, telephone service, service operation and human factors
Series F	Non-telephone telecommunication services
Series G	Transmission systems and media, digital systems and networks
Series H	Audiovisual and multimedia systems
Series I	Integrated services digital network
Series J	Cable networks and transmission of television, sound programme and other multimedia signals
Series K	Protection against interference
Series L	Environment and ICTs, climate change, e-waste, energy efficiency; construction, installation and protection of cables and other elements of outside plant
Series M	Telecommunication management, including TMN and network maintenance
Series N	Maintenance: international sound programme and television transmission circuits
Series O	Specifications of measuring equipment
Series P	Telephone transmission quality, telephone installations, local line networks
Series Q	Switching and signalling, and associated measurements and tests
Series R	Telegraph transmission
Series S	Telegraph services terminal equipment
Series T	Terminals for telematic services
Series U	Telegraph switching
Series V	Data communication over the telephone network
Series X	Data networks, open system communications and security
Series Y	Global information infrastructure, Internet protocol aspects, next-generation networks, Internet of Things and smart cities
Series Z	Languages and general software aspects for telecommunication systems