

International Telecommunication Union

ITU-T

TELECOMMUNICATION
STANDARDIZATION SECTOR
OF ITU

H.845.13

(07/2016)

SERIES H: AUDIOVISUAL AND MULTIMEDIA SYSTEMS

E-health multimedia services and applications –
Interoperability compliance testing of personal health
systems (HRN, PAN, LAN, TAN and WAN)

**Conformance of ITU-T H.810 personal health
devices: PAN/LAN/TAN interface Part 5M: Basic
electrocardiograph: Agent**

Recommendation ITU-T H.845.13



ITU-T H-SERIES RECOMMENDATIONS
AUDIOVISUAL AND MULTIMEDIA SYSTEMS

CHARACTERISTICS OF VISUAL TELEPHONE SYSTEMS	H.100–H.199
INFRASTRUCTURE OF AUDIOVISUAL SERVICES	
General	H.200–H.219
Transmission multiplexing and synchronization	H.220–H.229
Systems aspects	H.230–H.239
Communication procedures	H.240–H.259
Coding of moving video	H.260–H.279
Related systems aspects	H.280–H.299
Systems and terminal equipment for audiovisual services	H.300–H.349
Directory services architecture for audiovisual and multimedia services	H.350–H.359
Quality of service architecture for audiovisual and multimedia services	H.360–H.369
Telepresence	H.420–H.429
Supplementary services for multimedia	H.450–H.499
MOBILITY AND COLLABORATION PROCEDURES	
Overview of Mobility and Collaboration, definitions, protocols and procedures	H.500–H.509
Mobility for H-Series multimedia systems and services	H.510–H.519
Mobile multimedia collaboration applications and services	H.520–H.529
Security for mobile multimedia systems and services	H.530–H.539
Security for mobile multimedia collaboration applications and services	H.540–H.549
Mobility interworking procedures	H.550–H.559
Mobile multimedia collaboration inter-working procedures	H.560–H.569
BROADBAND, TRIPLE-PLAY AND ADVANCED MULTIMEDIA SERVICES	
Broadband multimedia services over VDSL	H.610–H.619
Advanced multimedia services and applications	H.620–H.629
Ubiquitous sensor network applications and Internet of Things	H.640–H.649
IPTV MULTIMEDIA SERVICES AND APPLICATIONS FOR IPTV	
General aspects	H.700–H.719
IPTV terminal devices	H.720–H.729
IPTV middleware	H.730–H.739
IPTV application event handling	H.740–H.749
IPTV metadata	H.750–H.759
IPTV multimedia application frameworks	H.760–H.769
IPTV service discovery up to consumption	H.770–H.779
Digital Signage	H.780–H.789
E-HEALTH MULTIMEDIA SERVICES AND APPLICATIONS	
Personal health systems	H.810–H.819
Interoperability compliance testing of personal health systems (HRN, PAN, LAN, TAN and WAN)	H.820–H.859
Multimedia e-health data exchange services	H.860–H.869

For further details, please refer to the list of ITU-T Recommendations.

Recommendation ITU-T H.845.13

Conformance of ITU-T H.810 personal health devices: PAN/LAN/TAN interface Part 5M: Basic electrocardiograph: Agent

Summary

Recommendation ITU-T H.845.13 is a transposition of Continua Test Tool DG2013, Test Suite Structure & Test Purposes, PAN-LAN-TAN Interface; Part 5M: Device Specializations. Agent (Basic Electrocardiograph) (Version 1.1, 2014-01-24), that was developed by the Continua Health Alliance. A number of versions of this specification existed before transposition.

This Recommendation includes an electronic attachment with the protocol implementation conformance statements (PICS) and the protocol implementation extra information for testing (PIXIT) required for the implementation of Annex A.

History

Edition	Recommendation	Approval	Study Group	Unique ID*
1.0	ITU-T H.845.13	2015-01-13	16	11.1002/1000/12273
2.0	ITU-T H.845.13	2016-07-14	16	11.1002/1000/12950

Keywords

Conformance testing, continua design guidelines, e-health, H.810, PAN/LAN/TAN interface, personal area network, personal connected health devices, touch area network.

* To access the Recommendation, type the URL <http://handle.itu.int/> in the address field of your web browser, followed by the Recommendation's unique ID. For example, <http://handle.itu.int/11.1002/1000/11830-en>.

FOREWORD

The International Telecommunication Union (ITU) is the United Nations specialized agency in the field of telecommunications, information and communication technologies (ICTs). The ITU Telecommunication Standardization Sector (ITU-T) is a permanent organ of ITU. ITU-T is responsible for studying technical, operating and tariff questions and issuing Recommendations on them with a view to standardizing telecommunications on a worldwide basis.

The World Telecommunication Standardization Assembly (WTSA), which meets every four years, establishes the topics for study by the ITU-T study groups which, in turn, produce Recommendations on these topics.

The approval of ITU-T Recommendations is covered by the procedure laid down in WTSA Resolution 1.

In some areas of information technology which fall within ITU-T's purview, the necessary standards are prepared on a collaborative basis with ISO and IEC.

NOTE

In this Recommendation, the expression "Administration" is used for conciseness to indicate both a telecommunication administration and a recognized operating agency.

Compliance with this Recommendation is voluntary. However, the Recommendation may contain certain mandatory provisions (to ensure, e.g., interoperability or applicability) and compliance with the Recommendation is achieved when all of these mandatory provisions are met. The words "shall" or some other obligatory language such as "must" and the negative equivalents are used to express requirements. The use of such words does not suggest that compliance with the Recommendation is required of any party.

INTELLECTUAL PROPERTY RIGHTS

ITU draws attention to the possibility that the practice or implementation of this Recommendation may involve the use of a claimed Intellectual Property Right. ITU takes no position concerning the evidence, validity or applicability of claimed Intellectual Property Rights, whether asserted by ITU members or others outside of the Recommendation development process.

As of the date of approval of this Recommendation, ITU had not received notice of intellectual property, protected by patents, which may be required to implement this Recommendation. However, implementers are cautioned that this may not represent the latest information and are therefore strongly urged to consult the TSB patent database at <http://www.itu.int/ITU-T/ipr/>.

© ITU 2016

All rights reserved. No part of this publication may be reproduced, by any means whatsoever, without the prior written permission of ITU.

Table of Contents

	Page
1 Scope.....	1
2 References.....	2
3 Definitions	2
3.1 Terms defined elsewhere	2
3.2 Terms defined in this Recommendation.....	2
4 Abbreviations and acronyms	2
5 Conventions	3
6 Test suite structure (TSS)	4
7 Electronic attachment	6
Annex A – Test purposes	7
A.1 TP definition conventions.....	7
A.2 Subgroup 1.3.13: Basic electrocardiograph (ECG).....	8
Bibliography.....	48

Electronic attachment: This Recommendation includes an electronic attachment with the protocol implementation conformance statements (PICS) and the protocol implementation extra information for testing (PIXIT) required for the implementation of Annex A.

Introduction

This Recommendation is a transposition of Continua Test Tool DG2013, Test Suite Structure & Test Purposes, PAN-LAN-TAN Interface; Part 5M: Device Specializations. Agent (Basic Electrocardiograph) (Version 1.1, 2014-01-24), that was developed by the Continua Health Alliance. A number of versions of this specification existed before transposition and these can be found in the table below.

Version	Date	Revision history
1.0	2013-05-24	Initial release for Test Tool DG2012.
1.1	2014-01-24	Initial release for Test Tool DG2013. This uses "TSS&TP_DG2012_PAN-LAN_PART_5M_v1.0.doc" as a baseline and adds new features included in [ITU-T H.810 (2015)]: <ul style="list-style-type: none">• Adds glucose meter BLE• Adds BLE SSP support• Adds NFC new transport• Adds INR device specialization

Recommendation ITU-T H.845.13

Conformance of ITU-T H.810 personal health devices: PAN/LAN/TAN interface Part 5M: Basic electrocardiograph: Agent

1 Scope

The scope of this Recommendation¹ is to provide a test suite structure and the test purposes (TSS & TP) for the PAN/LAN/TAN interface based on the requirements defined in the Continua Design Guidelines (CDG) [ITU-T H.810 (2015)]. The objective of this test specification is to provide a high probability of air interface interoperability between different devices.

The TSS and TP for the PAN/LAN/TAN interface document have been divided into ten parts. Each part is listed below:

- **Part 1:** Optimized exchange protocol [ISO/IEEE 11073-20601A] Agent
- **Part 2:** Optimized exchange protocol [ISO/IEEE 11073-20601A] Manager
- **Part 3:** Continua design guidelines. Agent
- **Part 4:** Continua design guidelines. Manager
- **Part 5:** Device specializations. Agent. This document is divided in 12 subparts:
 - **Part 5A:** Weighing scales
 - **Part 5B:** Glucose meter
 - **Part 5C:** Pulse oximeter
 - **Part 5D:** Blood pressure monitor
 - **Part 5E:** Thermometer
 - **Part 5F:** Cardiovascular fitness and activity monitor
 - **Part 5G:** Strength fitness equipment
 - **Part 5H:** Independent living activity hub
 - **Part 5I:** Adherence monitor
 - **Part 5J:** Insulin pump (Future development)
 - **Part 5K:** Peak flow
 - **Part 5L:** Body composition analyser
 - **Part 5M:** Basic electrocardiograph
 - **Part 5N:** International Normalized Ratio Monitor
- **Part 6:** Device specializations. Manager
- **Part 7:** Continua design guidelines. Agent BLE
- **Part 8:** Continua design guidelines. Manager BLE
- **Part 9:** Personal health devices transcoding whitepaper. Agent
- **Part 10:** Personal health devices transcoding whitepaper. Manager

¹ This Recommendation includes an electronic attachment with the protocol implementation conformance statements (PICS) and the protocol implementation extra information for testing (PIXIT) required for the implementation of Annex A.

2 References

The following ITU-T Recommendations and other references contain provisions which, through reference in this text, constitute provisions of this Recommendation. At the time of publication, the editions indicated were valid. All Recommendations and other references are subject to revision; users of this Recommendation are therefore encouraged to investigate the possibility of applying the most recent edition of the Recommendations and other references listed below. A list of the currently valid ITU-T Recommendations is regularly published. The reference to a document within this Recommendation does not give it, as a stand-alone document, the status of a Recommendation.

- [ITU-T H.810 (2015)] Recommendation ITU-T H.810 (2015), *Interoperability design guidelines for personal health systems*.
- [ITU-T H.810 (2016)] Recommendation ITU-T H.810 (2016), *Interoperability design guidelines for personal health systems*.
- [ISO/IEEE 11073-20601A] ISO/IEEE 11073-20601:2010, *Health informatics – Personal health device communication – Part 20601: Application profile – Optimized exchange protocol*, including ISO/IEEE 11073-20601:2010 Amd 1:2015.
<http://www.iso.org/iso/home/store/catalogue_tc/catalogue_detail.htm?csnumber=54331>
with
<http://www.iso.org/iso/home/store/catalogue_tc/catalogue_detail.htm?csnumber=63972>
- [ISO/IEEE 11073-104xx] ISO/IEEE 11073-104xx (in force), *Health informatics – Personal health device communication – Device specialization*.
NOTE – This is shorthand used to refer to the collection of device specialization standards that utilize [ISO/IEEE 11073-20601A], where xx can be any number from 01 to 99, inclusive.
- [ISO/IEEE 11073-10406] ISO/IEEE 11073-10406:2012, *Health informatics – Personal health device communication Part 10406: Device specialization – Basic electrocardiograph (ECG) (1- to 3-lead ECG)*.

3 Definitions

3.1 Terms defined elsewhere

This Recommendation uses the following terms defined elsewhere:

3.1.1 agent [ISO/IEEE 11073-20601A]: A node that collects and transmits personal health data to an associated manager.

3.1.2 manager [ISO/IEEE 11073-20601A]: A node receiving data from one or more agent systems. Some examples of managers include a cellular phone, health appliance, set top box, or a computer system.

3.2 Terms defined in this Recommendation

None.

4 Abbreviations and acronyms

This Recommendation uses the following abbreviations and acronyms:

DUT	Device Under Test
CDG	Continua Design Guidelines
INR	International Normalized Ratio

MDS	Medical Device System
NFC	Near Field Communication
PAN	Personal Area Network
PHD	Personal Healthcare Device
PHDC	Personal Healthcare Device Class
PICS	Protocol Implementation Conformance Statement
PIXIT	Protocol Implementation extra Information for Testing
SABTE	Sleep Apnoea Breathing Therapy Equipment
SDP	Service Discovery Protocol
SOAP	Simple Object Access Protocol
TP	Test Purpose
TSS	Test Suite Structure
USB	Universal Serial Bus

5 Conventions

The key words "SHALL", "SHALL NOT", "SHOULD", "SHOULD NOT", "MAY", "MAY NOT" in this Recommendation are to be interpreted as in [b-ETSI SR 001 262].

- SHALL is equivalent to 'must' or 'it is required to'.
- SHALL NOT is equivalent to 'must not' or 'it is not allowed'.
- SHOULD is equivalent to 'it is recommended to'.
- SHOULD NOT is equivalent to 'it is not recommended to'.
- MAY is equivalent to 'is permitted'.
- MAY NOT is equivalent to 'it is not required that'.

NOTE – The above-mentioned key words are capitalized for illustrative purposes only and they do not appear capitalized within this Recommendation.

Reference is made in the ITU-T H.800-series of Recommendations to different versions of the Continua Design Guidelines (CDG) by a specific designation. The list of terms that may be used in this Recommendation is provided in Table 1.

Table 1 – List of designations associated with the various versions of the CDG

CDG name	Transposed as	Version	Description	Designation
2016 plus errata	[ITU-T H.810 (2016)]	6.1	Release 2016 plus errata noting all ratified bugs [ITU-T H.810 (2016)].	–
2016	–	6.0	Release 2016 of the CDG including maintenance updates of the CDG 2015 and additional guidelines that cover new functionalities.	Iris
2015 plus errata	[ITU-T H.810 (2015)]	5.1	Release 2015 plus errata noting all ratified bugs [ITU-T H.810 (2015)].	–
2015	–	5.0	Release 2015 of the CDG including maintenance updates of the CDG 2013 and additional guidelines that cover new	Genome

Table 1 – List of designations associated with the various versions of the CDG

CDG name	Transposed as	Version	Description	Designation
			functionalities.	
2013 plus errata	[ITU-T H.810 (2013)]	4.1	Release 2013 plus errata noting all ratified bugs [b-ITU-T H.810 (2013)].	–
2013	–	4.0	Release 2013 of the CDG including maintenance updates of the CDG 2012 and additional guidelines that cover new functionalities.	Endorphin
2012 plus errata	–	3.1	Release 2012 plus errata noting all ratified bugs [b-CDG 2012].	–
2012	–	3.0	Release 2012 of the CDG including maintenance updates of the CDG 2011 and additional guidelines that cover new functionalities.	Catalyst
2011 plus errata	–	2.1	CDG 2011 integrated with identified errata.	–
2011	–	2.0	Release 2011 of the CDG including maintenance updates of the CDG 2010 and additional guidelines that cover new functionalities [b-CDG 2011].	Adrenaline
2010 plus errata	–	1.6	CDG 2010 integrated with identified errata	–
2010	–	1.5	Release 2010 of the CDG with maintenance updates of the CDG Version 1 and additional guidelines that cover new functionalities [b-CDG 2010].	1.5
1.0	–	1.0	First released version of the CDG [b-CDG 1.0].	–

6 Test suite structure (TSS)

The test purposes (TPs) for the PAN/LAN/TAN interface have been divided into the main subgroups specified below. Annex A describes the TPs for subgroup 1.3.13 (shown in bold).

- Group 1: Agent (AG)
 - Group 1.1: Transport (TR)
 - Subgroup 1.1.1: Design guidelines: Common (DGC)
 - Subgroup 1.1.2: USB design guidelines (UDG)
 - Subgroup 1.1.3: Bluetooth design guidelines (BDG)
 - Subgroup 1.1.4: Pulse oximeter design guidelines (PODG)
 - Subgroup 1.1.5: Cardiovascular design guidelines (CVDG)
 - Subgroup 1.1.6: Activity hub design guidelines (HUBDG)
 - Subgroup 1.1.7: ZigBee design guidelines (ZDG)
 - Subgroup 1.1.8: Glucose meter design guidelines (GLDG)
 - Subgroup 1.1.9: Bluetooth low energy design guidelines (BLEDG)
 - Subgroup 1.1.10: Basic electrocardiograph design guidelines (ECGDG)

- Subgroup 1.1.11: NFC design guidelines (NDG)
- Group 1.2: 20601: Optimized exchange protocol (OXP)
 - Subgroup 1.2.1: PHD domain information model (DIM)
 - Subgroup 1.2.2: PHD service model (SER)
 - Subgroup 1.2.3: PHD communication model (COM)
- Group 1.3: Devices class specializations (CLASS)
 - Subgroup 1.3.1: Weighing scales (WEG)
 - Subgroup 1.3.2: Glucose meter (GL)
 - Subgroup 1.3.3: Pulse oximeter (PO)
 - Subgroup 1.3.4: Blood pressure monitor (BPM)
 - Subgroup 1.3.5: Thermometer (TH)
 - Subgroup 1.3.6: Cardiovascular (CV)
 - Subgroup 1.3.7: Strength (ST)
 - Subgroup 1.3.8: Activity hub (HUB)
 - Subgroup 1.3.9: Adherence monitor (AM)
 - Subgroup 1.3.10: Insulin pump (IP) (Future development)
 - Subgroup 1.3.11: Peak flow (PF)
 - Subgroup 1.3.12: Body composition analyzer (BCA)
 - **Subgroup 1.3.13: Basic electrocardiograph (ECG)**
 - Subgroup 1.3.14: International normalized ratio (INR)
 - Subgroup 1.3.15: Sleep apnoea breathing therapy equipment (SABTE)
- Group 1.4: Personal health device transcoding whitepaper (PHDTW)
 - Subgroup 1.4.1: Whitepaper general requirements (GEN)
 - Subgroup 1.4.2: Whitepaper thermometer requirements (TH)
 - Subgroup 1.4.3: Whitepaper blood pressure requirements (BPM)
 - Subgroup 1.4.4: Whitepaper heart rate requirements (HR)
 - Subgroup 1.4.5: Whitepaper glucose meter requirements (GL)
 - Subgroup 1.4.6: Whitepaper weight scale requirements (WS)
- Group 2: Manager (MAN)
 - Group 2.1: Transport (TR)
 - Subgroup 2.1.1: Design guidelines: Common (DGC)
 - Subgroup 2.1.2: USB design guidelines (UDG)
 - Subgroup 2.1.3: Bluetooth design guidelines (BDG)
 - Subgroup 2.1.4: Cardiovascular design guidelines (CVDG)
 - Subgroup 2.1.5: Activity hub design guidelines (HUBDG)
 - Subgroup 2.1.6: ZigBee design guidelines (ZDG)
 - Subgroup 2.1.7: Bluetooth low energy design guidelines (BLEDG)
 - Subgroup 2.1.8: NFC design guidelines (NDG)

- Group 2.2: 20601: Optimized exchange protocol (OXP)
 - Subgroup 2.2.1: General (GEN)
 - Subgroup 2.2.2: PHD domain information model (DIM)
 - Subgroup 2.2.3: PHD service model (SER)
 - Subgroup 2.2.4: PHD communication model (COM)
- Group 2.3: Devices class specializations (CLASS)
 - Subgroup 2.3.1: Weighing scales (WEG)
 - Subgroup 2.3.2: Glucose meter (GL)
 - Subgroup 2.3.3: Pulse oximeter (PO)
 - Subgroup 2.3.4: Blood pressure monitor (BPM)
 - Subgroup 2.3.5: Thermometer (TH)
 - Subgroup 2.3.6: Cardiovascular (CV)
 - Subgroup 2.3.7: Strength (ST)
 - Subgroup 2.3.8: Activity hub (HUB)
 - Subgroup 2.3.9: Adherence monitor (AM)
 - Subgroup 2.3.10: Insulin pump (IP) (Future development)
 - Subgroup 2.3.11: Peak flow (PF)
 - Subgroup 2.3.12: Body composition analyser (BCA)
 - Subgroup 2.3.13: Basic electrocardiograph (ECG)
 - Subgroup 2.3.14: International normalized ratio (INR)
 - Subgroup 2.3.15: Sleep apnoea breathing therapy equipment (SABTE)
- Group 2.4: Personal health device transcoding whitepaper (PHDTW)
 - Subgroup 2.4.1: Whitepaper general requirements (GEN)
 - Subgroup 2.4.2: Whitepaper thermometer requirements (TH)
 - Subgroup 2.4.3: Whitepaper blood pressure requirements (BPM)
 - Subgroup 2.4.4: Whitepaper heart rate requirements (HR)
 - Subgroup 2.4.5: Whitepaper glucose meter requirements (GL)
 - Subgroup 2.4.6: Whitepaper weight scale requirements (WS)

7 Electronic attachment

The protocol implementation conformance statements (PICS) and the protocol implementation extra information for testing (PIXIT) required for the implementation of this annex can be downloaded from <http://handle.itu.int/11.1002/2000/12067>.

In the electronic attachment, letters "C" and "I" in the column labelled "Mandatory" are used to distinguish between "PICS" and "PIXIT" respectively during testing. If the cell is empty, the corresponding PICS is "independent". If the field contains a "C", the corresponding PICS is dependent on other PICS, and the logical expression is detailed in the "SCR_Expression" field. The static conformance review (SCR) is used in the test tool to assert whether the PICS selection is consistent.

Annex A

Test purposes

(This annex forms an integral part of this Recommendation.)

A.1 TP definition conventions

The test purposes (TPs) are defined according to the following rules:

- **TP Id:** It is a unique identifier (TP/<TT>/<DUT>/<GR>/<SGR>/<XX> – <NNN>). It is specified according to the naming convention defined below:
 - Each test purpose identifier is introduced by the prefix "TP".
 - <TT>: This is the test tool that will be used in the test case:
 - PAN: Personal area network (Bluetooth or USB)
 - LAN: Local area network (ZigBee)
 - PAN-LAN: Personal area network (Bluetooth or USB) - Local area network (ZigBee)
 - LP-PAN: Low power personal area network (Bluetooth low energy)
 - TAN: Touch area network (NFC)
 - PLT: Personal area network (Bluetooth or USB) – Local area network (ZigBee) – Touch area network (NFC)
 - <DUT>: This is the device under test:
 - AG: PAN/LAN Agent
 - MAN: PAN/LAN Manager
 - <GR>: This identifies a group of test cases.
 - <SGR>: This identifies a subgroup of test cases.
 - <XX>: This identifies the type of testing:
 - BV: Valid behaviour test
 - BI: Invalid behaviour test
 - <NNN>: This is a sequential number that identifies the test purpose.
- **TP label:** This is the TP's title.
- **Coverage:** This contains the specification reference and clause to be checked by the TP.
 - Spec: This indicates the earliest version of the specification from which the testable items to be checked by the TP were included.
 - Testable item: This contains testable items to be checked by the TP.
- **Test purpose:** It is a description about the requirements to be tested.
- **Applicability:** This contains the PICS items that define if the test case is applicable or not for a specific device. When a TP contains an "ALL" in this field it means that it applies to the device under test within that scope of the test (specialization, transport used, etc.).
- **Other PICS:** It contains additional PICS items (apart from the PICS specified in the Applicability row) which are used within the test case implementation and can modify the final verdict. When this row is empty, it means that only the PICS specified in the Applicability row are used within the test case implementation.
- **Initial condition:** This indicates the state to which the device under test (DUT) needs to be moved at the beginning of TC execution.

- **Test procedure:** This describes the steps to be followed in order to execute the test case.
- **Pass/Fail criteria:** This provides criteria to decide whether the DUT passes or fails the test case.

A.2 Subgroup 1.3.13: Basic electrocardiograph (ECG)

TP Id		TP/PLT/AG/CLASS/ECG/BV-000		
TP label		Get MDS Object for Basic ECG specialization/Heart Rate profile: Mandatory, Conditional and Optional Attributes		
Coverage	Spec	[ISO/IEEE 11073-10406]		
	Testable items	ECG_MDSAttr1; M	ECG_MDSAttr2; M	ECG_MDSAttr3; M
		ECG_MDSAttr4; M	ECG_MDSAttr5; M	ECG_MDSAttr6; R
		ECG_MDSAttr7; R	ECG_MDSAttr8; R	ECG_MDSAttr10; M
OperProc2; M				
Test purpose		<p>Check that:</p> <p>The Agent supports a Get command that requests all attributes [AND]</p> <p>The MDS Object contains the attributes specified for a Heart Rate Agent.</p>		
Applicability		C_AG_OXP_164 AND C_AG_OXP_000		
Other PICS		C_AG_OXP_181		
Initial condition		The simulated manager and the agent under test are in the operating state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated manager issues a "roiv-cmip-get" command with the handle set to 0 (to request for an MDS object) and the attribute-id-list set to 0 to indicate all attributes. 2. The agent under test responds with a "rors-cmip-get" service message in which the attribute-list contains a list of all implemented attributes of the MDS object: <p>MDS Attributes:</p> <ol style="list-style-type: none"> a. Not Recommended attribute System-Type. <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SYS_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <not relevant> b. Mandatory attribute System-Type-Spec-List <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SYS_TYPE_SPEC_LIST <input type="checkbox"/> attribute-type = TypeVerList <input type="checkbox"/> attribute-value.length = 4 bytes attribute-value = {MDC_DEV_SPEC_PROFILE_ECG, 1} and {MDC_DEV_SUB_SPEC_PROFILE_HR, 1} c. Mandatory attribute System-model <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_MODEL <input type="checkbox"/> attribute-type = SystemModel <input type="checkbox"/> attribute-value.length = <variable> <input type="checkbox"/> attribute-value = {Manufacturer, Model} 		

	<p>d. Mandatory attribute Dev-Configuration-Id</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_DEV_CONFIG_ID <input type="checkbox"/> attribute-type = ConfigId <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <ul style="list-style-type: none"> – IF NOT C_AG_OXP_181 then attribute-value = 0x0258 – ELSE attribute-value = < between 0x4000 and 0x7FFF > <p>e. Recommended attribute Power-Status</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_POWER_STAT <input type="checkbox"/> attribute-type = PowerStatus (BITS-16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <ul style="list-style-type: none"> ON_MAINS (0x8000) or ON_BATTERY(0x4000) Only one of the following may be active: <ul style="list-style-type: none"> ▪ chargingFull(8), ▪ chargingTrickle(9), ▪ chargingOff(10). ▪ The rest of the bits must not be set <p>f. Recommended attribute Battery-Level</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_VAL_BATT_CHARGE (0X09 0X9C) <input type="checkbox"/> attribute-type = INT-U16 <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <value between 0 and 100> If value >100, the meaning of the value is “undefined” <p>g. Recommended attribute Remain-Battery-Time</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_BATT_REMAIN (0X09 0X88) <input type="checkbox"/> attribute-type = BatMeasure <input type="checkbox"/> attribute-value.length = 6 bytes <input type="checkbox"/> attribute-value = <4 bytes to define the value. 2 remaining bytes to define the units, which shall be set to one of: MDC_DIM_MIN (0x08 0xA0), MDC_DIM_HR (0x08 0xC0) or MDC_DIM_DAY (0x08 0xE0) >
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/AG/CLASS/ECG/BV-001			
TP label	Get MDS Object for Basic ECG specialization/Simple ECG profile: Mandatory, Conditional and Optional Attributes			
Coverage	Spec	[ISO/IEEE 11073-10406]		
	Testable items	ECG_MDSAttr1; M	ECG_MDSAttr2; M	ECG_MDSAttr3; M
		ECG_MDSAttr4; M	ECG_MDSAttr5; M	ECG_MDSAttr6; R
		ECG_MDSAttr7; R	ECG_MDSAttr8; R	ECG_MDSAttr10; M
	OperProc2; M			

Test purpose	<p>Check that:</p> <p>The Agent supports a Get command that requests all attributes</p> <p>[AND]</p> <p>The MDS Object contains the attributes specified for a Simple ECG Agent.</p>
Applicability	C_AG_OXP_165 AND C_AG_OXP_000
Other PICS	C_AG_OXP_181
Initial condition	The simulated manager and the agent under test are in the operating state.
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager issues a "roiv-cmip-get" command with the handle set to 0 (to request for an MDS object) and the attribute-id-list set to 0 to indicate all attributes. 2. The agent under test responds with a "rors-cmip-get" service message in which the attribute-list contains a list of all implemented attributes of the MDS object: <ul style="list-style-type: none"> MDS Attributes: <ol style="list-style-type: none"> a. Not recommended attribute System-Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SYS_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <not relevant> b. Mandatory attribute System-Type-Spec-List <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SYS_TYPE_SPEC_LIST <input type="checkbox"/> attribute-type = TypeVerList <input type="checkbox"/> attribute-value.length = 4 bytes attribute-value = {MDC_DEV_SPEC_PROFILE_ECG, 1} and {MDC_DEV_SUB_SPEC_PROFILE_ECG, 1} c. Mandatory attribute System-model <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_MODEL <input type="checkbox"/> attribute-type = SystemModel <input type="checkbox"/> attribute-value.length = <variable> <input type="checkbox"/> attribute-value = {Manufacturer, Model} d. Mandatory attribute Dev-Configuration-Id <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_DEV_CONFIG_ID <input type="checkbox"/> attribute-type = ConfigId <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = < between 0x4000 and 0x7FFF> e. Recommended attribute Power-Status <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_POWER_STAT <input type="checkbox"/> attribute-type = PowerStatus (BITS-16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <ul style="list-style-type: none"> ON_MAINS (0x8000) or ON_BATTERY(0x4000) <p>Only one of the following may be active:</p> <ul style="list-style-type: none"> ▪ chargingFull(8), ▪ chargingTrickle(9), ▪ chargingOff(10). ▪ The rest of the bits must not be set

	<p>f. Recommended attribute Battery-Level</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_VAL_BATT_CHARGE (0X09 0X9C) <input type="checkbox"/> attribute-type = INT-U16 <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <value between 0 and 100> If value >100, the meaning of the value is “undefined” <p>g. Recommended attribute Remain-Battery-Time</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_BATT_REMAIN (0X09 0X88) <input type="checkbox"/> attribute-type = BatMeasure <input type="checkbox"/> attribute-value.length = 6 bytes <input type="checkbox"/> attribute-value = <4 bytes to define the value. 2 remaining bytes to define the units, which shall be set to one of: MDC_DIM_MIN (0x08 0xA0), MDC_DIM_HR (0x08 0xC0) or MDC_DIM_DAY (0x08 0xE0) >
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/AG/CLASS/ECG/BV-002			
TP label	MDS Configuration objects events for Basic ECG specialization/Heart Rate profile			
Coverage	Spec	[ISO/IEEE 11073-10406]		
	Testable items	ECG_MDSEvent1; M	ECG_NumGen1; M	ECG_RTSAgen1; M
		ECG_EnumGen1; M	HeartRate1; C	HeartRateProfile1; M
		HeartRateProfile2; O	HeartRateProfile3; O	HeartRateProfile4; O
		HeartRateProfile5; M	ConfigProc1; M	
Test purpose	<p>Check that:</p> <p>A Basic ECG Agent shall send the [MDS-Configuration-Event] using a [Confirmed] event report. The [MDS-Configuration-Event] shall include the event-info [ConfigReport].</p> <p>[AND]</p> <p>Check objects supported by the Heart Rate Agent (standard /extended configuration)</p>			
Applicability	C_AG_OXP_164 AND C_AG_OXP_000			
Other PICS	C_AG_OXP_010, C_AG_OXP_181, C_AG_ECG_001, C_AG_ECG_002, C_AG_ECG_003, C_AG_ECG_004, C_AG_ECG_005			
Initial condition	The simulated manager and the agent under test are in the unassociated state.			
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent responds with a “Remote Operation Invoke Confirmed Event Report” message with an MDC_NOTI_CONFIG event to send its configuration to the manager: <ol style="list-style-type: none"> a. APDU Type <ul style="list-style-type: none"> <input type="checkbox"/> field- type = PrstApdu <input type="checkbox"/> field-length =2 bytes <input type="checkbox"/> field-value =0xE7 0x00 			

	<ul style="list-style-type: none"> b. invoke-id <ul style="list-style-type: none"> <input type="checkbox"/> field- type = InvokeIDType <input type="checkbox"/> field-length =INT-U16 <input type="checkbox"/> field- value=<Not relevant for this test> c. message <ul style="list-style-type: none"> <input type="checkbox"/> field- type = roiv-cmip-confirmed-event-report <input type="checkbox"/> field-length =two bytes <input type="checkbox"/> field- value=0x01 0x01 (EventReportArgumentSimple) d. obj-handle (EventReportArgumentSimple) <ul style="list-style-type: none"> <input type="checkbox"/> field- type = HANDLE <input type="checkbox"/> field-length =INT-U16 e. event-time (EventReportArgumentSimple) <ul style="list-style-type: none"> <input type="checkbox"/> field- type = Relative Time <input type="checkbox"/> field-length =INT-U32 <input type="checkbox"/> field-value = <ul style="list-style-type: none"> • IF NOT C_AG_OXP_010 THEN value = 0xFF 0xFF 0xFF 0xFF f. event-type (EventReportArgumentSimple) <ul style="list-style-type: none"> <input type="checkbox"/> field- type = OID-Type <input type="checkbox"/> field-length =INT-U16 <input type="checkbox"/> field- value=0x 0D 0x 1C (MDC_NOTI_CONFIG) g. config-report-id (ConfigReport) <ul style="list-style-type: none"> <input type="checkbox"/> field- type = ConfigId <input type="checkbox"/> field-length = INT-U16 <input type="checkbox"/> field- value = <ul style="list-style-type: none"> – IF NOT C_AG_OXP_181 then attribute-value = 0x0258 – ELSE attribute-value = < between 0x4000 and 0x7FFF > h. obj-class (ConfigReport → ConfigObjectList (ConfigObject)). To check the objects that are supported by the Agent, Type Attribute will be checked in AttributeList. <ul style="list-style-type: none"> <input type="checkbox"/> field- type = OID-Type <input type="checkbox"/> field-length = INT-U16 <input type="checkbox"/> field- value = <ul style="list-style-type: none"> – One mandatory numeric object for Heart Rate. – One optional numeric object for R-R Interval. – One to three optional RT-SA objects for ECG Waveforms. – Two optional enumeration objects for Device Status and Context Data Trigger.
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/AG/CLASS/ECG/BV-003
TP label	MDS Configuration objects events for Basic ECG specialization/Simple ECG profile

Coverage	Spec	[ISO/IEEE 11073-10406]		
	Testable items	ECG_MDSEvent1; M	ECG_NumGen1; M	ECG_RTSAgen1; M
		ECG_EnumGen1; M	SimpleECGProfile1; M	SimpleECGProfile2; O
		SimpleECGProfile3; O	SimpleECGProfile4; O	SimpleECGProfile5; O
	ConfigProc1; M			
Test purpose	<p>Check that:</p> <p>A Basic ECG Agent shall send the [MDS-Configuration-Event] using a [Confirmed] event report. The [MDS-Configuration-Event] shall include the event-info [ConfigReport].</p> <p>[AND]</p> <p>Check objects supported by the Simple ECG Agent (standard /extended configuration)</p>			
Applicability	C_AG_OXP_165 AND C_AG_OXP_000			
Other PICS	C_AG_OXP_010, C_AG_OXP_181, C_AG_ECG_001, C_AG_ECG_002, C_AG_ECG_003, C_AG_ECG_004, C_AG_ECG_005			
Initial condition	The simulated manager and the agent under test are in the unassociated state.			
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent responds with a “Remote Operation Invoke Confirmed Event Report” message with an MDC_NOTI_CONFIG event to send its configuration to the manager: <ol style="list-style-type: none"> a. APDU Type <ul style="list-style-type: none"> <input type="checkbox"/> field- type = PrstApdu <input type="checkbox"/> field-length =2 bytes <input type="checkbox"/> field-value =0xE7 0x00 b. invoke-id <ul style="list-style-type: none"> <input type="checkbox"/> field- type = InvokeIDType <input type="checkbox"/> field-length =INT-U16 <input type="checkbox"/> field- value=<Not relevant for this test> c. message <ul style="list-style-type: none"> <input type="checkbox"/> field- type = roiv-cmip-confirmed-event-report <input type="checkbox"/> field-length =two bytes <input type="checkbox"/> field- value=0x01 0x01 (EventReportArgumentSimple) d. obj-handle (EventReportArgumentSimple) <ul style="list-style-type: none"> <input type="checkbox"/> field- type = HANDLE <input type="checkbox"/> field-length =INT-U16 e. event-time (EventReportArgumentSimple) <ul style="list-style-type: none"> <input type="checkbox"/> field- type = Relative Time <input type="checkbox"/> field-length =INT-U32 <input type="checkbox"/> field-value = <ul style="list-style-type: none"> • IF NOT C_AG_OXP_010 THEN value = 0xFF 0xFF 0xFF 0xFF f. event-type (EventReportArgumentSimple) <ul style="list-style-type: none"> <input type="checkbox"/> field- type = OID-Type <input type="checkbox"/> field-length =INT-U16 			

	<ul style="list-style-type: none"> <input type="checkbox"/> field- value=0x 0D 0x 1C (MDC_NOTI_CONFIG) g. config-report-id (ConfigReport) <ul style="list-style-type: none"> <input type="checkbox"/> field- type = ConfigId <input type="checkbox"/> field-length = INT-U16 <input type="checkbox"/> field- value = < between 0x4000 and 0x7FFF > h. obj-class (ConfigReport → ConfigObjectList (ConfigObject)). To check the objects that are supported by the agent, the Type Attribute will be checked in AttributeList. <ul style="list-style-type: none"> <input type="checkbox"/> field- type = OID-Type <input type="checkbox"/> field-length = INT-U16 <input type="checkbox"/> field- value = <ul style="list-style-type: none"> – One to three mandatory RT-SA objects for ECG Waveforms numeric objects for. – Two optional numeric objects, one for Heart Rate and other for R-R Interval. – Two optional enumeration objects, one for Device Status and the other for Context Data Trigger.
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id		TP/PLT/AG/CLASS/ECG/BV-004		
TP label		MDS objects events Basic ECG specialization		
Coverage	Spec	[ISO/IEEE 11073-10406]		
	Testable items	ECG_MDSEvent3; M	ECG_MDSEvents 4; M	ECG_MDSEvents 5; M
		ECG_MDSEvents 6; M	ObjAccServ1; M	ObjAccServ3; M
		ObjAccServ4; M	ObjAccServ5; O	ObjAccServ7; O
Test purpose		<p>Check that:</p> <p>The Agent sends the MDS-Dynamic-Data-Update-Fixed using a confirmed or unconfirmed event report and it includes the event-info ScanReportInfoFixed.</p> <p>[AND/OR]</p> <p>The Agent sends the MDS-Dynamic-Data-Update-Var using a confirmed or unconfirmed event report and it includes the event-info ScanReportInfoVar.</p> <p>[AND]</p> <p>Agent-initiated mode shall be supported for measurement data transmission.</p> <p>[AND]</p> <p>A Simple ECG or Heart Rate agent may support either one or both single-person and multi-person event reports</p> <p>[AND]</p> <p>A Heart Rate agent with standard configuration shall use the fixed format data update messages method for transmitting measurement data</p> <p>[AND]</p> <p>A Simple ECG or Heart Rate agent with extended configuration may use either fixed or variable format data update messages for transmitting measurement data.</p>		

Applicability	(C_AG_OXP_164 OR C_AG_OXP_165) AND C_AG_OXP_000 AND (C_AG_OXP_182 OR C_AG_OXP_183 OR C_AG_OXP_184 OR C_AG_OXP_189)
Other PICS	C_AG_OXP_181
Initial condition	The simulated manager and the agent under test are in the unassociated state.
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent under test responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 4. Check that the field Dev-Config-Id is set to the tested configuration. If it is not, the manager responds with an "unsupported-config" and waits for a new configuration. Repeat this step until a Dev-config-Id equal to the tested configuration is received. 5. Record the agent configuration. 6. Take Measurements for every supported object in the agent under test. 7. Wait to receive every event report and check: <ul style="list-style-type: none"> <input type="checkbox"/> field- type = Event Report <input type="checkbox"/> field-length = 2 bytes <input type="checkbox"/> field- value=0x01 0x01 (EventReportArgumentSimple, confirmed) This field identifies the type of message sent by the agent, for the confirmed event configuration, roiv-cmip-confirmed-event-report.
Pass/Fail criteria	<ul style="list-style-type: none"> • Check that every received MDS Event report is one of the following Data APDU and that it is confirmed. • For Standard Configuration (NOT C_AG_OXP_181): an MDS Event Report is sent by the agent under test to report measurements for every object: <ul style="list-style-type: none"> <input type="checkbox"/> MDC_NOTI_SCAN_REPORT_FIXED <input type="checkbox"/> MDC_NOTI_SCAN_REPORT_MP_FIXED • For Extended Configuration, an MDS Event Report is sent by the agent under test to report measurements for every object: <ul style="list-style-type: none"> <input type="checkbox"/> MDC_NOTI_SCAN_REPORT_FIXED <input type="checkbox"/> MDC_NOTI_SCAN_REPORT_MP_FIXED <input type="checkbox"/> MDC_NOTI_SCAN_REPORT_VAR <input type="checkbox"/> MDC_NOTI_SCAN_REPORT_MP_VAR
Notes	

TP Id	TP/PLT/AG/CLASS/ECG/BV-005			
TP label	Heart rate Object for Standard Configuration (0x0258)			
Coverage	Spec	[ISO/IEEE 11073-10406]		
	Testable items	HeartRate2; M	HeartRate4; M	HeartRate6; R
		HeartRate8; M	HeartRate10; R	HeartRate12; O
		HeartRate14; R	HeartRate16; R	HeartRate18; R
		HeartRate20; M	HeartRate22; M	HeartRate24; R
		HeartRate26; O	HeartRate28; O	HeartRate30; R

		HeartRate32; R	HeartRate34; C	HeartRate36; R
		HeartRate38; R	HeartRate40; R	HeartRate42; R
		HeartRate44; C	HeartRate46; R	HeartRate48; R
		HeartRate50; R	HeartRate52; R	HeartRate54; M
		HeartRate55; C	HeartRate56; C	ConfigProc2; M
Test purpose	Check that: Heart Rate Numeric Object contains the attributes specified for Standard Configuration (0x0258)			
Applicability	C_AG_OXP_164 AND (NOT C_AG_OXP_181) AND C_AG_OXP_000			
Other PICS				
Initial condition	The simulated manager and the agent under test are in the unassociated state.			
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 4. Check that the field Dev-Config-Id is set to 0x0258. If it is not, the manager responds with an "unsupported-config" and waits for a new configuration. Repeat this step until a Dev-config-Id equal to 0x0258 is received. 5. Once the agent under test sends a standard configuration, check Heart Rate object. 6. The Heart Rate object contents shall be: <ol style="list-style-type: none"> a. Mandatory attribute Handle <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_HANDLE <input type="checkbox"/> attribute-type = HANDLE <input type="checkbox"/> attribute-value = 0x00 0x01 b. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = 0x00 0x02 (MDC_PART_SCADA), 0x41 0x82 (MDC_ECG_HEART_RATE 16770) c. Mandatory attribute Metric-Spec-Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value ≠ 0x00 0x00 <ul style="list-style-type: none"> • Bit 1 (mss-avail-stored-data(1)) is set. • Bit 9 (mss-acc-agent-initiated(9)) is set. • IF bit 3 (mss-msmt-aperiodic) is set THEN bit 5 (mss-msmt-btb-metric) d. Mandatory attribute Unit-Code <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE <input type="checkbox"/> attribute-type = OID-Type <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_DIM_BEAT_PER_MIN 			

	<p>e. Mandatory attribute Attribute-Value-Map</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ATTRIBUTE_VAL_MAP <input type="checkbox"/> attribute-type = AttrValMap <input type="checkbox"/> attribute-count = 2 <input type="checkbox"/> attribute-value = (MDC_ATTR_NU_VAL_OBS_BASIC, 2 MDC_ATTR_TIME_STAMP_REL, 4) <p>7. Check that no other attributes are present in the initial configuration.</p>
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/AG/CLASS/ECG/BV-006			
TP label	Heart Rate Object for Extended Configuration			
Coverage	Spec	[ISO/IEEE 11073-10406]		
	Testable items	HeartRate3; M	HeartRate5; M	HeartRate7; R
		HeartRate9; M	HeartRate11; R	HeartRate13; O
		HeartRate15; R	HeartRate17; R	HeartRate19; R
		HeartRate21; M	HeartRate23; C	HeartRate25; R
		HeartRate27; O	HeartRate29; O	HeartRate31; C
		HeartRate33; C	HeartRate35; C	HeartRate37; C
		HeartRate39; R	HeartRate41; C	HeartRate43; C
		HeartRate45; C	HeartRate47; C	HeartRate49; C
		HeartRate51; C	HeartRate53; R	HeartRate55; C
HeartRate56; C				
Test purpose	Check that: Heart Rate Numeric Object contains the attributes specified for Extended Configuration			
Applicability	(C_AG_OXP_164 OR C_AG_OXP_165) AND C_AG_ECG_001 AND C_AG_OXP_181 AND C_AG_OXP_000			
Other PICS	C_AG_OXP_041, C_AG_OXP_046, C_AG_OXP_047, C_AG_OXP_183, C_AG_OXP_189			
Initial condition	The simulated manager and the agent under test are in the unassociated state.			
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent under test responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 4. Check that the field Dev-Config-Id is set to the tested extended configuration. If it is not, the manager responds with an "unsupported-config" and waits for a new configuration. Repeat this step until a Dev-config-Id equal to the extended configuration is received. 5. Once the agent under test sends the tested configuration, check the Heart Rate object. 			

6. The Heart Rate object contents shall be:
- a. Mandatory attribute Type
 - attribute-id = MDC_ATTR_ID_TYPE
 - attribute-type = TYPE
 - attribute-value = one of these values:
 - 0x00 0x02 (MDC_PART_SCADA), 0x41 0x82 (MDC_ECG_HEART_RATE 16770)
 - 0x00 0x80 (MDC_PART_DM 182), 0x55 0xDE (MDC_ECG_HEART_RATE_INSTANT 21982)
 - b. IF Not Recommended attribute Supplemental-Types
 - attribute-id = MDC_ATTR_SPPLEMENTAL_TYPES
 - attribute-type = SupplementalTypeList
 - attribute-value.length = <variable>Sequence of TYPE (TYPE.length= 4 bytes)
 - attribute-value = <Not relevant for this test>
 - c. Mandatory attribute Metric-Spec-Small
 - attribute-id = MDC_ATTR_METRIC_SPEC_SMALL
 - attribute-type = MetricSpecSmall
 - attribute-value.length = 2 bytes
 - attribute-value =
 - IF bit 3 (mss-msmt-aperiodic) is set THEN bit 5 (mss-msmt-btb-metric)
 - d. IF Not recommended attribute Metric-Structure-Small is present
 - attribute-id = MDC_ATTR_METRIC_STRUCTURE_SMALL
 - attribute-type = MetricStructureSmall
 - attribute-length = 2 bytes
 - attribute-value = <Not relevant for this test>
 - e. IF Optional attribute Measurement-Status is present
 - attribute-id = MDC_ATTR_MSMT_STAT
 - attribute-type = MeasurementStatus
 - attribute-value.length = 2 bytes
 - attribute-value = <Not relevant for this test>
 - f. IF Not recommended attribute Metric-Id is present
 - attribute-id = MDC_ATTR_ID_PHYSIO
 - attribute-type = OID-Type(INT-U16)
 - attribute-value.length =2 bytes
 - attribute-value = <Not relevant for this test>
 - g. IF Not Recommended attribute Metric-Id-List is present
 - attribute-id = MDC_ATTR_ID_PHYSIO_LIS
 - attribute-type = MetricIdList
 - attribute-value = <Not relevant for this test>
 - h. IF Not recommended attribute Metric-Id-Partition is present
 - attribute-id = MDC_ATTR_METRIC_ID_PART
 - attribute-type = NomPartition(INT-U16)
 - attribute-value.length = 2 bytes
 - attribute-value = <Not relevant for this test>

	<ul style="list-style-type: none"> i. Mandatory attribute Unit-Code <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE <input type="checkbox"/> attribute-type = OID-Type <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_DIM_BEAT_PER_MIN j. IF Not recommended attribute Source-Handle-Reference is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SOURCE_HANDLE_REF <input type="checkbox"/> attribute-type = HANDLE(INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> k. IF Not recommended attribute Measure-Active-Period <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_PD_MSMT_ACTIVE <input type="checkbox"/> attribute-type = FLOAT-Type (INT-U32) <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> l. IF Not Recommended attribute Accuracy is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_ACCUR_MSMT <input type="checkbox"/> attribute-type = FLOAT-Type (INT-U32) <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test>
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/AG/CLASS/ECG/BV-007			
TP label	R-R Interval Object for Extended Configuration			
Coverage	Spec	[ISO/IEEE 11073-10406]		
	Testable items	RRInterval1; M	RRInterval2; M	RRInterval3; R
		RRInterval4; M	RRInterval5; R	RRInterval6; O
		RRInterval7; R	RRInterval8; R	RRInterval9; R
		RRInterval10; M	RRInterval11; C	RRInterval12; R
		RRInterval13; O	RRInterval14; O	RRInterval15; C
		RRInterval16; C	RRInterval17; C	RRInterval18; C
		RRInterval19; R	RRInterval20; C	RRInterval21; C
		RRInterval22; C	RRInterval23; C	RRInterval24; C
		RRInterval25; C	RRInterval26; R	RRInterval27; M
Test purpose	Check that: R-R Interval Numeric Object contains the attributes specified for Extended Configuration			

Applicability	(C_AG_OXP_164 OR C_AG_OXP_165) AND C_AG_ECG_002 AND C_AG_OXP_181 AND C_AG_OXP_000
Other PICS	C_AG_OXP_041, C_AG_OXP_046, C_AG_OXP_047, C_AG_OXP_183, C_AG_OXP_189
Initial condition	The simulated manager and the agent under test are in the unassociated state.
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent under test responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 4. Check that the field Dev-Config-Id is set to the tested extended configuration. If it is not, the manager responds with an "unsupported-config" and waits for a new configuration. Repeat this step until a Dev-config-Id equal to the extended configuration is received. 5. Once the agent under test sends the tested configuration, check the R-R Interval object. 6. The R-R Interval object contents shall be: <ol style="list-style-type: none"> a. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = 0x00 0x02 (MDC_PART_SCADA), 0x3F 0x28 (MDC_ECG_TIME_PD_RR_GL 16168) b. IF Not Recommended attribute Supplemental-Types <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SPPLEMENTAL_TYPES <input type="checkbox"/> attribute-type = SupplementalTypeList <input type="checkbox"/> attribute-value.length = <variable>Sequence of TYPE (TYPE.length= 4 bytes) <input type="checkbox"/> attribute-value = <Not relevant for this test> c. Mandatory attribute Metric-Spec-Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <ul style="list-style-type: none"> • Bit 3 (mss-msmt-aperiodic) is set • Bit 5 (mss-msmt-btb-metric) is set d. IF Not recommended attribute Metric-Structure-Small is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_STRUCTURE_SMALL <input type="checkbox"/> attribute-type = MetricStructureSmall <input type="checkbox"/> attribute-length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> e. IF Optional attribute Measurement-Status is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_MSMT_STAT <input type="checkbox"/> attribute-type = MeasurementStatus <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> f. IF Not recommended attribute Metric-Id is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_PHYSIO

	<ul style="list-style-type: none"> <input type="checkbox"/> attribute-type = OID-Type(INT-U16) <input type="checkbox"/> attribute-value.length =2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> g. IF Not Recommended attribute Metric-Id-List is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_PHYSIO_LIS <input type="checkbox"/> attribute-type = MetricIdList <input type="checkbox"/> attribute-value = <Not relevant for this test> h. IF Not recommended attribute Metric-Id-Partition is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_ID_PART <input type="checkbox"/> attribute-type = NomPartition(INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> i. Mandatory attribute Unit-Code <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE <input type="checkbox"/> attribute-type = OID-Type <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_DIM_MILLI_SEC or MDC_DIM_TICK j. IF Not recommended attribute Source-Handle-Reference is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SOURCE_HANDLE_REF <input type="checkbox"/> attribute-type = HANDLE(INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> k. IF Not recommended attribute Measure-Active-Period <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_PD_MSMT_ACTIVE <input type="checkbox"/> attribute-type = FLOAT-Type (INT-U32) <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> l. IF Not Recommended attribute Accuracy is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_ACCUR_MSMT <input type="checkbox"/> attribute-type = FLOAT-Type (INT-U32) <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test>
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/AG/CLASS/ECG/BV-008		
TP label	Tick-Resolution attribute for R-R Interval Tick units		
Coverage	Spec	[ISO/IEEE 11073-10406]	
	Testable items	ECG_MDSAttr9; C	RRInterval28; C

Test purpose	Check that: If the agent implements the R-R interval object and uses MDC_DIM_TICK for the corresponding Unit-Code attribute, the Tick-Resolution attribute shall be implemented.
Applicability	(C_AG_OXP_164 OR C_AG_OXP_165) AND C_AG_ECG_002 AND C_AG_OXP_181 AND C_AG_OXP_000
Other PICS	
Initial condition	The simulated manager and the agent under test are in the unassociated state.
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent under test responds with a “Remote Operation Invoke Confirmed Event Report” message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 4. Check that the field Dev-Config-Id is set to the tested extended configuration. If it is not, the manager responds with an “unsupported-config” and waits for a new configuration. Repeat this step until a Dev-config-Id equal to the extended configuration is received. 5. Once the agent under test sends the tested configuration, check the R-R Interval object: <ol style="list-style-type: none"> a. Mandatory attribute Unit-Code <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE <input type="checkbox"/> attribute-type = OID-Type <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_DIM_MILLI_SEC or MDC_DIM_TICK 6. IF the Unit-code of the R-R Interval object is MDC_DIM_TICK THEN <ol style="list-style-type: none"> i. The simulated manager issues a “roiv-cmip-get” command with the handle set to 0 (to request for an MDS object) and the attribute-id-list set to 0 to indicate all attributes. ii. The agent under test responds with a “rors-cmip-get” service message in which the attribute-list contains a list of all implemented attributes of the MDS object: <ol style="list-style-type: none"> a. Conditional attribute Tick-Resolution is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TICK_RES <input type="checkbox"/> attribute-type = FLOAT-Type (INT-U32) <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test>
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/AG/CLASS/ECG/BV-009			
TP label	ECG waveform Object for Extended Configuration			
Coverage	Spec	[ISO/IEEE 11073-10406]		
	Testable items	Waveform1; M	Waveform2; M	Waveform3; R
		Waveform4; M	Waveform5; O	Waveform6; R
		Waveform7; R	Waveform8; R	Waveform9; M

	Waveform10; C	Waveform11; R	Waveform12; O
	Waveform13; O	Waveform14; C	Waveform15; C
	Waveform16; C	Waveform17; C	Waveform18; R
	Waveform19; M	Waveform20; M	Waveform21; M
	WaveforM22; M		
Test purpose	Check that: ECG waveform RT-SA Object contains the attributes specified for Extended Configuration		
Applicability	(C_AG_OXP_164 OR C_AG_OXP_165) AND C_AG_ECG_003 AND C_AG_OXP_181 AND C_AG_OXP_000		
Other PICS	C_AG_OXP_041, C_AG_OXP_046, C_AG_OXP_047, C_AG_OXP_180, C_AG_OXP_183, C_AG_OXP_189		
Initial condition	The simulated manager and the agent under test are in the unassociated state.		
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent under test responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 4. Check that the field Dev-Config-Id is set to the tested extended configuration. If it is not, the manager responds with an "unsupported-config" and waits for a new configuration. Repeat this step until a Dev-config-Id equal to the extended configuration is received. 5. Once the agent under test sends the tested configuration, check the ECG waveform object. 6. The ECG waveform object contents shall be: <ol style="list-style-type: none"> a. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE (0x09 0x2F) <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = one of these values: <ul style="list-style-type: none"> • 0x00 0x02 (MDC_PART_SCADA), 0x01 0x00 (MDC_ECG_ELEC_POTL 256) • 0x00 0x02 (MDC_PART_SCADA), 0x01 0x01 (MDC_ECG_ELEC_POTL_I 257) • 0x00 0x02 (MDC_PART_SCADA), 0x01 0x02 (MDC_ECG_ELEC_POTL_II 258) • 0x00 0x02 (MDC_PART_SCADA), 0x01 0x3D (MDC_ECG_ELEC_POTL_III 317) • 0x00 0x02 (MDC_PART_SCADA), 0x01 0x3E (MDC_ECG_ELEC_POTL_AVR 318) • 0x00 0x02 (MDC_PART_SCADA), 0x01 0x3F (MDC_ECG_ELEC_POTL_AVL 319) • 0x00 0x02 (MDC_PART_SCADA), 0x01 0x40 (MDC_ECG_ELEC_POTL_AVF 320) • 0x00 0x02 (MDC_PART_SCADA), 0x01 0x03 (MDC_ECG_ELEC_POTL_V1 259) • 0x00 0x02 (MDC_PART_SCADA), 0x01 0x04 (MDC_ECG_ELEC_POTL_V2 260) 		

- 0x00 0x02 (MDC_PART_SCADA), 0x01 0x05 (MDC_ECG_ELEC_POTL_V3 261)
- 0x00 0x02 (MDC_PART_SCADA), 0x01 0x06 (MDC_ECG_ELEC_POTL_V4 262)
- 0x00 0x02 (MDC_PART_SCADA), 0x01 0x07 (MDC_ECG_ELEC_POTL_V5 263)
- 0x00 0x02 (MDC_PART_SCADA), 0x01 0x08 (MDC_ECG_ELEC_POTL_V6 264)

- b. If Not Recommended attribute Supplemental-Types is present
- attribute-id = MDC_ATTR_SUPPLEMENTAL_TYPES (0x0A 0x61)
 - attribute-type = SupplementalTypeList
 - attribute.value.length= Sequence of TYPE (TYPE.length= 4 bytes)
 - attribute-value = <Not relevant for this test>
- c. Mandatory attribute Metric-Spec-Small
- attribute-id = MDC_ATTR_METRIC_SPEC_SMALL (0x0A 0x46)
 - attribute-type = MetricSpecSmall (2 bytes)
 - attribute-value = 0x00 0x40
 - Bit 9 (mss-acc-agent-initiated(9)) is set
- d. IF Optional attribute Measurement-Status is present
- attribute-id = MDC_ATTR_MSMT_STAT
 - attribute-type = MeasurementStatus
 - attribute-value.length = 2 bytes
 - attribute-value = <Not relevant for this test>
- e. IF Not recommended attribute Metric-Id is present
- attribute-id = MDC_ATTR_ID_PHYSIO
 - attribute-type = OID-Type(INT-U16)
 - attribute-value.length =2 bytes
 - attribute-value = <Not relevant for this test>
- f. IF Not Recommended attribute Metric-Id-List is present
- attribute-id = MDC_ATTR_ID_PHYSIO_LIS
 - attribute-type = MetricIdList
 - attribute-value = <Not relevant for this test>
- g. IF Not recommended attribute Metric-Id-Partition is present
- attribute-id = MDC_ATTR_METRIC_ID_PART
 - attribute-type = NomPartition(INT-U16)
 - attribute-value.length = 2 bytes
 - attribute-value = <Not relevant for this test>
- h. Mandatory attribute Unit-Code
- attribute-id = MDC_ATTR_UNIT_CODE (0x09 0x96)
 - attribute-type = OID-Type
 - attribute-value.length = 2 bytes
 - attribute-value = MDC_DIM_MILLI_VOLT
- i. IF Not recommended attribute Source-Handle-Reference is present
- attribute-id = MDC_ATTR_SOURCE_HANDLE_REF
 - attribute-type = HANDLE(INT-U16)

	<ul style="list-style-type: none"> <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> j. IF Not recommended attribute Measure-Active-Period <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_PD_MSMT_ACTIVE <input type="checkbox"/> attribute-type = FLOAT-Type (INT-U32) <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> k. Mandatory attribute Sample-Period <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_PD_SAMP <input type="checkbox"/> attribute-type = RelativeTime <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant in this test> l. Mandatory attribute Scale-and-Range-Specification <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SCALE_SPECN_I8 or MDC_ATTR_SCALE_SPECN_I16 or MDC_ATTR_SCALE_SPECN_I32 <input type="checkbox"/> attribute-type = ScaleRangeSpec8 OR ScaleRangeSpec16 OR ScaleRangeSpec32 <input type="checkbox"/> attribute-value.length = 1, 2 OR 4 bytes, depending on the type <input type="checkbox"/> attribute-value = <Not relevant in this test> m. Mandatory attribute Sa-Specification <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SA_SPECN <input type="checkbox"/> attribute-type = SaSpec <input type="checkbox"/> attribute-value.length = 6 bytes <input type="checkbox"/> attribute-value = <Not relevant in this test> 7. IF the agent under sends the ECG waveforms RT-SA observations through a scanner object THEN the simulated manager enables the scanner and receives the RT-SA event reports. The attribute of interest is: <ul style="list-style-type: none"> a. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SIMP_SA_OBS_VAL ((x0A 0x48) <input type="checkbox"/> attribute-type = OCTET STRING <input type="checkbox"/> attribute-value = <length must be even> 8. IF the agent under sends the ECG waveforms RT-SA observations through a PM-Store object THEN the simulated manager sends a request for PM-Sore data (TrigSegmDataXfer) and the agent sends the RT-SA stored data. The attribute of interest is: <ul style="list-style-type: none"> a. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SIMP_SA_OBS_VAL ((x0A 0x48) <input type="checkbox"/> attribute-type = OCTET STRING <input type="checkbox"/> attribute-value = <length must be even>
Pass/Fail Criteria	All checked values are as specified in the test procedure.
Notes	MetricSpecSmall must set bit mss-acc-agent-initiated(9) to TRUE because Scanner events are agent initiated by intent in [ISO/IEEE 11073-20601A] in spite of the fact that the manager enables/disables these objects (see bugzilla #856 for further details).

TP Id		TP/PLT/AG/CLASS/ECG/BV-010		
TP label		ECG waveform data availability		
Coverage	Spec	[ISO/IEEE 11073-10406]		
	Testable items	Waveform23; M		
Test purpose		<p>Check that:</p> <p>The ECG waveform data shall be made available only through a scanner object or PM-Store object</p>		
Applicability		(C_AG_OXP_164 OR C_AG_OXP_165) AND C_AG_ECG_003 AND C_AG_OXP_181 AND C_AG_OXP_000		
Other PICS				
Initial condition		The simulated manager and the agent under test are in the unassociated state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 4. Check that the field Dev-Config-Id is set to the tested extended configuration. If it is not, the manager responds with an "unsupported-config" and waits for a new configuration. Repeat this step until a Dev-config-Id equal to the extended configuration is received. 5. Check that ECG waveform RT-SA object/s is/are present and record its/their object handle/s. 6. If the agent supports PM-Store: <ol style="list-style-type: none"> a. The simulated manager sends a Get-Segment-Info object action for the PM-Store object with SegmSelection = all-segment b. The agent issues a response (rors-cmip-confirmed-action) with the PM-Segment attributes it supports in the SegmentInfoList structure 7. Check that all ECG waveform RT-SA object/s handle/s are referenced in the Scanner or PM-Store objects: <ol style="list-style-type: none"> a. If EpiCfgScanner object (MDC_MOC_SCAN_CFG_EPI) or PeriCfgScanner (MDC_MOC_SCAN_CFG_PERI) is present <ol style="list-style-type: none"> i. IF Attribute Scan-Handle-List is supported: <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SCAN_HANDLE_LIST <input type="checkbox"/> attribute-type = HANDLEList <input type="checkbox"/> attribute-value.length = <Variable> <input type="checkbox"/> attribute-value = It must include references to ECG waveform RT-SA objects handles ii. IF attribute Scan-Handle-Attr-Val-Map is supported: <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SCAN_HANDLE_ATTR_VAL_MAP <input type="checkbox"/> attribute-type = HANDLEAttrValMap <input type="checkbox"/> attribute-value.count = N <input type="checkbox"/> attribute-value.length = <Variable> <input type="checkbox"/> attribute-value = It must include references to ECG waveform RT-SA objects handles b. If the PM-Store object (MDC_MOC_VMO_PMSTORE) is present, then check the PM-Segment-Entry-Map of each PM-Segment 8. Check the MDS event reports sent by the agent under test. 		

Pass/Fail criteria	<ul style="list-style-type: none"> In step 7.a, all ECG waveform RT-SA objects implemented by the agent under test must be referenced in the Scan-Handle-List or Scan-Handle-Attr-Val-Map attributes. In step 7.b, all ECG waveform RT-SA objects implemented by the agent under test must be referenced (through the PM-Segment-Entry-Map attribute) at least one time in the set of PM-Segments implemented by PM-Store objects. In step 8, the MDS event report sent by the agent under test must not include the ECG waveform RT-SA object observations.
Notes	

TP Id	TP/PLT/AG/CLASS/ECG/BV-011			
TP label	Device Status Object for Extended Configuration			
Coverage	Spec	[ISO/IEEE 11073-10406]		
	Testable items	DeviceStatus1; M	DeviceStatus2; M	DeviceStatus3; R
		DeviceStatus4; M	DeviceStatus5; R	DeviceStatus6; O
		DeviceStatus7; R	DeviceStatus8; R	DeviceStatus9; R
		DeviceStatus10; R	DeviceStatus11; C	DeviceStatus12; R
		DeviceStatus13; O	DeviceStatus14; O	DeviceStatus15; C
		DeviceStatus16; C	DeviceStatus17; C	DeviceStatus18; C
		DeviceStatus19; R	DeviceStatus20; R	DeviceStatus21; R
		DeviceStatus22; M	DeviceStatus23; R	DeviceStatus24; R
		DeviceStatus25; R	DeviceStatus27; O	DeviceStatus28; M
Test purpose	Check that: Device Status Enumeration Object contains the attributes specified for Extended Configuration			
Applicability	(C_AG_OXP_164 OR C_AG_OXP_165) AND C_AG_ECG_004 AND C_AG_OXP_181 AND C_AG_OXP_000			
Other PICS	C_AG_OXP_041, C_AG_OXP_046, C_AG_OXP_047, C_AG_OXP_183, C_AG_OXP_189			
Initial condition	The simulated manager and the agent under test are in the unassociated state.			
Test procedure	<ol style="list-style-type: none"> The simulated manager receives an association request from the agent under test. The simulated manager responds with a result = accepted-unknown-config. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. Check that the field Dev-Config-Id is set to extended configuration. If it is not, the manager responds with an "unsupported-config" and waits for a new configuration. Repeat this step until a Dev-config-Id equal to tested extended configuration is received. Once the agent under test sends the tested configuration, check the Device Status object. The Device Status object contents shall be: <ol style="list-style-type: none"> Mandatory attribute Type 			

- attribute-id = MDC_ATTR_ID_TYPE
- attribute-type = TYPE
- attribute-value = MDC_PART_PHD_DM, MDC_ECG_DEV_STAT
- b. IF Not Recommended attribute Supplemental-Types
 - attribute-id = MDC_ATTR_SPPLEMENTAL_TYPES
 - attribute-type = SupplementalTypeList
 - attribute-value.length = <variable>Sequence of TYPE (TYPE.length= 4 bytes)
 - attribute-value = <Not relevant for this test>
- c. IF Not recommended attribute Metric-Structure-Small is present
 - attribute-id = MDC_ATTR_METRIC_STRUCTURE_SMALL
 - attribute-type = MetricStructureSmall
 - attribute-length = 2 bytes
 - attribute-value = <Not relevant for this test>
- d. IF Optional attribute Measurement-Status is present
 - attribute-id = MDC_ATTR_MSMT_STAT
 - attribute-type = MeasurementStatus
 - attribute-value.length = 2 bytes
 - attribute-value = <Not relevant for this test>
- e. IF Not recommended attribute Metric-Id is present
 - attribute-id = MDC_ATTR_ID_PHYSIO
 - attribute-type = OID-Type(INT-U16)
 - attribute-value.length =2 bytes
 - attribute-value = <Not relevant for this test>
- f. IF Not Recommended attribute Metric-Id-List is present
 - attribute-id = MDC_ATTR_ID_PHYSIO_LIS
 - attribute-type = MetricIdList
 - attribute-value = <Not relevant for this test>
- g. IF Not recommended attribute Metric-Id-Partition is present
 - attribute-id = MDC_ATTR_METRIC_ID_PART
 - attribute-type = NomPartition(INT-U16)
 - attribute-value.length = 2 bytes
 - attribute-value = <Not relevant for this test>
- h. IF Not recommended attribute Unit-Code is present
 - attribute-id = MDC_ATTR_UNIT_CODE
 - attribute-type = OID-Type(INT-U16)
 - attribute-value.length = 2 bytes
 - attribute-value = <Not relevant for this test>
- i. IF Not recommended attribute Source-Handle-Reference is present
 - attribute-id = MDC_ATTR_SOURCE_HANDLE_REF
 - attribute-type = HANDLE(INT-U16)
 - attribute-value.length = 2 bytes
 - attribute-value = <Not relevant for this test>

- j. IF Not recommended attribute Measure-Active-Period
 - attribute-id = MDC_ATTR_TIME_PD_MSMT_ACTIVE
 - attribute-type = FLOAT-Type (INT-U32)
 - attribute-value.length = 4 bytes
 - attribute-value = <Not relevant for this test>
- k. IF Not Recommended attribute Enum-Observed-Value-Simple-OID is present
 - attribute-id= MDC_ATTR_ENUM_OBS_VAL_SIM_OID
 - attribute-type = OID-Type (INT-U16)
 - attribute-value.length = 2 bytes
 - attribute-value = <Not relevant for this test>
- l. IF Not Recommended attribute Enum-Observed-Value-Simple-Bit-Str is present
 - attribute-id= MDC_ATTR_ENUM_OBS_VAL_SIM_BIT_STR
 - attribute-type = BITS-32
 - attribute-value.length = BITS-32
 - attribute-value= <Not relevant for this test>
- m. IF Agent supports fixed or variable format MDS event report and it does not support PM-Store or Scanner THEN Mandatory attribute Enum-Observed-Value-Basic-Bit-Str is present
 - attribute-id= MDC_ATTR_ENUM_OBS_VAL_BASIC_BIT_STR
 - attribute-type = BITS-16
 - attribute-value.length = 2 bytes
 - attribute-value = One of the following bits may be active:
 - leadwire-loss(0)
 - leadsignal-loss(1)
 - leadwire-loss-first-lead(2)
 - leadsignal-loss-first-lead(3)
 - leadwire-loss-second-lead(4)
 - leadsignal-loss-second-lead(5)
 - leadwire-loss-third-lead(6)
 - leadsignal-loss-third-lead(7)
 - The rest of the bits must not be set
- n. IF Not Recommended attribute Enum-Observed-Value-Simple-Str is present
 - attribute-id= MDC_ATTR_ENUM_OBS_VAL_SIM_STR
 - attribute-type = EnumPrintableString
 - attribute-value.length = <variable>
 - attribute-value = <Not relevant for this test>
- o. IF Not Recommended attribute Enum-Observed-Value is present
 - attribute-id= MDC_ATTR_VAL_ENUM_OBS
 - attribute-type = EnumObsValue
 - attribute-value.length = <variable>
 - attribute-value = <Not relevant for this test>
- p. IF Not recommended attribute Enum-Observed-Value-Partition is present
 - attribute-id= MDC_ATTR_ENUM_OBS_VAL_PART
 - attribute-type = NomPartition (INT-U16)

	<input type="checkbox"/> attribute-value-length=2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test>
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id		TP/PLT/AG/CLASS/ECG/BV-012		
TP label		Context Data Trigger Object for Extended Configuration		
Coverage	Spec	[ISO/IEEE 11073-10406]		
	Testable items	ContextDataTrig1; M	ContextDataTrig2; M	ContextDataTrig3; R
		ContextDataTrig4; M	ContextDataTrig5; R	ContextDataTrig6; O
		ContextDataTrig7; R	ContextDataTrig8; R	ContextDataTrig9; R
		ContextDataTrig10; R	ContextDataTrig11; C	ContextDataTrig12; R
		ContextDataTrig13; O	ContextDataTrig14; O	ContextDataTrig15; C
		ContextDataTrig16; C	ContextDataTrig17; C	ContextDataTrig18; C
		ContextDataTrig19; R	ContextDataTrig20; M	ContextDataTrig21; R
		ContextDataTrig22; R	ContextDataTrig23; R	ContextDataTrig24; R
		ContextDataTrig25; R		
Test purpose		Check that: Context Data Trigger Enumeration Object contains the attributes specified for Extended Configuration		
Applicability		(C_AG_OXP_164 OR C_AG_OXP_165) AND C_AG_ECG_005 AND C_AG_OXP_181 AND C_AG_OXP_000		
Other PICS		C_AG_OXP_041, C_AG_OXP_046, C_AG_OXP_047, C_AG_OXP_183, C_AG_OXP_189		
Initial condition		The simulated manager and the agent under test are in the unassociated state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 4. Check that the field Dev-Config-Id is set to extended configuration. If it is not, the manager responds with an "unsupported-config" and waits for a new configuration. Repeat this step until a Dev-config-Id equal to the tested extended configuration is received. 5. Once the agent under test sends the tested configuration, check the Context Data Trigger object. 6. The Context Data Trigger object contents shall be: <ol style="list-style-type: none"> a. Mandatory attribute Type <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = MDC_PART_PHD_DM, MDC_ECG_EVT_CTXT_GEN 		

- b. IF Not Recommended attribute Supplemental-Types
 - attribute-id = MDC_ATTR_SPPLEMENTAL_TYPES
 - attribute-type = SupplementalTypeList
 - attribute-value.length = <variable>Sequence of TYPE (TYPE.length= 4 bytes)
 - attribute-value = <Not relevant for this test>
- c. IF Not recommended attribute Metric-Structure-Small is present
 - attribute-id = MDC_ATTR_METRIC_STRUCTURE_SMALL
 - attribute-type = MetricStructureSmall
 - attribute-length = 2 bytes
 - attribute-value = <Not relevant for this test>
- d. IF Optional attribute Measurement-Status is present
 - attribute-id = MDC_ATTR_MSMT_STAT
 - attribute-type = MeasurementStatus
 - attribute-value.length = 2 bytes
 - attribute-value = <Not relevant for this test>
- e. IF Not recommended attribute Metric-Id is present
 - attribute-id = MDC_ATTR_ID_PHYSIO
 - attribute-type = OID-Type(INT-U16)
 - attribute-value.length = 2 bytes
 - attribute-value = <Not relevant for this test>
- f. IF Not Recommended attribute Metric-Id-List is present
 - attribute-id = MDC_ATTR_ID_PHYSIO_LIS
 - attribute-type = MetricIdList
 - attribute-value = <Not relevant for this test>
- g. IF Not recommended attribute Metric-Id-Partition is present
 - attribute-id = MDC_ATTR_METRIC_ID_PART
 - attribute-type = NomPartition(INT-U16)
 - attribute-value.length = 2 bytes
 - attribute-value = <Not relevant for this test>
- h. IF Not recommended attribute Unit-Code is present
 - attribute-id = MDC_ATTR_UNIT_CODE
 - attribute-type = OID-Type(INT-U16)
 - attribute-value.length = 2 bytes
 - attribute-value = <Not relevant for this test>
- i. IF Not recommended attribute Source-Handle-Reference is present
 - attribute-id = MDC_ATTR_SOURCE_HANDLE_REF
 - attribute-type = HANDLE(INT-U16)
 - attribute-value.length = 2 bytes
 - attribute-value = <Not relevant for this test>
- j. IF Not recommended attribute Measure-Active-Period
 - attribute-id = MDC_ATTR_TIME_PD_MSMT_ACTIVE
 - attribute-type = FLOAT-Type (INT-U32)
 - attribute-value.length = 4 bytes

	<p><input type="checkbox"/> attribute-value = <Not relevant for this test></p> <p>k. IF Agent supports fixed or variable format MDS event report and it does not support PM-Store or Scanner THEN Mandatory attribute Enum-Observed-Value-Simple-OID is present</p> <p><input type="checkbox"/> attribute-id= MDC_ATTR_ENUM_OBS_VAL_SIMP_OID</p> <p><input type="checkbox"/> attribute-type = OID-Type (INT-U16)</p> <p><input type="checkbox"/> attribute-value.length = 2 bytes</p> <p><input type="checkbox"/> attribute-value = One of these values:</p> <ul style="list-style-type: none"> • MDC_ECG_EVT_CTXT_USER (21978) • MDC_ECG_EVT_CTXT_PERIODIC (21979) • MDC_ECG_EVT_CTXT_DETECTED (21980) • MDC_ECG_EVT_CTXT_EXTERNAL (21981) <p>l. IF Not Recommended attribute Enum-Observed-Value-Simple-Bit-Str is present</p> <p><input type="checkbox"/> attribute-id= MDC_ATTR_ENUM_OBS_VAL_SIMP_BIT_STR</p> <p><input type="checkbox"/> attribute-type = BITS-32</p> <p><input type="checkbox"/> attribute-value.length = BITS-32</p> <p><input type="checkbox"/> attribute-value= <Not relevant for this test></p> <p>m. IF Not Recommended attribute Enum-Observed-Value-Basic-Bit-Str is present</p> <p><input type="checkbox"/> attribute-id= MDC_ATTR_ENUM_OBS_VAL_BASIC_BIT_STR</p> <p><input type="checkbox"/> attribute-type = BITS-16</p> <p><input type="checkbox"/> attribute-value.length = 2 bytes</p> <p><input type="checkbox"/> attribute-value = <Not relevant for this test></p> <p>n. IF Not Recommended attribute Enum-Observed-Value-Simple-Str is present</p> <p><input type="checkbox"/> attribute-id= MDC_ATTR_ENUM_OBS_VAL_SIM_STR</p> <p><input type="checkbox"/> attribute-type = EnumPrintableString</p> <p><input type="checkbox"/> attribute-value.length = <variable></p> <p><input type="checkbox"/> attribute-value = <Not relevant for this test></p> <p>o. IF Not Recommended attribute Enum-Observed-Value is present</p> <p><input type="checkbox"/> attribute-id= MDC_ATTR_VAL_ENUM_OBS</p> <p><input type="checkbox"/> attribute-type = EnumObsValue</p> <p><input type="checkbox"/> attribute-value.length = <variable></p> <p><input type="checkbox"/> attribute-value = <Not relevant for this test></p> <p>p. IF Not recommended attribute Enum-Observed-Value-Partition is present</p> <p><input type="checkbox"/> attribute-id= MDC_ATTR_ENUM_OBS_VAL_PART</p> <p><input type="checkbox"/> attribute-type = NomPartition (INT-U16)</p> <p><input type="checkbox"/> attribute-value-length=2 bytes</p> <p><input type="checkbox"/> attribute-value = <Not relevant for this test></p>
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/AG/CLASS/ECG/BV-013
TP label	PM-Store Object for Basic ECG specialization Extended Configuration. Disable agent-initiated transmissions (MDS Event Reports and Scanner objects)

Coverage	Spec	[ISO/IEEE 11073-10406]		
	Testable items	ECG_PMStoreGen2; M		
Test purpose	<p>Check that:</p> <p>Any configuration with a PM-store shall disable agent-initiated transmission as well as the use of scanner objects and support manager initiated transmission of data recorded in PM-stores.</p>			
Applicability	(C_AG_OXP_164 OR C_AG_OXP_165) AND C_AG_OXP_041 AND C_AG_OXP_181 AND C_AG_OXP_000			
Other PICS				
Initial condition	The simulated manager and the agent under test are in the operating state.			
Test procedure	<ol style="list-style-type: none"> 1. Check if the agent configuration includes scanner objects. 2. The simulated manager shall send a Get request for the PM-Store object with an attribute-id-list set to 0 to indicate all PM-Store attributes. 3. The simulated manager shall send a Get-Segment-Info object action for the PM-Segment object with SegmSelection = all-segments to indicate the PM-Segments attributes of all available PM-Segments. 4. The simulated manager asks for measurement. 5. Check event reports that are sent by the agent. 			
Pass/Fail criteria	<p>In step 1, the agent configuration shall not include scanner objects.</p> <p>In step 5, the agent shall not send the data with MDS event reports.</p>			
Notes				

TP Id	TP/PLT/AG/CLASS/ECG/BV-014			
TP label	PM-Store Object for Basic ECG specialization Extended Configuration. Periodic PM-Store			
Coverage	Spec	[ISO/IEEE 11073-10406]		
	Testable items	PerPMStoreAtt4; M	PerPMStoreAtt5; M	PerPMStoreAtt8; M
		PerPMStoreAtt9; M	PerPMStoreAtt14; M	
Test purpose	<p>Check that:</p> <p>The pmsc-epi-seg-entries bit of the [PM-Store-Capab] attribute shall not be set [AND]</p> <p>The pmsc-peri-seg-entries bit of the [PM-Store-Capab] shall be set [AND]</p> <p>[Store-Capacity-Count] attribute shall be present [AND]</p> <p>[Store-Usage-Count] attribute shall be present [AND]</p> <p>[Clear-Timeout] attribute shall be present</p>			
Applicability	(C_AG_OXP_164 OR C_AG_OXP_165) AND C_AG_OXP_041 AND C_AG_OXP_188 AND C_AG_OXP_181 AND C_AG_OXP_000			
Other PICS				

Initial condition	The simulated manager and the agent under test are in the unassociated state.
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent responds with a “Remote Operation Invoke Confirmed Event Report” message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 4. Record the handle for the PM-Store objects. 5. For each PM-Store objects: <ol style="list-style-type: none"> i. The simulated manager shall send a Get request for the PM-Store object with an attribute-id-list set to 0 to indicate all PM-Store attributes. ii. The agent issues a GET response with the PM-Store attributes it supports. IF the PmStoreCapab attribute - Bit 5 (pm-sc-peri-seg-entries) is set to TRUE THEN <ol style="list-style-type: none"> a. Mandatory attribute PM-Store-Capab <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_PM_STORE_CAPAB <input type="checkbox"/> attribute-type = PmStoreCapab <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <ul style="list-style-type: none"> • Bit 4 (pm-sc-epi-seg-entries) must be set to FALSE • Bit 5 (pm-sc-peri-seg-entries) must be set to TRUE b. Mandatory attribute Storage-Capacity-Count is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_STORE_CAPAC_CNT <input type="checkbox"/> attribute-type = INT-U32 <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = See relation with next attribute c. Mandatory attribute Storage-Usage-Count is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_STORE_USAGE_CNT <input type="checkbox"/> attribute-type = INT-U32 <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = always ≤ than Storage-Cpacity-Count d. Mandatory attribute Clear-Timeout is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_CLEAR_TIMEOUT <input type="checkbox"/> attribute-type = RelativeTime <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant in this test> <p>ELSE skip the PM-Store object and check the next one</p>
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/AG/CLASS/ECG/BV-015
TP label	PM-Store Object for Basic ECG specialization Extended Configuration. Episodic PM-Store

Coverage	Spec	[ISO/IEEE 11073-10406]		
	Testable items	AperPMStoreAtt4; M	AperPMStoreAtt5; M	AperPMStoreAtt8; M
		AperPMStoreAtt9; M	AperPMStoreAtt12; R	AperPMStoreAtt14; M
Test purpose	<p>Check that:</p> <p>The pmsc-epi-seg-entries bit of the [PM-Store-Capab] attribute shall be set [AND]</p> <p>The pmsc-peri-seg-entries bit of the [PM-Store-Capab] shall not be set [AND]</p> <p>[Store-Capacity-Count] attribute shall be present [AND]</p> <p>[Store-Usage-Count] attribute shall be present [AND]</p> <p>[Sample-Period] attribute is not recommended [AND]</p> <p>[Clear-Timeout] attribute shall be present</p>			
Applicability	(C_AG_OXP_164 OR C_AG_OXP_165) AND C_AG_OXP_041 AND C_AG_OXP_187 AND C_AG_OXP_181 AND C_AG_OXP_000			
Other PICS				
Initial condition	The simulated manager and the agent under test are in the unassociated state.			
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 4. Record the handle for the PM-Store objects. 5. For each PM-Store objects: <ol style="list-style-type: none"> i. The simulated manager shall send a Get request for the PM-Store object with an attribute-id-list set to 0 to indicate all PM-Store attributes. ii. The agent issues a GET response with the PM-Store attributes it supports <p>IF PmStoreCapab attribute - Bit 4 (pm-sc-epi-seg-entries) is set to TRUE THEN</p> <ol style="list-style-type: none"> a. Mandatory attribute PM-Store-Capab <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_PM_STORE_CAPAB <input type="checkbox"/> attribute-type = PmStoreCapab <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <ul style="list-style-type: none"> • Bit 4 (pm-sc-epi-seg-entries) must be set to TRUE • Bit 5 (pm-sc-peri-seg-entries) must be set to FALSE b. Mandatory attribute Storage-Capacity-Count is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_STORE_CAPAC_CNT <input type="checkbox"/> attribute-type = INT-U32 <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = See relation with next attribute 			

	<ul style="list-style-type: none"> c. Mandatory attribute Storage-Usage-Count is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_STORE_USAGE_CNT <input type="checkbox"/> attribute-type = INT-U32 <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = always ≤ than Storage-Cpacity-Count d. Not recommended attribute Sample-Period is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_PD_SAMP <input type="checkbox"/> attribute-type = RelativeTime <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant in this test> e. Mandatory attribute Clear-Timeout is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_CLEAR_TIMEOUT <input type="checkbox"/> attribute-type = RelativeTime <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant in this test> <p>ELSE skip the PM-Store object and check the next one</p>
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/AG/CLASS/ECG/BV-016		
TP label	Periodic PM-Store for Basic ECG specialization/Simple ECG profile		
Coverage	Spec	[ISO/IEEE 11073-10406]	
	Testable items	ECG_PersStoreM1; M	
Test purpose	<p>Check that:</p> <p>An agent that supports PM-store and that has a type value set to DEV_SUB_SPEC_PROFILE_ECG shall implement the periodic PM-store object</p>		
Applicability	C_AG_OXP_165 AND C_AG_OXP_041 AND C_AG_OXP_181 AND C_AG_OXP_000		
Other PICS	C_AG_OXP_187, C_AG_OXP_188		
Initial condition	The simulated manager and the agent under test are in the unassociated state.		
Test procedure	<ol style="list-style-type: none"> 1. Check PICS C_AG_OXP_187 and C_AG_OXP_188 values. 2. The simulated manager receives an association request from the agent under test. 3. The simulated manager responds with a result = accepted-unknown-config. 4. The agent responds with a “Remote Operation Invoke Confirmed Event Report” message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 5. Record the handle for the PM-Store objects. 6. For all PM-Store objects <ol style="list-style-type: none"> i. The simulated manager shall send a Get request for the PM-Store object with an attribute-id-list set to 0 to indicate all PM-Store attributes. ii. The agent issues a GET response with the PM-Store attributes it supports: 		

	<p>a. Mandatory attribute PM-Store-Capab</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_PM_STORE_CAPAB <input type="checkbox"/> attribute-type = PmStoreCapab <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <ul style="list-style-type: none"> • Bit 4 (pm-sc-epi-seg-entries) must be set to FALSE • Bit 5 (pm-sc-eri-seg-entries) must be set to TRUE
Pass/Fail criteria	<ul style="list-style-type: none"> • In step 1, the PICS C_AG_OXP_187 is set to FALSE and the PICS C_AG_OXP_188 is set to TRUE. • In step 6, checked values of PM-Store-Capab bits are as specified in the test procedure.
Notes	

TP Id	TP/PLT/AG/CLASS/ECG/BV-017		
TP label	Mandatory Clear-Segments (all-segments) method for Basic ECG specialization		
Coverage	Spec	[ISO/IEEE 11073-10406]	
	Testable items	PMStoreObjMeth1; M	ECG_PMStoreGen3; M
Test purpose	<p>Check that:</p> <p>A Basic electrocardiograph (ECG) (1- to 3-lead ECG) agent shall support the [Clear-Segments] method with [Confirmed] mode. The agent shall support the [Clear-Segments] method by setting the pm-sc-clear-segm-by-all-sup bit for the [PM-Store-Capab] attribute [AND]</p> <p>The data held in PM-store objects may be deleted by user actions via the manager or user interface on the device, and the capacity is limited only by the agent's data storage capabilities.</p>		
Applicability	(C_AG_OXP_164 OR C_AG_OXP_165) AND C_AG_OXP_041 AND C_AG_OXP_181 AND C_AG_OXP_000		
Other PICS	C_AG_OXP_071		
Initial condition	The simulated manager and the agent under test are in the operating state and the agent has at least one PM-Segment with data stored.		
Test procedure	<ol style="list-style-type: none"> 1. Check the PICS C_AG_OXP_071 value 2. Make sure the agent under test is not taking measurements which are stored in PM-Segments. 3. The simulated manager shall send a Get request for the PM-Store object with an attribute-id-list set to 0 to indicate all PM-Store attributes. 4. The agent under test issues a GET response with the PM-Store attributes. Check the values of the PM-Store-Capab attribute. <ol style="list-style-type: none"> a. PM-Store-Capab: <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_PM_STORE_CAPAB <input type="checkbox"/> attribute-type = PmStoreCapab <input type="checkbox"/> attribute-value = At least bit pm-sc-clear-segm-all-sup(10) is set to TRUE (this bit indicates that PM-Segments in the SegmSelection data type can be cleared by segment selection –all segments) 		

	<p>5. The simulated manager sends a Clear-Segment:</p> <p>a. Data APDU</p> <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action, <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_CLEAR <input type="checkbox"/> SegmSelection = all-segments <p>6. If the agent does not protect all segments, the agent under test operation response will be:</p> <p>a. Data APDU</p> <ul style="list-style-type: none"> <input type="checkbox"/> Type = Response Confirmed Action <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_CLEAR <input type="checkbox"/> Check the invoke-id of the response is mirrored from the request. <p>7. If the agent does protect all segments, the agent under test operation response will be:</p> <p>a. Data APDU</p> <ul style="list-style-type: none"> <input type="checkbox"/> Type = Roer <input type="checkbox"/> ErrorResult = no-allowed-by-object (24) and return code shall be MDC_RET_CODE_UNKNOWN. <input type="checkbox"/> Check the invoke-id of the response is mirrored from the request
Pass/Fail criteria	<ul style="list-style-type: none"> • In step 1, the PICS C_AG_OXP_071 is set to TRUE. • In step 6, the agent must send a confirmation if the agent does not protect any segments, otherwise the agent shall send a roer message (step 7).
Notes	

TP Id	TP/PLT/AG/CLASS/ECG/BV-018		
TP label	PM-Segment Start/Stop Time attributes (Absolute or Base Offset Time) for Basic ECG specialization		
Coverage	Spec	[ISO/IEEE 11073-10406]	
	Testable items	PerPMSegObj18; M	PerPMSegObj19; M
		AperPMSegObj17; M	
Test purpose	<p>Check that:</p> <p>For each implemented periodic session PM-segment object, an agent shall either implement the [Segment-Start-Abs-Time] attribute and the [Segment-End-Abs-Time] attribute or it shall implement the [Segment-Start-BO-Time] attribute and the [Segment-End-BO-Time] attribute.</p> <p>[AND]</p> <p>If [Segment-Start-Abs-Time] and [Segment-End-Abs-Time] are used, then absolute time stamps shall be used in the entries of the PM-segment.</p> <p>[AND]</p> <p>If [Segment-Start-BO-Time] and [Segment-End-BO-Time] are used, then base-offset time stamps shall be used in the entries of the PM-segment.</p>		
Applicability	(C_AG_OXP_164 OR C_AG_OXP_165) AND C_AG_OXP_041 AND C_AG_OXP_181 AND C_AG_OXP_000		

Other PICS	C_AG_OXP_009, C_AG_OXP_014
Initial condition	The simulated manager and the agent under test are in the operating state.
Test procedure	<p>1. The simulated manager shall send a Get-Segment-Info object action for the PM-Segment object with SegmSelection = all-segments to indicate the PM-Segments attributes of all available PM-Segments.</p> <p>2. The agent issues a “rors-cmip-confirmed-action” response with the PM-Segment attributes it supports:</p> <p>IF C_AG_OXP_009 = TRUE (agent supports Absolute Time) THEN</p> <p>a. Conditional attribute Segment-Start-Abs-Time shall be present</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_START_SEG <input type="checkbox"/> attribute-type = AbsoluteTime <input type="checkbox"/> attribute-value.length = 8 bytes <input type="checkbox"/> attribute-value = <ul style="list-style-type: none"> ▪ century = ▪ year ≤ 99 ▪ month ≤ 12 ▪ day ≤ 31 ▪ hour ≤ 24 ▪ minute ≤ 60 ▪ second ≤ 60 ▪ sec-fractions ≤ 100 <p>b. Conditional attribute Segment-End-Abs-Time shall be present</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_END_SEG <input type="checkbox"/> attribute-type = AbsoluteTime <input type="checkbox"/> attribute-value.length = 8 bytes <input type="checkbox"/> attribute-value = <ul style="list-style-type: none"> ▪ century = ▪ year ≤ 99 ▪ month ≤ 12 ▪ day ≤ 31 ▪ hour ≤ 24 ▪ minute ≤ 60 ▪ second ≤ 60 ▪ sec-fractions ≤ 100 <p>c. Conditional attribute Segment-Start-BO-Time shall not be present</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_START_SEG_BO <input type="checkbox"/> attribute-type = BaseOffsetTime <input type="checkbox"/> attribute-value.length = 8 bytes <input type="checkbox"/> attribute-value = <Not relevant in this test> <p>d. Conditional attribute Segment-End-BO-Time shall not be present</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_START_SEG_BO <input type="checkbox"/> attribute-type = BaseOffsetTime <input type="checkbox"/> attribute-value.length = 8 bytes <input type="checkbox"/> attribute-value = <Not relevant in this test>

	<p>e. Mandatory attribute PM-Segment-Entry-Map shall be present</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_PM_SEG_MAP <input type="checkbox"/> attribute-type = PmSegmentEntryMap <input type="checkbox"/> attribute-value = SEQUENCE <ul style="list-style-type: none"> ▪ segm-entry-header = seg-elem-hdr-absolute-time(0) <p>AND/OR</p> <ul style="list-style-type: none"> ▪ segm-entry-elem-list = The attr-val-map of all elements of this sequence includes MDC_ATTR_TIME_STAMP_ABS attribute <p>IF C_AG_OXP_014 = TRUE (the agent supports Base Offset Time) THEN</p> <p>a. Conditional attribute Segment-Start-Abs-Time shall not be present</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_START_SEG <input type="checkbox"/> attribute-type = AbsoluteTime <input type="checkbox"/> attribute-value.length = 8 bytes <input type="checkbox"/> attribute-value = <Not relevant in this test> <p>b. Conditional attribute Segment-End-Abs-Time shall not be present</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_END_SEG <input type="checkbox"/> attribute-type = AbsoluteTime <input type="checkbox"/> attribute-value.length = 8 bytes <input type="checkbox"/> attribute-value = <Not relevant in this test> <p>c. Conditional attribute Segment-Start-BO-Time shall be present</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_START_SEG_BO <input type="checkbox"/> attribute-type = BaseOffsetTime <input type="checkbox"/> attribute-value.length = 8 bytes <input type="checkbox"/> attribute-value = <Not relevant in this test> <p>d. Conditional attribute Segment-End-BO-Time shall be present</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_START_SEG_BO <input type="checkbox"/> attribute-type = BaseOffsetTime <input type="checkbox"/> attribute-value.length = 8 bytes <input type="checkbox"/> attribute-value = <Not relevant in this test> <p>e. Mandatory attribute PM-Segment-Entry-Map shall be present</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_PM_SEG_MAP <input type="checkbox"/> attribute-type = PmSegmentEntryMap <input type="checkbox"/> attribute-value = SEQUENCE <ul style="list-style-type: none"> ▪ segm-entry-header = seg-elem-hdr-bo-time(3) <p>AND/OR</p> <ul style="list-style-type: none"> ▪ segm-entry-elem-list = The attr-val-map of all elements of this sequence includes the MDC_ATTR_TIME_STAMP_BO attribute <p>3. Repeat step 1 and 2 for every PM-Store.</p>
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id		TP/PLT/AG/CLASS/ECG/BV-019		
TP label		Segment-entry-header for Basic ECG specialization with aperiodic PM-Store objects		
Coverage	Spec	[ISO/IEEE 11073-10406]		
	Testable items	AperPMSegObj18; M		
Test purpose		<p>Check that:</p> <p>For each entry in an implemented aperiodic PM-segment object, an agent shall include one of the time formats in the segm-entry-header.</p>		
Applicability		C_AG_OXP_164 AND C_AG_OXP_041 AND C_AG_OXP_181 AND C_AG_OXP_000		
Other PICS		C_AG_OXP_009, C_AG_OXP_014		
Initial condition		The simulated manager and the agent under test are in the operating state.		
Test procedure		<p>1. For all PM-Store objects:</p> <p>i. The simulated manager shall send a Get request for the PM-Store object with an attribute-id-list set to 0 to indicate all PM-Store attributes.</p> <p>ii. The agent issues a GET response with the PM-Store attributes it supports</p> <p>a. Mandatory attribute PM-Store-Capab</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_PM_STORE_CAPAB <input type="checkbox"/> attribute-type = PmStoreCapab <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <ul style="list-style-type: none"> • Check Bit 4 (pmsc-epi-seg-entries) value <p>2. For all PM-Store objects which its PM-Store-Capab Attribute – Bit4 (pmsc-epi-seg-entries) is set to TRUE, the simulated manager sends a Get-Segment-Info object action for the PM-Segment object with SegmSelection = all-segments to indicate the PM-Segments attributes of all available PM-Segments.</p> <p>3. The agent issues a "rors-cmip-confirmed-action" response with the PM-Segment attributes it supports:</p> <p>IF C_AG_OXP_009 = TRUE (Agent supports Absolute Time) THEN</p> <p>a. Mandatory attribute PM-Segment-Entry-Map shall be present</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_PM_SEG_MAP <input type="checkbox"/> attribute-type = PmSegmentEntryMap <input type="checkbox"/> attribute-value = SEQUENCE <ul style="list-style-type: none"> ▪ segm-entry-header = seg-elem-hdr-absolute-time(0) ▪ segm-entry-elem-list = <Not relevant for this test> <p>IF C_AG_OXP_014 = TRUE (Agent supports Base Offset Time) THEN</p> <p>a. Mandatory attribute PM-Segment-Entry-Map shall be present</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_PM_SEG_MAP <input type="checkbox"/> attribute-type = PmSegmentEntryMap <input type="checkbox"/> attribute-value = SEQUENCE <ul style="list-style-type: none"> ▪ segm-entry-header = seg-elem-hdr-bo-time(3) ▪ segm-entry-elem-list = <Not relevant for this test> <p>4. Repeat step 2 and 3 for every PM-Store.</p>		

Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/AG/CLASS/ECG/BV-020		
TP label	EpiCfgScanner Object for Basic ECG specialization. Mandatory attribute Min-Reporting-Interval		
Coverage	Spec	[ISO/IEEE 11073-10406]	
	Testable items	EpiScanObjAttr8; M	
Test purpose	Check that: For [Extended-Configuration] the [Min-Reporting-Interval] attribute shall be present.		
Applicability	(C_AG_OXP_164 OR C_AG_OXP_165) AND C_AG_OXP_047 AND C_AG_OXP_181 AND C_AG_OXP_000		
Other PICS	C_AG_OXP_144, C_AG_OXP_180		
Initial condition	The simulated manager and the agent under test are in the unassociated state.		
Test procedure	<ol style="list-style-type: none"> 1. Check the PICS C_AG_OXP_144 value. 2. The simulated manager receives an association request from the agent under test. 3. The simulated manager responds with a result = accepted-unknown-config. 4. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 5. The Configurable Episodic Scanner object (ConfigReport -> ConfigObject-> AttributeList) must include the following attribute: <ol style="list-style-type: none"> a. Mandatory attribute Min-Reporting-Interval shall be present: <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SCAN_REP_PD_MIN <input type="checkbox"/> attribute-type = RelativeTime <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> 		
Pass/Fail criteria	<ul style="list-style-type: none"> • In step 1, the PICS C_AG_OXP_144 is set to TRUE. • In step 5, all Episodic Scanners included in ConfigReport must include the attribute Min-Reporting-Interval. 		
Notes			

TP Id	TP/PLT/AG/CLASS/ECG/BV-021
TP label	Operating State. Manager to Agent Maximum APDU Size

Coverage	Spec	[ISO/IEEE 11073-20601A] Optimized exchange protocol		
	Testable items	CommonCharac 3; M		
	Spec	[ISO/IEEE 11073-10406]		
	Testable items	CommChar1;M	CommChar2;M	CommChar3;M
Test purpose	<p>Check that:</p> <p>The total size of the response do not exceed of the maximum APDU size established by the specialization</p> <p>[AND]</p> <p>A basic ECG (1- to 3-lead ECG) agent implementing only this device specialization shall be capable of receiving any APDU up to a size of Nrx.</p> <p>For this standard, Nrx shall be 256 octets.</p>			
Applicability	C_AG_OXP_000 AND (C_AG_OXP_164 OR C_AG_OXP_165)			
Other PICS	C_AG_OXP_041, C_AG_OXP_100			
Initial condition	The simulated manager and the agent are in the operating state.			
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager issues a "Remote Operation Invoke Get" command with: <ol style="list-style-type: none"> a. Obj-handle set to 0 (to request for MDS object) b. attribute-id-list.count = 119 c. attribute-id-list: (MDC_ATTR_ID_MODEL, MDC_ATTR_SYS_ID, MDC_ATTR_DEV_CONFIG_ID) repeated 39 times followed by an additional MDC_ATTR_ID_MODEL 2. Check the response of the agent. 3. The simulated manager issues a "Remote Operation Invoke Get" command with the handle set to 0 (to request for MDS object) and an empty attribute-id-list to indicate all attributes. 4. Check the response of the agent. 			
Pass/Fail criteria	<ul style="list-style-type: none"> • In step 2, the agent under test may respond with a rors-cmip-get listing all the requested attributes, or with a roer message. If PICS C_AG_OXP_100 =TRUE and the agent does not respond with a rors-cmip-get message, it responds with a roer message or rorej(resource-limitation) message, a WARNING will appear. <ul style="list-style-type: none"> ○ If the response is a get response, the total size of the response cannot exceed the sum of the APDU sizes of the supported specializations (limited to an absolute limit of 64 512 octets): <ul style="list-style-type: none"> ▪ Pulse oximeter -> 9216 octets ▪ Weighing scales -> 896 octets ▪ Glucose meter -> 5120 octets or 64 512 octets if the agent supports PM-Store ▪ Blood pressure -> 896 octets ▪ Thermometer -> 896 octets ▪ Independent activity hub -> 5120 octets ▪ Cardiovascular -> 64 512 octets or 6624 octets if the agent under test only supports Step Counter Profile ▪ Strength -> 64 512 octets: ▪ Adherence monitor -> 1024 octets ▪ Peak flow -> 2030 octets ▪ Body composition analyser -> 7730 octets 			

	<ul style="list-style-type: none"> ▪ Basic ECG/Simple ECG -> 7168 octets or 64512 octets if agent supports PM-Store ▪ Basic ECG/Heart rate -> 1280 octets or 64 512 octets if the agent supports PM-Store ▪ International normalized ratio -> 896 octets or 64 512 if the agent supports PM-Store ○ In the case where it responds with a roer, the reason must not be protocol-violation (23). • In step 4, the agent must respond with a rors-cmip-get message.
Notes	

TP Id		TP/PLT/AG/CLASS/ECG/BV-022		
TP label		Association Basic ECG Agent		
Coverage	Spec	[ISO/IEEE 11073-10406]		
	Testable items	AgProcAsReq1; M	AgProcAsReq2; M	AgProcAsReq3; M
		AgProcAsReq4; M	AgProcAsReq5; O	AgProcAsReq8; M
		AgProcAsReq9; M	AgProcAsReq10; M	AgProcAsReq11; M
		AgProcAsReq12; M	AgProcAsReq13; M	AgProcAsReq14; M
		AgProcAsReq15; M	ECG_MDSMethod7; M	
Test purpose		<p>Check that:</p> <p>During the association procedure, Body composition analyzer Agent sends the correct association request to the simulated Manager</p>		
Applicability		(C_AG_OXP_164 OR C_AG_OXP_165) AND C_AG_OXP_000		
Other PICS		C_AG_OXP_002, C_AG_OXP_017		
Initial condition		The simulated manager and the agent under test are in the unassociated state.		
Test procedure		<ol style="list-style-type: none"> 1. The agent sends a message to associate with the simulated manager, the expected fields sent by the agent are: <ol style="list-style-type: none"> a. APDU Type <ul style="list-style-type: none"> <input type="checkbox"/> field- type = AarqApdu <input type="checkbox"/> field-length =2 bytes <input type="checkbox"/> field-value =0xE2 0x00. b. assoc-version <ul style="list-style-type: none"> <input type="checkbox"/> field- type = AssociationVersion <input type="checkbox"/> field-length =BITS-32 <input type="checkbox"/> field- value=0x80 0x00 0x00 0x00 c. data-protoid <ul style="list-style-type: none"> <input type="checkbox"/> field- type = DataProtoid(INT-U16) <input type="checkbox"/> field-length =2 bytes <input type="checkbox"/> field- value=0x50 0x79 (20601) d. protocol-version <ul style="list-style-type: none"> <input type="checkbox"/> field- type = Protocol Version 		

- field-length = 4 bytes
- field- value= At least bit protocol-version2(1) is set to 1 (0x40 0x00 0x00 0x00 OR 0xC0 0x00 0x00 0x00)
- e. encoding rules
 - field- type = EncodingRules
 - field-length = 2 bytes
 - field- value=
 - Bit 0 must be set (support MDER)
 - Bits 1 and 2 may be set
 - The rest of the bits must be 0
- f. nomenclature-version
 - field- type = NomenclatureVersion
 - field-length = 4 bytes
 - field- value=0x80 0x00 0x00 0x00
 - This value indicates version1 is supported (nom-version1(0) is set).
- g. functional-units
 - field- type = FunctionalUnits
 - field-length = 4 bytes
 - field-value =
 - Bit 0 must not be set, only bit 1 or 2 may be set to 1.
- h. System type
 - field- type = SystemType
 - field-length = 4 bytes
 - field- value = 0x00 0x80 0x00 0x00 (sys-type-agent)
- i. System-Id
 - field- type = OCTET STRING
 - field-length = 8 bytes
 - field- value = 0xXX 0xXX 0xXX 0xXX 0xXX 0xXX 0xXX 0xXX (octet string length = 8 | EUI-64 manufacturer and device)
 - This value will be the System Id attribute of an MDS object and the received value will be compared with the value defined in PIXIT I_AG_OXP_001 and I_AG_OXP_002.
- j. dev-config-id
 - field- type = ConfigId(INT-U16)
 - field-length = 2 bytes
 - field- value =
 - <0x07D0> for Basic ECG/Heart Rate profile standard configuration
 - <between 0x40 0x00 and 0x7F 0xFF > for extended configuration.
- k. data-req-mode-flags (DataReqModeCapab)
 - field- type = DataReqModeFlags
 - field-length = 2 bytes
 - If Agent supports only Basic ECG specialization → Bit 15 is set (data-req-sup-init-agent(15))
- l. data-req-init-agent-count (DataReqModeCapab)
 - field- type = INT-U8
 - field-length = 2 bytes

	<ul style="list-style-type: none"> <input type="checkbox"/> field.value = 0x01 m. data-req-init-manager-count (DataReqModeCapab) <ul style="list-style-type: none"> <input type="checkbox"/> field- type = INT-U8 <input type="checkbox"/> field-length = 2 bytes <input type="checkbox"/> field.value = 0x00
Pass/Fail criteria	All checked attributes have proper values.
Notes	

TP Id		TP/PLT/AG/CLASS/ECG/BV-023		
TP label		Set Time (Absolute Time) Basic ECG Agent		
Coverage	Spec	[ISO/IEEE 11073-10406]		
	Testable items	ECG_MDSMethod2; M		
Test purpose		Check that: If the agent supports the [Absolute-Time-Stamp] attribute, the Set-Time method shall be implemented.		
Applicability		(C_AG_OXP_164 OR C_AG_OXP_165) AND C_AG_OXP_000 AND C_AG_OXP_009		
Other PICS				
Initial condition		The simulated manager and the agent under test are in the operating state.		
Test procedure		1. The simulated manager sends a SET action: <ul style="list-style-type: none"> <input type="checkbox"/> CHOICE = SetTimeInvoke <input type="checkbox"/> action-type = MDC_ACT_SET_TIME <input type="checkbox"/> the action-info-args are SetTimeInvoke <ul style="list-style-type: none"> ▪ date-time = <century, year ≤ 99 month ≤ 12 day ≤ 31 hour ≤ 24 minute ≤ 60 second ≤ 60 sec-fractions ≤ 100> ▪ accuracy = 0 2. The agent under test response shall be a rors-cmip-confirmed-action: <ul style="list-style-type: none"> <input type="checkbox"/> action-type = MDC_ACT_SET_TIME <input type="checkbox"/> action-info-args shall be empty. 		
Pass/Fail criteria		All checked values are as specified in the test procedure.		
Notes				

TP Id		TP/PLT/AG/CLASS/ECG/BV-024		
TP label		Set Time (Base Offset Time) Basic ECG Agent		
Coverage	Spec	[ISO/IEEE 11073-10406]		
	Testable items	ECG_MDSMethod4; M		

Test purpose	Check that: If the agent supports the [Base-Offset-Time-Stamp] attribute, the Set-Base-Offset-Time method shall be implemented
Applicability	(C_AG_OXP_164 OR C_AG_OXP_165) AND C_AG_OXP_000 AND C_AG_OXP_014
Other PICS	
Initial condition	The simulated manager and the agent under test are in the operating state.
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager sends a SET action: <ul style="list-style-type: none"> <input type="checkbox"/> CHOICE = SetBOTimeInvoke <input type="checkbox"/> action-type = MDC_ACT_SET_BO_TIME <input type="checkbox"/> the action-info-args are SetBOTimeInvoke <ul style="list-style-type: none"> ▪ date-time = bo-seconds = 0x00 0x00 0x00 0x00, bo-fractions = 0x00 0x00, bo-time-offset = 0x3C 2. The agent under test response shall be a rors-cmip-confirmed-action: <ul style="list-style-type: none"> <input type="checkbox"/> action-type = MDC_ACT_SET_BO_TIME <input type="checkbox"/> action-info-args shall be empty.
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

Bibliography

- [b-ITU-T H.810 (2013)] Recommendation ITU-T H.810 (2013), *Interoperability design guidelines for personal health systems*.
- [b-CDG 1.0] Continua Health Alliance, Continua Design Guidelines v1.0 (2008), *Continua Design Guidelines*.
- [b-CDG 2010] Continua Health Alliance, Continua Design Guidelines v1.5 (2010), *Continua Design Guidelines*.
- [b-CDG 2011] Continua Health Alliance, Continua Design Guidelines (2011), "Adrenaline", *Continua Design Guidelines*.
- [b-CDG 2012] Continua Health Alliance, Continua Design Guidelines (2012), "Catalyst", *Continua Design Guidelines*.
- [b-ETSI SR 001 262] ETSI SR 001 262 v1.8.1 (2003-12): *ETSI drafting rules*.

SERIES OF ITU-T RECOMMENDATIONS

Series A	Organization of the work of ITU-T
Series D	General tariff principles
Series E	Overall network operation, telephone service, service operation and human factors
Series F	Non-telephone telecommunication services
Series G	Transmission systems and media, digital systems and networks
Series H	Audiovisual and multimedia systems
Series I	Integrated services digital network
Series J	Cable networks and transmission of television, sound programme and other multimedia signals
Series K	Protection against interference
Series L	Environment and ICTs, climate change, e-waste, energy efficiency; construction, installation and protection of cables and other elements of outside plant
Series M	Telecommunication management, including TMN and network maintenance
Series N	Maintenance: international sound programme and television transmission circuits
Series O	Specifications of measuring equipment
Series P	Terminals and subjective and objective assessment methods
Series Q	Switching and signalling
Series R	Telegraph transmission
Series S	Telegraph services terminal equipment
Series T	Terminals for telematic services
Series U	Telegraph switching
Series V	Data communication over the telephone network
Series X	Data networks, open system communications and security
Series Y	Global information infrastructure, Internet protocol aspects and next-generation networks, Internet of Things and smart cities
Series Z	Languages and general software aspects for telecommunication systems