

International Telecommunication Union

ITU-T

TELECOMMUNICATION
STANDARDIZATION SECTOR
OF ITU

H.845.6

(07/2016)

SERIES H: AUDIOVISUAL AND MULTIMEDIA SYSTEMS

E-health multimedia services and applications –
Interoperability compliance testing of personal health
systems (HRN, PAN, LAN, TAN and WAN)

**Conformance of ITU-T H.810 personal health
devices: PAN/LAN/TAN interface Part 5F:
Cardiovascular fitness and activity monitor:
Agent**

Recommendation ITU-T H.845.6



ITU-T H-SERIES RECOMMENDATIONS
AUDIOVISUAL AND MULTIMEDIA SYSTEMS

CHARACTERISTICS OF VISUAL TELEPHONE SYSTEMS	H.100–H.199
INFRASTRUCTURE OF AUDIOVISUAL SERVICES	
General	H.200–H.219
Transmission multiplexing and synchronization	H.220–H.229
Systems aspects	H.230–H.239
Communication procedures	H.240–H.259
Coding of moving video	H.260–H.279
Related systems aspects	H.280–H.299
Systems and terminal equipment for audiovisual services	H.300–H.349
Directory services architecture for audiovisual and multimedia services	H.350–H.359
Quality of service architecture for audiovisual and multimedia services	H.360–H.369
Telepresence	H.420–H.429
Supplementary services for multimedia	H.450–H.499
MOBILITY AND COLLABORATION PROCEDURES	
Overview of Mobility and Collaboration, definitions, protocols and procedures	H.500–H.509
Mobility for H-Series multimedia systems and services	H.510–H.519
Mobile multimedia collaboration applications and services	H.520–H.529
Security for mobile multimedia systems and services	H.530–H.539
Security for mobile multimedia collaboration applications and services	H.540–H.549
Mobility interworking procedures	H.550–H.559
Mobile multimedia collaboration inter-working procedures	H.560–H.569
BROADBAND, TRIPLE-PLAY AND ADVANCED MULTIMEDIA SERVICES	
Broadband multimedia services over VDSL	H.610–H.619
Advanced multimedia services and applications	H.620–H.629
Ubiquitous sensor network applications and Internet of Things	H.640–H.649
IPTV MULTIMEDIA SERVICES AND APPLICATIONS FOR IPTV	
General aspects	H.700–H.719
IPTV terminal devices	H.720–H.729
IPTV middleware	H.730–H.739
IPTV application event handling	H.740–H.749
IPTV metadata	H.750–H.759
IPTV multimedia application frameworks	H.760–H.769
IPTV service discovery up to consumption	H.770–H.779
Digital Signage	H.780–H.789
E-HEALTH MULTIMEDIA SERVICES AND APPLICATIONS	
Personal health systems	H.810–H.819
Interoperability compliance testing of personal health systems (HRN, PAN, LAN, TAN and WAN)	H.820–H.859
Multimedia e-health data exchange services	H.860–H.869

For further details, please refer to the list of ITU-T Recommendations.

Recommendation ITU-T H.845.6

Conformance of ITU-T H.810 personal health devices: PAN/LAN/TAN interface Part 5F: Cardiovascular fitness and activity monitor: Agent

Summary

Recommendation ITU-T H.845.6 is a transposition of Continua Test Tool DG2013, Test Suite Structure & Test Purposes, PAN-LAN-TAN Interface; Part 5F: Device Specializations. Agent (Cardiovascular) (Version 1.4, 2014-01-24), that was developed by the Continua Health Alliance. A number of versions of this specification existed before transposition.

This Recommendation includes an electronic attachment with the protocol implementation conformance statements (PICS) and the protocol implementation extra information for testing (PIXIT) required for the implementation of Annex A.

History

Edition	Recommendation	Approval	Study Group	Unique ID*
1.0	ITU-T H.845.6	2015-01-13	16	11.1002/1000/12267
2.0	ITU-T H.845.6	2016-07-14	16	11.1002/1000/12943

Keywords

Conformance testing, continua design guidelines, e-health, H.810, WAN interface, personal area network, personal connected health devices, touch area network.

* To access the Recommendation, type the URL <http://handle.itu.int/> in the address field of your web browser, followed by the Recommendation's unique ID. For example, <http://handle.itu.int/11.1002/1000/11830-en>.

FOREWORD

The International Telecommunication Union (ITU) is the United Nations specialized agency in the field of telecommunications, information and communication technologies (ICTs). The ITU Telecommunication Standardization Sector (ITU-T) is a permanent organ of ITU. ITU-T is responsible for studying technical, operating and tariff questions and issuing Recommendations on them with a view to standardizing telecommunications on a worldwide basis.

The World Telecommunication Standardization Assembly (WTSA), which meets every four years, establishes the topics for study by the ITU-T study groups which, in turn, produce Recommendations on these topics.

The approval of ITU-T Recommendations is covered by the procedure laid down in WTSA Resolution 1.

In some areas of information technology which fall within ITU-T's purview, the necessary standards are prepared on a collaborative basis with ISO and IEC.

NOTE

In this Recommendation, the expression "Administration" is used for conciseness to indicate both a telecommunication administration and a recognized operating agency.

Compliance with this Recommendation is voluntary. However, the Recommendation may contain certain mandatory provisions (to ensure, e.g., interoperability or applicability) and compliance with the Recommendation is achieved when all of these mandatory provisions are met. The words "shall" or some other obligatory language such as "must" and the negative equivalents are used to express requirements. The use of such words does not suggest that compliance with the Recommendation is required of any party.

INTELLECTUAL PROPERTY RIGHTS

ITU draws attention to the possibility that the practice or implementation of this Recommendation may involve the use of a claimed Intellectual Property Right. ITU takes no position concerning the evidence, validity or applicability of claimed Intellectual Property Rights, whether asserted by ITU members or others outside of the Recommendation development process.

As of the date of approval of this Recommendation, ITU had not received notice of intellectual property, protected by patents, which may be required to implement this Recommendation. However, implementers are cautioned that this may not represent the latest information and are therefore strongly urged to consult the TSB patent database at <http://www.itu.int/ITU-T/ipr/>.

© ITU 2016

All rights reserved. No part of this publication may be reproduced, by any means whatsoever, without the prior written permission of ITU.

Table of Contents

	Page
1 Scope.....	1
2 References.....	2
3 Definitions	2
3.1 Terms defined elsewhere	2
3.2 Terms defined in this Recommendation.....	2
4 Abbreviations and acronyms	2
5 Conventions	3
6 Test suite structure (TSS)	4
7 Electronic attachment	7
Annex A – Test purposes	8
A.1 TP definition conventions.....	8
A.2 Subgroup 1.3.6: Cardiovascular (CV).....	9
Bibliography.....	76

Electronic attachment: This Recommendation includes an electronic attachment with the protocol implementation conformance statements (PICS) and the protocol implementation extra information for testing (PIXIT) required for the implementation of Annex A.

Introduction

This Recommendation is a transposition of Continua Test Tool DG2013, Test Suite Structure & Test Purposes, PAN-LAN-TAN Interface; Part 5F: Device Specializations. Agent (Cardiovascular) (Version 1.4, 2014-01-24), that was developed by the Continua Health Alliance. A number of versions of this specification existed before transposition and these can be found in the table below.

Version	Date	Revision history
1.2	2012-10-05	Initial release for Test Tool DG2011. This is the same version as "TSS&TP_1.5_PAN-LAN_PART_5F_v1.2.doc" because new features included in [b-CDG 2011] do not affect the test procedures specified in this document.
1.3	2013-05-24	Initial release for Test Tool DG2012. This uses "TSS&TP_DG2011_PAN-LAN_PART_5F_v1.2.doc" as a baseline and adds new features included in [b-CDG 2012]: Max APDU size for GM, BCA and ECG.
1.4	2014-01-24	Initial release for Test Tool DG2013. This uses "TSS&TP_DG2012_PAN-LAN_PART_5F_v1.4.doc" as a baseline and adds new features included in [ITU-T H.810 (2015)]: <ul style="list-style-type: none">• Adds glucose meter BLE• Adds BLE SSP support• Adds NFC new transport• Adds INR device specialization

Recommendation ITU-T H.845.6

Conformance of ITU-T H.810 personal health devices: PAN/LAN/TAN interface Part 5F: Cardiovascular fitness and activity monitor: Agent

1 Scope

The scope of this Recommendation¹ is to provide a test suite structure and the test purposes (TSS & TP) for the PAN/LAN/TAN interface based on the requirements defined in the Continua Design Guidelines (CDG) [ITU-T H.810 (2015)]. The objective of this test specification is to provide a high probability of air interface interoperability between different devices.

The TSS and TP for the PAN/LAN/TAN interface document have been divided into ten parts. Each part is listed below:

- **Part 1:** Optimized exchange protocol [ISO/IEEE 11073-20601A] Agent
- **Part 2:** Optimized exchange protocol [ISO/IEEE 11073-20601A] Manager
- **Part 3:** Continua design guidelines. Agent
- **Part 4:** Continua design guidelines. Manager
- **Part 5:** Device specializations. Agent. This document is divided in 14 subparts:
 - **Part 5A:** Weighing scales
 - **Part 5B:** Glucose Meter
 - **Part 5C:** Pulse oximeter
 - **Part 5D:** Blood pressure monitor
 - **Part 5E:** Thermometer
 - **Part 5F:** Cardiovascular fitness and activity monitor
 - **Part 5G:** Strength fitness equipment
 - **Part 5H:** Independent living activity hub
 - **Part 5I:** Adherence monitor
 - **Part 5J:** Insulin Pump (Future development)
 - **Part 5K:** Peak flow
 - **Part 5L:** Body composition analyser
 - **Part 5M:** Basic electrocardiograph
 - **Part 5N:** International normalized ratio monitor
- **Part 6:** Device specializations. Manager
- **Part 7:** Continua design guidelines. Agent BLE
- **Part 8:** Continua design guidelines. Manager BLE
- **Part 9:** Personal health devices transcoding white paper. Agent
- **Part 10:** Personal health devices transcoding white paper. Manager

¹ This Recommendation includes an electronic attachment with the protocol implementation conformance statements (PICS) and the protocol implementation extra information for testing (PIXIT) required for the implementation of Annex A.

2 References

The following ITU-T Recommendations and other references contain provisions which, through reference in this text, constitute provisions of this Recommendation. At the time of publication, the editions indicated were valid. All Recommendations and other references are subject to revision; users of this Recommendation are therefore encouraged to investigate the possibility of applying the most recent edition of the Recommendations and other references listed below. A list of the currently valid ITU-T Recommendations is regularly published. The reference to a document within this Recommendation does not give it, as a stand-alone document, the status of a Recommendation.

- [ITU-T H.810 (2015)] Recommendation ITU-T H.810 (2015), *Interoperability design guidelines for personal health systems*.
- [ITU-T H.810 (2016)] Recommendation ITU-T H.810 (2016), *Interoperability design guidelines for personal health systems*.
- [ISO/IEEE 11073-20601A] ISO/IEEE 11073-20601:2010, *Health informatics – Personal health device communication – Part 20601: Application profile – Optimized exchange protocol*, including ISO/IEEE 11073-20601:2010 Amd 1:2015.
<http://www.iso.org/iso/home/store/catalogue_tc/catalogue_detail.htm?csnumber=54331>
with
<http://www.iso.org/iso/home/store/catalogue_tc/catalogue_detail.htm?csnumber=63972>
- [ISO/IEEE 11073-104xx] ISO/IEEE 11073-104xx (in force), *Health informatics – Personal health device communication – Device specialization*.
NOTE – This is shorthand used to refer to the collection of device specialization standards that utilize [ISO/IEEE 11073-20601A], where xx can be any number from 01 to 99, inclusive.
- [ISO/IEEE 11073-10441] ISO/IEEE 11073-10441-2008, *Health informatics – Personal health device communication – Device specialization – Cardiovascular fitness and activity monitor*.

3 Definitions

3.1 Terms defined elsewhere

This Recommendation uses the following terms defined elsewhere:

3.1.1 agent [ISO/IEEE 11073-20601A]: A node that collects and transmits personal health data to an associated manager.

3.1.2 manager [ISO/IEEE 11073-20601A]: A node receiving data from one or more agent systems. Some examples of managers include a cellular phone, health appliance, set top box, or a computer system.

3.2 Terms defined in this Recommendation

None.

4 Abbreviations and acronyms

This Recommendation uses the following abbreviations and acronyms:

ATS	Abstract Test Suite
DUT	Device Under Test
CDG	Continua Design Guidelines

GUI	Graphical User Interface
INR	International Normalized Ratio
IUT	Implementation Under Test
MDS	Medical Device System
NFC	Near Field Communication
PAN	Personal Area Network
PCT	Protocol Conformance Testing
PCO	Point of Control and Observation
PHD	Personal Healthcare Device
PHDC	Personal Healthcare Device Class
PHM	Personal Health Manager
PICS	Protocol Implementation Conformance Statement
PIXIT	Protocol Implementation extra Information for Testing
SABTE	Sleep Apnoea Breathing Therapy Equipment
SDP	Service Discovery Protocol
SOAP	Simple Object Access Protocol
TCWG	Test and Certification Working Group
TP	Test Purpose
TSS	Test Suite Structure
USB	Universal Serial Bus
WDM	Windows Driver Model

5 Conventions

The key words "SHALL", "SHALL NOT", "SHOULD", "SHOULD NOT", "MAY", "MAY NOT" in this Recommendation are to be interpreted as in [b-ETSI SR 001 262].

- SHALL is equivalent to 'must' or 'it is required to'.
- SHALL NOT is equivalent to 'must not' or 'it is not allowed'.
- SHOULD is equivalent to 'it is recommended to'.
- SHOULD NOT is equivalent to 'it is not recommended to'.
- MAY is equivalent to 'is permitted'.
- MAY NOT is equivalent to 'it is not required that'.

NOTE – The above-mentioned key words are capitalized for illustrative purposes only and they do not appear capitalized within this Recommendation.

Reference is made in the ITU-T H.800-series of Recommendations to different versions of the Continua Design Guidelines (CDG) by a specific designation. The list of terms that may be used in this Recommendation is provided in Table 1.

Table 1 – List of designations associated with the various versions of the CDG

CDG name	Transposed as	Version	Description	Designation
2016 plus errata	[ITU-T H.810 (2016)]	6.1	Release 2016 plus errata noting all ratified bugs [ITU-T H.810 (2016)].	–
2016	–	6.0	Release 2016 of the CDG including maintenance updates of the CDG 2015 and additional guidelines that cover new functionalities.	Iris
2015 plus errata	[ITU-T H.810 (2015)]	5.1	Release 2015 plus errata noting all ratified bugs [ITU-T H.810 (2015)].	–
2015	–	5.0	Release 2015 of the CDG including maintenance updates of the CDG 2013 and additional guidelines that cover new functionalities.	Genome
2013 plus errata	[ITU-T H.810 (2013)]	4.1	Release 2013 plus errata noting all ratified bugs [b-ITU-T H.810 (2013)].	–
2013	–	4.0	Release 2013 of the CDG including maintenance updates of the CDG 2012 and additional guidelines that cover new functionalities.	Endorphin
2012 plus errata	–	3.1	Release 2012 plus errata noting all ratified bugs [b-CDG 2012].	–
2012	–	3.0	Release 2012 of the CDG including maintenance updates of the CDG 2011 and additional guidelines that cover new functionalities.	Catalyst
2011 plus errata	–	2.1	CDG 2011 integrated with identified errata.	–
2011	–	2.0	Release 2011 of the CDG including maintenance updates of the CDG 2010 and additional guidelines that cover new functionalities [b-CDG 2011].	Adrenaline
2010 plus errata	–	1.6	CDG 2010 integrated with identified errata	–
2010	–	1.5	Release 2010 of the CDG with maintenance updates of the CDG Version 1 and additional guidelines that cover new functionalities [b-CDG 2010].	1.5
1.0	–	1.0	First released version of the CDG [b-CDG 1.0].	–

6 Test suite structure (TSS)

The test purposes (TPs) for the PAN/LAN/TAN interface have been divided into the main subgroups specified below. Annex A describes the TPs for subgroup 1.3.6 (shown in bold).

- Group 1: Agent (AG)
 - Group 1.1: Transport (TR)
 - Subgroup 1.1.1: Design guidelines: Common (DGC)
 - Subgroup 1.1.2: USB design guidelines (UDG)

- Subgroup 1.1.3: Bluetooth design guidelines (BDG)
- Subgroup 1.1.4: Pulse oximeter design guidelines (PODG)
- Subgroup 1.1.5: Cardiovascular design guidelines (CVDG)
- Subgroup 1.1.6: Activity hub design guidelines (HUBDG)
- Subgroup 1.1.7: ZigBee design guidelines (ZDG)
- Subgroup 1.1.8: Glucose meter design guidelines (GLDG)
- Subgroup 1.1.9: Bluetooth low energy design guidelines (BLEDG)
- Subgroup 1.1.10: Basic electrocardiograph design guidelines (ECGDG)
- Subgroup 1.1.11: NFC design guidelines (NDG)
- Group 1.2: 20601: Optimized exchange protocol (OXP)
 - Subgroup 1.2.1: PHD domain information model (DIM)
 - Subgroup 1.2.2: PHD service model (SER)
 - Subgroup 1.2.3: PHD communication model (COM)
- Group 1.3: Devices class specializations (CLASS)
 - Subgroup 1.3.1: Weighing scales (WEG)
 - Subgroup 1.3.2: Glucose meter (GL)
 - Subgroup 1.3.3: Pulse oximeter (PO)
 - Subgroup 1.3.4: Blood pressure monitor (BPM)
 - Subgroup 1.3.5: Thermometer (TH)
 - **Subgroup 1.3.6: Cardiovascular (CV)**
 - Subgroup 1.3.7: Strength (ST)
 - Subgroup 1.3.8: Activity hub (HUB)
 - Subgroup 1.3.9: Adherence monitor (AM)
 - Subgroup 1.3.10: Insulin pump (IP) (Future development)
 - Subgroup 1.3.11: Peak flow (PF)
 - Subgroup 1.3.12: Body composition analyzer (BCA)
 - Subgroup 1.3.13: Basic electrocardiograph (ECG)
 - Subgroup 1.3.14: International normalized ratio (INR)
 - Subgroup 1.3.15: Sleep apnoea breathing therapy equipment (SABTE)
- Group 1.4: Personal health device transcoding whitepaper (PHDTW)
 - Subgroup 1.4.1: Whitepaper general requirements (GEN)
 - Subgroup 1.4.2: Whitepaper thermometer requirements (TH)
 - Subgroup 1.4.3: Whitepaper blood pressure requirements (BPM)
 - Subgroup 1.4.4: Whitepaper heart rate requirements (HR)
 - Subgroup 1.4.5: Whitepaper glucose meter requirements (GL)
 - Subgroup 1.4.6: Whitepaper weight scale requirements (WS)

- Group 2: Manager (MAN)
 - Group 2.1: Transport (TR)
 - Subgroup 2.1.1: Design guidelines: Common (DGC)
 - Subgroup 2.1.2: USB design guidelines (UDG)
 - Subgroup 2.1.3: Bluetooth design guidelines (BDG)
 - Subgroup 2.1.4: Cardiovascular design guidelines (CVDG)
 - Subgroup 2.1.5: Activity hub design guidelines (HUBDG)
 - Subgroup 2.1.6: ZigBee design guidelines (ZDG)
 - Subgroup 2.1.7: Bluetooth low energy design guidelines (BLEDG)
 - Subgroup 2.1.8: NFC design guidelines (NDG)
 - Group 2.2: 20601: Optimized exchange protocol (OXP)
 - Subgroup 2.2.1: General (GEN)
 - Subgroup 2.2.2: PHD domain information model (DIM)
 - Subgroup 2.2.3: PHD service model (SER)
 - Subgroup 2.2.4: PHD communication model (COM)
 - Group 2.3: Devices class specializations (CLASS)
 - Subgroup 2.3.1: Weighing scales (WEG)
 - Subgroup 2.3.2: Glucose meter (GL)
 - Subgroup 2.3.3: Pulse oximeter (PO)
 - Subgroup 2.3.4: Blood pressure monitor (BPM)
 - Subgroup 2.3.5: Thermometer (TH)
 - Subgroup 2.3.6: Cardiovascular (CV)
 - Subgroup 2.3.7: Strength (ST)
 - Subgroup 2.3.8: Activity hub (HUB)
 - Subgroup 2.3.9: Adherence monitor (AM)
 - Subgroup 2.3.10: Insulin pump (IP) (Future development)
 - Subgroup 2.3.11: Peak flow (PF)
 - Subgroup 2.3.12: Body composition analyzer (BCA)
 - Subgroup 2.3.13: Basic electrocardiograph (ECG)
 - Subgroup 2.3.14: International normalized ratio (INR)
 - Subgroup 2.3.15: Sleep apnoea breathing therapy equipment (SABTE)
 - Group 2.4: Personal health device transcoding whitepaper (PHDTW)
 - Subgroup 2.4.1: Whitepaper general requirements (GEN)
 - Subgroup 2.4.2: Whitepaper thermometer requirements (TH)
 - Subgroup 2.4.3: Whitepaper blood pressure requirements (BPM)
 - Subgroup 2.4.4: Whitepaper heart rate requirements (HR)
 - Subgroup 2.4.5: Whitepaper glucose meter requirements (GL)
 - Subgroup 2.4.6: Whitepaper weight scale requirements (WS)

7 Electronic attachment

The protocol implementation conformance statements (PICS) and the protocol implementation extra information for testing (PIXIT) required for the implementation of Annex A can be downloaded from <http://handle.itu.int/11.1002/2000/12067>

In the electronic attachment, letters "C" and "I" in the column labelled "Mandatory" are used to distinguish between "PICS" and "PIXIT" respectively during testing. If the cell is empty, the corresponding PICS is "independent". If the field contains a "C", the corresponding PICS is dependent on other PICS, and the logical expression is detailed in the "SCR_Expression" field. The static conformance review (SCR) is used in the test tool to assert whether the PICS selection is consistent.

Annex A

Test purposes

(This annex forms an integral part of this Recommendation.)

A.1 TP definition conventions

The test purposes (TPs) are defined according to the following rules:

- **TP Id:** This is a unique identifier (TP/<TT>/<DUT>/<GR>/<SGR>/<XX> – <NNN>). It is specified according to the naming convention defined below:
 - Each test purpose identifier is introduced by the prefix "TP".
 - <TT>: This is the test tool that will be used in the test case:
 - PAN: Personal area network (Bluetooth or USB)
 - LAN: Local area network (ZigBee)
 - PAN-LAN: Personal area network (Bluetooth or USB) – Local area network (ZigBee)
 - LP-PAN: Low power personal area network (Bluetooth low energy)
 - TAN: Touch area network (NFC)
 - PLT: Personal area network (Bluetooth or USB) – Local area network (ZigBee) – Touch area network (NFC)
 - <DUT>: This is the device under test:
 - AG: PAN/LAN Agent
 - MAN: PAN/LAN Manager
 - GR>: This identifies a group of test cases.
 - <SGR>: This identifies a subgroup of test cases.
 - <XX>: This identifies the type of testing:
 - BV: Valid behaviour test
 - BI: Invalid behaviour test
 - <NNN>: This is a sequential number that identifies the test purpose.
- **TP label:** This is the TP's title.
- **Coverage:** This contains the specification reference and clause to be checked by the TP
 - Spec: This indicates the earliest version of the specification from which the testable items to be checked by the TP were included.
 - Testable item: This contains testable items to be checked by the TP.
- **Test purpose:** It is a description about the requirements to be tested.
- **Applicability:** This contains the PICS items that define if the test case is applicable or not for a specific device. When a TP contains an "ALL" in this field it means that it applies to the device under test within that scope of the test (specialization, transport used, etc.).
- **Other PICS:** It contains additional PICS items (apart from the PICS specified in the Applicability row) which are used within the test case implementation and can modify the final verdict. When this row is empty, it means that only the PICS specified in the Applicability row are used within the test case implementation.
- **Initial condition:** This indicates the state to which the DUT needs to be moved at the beginning of TC execution.

- **Test procedure:** This describes the steps to be followed in order to execute the test case.
- **Pass/Fail criteria:** This provides criteria to decide whether the DUT passes or fails the test case.

A.2 Subgroup 1.3.6: Cardiovascular (CV)

TP Id		TP/PLT/AG/CLASS/CV/BV-000		
TP label		MDS Object for Cardiovascular fitness and activity monitor specialization		
Coverage	Spec	[ISO/IEEE 11073-10441]		
	Testable items	MDSAttr1; M	MDSAttr2; M	MDSAttr3; R
		MDSAttr4; R	MDSAttr5; R	MDSAttr6; M
		MDSAttr7; M	MDSAttr8; M	GETServ1; M
GETServ3; M		OperProc1; M		
Test purpose		<p>Check that:</p> <p>The MDS Object contains the attributes specified for a Cardiovascular Agent</p>		
Applicability		C_AG_OXP_000 AND C_AG_OXP_172		
Other PICS		C_AG_OXP_181		
Initial condition		The agent under test is in the operating state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated manager issues a "roiv-cmip-get" command with the handle set to 0 (to request an MDS object) and the attribute-id-list is set to 0 to indicate all attributes. 2. The agent responds with a "rors-cmip-get" service message in which the attribute-list contains a list of all implemented attributes of the MDS object: <ol style="list-style-type: none"> a. Mandatory attribute Dev-Configuration-Id <ul style="list-style-type: none"> <input type="checkbox"/> attribute-type = ConfigId <input type="checkbox"/> attribute-length = 2 bytes <input type="checkbox"/> attribute-value = < between 0x4000 and 0x7FFF > b. Attribute System-Type shall not be present. c. Mandatory attribute System-model <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_MODEL (0x09 0x28) <input type="checkbox"/> attribute-type = SystemModel <input type="checkbox"/> attribute-value.length = <Variable> <input type="checkbox"/> attribute-value = {Manufacturer, Model} d. Mandatory attribute System-Type-Spec-List <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SYS_TYPE_SPEC_LIST <input type="checkbox"/> attribute-type = TypeVerList <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = { MDC_DEV_SPEC_PROFILE_HF_CARDIO (0x10 0x29), 1} e. If Recommended Power-Status attribute is present: <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_POWER_STAT <input type="checkbox"/> attribute-type = PowerStatus <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <ul style="list-style-type: none"> ON_MAINS (0x8000) or ON_BATTERY(0x4000) • Only one of the following may be active: • chargingFull(8), 		

	<ul style="list-style-type: none"> • chargingTrickle(9), • chargingOff(10). <p>f. If Recommended Battery-Level attribute is present</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_VAL_BATT_CHARGE <input type="checkbox"/> attribute-type = BITS-16 <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <value between 0 and 100> If value >100, the meaning of the value is "undefined" <p>g. If Recommended Remaining-Battery-Time attribute is present:</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_BATT_REMAIN <input type="checkbox"/> attribute-type = BatMeasure <input type="checkbox"/> attribute-value.length = 6 bytes <input type="checkbox"/> attribute-value = <4 bytes to define the value. 2 remaining bytes to define the units, which shall be set to one of: MDC_DIM_MIN (0x08 0xA0), MDC_DIM_HR (0x08 0xC0), MDC_DIM_DAY (0x08 0xE0)>
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/AG/CLASS/CV/BV-001			
TP label	MDS Configuration objects events for Cardiovascular.			
Coverage	Spec	[ISO/IEEE 11073-10441]		
	Testable items	MDSEvent1; M	AltitudeGain1; O	AltitudeLoss1; O
		Altitude1; O	Distance1; O	AscentTime1; O
		DescentTime1; O	Latitude1; O	Longitude1; O
		Slopes1; O	Speed1; O	Cadence1; O
		Incline1; O	Heart rate1; O	Max user heart rate1; O
		Power1; O	Resistance1; O	Stride length1; O
		Breathing rate1; O	Energy expended1; O	Calories ingested1; O
		CarbohydrateCal1; O	SustainedPhysAct1; O	ActIntensity1; O
		BodyWeight1; O	Height1; O	Age1; O
		Session1; M	Sub-session1; O	ActivityTime1; O
	ProgramId1; O			
Test purpose	Check that: Cardiovascular Agent sends the MDS-Configuration-Event using a Confirmed event report and it includes the event-info ConfigReport			
Applicability	C_AG_OXP_000 AND C_AG_OXP_172			
Other PICS	C_AG_OXP_010, C_AG_CV_015, C_AG_CV_016, C_AG_CV_017, C_AG_CV_018, C_AG_CV_019, C_AG_CV_020, C_AG_CV_021, C_AG_CV_022, C_AG_CV_023, C_AG_CV_024, C_AG_CV_025, C_AG_CV_026, C_AG_CV_027, C_AG_CV_028, C_AG_CV_029, C_AG_CV_030, C_AG_CV_031, C_AG_CV_032, C_AG_CV_033, C_AG_CV_034, C_AG_CV_035, C_AG_CV_036, C_AG_CV_037, C_AG_CV_038, C_AG_CV_039, C_AG_CV_040, C_AG_CV_041, C_AG_CV_042, C_AG_CV_043			
Initial condition	The simulated manager and the agent under test are in the configuring state.			
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config 3. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager: 			

- a. APDU Type
 - field- type = PrstApdu
 - field-length =2 bytes
 - field-value =0xE7 0x00
- b. invoke-id
 - field- type = InvokeIDType
 - field-length =INT-U16
 - field- value=<Not relevant for this test>
- c. message
 - field- type = roiv-cmip-confirmed-event-report
 - field-length =two bytes
 - field- value=0x01 0x01 (EventReportArgumentSimple)
- d. obj-handle (EventReportArgumentSimple)
 - field- type = HANDLE
 - field-length =INT-U16
- e. event-time (EventReportArgumentSimple)
 - field- type = Relative Time
 - field-length =INT-U32
 - field-value =
 - IF NOT C_AG_OXP_010 THEN value = 0xFF 0xFF 0xFF 0xFF
- f. event-type (EventReportArgumentSimple)
 - field- type = OID-Type
 - field-length =INT-U16
 - field- value=0x 0D 0x 1C (MDC_NOTI_CONFIG)
- g. config-report-id (ConfigReport)
 - field- type = ConfigId
 - field-length = INT-U16
 - field- value = <Between 0x40 0x00 and 0x7F 0xFF>
- h. obj-class (ConfigReport → ConfigObjectList (ConfigObject))
 - field- type = OID-Type
 - field-length = INT-U16
 - field- value = Objects that will be checked:
 - The Session Enumeration Object must appear.
 - IF C_AG_CV_015 Then Altitude Gain Numeric Object is present, ELSE it is not present.
 - IF C_AG_CV_016 Then Program identifier Enumeration Object is present, ELSE it is not present.
 - IF C_AG_CV_017 Then Activity Time Enumeration Object is present, ELSE it is not present.
 - IF C_AG_CV_018 Then Age Numeric Object is present, ELSE it is not present.
 - IF C_AG_CV_019 Then Height Numeric Object is present, ELSE it is not present.
 - IF C_AG_CV_020 Then Body Weight Numeric Object is present, ELSE it is not present.
 - IF C_AG_CV_021 Then Activity Intensity Numeric Object is present, ELSE

	<p>it is not present.</p> <ul style="list-style-type: none"> • IF C_AG_CV_022 Then Sustained Phys activity threshold Numeric Object is present, ELSE it is not present. • IF C_AG_CV_023 Then Carbohydrate calories Numeric Object is present, ELSE it is not present. • IF C_AG_CV_024 Then Calories ingested Numeric Object is present, ELSE it is not present. • IF C_AG_CV_025 Then Energy Expended Numeric Object is present, ELSE it is not present. • IF C_AG_CV_026 Then Breathing Rate Numeric Object is present, ELSE it is not present. • IF C_AG_CV_027 Then Stride Length Numeric Object is present, ELSE it is not present. • IF C_AG_CV_028 Then Resistance Numeric Object is present, ELSE it is not present. • IF C_AG_CV_029 Then Power Numeric Object is present, ELSE it is not present. • IF C_AG_CV_030 Then Max User Heart Rate Numeric Object is present, ELSE it is not present. • IF C_AG_CV_031 Then Heart Rate Numeric Object is present, ELSE it is not present. • IF C_AG_CV_032 Then Altitude Loss Numeric Object is present, ELSE it is not present. • IF C_AG_CV_033 Then Incline Numeric Object is present, ELSE it is not present. • IF C_AG_CV_034 Then Cadence Numeric Object is present, ELSE it is not present. • IF C_AG_CV_035 Then Speed Numeric Object is present, ELSE it is not present. • IF C_AG_CV_036 Then Slopes Numeric Object is present, ELSE it is not present. • IF C_AG_CV_037 Then Longitude Numeric Object is present, ELSE it is not present. • IF C_AG_CV_038 Then Latitude Numeric Object is present, ELSE it is not present. • IF C_AG_CV_039 Then Altitude Numeric Object is present, ELSE it is not present. • IF C_AG_CV_040 Then Distance Numeric Object is present, ELSE it is not present. • IF C_AG_CV_041 Then Ascent time and Distance Numeric Object is present, ELSE it is not present. • IF C_AG_CV_042 Then Descent time and Distance Numeric Object is present, ELSE it is not present. • IF C_AG_CV_043 Then Sub-session Enumeration Object is present, ELSE it is not present.
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id		TP/PLT/AG/CLASS/CV/BV-002		
TP label		MDS object events for Cardiovascular fitness activity monitor agent.		
Coverage	Spec	[ISO/IEEE 11073-10441]		
	Testable items	MDSEvent3; M	MDSEvent4; M	MDSEvent5; M
		MDSEvent6; M	MDSEvent7; M	MDSEvent8; M
		MDSEvent9; M	MDSEvent10; M	
Test purpose		<p>Check that:</p> <p>Agent-initiated mode is supported for measurement data transmission and all types of event reports are used in confirmed mode</p> <p>[AND]</p> <p>The Agent sends the MDS-Dynamic-Data-Update-Fixed using a confirmed event report and it includes the event-info ScanReportInfoFixed</p> <p>[OR]</p> <p>The Agent sends the MDS-Dynamic-Data-Update-Var using a confirmed event report and it includes the event-info ScanReportInfoVar</p> <p>[OR]</p> <p>The Agent sends the MDS-Dynamic-Data-Update-MP-Fixed using a confirmed event report and it includes the event-info ScanReportInfoMPFixed</p> <p>[OR]</p> <p>The Agent sends the MDS-Dynamic-Data-Update-MP-Var using a confirmed event report and it includes the event-info ScanReportInfoMPVar</p>		
Applicability		C_AG_OXP_000 AND C_AG_OXP_172 AND (C_AG_OXP_182 OR C_AG_OXP_183 OR C_AG_OXP_184 OR C_AG_OXP_189)		
Other PICS				
Initial condition		The agent under test is in the operating state.		
Test procedure		<ol style="list-style-type: none"> 1. Take measurements for every supported object in the agent under test. 2. Wait to receive every event report and check: <ol style="list-style-type: none"> a. message <ul style="list-style-type: none"> <input type="checkbox"/> field- type = Event Report <input type="checkbox"/> field-length = 2 bytes <input type="checkbox"/> field- value=0x01 0x01 (EventReportArgumentSimple, confirmed) <p>This field identifies the type of message sent by the agent, for the confirmed event configuration, roiv-cmip-confirmed-event-report.</p>		
Pass/Fail criteria		<p>Check that every received report is a one of the following Data APDU and that it is confirmed:</p> <ul style="list-style-type: none"> • MDC_NOTI_SCAN_REPORT_FIXED • MDC_NOTI_SCAN_REPORT_MP_FIXED • MDC_NOTI_SCAN_REPORT_VAR • MDC_NOTI_SCAN_REPORT_MP_VAR 		
Notes				

TP Id		TP/PLT/AG/CLASS/CV/BV-005		
TP label		Altitude Gain Numeric Object		
Coverage	Spec	[ISO/IEEE 11073-10441]		
	Testable items	NumObj5; M	NumObj6; M	AltitudeGain1; O
		AltitudeGain2; M	AltitudeGain3; M	AltitudeGain4; M

	AltitudeGain5; M	AltitudeGain6; M	
Test purpose	Check that: The Altitude Gain Numeric object contains the attributes specified for Extended Configuration.		
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_015		
Other PICS			
Initial condition	The agent under test is in the unassociated state.		
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with an Association Response with result = "accepted-unknown-config" 3. The agent responds with a roiv-cmip-confirmed-event report message with a MDC_NOTI_CONFIG event to send its configuration to the manager. 4. The Altitude Gain object shall be: <ol style="list-style-type: none"> a. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = MDC_PART_PHD_HF MDC_HF_ALT_GAIN b. Mandatory attribute Metric-Spec_Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall (BITS-16) <input type="checkbox"/> attribute-value ≠ 0x00 0x00 <ul style="list-style-type: none"> • bit 0 (mss-avail-intermittent(0)) shall be set. • bit 1(mss-avail-stored-data(1)) shall be set. • bit 2 (mss-updt-aperiodic(2)) shall be set. • bit 3(mss-msmt-aperiodic(3)) shall be set • bit 9 (mss-acc-agent-initiated(9)) shall be set. • The other bits have to be 0. c. Mandatory attribute Unit-Code <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE <input type="checkbox"/> attribute-type = OID-Type (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_DIM_X_M or MDC_DIM_X_FOOT d. Mandatory attribute Source-Handle-Reference <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SOURCE_HANDLE_REF <input type="checkbox"/> attribute-type = HANDLE (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = It must be equal to the handle of any Session or Sub-session object in the configuration 5. Wait for the agent under test and the simulated manager to reach the operating state. 6. Take a measurement in the agent. 7. Wait until the manager receives an event report. 		
Pass/Fail criteria	<ul style="list-style-type: none"> • In step 4, all checked values are as specified. • In step 7, check that only non-negative values are used, with zero (0) indicating that no altitude was gained. 		
Notes			

TP Id		TP/PLT/AG/CLASS/CV/BV-005_A		
TP label		Altitude Gain, timestamp values		
Coverage	Spec	[ISO/IEEE 11073-10441]		
	Testable items	NumObj2; M	NumObj3; M	
Test purpose		<p>Check that:</p> <p>Altitude Gain Numeric object instance shall have a timestamp identical to its containing object instance (i.e. Session or Sub-Session).</p> <p>[AND]</p> <p>The timestamp attribute used for each object shall be the same as the one used for the associated Session or Sub-Session object instance</p>		
Applicability		C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_015		
Other PICS				
Initial condition		The simulated manager and the agent under test are in the operating state.		
Test procedure		<ol style="list-style-type: none"> 1. Take a measurement with the agent under test. 2. Wait for the simulated manager to receive it. Record the timestamp and the Measure-Active-Period of the Session and Sub-Session object and of the Altitude Gain object. 		
Pass/Fail criteria		<ul style="list-style-type: none"> • The timestamp attribute used for the Altitude Gain object shall be the same as that used for the associated Session or Sub-Session object instance. • The Altitude Gain instance shall have a timestamp identical to its associated Session or Sub-session object instance. 		
Notes				

TP Id		TP/PLT/AG/CLASS/CV/BV-006		
TP label		Altitude Loss Numeric Object		
Coverage	Spec	[ISO/IEEE 11073-10441]		
	Testable items	NumObj5; M	NumObj6; M	AltitudeLoss1; O
		AltitudeLoss2; M	AltitudeLoss3; M	AltitudeLoss4; M
Test purpose		<p>Check that:</p> <p>The Altitude Loss Numeric object contains the attributes specified for Extended Configuration.</p>		
Applicability		C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_032		
Other PICS				
Initial condition		The agent under test is in the unassociated state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with an Association Response with result = "accepted-unknown-config" 3. The agent responds with a roiv-cmip-confirmed-event report message with a MDC_NOTI_CONFIG event to send its configuration to the manager. 4. The Altitude Loss object shall be: <ol style="list-style-type: none"> a. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE 		

	<ul style="list-style-type: none"> <input type="checkbox"/> attribute-value = MDC_PART_PHD_HF MDC_HF_ALT_LOSS <p>b. Mandatory attribute Metric-Spec_Small</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall (BITS-16) <input type="checkbox"/> attribute-value ≠ 0x00 0x00 <ul style="list-style-type: none"> • bit 0 (mss-avail-intermittent(0)) shall be set. • bit 1(mss-avail-stored-data(1)) shall be set. • bit 2 (mss-updt-aperiodic(2)) shall be set. • bit 3(mss-msmt-aperiodic(3)) shall be set • bit 9 (mss-acc-agent-initiated(9)) shall be set. • The other bits have to be 0. <p>c. Mandatory attribute Unit-Code</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE <input type="checkbox"/> attribute-type = OID-Type (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_DIM_X_M or MDC_DIM_X_FOOT <p>d. Mandatory attribute Source-Handle-Reference</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SOURCE_HANDLE_REF <input type="checkbox"/> attribute-type = HANDLE (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = It must be equal to the handle of any Session or Sub-session object in the configuration <p>5. Wait for the agent under test and the simulated manager to reach the operating state.</p> <p>6. Take a measurement in the agent.</p> <p>7. Wait until the manager receives an event report.</p>
Pass/Fail criteria	<ul style="list-style-type: none"> • In step 4, all checked values are as specified. • In step 7, check that only non-negative values are used, with zero (0) indicating that no altitude was lost.
Notes	

TP Id	TP/PLT/AG/CLASS/CV/BV-006_A		
TP label	Altitude Loss, timestamp values		
Coverage	Spec	[ISO/IEEE 11073-10441]	
	Testable items	NumObj2; M	NumObj3; M
Test purpose	<p>Check that:</p> <p>Altitude Loss Numeric object instance shall have a timestamp identical to its containing object instance (i.e. Session or Sub-Session).</p> <p>[AND]</p> <p>The timestamp attribute used for each object shall be the same as the one used for the associated Session or Sub-Session object instance</p>		
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_032		
Other PICS			
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	1. Take a measurement with the agent under test.		

	2. Wait for the simulated manager to receive it. Record the timestamp and the Measure-Active-Period of the Session and Sub-Session object and of the Altitude Loss object.
Pass/Fail criteria	<ul style="list-style-type: none"> The timestamp attribute used for Altitude Loss object shall be the same as that used for the associated Session or Sub-Session object instance. The Altitude Loss instance shall have a timestamp identical to its associated Session or Sub-Session object instance.
Notes	

TP Id	TP/PLT/AG/CLASS/CV/BV-007			
TP label	Altitude Numeric Object Attributes			
Coverage	Spec	[ISO/IEEE 11073-10441]		
	Testable items	NumObj5; M	NumObj6; M	Altitude1; O
		Altitude2; M	Altitude3; M	Altitude4; M
Altitude5; M				
Test purpose	Check that: The Altitude Numeric object contains the attributes specified for Extended Configuration.			
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_039			
Other PICS				
Initial condition	The agent under test is in the unassociated state.			
Test procedure	<ol style="list-style-type: none"> The simulated manager receives an association request from the agent under test. The simulated manager responds with an Association Response with result = "accepted-unknown-config". The agent responds with a roiv-cmip-confirmed-event report message with a MDC_NOTI_CONFIG event to send its configuration to the manager. The Altitude object shall be: <ol style="list-style-type: none"> Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = MDC_PART_PHD_HF MDC_HF_ALT Mandatory attribute Metric-Spec_Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall (BITS-16) <input type="checkbox"/> attribute-value ≠ 0x00 0x00 <ul style="list-style-type: none"> bit 0 (mss-avail-intermittent(0)) shall be set. 1(mss-avail-stored-data(1)) shall be set. bit 2 (mss-updt-aperiodic(2)) shall be set. bit 3(mss-msmt-aperiodic(3)) shall be set bit 9 (mss-acc-agent-initiated(9)) shall be set. The other bits have to be 0. Mandatory attribute Unit-Code <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE <input type="checkbox"/> attribute-type = OID-Type (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_DIM_X_M or MDC_DIM_X_FOOT Mandatory attribute Source-Handle-Reference 			

	<input type="checkbox"/> attribute-id = MDC_ATTR_SOURCE_HANDLE_REF <input type="checkbox"/> attribute-type = HANDLE (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = It must be equal to the handle of any Session or Sub-session object in the configuration
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/AG/CLASS/CV/BV-007_A		
TP label	Altitude, timestamp values		
Coverage	Spec	[ISO/IEEE 11073-10441]	
	Testable items	NumObj2; M	NumObj3; M
Test purpose	Check that: Altitude Numeric object instance shall have a timestamp identical to its containing object instance (i.e. Session or Sub-Session). [AND] The timestamp attribute used for each object shall be the same as the one used for the associated Session or Sub-Session object instance		
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_039		
Other PICS			
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<ol style="list-style-type: none"> 1. Take a measurement with the agent under test. 2. Wait for the simulated manager to receive it. Record the timestamp and the Measure-Active-Period of the Session and Sub-Session object and of the Altitude object. 		
Pass/Fail criteria	<ul style="list-style-type: none"> • The timestamp attribute used for the Altitude object shall be the same as that used for the associated Session or Sub-Session object instance. • The Altitude instance shall have a timestamp identical to its associated Session or Sub-Session object instance. 		
Notes			

TP Id	TP/PLT/AG/CLASS/CV/BV-008			
TP label	Distance Numeric Object Attributes			
Coverage	Spec	[ISO/IEEE 11073-10441]		
	Testable items	NumObj5; M	NumObj6; M	Distance1; O
		Distance2; M	Distance 3; M	Distance 4; M
Distance 5; M		Distance 6; M		
Test purpose	Check that: The Distance Numeric object contains the attributes specified for Extended Configuration.			
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_040			
Other PICS				
Initial condition	The agent under test is in the unassociated state.			
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with an Association Response with result = "accepted-unknown-config". 3. The agent responds with a roiv-cmip-confirmed-event report message with a 			

	<p>MDC_NOTI_CONFIG event to send its configuration to the manager.</p> <p>4. The Distance object shall be:</p> <p>a. Mandatory attribute Type</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = MDC_PART_PHD_HF MDC_HF_DISTANCE <p>b. Mandatory attribute Metric-Spec_Small</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall (BITS-16) <input type="checkbox"/> attribute-value ≠ 0x00 0x00 <ul style="list-style-type: none"> • bit 0 (mss-avail-intermittent(0)) shall be set. • bit 1(mss-avail-stored-data(1)) shall be set. • bit 2 (mss-updt-aperiodic(2)) shall be set. • bit 3(mss-msmt-aperiodic(3)) shall be set • bit 9 (mss-acc-agent-initiated(9)) shall be set. • The other bits have to be 0. <p>c. Mandatory attribute Unit-Code</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE <input type="checkbox"/> attribute-type = OID-Type (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_DIM_X_M or MDC_DIM_X_FOOT or MDC_DIM_X_STEP <p>d. Mandatory attribute Source-Handle-Reference</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SOURCE_HANDLE_REF <input type="checkbox"/> attribute-type = HANDLE (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = It must be equal to the handle of any Session or Sub-session object in the configuration <p>5. Wait for the agent under test and the simulated manager to reach the operating state.</p> <p>6. Take a measurement in the agent.</p> <p>7. Wait until the manager receives an event report.</p>
Pass/Fail criteria	<ul style="list-style-type: none"> • In step 4, all checked values are as specified. • In step 7, check that only non-negative values are used.
Notes	

TP Id	TP/PLT/AG/CLASS/CV/BV-008_A		
TP label	Distance, timestamp values		
Coverage	Spec	[ISO/IEEE 11073-10441]	
	Testable items	NumObj2; M	NumObj3; M
Test purpose	<p>Check that:</p> <p>Distance Numeric object instance shall have a timestamp identical to its containing object instance (i.e. Session or Sub-Session).</p> <p>[AND]</p> <p>The timestamp attribute used for each object shall be the same as the one used for the</p>		

	associated Session or Sub-Session object instance
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_040
Other PICS	
Initial condition	The simulated manager and the agent under test are in the operating state.
Test procedure	<ol style="list-style-type: none"> 1. Take a measurement with the agent under test. 2. Wait for the simulated manager to receive it. Record the timestamp and the Measure-Active-Period of the Session and Sub-Session object and of the Distance object
Pass/Fail criteria	<ul style="list-style-type: none"> • The timestamp attribute used for the Distance object shall be the same as that used for the associated Session or Sub-Session object instance. • The Distance instance shall have a timestamp identical to its associated session or sub-session object instance.
Notes	

TP Id	TP/PLT/AG/CLASS/CV/BV-009		
TP label	Ascent Time and Distance Numeric Object Attributes		
Coverage	Spec	[ISO/IEEE 11073-10441]	
	Testable items	NumObj5; M	NumObj6; M
		AscentTime2; M	AscentTime3; M
	AscentTime5; M	AscentTime6; R	AscentTime7; M
Test purpose	<p>Check that:</p> <p>The Ascent Time and Distance Numeric object contains the attributes specified for Extended Configuration.</p>		
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_041		
Other PICS			
Initial condition	The agent under test is in the unassociated state.		
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with an Association Response with result = "accepted-unknown-config". 3. The agent responds with a roiv-cmip-confirmed-event report message with a MDC_NOTI_CONFIG event to send its configuration to the manager. 4. The Ascent time and Distance object shall be: <ol style="list-style-type: none"> a. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = MDC_PART_PHD_HF MDC_HF_ASC_TIME_DIST b. Mandatory attribute Metric-Spec_Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall (BITS-16) <input type="checkbox"/> attribute-value ≠ 0x00 0x00 <ul style="list-style-type: none"> • bit 0 (mss-avail-intermittent(0)) shall be set. • bit 1(mss-avail-stored-data(1)) shall be set. • bit 2 (mss-updt-aperiodic(2)) shall be set. • bit 3(mss-msmt-aperiodic(3)) shall be set • bit 9 (mss-acc-agent-initiated(9)) shall be set. • The other bits have to be 0. c. Mandatory attribute Unit-Code 		

	<ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE <input type="checkbox"/> attribute-type = OID-Type (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_DIM_X_M or MDC_DIM_X_FOOT or MDC_DIM_X_STEP <p>d. Mandatory attribute Source-Handle-Reference</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SOURCE_HANDLE_REF <input type="checkbox"/> attribute-type = HANDLE (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = It must be equal to the handle of any Session or Sub-session object in the configuration <p>e. Recommended attribute Measure-Active-Period</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_PD_MSMT_ACTIVE <input type="checkbox"/> attribute-type = FLOAT-Type (INT-U32) <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> <p>5. Wait for the agent under test and the simulated manager to reach the operating state.</p> <p>6. Take a measurement in the agent.</p> <p>7. Wait until the manager receives an event report.</p>
Pass/Fail criteria	<ul style="list-style-type: none"> • In step 4, all checked values are as specified. • In step 7, check that only non-negative values are used (for observed values of the ascent time and distance object).
Notes	

TP Id	TP/PLT/AG/CLASS/CV/BV-009_A		
TP label	Ascent time and distance, timestamp values		
Coverage	Spec	[ISO/IEEE 11073-10441]	
	Testable items	NumObj2; M	NumObj3; M
Test purpose	<p>Check that:</p> <p>Ascent Time and Distance Numeric object instance shall have a timestamp identical to its containing object instance (i.e. Session or Sub-Session).</p> <p>[AND]</p> <p>The timestamp attribute used for each object shall be the same as the one used for the associated Session or Sub-Session object instance</p>		
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_041		
Other PICS			
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<ol style="list-style-type: none"> 1. Take a measurement with the agent under test. 2. Wait for the simulated manager to receive it. Record the timestamp and the Measure-Active-Period of the Session and Sub-Session object and of the Ascent time and distance Object. 		
Pass/Fail criteria	<ul style="list-style-type: none"> • The timestamp Attribute used for Ascent time and distance object shall be the same as that used for the associated Session or Sub-Session object instance. • Ascent time and distance instance shall have a timestamp identical to its associated Session or Sub-Session object instance. 		
Notes			

TP Id		TP/PLT/AG/CLASS/CV/BV-010		
TP label		Descent Time and Distance Numeric Object Attributes		
Coverage	Spec	[ISO/IEEE 11073-10441]		
	Testable items	NumObj5; M	NumObj6; M	DescentTime1; O
		DescentTime2; M	DescentTime3; M	DescentTime4; M
		DescentTime5; M	DescentTime6; R	DescentTime7; M
Test purpose		<p>Check that:</p> <p>The Descent Time and Distance Numeric object contains the attributes specified for Extended Configuration.</p>		
Applicability		C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_042		
Other PICS				
Initial condition		The agent under test is in the unassociated state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with an Association Response with result = "accepted-unknown-config". 3. The agent responds with a roiv-cmip-confirmed-event report message with a MDC_NOTI_CONFIG event to send its configuration to the manager. 4. The object shall be: <ol style="list-style-type: none"> a. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = MDC_PART_PHD_HF MDC_HF_DESC_TIME_DIST b. Mandatory attribute Metric-Spec_Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall (BITS-16) <input type="checkbox"/> attribute-value ≠ 0x00 0x00 <ul style="list-style-type: none"> • bit 0 (mss-avail-intermittent(0)) shall be set. • bit 1(mss-avail-stored-data(1)) shall be set. • bit 2 (mss-updt-aperiodic(2)) shall be set. • bit 3(mss-msmt-aperiodic(3)) shall be set • bit 9 (mss-acc-agent-initiated(9)) shall be set. • The other bits have to be 0. c. Mandatory attribute Unit-Code <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE <input type="checkbox"/> attribute-type = OID-Type (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_DIM_X_M or MDC_DIM_X_FOOT or MDC_DIM_X_STEP d. Mandatory attribute Source-Handle-Reference <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SOURCE_HANDLE_REF <input type="checkbox"/> attribute-type = HANDLE (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = It must be equal to the handle of any Session or Sub-session object in the configuration e. Recommended attribute Measure-Active-Period 		

	<ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_PD_MSMT_ACTIVE <input type="checkbox"/> attribute-type = FLOAT-Type (INT-U32) <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> <ol style="list-style-type: none"> 5. Wait for the agent under test and the simulated manager to reach the operating state. 6. Take a measurement in the agent. 7. Wait until the manager receives an event report.
Pass/Fail criteria	<ul style="list-style-type: none"> • In step 4, all checked values are as specified. • In step 7, check that only non-negative values are used.
Notes	

TP Id	TP/PLT/AG/CLASS/CV/BV-010_A		
TP label	Descent time and distance, timestamp values		
Coverage	Spec	[ISO/IEEE 11073-10441]	
	Testable items	NumObj2; M	NumObj3; M
Test purpose	<p>Check that:</p> <p>Descent Time and Distance Numeric object instance shall have a timestamp identical to its containing object instance (i.e. Session or Sub-Session).</p> <p>[AND]</p> <p>The timestamp attribute used for each object shall be the same as the one used for the associated Session or Sub-Session object instance</p>		
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_042		
Other PICS			
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<ol style="list-style-type: none"> 1. Take a measurement with the agent under test. 2. Wait for the simulated manager to receive it. Record the timestamp and the Measure-Active-Period of the Session and Sub-Session object and of the Descent Time and Distance object. 		
Pass/Fail criteria	<ul style="list-style-type: none"> • The timestamp attribute used for Descent Time and Distance object shall be the same as the used for the associated Session or Sub-Session object instance. • The Descent Time and Distance instance shall have a timestamp identical to its associated Session or Sub-Session object instance. 		
Notes			

TP Id	TP/PLT/AG/CLASS/CV/BV-011			
TP label	Latitude Numeric Object Attributes			
Coverage	Spec	[ISO/IEEE 11073-10441]		
	Testable items	NumObj5; M	NumObj6; M	Latitude1; O
		Latitude2; M	Latitude3; M	Latitude4; R
Latitude5; M		Latitude6; M		
Test purpose	<p>Check that:</p> <p>The Latitude Numeric object contains the attributes specified for Extended Configuration.</p>			
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_038			
Other PICS				

Initial condition	The agent under test is in the unassociated state.
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with an Association Response with result = "accepted-unknown-config". 3. The agent responds with a roiv-cmip-confirmed-event report message with a MDC_NOTI_CONFIG event to send its configuration to the manager. 4. The Latitude object shall be: <ol style="list-style-type: none"> a. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = MDC_PART_PHD_HF MDC_HF_LATITUDE b. Mandatory attribute Metric-Spec_Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall (BITS-16) <input type="checkbox"/> attribute-value ≠ 0x00 0x00 <ul style="list-style-type: none"> • bit 0 (mss-avail-intermittent (0)) shall be set. • bit 1(mss-avail-stored-data (1)) shall be set. • bit 2 (mss-updt-aperiodic (2)) shall be set. • bit 3(mss-msmt-aperiodic(3)) shall be set • bit 9 (mss-acc-agent-initiated (9)) shall be set. • The other bits have to be 0. c. Not Recommended attribute Unit-Code <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE <input type="checkbox"/> attribute-type = OID-Type (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_DIM_ANG_DEG d. Mandatory attribute Source-Handle-Reference <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SOURCE_HANDLE_REF <input type="checkbox"/> attribute-type = HANDLE (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = It must be equal to the handle of any Session or Sub-session object in the configuration 5. Wait for the agent under test and the simulated manager to reach the operating state. 6. Take a measurement in the agent. 7. Wait until the manager receives an event report.
Pass/Fail criteria	<ul style="list-style-type: none"> • In step 4, all checked values are as specified. In step 7, check that the values are limited to -180 to 180.
Notes	

TP Id	TP/PLT/AG/CLASS/CV/BV-011_A		
TP label	Latitude, timestamp values		
Coverage	Spec	[ISO/IEEE 11073-10441]	
	Testable items	NumObj2; M	NumObj3; M
Test purpose	Check that:		

	<p>Latitude Numeric object instance shall have a timestamp identical to its containing object instance (i.e. Session or Sub-Session).</p> <p>[AND]</p> <p>The timestamp attribute used for each object shall be the same as the one used for the associated Session or Sub-Session object instance</p>
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_038
Other PICS	
Initial condition	The simulated manager and the agent under test are in the operating state.
Test procedure	<ol style="list-style-type: none"> 1. Take a measurement with the agent under test. 2. Wait for the simulated manager to receive it. Record the timestamp and the Measure-Active-Period of the Session and Sub-Session object and of the Latitude object.
Pass/Fail criteria	<ul style="list-style-type: none"> • The timestamp attribute used for the Latitude object shall be the same as that used for the associated Session or Sub-Session object instance. • The Latitude instance shall have a timestamp identical to its associated session or sub-session object instance.
Notes	

TP Id	TP/PLT/AG/CLASS/CV/BV-012		
TP label	Longitude Numeric Object Attributes		
Coverage	Spec	[ISO/IEEE 11073-10441]	
	Testable items	NumObj5; M	NumObj6; M
		Longitude2; M	Longitude3; M
		Longitude5; M	Longitude6; M
Test purpose	<p>Check that:</p> <p>The Longitude Numeric object contains the attributes specified for Extended Configuration.</p>		
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_037		
Other PICS			
Initial condition	The agent under test is in the unassociated state.		
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with an Association Response with result = "accepted-unknown-config". 3. The agent responds with a roiv-cmip-confirmed-event report message with a MDC_NOTI_CONFIG event to send its configuration to the manager. 4. The Longitude object shall be: <ol style="list-style-type: none"> a. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = MDC_PART_PHD_HF MDC_HF_LONGITUDE b. Mandatory attribute Metric-Spec_Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall (BITS-16) <input type="checkbox"/> attribute-value ≠ 0x00 0x00 <ul style="list-style-type: none"> • bit 0 (mss-avail-intermittent (0)) shall be set. • 1(mss-avail-stored-data (1)) shall be set. • bit 2 (mss-updt-aperiodic (2)) shall be set. • bit 3(mss-msmt-aperiodic(3)) shall be set 		

	<ul style="list-style-type: none"> • bit 9 (mss-acc-agent-initiated (9)) shall be set. • The other bits have to be 0. <p>c. Not Recommended attribute Unit-Code</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE <input type="checkbox"/> attribute-type = OID-Type (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_DIM_ANG_DEG <p>d. Mandatory attribute Source-Handle-Reference</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SOURCE_HANDLE_REF <input type="checkbox"/> attribute-type = HANDLE (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = It must be equal to the handle of any Session or Sub-session object in the configuration <p>5. Wait for the agent under test and the simulated manager to reach the operating state.</p> <p>6. Take a measurement in the agent.</p> <p>7. Wait until the manager receives an event report.</p>
Pass/Fail criteria	<ul style="list-style-type: none"> • In step 4, all checked values are as specified. • In step 7, check that the values are limited to -180 to 180.
Notes	

TP Id		TP/PLT/AG/CLASS/CV/BV-012_A	
TP label		Longitude, timestamp values	
Coverage	Spec	[ISO/IEEE 11073-10441]	
	Testable items	NumObj2; M	NumObj3; M
Test purpose		<p>Check that:</p> <p>Longitude Numeric object instance shall have a timestamp identical to its containing object instance (i.e. Session or Sub-Session).</p> <p>[AND]</p> <p>The timestamp attribute used for each object shall be the same as the one used for the associated Session or Sub-Session object instance</p>	
Applicability		C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_037	
Other PICS			
Initial condition		The simulated manager and the agent under test are in the operating state.	
Test procedure		<ol style="list-style-type: none"> 1. Take a measurement with the agent under test. 2. Wait for the simulated manager to receive it. Record the Time Stamp and the Measure-Active-Period of the Session and Sub-Session object and of the Longitude object. 	
Pass/Fail criteria		<ul style="list-style-type: none"> • The timestamp attribute used for the Longitude object shall be the same as that used for the associated Session or Sub-Session object instance. • The Longitude instance shall have a timestamp identical to its associated Session or Sub-Session object instance. 	
Notes			

TP Id	TP/PLT/AG/CLASS/CV/BV-013
TP label	Slopes Numeric Object Attributes

Coverage	Spec	[ISO/IEEE 11073-10441]		
	Testable items	NumObj5; M	NumObj6; M	Slopes1; O
		Slopes2; M	Slopes3; M	Slopes4; R
		Slopes5; M	Slopes6; M	
Test purpose	Check that: The Slopes Numeric object contains the attributes specified for Extended Configuration.			
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_036			
Other PICS				
Initial condition	The agent under test is in the unassociated state.			
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with an Association Response with result = "accepted-unknown-config". 3. The agent responds with a roiv-cmip-confirmed-event report message with a MDC_NOTI_CONFIG event to send its configuration to the manager. 4. The Slopes object shall be: <ol style="list-style-type: none"> a. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = MDC_PART_PHD_HF MDC_HF_SLOPES b. Mandatory attribute Metric-Spec_Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall (BITS-16) <input type="checkbox"/> attribute-value ≠ 0x00 0x00 <ul style="list-style-type: none"> • bit 0 (mss-avail-intermittent(0)) shall be set. • bit 1(mss-avail-stored-data(1)) shall be set. • 2 (mss-updt-aperiodic(2)) shall be set. • bit 3(mss-msmt-aperiodic(3)) shall be set • bit 9 (mss-acc-agent-initiated(9)) shall be set. • The other bits have to be 0. c. Not Recommended attribute Unit-Code <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE <input type="checkbox"/> attribute-type = OID-Type (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_DIM_DIMLESS d. Mandatory attribute Source-Handle-Reference <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SOURCE_HANDLE_REF <input type="checkbox"/> attribute-type = HANDLE (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = It must be equal to the handle of any Session or Sub-session object in the configuration 5. Wait for the agent under test and the simulated manager to reach the operating state. 6. Take a measurement in the agent. 7. Wait until the manager receives an event report. 			
Pass/Fail criteria	<ul style="list-style-type: none"> • In step 4, all checked values are as specified. • In step 7, check that only non-negative values are used. 			
Notes				

TP Id		TP/PLT/AG/CLASS/CV/BV-013_A		
TP label		Slopes, timestamp values		
Coverage	Spec	[ISO/IEEE 11073-10441]		
	Testable items	NumObj2; M	NumObj3; M	
Test purpose		<p>Check that:</p> <p>Slopes Numeric object instance shall have a timestamp identical to its containing object instance (i.e. Session or Sub-Session).</p> <p>[AND]</p> <p>The timestamp attribute used for each object shall be the same as the one used for the associated Session or Sub-Session object instance</p>		
Applicability		C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_036		
Other PICS				
Initial condition		The simulated manager and the agent under test are in the operating state.		
Test procedure		<ol style="list-style-type: none"> 1. Take a measurement with the agent under test. 2. Wait for the simulated manager to receive it. Record the Time Stamp and the Measure-Active-Period of the Session and Sub-Session object and of the Slopes object. 		
Pass/Fail criteria		<ul style="list-style-type: none"> • The timestamp attribute used for the Slopes object shall be the same as that used for the associated Session or Sub-Session object instance. • The Slopes instance shall have a timestamp identical to its associated Session or Sub-Session object instance. 		
Notes				

TP Id		TP/PLT/AG/CLASS/CV/BV-014		
TP label		Speed Numeric Object Attributes		
Coverage	Spec	[ISO/IEEE 11073-10441]		
	Testable items	NumObj5; M	NumObj6; M	Speed1; O
		Speed2; M	Speed3; M	Speed4; M
		Speed5; M	Speed6; M	
Test purpose		<p>Check that:</p> <p>The Speed Numeric object contains the attributes specified for Extended Configuration.</p>		
Applicability		C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_035		
Other PICS				
Initial condition		The agent under test is in the unassociated state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with an Association Response with result = "accepted-unknown-config". 3. The agent responds with a roiv-cmip-confirmed-event report message with a MDC_NOTI_CONFIG event to send its configuration to the manager. 4. The Speed object shall be: <ol style="list-style-type: none"> a. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = MDC_PART_PHD_HF MDC_HF_SPEED b. Mandatory attribute Metric-Spec_Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL 		

	<ul style="list-style-type: none"> <input type="checkbox"/> attribute-type = MetricSpecSmall (BITS-16) <input type="checkbox"/> attribute-value ≠ 0x00 0x00 <ul style="list-style-type: none"> • bit 0 (mss-avail-intermittent(0)) shall be set. • bit 1(mss-avail-stored-data(1)) shall be set. • bit 2 (mss-updt-aperiodic(2)) shall be set. • bit 3(mss-msmt-aperiodic(3)) shall be set • bit 9 (mss-acc-agent-initiated(9)) shall be set. • The other bits have to be 0. <p>c. Mandatory attribute Metric-Id</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_PHYSIO <input type="checkbox"/> attribute-type = OID-Type <input type="checkbox"/> attribute-value.length =INT-U16 <input type="checkbox"/> attribute-value = MDC_HF_MEAN_NULL_EXCLUDE or MDC_HF_MEAN_NULL_INCLUDE or MDC_HF_MEAN_NULL_EXCLUDE or MDC_HF_MAX or MDC_HF_MIN <p>d. Mandatory attribute Unit-Code</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE <input type="checkbox"/> attribute-type = OID-Type (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_DIM_X_M_PER_MIN or MDC_DIM_X_FOOT_PER_MIN or MDC_DIM_X_INCH_PER_MIN or MDC_DIM_X_STEP_PER_MIN <p>e. Mandatory attribute Source-Handle-Reference</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SOURCE_HANDLE_REF <input type="checkbox"/> attribute-type = HANDLE (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = It must be equal to the handle of any Session or Sub-session object in the configuration <p>5. Wait for the agent under test and the simulated manager to reach the operating state.</p> <p>6. Take a measurement in the agent.</p> <p>7. Wait until the manager receives an event report.</p>
Pass/Fail criteria	<ul style="list-style-type: none"> • In step 4, all checked values are as specified. • In step 7, check that only non-negative values are used.
Notes	

TP Id	TP/PLT/AG/CLASS/CV/BV-014_A		
TP label	Speed, timestamp values		
Coverage	Spec	[ISO/IEEE 11073-10441]	
	Testable items	NumObj2; M	NumObj3; M
Test purpose	<p>Check that:</p> <p>Speed Numeric object instance shall have a timestamp identical to its containing object instance (i.e. Session or Sub-Session).</p> <p>[AND]</p> <p>The timestamp attribute used for each object shall be the same as the one used for the associated Session or Sub-Session object instance</p>		

Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_035
Other PICS	
Initial condition	The simulated manager and the agent under test are in the operating state.
Test procedure	<ol style="list-style-type: none"> 1. Take a measurement with the agent under test. 2. Wait for the simulated manager to receive it. Record the timestamp and the Measure-Active-Period of the Session and Sub-Session object and of the Speed object.
Pass/Fail criteria	<ul style="list-style-type: none"> • The timestamp attribute used for Speed object shall be the same as that used for the associated Session or Sub-Session object instance. • The Speed instance shall have a timestamp identical to its associated Session or Sub-Session object instance.
Notes	

TP Id	TP/PLT/AG/CLASS/CV/BV-015			
TP label	Cadence Numeric Object Attributes			
Coverage	Spec	[ISO/IEEE 11073-10441]		
	Testable items	NumObj5; M	NumObj6; M	Cadence1; O
		Cadence2; M	Cadence3; M	Cadence4; M
Cadence5; R		Cadence6; M	Cadence7; M	
Test purpose	<p>Check that:</p> <p>The Cadence Numeric object contains the attributes specified for Extended Configuration.</p>			
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_034			
Other PICS				
Initial condition	The agent under test is in the unassociated state.			
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with an Association Response with result = "accepted-unknown-config". 3. The agent responds with a roiv-cmip-confirmed-event report message with a MDC_NOTI_CONFIG event to send its configuration to the manager. 4. The Cadence object shall be: <ol style="list-style-type: none"> a. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = MDC_PART_PHD_HF MDC_HF_CAD b. Mandatory attribute Metric-Spec_Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall (BITS-16) <input type="checkbox"/> attribute-value ≠ 0x00 0x00 <ul style="list-style-type: none"> • bit 0 (mss-avail-intermittent(0)) shall be set. • bit 1(mss-avail-stored-data(1)) shall be set. • bit 2 (mss-updt-aperiodic(2)) shall be set. • bit 3(mss-msmt-aperiodic(3)) shall be set • bit 9 (mss-acc-agent-initiated(9)) shall be set. • The other bits have to be 0. c. Mandatory attribute Metric-Id <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_PHYSIO 			

	<ul style="list-style-type: none"> <input type="checkbox"/> attribute-type = OID-Type (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_HF_MEAN_NULL_EXCLUDE or MDC_HF_MEAN_NULL_INCLUDE or MDC_HF_MEAN_NULL_EXCLUDE or MDC_HF_MAX or MDC_HF_MIN <p>d. Not Recommended attribute Unit-Code</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE <input type="checkbox"/> attribute-type = OID-Type (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_DIM_RPM <p>e. Mandatory attribute Source-Handle-Reference</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SOURCE_HANDLE_REF <input type="checkbox"/> attribute-type = HANDLE (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = It must be equal to the handle of any Session or Sub-session object in the configuration <p>5. Wait for the agent under test and the simulated manager to reach the operating state.</p> <p>6. Take a measurement in the agent.</p> <p>7. Wait until the manager receives an event report.</p>
Pass/Fail criteria	<ul style="list-style-type: none"> • In step 4, all checked values are as specified. • In step 7, check that only non-negative values are used.
Notes	

TP Id	TP/PLT/AG/CLASS/CV/BV-015_A		
TP label	Cadence, timestamp values		
Coverage	Spec	[ISO/IEEE 11073-10441]	
	Testable items	NumObj2; M	NumObj3; M
Test purpose	<p>Check that:</p> <p>Cadence Numeric object instance shall have a timestamp identical to its containing object instance (i.e. Session or Sub-Session).</p> <p>[AND]</p> <p>The timestamp attribute used for each object shall be the same as the one used for the associated Session or Sub-Session object instance</p>		
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_034		
Other PICS			
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<ol style="list-style-type: none"> 1. Take a measurement with the agent under test. 2. Wait for the simulated manager to receive it. Record the timestamp and the Measure-Active-Period of the Session and Sub-Session object and of the Cadence object. 		
Pass/Fail criteria	<ul style="list-style-type: none"> • The timestamp attribute used for Cadence object shall be the same as that used for the associated Session or Sub-Session object instance. • The Cadence instance shall have a timestamp identical to its associated session or sub-session object instance. 		
Notes			

TP Id		TP/PLT/AG/CLASS/CV/BV-016		
TP label		Incline Numeric Object Attributes		
Coverage	Spec	[ISO/IEEE 11073-10441]		
	Testable items	NumObj5; M	NumObj6; M	Incline1; O
		Incline2; M	Incline3; M	Incline4; M
		Incline5; M	Incline6; M	
Test purpose		<p>Check that:</p> <p>The Incline Numeric object contains the attributes specified for Extended Configuration.</p>		
Applicability		C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_033		
Other PICS				
Initial condition		The agent under test is in the unassociated state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with an Association Response with result = "accepted-unknown-config". 3. The agent responds with a roiv-cmip-confirmed-event report message with a MDC_NOTI_CONFIG event to send its configuration to the manager. 4. The Incline object shall be: <ol style="list-style-type: none"> a. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = MDC_PART_PHD_HF MDC_HF_INCLINE b. Mandatory attribute Metric-Spec_Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall (BITS-16) <input type="checkbox"/> attribute-value ≠ 0x00 0x00 <ul style="list-style-type: none"> • bit 0 (mss-avail-intermittent(0)) shall be set. • bit 1(mss-avail-stored-data(1)) shall be set. • bit 2 (mss-updt-aperiodic(2)) shall be set. • bit 3(mss-msmt-aperiodic(3)) shall be set • bit 9 (mss-acc-agent-initiated(9)) shall be set. • The other bits have to be 0. c. Mandatory attribute Metric-Id <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_PHYSIO <input type="checkbox"/> attribute-type = OID-Type (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_HF_MEAN_NULL_EXCLUDE or MDC_HF_MEAN_NULL_INCLUDE or MDC_HF_MEAN_NULL_EXCLUDE or MDC_HF_MAX or MDC_HF_MIN d. Mandatory attribute Unit-Code <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE <input type="checkbox"/> attribute-type = OID-Type (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_DIM_PERCENT or MDC_DIM_ANG_DEG e. Mandatory attribute Source-Handle-Reference <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SOURCE_HANDLE_REF 		

	<ul style="list-style-type: none"> <input type="checkbox"/> attribute-type = HANDLE <input type="checkbox"/> attribute-value.length = INT-U16 <input type="checkbox"/> attribute-value = It must be equal to the handle of any Session or Sub-session object in the configuration <ol style="list-style-type: none"> 5. Wait for the agent under test and the simulated manager to reach the operating state. 6. Take a measurement in the agent. 7. Wait until the manager receives an event report.
Pass/Fail criteria	<ul style="list-style-type: none"> • In step 4, all checked values are as specified. • In step 7, check values.
Notes	

TP Id	TP/PLT/AG/CLASS/CV/BV-016_A		
TP label	Incline, timestamp values		
Coverage	Spec	[ISO/IEEE 11073-10441]	
	Testable items	NumObj2; M	NumObj3; M
Test purpose	<p>Check that:</p> <p>Incline Numeric object instance shall have a timestamp identical to its containing object instance (i.e. Session or Sub-Session).</p> <p>[AND]</p> <p>The timestamp attribute used for each object shall be the same as the one used for the associated Session or Sub-Session object instance</p>		
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_033		
Other PICS			
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<ol style="list-style-type: none"> 1. Take a measurement with the agent under test. 2. Wait for the simulated manager to receive it. Record the timestamp and the Measure-Active-Period of the Session and Sub-Session object and of the Incline object. 		
Pass/Fail criteria	<ul style="list-style-type: none"> • The Timestamp attribute used for Incline object shall be the same as that used for the associated Session or Sub-Session object instance. • The Incline instance shall have a timestamp identical to its associated session or sub-session object instance. 		
Notes			

TP Id	TP/PLT/AG/CLASS/CV/BV-017			
TP label	Heart Rate Numeric Object Attributes			
Coverage	Spec	[ISO/IEEE 11073-10441]		
	Testable items	NumObj5; M	NumObj6; M	Heart rate1; O
		Heart rate2; M	Heart rate3; M	Heart rate4; M
Heart rate5; R		Heart rate6; M	Heart rate7 M	
Test purpose	<p>Check that:</p> <p>The Heart Rate Numeric object contains the attributes specified for Extended Configuration.</p>			
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_031			
Other PICS				
Initial condition	The agent under test is in the unassociated state.			
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 			

	<ol style="list-style-type: none"> 2. The simulated manager responds with an Association Response with result = "accepted-unknown-config". 3. The agent responds with a roiv-cmip-confirmed-event report message with a MDC_NOTI_CONFIG event to send its configuration to the manager. 4. The Heart rate object shall be: <ol style="list-style-type: none"> a. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = MDC_PART_PHD_HF MDC_HF_HR b. Mandatory attribute Metric-Spec_Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall (BITS-16) <input type="checkbox"/> attribute-value ≠ 0x00 0x00 <ul style="list-style-type: none"> • bit 0 (mss-avail-intermittent(0)) shall be set. • bit 1(mss-avail-stored-data(1)) shall be set. • bit 2 (mss-updt-aperiodic(2)) shall be set. • bit 3(mss-msmt-aperiodic(3)) shall be set • bit 9 (mss-acc-agent-initiated(9)) shall be set. • The other bits have to be 0. c. Mandatory attribute Metric-Id <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_PHYSIO <input type="checkbox"/> attribute-type = OID-Type (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_HF_MEAN_NULL_EXCLUDE or MDC_HF_MEAN_NULL_INCLUDE or MDC_HF_MEAN_NULL_EXCLUDE or MDC_HF_MAX or MDC_HF_MIN d. Not Recommended attribute Unit-Code <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE <input type="checkbox"/> attribute-type = OID-Type (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_DIM_BEAT_PER_MIN e. Mandatory attribute Source-Handle-Reference <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SOURCE_HANDLE_REF <input type="checkbox"/> attribute-type = HANDLE (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = It must be equal to the handle of any Session or Sub-session object in the configuration 5. Wait for the agent under test and the simulated manager to reach the operating state. 6. Take a measurement in the agent. 7. Wait until the manager receives an event report.
Pass/Fail criteria	<ul style="list-style-type: none"> • In step 4, all checked values are as specified. • In step 7, check that only non-negative values are used.
Notes	

TP Id		TP/PLT/AG/CLASS/CV/BV-017_A		
TP label		Heart Rate, timestamp values		
Coverage	Spec	[ISO/IEEE 11073-10441]		
	Testable items	NumObj2; M	NumObj3; M	
Test purpose		<p>Check that:</p> <p>Heart Rate Numeric object instance shall have a timestamp identical to its containing object instance (i.e. Session or Sub-Session).</p> <p>[AND]</p> <p>The timestamp attribute used for each object shall be the same as the one used for the associated Session or Sub-Session object instance</p>		
Applicability		C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_031		
Other PICS				
Initial condition		The simulated manager and the agent under test are in the operating state.		
Test procedure		<ol style="list-style-type: none"> 1. Take a measurement with the agent under test. 2. Wait for the simulated manager to receive it. Record the Time Stamp and the Measure-Active-Period of the Session and Sub-Session object and of the Heart rate object. 		
Pass/Fail criteria		<ul style="list-style-type: none"> • The timestamp attribute used for Heart Rate object shall be the same as that used for the associated Session or Sub-Session object instance. • The Heart rate instance shall have a timestamp identical to its associated session or sub-session object instance. 		
Notes				

TP Id		TP/PLT/AG/CLASS/CV/BV-018		
TP label		Max user Heart Rate Numeric Object Attributes		
Coverage	Spec	[ISO/IEEE 11073-10441]		
	Testable items	NumObj5; M	NumObj6; M	Max user heart rate1; O
		Max user heart rate2; M	Max user heart rate3; M	Max user heart rate4; X
		Max user heart rate5; R	Max user heart rate6; M	Max user heart rate7; M
Test purpose		<p>Check that:</p> <p>The Max User Heart Rate Numeric object contains the attributes specified for Extended Configuration.</p>		
Applicability		C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_030		
Other PICS				
Initial condition		The agent under test is in the unassociated state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with an Association Response with result = "accepted-unknown-config". 3. The agent responds with a roiv-cmip-confirmed-event report message with a MDC_NOTI_CONFIG event to send its configuration to the manager. 4. The Max user Heart rate object shall be: <ol style="list-style-type: none"> a. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = MDC_PART_PHD_HF MDC_HF_HR_MAX_USER b. Mandatory attribute Metric-Spec_Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL 		

	<ul style="list-style-type: none"> <input type="checkbox"/> attribute-type = MetricSpecSmall (BITS-16) <input type="checkbox"/> attribute-value ≠ 0x00 0x00 <ul style="list-style-type: none"> • bit 0 (mss-avail-intermittent(0)) shall be set. • bit 1(mss-avail-stored-data(1)) shall be set. • bit 2 (mss-updt-aperiodic(2)) shall be set. • bit 3(mss-msmt-aperiodic(3)) shall be set • bit 9 (mss-acc-agent-initiated(9)) shall be set. • The other bits have to be 0. <p>c. Mandatory attribute Metric-Id</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_PHYSIO <input type="checkbox"/> attribute-type = OID-Type (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <p>d. Not Recommended attribute Unit-Code</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE <input type="checkbox"/> attribute-type = OID-Type (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_DIM_BEAT_PER_MIN <p>e. Mandatory attribute Source-Handle-Reference</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SOURCE_HANDLE_REF <input type="checkbox"/> attribute-type = HANDLE (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = It must be equal to the handle of any Session or Sub-session object in the configuration <p>5. Wait for the agent under test and the simulated manager to reach the operating state.</p> <p>6. Take a measurement in the agent.</p> <p>7. Wait until the manager receives an event report.</p>
Pass/Fail criteria	<ul style="list-style-type: none"> • In step 4, all checked values are as specified. • In step 7, check that only non-negative values are used.
Notes	<p>Metric-Id has been considered as mandatory, but its qualifier has to be clarified by Continua.</p> <p>Opened bug :http://certification.continuaalliance.org/bugzilla/show_bug.cgi?id=465</p>

TP Id	TP/PLT/AG/CLASS/CV/BV-018_A		
TP label	Max user heart rate, timestamp values		
Coverage	Spec	[ISO/IEEE 11073-10441]	
	Testable items	NumObj2; M	NumObj3; M
Test purpose	<p>Check that:</p> <p>Max User Heart Rate Numeric object instance shall have a timestamp identical to its containing object instance (i.e. Session or Sub-Session).</p> <p>[AND]</p> <p>The timestamp attribute used for each object shall be the same as the one used for the associated Session or Sub-Session object instance</p>		
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_030		
Other PICS			

Initial condition	The simulated manager and the agent under test are in the operating state.
Test procedure	<ol style="list-style-type: none"> 1. Take a measurement with the agent under test. 2. Wait for the simulated manager to receive it. Record the timestamp and the Measure-Active-Period of the Session and Sub-Session object and of the Max user heart rate object.
Pass/Fail criteria	<ul style="list-style-type: none"> • The timestamp attribute used for the Max user heart rate object shall be the same as that used for the associated Session or Sub-Session object instance. • The Max user heart rate instance shall have a timestamp identical to its associated session or sub-session object instance.
Notes	

TP Id	TP/PLT/AG/CLASS/CV/BV-019			
TP label	Power Numeric Object Attributes			
Coverage	Spec	[ISO/IEEE 11073-10441]		
	Testable items	NumObj5; M	NumObj6; M	Power1; O
		Power2; M	Power3; M	Power4; M
		Power5; R	Power6; M	Power7; M
Test purpose	<p>Check that:</p> <p>The Power Numeric object contains the attributes specified for Extended Configuration.</p>			
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_029			
Other PICS				
Initial condition	The agent under test is in the unassociated state.			
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with an Association Response with result = "accepted-unknown-config". 3. The agent responds with a roiv-cmip-confirmed-event report message with a MDC_NOTI_CONFIG event to send its configuration to the manager. 4. The Power object shall be: <ol style="list-style-type: none"> a. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = MDC_PART_PHD_HF MDC_HF_POWER b. Mandatory attribute Metric-Spec_Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall (BITS-16) <input type="checkbox"/> attribute-value ≠ 0x00 0x00 <ul style="list-style-type: none"> • bit 0 (mss-avail-intermittent(0)) shall be set. • bit 1(mss-avail-stored-data(1)) shall be set. • bit 2 (mss-updt-aperiodic(2)) shall be set. • bit 3(mss-msmt-aperiodic(3)) shall be set • bit 9 (mss-acc-agent-initiated(9)) shall be set. • The other bits have to be 0. c. Mandatory attribute Metric-Id <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_PHYSIO <input type="checkbox"/> attribute-type = OID-Type (INT-U16) 			

	<ul style="list-style-type: none"> <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_HF_MEAN_NULL_EXCLUDE or MDC_HF_MEAN_NULL_INCLUDE or MDC_HF_MEAN_NULL_EXCLUDE or MDC_HF_MAX or MDC_HF_MIN <p>d. Not Recommended attribute Unit-Code</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE <input type="checkbox"/> attribute-type = OID-Type (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_DIM_X_WATT <p>e. Mandatory attribute Source-Handle-Reference</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SOURCE_HANDLE_REF <input type="checkbox"/> attribute-type = HANDLE (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = It must be equal to the handle of any Session or Sub-session object in the configuration <p>5. Wait for the agent under test and the simulated manager to reach the operating state.</p> <p>6. Take a measurement in the agent.</p> <p>7. Wait until the manager receives an event report.</p>
Pass/Fail criteria	<ul style="list-style-type: none"> • In step 4, all checked values are as specified. • In step 7, check that only non-negative values are used.
Notes	

TP Id	TP/PLT/AG/CLASS/CV/BV-019_A		
TP label	Power, timestamp values		
Coverage	Spec	[ISO/IEEE 11073-10441]	
	Testable items	NumObj2; M	NumObj3; M
Test purpose	<p>Check that:</p> <p>Power Numeric object instance shall have a timestamp identical to its containing object instance (i.e. Session or Sub-Session).</p> <p>[AND]</p> <p>The timestamp attribute used for each object shall be the same as the one used for the associated Session or Sub-Session object instance</p>		
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_029		
Other PICS			
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<ol style="list-style-type: none"> 1. Take a measurement with the agent under test. 2. Wait for the simulated manager to receive it. Record the Time Stamp and the Measure-Active-Period of the Session and Sub-Session object and of the Power object. 		
Pass/Fail criteria	<ul style="list-style-type: none"> • The Timestamp attribute used for the power object shall be the same as that used for the associated Session or Sub-Session object instance. • The Power instance shall have a timestamp identical to its associated session or sub-session object instance. 		
Notes			

TP Id		TP/PLT/AG/CLASS/CV/BV-020		
TP label		Resistance Numeric Object Attributes		
Coverage	Spec	[ISO/IEEE 11073-10441]		
	Testable items	NumObj5; M	NumObj6; M	Resistance1; O
		Resistance2; M	Resistance3; M	Resistance4; M
		Resistance5; R	Resistance6; M	
Test purpose		<p>Check that:</p> <p>The Resistance Numeric object contains the attributes specified for Extended Configuration.</p>		
Applicability		C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_028		
Other PICS				
Initial condition		The agent under test is in the unassociated state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with an Association Response with result = "accepted-unknown-config". 3. The agent responds with a roiv-cmip-confirmed-event report message with a MDC_NOTI_CONFIG event to send its configuration to the manager. 4. The Resistance object shall be: <ol style="list-style-type: none"> a. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = MDC_PART_PHD_HF MDC_HF_RESIST b. Mandatory attribute Metric-Spec-Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall (BITS-16) <input type="checkbox"/> attribute-value ≠ 0x00 0x00 <ul style="list-style-type: none"> • bit 0 (mss-avail-intermittent(0)) shall be set. • bit 1 (mss-avail-stored-data(1)) shall be set. • bit 2 (mss-updt-aperiodic(2)) shall be set. • bit 3 (mss-msmt-aperiodic(3)) shall be set • bit 9 (mss-acc-agent-initiated(9)) shall be set. • The other bits have to be 0. c. Mandatory attribute Metric-Id <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_PHYSIO <input type="checkbox"/> attribute-type = OID-Type (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_HF_MEAN_NULL_EXCLUDE or MDC_HF_MEAN_NULL_INCLUDE or MDC_HF_MEAN_NULL_EXCLUDE or MDC_HF_MAX or MDC_HF_MIN d. Not recommended attribute Unit-Code <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE <input type="checkbox"/> attribute-type = OID-Type (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes e. Mandatory attribute Source-Handle-Reference <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SOURCE_HANDLE_REF 		

	<input type="checkbox"/> attribute-type = HANDLE (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = It must be equal to the handle of any Session or Sub-session object in the configuration
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/AG/CLASS/CV/BV-020_A		
TP label	Resistance, timestamp values		
Coverage	Spec	[ISO/IEEE 11073-10441]	
	Testable items	NumObj2; M	NumObj3; M
Test purpose	Check that: Resistance Numeric object instance shall have a timestamp identical to its containing object instance (i.e. Session or Sub-Session). [AND] The timestamp attribute used for each object shall be the same as the one used for the associated Session or Sub-Session object instance		
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_028		
Other PICS			
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<ol style="list-style-type: none"> Take a measurement with the agent under test. Wait for the simulated manager to receive it. Record the Time Stamp and the Measure-Active-Period of the Session and Sub-Session object and of the Resistance object. 		
Pass/Fail criteria	<ul style="list-style-type: none"> The timestamp attribute used for resistance object shall be the same as the used for the associated Session or Sub-Session object instance. The Resistance instance shall have a timestamp identical to its associated session or sub-session object instance. 		
Notes			

TP Id	TP/PLT/AG/CLASS/CV/BV-021			
TP label	Stride Numeric Object Attributes			
Coverage	Spec	[ISO/IEEE 11073-10441]		
	Testable items	NumObj5; M	NumObj6; M	Stride length1; O
		Stride length2; M	Stride length3; M	Stride length4; M
Stride length5; M		Stride length6; M		
Test purpose	Check that: The Stride length Numeric object contains the attributes specified for Extended Configuration.			
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_027			
Other PICS				
Initial condition	The agent under test is in the unassociated state.			
Test procedure	<ol style="list-style-type: none"> The simulated manager receives an association request from the agent under test. The simulated manager responds with an Association Response with result = "accepted-unknown-config". The agent responds with a roiv-cmip-confirmed-event report message with a MDC_NOTI_CONFIG event to send its configuration to the manager. 			

	<p>4. The Stride object shall be:</p> <p>a. Mandatory attribute Type</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = MDC_PART_PHD_HF MDC_HF_STRIDE <p>b. Mandatory attribute Metric-Spec_Small</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall (BITS-16) <input type="checkbox"/> attribute-value ≠ 0x00 0x00 <ul style="list-style-type: none"> • bit 0 (mss-avail-intermittent(0)) shall be set. • bit 1(mss-avail-stored-data(1)) shall be set. • bit 2 (mss-updt-aperiodic(2)) shall be set. • bit 3(mss-msmt-aperiodic(3)) shall be set • bit 9 (mss-acc-agent-initiated(9)) shall be set. • The other bits have to be 0. <p>c. Mandatory attribute Metric-Id</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_PHYSIO <input type="checkbox"/> attribute-type = OID-Type (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_HF_MEAN_NULL_EXCLUDE or MDC_HF_MEAN_NULL_INCLUDE or MDC_HF_MEAN_NULL_EXCLUDE or MDC_HF_MAX or MDC_HF_MIN <p>d. Mandatory attribute Unit-Code</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE <input type="checkbox"/> attribute-type = OID-Type (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_DIM_X_M or MDC_DIM_X_INCH <p>e. Mandatory attribute Source-Handle-Reference</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SOURCE_HANDLE_REF <input type="checkbox"/> attribute-type = HANDLE (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = It must be equal to the handle of any Session or Sub-session object in the configuration
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/AG/CLASS/CV/BV-021_A		
TP label	Stride, timestamp values		
Coverage	Spec	[ISO/IEEE 11073-10441]	
	Testable items	NumObj2; M	NumObj3; M
Test purpose	Check that: Stride length Numeric object instance shall have a timestamp identical to its containing object instance (i.e. Session or Sub-Session). [AND]		

	The timestamp attribute used for each object shall be the same as the one used for the associated Session or Sub-Session object instance
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_027
Other PICS	
Initial condition	The simulated manager and the agent under test are in the operating state.
Test procedure	<ol style="list-style-type: none"> 1. Take a measurement with the agent under test. 2. Wait for the simulated manager to receive it. Record the timestamp and the Measure-Active-Period of the Session and Sub-Session object and of the Stride object.
Pass/Fail criteria	<ul style="list-style-type: none"> • The Timestamp attribute used for stride object shall be the same as the used for the associated Session or Sub-Session object instance. • The Stride instance shall have a timestamp identical to its associated session or sub-session object instance.
Notes	

TP Id	TP/PLT/AG/CLASS/CV/BV-022			
TP label	Breathing Numeric Object Attributes			
Coverage	Spec	[ISO/IEEE 11073-10441]		
	Testable items	NumObj5; M	NumObj6; M	Breathing rate1; O
		Breathing rate2; M	Breathing rate3; M	Breathing rate4; M
		Breathing rate5; R	Breathing rate6; M	Breathing rate7; M
Test purpose	<p>Check that:</p> <p>The Breathing rate Numeric object contains the attributes specified for Extended Configuration.</p>			
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_026			
Other PICS				
Initial condition	The agent under test is in the unassociated state.			
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with an Association Response with result = "accepted-unknown-config". 3. The agent responds with a roiv-cmip-confirmed-event report message with a MDC_NOTI_CONFIG event to send its configuration to the manager. 4. The Breathing object shall be: <ol style="list-style-type: none"> a. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = MDC_PART_PHD_HF MDC_RESP_RATE b. Mandatory attribute Metric-Spec_Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall (BITS-16) <input type="checkbox"/> attribute-value ≠ 0x00 0x00 <ul style="list-style-type: none"> • bit 0 (mss-avail-intermittent(0)) shall be set. • bit 1(mss-avail-stored-data(1)) shall be set. • bit 2 (mss-updt-aperiodic(2)) shall be set. • bit 3(mss-msmt-aperiodic(3)) shall be set • bit 9 (mss-acc-agent-initiated(9)) shall be set. 			

	<ul style="list-style-type: none"> • The other bits have to be 0. <p>c. Mandatory attribute Metric-Id</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_PHYSIO <input type="checkbox"/> attribute-type = OID-Type <input type="checkbox"/> attribute-value.length = INT-U16 <input type="checkbox"/> attribute-value = MDC_HF_MEAN_NULL_EXCLUDE or MDC_HF_MEAN_NULL_INCLUDE or MDC_HF_MEAN_NULL_EXCLUDE or MDC_HF_MAX or MDC_HF_MIN <p>d. Not Recommended attribute Unit-Code</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE <input type="checkbox"/> attribute-type = OID-Type <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_DIM_RESP_PER_MIN <p>e. Mandatory attribute Source-Handle-Reference</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SOURCE_HANDLE_REF <input type="checkbox"/> attribute-type = HANDLE <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = It must be equal to the handle of any Session or Sub-session object in the configuration <p>5. Wait for the agent under test and the simulated manager to reach the operating state.</p> <p>6. Take a measurement in the agent.</p> <p>7. Wait until the manager receives an event report.</p>
Pass/Fail criteria	<ul style="list-style-type: none"> • In step 4, all checked values are as specified. In step 7, check that only non-negative values are used.
Notes	

TP Id	TP/PLT/AG/CLASS/CV/BV-022_A		
TP label	Breathing rate, timestamp values		
Coverage	Spec	[ISO/IEEE 11073-10441]	
	Testable items	NumObj2; M	NumObj3; M
Test purpose	<p>Check that:</p> <p>Breathing rate Numeric object instance shall have a timestamp identical to its containing object instance (i.e. Session or Sub-Session).</p> <p>[AND]</p> <p>The timestamp attribute used for each object shall be the same as the one used for the associated Session or Sub-Session object instance</p>		
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_026		
Other PICS			
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<ol style="list-style-type: none"> 1. Take a measurement with the agent under test. 2. Wait for the simulated manager to receive it. Record the timestamp and the Measure-Active-Period of the Session and Sub-Session object and of the Breathing rate object. 		
Pass/Fail criteria	<ul style="list-style-type: none"> • The timestamp attribute used for the Breathing rate object shall be the same as that used for the associated Session or Sub-Session object instance. • The Breathing rate instance shall have a timestamp identical to its associated session or sub-session object instance. 		

Notes	
--------------	--

TP Id	TP/PLT/AG/CLASS/CV/BV-023		
TP label	Energy Numeric Object Attributes		
Coverage	Spec	[ISO/IEEE 11073-10441]	
	Testable items	NumObj5; M	NumObj6; M
		Energy expended2; M	Energy expended3; M
	Energy expended5; M	Energy expended6; M	
Test purpose	Check that: The Energy Numeric object contains the attributes specified for Extended Configuration.		
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_025		
Other PICS			
Initial condition	The agent under test is in the unassociated state.		
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with an Association Response with result = "accepted-unknown-config". 3. The agent responds with a roiv-cmip-confirmed-event report message with a MDC_NOTI_CONFIG event to send its configuration to the manager. 4. The Energy object shall be: <ol style="list-style-type: none"> a. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = MDC_PART_PHD_HF MDC_HF_ENERGY b. Mandatory attribute Metric-Spec_Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall (BITS-16) <input type="checkbox"/> attribute-value ≠ 0x00 0x00 <ul style="list-style-type: none"> • bit 0 (mss-avail-intermittent(0)) shall be set. • bit 1(mss-avail-stored-data(1)) shall be set. • bit 2 (mss-updt-aperiodic(2)) shall be set. • bit 3(mss-msmt-aperiodic(3)) shall be set • bit 9 (mss-acc-agent-initiated(9)) shall be set. • The other bits have to be 0. c. Mandatory attribute Unit-Code <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE <input type="checkbox"/> attribute-type = OID-Type (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_DIM_X_CAL or MDC_DIM_X_JOULES d. Mandatory attribute Source-Handle-Reference <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SOURCE_HANDLE_REF <input type="checkbox"/> attribute-type = HANDLE (NT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = It must be equal to the handle of any Session or Sub-session object in the configuration 		

	<ol style="list-style-type: none"> 5. Wait for the agent under test and the simulated manager to reach the operating state. 6. Take a measurement in the agent. 7. Wait until the manager receives an event report.
Pass/Fail criteria	<ul style="list-style-type: none"> • In step 4, all checked values are as specified. • In step 7, check that only non-negative values are used.
Notes	

TP Id	TP/PLT/AG/CLASS/CV/BV-023_A		
TP label	Energy expended, timestamp values		
Coverage	Spec	[ISO/IEEE 11073-10441]	
	Testable items	NumObj2; M	NumObj3; M
Test purpose	<p>Check that:</p> <p>Energy expended Numeric object instance shall have a timestamp identical to its containing object instance (i.e. Session or Sub-Session).</p> <p>[AND]</p> <p>The timestamp attribute used for each object shall be the same as the one used for the associated Session or Sub-Session object instance</p>		
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_025		
Other PICS			
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<ol style="list-style-type: none"> 1. Take a measurement with the agent under test. 2. Wait for the simulated manager to receive it. Record the Time Stamp and the Measure-Active-Period of the Session and Sub-Session object and of the Energy expended object. 		
Pass/Fail criteria	<ul style="list-style-type: none"> • The Timestamp attribute used for the Energy expended object shall be the same as that used for the associated Session or Sub-Session object instance. • The Energy expended instance shall have a timestamp identical to its associated Session or Sub-Session object instance. 		
Notes			

TP Id	TP/PLT/AG/CLASS/CV/BV-024			
TP label	Calories Ingested Numeric Object Attributes			
Coverage	Spec	[ISO/IEEE 11073-10441]		
	Testable items	NumObj5; M	NumObj6; M	Calories ingested1; O
		Calories ingested2; M	Calories ingested3; M	Calories ingested4; R
Calories ingested5; M		Calories ingested6; M		
Test purpose	<p>Check that:</p> <p>The Calories Ingested Numeric object contains the attributes specified for Extended Configuration.</p>			
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_024			
Other PICS				
Initial condition	The agent under test is in the unassociated state.			
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with an Association Response with result = "accepted-unknown-config". 			

	<ol style="list-style-type: none"> 3. The agent responds with a roiv-cmip-confirmed-event report message with a MDC_NOTI_CONFIG event to send its configuration to the manager. 4. The Calories Ingested object shall be: <ol style="list-style-type: none"> a. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = MDC_PART_PHD_HF MDC_HF_CAL_INGEST b. Mandatory attribute Metric-Spec_Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall (BITS-16) <input type="checkbox"/> attribute-value ≠ 0x00 0x00 <ul style="list-style-type: none"> • bit 0 (mss-avail-intermittent(0)) shall be set. • bit 1(mss-avail-stored-data(1)) shall be set. • bit 2 (mss-updt-aperiodic(2)) shall be set. • bit 3(mss-msmt-aperiodic(3)) shall be set • bit 9 (mss-acc-agent-initiated(9)) shall be set. • The other bits have to be 0. c. Not Recommended attribute Unit-Code <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE <input type="checkbox"/> attribute-type = OID-Type (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_DIM_X_CAL d. Mandatory attribute Source-Handle-Reference <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SOURCE_HANDLE_REF <input type="checkbox"/> attribute-type = HANDLE (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = It must be equal to the handle of any Session or Sub-session object in the configuration 5. Wait for the agent under test and the simulated manager to reach the operating state. 6. Take a measurement in the agent. 7. Wait until the manager receives an event report.
Pass/Fail criteria	<ul style="list-style-type: none"> • In step 4, all checked values are as specified. • In step 7, check that only non-negative values are used.
Notes	

TP Id		TP/PLT/AG/CLASS/CV/BV-024_A	
TP label		Calories ingested, timestamp values	
Coverage	Spec	[ISO/IEEE 11073-10441]	
	Testable items	NumObj2; M	NumObj3; M
Test purpose		Check that: Calories Ingested Numeric object instance shall have a timestamp identical to its containing object instance (i.e. Session or Sub-Session). [AND] The timestamp attribute used for each object shall be the same as the one used for the	

	associated Session or Sub-Session object instance
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_024
Other PICS	
Initial condition	The simulated manager and the agent under test are in the operating state.
Test procedure	<ol style="list-style-type: none"> 1. Take a measurement with the agent under test. 2. Wait for the simulated manager to receive it. Record the Time Stamp and the Measure-Active-Period of the Session and Sub-Session object and of the Calories ingested object.
Pass/Fail criteria	<ul style="list-style-type: none"> • The Timestamp attribute used for the Calories ingested object shall be the same as that used for the associated Session or Sub-Session object instance. • The Calories ingested instance shall have a timestamp identical to its associated session or sub-session object instance.
Notes	

TP Id	TP/PLT/AG/CLASS/CV/BV-025			
TP label	Carbohydrate Calories Ingested Numeric Object Attributes			
Coverage	Spec	[ISO/IEEE 11073-10441]		
	Testable items	NumObj5; M	NumObj6; M	CarbohydrateCal1; O
		CarbohydrateCal2; M	CarbohydrateCal3; M	CarbohydrateCal4; R
		CarbohydrateCal5; M	CarbohydrateCal6; M	
Test purpose	<p>Check that:</p> <p>The Carbohydrate Calories Ingested Numeric object contains the attributes specified for Extended Configuration.</p>			
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_023			
Other PICS				
Initial condition	The agent under test is in the unassociated state.			
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with an Association Response with result = "accepted-unknown-config". 3. The agent responds with a roiv-cmip-confirmed-event report message with a MDC_NOTI_CONFIG event to send its configuration to the manager. 4. The Carbohydrate Calories Ingested object shall be: <ol style="list-style-type: none"> a. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = MDC_PART_PHD_HF MDC_HF_CAL_INGEST_CARB b. Mandatory attribute Metric-Spec_Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall (BITS-16) <input type="checkbox"/> attribute-value ≠ 0x00 0x00 <ul style="list-style-type: none"> • bit 0 (mss-avail-intermittent(0)) shall be set. • bit 1(mss-avail-stored-data(1)) shall be set. • bit 2 (mss-updt-aperiodic(2)) shall be set. • bit 3(mss-msmt-aperiodic(3)) shall be set • bit 9 (mss-acc-agent-initiated(9)) shall be set. • The other bits have to be 0. 			

	<p>c. Not Recommended attribute Unit-Code</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE <input type="checkbox"/> attribute-type = OID-Type (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_DIM_X_CAL <p>d. Mandatory attribute Source-Handle-Reference</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SOURCE_HANDLE_REF <input type="checkbox"/> attribute-type = HANDLE (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = It must be equal to the handle of any Session or Sub-session object in the configuration <p>5. Wait for the agent under test and the simulated manager to reach the operating state.</p> <p>6. Take a measurement in the agent.</p> <p>7. Wait until the manager receives an event report.</p>
Pass/Fail criteria	<ul style="list-style-type: none"> • In step 4, all checked values are as specified. In step 7, check that only non-negative values are used.
Notes	

TP Id	TP/PLT/AG/CLASS/CV/BV-025_A		
TP label	Carbohydrate calories ingested, timestamp values		
Coverage	Spec	[ISO/IEEE 11073-10441]	
	Testable items	NumObj2; M	NumObj3; M
Test purpose	<p>Check that:</p> <p>Carbohydrate Calories Ingested Numeric object instance shall have a timestamp identical to its containing object instance (i.e. Session or Sub-Session).</p> <p>[AND]</p> <p>The timestamp attribute used for each object shall be the same as the one used for the associated Session or Sub-Session object instance</p>		
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_023		
Other PICS			
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<ol style="list-style-type: none"> 1. Take a measurement with the agent under test 2. Wait for the simulated manager to receive it. Record the Time Stamp and the Measure-Active-Period of the Session and Sub-Session object and of the Carbohydrate calories ingested object. 		
Pass/Fail criteria	<ul style="list-style-type: none"> • The timestamp attribute used for the Carbohydrate Calories Ingested object shall be the same as that used for the associated Session or Sub-Session object instance. • The Carbohydrate calories ingested instance shall have a timestamp identical to its associated session or sub-session object instance. 		
Notes			

TP Id	TP/PLT/AG/CLASS/CV/BV-026
TP label	Sustained Phys Activity Threshold Numeric Object Attributes

Coverage	Spec	[ISO/IEEE 11073-10441]		
	Testable items	NumObj5; M	NumObj6; M	SustainedPhysAct1; O
		SustainedPhysAct2; M	SustainedPhysAct3; M	SustainedPhysAct4; R
		SustainedPhysAct5; M	SustainedPhysAct6; M	
Test purpose	<p>Check that:</p> <p>The Sustained Phys Activity Threshold Numeric object contains the attributes specified for Extended Configuration.</p>			
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_022			
Other PICS				
Initial condition	The agent under test is in the unassociated state.			
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with an Association Response with result = "accepted-unknown-config". 3. The agent responds with a roiv-cmip-confirmed-event report message with a MDC_NOTI_CONFIG event to send its configuration to the manager. 4. The Sustained Phys Activity Threshold object shall be: <ol style="list-style-type: none"> a. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = MDC_PART_PHD_HF MDC_HF_SUS_PA_THRESHOLD b. Mandatory attribute Metric-Spec_Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall (BITS-16) <input type="checkbox"/> attribute-value ≠ 0x00 0x00 <ul style="list-style-type: none"> • bit 0 (mss-avail-intermittent(0)) shall be set. • bit 1(mss-avail-stored-data(1)) shall be set. • bit 2 (mss-updt-aperiodic(2)) shall be set. • bit 3(mss-msmt-aperiodic(3)) shall be set • bit 9 (mss-acc-agent-initiated(9)) shall be set. • The other bits have to be 0. c. Not Recommended attribute Unit-Code <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE <input type="checkbox"/> attribute-type = OID-Type (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_DIM_MIN d. Mandatory attribute Source-Handle-Reference <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SOURCE_HANDLE_REF <input type="checkbox"/> attribute-type = HANDLE (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = It must be equal to the handle of any Session or Sub-session object in the configuration 5. Wait for the agent under test and the simulated manager to reach the operating state. 6. Take a measurement in the agent. 7. Wait until the manager receives an event report. 			
Pass/Fail criteria	<ul style="list-style-type: none"> • In step 4, all checked values are as specified. 			

	<ul style="list-style-type: none"> In step 7, check that only non-negative values are used.
Notes	

TP Id	TP/PLT/AG/CLASS/CV/BV-026_A		
TP label	Sustained phys activity threshold, timestamp values		
Coverage	Spec	[ISO/IEEE 11073-10441]	
	Testable items	NumObj2; M	NumObj3; M
Test purpose	<p>Check that:</p> <p>Sustained phys activity threshold Numeric object instance shall have a timestamp identical to its containing object instance (i.e. Session or Sub-Session).</p> <p>[AND]</p> <p>The timestamp attribute used for each object shall be the same as the one used for the associated Session or Sub-Session object instance</p>		
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_022		
Other PICS			
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<ol style="list-style-type: none"> Take a measurement with the agent under test. Wait for the simulated manager to receive it. Record the Time Stamp and the Measure-Active-Period of the Session and Sub-Session object and of the Sustained phys activity threshold object. 		
Pass/Fail criteria	<ul style="list-style-type: none"> The Timestamp attribute used for the Sustained phys activity threshold object shall be the same as that used for the associated Session or Sub-Session object instance. The Sustained phys activity threshold instance shall have a timestamp identical to its associated session or sub-session object instance. 		
Notes			

TP Id	TP/PLT/AG/CLASS/CV/BV-027		
TP label	Activity Intensity Numeric Object Attributes		
Coverage	Spec	[ISO/IEEE 11073-10441]	
	Testable items	NumObj5; M	NumObj6; M
		ActIntensity2; M	ActIntensity3; M
	ActIntensity5; M	ActIntensity6; M	
Test purpose	<p>Check that:</p> <p>The Activity Intensity Numeric object contains the attributes specified for Extended Configuration.</p>		
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_021		
Other PICS			
Initial condition	The agent under test is in the unassociated state.		
Test procedure	<ol style="list-style-type: none"> The simulated manager receives an association request from the agent under test. The simulated manager responds with an Association Response with result = "accepted-unknown-config". The agent responds with a roiv-cmip-confirmed-event report message with a MDC_NOTI_CONFIG event to send its configuration to the manager. The Activity Intensity object shall be: <ol style="list-style-type: none"> Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE 		

	<ul style="list-style-type: none"> <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = MDC_PART_PHD_HF MDC_HF_ACTIVITY_INTENSITY <p>b. Mandatory attribute Metric-Spec_Small</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall (BITS-16) <input type="checkbox"/> attribute-value ≠ 0x00 0x00 <ul style="list-style-type: none"> • bit 0 (mss-avail-intermittent(0)) shall be set. • bit 1(mss-avail-stored-data(1)) shall be set. • bit 2 (mss-updt-aperiodic(2)) shall be set. • bit 3(mss-msmt-aperiodic(3)) shall be set • bit 9 (mss-acc-agent-initiated(9)) shall be set. • The other bits have to be 0. <p>c. Not Recommended attribute Unit-Code</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE <input type="checkbox"/> attribute-type = OID-Type (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_DIM_PERCENT <p>d. Mandatory attribute Source-Handle-Reference</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SOURCE_HANDLE_REF <input type="checkbox"/> attribute-type = HANDLE (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = It must be equal to the handle of any Session or Sub-session object in the configuration <p>5. Wait for the agent under test and the simulated manager to reach the operating state.</p> <p>6. Take a measurement in the agent.</p> <p>7. Wait until the manager receives an event report.</p>
Pass/Fail criteria	<ul style="list-style-type: none"> • In step 4, all checked values are as specified. • In step 7, check that only values between zero (0) and 100 are used. The observed value reported in this object is the percentage of maximal intensity effort expended during the measurement period, as defined by the associated Session or Sub-Session object.
Notes	

TP Id	TP/PLT/AG/CLASS/CV/BV-027_A		
TP Label	Activity intensity, timestamp values		
Coverage	Spec	[ISO/IEEE 11073-10441]	
	Testable items	NumObj2; M	NumObj3; M
Test purpose	Check that: Activity Intensity Numeric object instance shall have a timestamp identical to its containing object instance (i.e. Session or Sub-Session). [AND] The timestamp attribute used for each object shall be the same as the one used for the associated Session or Sub-Session object instance		
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_021		

Other PICS	
Initial condition	The simulated manager and the agent under test are in the operating state.
Test procedure	<ol style="list-style-type: none"> 1. Take a measurement with the agent under test. 2. Wait for the simulated manager to receive it. Record the timestamp and the Measure-Active-Period of the Session and Sub-Session object and of the Activity Intensity object.
Pass/Fail criteria	<ul style="list-style-type: none"> • The timestamp attribute used for the Activity Intensity object shall be the same as that used for the associated Session or Sub-Session object instance. • The Activity intensity instance shall have a timestamp identical to its associated Session or Sub-Session object instance.
Notes	

TP Id	TP/PLT/AG/CLASS/CV/BV-028			
TP label	Body Weight Numeric Object Attributes			
Coverage	Spec	[ISO/IEEE 11073-10441]		
	Testable items	NumObj5; M	NumObj6; M	BodyWeight1; O
		BodyWeight2; M	BodyWeight3; M	BodyWeight4; M
BodyWeight5; M		BodyWeight6; M		
Test purpose	<p>Check that:</p> <p>The Body Weight Numeric object contains the attributes specified for Extended Configuration.</p>			
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_020			
Other PICS				
Initial condition	The agent under test is in the unassociated state.			
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with an Association Response with result = "accepted-unknown-config". 3. The agent responds with a roiv-cmip-confirmed-event report message with a MDC_NOTI_CONFIG event to send its configuration to the manager. 4. The Body Weight object shall be: <ol style="list-style-type: none"> a. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = MDC_PART_SCADA MDC_MASS_BODY_ACTUAL b. Mandatory attribute Metric-Spec_Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall (BITS-16) <input type="checkbox"/> attribute-value ≠ 0x00 0x00 <ul style="list-style-type: none"> • bit 0 (mss-avail-intermittent(0)) shall be set. • bit 1(mss-avail-stored-data(1)) shall be set. • bit 2 (mss-updt-aperiodic(2)) shall be set. • bit 3(mss-msmt-aperiodic(3)) shall be set • bit 9 (mss-acc-agent-initiated(9)) shall be set. • The other bits have to be 0. c. Mandatory attribute Unit-Code <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE 			

	<ul style="list-style-type: none"> <input type="checkbox"/> attribute-type = OID-Type (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_DIM_X_G or MDC_DIM_X_LB <p>d. Mandatory attribute Source-Handle-Reference</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SOURCE_HANDLE_REF <input type="checkbox"/> attribute-type = HANDLE (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = It must be equal to the handle of any Session or Sub-Session object in the configuration <p>5. Wait for the agent under test and the simulated manager to reach the operating state.</p> <p>6. Take a measurement in the agent.</p> <p>7. Wait until the manager receives an event report.</p>
Pass/Fail criteria	<ul style="list-style-type: none"> • In step 4, all checked values are as specified. • In step 7, check that only non-negative values are used.
Notes	

TP Id	TP/PLT/AG/CLASS/CV/BV-028_A		
TP label	Body weight, timestamp values		
Coverage	Spec	[ISO/IEEE 11073-10441]	
	Testable items	NumObj2; M	NumObj3; M
Test purpose	<p>Check that:</p> <p>Body Weight Numeric object instance shall have a timestamp identical to its containing object instance (i.e. Session or Sub-Session).</p> <p>[AND]</p> <p>The timestamp attribute used for each object shall be the same as the one used for the associated Session or Sub-Session object instance</p>		
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_020		
Other PICS			
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<ol style="list-style-type: none"> 1. Take a measurement with the agent under test. 2. Wait for the simulated manager to receive it. Record the Time Stamp and the Measure-Active-Period of the Session and Sub-Session object and of the Body Weight object. 		
Pass/Fail criteria	<ul style="list-style-type: none"> • The Timestamp attribute used for the Body Weight object shall be the same as that used for the associated Session or Sub-Session object instance. • Body weight instance shall have a timestamp identical to its associated Session or Sub-Session object instance. 		
Notes			

TP Id	TP/PLT/AG/CLASS/CV/BV-029			
TP label	Height Numeric Object Attributes			
Coverage	Spec	[ISO/IEEE 11073-10441]		
	Testable items	NumObj5; M	NumObj6; M	Height1; O
		Height2; M	Height3; M	Height4; M
		Height5; M	Height6; M	
Test purpose	Check that:			

	The Body Height Numeric object contains the attributes specified for Extended Configuration.
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_019
Other PICS	
Initial condition	The agent under test is in the unassociated state.
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with an Association Response with result = "accepted-unknown-config". 3. The agent responds with a roiv-cmip-confirmed-event report message with a MDC_NOTI_CONFIG event to send its configuration to the manager. 4. The Height object shall be: <ol style="list-style-type: none"> a. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = MDC_PART_SCADA MDC_LEN_BODY_ACTUAL b. Mandatory attribute Metric-Spec_Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall (BITS-16) <input type="checkbox"/> attribute-value ≠ 0x00 0x00 <ul style="list-style-type: none"> • bit 0 (mss-avail-intermittent(0)) shall be set. • bit 1(mss-avail-stored-data(1)) shall be set. • bit 2 (mss-updt-aperiodic(2)) shall be set. • bit 3(mss-msmt-aperiodic(3)) shall be set • bit 9 (mss-acc-agent-initiated(9)) shall be set. • The other bits have to be 0. c. Mandatory attribute Unit-Code <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE <input type="checkbox"/> attribute-type = OID-Type (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_DIM_X_M or MDC_DIM_X_FOOT d. Mandatory attribute Source-Handle-Reference <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SOURCE_HANDLE_REF <input type="checkbox"/> attribute-type = HANDLE (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = It must be equal to the handle of any Session or Sub-session object in the configuration 5. Wait for the agent under test and the simulated manager to reach the operating state. 6. Take a measurement in the agent. 7. Wait until the manager receives an event report.
Pass/Fail criteria	<ul style="list-style-type: none"> • In step 4, all checked values are as specified. • In step 7, check that only non-negative values are used.
Notes	

TP Id	TP/PLT/AG/CLASS/CV/BV-029_A
TP label	Height, timestamp values

Coverage	Spec	[ISO/IEEE 11073-10441]		
	Testable items	NumObj2; M	NumObj3; M	
Test purpose	<p>Check that:</p> <p>Body Height Numeric object instance shall have a timestamp identical to its containing object instance (i.e. Session or Sub-Session).</p> <p>[AND]</p> <p>The timestamp attribute used for each object shall be the same as the one used for the associated Session or Sub-Session object instance</p>			
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_019			
Other PICS				
Initial condition	The simulated manager and the agent under test are in the operating state.			
Test procedure	<ol style="list-style-type: none"> 1. Take a measurement with the agent under test. 2. Wait for the simulated manager to receive it. Record the Time Stamp and the Measure-Active-Period of the Session and Sub-Session object and of the Height object. 			
Pass/Fail criteria	<ul style="list-style-type: none"> • The timestamp attribute used for the Height object shall be the same as that used for the associated Session or Sub-Session object instance. • The Height instance shall have a timestamp identical to its associated Session or Sub-Session object instance. 			
Notes				

TP Id	TP/PLT/AG/CLASS/CV/BV-030			
TP label	Age Numeric Object Attributes			
Coverage	Spec	[ISO/IEEE 11073-10441]		
	Testable items	NumObj5; M	NumObj6; M	Age1; O
		Age2; M	Age3; M	Age4; R
		Age5; M	Age6; M	
Test purpose	<p>Check that:</p> <p>The Age Numeric object contains the attributes specified for Extended Configuration.</p>			
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_018			
Other PICS				
Initial condition	The agent under test is in the unassociated state.			
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with an Association Response with result = "accepted-unknown-config". 3. The agent responds with a roiv-cmip-confirmed-event report message with a MDC_NOTI_CONFIG event to send its configuration to the manager. 4. The Age object shall be: <ol style="list-style-type: none"> a. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = MDC_PART_PHD_HF MDC_HF_AGE b. Mandatory attribute Metric-Spec_Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall (BITS-16) <input type="checkbox"/> attribute-value ≠ 0x00 0x00 			

	<ul style="list-style-type: none"> • bit 0 (mss-avail-intermittent(0)) shall be set. • bit 1(mss-avail-stored-data(1)) shall be set. • bit 2 (mss-updt-aperiodic(2)) shall be set. • bit 3(mss-msmt-aperiodic(3)) shall be set • bit 9 (mss-acc-agent-initiated(9)) shall be set. • The other bits have to be 0. <p>c. Not Recommended attribute Unit-Code</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE <input type="checkbox"/> attribute-type = OID-Type (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_DIM_YR <p>d. Mandatory attribute Source-Handle-Reference</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SOURCE_HANDLE_REF <input type="checkbox"/> attribute-type = HANDLE (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = It must be equal to the handle of any Session or Sub-Session object in the configuration <p>5. Wait for the agent under test and the simulated manager to reach the operating state.</p> <p>6. Take a measurement in the agent.</p> <p>7. Wait until the manager receives an event report.</p>
Pass/Fail criteria	<ul style="list-style-type: none"> • In step 4, all checked values are as specified. • In step 7, check that only non-negative values are used.
Notes	

TP Id	TP/PLT/AG/CLASS/CV/BV-030_A		
TP label	Age, timestamp values		
Coverage	Spec	[ISO/IEEE 11073-10441]	
	Testable items	NumObj2; M	NumObj3; M
Test purpose	<p>Check that:</p> <p>Age Numeric object instance shall have a timestamp identical to its containing object instance (i.e. Session or Sub-Session).</p> <p>[AND]</p> <p>The timestamp attribute used for each object shall be the same as the one used for the associated Session or Sub-Session object instance</p>		
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_018		
Other PICS			
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<ol style="list-style-type: none"> 1. Take a measurement with the agent under test. 2. Wait for the simulated manager to receive it. Record the Time Stamp and the Measure-Active-Period of the Session and Sub-Session object and of the Age object. 		
Pass/Fail criteria	<ul style="list-style-type: none"> • The Timestamp attribute used for Age object shall be the same as that used for the associated Session or Sub-Session object instance. • The Age instance shall have a timestamp identical to its associated Session or Sub-Session object instance. 		
Notes			

TP Id		TP/PLT/AG/CLASS/CV/BV-031		
TP label		Session Enumeration Object Attributes		
Coverage	Spec	[ISO/IEEE 11073-10441]		
	Testable items	Session1; M	Session2; M	Session3; M
		Session4; R	Session5; R	Session6; M
		Session7; M	Session8; R	Session9; R
		Session11; M	Session12; M	
Test purpose		<p>Check that:</p> <p>The Session Enumeration object contains the attributes specified for Extended Configuration.</p>		
Applicability		C_AG_OXP_000 AND C_AG_OXP_172		
Other PICS				
Initial condition		The agent under test is in the unassociated state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with an Association Response with result = "accepted-unknown-config". 3. The agent responds with a roiv-cmip-confirmed-event report message with a MDC_NOTI_CONFIG event to send its configuration to the manager. 4. The Session object shall be: <ol style="list-style-type: none"> a. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value.length = Sequence of partition (NomPartition (INT-U16)) and code (OID-Type) <input type="checkbox"/> attribute-value = MDC_PART_PHD_HF MDC_HF_SESSION b. Mandatory attribute Metric-Spec_Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall (BITS-16) <input type="checkbox"/> attribute-value ≠ 0x00 0x00 <ul style="list-style-type: none"> • bit 0 (mss-avail-intermittent(0)) shall be set. • bit 1(mss-avail-stored-data(1)) shall be set. • bit 2 (mss-updt-aperiodic(2)) shall be set. • bit 3(mss-msmt-aperiodic(3)) shall be set • bit 9 (mss-acc-agent-initiated(9)) shall be set. • The other bits have to be 0. c. Not Recommended attribute Unit-Code <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE <input type="checkbox"/> attribute-type = OID-Type (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes d. Not Recommended attribute Unit-LabelString <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_LABEL_STRING <input type="checkbox"/> attribute-type = OCTET STRING <input type="checkbox"/> attribute-value.length = <variable> e. Optional Label-String: 		

	<ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_LABEL_STRING <input type="checkbox"/> attribute-type = OCTET STRING <input type="checkbox"/> attribute-value = If an existing acceptable nomenclature term (for activity defined in Enum-Observed-Value-Simple-Oid) is not available → attribute-value = MDC_HF_ACT_UNKONWN and an appropriate clarifying text in the Label-String attribute <p>f. Mandatory attribute Measure-Active-Period</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_PD_MSMT_ACTIVE <input type="checkbox"/> attribute-type = FLOAT-Type (INT-U32) <input type="checkbox"/> attribute-value.length = 4 bytes <p>g. Mandatory attribute Enum-Observed-Value-Simple-OID</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ENUM_OBS_VAL_SIM_OID <input type="checkbox"/> attribute-type = OID-Type (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <ul style="list-style-type: none"> • Valid values→ MDC_HF_ACT_UNKNOWN or MDC_HF_ACT_MONITOR or MDC_HF_ACT_SKI or MDC_HF_ACT_RUN or MDC_HF_ACT_BIKE or MDC_HF_ACT_STAIR or MDC_HF_ACT_ROW or MDC_HF_ACT_HOME or MDC_HF_ACT_WORK or MDC_HF_ACT_WALK • If there are multiple sub-sessions associated: attribute-value = MDC_HF_ACT_MULTIPLE <p>h. Not Recommended attribute Enum-Observed-Value-Simple-Str</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id= MDC_ATTR_ENUM_OBS_VAL_SIM_STR <input type="checkbox"/> attribute-type = EnumPrintableString <input type="checkbox"/> attribute-value.length= <variable> <p>i. Not Recommended attribute Enum-Observed-Value</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id= MDC_ATTR_VAL_ENUM_OBS <input type="checkbox"/> attribute-type = EnumObsValue <input type="checkbox"/> attribute-value.length= <variable>
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/AG/CLASS/CV/BV-032		
TP label	Session and associated Sub-Session 1		
Coverage	Spec	[ISO/IEEE 11073-10441]	
	Testable items	Sub-session10; M	Sub-session13; M
Test purpose	Check that: The timestamp attribute of the Sub-Session shall fall within the time span specified by the Session to which it is associated. [AND] Metrics that represent observations for the Sub-Session shall have a timestamp equal to the associated Sub-Session object's timestamp.		
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_043		
Other PICS			
Initial condition	The agent under test is in the operating state.		
Test procedure	1. Take Measurements for the Session and Sub-Session Objects in the agent under test.		

	2. Wait to receive event reports and record the Session and the Sub-session objects for later comparison.
Pass/Fail criteria	<ul style="list-style-type: none"> If the Session object has a timestamp, the associated Sub-Session objects shall have the same type of timestamp.
Notes	

TP Id	TP/PLT/AG/CLASS/CV/BV-032_A		
TP label	Session and associated Sub-Session 2		
Coverage	Spec	[ISO/IEEE 11073-10441]	
	Testable items	Session13; M	Session14; M
Test purpose	<p>Check that:</p> <p>All Sub-Sessions that are contained by the Session shall have a timestamp that falls within the time span that begins with the Session's timestamp and lasts for the Measure-Active-Period attribute.</p> <p>[AND]</p> <p>The sum of all contained Sub-Session Measure-Active-Period attributes shall be equal to the Measure-Active-Period attribute of the containing Session</p>		
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_043		
Other PICS			
Initial condition	The agent under test is in the operating state.		
Test procedure	<ol style="list-style-type: none"> Take Measurements for the Session and Sub-Session Objects in the agent under test. Wait to receive event reports and record the Session and the Sub-Session objects for later comparison. 		
Pass/Fail criteria	<ul style="list-style-type: none"> The sum of the Measure-Active-Period of the Sub-Sessions has to be equal to the Measure-Active-Period of the Session. If the Sub-session objects have a timestamp, then it shall fall in the period defined between the timestamp and the Measure-Active-Period of the session object. 		
Notes			

TP Id	TP/PLT/AG/CLASS/CV/BV-033			
TP label	Sub-Session Enumeration Object Attributes			
Coverage	Spec	[ISO/IEEE 11073-10441]		
	Testable items	Session16; M	Sub-session1; O	Sub-session2; M
		Sub-session3; M	Sub-session4; R	Sub-session5; R
		Sub-session6; M	Sub-session7; M	Sub-session8; R
Sub-session9; R		Sub-session11; R	Sub-session14; M	
Test purpose	<p>Check that:</p> <p>The Sub-Session Enumeration object contains the attributes specified for Extended Configuration.</p>			
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_043			
Other PICS				
Initial condition	The agent under test is in the unassociated state.			
Test procedure	<ol style="list-style-type: none"> The simulated manager receives an association request from the agent under test. The simulated manager responds with an Association Response with result = "accepted-unknown-config". The agent responds with a roiv-cmip-confirmed-event report message with a 			

MDC_NOTI_CONFIG event to send its configuration to the manager.

4. The Sub-Session object shall be:
 - a. Mandatory attribute Type
 - attribute-id = MDC_ATTR_ID_TYPE
 - attribute-type = TYPE
 - attribute-value.length = Sequence of partition (NomPartition (INT-U16)) and code (OID-Type)
 - attribute-value = MDC_PART_PHD_HF | MDC_HF_SUBSESSION
 - b. Mandatory attribute Metric-Spec_Small
 - attribute-id = MDC_ATTR_METRIC_SPEC_SMALL
 - attribute-type = MetricSpecSmall (BITS-16)
 - attribute-value ≠ 0x00 0x00
 - bit 0 (mss-avail-intermittent(0)) shall be set.
 - bit 1 (mss-avail-stored-data(1)) shall be set.
 - bit 2 (mss-updt-aperiodic(2)) shall be set.
 - bit 3 (mss-msmt-aperiodic(3)) shall be set
 - bit 9 (mss-acc-agent-initiated(9)) shall be set.
 - The other bits have to be 0.
 - c. Not Recommended attribute Unit-Code
 - attribute-id = MDC_ATTR_UNIT_CODE
 - attribute-type = OID-Type (INT-U16)
 - attribute-value.length = 2 bytes
 - d. Not Recommended attribute Unit-LabelString
 - attribute-id = MDC_ATTR_UNIT_LABEL_STRING
 - attribute-type = OCTET STRING
 - attribute-value.length = <variable>
 - j. Optional Label-String:
 - attribute-id = MDC_ATTR_UNIT_LABEL_STRING
 - attribute-type = OCTET STRING
 - attribute-value = If an existing acceptable nomenclature term (for activity defined in Enum-Observed-Value-Simple-Oid) is not available → attribute-value = MDC_HF_ACT_UNKONWN and an appropriate clarifying text in the Label-String attribute
 - e. Mandatory attribute Measure-Active-Period
 - attribute-id = MDC_ATTR_TIME_PD_MSMT_ACTIVE
 - attribute-type = FLOAT-Type (INT-U32)
 - attribute-value.length = 4 bytes
 - f. Mandatory attribute Enum-Observed-Value-Simple-OID
 - attribute-id = MDC_ATTR_ENUM_OBS_VAL_SIM_OID
 - attribute-type = OID-Type (INT-U16)
 - attribute-value.length = 2 bytes
 - attribute-value = MDC_HF_ACT_UNKNOWN or MDC_HF_ACT_MONITOR or MDC_HF_ACT_SKI or MDC_HF_ACT_RUN or MDC_HF_ACT_BIKE or MDC_HF_ACT_STAIR or MDC_HF_ACT_ROW or MDC_HF_ACT_HOME or MDC_HF_ACT_WORK or MDC_HF_ACT_WALK
 - g. Not Recommended attribute Enum-Observed-Value-Simple-Str
 - attribute-id= MDC_ATTR_ENUM_OBS_VAL_SIM_STR

	<ul style="list-style-type: none"> <input type="checkbox"/> attribute-type = EnumPrintableString <input type="checkbox"/> attribute-value.length= <variable> <p>h. Not Recommended attribute Enum-Observed-Value</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id= MDC_ATTR_VAL_ENUM_OBS <input type="checkbox"/> attribute-type = EnumObsValue <input type="checkbox"/> attribute-value.length= <variable>
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	http://continua.plugfests.com/show_bug.cgi?id=448

TP Id	TP/PLT/AG/CLASS/CV/BV-034			
TP label	Activity Time Object Attributes			
Coverage	Spec	[ISO/IEEE 11073-10441]		
	Testable items	Session16; M	ActivityTime1; O	ActivityTime2; M
		ActivityTime3; M	ActivityTime4; R	ActivityTime5; R
		ActivityTime6; R	ActivityTime7; M	ActivityTime8; M
		ActivityTime9; R	ActivityTime10; R	ActivityTime11; M
ActivityTime12; M				
Test purpose	<p>Check that:</p> <p>The Activity Time Enumeration Object contains the attributes specified for Extended Configuration.</p>			
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_017			
Other PICS				
Initial condition	The agent under test is in the unassociated state.			
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with an Association Response with result = "accepted-unknown-config". 3. The agent responds with a roiv-cmip-confirmed-event report message with a MDC_NOTI_CONFIG event to send its configuration to the manager. 4. The Activity Time object shall be: <ol style="list-style-type: none"> a. Mandatory attribute Type <ol style="list-style-type: none"> a. attribute-id = MDC_ATTR_ID_TYPE b. attribute-type = TYPE c. attribute-value.length =Sequence of partition (NomPartition (INT-U16)) and code (OID-Type)) d. attribute-value = MDC_PART_PHD_HF MDC_HF_ACTIVITY_TIME b. Mandatory attribute Metric-Spec_Small <ol style="list-style-type: none"> a. attribute-id = MDC_ATTR_METRIC_SPEC_SMALL b. attribute-type = MetricSpecSmall (BITS-16) c. attribute-value ≠ 0x00 0x00 <ul style="list-style-type: none"> • bit 0 (mss-avail-intermittent(0)) shall be set. • bit 1(mss-avail-stored-data(1)) shall be set. • bit 2 (mss-updt-aperiodic(2)) shall be set. • bit 3(mss-msmt-aperiodic(3)) shall be set • bit 9 (mss-acc-agent-initiated(9)) shall be set. • The other bits have to be 0. 			

	<ul style="list-style-type: none"> c. Not Recommended attribute Unit-Code <ul style="list-style-type: none"> a. attribute-id = MDC_ATTR_UNIT_CODE b. attribute-type = OID-Type (INT-U16) c. attribute-value.length = 2 bytes d. Mandatory attribute Source-Handle-Reference <ul style="list-style-type: none"> a. attribute-id = MDC_ATTR_SOURCE_HANDLE_REF b. attribute-type = HANDLE (INT-U16) c. attribute-value = It must be equal to the handle of any Session or Sub-session object in the configuration e. Not Recommended attribute Absolute-Time-Stamp <ul style="list-style-type: none"> a. attribute-id = MDC_ATTR_TIME_STAMP_ABS b. attribute-type = AbsoluteTime c. attribute-value.length = 8 bytes f. Not Recommended attribute Unit-LabelString <ul style="list-style-type: none"> a. attribute-id = MDC_ATTR_UNIT_LABEL_STRING b. attribute-type = OCTET STRING c. attribute-value.length = <variable> d. If an existing acceptable nomenclature term (for activity defined in Enum-Observed-Value-Simple-Oid) is not available → attribute-value = MDC_HF_ACT_UNKONWN g. Mandatory attribute Measure-Active-Period <ul style="list-style-type: none"> a. attribute-id = MDC_ATTR_TIME_PD_MSMT_ACTIVE b. attribute-type = FLOAT-Type (INT-U32) c. attribute-value.length = 4 bytes h. Mandatory attribute Enum-Observed-Value-Simple-OID <ul style="list-style-type: none"> a. attribute-id = MDC_ATTR_ENUM_OBS_VAL_SIM_OID b. attribute-type = OID-Type (INT-U16) c. attribute-value.length = 2 bytes d. attribute-value = MDC_HF_ACT_AMB or MDC_HF_ACT_REST, MDC_HF_ACT_MOTOR or MDC_HF_ACT_LYING or MDC_HF_ACT_SPEEP or MDC_HF_ACT_PHYS or MDC_HF_ACT_SUS_PHYS or MDC_HF_ACT_UNKNOWN i. Not Recommended attribute Enum-Observed-Value-Simple-Str <ul style="list-style-type: none"> a. attribute-id= MDC_ATTR_ENUM_OBS_VAL_SIM_STR b. attribute-type = EnumPrintableString c. attribute-value.length= <variable> j. Not Recommended attribute Enum-Observed-Value <ul style="list-style-type: none"> a. attribute-id= MDC_ATTR_VAL_ENUM_OBS b. attribute-type = EnumObsValue c. attribute-value.length= <variable>
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/AG/CLASS/CV/BV-034_A	
TP label	Activity time, timestamp values	
Coverage	Spec	[ISO/IEEE 11073-10441]

	Testable items	Session15; M	ActivityTime11; M	
Test purpose		Check that: Metrics that represent observations for the Session as a whole shall have a timestamp equal to the associated Session object's timestamp.		
Applicability		C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_017		
Other PICS				
Initial condition		The simulated manager and the agent under test are in the operating state.		
Test procedure		<ol style="list-style-type: none"> 1. Take a measurement with the agent under test. 2. Wait for the simulated manager to receive it. Record the timestamp and the Measure-Active-Period of the Session and Sub-Session object and of the Activity time object. 		
Pass/Fail criteria		<ul style="list-style-type: none"> • The timestamp attribute used for the Activity time object shall be the same as that used for the associated Session or Sub-Session object instance. • The Activity time instance shall have a timestamp identical to its associated session or sub-session object instance. 		
Notes				

TP Id		TP/PLT/AG/CLASS/CV/BV-035		
TP label		Program Identifier Object Attributes		
Coverage	Spec	[ISO/IEEE 11073-10441]		
	Testable items	Session16; M	ProgramId1; O	ProgramId2; M
		ProgramId3; M	ProgramId4; R	ProgramId5; R
		ProgramId6; R	ProgramId7; R	ProgramId8; M
		ProgramId9; R	ProgramId10; M	
Test purpose		Check that: The Program Identifier Enumeration object contains the attributes specified for Extended Configuration.		
Applicability		C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_016		
Other PICS				
Initial condition		The agent under test is in the unassociated state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with an Association Response with result = "accepted-unknown-config". 3. The agent responds with a roiv-cmip-confirmed-event report message with a MDC_NOTI_CONFIG event to send its configuration to the manager. 4. The Program Identifier object shall be: <ol style="list-style-type: none"> a. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value.length =Sequence of partition (NomPartition (INT-U16)) and code (OID-Type) <input type="checkbox"/> attribute-value = MDC_PART_PHD_HF MDC_HF_PROGRAM_ID b. Mandatory attribute Metric-Spec_Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall (BITS-16) <input type="checkbox"/> attribute-value ≠ 0x00 0x00 <ul style="list-style-type: none"> • bit 0 (mss-avail-intermittent(0)) shall be set. 		

	<ul style="list-style-type: none"> • bit 1(mss-avail-stored-data(1)) shall be set. • bit 2 (mss-updt-aperiodic(2)) shall be set. • bit 3(mss-msmt-aperiodic(3)) shall be set • bit 9 (mss-acc-agent-initiated(9)) shall be set. • The other bits have to be 0. <p>c. Not Recommended attribute Unit-Code</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE <input type="checkbox"/> attribute-type = OID-Type (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <p>d. Mandatory attribute Source-Handle-Reference</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SOURCE_HANDLE_REF <input type="checkbox"/> attribute-type = HANDLE (INT-U16) <input type="checkbox"/> attribute-value = It must be equal to the handle of any Session or Sub-session object in the configuration <p>e. Not Recommended attribute Absolute-Time-Stamp</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_STAMP_ABS <input type="checkbox"/> attribute-type = AbsoluteTime <input type="checkbox"/> attribute-value.length = 8 bytes <p>f. Not Recommended attribute Unit-LabelString</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_LABEL_STRING <input type="checkbox"/> attribute-type = OCTET STRING <input type="checkbox"/> attribute-value.length = <variable> <p>g. Mandatory attribute Measure-Active-Period</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_PD_MSMT_ACTIVE <input type="checkbox"/> attribute-type = FLOAT-Type (INT-U32) <input type="checkbox"/> attribute-value.length = 4 bytes <p>h. Not Recommended attribute Enum-Observed-Value-Simple-OID</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ENUM_OBS_VAL_SIM_OID <input type="checkbox"/> attribute-type = OID-Type (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <p>i. Mandatory attribute Enum-Observed-Value-Simple-Str</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id= MDC_ATTR_ENUM_OBS_VAL_SIM_STR <input type="checkbox"/> attribute-type = EnumPrintableString <input type="checkbox"/> attribute-value.length= <variable> <p>j. Not Recommended attribute Enum-Observed-Value</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id= MDC_ATTR_VAL_ENUM_OBS <input type="checkbox"/> attribute-type = EnumObsValue <input type="checkbox"/> attribute-value.length= <variable>
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/AG/CLASS/CV/BV-035_A
TP label	Program identifier, timestamp values

Coverage	Spec	[ISO/IEEE 11073-10441]		
	Testable items	Session15; M	ProgramId10; M	
Test purpose	Check that: Metrics that represent observations for the Session as a whole shall have a timestamp equal to the associated Session object's timestamp.			
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_016			
Other PICS				
Initial condition	The simulated manager and the agent under test are in the operating state.			
Test procedure	<ol style="list-style-type: none"> 1. Take a measurement with the agent under test. 2. Wait for the simulated manager to receive it. Record the timestamp and the Measure-Active-Period of the Session and Sub-Session object and of the Program Identifier object. 			
Pass/Fail criteria	<ul style="list-style-type: none"> • The timestamp attribute used for Program identifier object shall be the same as the used for the associated Session or Sub-Session object instance. • The Program identifier instance shall have a timestamp identical to its associated session or sub-session object instance. 			
Notes				

TP Id	TP/PLT/AG/CLASS/CV/BV-036			
TP label	Association Request			
Coverage	Spec	[ISO/IEEE 11073-10441]		
	Testable items	MDSMethod4; M	AssocReq1; M	AssocReq2; M
		AssocReq3; M	AssocReq4; M	AssocReq5; M
		AssocReq6; M	AssocReq7; M	AssocReq8; M
		AssocReq9; M	AssocReq10; M	AssocReq11; M
AssocReq12; M				
Test purpose	Check that: The association procedure data exchange is correct.			
Applicability	C_AG_OXP_000 AND C_AG_OXP_172			
Other PICS	C_AG_OXP_017			
Initial condition	The simulated manager and the agent under test are in the unassociated state.			
Test procedure	<ol style="list-style-type: none"> 1. The agent under test sends an AARQ message to the simulated manager. The expected fields sent by the agent are: <ol style="list-style-type: none"> a. APDU Type <ul style="list-style-type: none"> <input type="checkbox"/> field-length =2 bytes <input type="checkbox"/> field-value =0xE2 0x00 (AareApdu) b. assoc-version <ul style="list-style-type: none"> <input type="checkbox"/> field-type = AssociationVersion <input type="checkbox"/> field-length =BITS-32 <input type="checkbox"/> field-value =0x80 0x00 0x00 0x00 <input type="checkbox"/> assoc-version = 0x80 0x00 0x00 0x00 (asassoc-version1(0) set) indicates that version 1 of the association protocol is supported. c. data-proto-id <ul style="list-style-type: none"> <input type="checkbox"/> field-type = DataProtold <input type="checkbox"/> field-length =INT-U16 			

- field-value = 0x50 0x79 (20601)
- data-proto-id=20601 indicates exchange protocol follows this standard, and data-proto-info field shall contain PhdAssociationInformation.

d. protocol-version

- field-type = Protocol Version
- field-length =BITS-32
- field-value = 0x80 0x00 0x00 0x00
- This value shows that version 1 of the data exchange protocol is supported (assoc-version1(0)=1),.

e. encoding rules

- field-type = EncodingRules
- field-length = BITS-16
- field-value= depends on the encoding rules supported/selected. mder(0) always is set (MDER always is supported) and xer(1) or/and per(2) may be set (optional).

f. nomenclature version

- field-type = NomenclatureVersion
- field-length =BITS-32
- field-value = 0x80 0x00 0x00 0x00
- This value indicates version 1 is supported (nom-version1(0) is set).

g. functional-units

- field-type = FunctionalUnits
- field-length = BITS-32
- If the agent has no Test Association capabilities: field-value = 0x00 0x00 0x00 0x00
- If the agent has tested capabilities that can be used within Test Association: field-value = 0x40 0x00 0x00 0x00
- If the agent has tested capabilities that can be used within Test Association and requires that the Manager establishes a Test Association: field-value = 0x60 0x00 0x00 0x00

h. system type

- field-type = SystemType
- field-length = BITS-32
- field- value = 0x00 0x80 0x00 0x00 (sys-type-agent)

i. system-id

- field-type = OCTET STRING
- field-length = 0x00 0x0A
- field-value = 0xXX 0xXX 0xXX 0xXX 0xXX 0xXX 0xXX 0xXX (EUI-64 manufacturer and device). This value will be System Id attribute of MDS Object.

j. dev-config-id

- field-type = ConfigId
- field-length = INT-U16
- field-value = <between 0x40 0x00 and 0x7F 0xFF>

k. data-req-mode-flags (DataReqModeCapab):

- field-type = DataReqModeFlags (BITS-16)
- field-length = INT-U16
- If the Agent implements only this Device Specialization: field-value = 0x00 0x01 – Agent initiated data request/flows

	l. data-req-init-agent-count (DataReqModeCapab) <ul style="list-style-type: none"> <input type="checkbox"/> field-type = INT-U8 <input type="checkbox"/> field-length = 1 byte <input type="checkbox"/> If the Agent implements only this Device Specialization: field-value = 0x01 m. data-req-init-manager-count (DataReqModeCapab) <ul style="list-style-type: none"> <input type="checkbox"/> field-type = INT-U8 <input type="checkbox"/> field-length = 1 byte <input type="checkbox"/> If the Agent implements only this Device Specialization: field-value = 0x00
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/AG/CLASS/CV/BV-041		
TP label	Config Changes Service. Altitude Gain Contextual Attribute.		
Coverage	Spec	[ISO/IEEE 11073-10441]	
	Testable items	NumObj1; M	
	Spec	[ITU-T H.810 (2015)]	
	Testable items	Communication 8; M	
Test purpose	Check that: Whenever a Contextual Attribute changes, the Agent shall report these changes to the Manager using an MDS object event prior to reporting any of the dependent values.		
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_015 AND C_AG_CV_044		
Other PICS			
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<ol style="list-style-type: none"> 1. If the attribute that is going to be changed is reported in a Fixed format event report, take some measurements with the agent under test. 2. Make a change to the Contextual Attribute Unit-Code for Altitude Gain Object (meters to feet or feet to meters.) 3. The agent shall send an MDS event report indicating the new Contextual Attribute value. 4. Take some more measurements. 5. Wait for the manager to receive new event reports from the agent, which report the measurements from step 4. 		
Pass/Fail criteria	<ul style="list-style-type: none"> • The agent sends an MDS event report to inform about the Contextual Attribute that has been changed. • Data has changed accordingly to a new Contextual Attribute. 		
Notes			

TP Id	TP/PLT/AG/CLASS/CV/BV-042		
TP label	Config Changes Service. Altitude Loss Contextual Attribute		
Coverage	Spec	[ISO/IEEE 11073-10441]	
	Testable items	NumObj1; M	
	Spec	[ITU-T H.810 (2015)]	
	Testable items	Communication 8; M	

Test purpose	Check that: Whenever a Contextual Attribute changes, the Agent shall report these changes to the Manager using an MDS object event prior to reporting any of the dependent values.
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_032 AND C_AG_CV_045
Other PICS	
Initial condition	The simulated manager and the agent under test are in the operating state.
Test procedure	<ol style="list-style-type: none"> 1. If the attribute that is going to be changed is reported in a Fixed format event report, take some measurements with the agent under test. 2. Make a change to the Contextual Attribute Unit-Code for the Altitude Loss object (meters to feet or feet to meters). 3. The agent shall send an MDS event report indicating the new Contextual Attribute value. 4. Take some more measurements. 5. Wait for the manager to receive new event reports from the agent, which report the measurements from step 4.
Pass/Fail criteria	<ul style="list-style-type: none"> • The agent sends an MDS event report to inform about the Contextual Attribute that has been changed. • Data has changed accordingly to a new Contextual Attribute.
Notes	

TP Id	TP/PLT/AG/CLASS/CV/BV-043		
TP label	Config Changes Service. Altitude Contextual Attribute.		
Coverage	Spec	[ISO/IEEE 11073-10441]	
	Testable items	NumObj1; M	
	Spec	[ITU-T H.810 (2015)]	
	Testable items	Communication 8; M	
Test purpose	Check that: Whenever a Contextual Attribute changes, the Agent shall report these changes to the Manager using an MDS object event prior to reporting any of the dependent values.		
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_039 AND C_AG_CV_046		
Other PICS			
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<ol style="list-style-type: none"> 1. If the attribute that is going to be changed is reported in a Fixed format event report, take some measurements with the agent under test. 2. Make a change to the Contextual Attribute Unit-Code for the Altitude Object (meters to feet or feet to meters). 3. The agent shall send an MDS event report indicating the new Contextual Attribute value. 4. Take some more measurements. 5. Wait for the manager to receive new event reports from the agent, which report the measurements from step 4. 		
Pass/Fail criteria	<ul style="list-style-type: none"> • The agent sends an MDS event report to inform about the Contextual Attribute that has been changed. • Data has changed accordingly to a new Contextual Attribute. 		
Notes			

TP Id		TP/PLT/AG/CLASS/CV/BV-044		
TP label		Config Changes Service. Distance Contextual Attribute.		
Coverage	Spec	[ISO/IEEE 11073-10441]		
	Testable items	NumObj1; M		
	Spec	[ITU-T H.810 (2015)]		
	Testable items	Communication 8; M		
Test purpose		<p>Check that:</p> <p>Whenever a Contextual Attribute changes, the Agent shall report these changes to the Manager using an MDS object event prior to reporting any of the dependent values.</p>		
Applicability		C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_040 AND C_AG_CV_047		
Other PICS				
Initial condition		The simulated manager and the agent under test are in the operating state.		
Test procedure		<ol style="list-style-type: none"> 1. If the attribute that is going to be changed is reported in a Fixed format event report, take some measurements with the agent under test. 2. Make a change to the Contextual Attribute Unit-Code for Distance object (meters to feet, feet to meters, meters to steps, steps to meters, feet to steps or steps to feet). 3. The agent shall send an MDS event report indicating the new Contextual Attribute value. 4. Take some more measurements. 5. Wait for the manager to receive new event reports from the agent, which report the measurements from step 4. 		
Pass/Fail criteria		<ul style="list-style-type: none"> • The Agent sends an MDS event report to inform about the Contextual Attribute that has been changed. • Data has changed accordingly to new Contextual Attribute. 		
Notes				

TP Id		TP/PLT/AG/CLASS/CV/BV-045		
TP label		Config Changes Service. Ascent Time and Distance Contextual Attribute.		
Coverage	Spec	[ISO/IEEE 11073-10441]		
	Testable items	NumObj1; M		
	Spec	[ITU-T H.810 (2015)]		
	Testable items	Communication 8; M		
Test purpose		<p>Check that:</p> <p>Whenever a Contextual Attribute changes, the Agent shall report these changes to the Manager using an MDS object event prior to reporting any of the dependent values.</p>		
Applicability		C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_041 AND C_AG_CV_048		
Other PICS				
Initial condition		The simulated manager and the agent under test are in the operating state.		
Test procedure		<ol style="list-style-type: none"> 1. If the attribute that is going to be changed is reported in a Fixed format event report, take some measurements with the agent under test. 2. Make a change to the Contextual Attribute Unit-Code for Ascent Time and Distance object (meters to feet, feet to meters, meters to steps, steps to meters, feet to steps or steps to feet). 3. The agent shall send an MDS event report indicating the new Contextual Attribute value. 		

	<ol style="list-style-type: none"> 4. Take some more measurements. 5. Wait for the manager to receive new event reports from the agent, which report the measurements from step 4.
Pass/Fail criteria	<ul style="list-style-type: none"> • The agent sends an MDS event report to inform about the Contextual Attribute that has been changed. • Data has changed accordingly to new Contextual Attribute.
Notes	

TP Id	TP/PLT/AG/CLASS/CV/BV-046		
TP label	Config Changes Service. Descent Time and Distance Contextual Attribute.		
Coverage	Spec	[ISO/IEEE 11073-10441]	
	Testable items	NumObj1; M	
	Spec	[ITU-T H.810 (2015)]	
	Testable items	Communication 8; M	
Test purpose	<p>Check that:</p> <p>Whenever a Contextual Attribute changes, the Agent shall report these changes to the Manager using an MDS object event prior to reporting any of the dependent values.</p>		
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_042 AND C_AG_CV_049		
Other PICS			
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<ol style="list-style-type: none"> 1. If the attribute that is going to be changed is reported in a Fixed format event report, take some measurements with the agent under test. 2. Make a change to the Contextual Attribute Unit-Code for Descent Time and Distance object (meters to feet, feet to meters, meters to steps, steps to meters, feet to steps or steps to feet). 3. The agent shall send an MDS event report indicating the new Contextual Attribute value. 4. Take some more measurements. 5. Wait for the manager to receive new event reports from the agent, which report the measurements from step 4. 		
Pass/Fail criteria	<ul style="list-style-type: none"> • The agent sends an MDS event report to inform about the Contextual Attribute that has been changed. • Data has changed accordingly to new Contextual Attribute. 		
Notes			

TP Id	TP/PLT/AG/CLASS/CV/BV-047		
TP label	Config Changes Service. Speed Contextual Attribute.		
Coverage	Spec	[ISO/IEEE 11073-10441]	
	Testable items	NumObj1; M	
	Spec	[ITU-T H.810 (2015)]	
	Testable items	Communication 8; M	
Test purpose	<p>Check that:</p> <p>Whenever a Contextual Attribute changes, the Agent shall report these changes to the Manager using an MDS object event prior to reporting any of the dependent values.</p>		
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_035 AND C_AG_CV_050		

Other PICS	
Initial Condition	The simulated manager and the agent under test are in the operating state.
Test procedure	<ol style="list-style-type: none"> 1. If attribute that is going to be changed is reported in a Fixed format event report, take some measurements with the agent under test. 2. Make a change to the Contextual Attribute Unit-Code for Speed Object (meters per minute to feet, inches or steps per minute, feet per minute to meters, inches or steps per minute, steps per minute to meters, feet or inches per minute, or inches per minute to meters, feet or steps per minute). 3. The agent shall send an MDS event report indicating the new Contextual Attribute value. 4. Take some more measurements. 5. Wait for the manager to receive new event reports from the agent, which report the measurements from step 4.
Pass/Fail criteria	<ul style="list-style-type: none"> • The agent sends an MDS event report to inform about the Contextual Attribute that has been changed. • Data has changed accordingly to new Contextual Attribute.
Notes	

TP Id	TP/PLT/AG/CLASS/CV/BV-048		
TP label	Config Changes Service. Incline Contextual Attribute.		
Coverage	Spec	[ISO/IEEE 11073-10441]	
	Testable items	NumObj1; M	
	Spec	[ITU-T H.810 (2015)]	
	Testable items	Communication 8; M	
Test purpose	<p>Check that:</p> <p>Whenever a Contextual Attribute changes, the Agent shall report these changes to the Manager using an MDS object event prior to reporting any of the dependent values.</p>		
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_033 AND C_AG_CV_051		
Other PICS			
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<ol style="list-style-type: none"> 1. If the attribute that is going to be changed is reported in a Fixed format event report, take some measurements with the agent under test. 2. Make a change to the Contextual Attribute Unit-Code for Incline object (percent to angle degrees or angle degrees to percent). 3. The agent shall send an MDS event report indicating the new Contextual Attribute value. 4. Take some more measurements. 5. Wait for the manager to receive new event reports from the agent, which report the measurements from step 4. 		
Pass/Fail criteria	<ul style="list-style-type: none"> • The agent sends an MDS event report to inform about the Contextual Attribute that has been changed. • Data has changed accordingly to new Contextual Attribute. 		
Notes			

TP Id		TP/PLT/AG/CLASS/CV/BV-049		
TP label		Config Changes Service. Stride-Length Contextual Attribute.		
Coverage	Spec	[ISO/IEEE 11073-10441]		
	Testable items	NumObj1; M		
	Spec	[ITU-T H.810 (2015)]		
	Testable items	Communication 8; M		
Test purpose		<p>Check that:</p> <p>Whenever a Contextual Attribute changes, the Agent shall report these changes to the Manager using an MDS object event prior to reporting any of the dependent values.</p>		
Applicability		C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_027 AND C_AG_CV_052		
Other PICS				
Initial condition		The simulated manager and the agent under test are in the operating state.		
Test procedure		<ol style="list-style-type: none"> 1. If the attribute that is going to be changed is reported in a Fixed format event report, take some measurements with the agent under test. 2. Make a change to the Contextual Attribute Unit-Code for Stride Length object (meters to inches or inches to meters). 3. The Agent shall send an MDS event report indicating the new Contextual Attribute value. 4. Take some more measurements. 5. Wait for the manager to receive new event reports from the Agent, which report the measurements from step 4. 		
Pass/Fail criteria		<ul style="list-style-type: none"> • The agent sends an MDS event report to inform about the Contextual Attribute that has been changed. • Data has changed accordingly to new Contextual Attribute. 		
Notes				

TP Id		TP/PLT/AG/CLASS/CV/BV-050		
TP label		Config Changes Service. Energy Expended Contextual Attribute.		
Coverage	Spec	[ISO/IEEE 11073-10441]		
	Testable items	NumObj1; M		
	Spec	[ITU-T H.810 (2015)]		
	Testable items	Communication 8; M		
Test purpose		<p>Check that:</p> <p>Whenever a Contextual Attribute changes, the Agent shall report these changes to the Manager using an MDS object event prior to reporting any of the dependent values.</p>		
Applicability		C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_025 AND C_AG_CV_053		
Other PICS				
Initial condition		The simulated manager and the agent under test are in the operating state.		
Test procedure		<ol style="list-style-type: none"> 1. If the attribute that is going to be changed is reported in a Fixed format event report, take some measurements with the agent under test. 2. Make a change to the Contextual Attribute Unit-Code for Energy Expended object (calories to joules or joules to calories). 3. The agent shall send an MDS event report indicating the new Contextual Attribute value. 		

	<ol style="list-style-type: none"> 4. Take some more measurements. 5. Wait for the manager to receive new event reports from the agent, which report the measurements from step 4.
Pass/Fail criteria	<ul style="list-style-type: none"> • The agent sends an MDS event report to inform about the Contextual Attribute that has been changed. • Data has changed accordingly to new Contextual Attribute.
Notes	

TP Id	TP/PLT/AG/CLASS/CV/BV-051		
TP label	Config Changes Service. Body Weight Contextual Attribute.		
Coverage	Spec	[ISO/IEEE 11073-10441]	
	Testable items	NumObj1; M	
	Spec	[ITU-T H.810(2015)]	
	Testable items	Communication 8; M	
Test purpose	<p>Check that:</p> <p>Whenever a Contextual Attribute changes, the Agent shall report these changes to the Manager using an MDS object event prior to reporting any of the dependent values.</p>		
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_020 AND C_AG_CV_054		
Other PICS			
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<ol style="list-style-type: none"> 1. If the attribute that is going to be changed is reported in a Fixed format event report, take some measurements with the agent under test. 2. Make a change to the Contextual Attribute Unit-Code for Body Weight object (grams to pounds or pounds to grams). 3. The agent shall send an MDS event report indicating the new Contextual Attribute value. 4. Take some more measurements. 5. Wait for the manager to receive new event reports from the agent, which report the measurements from step 4. 		
Pass/Fail criteria	<ul style="list-style-type: none"> • The agent sends an MDS event report to inform about the Contextual Attribute that has been changed. • Data has changed accordingly to new Contextual Attribute. 		
Notes			

TP Id	TP/PLT/AG/CLASS/CV/BV-052		
TP label	Config Changes Service. Height Contextual Attribute.		
Coverage	Spec	[ISO/IEEE 11073-10441]	
	Testable items	NumObj1; M	
	Spec	[ITU-T H.810 (2015)]	
	Testable items	Communication 8; M	
Test purpose	<p>Check that:</p> <p>Whenever a Contextual Attribute changes, the Agent shall report these changes to the Manager using an MDS object event prior to reporting any of the dependent values.</p>		
Applicability	C_AG_OXP_000 AND C_AG_OXP_172 AND C_AG_CV_020 AND C_AG_CV_054		

Other PICS	
Initial condition	The simulated manager and the agent under test are in the operating state.
Test procedure	<ol style="list-style-type: none"> 1. If the attribute that is going to be changed is reported in a Fixed format event report, take some measurements with the agent under test. 2. Make a change to the Contextual Attribute Unit-Code for Height Object (meters to feet or feet to meters). 3. The agent shall send an MDS event report indicating the new Contextual Attribute value. 4. Take some more measurements. 5. Wait for the Manager to receive new event reports from the Agent, which report the measurements from step 4.
Pass/Fail criteria	<ul style="list-style-type: none"> • The agent sends an MDS event report to inform about the Contextual Attribute that has been changed. • Data has changed accordingly to new Contextual Attribute.
Notes	

TP Id	TP/PLT/AG/CLASS/CV/BV-053		
TP label	Operating State. Manager to Agent Maximum APDU Size		
Coverage	Spec	[ISO/IEEE 11073-20601A]	
	Testable items	CommonCharac 3; M	
	Spec	[ITU-T H.810 (2015)]	
	Testable items	Cardio_DG 1; M	
Test purpose	<p>Check that:</p> <p>Check that the total size of the response do not exceed of the maximum APDU size established by the specialization</p> <p>[AND]</p> <p>Continua PAN step counter service components shall be able to support a maximum APDU size of 224 octets from Continua PAN client components.</p>		
Applicability	C_AG_OXP_000 AND C_AG_OXP_172		
Other PICS	C_AG_OXP_041, C_AG_OXP_100, C_AG_CV_001		
Initial condition	The simulated manager and the agent are in the operating state.		
Test procedure	<ol style="list-style-type: none"> 1. IF the agent supports Step Counter sub-specialization (C_AG_CV_001=TRUE) THEN the simulated manager issues a "Remote Operation Invoke Get" command with: <ol style="list-style-type: none"> a. Obj-handle set to 0 (to request for MDS object) b. attribute-id-list.count = 103 c. attribute-id-list: (MDC_ATTR_ID_MODEL, MDC_ATTR_SYS_ID, MDC_ATTR_DEV_CONFIG_ID) repeated 34 times followed by an additional MDC_ATTR_ID_MODEL <p>ELSE (the agent does not support Step Counter sub-specialization) THEN the simulated manager issues a "Remote Operation Invoke Get" command with:</p> <ol style="list-style-type: none"> d. Obj-handle set to 0 (to request for MDS object) e. attribute-id-list.count = 4087 f. attribute-id-list: (MDC_ATTR_ID_MODEL, MDC_ATTR_SYS_ID, MDC_ATTR_DEV_CONFIG_ID) repeated 1362 times followed by an additional MDC_ATTR_ID_MODEL 2. Check the response of the agent. 3. The simulated manager issues a "Remote Operation Invoke Get" command with the 		

	<p>handle set to 0 (to request for MDS object) and an empty attribute-id-list to indicate all attributes.</p> <p>4. Check the response of the agent.</p>
Pass/Fail criteria	<ul style="list-style-type: none"> • In step 2, the agent under test may respond with a rors-cmip-get listing all the requested attributes, or with a roer message. If PICS C_AG_OXP_100 =TRUE and the agent does not respond with a rors-cmip-get message, it responds with a roer message or roej(resource-limitation) message, a WARNING will appear. <ul style="list-style-type: none"> ○ If the response is a get response, the total size of the response cannot exceed the sum of the APDU sizes of the supported specializations (limited to an absolute limit of 64512 octets): <ul style="list-style-type: none"> ▪ Pulse oximeter → 9216 octets ▪ Weighing scales → 896 octets ▪ Glucose meter →5120 octets or 64512 octets if agent supports PM-Store ▪ Blood pressure →896 octets ▪ Thermometer →896 octets ▪ Independent activity hub →5120 octets ▪ Cardiovascular →64512 octets or 6624 octets if the agent under test only supports Step Counter Profile ▪ Strength →64512 octets: ▪ Adherence monitor →1024 octets ▪ Peak flow →2030 octets ▪ Body composition analyser →7730 octets ▪ Basic ECG/Simple ECG →7168 octets or 64512 octets if the agent supports PM-Store ▪ Basic ECG/Heart Rate →1280 octets or 64512 octets if the agent supports PM-Store ▪ International normalized ratio →896 octets or 64512 if the agent supports PM-Store ○ In the case where it responds with a roer, the reason must not be protocol-violation (23) • In step 4, the agent must respond with a rors-cmip-get message.
Notes	

Bibliography

- [b-ITU-T H.810 (2013)] Recommendation ITU-T H.810 (2013), *Interoperability design guidelines for personal health systems*.
- [b-CDG 1.0] Continua Health Alliance, Continua Design Guidelines v1.0. (2008), *Continua Design Guidelines*.
- [b-CDG 2010] Continua Health Alliance, Continua Design Guidelines v1.5 (2010), *Continua Design Guidelines*.
- [b-CDG 2011] Continua Health Alliance, Continua Design Guidelines (2011), *"Adrenaline", Continua Design Guidelines*.
- [b-CDG 2012] Continua Health Alliance, Continua Design Guidelines (2012), *"Catalyst ", Continua Design Guidelines*.
- [b-ETSI SR 001 262] ETSI SR 001 262 v1.8.1 (2003-12): *ETSI drafting rules*.
- [b-IEEE 11073-20601] IEEE Std 11073-20601™ (2008): *Optimized exchange protocol*.

SERIES OF ITU-T RECOMMENDATIONS

Series A	Organization of the work of ITU-T
Series D	General tariff principles
Series E	Overall network operation, telephone service, service operation and human factors
Series F	Non-telephone telecommunication services
Series G	Transmission systems and media, digital systems and networks
Series H	Audiovisual and multimedia systems
Series I	Integrated services digital network
Series J	Cable networks and transmission of television, sound programme and other multimedia signals
Series K	Protection against interference
Series L	Environment and ICTs, climate change, e-waste, energy efficiency; construction, installation and protection of cables and other elements of outside plant
Series M	Telecommunication management, including TMN and network maintenance
Series N	Maintenance: international sound programme and television transmission circuits
Series O	Specifications of measuring equipment
Series P	Terminals and subjective and objective assessment methods
Series Q	Switching and signalling
Series R	Telegraph transmission
Series S	Telegraph services terminal equipment
Series T	Terminals for telematic services
Series U	Telegraph switching
Series V	Data communication over the telephone network
Series X	Data networks, open system communications and security
Series Y	Global information infrastructure, Internet protocol aspects and next-generation networks, Internet of Things and smart cities
Series Z	Languages and general software aspects for telecommunication systems